



EC Declaration of Conformity to: Medical Devices Directive 93/42/EEC

Legal Manufacturer:	BD Switzerland Sàrl Route de Crassier 17, Business Park Terre-Bonne, Batiment A4, 1262 Eysins Switzerland
EU Authorised Representative:	Becton Dickinson Ireland Ltd. Donore Road, Drogheda Co. Louth, A92 YW26 Ireland
Manufacturing Site (s):	Flextronics Romania SRL Calea Torontalului DN6, km 5.7 Timisoara, 300000, Romania
Device Description/Family:	Alaris™ GP Volumetric Pump <i>(See attached Product Schedule)</i>
EC Product Classification:	Class IIb, Annex IX, Rule 11
GMDN:	<i>13215 – Infusion Pump, general purpose</i> A mains electricity (AC-powered) device designed to facilitate the accurate and consistent administration of drugs and solutions which can be delivered via intravenous, subcutaneous, arterial, epidural, and intracavitary routes using a dedicated infusion set. It is used to supply higher pressures than those provided by manually clamped gravity infusion sets or infusion controllers. The device has a typical flow range of 1 to 999 ml/hour and delivers solutions from a standard infusion bag or bottle of fluid. It typically has internal batteries that enable operation for a short period when no mains electricity is available (e.g. during transportation or a power outage).

We herewith declare that the product(s) listed above and detailed in the attached Product Schedule meet the provisions of the Council Directive 93/42/EEC of 14 June 1993 concerning medical devices.

Unofficial Copy if Printed – See Intranet for Latest Released Version	
For internal use only. This document contains confidential, proprietary information of CareFusion or one of its subsidiaries. It may not be copied or reproduced without prior written permission from CareFusion.	Page 1 of 4



	<i>GP Plus (all variants)</i>	<ul style="list-style-type: none"> - Medical Device Directive 93/42/EEC - EMC Directive 2014/30/EU - RoHS Directive 2011/65/EU - Machinery Directive 2006/42/EC - Waste electrical and electronic equipment 2012/19/EU - Product Liability 85/374/EEC - REACH 1907/2006 - Packaging and Packaging Waste Directive 94/62/EC - Battery 2006/66/EC - Electronic Instructions for Use of Medical Devices 207/2012 - EN ISO 13485:2016 - EN ISO 14971:2012 - IEC 60601-1:2005+A1:2012 - IEC 60601-1-2:2014 - IEC 60601-1-6:2010 - IEC 62366:2007 - IEC 60601-1-8:2006 +A1:2012 - IEC 60601-2-24:2012 - IEC 62304:2006 - EN ISO 15223-1:2016 - EN 1041:2008+A1:2013 - IEC 60529:1991+A1:2000 - EN 1789:2007 +A1:2010 - ISTA-1A-2014
	<i>neXus GP</i>	<ul style="list-style-type: none"> - Medical Device Directive 93/42/EEC - EMC Directive 2014/30/EU - RoHS Directive 2011/65/EU - Machinery Directive 2006/42/EC - Waste Electrical and Electronic Equipment 2012/19/EU - Product Liability 85/374/EEC - REACH 1907/2006 - Packaging and Packaging Waste Directive 94/62/EC - Battery 2006/66/EC - Electronic Instructions for Use of Medical Devices 207/2012 - Radio Equipment Directive 2014/53/EU - EN ISO 13485:2016 - EN ISO 14971:2012 - EN ISO 15223-1:2016 - EN 1041:2008+A1:2013 - IEC 60601-1:2005+A1:2012 - IEC 60601-1-2:2014 - IEC 60601-1-6:2010+A1:2013 - IEC 62366-1:2015 - IEC 60601-1-8:2006+A1:2012 - IEC 60601-2-24:2012 - IEC 62304:2006+A1:2015 - EN 60529:1992+A2:2013 - ISTA-2A-2011 - ETSI EN 300 328 V2.1.1 (2016-11) - ETSI EN 301 893 V2.1.1 (2017-05) - ETSI EN 301 489-17 V3.1.1 (2016-11) - Medical Electrical Equipment & System Electromagnetic Immunity Test for RFID Readers AIM 7351731

Unofficial Copy if Printed – See Intranet for Latest Released Version	
For internal use only. This document contains confidential, proprietary information of CareFusion or one of its subsidiaries. It may not be copied or reproduced without prior written permission from CareFusion.	Page 2 of 4




Notified Body:	BSI Group The Netherlands B.V. Say Building, John M. Keynesplein 9, 1066 EP Amsterdam. Notified Body Number: 2797 Former Notified Body: BSI Group, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes, MK5 8PP. Notified body number 0086
CE Certificate Number:	<i>Annex II (EC Certificate No. 502238)</i>
Date of issuance of original CE certificate:	16 November 2005

STED File: 004

Issue Level: 33

Signed:

Ondina Bennaim

DocuSigned by:
Ondina Bennaim
 Signer Name: Ondina Bennaim
Signing Reason: I approve this document
Signing Time: 25-May-2023 | 1:37:07 AM PDT
0EF48A97D3A94213AA44C0E205189A5C

Date: **25-May-2023**

Senior Director Regulatory Affairs
International Infusion

Unofficial Copy if Printed – See Intranet for Latest Released Version	
For internal use only. This document contains confidential, proprietary information of CareFusion or one of its subsidiaries. It may not be copied or reproduced without prior written permission from CareFusion.	Page 3 of 4



Product Schedule Alaris™ GP Volumetric Pump

GMDN Number: 13215

Part Number	Description	EC Product Class
9002TIG03	Alaris™ GP Volumetric Pump with Plus software	IIb
9002TIG03-G	Alaris™ GP Guardrails™ Volumetric Pump with Plus software	IIb
GPneXus1	BD Alaris™ neXus GP	IIb

Unofficial Copy if Printed – See Intranet for Latest Released Version

For internal use only. This document contains confidential, proprietary information of CareFusion or one of its subsidiaries. It may not be copied or reproduced without prior written permission from CareFusion.

Page
4 of 4