



NINGBO GREETMED MEDICAL INSTRUMENTS CO., LTD.

CERTIFICATE OF QUALITY ANALYSIS

Description: alcohol prep pad, quantity: 200,000pcs

packing: 200pcs/box,100box/ctn, 4ctns, Total: 4ctns,

Lot No.: 201807 MFG Date: 201807, Expired Date: 202307

specification: Bag size: $60 \pm 2\text{mm} \times 60 \pm 2\text{mm}$,

Non-woven materials : $65 \times 30\text{mm}$, $70 \pm 2\text{g/M}^2$

alcohol: 70%, weight: $0.5 \pm 0.05\text{g}$

FINAL INSPECTION RESULT: Samples qty:150pcs

No	Inspection items	Standard requirements	Defects quantity	result	Testing Instruments No.
1	Packing	According to Contract.	0	OK!	By visual Inspection
2.	Printing quality	Nice print, very clear	0	OK!	By visual Inspection
3	Size of bag, non-woven.	$60 \pm 2\text{mm} \times 60 \pm 2\text{mm}$ $70 \pm 2\text{g/M}^2$	0	OK!	Digital scale Vernier caliper
4.	Weight of 70% alcohol	$0.5 \pm 0.05\text{g}$	0	OK!	Digital scale
5	Cutting of sides	No obviously burr	0	OK	By visual Inspection

Final Result: acceptable

Inspection: ShenQingSong

Approved by: Ding Yan Ping Jul 04th, 2018

Place, Date of Issue: Ningbo, jul 04, 2018

Signature: 李桂蓉

Name: Li Guirong

Position: General Manager





NINGBO GREETMED MEDICAL INSTRUMENTS CO., LTD.
