

TÜV Rheinland LGA Products GmbH • 51105 Köln

*Grena Ltd*  
*1000 Great West Road*  
*Brentford, Middlesex*  
*TW8 9HH*  
*United Kingdom*

Contact

Tel. +49 911 655-5225  
Mail: [medical-products@de.tuv.com](mailto:medical-products@de.tuv.com)

Date April 30, 2024

### **Notified Body Confirmation Letter**

Reference. : PLA\_HZ\_2024-04-29

To whom it may concern,

### **Confirmation of the status of a formal application, written agreement, and appropriate surveillance in the framework of Regulation EU 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices.**

This letter confirms that **TÜV Rheinland LGA Products GmbH**, a Notified Body (NB) designated against Regulation (EU) 2017/745 (MDR) and identified by the number **0197** on NANDO, has received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the following manufacturer:

Grena Ltd  
1000 Great West Road  
Brentford, Middlesex  
TW8 9HH  
United Kingdom  
SRN Number (if available): GB-MF-000007429

The devices covered by the formal application and the written agreement mentioned above are identified in the tables below. Table 1 identifies the devices for which an MDR application has been received, written agreement concluded and for which the NB is also responsible for appropriate surveillance under the applicable Directive. Table 2 identifies the devices for which an MDR application has been received and a written agreement concluded, but the NB has not yet taken the responsibility for appropriate surveillance of the corresponding devices under the applicable Directive.

In the case of devices covered by certificates issued under Directive 90/385/EEC (AIMDD) or Directive 93/42/EEC (MDD) that expired after May 26, 2021 but before March 20, 2023 without having been withdrawn, this letter also confirms that the manufacturer either signed the written agreement under MDR by the date of MDD/AIMDD certificate expiry; or provided evidence that a competent authority of a Member State had granted a derogation or exemption from the applicable conformity assessment procedure in accordance with Article 59(1) of MDR or Article 97(1) of the MDR respectively, by March 20, 2023 for the relevant devices.

TÜV Rheinland  
LGA Products GmbH

Am Grauen Stein  
51105 Köln  
Germany

Headquarter

Tillystraße 2  
90431 Nuremberg

Phone. +49 911 655 5225  
Fax +49 911 655 5226  
[service@de.tuv.com](mailto:service@de.tuv.com)  
[www.tuv.com/safety](http://www.tuv.com/safety)

Board of Management

Dipl.-Ing.  
Thomas Weigand, Spokesman

Dipl.-Kfm.  
Dr. Jörg Schlösser

Nuremberg HRB 26013  
VAT No.: DE 811835490

Chairman of the  
Supervisory Board

Dr.-Ing. Michael Fübi

The transition timelines that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120.3c of MDR (as amended by (EU) 2023/607), are shown below:

- May 26, 2026 for Class III custom-made implantable devices
- December 31, 2027 for Class III devices and Class IIb implantable devices excluding Well-established technologies (WET - sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- December 31, 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition or have a measuring function
- December 31, 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

On behalf of the Notified Body

Daniel Swiatko  
Certification body

**Table 1: Devices covered by this letter and for which the NB is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive:**

| Device name or Basic UDI-DI (under MDR application)   | MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage) | If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device | MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification |
|---|---|--|--|
| <b>Click'aV® Ligating Clips, Click'aV Plus™ Ligating Clips</b><br><br><b>Basic UDI-DI: 50601712800654</b> | Class IIb implantable (WET)   | N/A  | HD 1554225-1; NB 0197  |
| <b>Vclip® Ligating Clips, LigaV® Ligating Clips</b><br><br><b>Basic UDI-DI: 50601712801555</b>            | Class IIb implantable (WET)   | N/A  | HD 1554225-1; NB 0197  |
| <b>Disposable Endoscopic Instruments</b><br><br><b>Basic UDI-DI: 5060171280114V</b>                       | Class IIb non-implantable   | N/A  | HD 1554225-1; NB 0197  |
| <b>Reusable Limited Use Endoscopic Instruments</b><br><br><b>Basic UDI-DI:</b>                            | Class IIb non-implantable   | N/A  | HD 1554225-1; NB 0197  |

| Device name or Basic UDI-DI (under MDR application)   | MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage) | If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device | MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification |
|---|---|--|--|
| <b>5060171280124X</b>   |   |  |  |
| <b>Reusable Detachable Endoscopic Instruments</b><br><br>Basic UDI-DI:<br>50601712801453          | Class IIb<br>non-implantable  | N/A  | HD 1554225-1;<br>NB 0197   |
| <b>Retrieval Bags</b><br><br>Basic UDI-DI:<br>50601712800858                                      | Class IIa   | N/A  | DD 60147049 0001;<br>NB 0197   |
| <b>Ring Protect™ Disposable Wound Protector/ Retractor</b><br><br>Basic UDI-DI:<br>5060171280095A | Class IIa   | N/A  | DD 60147049 0001;<br>NB 0197   |
| <b>Veress Needles</b><br><br>Basic UDI-DI:<br>50601712800756                                      | Class IIa   | N/A  | DD 60147049 0001;<br>NB 0197   |
| <b>Reusable Trocars</b><br><br>Basic UDI-DI:<br>5060171280024U                                    | Class IIa   | N/A  | DD 60147049 0001;<br>NB 0197   |

**Table 2: Devices covered by this letter and for which the NB is NOT responsible for appropriate surveillance of the corresponding devices under the applicable Directive:**

| Device name or Basic UDI-DI (under MDR application) | MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage) | If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device | MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification |
|---|---|--|--|
| N/A   | N/A   | N/A  | N/A  |

**Confirmation Letter Revision History**

| Date       | NB internal reference traceable to each version of the letter | Action        |
|------------|---|---------------|
| 2024/04/30 | 1   | Initial issue |