

Warsaw, October 24, 2023

MANUFACTURER STATEMENT

On behalf of „MPW MED. INSTRUMENTS” SPÓŁDZIELNIA PRACY, 46 Boremlowska Street,
04-347 Warsaw, Poland

we declare that, for centrifuge MPW-55 with rotor 11202, the fastest acceleration is 20
seconds and the fastest deceleration is 10 seconds.

Member of Management Board

Wojciech Anisiewicz

„MPW MED. INSTRUMENTS”
SPÓŁDZIELNIA PRACY
w Warszawie

„MPW MED. INSTRUMENTS” SPÓŁDZIELNIA PRACY

ul. Boremlowska 46, 04-347 Warszawa
tel.: +48 22 610 56 21 / fax: +48 22 610 55 36
mail: mpw@mpw.pl / www.mpw.pl

REGON: 000401029
KRS: 0000128941
VAT UE: PL 5250001753
BDO 000042924

Alior Bank S.A.
16 2490 0005 0000 4530 5256 8362 **PLN**
PL 44 2490 0005 0000 4600 2441 7250 **EUR**

PN-EN ISO 9001
PN-EN ISO 13485