



## DECLARATION OF CONFORMITY TO CE MARK

### THE MANUFACTURER:

KW APPARECCHI SCIENTIFICI S.R.L.

Via della Resistenza 119, 53035 Monteriggioni (SI) – ITALY

Tel. +39 0577 309144 – Fax. +39 0577 309143

e-mail: [kw@kwkw.it](mailto:kw@kwkw.it) – web: [www.kwkw.it](http://www.kwkw.it)

### DECLARES THAT THE MODELS:

K64 HPL

### ARE COMPLIANT WITH THE FOLLOWING EUROPEAN COMMUNITY GUIDELINES:

- MEDICAL DEVICES Class IIA 93/42 - 2007/47/CE Class IIa (rule 2, attachment V)
- ELECTROMAGNETIC COMPATIBILITY 2014/30/UE
- LOW VOLTAGE 2014/35/UE

### ARE COMPLIANT WITH THE FOLLOWING TECHNICAL STANDARDS:

- SAFETY CEI EN 61010 – 1: 2010 / A1:2019
- ELECTROMAGNETIC COMPATIBILITY CEI EN 61326 – 1: 2013

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Ing. Fabiani Stefano  
President and CEO

Monteriggioni, 01/01/2024