

CERTIFICATE OF ANALYSIS

4408399 **GeneScan™ 600 LIZ™ dye Size Standard v2.0**

Packaging Lot: 91297284

Expire Date: 09.09.2023 (DD.MM.YYYY)

Filling lots for components in package:

Lot	Description
91268746	GeneScan™ - 600 LIZ™ Size Std. v2.0

QUALITY CONTROL

TEST	SPECIFICATION	RESULT
PHYSICAL INSPECTION		
Pouches and tubes are visually inspected for correct label and lot number. Caps are tight with no signs of leakage.		
Visual inspection of labels, print quality		Pass
Random visual inspection of pouches		Pass
Inspection of fill, cap torque, and labeling		Pass
Random visual inspection of tubes		Pass
USE TEST		
Each lot is tested for the correct number of peaks and the absence of extra peaks on a multi-capillary Applied Biosystems Genetic Analyzer.		
Number of Peaks		Pass
Absence of extra peaks		Pass
FUNCTIONAL TEST		
Each lot is tested for overall average peak height on a multi-capillary Applied Biosystems Genetic Analyzer.		
Average Peak Height	1000 - 3000 RFU	Pass

For Research Use Only. Not for use in diagnostic procedures.

ISO CERTIFICATION

Manufactured in compliance with ISO 9001 and ISO 13485 certified quality management systems.

Quality authorized by QC: **J. Žilinskienė**




Manufactured for Thermo Fisher Scientific