



Technical Data

Endo Agar, Special

M029R

Endo Agar, Special is recommended for the detection of coliform and other enteric organisms.

Composition**

Ingredients	Gms / Litre
Peptone, special	11.500
Lactose	12.900
Dipotassium phosphate	0.480
Monopotassium phosphate	0.220
Sodium chloride	3.600
Sodium sulphite	0.860
Sodium lauryl sulphate	0.010
Basic fuchsin	0.830
Agar	9.600
Final pH (at 25°C)	7.3±0.2

**Formula adjusted, standardized to suit performance parameters

Directions

Suspend 40.0 grams in 1000 ml distilled water. Boil to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool to 45-50°C. Mix well and pour into sterile Petri plates.

Caution: Basic Fuchsin is a potential Carcinogen and care should be taken to avoid inhalation of the powdered dye and contamination of the skin.

Principle And Interpretation

Endo (1) had first developed a culture medium for differentiation of lactose fermentors and non-fermenters and further developed as today's Endo Agar (2). Endo agar is used for microbiological examination of potable water and waste water, dairy products and food (3,4,5).

Sodium sulphite and basic fuchsin has inhibitory effect on gram-positive microorganisms. Sodium Lauryl sulphate inhibits many organisms other than coliforms. Lactose fermenting coliforms produce aldehyde and acid. The aldehyde in turn liberates fuchsin from the fuchsin-sulphite complex, giving rise to a red colouration of colonies. With *Escherichia coli* this reaction is very pronounced that the fuchsin crystallises, exhibiting to the colonies a permanent greenish metallic lustre (fuchsin lustre). The phosphates buffer the medium. Peptone special provides essential nutrients especially nitrogenous for the coliforms.

Quality Control

Appearance

Light pink to purple homogeneous free flowing powder

Gelling

Firm, comparable with 0.96% Agar gel.

Colour and Clarity of prepared medium

Pink Clear to slightly opalescent gel with a slight precipitate forms in Petri plates.

Reaction

Reaction of 4.0% w/v aqueous solution at 25°C. pH : 7.3±0.2

pH

7.10-7.50

Cultural Response

Cultural characteristics observed after an incubation at 35-37°C for 18-24 hours.

Cultural Response

Organism	Growth	Inoculum (CFU)	Recovery	Colour of Colony
Cultural Response				
<i>Bacillus subtilis</i> ATCC 6633	inhibited	$\geq 10^3$	0%	
<i>Enterobacter aerogenes</i> ATCC 13048	good-luxuriant	50-100	$\geq 50\%$	pink
<i>Enterococcus faecalis</i> ATCC 29212	none-poor	50-100	$\leq 10\%$	pink, small
<i>Escherichia coli</i> ATCC 25922	good-luxuriant	50-100	$\geq 50\%$	pink to rose red with metallic sheen
<i>Klebsiella pneumoniae</i> ATCC 13883	good-luxuriant	50-100	$\geq 50\%$	pink, mucoid
<i>Salmonella Typhi</i> ATCC 6539	good-luxuriant	50-100	$\geq 50\%$	colourless to pale pink
<i>Staphylococcus aureus</i> ATCC 25923	inhibited	$\geq 10^3$	0%	
<i>Pseudomonas aeruginosa</i> ATCC 27853	good-luxuriant	50-100	$\geq 50\%$	colourless, irregular
<i>Proteus vulgaris</i> ATCC 13315	good-luxuriant	50-100	$\geq 50\%$	colourless to pale pink
<i>Shigella sonnei</i> ATCC 25931	good-luxuriant	50-100	$\geq 50\%$	colourless to pale pink

Storage and Shelf Life

Store below 30°C in tightly closed container and prepared medium at 2 – 8°C away from light to avoid photo-oxidation. Use before expiry date on the label.

Reference

1. Endo, 1904, Zentralbl. Bakteriolog., Abt. 1., Orig., 35:109.
2. Levin and Schoenlein, 1930, A Compilation of Culture Media for the Cultivation of Microorganisms, Williams and Wilkins, Baltimore.
3. Greenberg, Trussell and Clesceri (ed.), 1998, Standard Methods for the Examination of Water and Wastewater, 20th ed., APHA, Washington, D.C.
4. Richardson (ed.), 1992, Standard Methods for the Examination of Dairy Products, 16th ed., APHA, Washington, D.C.
5. Speck M., 1984, Compendium of Methods for the Microbiological Examination of Foods, 3rd ed., APHA, Washington, D.C.

Revision : 01 / 2014



Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.