

EU Declaration of Conformity

We the Manufacturer; GRANT INSTRUMENTS (CAMBRIDGE) LTD
Evolution House
Unit 2, Durham Way,
Royston Gateway,
Royston, SG8 5GX.
United Kingdom

Declare under our sole responsibility that the following Product(s);

Description of equipment	Unstirred laboratory water baths
Model(s)	JBA5, JBA5 AUS, JBA12, JBA12 AUS, JBA18, JBA18 AUS, JBN5, JBN5 AUS, JBN12, JBN12 AUS, JBN18, JBN18 AUS, JBN26, JBN26 AUS, SAP2, SAP2 AUS, SAP2S, SAP2S AUS, SAP5, SAP5 AUS, SAP12, SAP12 AUS, SAP18, SAP18 AUS, SAP26, SAP26 AUS, SAP34, SAP34 AUS, SAPD & SAPD AUS
Including Accessories	Plastic tray, plastic lid and mains cord-set.

The object of the declaration described above is in conformity with the relevant Union harmonisation legislation;

Directive(s)	EMC Directive 2014/30/EU Low Voltage Directive (LVD) 2014/35/EU ROHS Directive 2011/65/EU (including 2015/863)
--------------	--

Conformity was checked in accordance with the following harmonized standards;

EMC	BS EN IEC 61326-1:2021
Safety	BS EN 61010-1:2010+A1:2019 BS EN 61010-2-010:2020
RoHS	BS EN IEC 63000:2018



.....
Mark Davison
Chief Executive Officer (CEO)
Grant Instruments (Cambridge) Ltd.

Dated: 07-Feb-2024

CE mark first applied: 2013



NOTE: Product complies with the listed directives only when used with the supplied accessories.

This device complies with Part 15 of the FCC rules. Operation is subject to the following two conditions: (1) this device may not cause harmful interference and (2) this device must accept any interference received, including interference that may cause undesired operation.