

**EU declaration of conformity**

Manufacturer according to Regulation 2017/745	Schülke & Mayr GmbH Robert-Koch-Str. 2 22851 Norderstedt Germany	
Registration Number acc. to Art. 31 2017/745	DE-MF-000005701	
<b>Product name</b>	<b>MEtherm 55</b>	
basic UDI-DI	4032651-BSC00000005-CW	
Code acc. to Art. 26 2017/745	V9099	
Intended Purpose	Neutralizing Agent	
Risk Class according to Regulation 2017/745	I	
	annex	VIII
	rule	1
Standards applied	EN ISO 13485 additional standards see technical documentation Schülke & Mayr GmbH	
Conformity Assessment Procedure according to Regulation 2017/745	annex	IV / V
Certificate	EN ISO 13485	004567 MP2016
Version	2-0	

Schülke & Mayr GmbH herewith declares that the device covered by this declaration is in conformity with the Regulation 2017/745 concerning medical devices.

Schülke & Mayr GmbH declares that Schülke & Mayr GmbH bears the sole responsibility for issuing this Declaration

Norderstedt

09. Nov. 2021

ppa.

  
Dr. Uwe Berlekamp  
Schülke & Mayr GmbH  
Director Business Line Healthcare

ppa.

  
Jörn Ahlsdorff  
Schülke & Mayr GmbH  
Director Industrial Operations  
International Industrial Operations