



KIMASED TUBES (ESR)



STERILE R

Recommended use

Mixing indications: 1 - immediately after blood collection gently invert the sample 8-10 times.
2 - before testing mix again manually or mechanically for 3-5 minutes.



Sample preservation: at room temperature (maximum 4 hours).

Storage temperatures	Up to a 24°C	at 2 - 4°C
Maximum preservation time	4 hours	< 24 hours