

17.09.2021

We ILDAM CAM LABORATUVAR MALZ. TIC. VE SAN. A.S. located at 1235th Street No: 18 Ostim 06370 ANKARA do confirm, that offered incubator **ILD-IN-120** do comply with the following technical specification for 2 units:

Parameter Specification Description of the thermostat used in the laboratory **YES**

Capacity 80 - 100 l **YES**

Programmable temperature range 5 - 70 degrees C **YES**

Programming resolution  $\leq 1$  degree **YES**

Operating time 1 - 999 min, uninterrupted **YES**

LCD or LED display **YES**

Microprocessor control **YES**

Use double Internal glass door **YES**

Stainless steel interior construction **YES**

Visual and audible alarm divergence from the set temperature **YES**

220 V power supply, 50 Hz " **YES**

**ILDAM CAM LABORATUVAR MALZEMELERİ**  
**TİCARET VE SANAYİ A.Ş.**  
58. Sok. No: 18 Ostim/ANKARA  
Tel: 0312 354 28 65 Fax: 0312 354 31 67  
Ticaret Sic. No: 26972 Ostim/İD: 473 003 1733  
Ödenmiş Sermayesi: 550.000,00 TL (Beşyüzelli Bin)  
www.ildam.com.tr e-mail: ildam@ildam.com.tr