

BIOBASE

Biobase Disinfection(Shandong)Co.,Ltd.

ADD: Room 303, R&D Building,OLABO Intelligent Manufacturing Industrial Park, No. 1, Biobase Road, Ancheng Town, Pingyin County, Jinan City, Shandong

DECLARATION OF CONFORMITY

Technical file of the company mentioned below has been inspected and audit has been completed successfully

2014/35/EU Low Voltage Directive and 2014/30/EU Electromagnetic Compatibility Directive have been taken as reference for these processes

Company Name:

Biobase Disinfection(Shandong)Co.,Ltd.

Room 303, R&D Building,OLABO Intelligent Manufacturing Industrial Park, No. 1, Biobase Road, Ancheng Town, Pingyin County, Jinan City, Shandong

Brand:

BIOBASE

Related Directives and Annex:

**2014/35/EU Low Voltage Directive (LVD)
2014/30/EU Electromagnetic Compatibility (EMC)**

Related Standards:

EN 61010-1:2010/A1:2019 (Safety requirements for electrical equipment for measurement, control, and laboratory use Part 1: General requirements)
EN IEC 61326-1:2021 (Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements)

Product(s):

Vertical Autoclave

Type(s)/Model(s):

BKQ-L60; BKQ-L80; BKQ-L100; BKQ-L120; BKQ-B50J; BKQ-B75J
BKQ-B100J; BKQ-B120J

Classification:

Laboratory Equipment

Examination Period:

June 10, 2026

Date of Expiry:

June 9, 2031

Review Result:

We, Biobase Disinfection (Shandong) Co.,Ltd., declare that during the self-testing and performance evaluation, no Non-compliance according to the requirements of the Low Voltage Directive 2014/35/EU and Electromagnetic Compatibility Directive 2014/30/EU was detected.

Year of DOC marking

For and on behalf of
2026
BIOBASE DISINFECTION
(SHANDONG) CO., LTD.

RA Specialist:

Roxanne Yan

Document No: BKSD-R26061001

Authorized Signature(s)

