



ASAP INTERNATIONAL SDN. BHD. (1157980-X)

No. 1, Jalan Sitar 33/6, Seksyen 33, 40400 Shah Alam, Selangor, Malaysia.

Tel: +6 03 5191 0166

Fax: +6 03 5191 0702

Website: www.whyasap.com

Email: info@whyasap.com

PRODUCT SPECIFICATION

Product Series	: ASAP Nitrile Powder Free Examination Glove
Material	: Nitrile
Features	: Nitrile Examination Glove, Powder Free, Non-Sterile, Ambidextrous, Finger Textured, Beaded Cuff, Single-Used Only
Color	: Black / Light Blue / Light Violet
Available Sizes	: Extra Small, Small, Medium, Large, Extra-Large
Packing Mode	: 100 pieces per box, 10 boxes per carton
Shelf Life	: 3 years

Test	Particular		Specification	Standard Compliance
Dimension (mm)	Palm Width	Extra Small	75 ± 5	EN 455 Part 2: 2015
		Small	85 ± 5	
		Medium	95 ± 5	
		Large	105 ± 5	
		Extra Large	115 ± 5	
	Length		Min. 240	
	Thickness (Single wall)	Cuff (25 ± 5 from bead)	0.06 ± 0.02	In-house
		Palm (center of palm)	0.07 ± 0.02	In-house
Finger (13 ± 3 from tip)		0.09 ± 0.02	In-house	
Physical Properties*	Force at Break (N)	Before Ageing	Min. 6	EN 455 Part 2: 2015
		After Ageing	Min. 6	
Powder Content*	Powder free (mg/glove)		Max. 2	EN 455 Part 2: 2015
Physical Appearance	Rolled beaded cuff Finger Textured Ambidextrous			

****Remark:** Results applicable upon request

Handling and Storage

Store in a cool, dry and dust free facility or at room temperature. Exposed product should be shielded from humidity, direct sunlight, and fluorescent lighting.

Caution

This product contains substance which may cause an allergic reaction in some individuals. Do not expose these gloves to any person with a known or suspected sensitivity to nitrile rubber. If allergic reaction occurs, discontinue use immediately and consult a physician. This product is not made with natural rubber latex.

