

Declaration Ref No: DC21-0035

CE Declaration of Conformity

According to Annex III of the IVD Directive 98/79/EC

We,

Atlas Medical

Head office: Ludwig-Erhard-Ring 3
Blankenfelde-Mahlow, Germany.
Tel: +49 - 33708 – 3550 30
Email: info@atlas-medical.com

Middle East Site: Sahab Free Zone Area, P. O. Box 212555, Amman, Jordan.
Tel.: +962 6 4026468
Fax: +962 6 4022588
Email: info@atlas-medical.com

Declare our responsibility that the following product:


See Attached list


- Comply with all essential requirements (Annex I) of the IVD Directive 98/79/EC. This compliance has been properly documented and covers the items listed in Annex I of the IVD Directive.
- This product is produced under Atlas quality system (ISO13485:2016) issued by GMED:
Certificate N°: 36655 rev 1
Expiry Date: October 8th.2023
- Comply with the essential requirements of following standards (EN 18113-1, -2, -4:2011, EN ISO 15223:2016, EN ISO 23640:2015, EN ISO 14971:2019, ISO 2859/1:1999, EN ISO 13612:2002, EN ISO 13641:2002).

And

Intended for In-Vitro Professional use only.

Manufacturer
Atlas Medical
Ludwig-Erhard-Ring 3
Blankenfelde-Mahlow , Germany.



Atlas Medical	Issue date	Date of review	Management approval	MRXDO10F.10 08.02.2011
	March.2021	09.03.2021		

CE Declaration of Conformity

According to Annex III of the IVD Directive 98/79/EC

Product Description
8.00.02.0.0100 : ASO Latex Kit, 100 Tests (4ml Latex, 2x1.0ml controls).
8.00.00.0.0100: CRP Latex Kit, 100 Tests (4 ml Latex, 2x1.0 ml Controls)
8.00.04.0.0100: RF Latex Kit, 100 Tests (4ml Latex, 2x1.0ml controls)
8.00.17.0.0100: D-Dimer Latex Kit, 100 Tests
8.00.13.0.0300 : Streptococcus Latex Kit, 6 Groups, 6x50 Tests (5x1.5ml Latex (A,B,C,G,F), 1x3ml Latex(D), 1x1.0ml Positive Control, 1x2ml Extraction Reagent E, 1x1.5ml Extraction Reagent 1, 1x1.5ml Extraction Reagent 2, 2x2.5ml Extraction Reagent 3, Stirring Sticks, Glass Slide).
8.00.18.3.0500 : RPR Syphilis (Coarse Grain) Kit, 500 Tests (10 ml latex, 2x1ml control) Without card, stirring sticks.
8.00.18.3.1000 RPR Carbon Antigen (Coarse Grain) Kit, 1000 Tests (Reagent only).



Declaration Ref No: DC22-0065

CE Declaration of Conformity

According to Annex III of the IVD Directive 98/79/EC

We,

Atlas Medical GmbH

Head office: Ludwig-Erhard-Ring 3
Blankenfelde-Mahlow, Germany.

Tel: +49 - 33708 – 3550 30

Email: info@atlas-medical.com

Manufacturing Site: Sahab Free Zone Area, P. O. Box 204, Amman 11512, Jordan.

Tel.: +962 6 4026468

Fax: +962 6 4022588

Email: info@atlas-medical.com

Declare our responsibility that the following product:

See Attached list


- Comply with all essential requirements (Annex I) of the IVD Directive 98/79/EC. This compliance has been properly documented and covers the items listed in Annex I of the IVD Directive.
- This product is produced under Atlas quality system (ISO13485:2016) issued by GMED:
Certificate N°: 36655 rev 1
Expiry Date: October 8th.2023
- Comply with the essential requirements of following standards (EN 18113-1, -2,-4:2011, EN ISO 15223:2016 , EN ISO 23640:2015, EN ISO 14971:2019, ISO 2859/1:1999, EN ISO 13612:2002, EN ISO 13641:2002.

And

Intended for In-Vitro Professional use only.

Manufacturer
Atlas Medical
Ludwig-Erhard-Ring 3
Blankenfelde-Mahlow , Germany.



Atlas Medical	Issue date	Date of review	Management approval	MRXDO10F.10 08.02.2011
	May.2022	21.05.2022		

CE Declaration of Conformity

According to Annex III of the IVD Directive 98/79/EC

Item code	Product Description
8.00.01.0.0100	Atlas CRP Latex Kit with Buffer (100 Tests)
8.00.05.0.0100	Atlas RF Latex kit with Buffer(100 Tests)
8.00.11.0.0050	Atlas SLE Latex kit (50 Tests)
8.00.11.0.0100	Atlas SLE Latex kit (100 Tests)
8.00.12.0.0100	Atlas Staphylococcus Latex Kit (100 Tests)
8.00.17.0.0050	Atlas D-Dimer Latex Kit (50 Tests)
8.00.19.3.0100	Atlas TPHA Kit (100 Tests)
8.00.19.3.0200	Atlas TPHA Kit (200 Tests)
8.00.20.3.2500	Atlas VDRL Kit, 5ml+55ml buffer
8.04.38.0.0020	Atlas Fecal Occult Blood Test (FOB) Test Cassette , 20 Tests/Box
8.04.85.0.0050	Atlas Fecal Occult Blood Test (FOB) Test Strip, 50 Tests/Box
8.04.109.0.0020	Atlas Procalcitonin test (PCT) , 20 Tests/Box
8.16.78.0.0025	Atlas Calprotectin Test Cassette , 25 Tests/Box
8.04.45.0.0001	Atlas Troponin I Test Cassette, Bulk
8.04.45.0.0020	Atlas Troponin I Test Cassette , 20 Tests/Box.
8.04.45.0.0030	Atlas Troponin I Test Cassette , 30 Tests/Box.
8.04.46.0.0001	Atlas Myoglobin Test Cassette, Bulk
8.04.46.0.0020	Atlas Myoglobin Test Cassette , 20 Tests/Box.
8.04.46.0.0030	Atlas Myoglobin Test Cassette , 30 Tests/Box.
8.04.47.0.0001	Atlas CK-MB Test Cassette , Bulk.
8.04.47.0.0020	Atlas CK-MB Test Cassette , 20 Tests/Box.
8.04.47.0.0030	Atlas CK-MB Test Cassette , 30 Tests/Box.
8.04.48.0.0001	Atlas Cardiac Triple Tests Cassette (Troponin I, CK-MB, Myoglobin), Bulk.
8.04.48.0.0020	Atlas Cardiac Triple Tests Cassette (Troponin I, CK-MB, Myoglobin), 20 Tests/Box.
8.04.48.0.0030	Atlas Cardiac Triple Tests Cassette (Troponin I, CK-MB, Myoglobin), 30 Tests/Box.
8.14.19.1.0096	Helicobacter pylori Antigen ELISA, 96 Tests.
8.51.00.0.0096	25-OH VITAMIN D Elisa Kit, 96 Tests.
8.57.00.0.0096	Vitamin B12 Elisa Kit, 96 Tests

LA

GMED certifie que le système de management de la qualité développé par
GMED certifies that the quality management system developed by

ATLAS MEDICAL GmbH
Ludwig-Erhard-Ring 3
15827 Blankenfelde-Mahlow GERMANY

pour les activités
for the activities

Conception et développement, fabrication et vente de dispositifs médicaux de diagnostic in vitro .

Design and Development, Manufacturing and Sales of in vitro diagnostic medical devices.

réalisées sur le(s) site(s) de
performed on the location(s) of

Voir addendum

See addendum

est conforme aux exigences des normes internationales
complies with the requirements of the international standards

ISO 13485: 2016

Début de validité / Effective date October 9th, 2023 (included)

Valable jusqu'au / Expiry date : October 8th, 2026 (included)

Etabli le / Issued on : October 9th, 2023




**CERTIFICATION
DE SYSTEMES
DE MANAGEMENT**
Accréditation n°4-0608
Liste des sites accrédités
et portée disponible sur
www.cofrac.fr

GMED N° 36655-2

Ce certificat est délivré selon les règles de certification GMED / This certificate is issued according to the rules of GMED certification

Renouvelle le certificat 36655-1



On behalf of the President
Béatrice LYS
Technical Director

Ce certificat couvre les activités et les sites suivants :
This certificate covers the following activities and sites:

French version :

Conception et développement, fabrication et vente de dispositifs médicaux de diagnostic *in vitro* à usage professionnel et/ ou d'autodiagnostic, dans les domaines du groupage sanguin, de la microbiologie, de la biochimie, de la toxicologie, de l'oncologie, de la cardiologie, de l'histologie, de l'endocrinologie et des maladies infectieuses, dans les techniques d'Agglutination/ ELISA/ Tests rapides/ Colorimétrie/ Disques antibiotiques.

English version:

Design and Development, Manufacturing and Sales of in vitro diagnostic medical devices for professional use and/or for self-testing, in the field of Immunohematology, Microbiology, Biochemistry, Toxicology, Oncology, Cardiology, Histology, Endocrinology Biosensors and Infectious diseases, in techniques of Agglutination/ ELISA/ Rapid tests/ Colorimetry/Antibiotic disks.

**ATLAS MEDICAL GmbH
Ludwig-Erhard-Ring 3
15827 Blankenfelde-Mahlow
GERMANY**

French version:

Siège social, responsable de la mise sur le marché

English version:

Headquarter, legal manufacturer

**Sahab Industrial Zone Area
King Abdullah II Industrial City
Amman 11512
JORDAN**


French version:

Conception, fabrication et contrôle final

English version:

Design, manufacture and final control

2 sites / 2 sites

DocuSigned by:
Beatrice Lys
FF33BDA98AA04A3...


**On behalf of the President
Béatrice LYS
Technical Director**

Date: 05/Jan/2023

STATEMENT

We, Atlas Medical having a registered office at Ludwig-Erhard-Ring 3, 15827 Blankenfelde-Mahlow, Berlin, Germany assign SRL Sanmedico having a registered office at A. Corobceanu Street 7A, apt.9, Chisinau MD-2012, Moldova, as authorized representative in correspondence with the conditions of directive 98/79/EEC.

We declare that the company mentioned above is authorized to register, notify, renew or modify the registration of medical devices on the territory of the Republic of Moldova.

On Behalf of Manufacturer:

General Manager

Haya Amawi

Signature: _____

Date: _____

Atlas Medical GmbH
Ludwig - Erhard Ring 3
15827 Blankenfelde - Mahlow
Tel. (0049) 33708 - 355030

Atlas Medical: Ludwig-Erhard-Ring 3, 15827 Blankenfelde-Mahlow, Berlin, Germany,
Tel: +4933708355030

Regulatory Office: William James House, Cowley Rd, Cambridge, CB4 0WX, United Kingdom
Tel: +44 (0) 1223 858 910

Middle East Site: P.O Box 204, King Abdullah II Industrial Estate, Amman, 11512, Jordan
Tel: +962 6 4026468

ATLAS C-REACTIVE PROTEIN (CRP) LATEX KIT

For the qualitative and semi-quantitative measurement of C-reactive protein (CRP) in human serum.

IVD For in -vitro diagnostic and professional use only

2°C  8°C
Store at 2-8°C

INTENDED USE

Atlas C-Reactive Protein (CRP) is used to measure the CRP in human serum qualitatively and semi- quantitatively.

INTRODUCTION

C-reactive protein (CRP), the classic acute-phase of human serum, is synthesized by hepatocytes. Normally, it is present only in trace amounts in serum, but it can increase as much as 1,000-fold in response to injury or infection. The clinical measurement of CRP in serum therefore appears to be a valuable screening test for organic disease and a sensitive index of disease activity in inflammatory, infective and ischemic conditions. MacLeod and Avery found that antibody produced against purified CRP provided a more sensitive test than the C-polysaccharide assay. Since that time a number of immunological assays have been devised to measure CRP such as capillary precipitation, double immunodiffusion and radical immunodiffusion.

The CRP reagent kit is based on the principle of the latex agglutination assay described by Singer and Plotz. The major advantage of this method is the rapid two (2) minute reaction time.

PRINCIPLE

The CRP reagent kit is based on an immunological reaction between CRP Antisera bound to biologically inert latex particles and CRP in the test specimen. When serum containing greater than 6 mg/L CRP is mixed with the latex reagent, visible agglutination occurs.

MATERIALS

MATERIALS PROVIDED

- CRP Latex Reagent: Latex particles coated with goat IgG anti-human CRP, pH 8.2 **MIX WELL BEFORE USE.**

- CRP Positive Control Serum: A stabilized pre-diluted human serum containing >20mg/L CRP.
- CRP Negative Control Serum: A stabilized pre-diluted animal serum.
- Glass Slides.
- Stirring Sticks.

MATERIALS REQUIRED BUT NOT PROVIDED

- Mechanical rotator with adjustable speed at 80-100 r.p.m.
- Vortex mixer.
- Pipettes 50 µL.
- Glycine Buffer (20x): add one part to nineteen parts of distilled water before use.

PRECAUTIONS

- Reagents containing sodium azide may be combined with copper and lead plumbing to form highly explosive metal azides. Dispose of reagents by flushing with large amounts of water to prevent azide buildup.
- For In Vitro diagnostic use.
- Positive and negative controls prepared using human serum found negative for hepatitis B surface antigen (HBsAg) by FDA required test; however, handle controls as if potentially infectious.
- Accuracy of the test depends on the drop size of the latex reagent (40µl). Use only the dropper provided with the latex and hold perpendicularly when dispensing.
- Glass slides should be thoroughly rinsed with water and wiped with lint-free tissue after each use.

STORAGE AND STABILITY

- Reagents are stable until specified expiry date on bottle label when stored refrigerated (2 - 8°C). **DO NOT FREEZE.**
- The CRP latex reagent, once shaken must be uniform without visible clumping. When stored refrigerated, a slight sedimentation may occur and should be considered normal.
- Do not use the latex reagent or controls if they become contaminated.

SPECIMEN COLLECTION AND STORAGE

- Use fresh serum collected by centrifuging clotted blood.

- If the test cannot be carried out on the same day, store the specimen for 7 days at 2-8 °C and for 3 months at -20°C.
- For longer periods the sample must be frozen.
- As in all serological tests, hemolytic or contaminated serum must not be used.
- Do not use plasma.

PROCEDURE

A.QUALITATIVE TEST:

- Allow the reagents and samples to reach room temperature. The sensitivity of the test may be reduced at low temperatures.
- Place 40 µL of the sample and one drop of each Positive and Negative controls into separate circles on the slide test.
- Mix the CRP-latex reagent vigorously or on a vortex mixer before using and add one drop (40 µL) next to the samples to be tested.
- Mix the drops with a stirrer, spreading them over the entire surface of the circle. Use different stirrers for each sample.
- Place the slide on a mechanical rotator at 80-100 r.p.m. for 2 minutes. False positive results could appear if the test is read later than two minutes.

B.SEMI-QUANTITATIVE TEST:

- Make serial two fold dilutions of the sample in 9 g/L saline solution.
- Proceed for each dilution as in the qualitative method.

QUALITY CONTROL

Positive and Negative controls are recommended to monitor the performance of the procedure, as well as a comparative pattern for a better result interpretation.


All result different from the negative control result, will be considered as a positive.

INTERPRETATION OF RESULTS

A.QUALITATIVE TEST:

A **negative** reaction is indicated by a uniform milky suspension with no agglutination as observed with the CRP Negative Control.

A **positive** reaction is indicated by any observable agglutination in the reaction mixture. The specimen reaction should be compared to the CRP Negative Control (Fig. 1).



Positive Negative

Figure 1

B. Semi-QUANTITATIVE TEST:

The approximate CRP concentration in the patient sample is calculated as follow:

6×CRP titer = ---- mg/L

INTERFERENCES

NONE INTERFERING SUBSTANCES:

- Hemoglobin (10g/dl)
- Bilirubin(20mg/dl)
- Lipemia(10g/dl)
- Other substances interfere, such as RF (100IU/ml).

NOTE

- High CRP concentration samples may give negative results .Retest the sample again using a drop of 20µl.
- The strength of agglutination is not indicative of the CRP concentration in the samples tested.
- Clinical diagnosis should not be made on findings of a single test result, but should integrate both clinical and laboratory data.

LIMITATIONS

1. Reaction time is critical. If reaction time exceeds two (2) minutes, drying of the reaction mixture may cause false positive results.
2. Freezing the CRP Latex Reagent will result in spontaneous agglutination.
3. Intensity of agglutination is not necessarily indicative of relative CRP concentration; therefore, screening reactions should not be graded.
4. A false negative can be attributed to a prozone phenomenon (antigen excess). It is recommended, therefore, to check all negative sera by retesting at a 1:10 dilution with glycine buffer.

REFERENCE VALUES

Up to 6 mg/L. Each laboratory should establish its own reference range.

PERFORMANCE CHARACTERISTICS

- **Sensitivity:** 6(5-10) mg/L
- **Prozone effect:** No prozone effect was detected up to 1600 mg/L
- **Diagnostic sensitivity:** 95.6 %.
- **Diagnostic specificity:** 96.2 %.

REFERENCES

1. Pepys, M.B.. Lancet 1:653 (1981).
2. Werner, M.. Clin.Chem. Acta 25:299 (1969).
3. MacLeod, C.M., et. al.. J. Exp. Med 73:191 (1941).
4. Wood, HF., et. al.. J. Clin. Invest. 30: 616 (1951).
5. Mancini, G., et. al. Immunochemistry 2:235 (1965).
6. Singer, J.M., et. al.. Am. J. Med 21: 888 (1956).
7. Fischer, C.L., Gill,. C.W.. In Serum Protein Abnormalities. Boston, Little, Brown and Co., (1975).



ATLAS MEDICAL














William James House,
Cowley Road, Cambridge,
CB4 0WX, UK

Tel: +44 (0) 1223 858 910

Fax: +44 (0) 1223 858 524

PPI005A01

Rev H (06.06.2017)

	Catalogue Number		Store at
	For In-Vitro Diagnostic use		Caution
	Number of tests in the pack		Read product insert before use
	Lot (batch) number		Manufacturer
	Fragile, handle with care		Expiry date
	Manufacturer fax number		Do not use if package is damaged
	Manufacturer telephone number		

ATLAS RHEUMATOID FACTOR (RF) LATEX KIT

latex slide test for the qualitative and semi-quantitative measurement of RF in human serum.

IVD For In-Vitro diagnostic and professional use only



Store at 2-8°C

INTENDED USE

A latex slide test for the qualitative and semi-quantitative measurement of RF in human serum.

INTRODUCTION

Rheumatoid factors (RF) are antibodies directed against antigenic sites in the Fc fragment of human and animal IgG. Their frequent occurrence in rheumatoid arthritis makes them useful for diagnosis and monitoring of the disease.

One method used for rheumatoid factor detection is based on the ability of rheumatoid arthritis sera to agglutinate sensitized sheep red cells, as observed by Waaler and Rose. A more sensitive reagent consisting of biologically inert latex beads coated with human gamma globulin was later described by Singer and Plotz. The RF kit is based on the principle of the latex agglutination assay of Singer and Plotz. The major advantage of this method is rapid performance (2 minute reaction time) and lack of heterophile antibody interference.

PRINCIPLE

The RF reagent is based on an immunological reaction between human IgG bound to biologically inert latex particles and rheumatoid factors in the test specimen. When serum containing rheumatoid factors is mixed with the latex reagent, visible agglutination occurs.

MATERIALS

MATERIALS PROVIDED

- RF Latex Reagent: Latex particles coated with human gamma-globulin, pH, 8.2. Preservative. Contains N, N-dimethylformamide.
- RF Positive Control Serum: Human serum with a RF concentration > 30 IU/mL. Preservative.

- RF Negative Control Serum: Animal serum. Preservative.
- Reaction Slide
- Stirring sticks

MATERIALS REQUIRED BUT NOT PROVIDED

- Timer
- Test Tubes (for dilution)
- Serological pipettes (for sample addition and for dilution)
- Rotator (optional)
- Glycine Buffer (20x): add one part to nineteen parts of distilled water before use.

PRECAUTIONS

- All reagents contain 0.1 % (w/v) sodium azide as a preservative.
- Reagents containing sodium azide may be combined with copper and lead plumbing to form highly explosive metal azides. Dispose of reagents by flushing with large amounts of water to prevent azide buildup.
- For In Vitro diagnostic use.
- Positive and negative controls prepared using human serum found negative for hepatitis B surface antigen (HBsAg) by FDA required test; however, handle controls as if potentially infectious.
- Accuracy of the test depends on the drop size of the latex reagent (40 µL). Use only the dropper supplied with latex and hold it perpendicularly when dispensing.
- Use a clean pipette tip and stirring stick for each specimen, and glass slides should be thoroughly rinsed with water and wiped with lint-free tissue after each use.
- Check reactivity of the reagent using the controls provided.

STORAGE AND STABILITY

- Reagents are stable until specified expiry date on bottle label when stored refrigerated (2-8°C).
- Do not freeze.
- The RF latex reagent, once shaken must be uniform without visible clumping. When stored refrigerated, a slight sedimentation may occur and should be considered normal.
- Do not use the latex reagent or controls if they become contaminated.

SPECIMEN COLLECTION AND STORAGE

- Use fresh serum collected by centrifuging clotted blood.
- If the test cannot be carried out on the same day, store the specimen for 7 days at 2-8°C and for 3 months at -20°C.
- As in all serological tests, hemolytic or contaminated serum must not be used.
- Do not use PLASMA.

PROCEDURE

Qualitative method

- Allow the reagents and samples to reach room temperature. The sensitivity of the test may be reduced at low temperatures.
- Place 50 µL of the sample and one drop of each Positive and Negative controls into separate circles on the slide test.
- Mix the RF-latex reagent rigorously or on a vortex mixer before using and add one drop (50 µL) next to the sample to be tested.
- Mix the drops with a stirrer, spreading them over the entire surface of the circle. Use different stirrers for each sample.
- Place the slide on a mechanical rotator at 80-100 r.p.m. for 2 minutes. False positive results could appear if the test is read later than two minutes.

Semi-quantitative method

- Make serial two fold dilutions of the sample in 9 g/L saline solution.
- Proceed for each dilution as in the qualitative method.

READING AND INTERPRETATION

Examine macroscopically the presence or absence of visible agglutination immediately after removing the slide from the rotator. The presence of agglutination indicates a RF concentration equal or greater than 8 IU/mL (Note 1). The titer, in the semi-quantitative method, is defined as the highest dilution showing a positive result.

CALCULATIONS

The approximate RF concentration in the patient sample is calculated as follows:

$$8 \times \text{RF Titer} = \text{IU/mL}$$

INTERFERENCES

NON INTERFERING SUBSTANCES:

- Hemoglobin (10g/dl)
- Bilirubin(20mg/dl)
- Lipemia(10g/dl)

Other substances may interfere.

QUALITY CONTROL

1. RF Positive and Negative Control should be included in each test batch.
2. Acceptable performance is indicated when a uniform milky suspension with no agglutination is observed with the RF Negative Control and agglutination with large aggregates is observed with the RF Positive Control.

PERFORMANCE CHARACTERISTICS

Analytical sensitivity

8(6-16) IU/ml, under the described assay conditions.

PROZONE EFFECT

No prozone effect was detected up to 1500 IU/ml.

DIAGNOSTIC SENSITIVITY

100%.

DIAGNOSTIC SPECIFICITY

100%.

The diagnostic sensitivity and specificity have been obtained using 118 samples compared with the same method of a computer.

LIMITATIONS

- Reaction time is critical. If reaction time exceeds 2 minutes, drying of the reaction mixture may cause false positive result.
- Freezing the RF Latex Reagent will result in spontaneous agglutination.
- Intensity of agglutination is not necessarily indicative of relative RF concentration; therefore, screening reactions should not be graded.
- Increased levels of RF may be found in some diseases other than rheumatoid arthritis such as infectious mononucleosis, sarcoidosis, lupus erythematosus, Sjogren's syndrome.
- Certain patients with rheumatoid arthritis will not have the RF present in their serum.

- The incidence of false positive results is about 3-5 %.Individuals suffering from infectious mononucleosis, hepatitis, syphilis as well as elderly people may give positive results.
- Diagnosis should not be solely based on the results of latex method but also should be complemented with a Waaler Rose test along with the clinical examination.

REFERENCE VALUES

Up to 8 IU/mL. Each laboratory should establish its own reference range.

NOTES

1. Results obtained with a latex method do not compare with those obtained with Waaler Rose test. Differences in the results between methods do not reflect differences in the ability to detect rheumatoid factors.

REFERENCES

1. Robert W Dorner et al. Clinica Chimica Acta 1987; 167: 1 – 21.
2. Frederick Wolfe et al. Arthritis and Rheumatism 1991; 34: 951- 960.
3. Robert H Shmerling et al. The American Journal of Medicine 1991; 91: 528 –534.
4. Adalbert F. Schubart et al. The New England Journal of Medicine 1959; 261: 363 – 368.
5. Charles M. Plotz 1956; American Journal of Medicine; 21:893 – 896.
6. Young DS. Effects of drugs on clinical laboratory test, 4th ed. AACC Press, 1995.














ATLAS MEDICAL

William James Hous Cowley Road,
Cambridge, CB4 0WX, UK.

Tel: +44 (0) 1223 858 910

Fax: +44 (0) 1223 858 524

PPI008A01, Rev H (17.06.2017)

	Catalogue Number		Store at
	For In-Vitro Diagnostic use		Caution
	Number of tests in the pack		Read product insert before use
	Lot (batch) number		Manufacturer
	Fragile, handle with care		Expiry date
	Manufacturer fax number		Do not use if package is damaged
	Manufacturer telephone number		



ATLAS SLE LATEX TEST

A latex agglutination slide test for the qualitative and semi-quantitative detection of DNP antibodies associated with Systemic Lupus Erythematosus (SLE) in human serum

IVD For In-Vitro diagnostic and professional use only

2°C  8°C
Store at 2°-8°C

INTENDED USE

The SLE TEST is intended to be used as an aid in the diagnosis of Systemic Lupus Erythematosus (SLE) through the detection and quantitation of serum antinucleoprotein factors associated with SLE..

INTRODUCTION AND PRINCIPLE

The detection of antinuclear antibodies by laboratory methods include immunofluorescence, LE cell test and agglutination of coated particles. The antibodies that are believed to be most characteristic of SLE are those that are directed against deoxyribonucleoprotein (DNP). These antibodies are believed to cause the formation of the LE cell in vitro, with this unusual event occurring in 75-80% of those patients diagnosed as having SLE. It is not necessary to have a positive LE cell test for the diagnosis of SLE as this test had been found negative in certain individuals having symptoms suggestive for SLE. In these individuals, antinuclear antibodies may be demonstrated by methods other than the LE cell test.

The principle of the SLE TEST is based on the agglutination reaction between latex particles coated with DNP being brought into contact with a serum, which contains antinuclear antibodies. Agglutination indicates a positive reaction. The reaction time for this occurrence is within one minute.

MATERIALS

MATERIALS PROVIDED

- SLE Latex Reagent: polystyrene latex particles coated with DNP extracted from fetal calf thymus. Sodium azide (0.1%) is used as preservative. Shake well prior to use.
- SLE Positive Control: Human serum that has been diluted and stabilized with buffers and contains sodium azide (0.1%) as a preservative.
- SLE Negative Control: Human serum that has been diluted and stabilized with buffers and contains sodium azide (0.1%) as a preservative.
- Disposable stirring sticks.
- Glass slide.

MATERIALS NEEDED BUT NOT PROVIDED

- Timer.
- Micropipette.
- Physiological saline (0.9%NaCl).
- Test tubes 12x75mm.
- Serological pipettes (1ml delivery).
- Lab rotator (optional).

PRECAUTIONS

- For In Vitro Diagnostic Use Only.
- Even though the control sera supplied in the SLE TEST Kit have been tested by an FDA approved method for the presence of Hepatitis B Surface Antigen (HBsAg) and HTLV-III antibodies and found to be non-reactive, all human serum products and patient specimens should be considered potentially hazardous and handled in the same manner as an infectious agent.
- The preservative sodium azide may react with metal plumbing to form explosive metal oxides.
- In disposal, flush with a large volume of water to prevent metal azide build up.

STORAGE & STABILITY

- When not in use, store reagent and controls at 2-8°C.
- DO NOT FREEZE.
- Prior to use, allow reagents and controls to warm up to room temperature.

- Expiration date is specified on the kit label and on each vial. Biological indication of product instability is positive and negative controls.

SPECIMEN COLLECTION

- The test should be performed on serum.
- The test sera and controls should not be heat inactivated.
- Fresh specimens (less than 24 hours) should be used in performing the test.
- If testing is delayed, specimens should be refrigerated (or frozen where applicable).
- Bacterial contamination may cause false positive agglutination.

PROCEDURES

A. Method I (Qualitative)

1. Bring all reagents and serum samples to room temperature.
2. Positive and Negative Controls should be tested with each series of test sera. Using micropipette, place 0.040ml of test serum on one circle of the test slide. Use separate pipette tip for each test serum.
3. Important: The SLE Latex Reagent must be shaken vigorously for 30 seconds prior to using on each day's testing. This is to insure that there is no aggregation of the latex particles which may occur upon standing. Do not use a vortex mixer.
4. Deliver one drop of SLE Latex to each circle that contains specimen on the slide. Spread the resulting mixture by using the plastic stick provided. Do not use the same plastic stick to mix each test serum or control as this will cause cross-contamination.
5. Gently tilt and rotate slide by hand for one minute (rotator can be used).
6. Observe for macroscopic clumping using the indirect oblique light source. The reaction of the test serum is compared to the SLE positive and negative control sera.
7. Observe for agglutination no longer than one minute.

* Sera that are positive in the screening test should be retested in the titration test (semi-quantitative test) to provide verification for borderline interpretations.

B. Method II (Semi-Quantitative)

1. For each test serum to be titrated, label 6 test tubes (12x75 mm).
2. To each tube add 0.2 ml physiological saline.
3. To Tube No.1 add 0.2 ml of undiluted test serum.
4. Serially make two-fold dilutions by mixing contents of tube No.1 with a pipette and transferring 0.2 ml to tube No.2. Repeat serial transfers for each tube. For the 6 tubes, the dilutions range from 1:2 to 1:64. If required, additional serum dilutions can be added.
5. Repeat Steps 3 to 7 as given in Method I (Qualitative).

RESULTS:

1. Positive Result:

Presence of agglutination within 1 minute.

2. Negative Result:

Smooth milky suspension within 1 minute.

LIMITATION

Those patients with scleroderma, rheumatoid arthritis, dermatomyositis, and a variety of connective tissue diseases may show reactivity when their serum is tested with the SLE TEST latex. In recent studies, it has been reported that many widely used drugs such as hydralazine, isoniazid, procainamide and a number of anticonvulsant drugs can induce a systemic lupus erythmatosis (SLE) syndrome.

BIBLIOGRAPHY

1. Christian C.L., R. Mendez-Bryan, and D.L. Larson, 1958. Proc. Soc. Exptl. Biol. Med. 98, 820-823.
2. Friou, G.J., S.C. Finch, and K.D. Detre, 1958. J. Immunol. 80, 324-329.
3. Hargraves, M.M., H. Richmond, and R. Morton, 1948. Proc. Mayo Clin. 23, 25-28.
4. Holman, H.R., and H.G. Kunkel, 1957. Science 126,163.

5. Miescher, P.A., and R. Strassel, 1957. Vox. Sang, 2, 283-287.
6. Miescher, P.A., N. Rothfield, and A. Miescher, 1966. Lupus Erythematosus, E.L. Dubois, Ed., Blakiston Co., N.Y.
7. Rothfield, N.F., J.J. Phythyon, C. McEwen, and P. Miescher, 1961. Arthritis Rheum. 4, 223-229.



ATLAS Medical

William James House,

Cowley Road, Cambridge, CB4 0WX, UK

Tel: ++44 (0) 1223 858 910

Fax: ++44 (0) 1223 858 524

PPI040A01

Rev C (24.10.2015)

	Catalogue Number		Store at
	For In-Vitro Diagnostic use		Caution
	Number of tests in the pack		Read product insert before use
	Lot (batch) number		Manufacturer
	Fragile, handle with care		Expiry date
	Manufacturer fax number		Do not use if package is damaged
	Manufacturer telephone number		

ANTISTREPTOLYSIN-O (ASO) LATEX SLIDE TEST

For the qualitative and quantitative measurement of antibodies to Antistreptolysin-O in human serum.

IVD For in -vitro diagnostic and professional use only



INTENDED USE

ATLAS ANTISTREPTOLYSIN-O (ASO) latex slide Test is used for the qualitative and quantitative measurement of antibodies to Antistreptolysin-O in human serum.

INTRODUCTION

The group A β -hemolytic streptococci produces various toxins that can act as antigens. One of these exotoxins streptolysin-O, was discovered by Todd in 1932.

A person infected with group A -hemolytic streptococci produces specific antibodies against these exotoxins, one of which is antistreptolysin-O. The quantity of this antibody in a patient's serum will establish the degree of infection due to the -hemolytic streptococcal.

The usual procedure for the determination of the antistreptolysin titer is based on the inhibitory effect that the patient's serum produces on the hemolytic power of a pre-titrated and reduced streptolysin-O. However, the antigen-antibody reaction occurs independently of the hemolytic activity of streptolysin-O. This property enables the establishment of a qualitative and quantitative test for the determination of the antistreptolysin-O by agglutination of latex particles on slide.

PRINCIPLE

ASO test method is based on an immunologic reaction between streptococcal exotoxins bound to biologically inert latex particles and streptococcal antibodies in the test sample. Visible agglutination occurs when increased antibody level, are present in the test specimen.

MATERIALS

MATERIALS PROVIDED

- ASO Latex Reagent: Latex particles coated with streptolysin O, pH, 8,2. Preservative
- ASO Positive Control(Red cap): Human serum with an ASO concentration > 200 IU/mL.Preservative
- ASO Negative Control (Blue cap) Animal serum. Preservative
- Reaction Slide.
- Stirring Sticks.

MATERIALS REQUIRED BUT NOT PROVIDED

- Timer.
- Test Tubes 12x75mm.
- Test Tube Rack.
- Serological pipettes.
- High intensity light.
- Saline Solution, 0.9% NaCL.

PRECAUTIONS

- All reagents contain 0.1% (w/v) sodium azide as a preservative. Store all reagents at 2-8°C. **DO NOT FREEZE.**
- Reagents containing sodium azide may be combined with copper and lead plumbing to form highly explosive metal azides. Dispose of reagents by flushing with large amounts of water to prevent azide build-up.
- For In Vitro diagnostic use.
- Positive and negative controls prepared using human serum found negative for hepatitis B surface antigen (HBsAg) and HIV-III by FDA required test; however, handle controls as if potentially infectious.

REAGENT STORAGE AND STABILITY

- Reagents are stable until specified expiry date on bottle label when stored refrigerated (2-8°C).
- **DO NOT FREEZE.**
- The ASO Latex Reagent, once shaken must be uniform without visible clumping. When stored refrigerated, a slight sedimentation may occur and should be considered normal.
- Do not use the latex reagent or controls if they become contaminated.

SPECIMEN COLLECTION AND STORAGE

- Use fresh serum collected by centrifuging clotted blood.
- If the test cannot be carried out on the same day, store the specimen for 7 days at 2-8°C and for 3 months at -20°C.

- For longer periods the sample must be frozen.
- As in all serological tests, hemolytic or contaminated serum must not be used.
- **DO NOT USE PLASMA.**

PROCEDURE

Qualitative method

1. Allow the reagents and samples to reach room temperature. The sensitivity of the test may be reduced at low temperatures.
2. Place 50 μ L of the sample and one drop of each Positive and Negative controls into separate circles on the slide test.
3. Mix the ASO-latex reagent vigorously or on a vortex mixer before using and add one drop (50 μ L) next to the sample to be tested.
4. Mix the drops with a stirrer, spreading them over the entire surface of the circle. Use different stirrers for each sample.
5. Place the slide on a mechanical rotator at 80-100 r.p.m. for 2 minutes. False positive results could appear if the test is read later than two minutes.

Semi-quantitative method

1. Make serial two fold dilutions of the sample in 9 g/L saline solution.
2. Proceed for each dilution as in the qualitative method.

QUALITY CONTROL

Positive and Negative Controls should be included in each test batch.


Acceptable performance is indicated when a uniform milky suspension with no agglutination is observed with the ASO Negative Control and agglutination with large aggregates is observed with the ASO Positive Control.

RESULTS

A.QUALITATIVE TEST:

A negative reaction is indicated by a uniform milky suspension with no agglutination as observed with the ASO Negative Control.

A positive reaction is indicated by any observable agglutination in the reaction mixture. The specimen reaction should be compared to the ASO Negative Control (Fig. 1).



Positive Negative

Figure 1

B. QUANTITATIVE TEST

A positive reaction is indicated by any observable agglutination in the reaction mixture. Record the last dilution showing a positive reaction. Concentration of ASO can be determined by multiplying the last positive dilution factor of the sample with the concentration of the positive control (200 IU/ml).

The titer of the serum is the reciprocal of the highest dilution which exhibits a positive reaction.

IU/ml of sample = conc. of positive control (200) x specimen titer

<u>DILUTION</u>	<u>IU/ml</u>
1:1	200
1:2	400
1:4	800
1:8	1600
Etc.	

REFERENCE VALUES

Up to 200 IU/mL (adults) and 100 IU/mL (children < 5 years old)⁶. Each laboratory should establish its own reference range.

PERFORMANCE CHARACTERISTICS

Analytical sensitivity:

200 (±50) IU/ml.

PROZONE EFFECT

No prozone effect was detected up to 1500 IU/ml.

SENSITIVITY

98%.

SPECIFICITY

97%.

INTERFERENCES

NON INTERFERING SUBSTANCES:

- Hemoglobin (10g/dl)
- Bilirubin (20mg/dl)
- Lipemia (10g/dl)

Other substances may interfere

REFERENCES

1. Haffejee . Quarterly Journal of Medicine 1992. New series 84; 305: 641-658.
2. Ahmed Samir et al. Pediatric Annals 1992; 21: 835-842.
3. Spaun J et al. Bull Wld Hlth Org 1961; 24: 271-279.
4. The association of Clinical Pathologists 1961. Broadsheet 34.
5. Picard B et al. La Presse Medicale 1983; 23: 2-6.
6. Klein GC. Applied Microbiology 1971; 21: 999-1001.
7. Young DS. Effects of drugs on clinical laboratory test, 4th ed. AACC Press, 1995.



ATLAS Medical

William James House,














Cowley Road, Cambridge, CB4 4WX, UK

Tel: ++44 (0) 1223 858 910

Fax: ++44 (0) 1223 858 524

PPI003A01

Rev H (09.09.2017)

	Catalogue Number		Store at
	For In-Vitro Diagnostic use		Caution
	Number of tests in the pack		Read product insert before use
	Lot (batch) number		Manufacturer
	Fragile, handle with care		Expiry date
	Manufacturer fax number		Do not use if package is damaged
	Manufacturer telephone		



浙江东方基因生物制品股份有限公司
Zhejiang Orient Gene Biotech Co., LTD



CE-DOC-OG038
Version 2.0

EC Declaration of Conformity

In accordance with Directive 98/79/EC

Legal Manufacturer: *Zhejiang Orient Gene Biotech Co., Ltd*

Legal Manufacturer Address: *3787#, East Yangguang Avenue, Dipu Street,
Anji 313300, Huzhou, Zhejiang, China*

Declares, that the products
Product Name and Model(s)

Troponin I Rapid Test Cassette (Whole Blood/Serum/Plasma)	GDTRO-402a
Troponin I Rapid Test Cassette (Whole Blood/Serum/Plasma)	GDTRO-402b

Classification: *Other*
Conformity assessment route: *Annex III (EC DECLARATION OF CONFORMITY)*

We, the Manufacturer, herewith declare with sole responsibility that our product/s mentioned above meet/s the provisions of the Directive 98/79/EC of the European Parliament and of the Council on In-Vitro Diagnostic Medical Devices.

We hereby explicitly appoint

EC Representative's Name: *Shanghai International Holding Corp. GmbH (Europe)*

EC Representative's Address: *Eiffestrasse 80, 20537 Hamburg, Germany*

to act as our European Authorized Representative as defined in the aforementioned Directive.

I, the undersigned, hereby declare that the medical devices specified above conform with the directive 98/79/EC on in vitro diagnostic medical devices and pertinent essential requirements

Date Signed: August 11, 2020

Name of authorized signatory: *Joyce Pang*
Position held in the company: *Vice-President*



EC Declaration of Conformity

In accordance with Directive 98/79/EC

Legal Manufacturer: Healgen Scientific Limited Liability Company

Legal Manufacturer Address: 3818 Fuqua Street, Houston, TX 77047, USA.

Declares, that the products
Product Name and Model(s)

Orient Gene HCV Hepatitis C Virus Rapid Test (Serum/Plasma) (Cassette)	GCHCV-302a
Orient Gene HCV Hepatitis C Virus Rapid Test (Whole blood/Serum/Plasma)(Cassette)	GCHCV-402a

EDMA Code: 15 70 02 02

Classification: Annex II List A
Conformity assessment route: Annex IV (Full Quality Assurance)

Compliance of the designated product with the Directive 98/79/EC has been assessed and certified by the Notified Body

Notified Body: TÜV SÜD Product Service GmbH

Notified Body Address: Munich Branch Ridlerstraße 65 80339 München Germany

EC Certificate No.: V1 092378 0004 Rev. 02 Valid until: 2025-05-26

EC Design-Examination Certificate No.: V7 092378 0009 Rev. 00 Valid until: 2025-05-26

It bears the mark

CE 0123

We, the Manufacturer, herewith declare with sole responsibility that our product/s mentioned above meet/s the provisions of the Directive 98/79/EC of the European Parliament and of the Council on In-Vitro Diagnostic Medical Devices.


We hereby explicitly appoint

EC Representative Name: QARAD b.v.b.a.

EC Representative Address: Cipalstraat 3, B-2440 Geel, Belgium

to act as our European Authorized Representative as defined in the aforementioned Directive.

I, the undersigned, hereby declare that the medical devices specified above conform with the directive 98/79/EC on in vitro diagnostic medical devices and pertinent essential requirements

Signature: 
Name of authorized signatory: Joyce Pang
Position held in the company: Vice-President
Date: 2022.4.22



Certificate

No. Q5 092305 0001 Rev. 01

Holder of Certificate: **Zhejiang Orient Gene Biotech Co., Ltd.**
3787#, East Yangguang Avenue, Dipu Street Anji
313300 Huzhou, Zhejiang
PEOPLE'S REPUBLIC OF CHINA

Certification Mark:



Scope of Certificate: **Design and Development, Production and Distribution of In Vitro Diagnostic Reagent and Instrument for the Detection of Drugs of Abuse, Fertility, Infectious Diseases, Oncology, Biochemistry, Cardiac Diseases, Allergic Disease based on Rapid Test, PCR and Liquid Biochip Method.**

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system, which meets the requirements of the listed standard(s). All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with. For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:Q5 092305 0001 Rev. 01

Report No.: SH2198802

Valid from: 2022-04-11

Valid until: 2024-03-16

Date, 2022-04-11

Christoph Dicks

Head of Certification/Notified Body

Certificate

No. Q5 092305 0001 Rev. 01

Applied Standard(s):

EN ISO 13485:2016
Medical devices - Quality management systems -
Requirements for regulatory purposes
(ISO 13485:2016)
DIN EN ISO 13485:2016

Facility(ies):

Zhejiang Orient Gene Biotech Co., Ltd.
3787#, East Yangguang Avenue, Dipu Street Anji, 313300
Huzhou, Zhejiang, PEOPLE'S REPUBLIC OF CHINA

See Scope of Certificate



浙江东方基因生物制品股份有限公司
Zhejiang Orient Gene Biotech Co.,LTD

STATEMENT

We, Zhejiang Orient Gene Biotech Co., Ltd , having a registered office at 3787#, East Yangguang Avenue, Dipu Street Anji 313300, Huzhou, Zhejiang, China assign SRL SANMEDICO having a registered office at A. Corobceanu street 7A, apt. 9, Chişinău MD-2012, Moldova, as non-exclusive authorized representative for Orient Gene Brand product in correspondence with the conditions of directive 98/79/EEC.

We declare that the company mentioned above is authorized to register, notify, renew or modify the registration of medical devices on the territory of the Republic of Moldova.

This Statement letter will be valid from Feb.21th,2023 to Feb.20th, 2024.

Zhejiang Orient Gene Biotech Co., Ltd

General Manager:

Date:2023/2/21



[Handwritten signature in blue ink]

地址：浙江省湖州市安吉县递铺镇阳光大道东段 3787 号
Add: 3787#, East Yangguang Avenue, Dipu Street Anji 313300, Huzhou, Zhejiang, China
电话 Tel:+86-572-5226111 传真 Fax: +86-572-5226222 邮编 P.C.:313300

HCV Ab Rapid Test Cassette (Whole Blood/Serum/Plasma)

INTENDED USE

The HCV Ab Rapid Test Cassette (Whole Blood/Serum/Plasma) is a sandwich lateral flow chromatographic immunoassay for the qualitative detection of antibodies (IgG, IgM, and IgA) anti- Hepatitis C virus (HCV) in human whole blood, serum or plasma. It is intended to be used as a screening test and as an aid in the diagnosis of infection with HCV. Any reactive specimen with the HCV Ab Rapid Cassette must be confirmed with alternative testing method(s) and clinical findings.

INTRODUCTION

Hepatitis C Virus (HCV) is a small, enveloped, positive-sense, single-stranded RNA Virus. Antibody to HCV is found in over 80% of patients with well-documented non-A, non-B hepatitis. Conventional methods fail to isolate the virus in cell culture or visualize it by electron microscope. Cloning the viral genome has made it possible to develop serologic assays that use recombinant antigens ^(1, 2). Compared to the first generation HCV EIAs using single recombinant antigen, multiple antigens using recombinant protein and/or synthetic peptides have been added in new serologic tests to avoid nonspecific cross-reactivity and to increase the sensitivity of the HCV antibody tests ^(3, 4).

HCV Ab Rapid Test Cassette (Whole Blood/Serum/Plasma) is a rapid test to qualitatively detect the presence of antibody to HCV in a whole blood, serum or plasma specimen. The test utilizes a combination of recombinant antigen to selectively detect elevated levels of HCV antibodies in whole blood, serum or plasma.

PRINCIPLE

The HCV Ab Rapid Test Cassette is a lateral flow chromatographic immunoassay based on the principle of the double antigen–sandwich technique. The test cassette consists of: 1) a burgundy colored conjugate pad containing HCV antigens conjugated with colloidal gold (HCV Ag conjugates) and rabbit IgG–gold conjugates, 2) a nitrocellulose membrane strip containing a test band (T band) and a control band (C band). The T band is pre-coated with non-conjugated HCV antigens, and the C band is pre-coated with goat anti-rabbit IgG. When an adequate volume of test specimen is dispensed into the sample well of the cassette, the specimen migrates by capillary action across the cassette. The antibodies: either the IgG, the IgM, or the IgA, to HCV if present in the specimen will bind to the HCV Ag conjugates. The immunocomplex is then captured on the membrane by the pre-coated HCV antigens, forming a burgundy colored T band, indicating a HCV Ab positive test result. Absence of the T band suggests a negative result. The test contains an internal control (C band) which should exhibit a burgundy colored band of the immunocomplex of goat anti-rabbit IgG/rabbit IgG–gold conjugate regardless the presence of any antibodies to HCV. Otherwise, the test result is invalid and the specimen must be retested with another device.

PRODUCT CONTENTS

HCV Ab Rapid Test Cassette (Whole Blood/Serum/Plasma) containing HCV antigen coated particles and HCV antigen coated on the membrane.

MATERIALS SUPPLIED

1. Test Strip 2. Pipette Dropper 3.Desiccant 4.Buffer 5.Package Insert

MATERIAL REQUIRED BUT NOT PROVIDED

1.Specimen collection containers 2.Lancets (for fingerstick whole blood only)
3.Centrifuge (for plasma only) 4.Timer
5.Heparinized capillary tubes and dispensing bulb (for fingerstick whole blood only)

STORAGE AND STABILITY

The kit can be stored at room temperature or refrigerated (2-30°C). The test device is stable through the expiration date printed on the sealed pouch. The test device must remain in the sealed pouch until use. DO NOT FREEZE. Do not use beyond the expiration date.

WARNINGS AND PRECAUTIONS

1.For professional In Vitro diagnostic use only. Do not use after expiration date.
2.Warning: the reagents in this kit contain sodium azide which may react with lead or copper plumbing to form potentially explosive metal azides. When disposing of such reagents, always flush with large volumes of water to

prevent azide build-up.

3. Do not use it if the tube/pouch is damaged or broken.

4.Test is for single use only. Do not re-use under any circumstances.

5.Handle all specimens as if they contain infectious agents.Observe established precautions against microbiological hazards throughout testing and follow the standard procedures for proper disposal of specimens.

6. Wear protective clothing such as laboratory coats, disposable gloves and eye protection when specimens are assayed.

7. Humidity and temperature can adversely affect results .

SPECIMEN COLLECTION

1.The HCV Rapid Test Cassette (Whole Blood/Serum/Plasma) can be performed using whole blood (from venipuncture or fingerstick), serum or plasma.

2.To collect Fingerstick Whole Blood specimens:

- Wash the patient's hand with soap and warm water or clean with an alcohol swab. Allow to dry.

- Massage the hand without touching the puncture site by rubbing down the hand towards the fingertip of the middle or ring finger.

- Puncture the skin with a new sterile lancet for each person. Wipe away the first sign of blood.

- Gently rub the hand from wrist to palm to finger to form a rounded drop of blood over the puncture site.

- Add the Fingerstick Whole Blood specimen to the test device by using a capillary tube:

- Touch the end of the capillary tube to the blood until filled to approximately 50 µL. Avoid air bubbles.

- Place the bulb onto the top end of the capillary tube, then squeeze the bulb to dispense the whole blood into the specimen well (S) of the test device.

- Add the Fingerstick Whole Blood specimen to the test device by using hanging drops:

- Position the patient's finger so that the drop of blood is just above the specimen well (S) of the test device.

- Allow 2 hanging drops of fingerstick whole blood to fall into the center of specimen well (S) on the test device or, move the patient's finger so that the hanging drop touches the center of the specimen well (S). Avoid touching the finger directly to the specimen well (S).

3.Separate serum or plasma from blood as soon as possible to avoid hemolysis. Use only clear, non-hemolyzed specimens.

4.Testing should be performed immediately after specimen collection. Do not leave the specimens at room temperature for prolonged periods. Serum and plasma specimens may be stored at 2-8°C for up to 3 days. For long term storage, specimens should be kept below -20°C. Whole blood collected by venipuncture should be stored at 2-8°C if the test is to be run within 2 days of collection. Do not freeze whole blood specimens. Whole blood collected by fingerstick should be tested immediately.

5.Bring specimens to room temperature prior to testing. Frozen specimens must be completely thawed and mixed well prior to testing. Specimens should not be frozen and thawed repeatedly.

6.If specimens are to be shipped, they should be packed in compliance with local regulations covering the transportation of etiologic agents.

TEST PROCEDURE

Allow test device, specimen, buffer and/or controls to equilibrate to room temperature (15-30°C) prior to testing.

1.Remove the test device from the foil pouch and use it as soon as possible. Best results will be obtained if the assay is performed within one hour.

2. Place the test device on a clean and level surface.


For Serum or Plasma Specimens: Hold the dropper vertically and transfer 1 drop of serum or plasma (approximately 30 µL) to the specimen well (S) of the test device, then add 1 drop of buffer (approximately 40 µL) and start the timer. See illustration below.

For Venipuncture Whole Blood Specimens: Hold the dropper vertically and transfer 2 drops of venipuncture whole blood (approximately 50µL) to the specimen well (S) of the test device, then add 1 drop of buffer (approximately 40 µL) and start the timer. See illustration below.

For Fingerstick Whole Blood Specimens: Allow 2 hanging drops of fingerstick whole blood (approximately 50 µ L) to fall into the center of the specimen well (S) on the test device, then add 1 drop of buffer (approximately 40 µ L) and start the timer. See illustration below.

HCV Ab Rapid Test Cassette (Whole Blood/Serum/Plasma)

3. Wait for the red line(s) to appear. The result should be read in 15 minutes. Do not interpret the result after 15 minutes.



INTERPRETATION OF RESULTS

(please refer to the illustration above)

Positive: Two lines appear. One colored line should be in the control line region (C) and another apparent colored line should be in the test line region (T).

Negative: One colored line appears in the control line region (C). No line appears in the test line region (T).

Invalid: Control line fails to appear. Insufficient specimen volume or incorrect procedural techniques are the most likely reasons for control line failure. Review the procedure and repeat the test with a new test device. If the problem persists, discontinue using the test kit immediately and contact your local distributor.

QUALITY CONTROL

A procedural control is included in the test. A red line appearing in the control region (C) is the internal procedural control. It confirms sufficient specimen volume and correct procedural technique. Control standards are not supplied with this test. However, it is recommended that positive and negative controls are sourced from a local competent authority and tested as a good laboratory practice, to confirm the test procedure and verify the test performance.

LIMITATIONS

1. The HCV Ab Rapid Test Cassette (Whole Blood/ Serum/Plasma) is for in vitro diagnostic use only. This test should be used for the detection of antibodies to HCV in whole blood, serum or plasma specimen.
2. The HCV Ab Rapid Test Cassette (Whole Blood/Serum/Plasma) will only indicate the presence of antibodies to HCV in the specimen and should not be used as the sole criteria for the diagnosis of Hepatitis C viral infection.
3. As with all diagnostic tests, all results must be considered with other clinical information available to the physician.
4. If the test result is negative and clinical symptoms persist, additional follow-up testing using other clinical methods is recommended. A negative result at any time does not preclude the possibility of Hepatitis C Virus infection.
5. A negative result can occur if the quantity of the antibodies to HCV present in the specimen is below the detection limits of the assay, or the antibodies that are detected are not present during the stage of disease in which a sample is collected.
6. Some specimens containing unusually high titer of heterophile antibodies or rheumatoid factor may affect expected results.

PERFORMANCE CHARACTERISTICS

Sensitivity: HCV Ab Rapid Test Cassette (Whole Blood/ Serum/Plasma) has passed a seroconversion panel and compared with leading commercial HCV EIA test using clinical specimens.

Specificity: The recombinant antigens used for HCV Ab Rapid Test Cassette (Whole Blood/Serum/Plasma) are encoded by genes for both structural (nucleocapsid) and non-structural proteins. HCV Ab Rapid Test Cassette (Whole Blood/Serum/Plasma) is highly specific for antibodies to Hepatitis C Virus compared with a leading

commercial HCV EIA test.

The HCV Ab Rapid Test Cassette vs EIA test

Method		EIA		Total Results
HCV Ab RapidTest	Results	Positive	Negative	
	Positive	105	19	
	Negative	2	1760	
Total Results		107	1779	1886

Relative sensitivity: 98.1%

Relative specificity: 98.9%

Accuracy: 98.9%

REFERENCE

1. Choo, Q.L., G.Kuo,A.J. Weiner, L.R. Overby,D.W. Bradley, andM. Houghton. Isolation of a cDNA clone derived from a blood-borne non-A, non-B viral hepatitis genome Science 189;244:359
2. Kuo, G., Q.L. Choo, H.J. Alter, and M. Houghton. An assay for circulating antibodies to a major etiolog Virus of human non-A, non-B hepatitis. Science 1989; 244:362.
3. Van der Poel, C.L., H.T.M. Cuypers, H.W. Reesink, and P.N. Lelie .Confirmation of hepatitis C Virus infection by new four- antigen recombinant immunoblot assay. Lancet 1991;337:317
4. Wilber, J.C.Development and use of laboratory tests for hepatitis C infection: a review.J. Clin. Immunoassy 1993;16:204.

Troponin I

Troponin I Rapid Test Device (Whole Blood/Serum/Plasma) Package Insert

A rapid visual immunoassay for the qualitative presumptive detection of cardiac Troponin I in human whole blood, serum, or plasma specimens.
For professional in vitro diagnostic use only.

INTENDED USE

The Troponin I Rapid Test Device (Whole Blood/Serum/Plasma) is a rapid visual immunoassay for the qualitative presumptive detection of cardiac Troponin I in human whole blood, serum, or plasma specimens. This kit is intended to be used as an aid in the diagnosis of myocardial infarction (MI).

SUMMARY

Cardiac Troponin I (cTnI) is a protein found in cardiac muscle with a molecular weight of 22.5 kDa.¹ Troponin I is part of a three subunit complex comprising of Troponin T and Troponin C. Along with tropomyosin, this structural complex forms the main component that regulates the calcium sensitive ATPase activity of actomyosin in striated skeletal and cardiac muscle.² After cardiac injury occurs, Troponin I is released into the blood 4-6 hours after the onset of pain. The release pattern of cTnI is similar to CK-MB, but while CK-MB levels return to normal after 72 hours, Troponin I remains elevated for 6-10 days, thus providing for a longer window of detection for cardiac injury. The high specificity of cTnI measurements for the identification of myocardial damage has been demonstrated in conditions such as the perioperative period, after marathon runs, and blunt chest trauma.³ cTnI release has also been documented in cardiac conditions other than acute myocardial infarction (AMI) such as unstable angina, congestive heart failure, and ischemic damage due to coronary artery bypass surgery.⁴ Because of its high specificity and sensitivity in the myocardial tissue, Troponin I has recently become the most preferred biomarker for myocardial infarction.⁵

PRINCIPLE

The Troponin I Rapid Test Device (Whole Blood/Serum/Plasma) has been designed to detect cardiac Troponin I through visual interpretation of color development in the strip. The membrane was immobilized with anti-cTnI antibodies on the test region. During the test, the specimen is allowed to react with colored anti-cTnI antibodies colloidal dog conjugates, which were precoated on the sample pad of the test. The mixture then moves on the membrane by a capillary action, and interact with reagents on the membrane. If there were enough cTnI in specimens, a colored band will form at the test region of the membrane.

Presence of this colored band indicates a positive result, while its absence indicates a negative result. Appearance of a colored band at the control region serves as a procedural control. This indicates that proper volume of specimen has been added and membrane wicking has occurred.

PRECAUTIONS

- For professional In Vitro diagnostic use only.
- Warning: the reagents in this kit contain sodium azide which may react with lead or copper plumbing to form potentially explosive metal azides. When disposing of such reagents, always flush with large volumes of water to prevent azide build-up.
- Do not use it if the tube/pouch is damaged or broken.
- Test is for single use only. Do not re-use under any circumstances.
- Handle all specimens as if they contain infectious agents. Observe established standard procedure for proper disposal of specimens
- Wear protective clothing such as laboratory coats, disposable gloves and eye protection when specimens are assayed.
- Humidity and temperature can adversely affect results

STORAGE AND STABILITY

All reagents are ready to use as supplied. Store unused test device unopened at 2°C-30°C. If stored at 2°C-8°C, ensure that the test device is brought to room temperature before opening. The test is not stable out off the expiration date printed on the sealed pouch. Do not freeze the kit or expose the kit over 30°C.

SPECIMEN COLLECTION AND PREPARATION

- The Troponin I Rapid Test Device (Whole Blood/Serum/Plasma) is intended only for use with human whole blood, serum, or plasma specimens.
- Only clear, non-hemolyzed specimens are recommended for use with this test.
- Serum or plasma should be separated with soonest possible opportunity to avoid hemolysis.
- Perform the testing immediately after the specimen collection. Do not leave the specimens at room temperature for prolonged periods. Specimens may be stored at 2-8°C for up to 3 days. For long term storage, specimens should be kept below -20°C.
- Bring specimens to room temperature prior to testing. Frozen specimens must be completely thawed and mixed well prior to testing. Avoid repeated freezing and thawing of specimens.

- Pack the specimens in compliance with applicable regulations for transportation of etiological agents, in case they need to be shipped.
- Icteric, lipemic, hemolyzed, heat treated and contaminated sera may cause erroneous results.
- There is a slight possibility that some whole blood specimens with very high viscosity or which have been stored for more than 2 days may not run properly on the test device. Repeat the test with a serum or plasma specimen from the same patient using a new test device.

MATERIALS

Materials Provided

- Test devices
- Buffer

- Disposable Droppers
- Package insert

Materials Required But Not Provided

- Specimen collection containers
- Centrifuge (for plasma only)

- Clock or Timer

DIRECTIONS FOR USE

Allow test device, specimen, buffer and/or controls to equilibrate to room temperature (15-30°C) prior to testing.

- Remove the test from its sealed pouch, and place it on a clean, level surface. Label the device with patient or control identification. To obtain a best result, the assay should be performed within one hour.

- Transfer 2 drops of serum or plasma to the specimen well of the device with a disposable pipette provided in the kit, and then start the timer.

OR


Transfer 3 drops of whole blood specimen to the specimen well of the device with a disposable pipette provided in the kit, then add 1 drop of buffer, and start the timer.

OR

Allow 3 hanging drops of fingerstick whole blood specimen to fall into the center of the specimen well (S) on the device, then add 1 drop of buffer, and start the timer. Avoid trapping air bubbles in the specimen well (S), and do not drop any solution in observation window.

As the test begins to work, you will see color move across the membrane.

- Wait for the colored band(s) to appear. The result should be read at 10 minutes. Do not interpret the result after 20 minutes.



INTERPRETATION OF RESULTS

(Please refer to the illustration above)

POSITIVE: Two colored bands appear on the membrane. One band appears in the control region (C) and another band appears in the test region (T).

NEGATIVE: Only one colored band appears in the control region (C). No apparent colored band appears in the test region (T).

INVALID: Control band fails to appear. Results from any test which has not produced a control band at the specified reading time must be discarded.

Please review the procedure and repeat with a new test. If the problem persists, discontinue using the kit immediately and contact your local distributor.

NOTE:

- The intensity of the color in test region (T) may vary depending on the concentration of aimed substances present in the specimen. Therefore, any shade of color in the test region should be considered positive. Besides, the substances level can not be determined by this qualitative test.
- Insufficient specimen volume, incorrect operation procedure, or performing expired tests are the most likely reasons for control band failure.

QUALITY CONTROL

Internal procedural controls are included in the test. A colored band appearing in the control region (C) is considered an internal positive procedural control. It confirms sufficient specimen volume and correct procedural technique.

External controls are not supplied with this kit. It is recommended that positive and negative controls be tested as a good laboratory practice to confirm the test procedure and to verify proper test performance.

LIMITATIONS

- The Troponin I Rapid Test Device (Whole Blood/Serum/Plasma) is for professional in vitro diagnostic use, and should be used for the qualitative detection of cardiac Troponin I only. There is no meaning attributed to linen color intensity or width.
- The Troponin I Rapid Test Device (Whole Blood/Serum/Plasma) will only indicate the presence of Troponin I in the specimen and should not be used as the sole criteria for the diagnosis of tuberculosis.
- If the test result is negative and clinical symptoms persist, additional testing using other clinical methods is recommended. The test cannot detect less than 0.5 ng/mL of cTnI in specimens. Thus, a negative result does not at anytime rule out the existence of Troponin I in blood, because the antibodies may be absent or below the minimum detection level of the test.
- Like with all diagnostic tests, a confirmed diagnosis should only be made by a physician after all clinical and laboratory findings have been evaluated.
- Some specimens containing unusually high titers of heterophile antibodies or rheumatoid factor (RF) may affect expected results. Even if the test results are positive, further clinical evaluation should be considered with other clinical information available to the physician.

PERFORMANCE CHARACTERISTICS

Table: Troponin I Rapid Test vs. EIA

Method	Troponin I Rapid Test Device		Total Results
	Results	Positive	Negative
EIA	Positive	138	2
	Negative	1	315
Total Results		139	317
		456	

Relative Sensitivity: 98.6% (94.9%-99.8%)*

Relative Specificity: 99.7% (98.3%-99.9%)*

Overall Agreement: 99.3% (98.1%-99.9%)*

*95% Confidence Interval

BIBLIOGRAPHY

- Adams, et al. Biochemical markers of myocardial injury, Immunoassay Circulation 88: 750-763, 1993.
- Mehegan JP, Tobacman LS. Cooperative interaction between troponin molecules bound to the cardiac thin filament. J.Biol.Chem. 266:966, 1991.
- Adams, et al. Diagnosis of Perioperative myocardial infarction with measurements of cardiac troponin I. N.Eng.J.Med 330:670, 1994.
- Hossein-Nia M, et al. Cardiac troponin I release in heart transplantation. Ann. Thorac. Surg. 61: 227, 1996.
- Alpert JS, et al. Myocardial Infarction Redefined, Joint European Society of Cardiology American College of Cardiology: J. Am. Coll. Cardio., 36(3):959, 2000.



ФЕДЕРАЛЬНАЯ СЛУЖБА ПО НАДЗОРУ В СФЕРЕ ЗДРАВООХРАНЕНИЯ
(РОСЗДРАВНАДЗОР)

РЕГИСТРАЦИОННОЕ УДОСТОВЕРЕНИЕ НА МЕДИЦИНСКОЕ ИЗДЕЛИЕ

от 07 мая 2019 года № РЗН 2019/8352

На медицинское изделие

**Индикаторы химические для контроля процесса паровой и воздушной
стерилизации по ТУ 20.59.52-001-35927791-2017**

Настоящее регистрационное удостоверение выдано

**Общество с ограниченной ответственностью "Научно-Производственное
Объединение "Маркер" (ООО "НПО "Маркер"), Россия,
117292, Москва, ул. Профсоюзная, д. 26/44**

Производитель

**Общество с ограниченной ответственностью "Научно-Производственное
Объединение "Маркер" (ООО "НПО "Маркер"), Россия,
117292, Москва, ул. Профсоюзная, д. 26/44**

Место производства медицинского изделия

**ООО «НПО Маркер», Россия, 300013, г. Тула, Привокзальный р-н,
ул. Болдина, д. 98а, лит. Е**

Номер регистрационного досье № РД-25642/72833 от 30.01.2019

Класс потенциального риска применения медицинского изделия 1

Код Общероссийского классификатора продукции по видам экономической
деятельности 32.50.50.000

Настоящее регистрационное удостоверение имеет приложение на 2 листах

приказом Росздравнадзора от 07 мая 2019 года № 3413
допущено к обращению на территории Российской Федерации

**Врио руководителя Федеральной службы
по надзору в сфере здравоохранения**



Д.В. Пархоменко

0039607

ПРИЛОЖЕНИЕ К РЕГИСТРАЦИОННОМУ УДОСТОВЕРЕНИЮ НА МЕДИЦИНСКОЕ ИЗДЕЛИЕ

от 07 мая 2019 года

№ РЗН 2019/8352

Лист 1

На медицинское изделие

Индикаторы химические для контроля процесса паровой и воздушной стерилизации по ТУ 20.59.52-001-35927791-2017, в вариантах исполнения:

1. Индикаторы химические для контроля процесса паровой и воздушной стерилизации, в составе:

1.1. Интегрирующий индикатор «Маркер», 5 класс для контроля процесса паровой и воздушной стерилизации.

1.2. Многопеременный индикатор «ХимТест», 4 класс для контроля параметров паровой стерилизации для режимов: 121 °C /20 мин, 126 °C /10 мин, 134 °C /5 мин.

1.3. Многопеременный индикатор «ХимТест», 4 класс для контроля параметров воздушной стерилизации для режимов: 160 °C /150 мин, 180 °C /60 мин, 200 °C /30 мин.

1.4. Имитирующий индикатор «Маркер-Прион», 6 класс для контроля параметров паровой стерилизации для режима: 134 °C /18 мин.

2. Индикаторы химические для контроля процесса паровой и воздушной стерилизации лекарственных средств, в составе:

2.1. Многопеременный индикатор «Маркер-Фарм», 4 класс для контроля параметров паровой и воздушной стерилизации для режимов: 100 °C /30 мин, 110 °C /20 мин, 120 °C /15 мин, 180 °C /30 мин.

2.2. Многопеременный индикатор «ХимТест-Фарм-1», 4 класс для контроля параметров паровой стерилизации для режимов: 100 °C /15 мин, 110 °C /10 мин, 120 °C /8 мин.

2.3. Многопеременный индикатор «ХимТест-Фарм-2», 4 класс для контроля параметров паровой стерилизации для режимов: 110 °C /15 мин, 120 °C /12 мин.

2.4. Многопеременный индикатор «ХимТест-Фарм-3», 4 класс для контроля параметров паровой стерилизации для режимов: 100 °C /30 мин, 110 °C /20 мин, 120 °C /15 мин.

2.5. Многопеременный индикатор «ХимТест-Фарм-4», 4 класс для контроля параметров паровой стерилизации для режимов: 112 °C /20 мин, 121 °C /15 мин.

2.6. Многопеременный индикатор «ХимТест-Фарм-5», 4 класс для контроля параметров паровой стерилизации для режимов: 130 °C /30 мин, 121 °C /20 мин.

2.7. Многопеременный индикатор «ХимТест-Фарм-6», 4 класс для контроля параметров паровой стерилизации для режима: 120 °C /30 мин.

2.8. Многопеременный индикатор «ХимТест-Фарм-7», 4 класс для контроля

**Врио руководителя Федеральной службы
по надзору в сфере здравоохранения**

Д.В. Пархоменко

0055896

**ПРИЛОЖЕНИЕ
К РЕГИСТРАЦИОННОМУ УДОСТОВЕРЕНИЮ
НА МЕДИЦИНСКОЕ ИЗДЕЛИЕ**

от 07 мая 2019 года

№ РЗН 2019/8352

Лист 2

параметров воздушной стерилизации для режима: 180 °С /30 мин.

2.9. Многопеременный индикатор «ХимТест-Фарм-8», 4 класс для контроля параметров воздушной стерилизации для режима: 180 °С /45 мин.

3. Индикаторы химические для контроля процесса стерилизации (парового обеззараживания) медицинских отходов, в составе:

3.1. Многопеременный индикатор «ХимТест-О-1», для контроля параметров парового обеззараживания для режимов: 120 °С /90 мин, 126 °С /60 мин, 132 °С /45 мин, 134 °С /27 мин.

3.2. Многопеременный индикатор «ХимТест-О-2», для контроля параметров парового обеззараживания для режимов: 120 °С /120 мин, 126 °С /90 мин, 132 °С /60 мин, 134 °С /35 мин.

3.3. Многопеременный индикатор «ХимТест-О-3», для контроля параметров парового обеззараживания для режимов: 132 °С /90 мин, 134 °С /60 мин.

Врио руководителя Федеральной службы
по надзору в сфере здравоохранения



Д.В. Пархоменко

0055897

ООО «Научно-Производственное Объединение Маркер»

ИНН: 7728890217

КПП: 772801001

ОГРН: 5147746104182

117292, г. Москва, ул. Профсоюзная, д. 26/44

тел.: +7 (495) 178-02-08; e-mail: info@npomarker.ru

Индикаторы химические для контроля процесса паровой и воздушной стерилизации
ТУ 20.59.52-001-35927791-2017

ПАСПОРТ

03.03.2020

Индикаторы химические для контроля процесса паровой и воздушной стерилизации:
многопеременный индикатор «ХимТест», 4 класс для контроля параметров воздушной
стерилизации для режимов: 160 °C /150 мин, 180 °C /60 мин, 200 °C /30 мин;

Партия № 2503/2

Дата изготовления: март 2020 г.

Годен до: март 2025 г.

Вид исполнения: листы с индикаторами

Результаты приемосдаточных испытаний

Наименование испытаний (проверок)	№№ пунктов ТУ (технических требований)	Результат испытаний
Проверка соответствия комплекту документации	1.1.1	соответствует
Проверка исполнений, общего внешнего вида, конструкции, формы, материалов, основных размеров, массы	1.2.1-1.2.3	соответствует
Проверка условий достижения конечного состояния	1.2.4, 1.2.5	соответствует
Проверка условий не достижения конечного состояния	1.2.6	соответствует
Проверка комплектности, маркировки и упаковки	1.3, 1.4, 1.5	соответствует

Генеральный директор ООО «НПО Маркер»

И.П. Антонова



SYNTESYS



SYNTESYS S.A.S. DI RINALDO R. & C.
VIA G. GALILEI, 10/3
35037 Z.I. SELVE DI TEOLO (PD)
TEL. +39 049 9903866 R.A. FAX +39 049 9903867
COD.FISCALE P.IVA N.REG.IMP. PADOVA 03573950288
E-MAIL INFO@SYNTESYS.IT - WEB WWW.SYNTESYS.IT

DICHIARAZIONE DI CONFORMITA'
Conformity declaration



Il sottoscritto, Rinaldo Ruggero legale rappresentante della ditta:
The undersigned, Rinaldo Ruggero legal representative of the company:

produttore/manufacturer

SYNTESYS S.a.s. di Rinaldo Ruggero & C.
indirizzo/address

Via G. Galilei, 10/3 35037 Zona Industriale SELVE DI TEOLO (PADOVA) ITALY

o rappresentante il mandatario autorizzato entro la Unione Europea or representing the authorized mandatary within the European Community

Mandatario autorizzato/authorized mandatary

indirizzo/address

Dichiara sotto la propria responsabilità che il prodotto/*declares under his own responsibility that the product:*

Denominazione degli
articoli
prodotti/*Description of
Manufacturer*

Contenitori per urina, contenitori per feci,
contenitori universali, Pipette Pasteur, Piastre di
Petri, Anse Sterili per batteriologia, Aste a "L",
Puntali Eppendorf gialli e blue, cuvette per
spettrofotometro, tazzine per campionamento siero,
bacchette per distacco ed estrazione del coagulo,
pinzette in polistirolo monouso, provette monouso in
plastica, tappi alettati per provette diam. 12 mm e
16mm, provette con granuli ed acceleratore, provette
sottovuoto per prelievo, Sistema SEDIPLAST,
Microprovette, Portavetrini, Vetrini precolorati,
Portaprovette, supporti per microprovette, bottiglie
per raccolta urine.

*Urine container, faeces container, universal
container, Pasteur pipette, Petri dishes, Sterile
loops, Sterile loops open "L", Eppendorf tips yellow
and blue, cuvettes for spectrophotometer, samples
cups, Rod to detach clot, disposable forceps,
Disposable plastic tubes, winged stoppers for tubes
diam. 12mm & 16mm, Test tube with granules and clot
activator, vacuum test tube, SEDIPLAST system,
micro test tubes, Slides Mailer, "TESTSIMPLETS" slide
rack for test tubes, rack for micro test tubes,
Bottles for urine collection.*



SYNTESYS



SYNTESYS S.A.S. DI RINALDO R. & C.
VIA G. GALILEI, 10/3
35037 Z.I. SELVE DI TEOLO (PD)

TEL. +39 049 9903866 R.A. FAX +39 049 9903867

COD.FISCALE P.IVA N.REG.IMP. PADOVA 03573950288
E-MAIL INFO@SYNTESYS.IT - WEB WWW.SYNTESYS.IT

Materiale/Material

**Polipropilene, Polistirolo, Polietilene e
Polimetilmetacrilato**

***Polypropylene, Polystyrene, Polyethylene and
Polymethylmetacrylate***

È conforme alle disposizioni della direttiva 98/79/CE concernente i dispositivi medici diagnostici in vitro e recepito in Italia con D.L. del 08/09/2000 n° 332 allegato 1 (requisiti essenziali) ed è fabbricato in accordo ai requisiti di cui all'Allegato III della sopra citata direttiva / *It meets the CE Directive 98/79 CE about in vitro diagnostic device specifications established by the Italian law n. 332, dated 8th September 2000. The device is made according to the specifications of the III attached of the above-mentioned directive.*

Dichiara inoltre che la documentazione tecnica di supporto alla presente dichiarazione di conformità è conservata presso gli uffici dell'azienda e sarà posta alla disposizione di chi la richiede/declares that all technical documents attached to this conformity statement are filed in our company and can be consulted by any authorized body on demand.

Data 07/01/2016

Issued on January 7th 2016

SYNTESYS S.a.s.
Il legale rappresentante
Rinaldo Ruggero



CISQ is a member of



The International Certification Network
www.iqnet-certification.com

CERTIFICATO N. **6574/3**
CERTIFICATE No. _____

SI CERTIFICA CHE IL SISTEMA DI GESTIONE PER LA QUALITÀ DI
WE HEREBY CERTIFY THAT THE QUALITY MANAGEMENT SYSTEM OPERATED BY

SYNTESYS S.R.L.

Sede e Unità Operativa

Via G. Galilei, 10/1-2-3 - Zona Industriale - 35037 Selve di Teolo (PD) – Italia

Commercializzazione di prodotti per analisi di laboratorio. Produzione di prodotti per analisi di laboratorio e articoli sanitari. Progettazione e gestione della produzione di tamponi sterili per la raccolta e la conservazione di campioni biologici, anche in ambito chirurgico, con o senza terreno di trasporto.

Unità Operative

Via G. Galilei, 16/1 - Zona Industriale - 35037 Selve di Teolo (PD) – Italia *

Via San Benedetto, 48/A - Zona Industriale - 35037 Selve di Teolo (PD) – Italia *

Via G. Galilei, 3 - Zona Industriale - 35037 Selve di Teolo (PD) – Italia *

* Magazzino.

È CONFORME ALLA NORMA / IS IN COMPLIANCE WITH THE STANDARD

UNI EN ISO 9001:2015

Sistema di Gestione per la Qualità / Quality Management System

PER LE SEGUENTI ATTIVITÀ / FOR THE FOLLOWING ACTIVITIES

EA: 29 - 14

Commercializzazione di prodotti per analisi di laboratorio. Produzione di prodotti per analisi di laboratorio e articoli sanitari. Progettazione e gestione della produzione di tamponi sterili per la raccolta e la conservazione di campioni biologici, anche in ambito chirurgico, con o senza terreno di trasporto.

Trading of products for laboratory analysis. Manufacturing of products for laboratory analysis and sanitary products. Design and production management of sterile swabs for the collection and the preservation of biological samples, also for surgical application, with or without transport medium.

Riferirsi alla documentazione del Sistema di Gestione per la Qualità aziendale per l'applicabilità dei requisiti della norma di riferimento.

Refer to the documentation of the Quality Management System for details of application to reference standard requirements.

Il presente certificato è soggetto al rispetto del documento ICIM "Regolamento per la certificazione dei sistemi di gestione" e al relativo Schema specifico.

The use and the validity of this certificate shall satisfy the requirements of the ICIM document "Rules for the certification of company management systems" and specific Scheme.

Per informazioni puntuali e aggiornate circa eventuali variazioni intervenute nello stato della certificazione di cui al presente certificato,

si prega di contattare il n° telefonico +39 02 725341 o indirizzo e-mail info@icim.it.

For timely and updated information about any changes in the certification status referred to in this certificate, please contact the number +39 02 725341 or email address info@icim.it.

DATA EMISSIONE
FIRST ISSUE
05/06/2013

EMISSIONE CORRENTE
CURRENT ISSUE
05/06/2022

DATA DI SCADENZA
EXPIRING DATE
04/06/2025


Vincenzo Delacqua
Rappresentante Direzione / Management Representative
ICIM S.p.A.

Piazza Don Enrico Mapelli, 75 – 20099 Sesto San Giovanni (MI)
www.icim.it



SGQ N° 004 A



www.cisq.com

CISQ è la Federazione Italiana di Organismi di
Certificazione dei sistemi di gestione aziendale. CISQ
is the Italian Federation of management system
Certification Bodies.



CISQ is a member of



The International Certification Network
www.iqnet-certification.com

CERTIFICATO n. **7111/3**
CERTIFICATE No. _____

SI CERTIFICA CHE IL SISTEMA DI GESTIONE PER LA QUALITÀ DI
WE HEREBY CERTIFY THAT THE QUALITY MANAGEMENT SYSTEM OPERATED BY

SYNTESYS S.R.L.

Sede e Unità Operativa

Via G. Galilei, 10/1-2-3 - Zona Industriale - 35037 Selve di Teolo (PD) – Italia
Commercializzazione di prodotti per analisi di laboratorio. Produzione di prodotti per analisi di laboratorio e articoli sanitari. Progettazione e gestione della produzione di tamponi sterili per la raccolta e la conservazione di campioni biologici, anche in ambito chirurgico, con o senza terreno di trasporto.

Unità Operative

Via G. Galilei, 16/1 - Zona Industriale - 35037 Selve di Teolo (PD) – Italia *
Via San Benedetto, 48/A - Zona Industriale - 35037 Selve di Teolo (PD) – Italia *
Via G. Galilei, 3 - Zona Industriale - 35037 Selve di Teolo (PD) – Italia *
* Magazzino.

È CONFORME ALLA NORMA / IS IN COMPLIANCE WITH THE STANDARD

UNI CEI EN ISO 13485:2016

Sistema di Gestione per la Qualità / Quality Management System

PER LE SEGUENTI ATTIVITÀ / FOR THE FOLLOWING ACTIVITIES

Commercializzazione di prodotti per analisi di laboratorio. Produzione di prodotti per analisi di laboratorio e articoli sanitari. Progettazione e gestione della produzione di tamponi sterili per la raccolta e la conservazione di campioni biologici, anche in ambito chirurgico, con o senza terreno di trasporto.

Trading of products for laboratory analysis. Manufacturing of products for laboratory analysis and sanitary products. Design and production management of sterile swabs for the collection and the preservation of biological samples, also for surgical application, with or without transport medium.

Riferirsi alla documentazione del Sistema di Gestione per la Qualità aziendale per l'applicabilità dei requisiti della norma di riferimento.
Refer to the documentation of the Quality Management System for details of application to reference standard requirements.

Il presente certificato è soggetto al rispetto del documento ICIM "Regolamento per la certificazione dei sistemi di gestione" e al relativo Schema specifico.
The use and the validity of this certificate shall satisfy the requirements of the ICIM document "Rules for the certification of company management systems" and Specific Scheme.

Per informazioni puntuali e aggiornate circa eventuali variazioni intervenute nello stato della certificazione di cui al presente certificato, si prega di contattare il n° telefonico +39 02 725341 o indirizzo e-mail info@icim.it.

For timely and updated information about any changes in the certification status referred to in this certificate, please contact the number +39 02 725341 or email address info@icim.it.

DATA EMISSIONE
FIRST ISSUE
21/06/2014

EMISSIONE CORRENTE
CURRENT ISSUE
05/06/2022

DATA DI SCADENZA
EXPIRING DATE
04/06/2025


Vincenzo Delacqua
Rappresentante Direzione / Management Representative
ICIM S.p.A.
Piazza Don Enrico Mapelli, 75 – 20099 Sesto San Giovanni (MI)
www.icim.it



SGQ N° 004 A



www.cisq.com

CISQ è la Federazione Italiana di Organismi di
Certificazione dei sistemi di gestione aziendale. CISQ
is the Italian Federation of management system
Certification Bodies.

Certificate

CISQ/ICIM S.P.A. has issued an IQNet recognized certificate that the organization:

SYNTESYS S.R.L.

Head Office and Operative Unit

Via G. Galilei, 10/1-2-3 - Zona Industriale - I-35037 Selve di Teolo (PD)

Operative Units

Via G. Galilei, 16/1 - Zona Industriale - I-35037 Selve di Teolo (PD)

Via San Benedetto, 48/A - Zona Industriale - I-35037 Selve di Teolo (PD)

Via G. Galilei, 3 - Zona Industriale - I-35037 Selve di Teolo (PD)

has implemented and maintains a/an

Quality Management System

for the following scope:

Trading of products for laboratory analysis. Manufacturing of products for laboratory analysis and sanitary products. Design and production management of sterile swabs for the collection and the preservation of biological samples, also for surgical application, with or without transport medium.

which fulfils the requirements of the following standard:

ISO 9001:2015

Issued on: **2022-06-05**

First issued on: **2013-06-05**

Expires on: **2025-06-04**

This attestation is directly linked to the IQNet Partner's original certificate and shall not be used as a stand-alone document.

Registration Number: **IT-83562**



Alex Stoichitoiu
President of IQNET



Mario Romersi
President of CISQ



This attestation is directly linked to the IQNET Member's original certificate and shall not be used as a stand-alone document.

IQNET Members*:

AENOR Spain **AFNOR Certification** France **APCER** Portugal **CCC** Cyprus **CISQ** Italy **CQC** China **CQM** China **CQS** Czech Republic
Cro Cert Croatia **DQS Holding GmbH** Germany **EAGLE Certification Group** USA **FCAV** Brazil **FONDONORMA** Venezuela **ICONTEC**
Colombia **ICS** Bosnia and Herzegovina **Inspecta Sertifointi Oy** Finland **INTECO** Costa Rica **IRAM** Argentina **JQA** Japan **KFQ** Korea
LSQA Uruguay **MIRTEC** Greece **MSZT** Hungary **Nemko AS** Norway **NSAI** Ireland **NYCE-SIGE** México **PCBC** Poland **Quality Austria**
Austria **SII** Israel **SIQ** Slovenia **SIRIM QAS International** Malaysia **SQS** Switzerland **SRAC** Romania **TSE** Turkey **YUQS** Serbia

* The list of IQNET Members is valid at the time of issue of this certificate. Updated information is available under www.iqnet-certification.com

Certificate

CISQ/ICIM S.P.A. has issued an IQNet recognized certificate that the organization:

SYNTESYS S.R.L.

Head Office and Operative Unit

Via G. Galilei, 10/1-2-3 - Zona Industriale - I-35037 Selve di Teolo (PD)

Operative Units

Via G. Galilei, 16/1 - Zona Industriale - I-35037 Selve di Teolo (PD)

Via San Benedetto, 48/A - Zona Industriale - I-35037 Selve di Teolo (PD)

Via G. Galilei, 3 - Zona Industriale - I-35037 Selve di Teolo (PD)

has implemented and maintains a/an

Quality Management System

for the following scope:

Trading of products for laboratory analysis. Manufacturing of products for laboratory analysis and sanitary products. Design and production management of sterile swabs for the collection and the preservation of biological samples, also for surgical application, with or without transport medium.

which fulfils the requirements of the following standard:

ISO 13485:2016

Issued on: **2022-06-05**

First issued on: **2014-06-21**

Expires on: **2025-06-04**

This attestation is directly linked to the IQNet Partner's original certificate and shall not be used as a stand-alone document.

Registration Number: **IT-93779**



Alex Stoichitoiu
President of IQNET



Mario Romersi
President of CISQ



This attestation is directly linked to the IQNET Member's original certificate and shall not be used as a stand-alone document.

IQNET Members*:

AENOR Spain **AFNOR Certification** France **APCER** Portugal **CCC** Cyprus **CISQ** Italy **CQC** China **CQM** China **CQS** Czech Republic
Cro Cert Croatia **DQS Holding GmbH** Germany **EAGLE Certification Group** USA **FCAV** Brazil **FONDONORMA** Venezuela **ICONTEC**
Colombia **ICS** Bosnia and Herzegovina **Inspecta Sertifointi Oy** Finland **INTECO** Costa Rica **IRAM** Argentina **JQA** Japan **KFQ** Korea
LSQA Uruguay **MIRTEC** Greece **MSZT** Hungary **Nemko AS** Norway **NSAI** Ireland **NYCE-SIGE** México **PCBC** Poland **Quality Austria**
Austria **SII** Israel **SIQ** Slovenia **SIRIM QAS International** Malaysia **SQS** Switzerland **SRAC** Romania **TSE** Turkey **YUQS** Serbia

* The list of IQNET Members is valid at the time of issue of this certificate. Updated information is available under www.iqnet-certification.com

AUTHORIZATION LETTER

We, **Syntesys S.R.L.** having a registered office at Via G. Galilei 10/3, 35037 Selve di Teolo - PD - Italy, assign **Sanmedico SRL** having a registered office at A.Corobceanu str., apt. 9, Chişinău MD-2012, Moldova, as authorized representative.

We declare that the company mentioned above is authorized to register, notify, renew or modify the registration of medical devices on the territory of the Republic of Moldova.

This letter is valid till 28.08.2024

Teolo, 28.08.2023



SYNTESYS S.R.L.
UNIPERSONALE

Via G. Galilei, 10/3 - 35037 Z.I. Selve - Teolo (PD)
C.F./P.I./R.I. PD: 03573950288 - Cap. Soc. 20.700,00 €
Tel. 049 9903866 - Fax 049 9903867



Rinaldo Ruggero
CEO and Legal Representative
SYNTESYS S.R.L.