


Test Report

– Eppendorf PCR clean –

PCR Tube 0.5 mL



500 Tubes, thin walled
PCR clean

LOT K198395M
3534

 2026-07-28

Order no.: 0030 124.537
Eppendorf AG
22331 Hamburg · Germany
Email: eppendorf@eppendorf.com

Cat. no.: 0030124537
Distributed in North America by
Eppendorf North America, Inc. · USA
Email: info@eppendorf.com

• **Human DNA**

Method: *real-time* PCR acc. to SOP 99880/V2
Limit: < 2 pg; less than one human cell

• **DNase**

Method: DNA Digestion acc. to SOP 100570/V1
Limit: Not detectable (LOD: 1.0×10^{-6} Kunitz units)

• **PCR inhibition**

Method: *real-time* PCR acc. to SOP 100559/V1
Limit: Less than 10 targets amplifiable

• **RNase**

Method: RNA Digestion acc. to SOP 100571/V1
Limit: Not detectable (LOD: 1.0×10^{-9} Kunitz units)

Customer: Eppendorf AG, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 31.08.21

Testing of the above described lot showed conformity within the limits of detection.

01.09.21

Date of release

Epp 2108 9778

Registration number



Dipl.-Biol. Thomas Weigel (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.