



GE HealthCare

LOGIQ Fortis™

User Manual



Digitally signed by Chicu Natalia
Date: 2025.06.20 18:45:41 EEST
Reason: MoldSign Signature
Location: Moldova

MOLDOVA EUROPEANĂ



5919773-1EN Rev. 2

Version R4

GENERAL USER DOCUMENTATION

© 2023 By GE HealthCare



GE HealthCare

Regulatory Requirement

The LOGIQ Fortis complies with regulatory requirements of the following European Regulation 2017/745 EU concerning medical devices.



First CE Marked in 2021.

This User Manual is a reference for the LOGIQ Fortis. It applies to Version R4 software for the LOGIQ Fortis ultrasound system.

GE Healthcare
P.O. Box 414, Milwaukee, Wisconsin 53201 U.S.A.
(Asia, Pacific, Latin America, North America)

GE Healthcare GmbH
Beethovenstrasse 239
Postfach 11 05 60
D-42655 Solingen GERMANY
TEL: 49 212.28.02.208; FAX: 49 212.28.02.380

Revision History

Table -1 • Reason for Change

Revision	Date (YYYY/MM/DD)	Reason for change
Rev. 2	2023/10/10	Initial Release

Please verify that you are using the latest revision of this document. Information pertaining to this document is maintained on MyWorkshop. If you need to know the latest revision, contact your distributor, local GE Sales Representative or in the USA call the GE Ultrasound Clinical Answer Center at 1 800 682 5327 or 1 262 524 5698.

Regulatory Requirements

Conformance Standards


The following classifications are in accordance with the IEC/EN 60601-1:6.8.1:

- According to 2017/745 EU Medical Device Regulations, this is Class IIa Medical Device.
- According to IEC/EN 60601-1:
Equipment is Class I, Type B with BF or CF Applied Parts.
- According to CISPR 11:
Equipment is Group 1, Class A ISM Equipment.

This product complies with the regulatory requirement of the following:

- The LOGIQ Fortis console and related probes are an Active Diagnostic Medical Product belonging to Class IIa according to the MDR - REGULATION (EU) 2017/745 for use on human patients.
- Medical Device Regulation (EU) 2017/745: the CE label affixed to the product testifies compliance to the Regulation.

The location of the CE marking is shown in the Safety chapter of this manual.



Authorized EU Representative

European registered place of business:

GE Medical Systems SCS

283 rue de la Minière

78530 BUC, France



Authorized Swiss Representative

GE Medical Systems (Schweiz) AG

Europa-Strasse 31

8152 Glattbrugg

Switzerland

- International Electrotechnical Commission (IEC).
 - IEC/EN 60601-1 Medical Electrical Equipment, Part 1 General Requirements for Safety.
 - IEC/EN 60601-1-2 Electromagnetic disturbances - Requirements and tests.
 - IEC 60601-1-6 (Usability)

- IEC 60601-2-37 Medical electrical equipment. Particular requirements for the safety of ultrasonic medical diagnostic and monitoring equipment.
- International Organization of Standards (ISO)
 - ISO 10993-1 Biological evaluation of medical devices.
 - EN ISO 20417 (Information supplied with medical devices)
- Underwriters' Laboratories, Inc. (UL), an independent testing laboratory.
 - UL 60601-1 Medical Electrical Equipment, Part 1 General Requirements for Safety.
- ANSI/AAMI ES60601-1 Medical electrical equipment -Part 1: General requirements for basic safety and essential performance.
- Canadian Standards Association (CSA).
 - CSA 22.2, 60601-1 Medical Electrical Equipment, Part 1 General Requirements for Safety.
- Medical Device Good Manufacturing Practice Manual issued by the FDA (Food and Drug Administration, Department of Health, USA).
- EN/ISO 15223-1 Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements.

Certifications

GE Ultrasound Korea, Ltd. is ISO 13485 certified.

Original documentation

- The original document was written in English.

Importer Information

Turkey

GE Medical Systems Türkiye Ltd. Şti.
Esentepe Mah. Harman Sok. No: 8
34394 Şişli İstanbul Türkiye

Table of Contents

Chapter 1 - Introduction

System Overview	24
Attention	24
Documentation	24
Principles of Operation	27
Indications for Use	28
Contraindication	28
Prescription Device	29
Clinical Benefit (System and Probes)	29
Patient Population (System and Probes)	29
Contact Information	30
Contacting GE HealthCare Ultrasound	30
Manufacturer	35
Factory Sites	35

Chapter 2 - Safety

Owner Responsibility	38
Owner requirements	38
Notice against user modification	38
Reporting	38
Safety Precautions	39
Precaution Levels	39
Hazard Symbols	39
Patient Safety Related Hazards	41
Equipment and Personnel Safety	43
Classifications	47
EMC (Electromagnetic Compatibility)	48
Patient Environmental Devices	58
Acoustic Output	60
RoHS Hazardous Substances	62

Table of Contents

WEEE Passport	63
Safe Product and Packaging Disposal	64
Device Labels	65
Label Icon Description	65
Probe Labeling	71

Chapter 3 - Preparing the System for Use

Site Requirements	76
Before the system arrives	76
Environmental Requirements	76
Console Overview	78
Console Part Identification	78
Peripheral/Accessory Connection	81
Powering the System	88
Connect System to Power Supply	88
Circuit breaker	89
Power On	90
Power Up Sequence	90
Password Protection	91
Power Off	100
System Language Configuration	102
Select the System Language and Date/Time Format	102
Preset Restore User Interface Corruption	104
Probe Connections	107
Concept	107
Connecting the Probe	107
Cable Handling	109
Probe Management	109
Selecting Probes	110
Deactivating the Probe	110
Disconnecting the Probe	111
Storing the Probe	111
Storage/Transportation	111

Operator Controls	113
Control Panel Map	113
Control Panel Adjustment	114
Keyboard	115
Touch Panel	117
Touch Panel or Operator Panel Reset	119
Mode, Display and Print	121
Measurement and Annotation	121
Monitor	123
Moving the Wide Screen Monitor	123
Adjusting the Monitor	125
Wide Screen Monitor Display Settings	126
Test Patterns	130
Monitor Display	132
Live/Freeze Indicator	133
Monitor Display Layout	133
Monitor Display Image Management Controls	137
System Positioning and Transporting	139
Moving the System	139
Wheels	141
Transporting the System	143
Acclimation Time	144

Chapter 4 - Preparing for an Exam

Beginning an Exam	146
Patient Screen	146
OB Exam	150
Starting an Examination	153
Warnings and Cautions	153
Creating a New Patient Record	153
Perform an Exam	154
Ending an Exam	155
Scanning Without Entering Any Patient Data	156

Table of Contents

Starting a New Exam on an Existing Patient	156
Retrieve Patient/Exam Information from Worklist	156
Sending Multiple Exams and Patients with Single Click to PACS	157
Change Patient Information except Patient ID and scan	158
Change Patient ID of the Existing Patient (Edit & Copy)	158
Retrieving and Editing Archived Information	162
Searching for an Existing Patient	162
Changing Patient Information or an Exam	163
Deleting the Existing Patient/Exam/Image	164
MyPreset	166
Activating MyPreset	166
Arranging MyPreset Tab	167

Chapter 5 - Optimizing the Image

Optimizing B-Mode	170
Intended Use	170
B-Mode Scanning Hints	170
B-Mode Controls	171
Continuous Auto Tissue Optimization	174
Continuous Tissue Optimization (CTO)	175
SRI-HD	175
CrossXBeam	177
B-Flow	178
LOGIQView	181
BSteer+ (Option)	182
Touch TGC	186
Speed of Sound (SoS) Tissue Imaging	188
Minimizing Grating Lobe/Side Lobe Artifacts	189
Optimizing M-Mode	191
Intended Use	191
M-Mode Display	191
Typical Exam Protocol	191
M-Mode Controls	192

Anatomical M-Mode (AMM) and Anatomical Color M-Mode (ACMM)	192
Optimizing Color Flow	195
Intended Use	195
Typical Exam Protocol	195
Activating Color Flow	195
Color Flow Mode Controls	196
Radiantflow	198
Flow Model Shortcuts	198
Power Doppler Imaging (PDI)	200
Optimizing M Color Flow	204
Optimizing Spectral Doppler	205
Intended Use	205
Spectral Doppler Display	206
PW Doppler Mode Display	207
Doppler Mode Controls	208
Using 3D	212
3D Acquisition	212
3D Acquisition Parameter Description	213
Manipulating the Volume of Interest	215
Easy 3D	217
Advanced 3D	218
Movie 3D	221
Tru3D and Volume Measurement	222

Chapter 6 - Scanning/Display Functions

Freezing an Image	230
CINE	231
Cine Gauge	231
Touch Panel	231
Using CINE	232
Mark CINE	234
Preview	235
Background Store	235

Table of Contents

Image Storage Hints	236
CINE Capture	236
Image Zoom	239
Location of Zoom Control	239
HD Zoom	239
Pan Zoom	240
Magnification Zoom	241
Split Screen	242
Dual Screen	242
Quad Screen	242
Simultaneous Mode	243
Dual Caliper	243
Annotating an Image	247
Adding Comments to an Image	248
Body Patterns	254
Using the Fast Key	262
Create a Fast Key	262
Start a Fast Key	263
Backup and Restore the Fast Key	263

Chapter 7 - General Measurements and Calculations

General Measurements and Calculations Introduction	266
Basic Operation	266
Measurement Controls	268
Touch Panel	269
Exam Category/Study/Measurement	272
General Instructions	272
Measurement Cursor	274
Copy, Move and Paste Measurement Tools	275
Performing Measurements on Saved Images	276
Generic Measurements	278
B-Mode Measurements	278
M-Mode Measurements	285

Doppler Mode Measurements	287
Helpful Hints	291
Mode Measurements	292
B-Mode Measurements	292
Doppler Mode Measurements	296
M-Mode Measurements	304
Worksheet	306
To View a Worksheet	306
To Edit a Worksheet	307
Recording Worksheet	310
OB Worksheet	310
Vascular Worksheet	311
Anatomical Survey	315
Editing	315
Measurement and Calculation Setup	318
Starting Study and Measurement Setup	318
Display/Hide a Folder or a Measurement in the Touch Panel	321
Setting up an Automatic Measurement Flow	325
Change the Tool Used to Make a Measurement	326
Adding Folders and Measurements	326
M & A Advanced Preset	335
Doppler Tab - Modify Calculation	338
Application Measurement Preset	339

Chapter 8 - Application M&A

General Information	342
General Guidelines	342
Abdomen	343
Small Parts	344
B-Mode Measurements	344
OB	346
OB Type change	346
To Start an Obstetrics Exam	347

Table of Contents

OB Measurements Performed Over Multiple Planes	350
OB Calculation	352
OB Graphs	353
OB-Multigestational	360
OB Table Editor	362
GYN	371
B-Mode Measurements	371
Cardiac	373
Naming Format for Cardiac Measurements	373
Cardiac Doppler Measurements	376
Vascular	377
IMT Measurement	377
Auto Vascular Calculation	386
Manual Vascular Calculation	388
Intravessel Ratio	389
Bypass Graft Anastomosis Graph	390
Urology	393
Bladder Volume	393
Renal Volume	393
Prostate Volume	394
Pelvic Floor Measurements	395
Pediatrics	397
Pediatrics Hip	397

Chapter 9 - Recording Images

Set Up to Record Images	402
Image Management Guide	402
Adding Devices	403
Adding a Dataflow	403
Adding Devices to a Print Button	404
Formatting Removable Media	404
Storing Images and Cineloops	405
Storing an Image	405

Storing a CINE Loop	405
Review Images in Archive	409
Review the Patient Exam/Image	409
Active Images	409
Image Tags	411
Image Reorder	415
Image History	416
Clipboard	418
Clipboard Icon	418
Saving the Image/CINE to the Clipboard	419
Previewing Clipboard Images	419
Recalling Images from the Clipboard	419
To Delete an Image from the Clipboard	419
SaveAs	420
Using SaveAs	420
'SaveAs' Images	423
Notes on Transfer to CD/DVD	423
Storing Images with More Resolution	424
EZBackup	425
Perform EZBackup	427
To Review EZBacked Up Images	428
Data Transfer	430
Export/Import	430
Worklist (Search and Retrieve the Patient/Exam Information)	431
Unified Background Export	432
DICOM Viewer	436
External Drives	438
Intended Use	438
DVD Drive	438
USB Hard Disk Drive and USB Flash Drive	439
USB Quick Save	442
Printing Options	444
Setting up Digital Peripherals	444
External Paper Printer	449

Table of Contents

Portable Exam	452
Perform a Portable Exam (Using the Worklist)	452

Chapter 10 - Customizing Your System

Presets	454
System Presets	456
Changing system parameters	456
System/General Preset Menu	456
System/System Display Preset Menu	459
System/System Imaging Preset Menu	461
System/System Measure Preset Menu	464
System Backup and Restore Preset Menu	467
System/Peripherals Preset Menu	468
System/User Configurable Key	470
System/About Preset Menu	472
Licenses	472
Imaging Presets	473
Changing imaging presets	473
General	474
Comments Libraries Presets	476
Comments Libraries/Libraries Preset Menu	476
Comments Libraries/Comments Preset Menu	478
Comments Libraries/Applications Preset Menu	479
Comments Libraries/Mapping Preset Menu	480
Body Patterns Presets	483
Body Pattern Libraries/Libraries Preset Menu	483
Body Pattern Libraries/Body Patterns Preset Menu	484
Body Pattern Libraries/Applications Preset Menu	485
Application Presets	487
Settings	487
Print Controls	489
Imaging Controls	490
Comments and Body Patterns	491

Measurements	491
User Specific	492
Test Patterns	493
3D/4D	494
4D Presets	494
Configuring Connectivity	497
Connection Manager	497
Scanner Page	497
Network Page	499
MyComputer Device Page	505
Dicom Page	512
Special Devices Page	522
Print Button Page	542
Dataflow Page	545
Advanced Settings Page	546
Connection Manager Summary Page	559
Measure	561
Reports	562
System Administration	563
Administrator Tasks	563
Privacy and Security	563
System Admin	568
Users	571
Logon	572
Groups	578
LDAP Configuration	581
System Password	586
Disk Encryption	586
Audit Report and System Log Server Configuration	591
Imaging Preset Manager	594
Creating a User-Defined Application Preset	594
Sharing User Presets between LOGIQ Fortis Systems	596
Retain Field of View	597
Backup and Restore	599

Table of Contents

Backup	599
Restore from Media or Local Backup	601
Preset Synchronization Using Media	603
Search	604

Chapter 11 - Probes and Biopsy

Probe Overview	606
Ergonomics	606
Supported Probes	606
Probe orientation	616
Probe Naming Conventions	617
Probe Safety	618
Care and Handling	618
Handling precautions	619
Electrical Shock Hazard	620
Special handling instructions	620
Probe handling and infection control	622
Care and Maintenance	623
Inspecting probes	623
Environmental Requirements	623
Probe Processing	624
Probe Cleaning and Disinfecting Notes	644
Probe Disinfectants	646
Coupling Gels	646
V Nav Cleaning Requirements	648
Planned Maintenance	649
Automatic Probe Diagnostics	650
Probe Check (not available in all countries)	650
Automatic Probe Check	652
Returning/Shipping Probes and Repair Parts	652
Biopsy Special Concerns	653
Precautions Concerning the Use of Biopsy Procedures	653
Biopsy Guide Sterilization	654

Performing a Biopsy	655
Displaying the Guidezone	655
Preparing the Biopsy Guide Attachment	658
4D Probe Biopsy Needle Path Selection	672
Biopsy Needle Path Verification	672
The Biopsy Procedure	673
Post Biopsy	673
Probe Biopsy Reprocessing	673
Surgery/Intra-operative Use	676
Preparing for Surgery/Intra-operative Procedures	676

Chapter 12 - User Maintenance

System Data	678
Features/Specifications	678
Clinical Measurement Accuracy	684
Privacy and Security	688
Privacy & Security Environment	688
Identity Provisioning	688
Network Infrastructure	690
Audit Logs	690
Anti-Virus Software Note	691
Security	691
Loading Windows Patches	693
Loading a Windows Patch	693
Performing Software Patch Verification	693
Patch Installation Notes	695
Reloading Software	695
Software Download	696
Overview	696
Software Download Available	696
Base Image and Software Load	700
Base Image and Software Load Procedure	700
Copyrighted Material	701

Table of Contents

Viewing Copyrighted Third Party Software License Information	701
System Care and Maintenance	702
Overview	702
Maintenance Schedule	702
Cleaning and Disinfecting the System	703
Other Maintenance	710
Quality Assurance	715
Introduction	715
Typical Tests to Perform	715
Baselines	717
Periodic Checks	717
Results	717
System Setup	718
Test Procedures	718
Setting up a Record Keeping System	723
Ultrasound Quality Assurance Checklist	724
Image Quality Check	726
Image Quality Check (IQC)	726
Assistance	727
Supplies/Accessories	727

Chapter 13 - Advanced Features

Using 4D	737
Features supported with 4D	737
4D Principles of Operation	737
4D Operational Controls	743
4D Presets	745
Performing a 4D Scan	746
Tomographic Ultrasound Imaging (TUI)	762
SonoRenderlive	764
SonoAVC™	766
OmniView	770
STIC (Spatio-Temporal Image Correlation)	772

Contrast Imaging	776
Overview	776
Contrast Imaging Overview	776
Mode	778
Contrast MVI Mode	778
Contrast Presets	778
Sonazoid™ Contrast Agent	779
Contrast Controls	779
Static 3D with Contrast	782
CEUS LI-RADS Feature	783
Time Intensity Curve (TIC) Analysis	784
Parametric Imaging	805
Strain Elastography	816
Description	816
Using Strain Elastography	816
Clinical Applications	818
Strain Elastography Controls	818
Application Parameters	819
General Imaging Parameters by Application/Probe/Feature	820
Elastography Analysis	821
Overview	821
Using Elastography Analysis	821
Additional Notes for Elastography Analysis	823
Shear Wave Elastography	824
Overview	824
Intended Uses	824
Configuring Shear Wave	825
Activating Shear Wave	827
Shear Wave Display	827
Shear Wave Touch Panel	832
Using Shear Wave (SW)	835
Typical Exam Protocol (Liver)	836
Shear Wave Measurements	837
Scanning Hints	842

Table of Contents

Ultrasound-Guided Attenuation Parameter (UGAP) Option	843
Overview	843
Activation	843
Measurement and Worksheet	844
2D Color Map	846
Continuous Wave Doppler (CWD)	848
Overview	848
Steerable	848
Non-Imaging	848
Activating CW Doppler	848
Exiting CW Doppler	848
Tissue Velocity Imaging (TVI)	849
Intended Use	849
Activating TVI	849
Optimizing TVI	849
TVI and TVD	850
Quantitative Analysis (QAnalysis)	851
Overview	851
Quantitative Flow Analysis	851
Selecting QAnalysis Image Range	855
Activating QAnalysis	855
Common QAnalysis Function	856
Stress Echo	866
Introduction	866
Getting started with a stress study	866
Image acquisition	868
Continuous Capture mode	870
Analysis	877
Editing/Creating template	879
Wall Motion Segment Setup	884
Utility Application Settings for Protocol	884
Report	885
Cardiac Automated Functional Imaging (Cardiac Strain)	887
Cardiac Strain	887

Data Requirements	887
Recommended Probes (RAW Data only)	887
Image characteristics	888
Acquisition	888
Sequential Acquisition	888
Heart rate and cycles	888
Configuring the tool	888
Configurable Workflow Controls	889
Configurable analysis settings	890
Reference Layer for Strain	890
Starting the Tool	890
Navigating the Tool	891
Cardiac Strain on the APLAX View	892
Cardiac Strain on A4-Ch and A2-Ch views	893
AutoEF Layout	894
Tracking	894
ROI Adjustment	895
To Create a New ROI	895
Event Timing	898
Inspecting Results	899
Peak Detection	900
Exiting AFI	900
Auto EF	902
Auto EF Measurements	902
Acquisition	902
Starting AutoEF	902
Tracking Validation	904
Possible Causes of Poor Tracking	904
To Create a New ROI Trace	905
Exiting AutoEF	906
ECG	907
Overview	907
ECG Cable/Respiration	908
Physiological Trace Monitor Display	909

Table of Contents

ECG Touch Panel	910
Volume Navigation	913
Introduction	913
Load the Volume Dataset	916
Query/Retrieve	916
Setting Up V Nav sensors	924
Activate V-Nav	926
Using V-Nav	930
Perform the registration	932
Advanced GPS Markers	940
V Nav Trackers	947
Vscan Air™ CL (Option)	956
Prepare the Vscan Air CL Probe for Use	957
Vscan Air On Console Charger	958
Pair Vscan Air CL to the Ultrasound System	961
Vscan Air CL Battery Indicator	962
Wireless Connection Quality Indicator	963
Vscan Air CL Temperature Indicator and Thermal Management	963
Scanning	964
Preparing for a Guided Procedure with Vscan Air CL	965
Disconnect the Vscan Air CL Probe	967
Supported Features	967
Charger Cleaning and Disinfection	968
Regulatory Requirements	969
Vscan Air On Console Charger Rating Plate	970
Acoustic Output Reporting Tables for Track 3/EN/IEC 60601-2-37	970
Breast Productivity Package	971
Overview	971
Breast Lesion M&A	971
Breast Measure Assistant (Auto Contour)	973
Breast Assistant, Powered by Koios DS (not available in all countries)	974
Thyroid Productivity Package	976
Overview	976
Worksheet and Summary Worksheets	979

Thyroid Assistant, Powered by Koios DS (not available in all countries)	980
Start Assistant	982
Introduction	982
Start Assistant Mapping Editor	983
Scan Assistant	986
Introduction	986
Scan Assistant Definitions	986
Scan Assistant Description	987
Setting up Scan Assistant	987
Using Scan Assistant	988
Exporting Scan Assistant Programs to Another LOGIQ Fortis	992
Scan Assistant Creator	992
Compare Assistant	1012
Overview	1012
Workflow example	1013
Setting Up Compare Assistant	1014
OB Measure Assistant	1020
Overview	1020
Auto Renal Measure Assistant	1021
Availability	1021
Auto Renal Configurations	1021
Activate Measure Assistant	1022
Perform Measurements	1023
Summary of Results	1023
Auto Preset Assistant	1025
Auto Preset Assistant Configuration in Imaging Preset Manager	1025
Auto Preset Assistant Configuration in System Imaging	1026
Auto Preset Assistant Configuration in User Defined Trackball Set Key Menu	1027
Using Auto Preset Assistant	1028
Auto Abdominal Color Assistant	1030
Auto Abdominal Color Assistant Configuration for System Imaging	1030
Auto Abdominal Color Assistant Configuration in User Defined Trackball Set Key Menu	1031
Auto Flow Model Selection Configuration in Imaging Menu	1032

Table of Contents

Using Auto Abdominal Color Assistant	1033
Hepatic Assistant	1034
Overview	1034
Hepatic Assistant Availability	1034
Activate Hepatic Assistant	1034
Visualization	1036
EZ Imaging	1040
Overview	1040
Activate EZ Touch Panel	1040
MyPreset Shortcuts “By Probe”	1043
MyPreset Shortcuts “By Category”	1044
Quick Patient Change	1044
Anonymize the patient	1046
Overview	1046
Anonymize the patient	1046
Report Writer	1048
Introduction	1048
Creating a report	1048
Activating the Report	1049
Editing a Report	1051
Accessing Worksheet, OB Graph and Anatomical Survey Pages	1059
Storing the Report	1059
Retrieving an Archived Report	1060
Deleting a Report from Archive	1060
Printing the Report	1060
Exporting the Report to Media	1061
Exiting the report	1061
Designing Your Own Template	1061
Direct Report	1079
Report Presets	1082
Configuring the Wireless Network	1088
Wireless LAN (WLAN)	1088
Wireless LAN (WLAN) Specifications	1088
Connecting to the WLAN	1089

Network and Spooler Status Icons	1090
Adding a Wireless Network	1090
Removing a WLAN	1091
Customizing Wireless Network Settings	1092
Setting an IP Address	1093
Refreshing a WLAN	1093
Managing Connectivity to a Wireless Network	1093
Monitoring the WLAN	1094
WLAN Diagnostics	1094
Repairing the WLAN	1094
Available WLAN Channels	1095
Disconnecting from the WLAN	1095
Device Mgmt	1096
Overview	1096
Upload For Fleet and Manual Backup (Cloud Backup)	1097
Backup Automatically (Local and Cloud Backup)	1098
Manual Backup (Cloud Backup)	1099
Assign Configuration to Fleet or to the Device (on Cloud)	1099
Receive Installation Notification	1099
Installation Dialog	1100
Detailed Restore from Cloud	1106
Cancel a Failed Job	1107
Icon and Notification	1108
Smart Device Apps	1110
LOGIQ Apps	1110
Tricefy Uplink	1117
Introduction	1117
Uploading Exam Information to the Tricefy Cloud	1118
Tricefy Icons	1119
Tricefy Activation	1119
Digital Expert	1120
Digital Expert Remote Training	1120
Data Streaming (Option)	1121
User Setup for Data Streaming	1121

Table of Contents

Enable Data Streaming	1122
Starting Data Streaming	1122
Data Streaming in Process	1123
Running Scanner Apps	1125
App Launchpad	1125
Scanner Apps	1125
Starting a Scanner App from the App Launchpad	1125
Service and Applications Support	1127
GE HealthCare Connect Guide	1127
Support Requests	1130
Service Desktop	1133
Overview	1133
Accessing the Service Desktop	1133
Battery Power Mode	1134
Overview	1134
Power Assistant Battery Option	1134
Scan on Battery Option (ScoB)	1135
View current battery status	1138
Battery charging	1140
Refreshing the battery	1141
Battery Deterioration	1141
Battery Disposal	1142
Magstripe Card Reader	1143
Magstripe Magnetic Card Reader	1143
Footswitch	1144
Wired Footswitch	1144
DVR	1145
Setting up the DVR	1145
Using the DVR	1145

Chapter 1

Introduction

In this section

System Overview.	24
Contact Information.	30

System Overview

Attention

This manual contains necessary and sufficient information to operate the system safely. Advanced equipment training may be provided by a factory trained Applications Specialist for the agreed-upon time period.

Read and understand all instructions in this manual before attempting to use the ultrasound system.

Keep this manual with the equipment at all times. Periodically review the procedures for operation and safety precautions.

Disregarding information on safety is considered abnormal use.

Not all features, products, probes, or peripherals described in this document may be available or cleared for sale in all markets. Please contact your local GE HealthCare Ultrasound representative to get the latest information.

NOTE

Please note that orders are based on the individually agreed upon specifications and may not contain all features listed in this manual.

NOTE

All references to standards/regulations and their revisions are valid at the time of publication of the user manual.

Documentation

Safety instructions must be reviewed before operating the system.

The ultrasound system manuals are written for users who are familiar with basic ultrasound principles and techniques. They do not include sonographic training or detailed clinical procedures.

Documentation is provided in the following ways:

- Online Help PDFs viewable on the system
- Offline PDFs viewable on a Windows PC
- PDFs on the internet at the GE Customer Documentation Portal
- Paper copies (orderable via H-Cat)

Table 1-1 • Documentation

Publication	Translated	Available via F1	Available via Media in the eIFU Kit	Available on Paper (if purchased via H-Cat)
User Manual Provides information needed by the user to operate the system safely.	Yes (all required languages)	Yes	Yes	Yes
Release Notes Provides precautions and instructions that supplement the User Manual	Yes	No	Yes	Yes
Advanced Reference Manual Provides Acoustic Output Data and System Measurement and Analysis Tables and Formulas.	English and French	No	Yes	Yes
Privacy and Security Manual Describes Privacy and Security considerations and capabilities, and how they may be configured and used appropriately on the system.	Yes (all required languages)	No	No	No
Basic Service Manual Supplies block diagrams, spare parts, adjustments, instructions to help qualified technical personnel repair the system.	No	No	No	Yes

NOTE

The Basic Service Manual referenced in this manual is part number 5874373.

NOTE

An AIUM Booklet is shipped with systems shipped in the United States and Canada.

NOTE

Dates on screenshots are represented in MM/DD/YYYY format throughout the manual. Information on how to change the system's date can be found in Customizing Your System.

NOTE

The screen graphics in this manual are only for illustrational purposes. Actual screen output may differ.

Online Help PDFs

Online Help PDFs are available on the system by pressing the F1 key or via the eIFU icon on the second Utility Page.

Viewing Online Help in a Language Different from the System Language

On the Utility > System > General Preset page, select the language in which to view Online Help from the Online Help Language dropdown selections.

If the translated Online Help is not available, the default language (English) is viewable.

Translated Online Help files can be updated via the eIFU USB Flash Drive provided with the ultrasound system in the eIFU Kit. To order another kit, contact your GE HealthCare Representative.

Offline PDFs

To view user documentation PDFs on a Windows PC:

1. Insert the media into the media drive.
2. Open the media drive on your desktop.
3. Double click the HTML document for the ultrasound system.
4. Select the item you want to view (click on the blue, underlined link in the File Name column).

To close the window, click on the 'X' in the upper, right-hand corner of the browser window.

NOTE

If your PC does not have Adobe Reader, a free download is available on the Adobe website at <http://www.adobe.com>.

Updating Documentation on the Ultrasound Scanner Via the USB

The latest version of the Online Help is located on the USB flash drive. To update to the latest version:

1. Power down the ultrasound system and insert the eIFU USB flash drive into a rear USB port.

NOTE

Ensure that the system is USB Device Enabled (check setting on System Admin Utility Page).

2. Power on the ultrasound system and follow the screen prompts.
 1. Select Install SW ... on the Start Application screen.
 2. Select OK on the first StartLoader screen.
 3. Select the package and then select Install on the second StartLoader screen; software installation begins. After "Installation has completed" message appears and system restarts, remove USB flash drive.
 4. As you verify each feature works correctly, select "Passed." If all features work correctly and "Passed" is selected for all features, the signature field is enabled at the bottom of the New Software Verification Checklist. Type your signature (minimum of three characters) and press OK. The system is now ready for use.

NOTE

You can search through a document, use hyperlinks in the Table of Contents and Index to locate topics, and navigate via bookmarks.

NOTE

In addition to viewing documentation on the Ultrasound system, the Documentation media can be read on any PC.

To exit, press the 'X' in the upper, right-hand corner of the documentation window.

Paper Documentation

Paper manuals can be ordered via H-Cat.

Online Customer Documentation Portal

Documentation is available in the Customer Documentation Portal on the internet at the GE HealthCare Support Documentation Library at:

<https://www.gehealthcare.com/documentation>

Navigate to the Customer Documentation Portal on the website and enter the following information to search for the desired manual:

- Document Number

OR

- Modality
- Product(s)
- Document Type
- Keyword

The desired manual can be downloaded from the website or the manual url can be copied to share the link.

Principles of Operation

Medical ultrasound images are created by computer and digital memory from the transmission and reception of mechanical high-frequency waves applied through a transducer. The mechanical ultrasound waves spread through the body, producing an echo where density changes occur. For example, in the case of human tissue, an echo is created where a signal passes from an adipose tissue (fat) region to a muscular tissue region. The echoes return to the transducer where they are converted back into electrical signals.

These echo signals are highly amplified and processed by several analog and digital circuits having filters with many frequency and time response options, transforming the high-frequency electrical signals into a series of digital image signals which are stored in memory. Once in memory, the image can be displayed in real-time on the image monitor. All signal transmission, reception and processing characteristics are controlled by the main computer. By selection from the system control panel, the user can alter the characteristics and features of the system, allowing a wide range of uses, from obstetrics to peripheral vascular examinations.

Introduction

Transducers are accurate, solid-state devices, providing multiple image formats. The digital design and use of solid-state components provides highly stable and consistent imaging performance with minimal required maintenance. Sophisticated design with computer control offers a system with extensive features and functions which is user-friendly and easy to use.

Indications for Use

The LOGIQ Fortis is intended for use by a qualified physician for ultrasound evaluation.

Indications for Use Statement

LOGIQ Fortis is a general purpose diagnostic ultrasound system intended for use by qualified and trained healthcare professionals for ultrasound imaging, measurement, display and analysis of the human body and fluid.

The ultrasound system is intended to be used in a hospital or medical clinic.

Ultrasound system clinical applications include : Fetal / Obstetrics; Abdominal (including Renal, Gynecology/Pelvic); Pediatric; Small Organ (Breast, Testes, Thyroid); Neonatal Cephalic; Adult Cephalic; Cardiac (Adult and Pediatric); Peripheral Vascular; Musculoskeletal Conventional and Superficial; Urology (including Prostate); Transrectal; Transvaginal; Transesophageal and Intraoperative (Abdominal and Vascular).

Modes of operation include: B, M, PW Doppler, CW Doppler, Color Doppler, Color M Doppler, Power Doppler, Harmonic Imaging, Coded Pulse, 3D/4D Imaging mode, Elastography, Shear Wave Elastography, Attenuation Imaging and Combined modes: B/M, B/Color, B/PWD, B/Color/PWD, B/Power/PWD.

Operator Profile

- Qualified and trained physicians or sonographers with at least basic ultrasound knowledge.
- The operator must have read and understood the user manual.

NOTE

Only qualified physicians or sonographers should perform ultrasound scanning on human subjects for medical diagnostic reasons. Request training, if needed.

Types of Use

Multiple patient, multiple use.

Frequency of Use

Daily (Typically 8 hours)

Contraindication

The ultrasound system is not intended for ophthalmic use or any use causing the acoustic beam to pass through the eye.

Prescription Device



CAUTION: United States law restricts this device to sale or use by, or on the order of a physician.

Clinical Benefit (System and Probes)

The clinical benefit of a diagnostic ultrasound device is to help healthcare professionals provide accurate diagnostic information (visualize human tissue/internal structure) that enhances the diagnostic and treatment care pathways of the patient for a variety of diseases and conditions.

Patient Population (System and Probes)

- Age: all ages (including embryos and fetuses)
- Location: worldwide
- Sex: male and female
- Weight: all weight categories
- Height: no limitations

NOTE

Extreme obesity may affect the operation of the device.



WARNING

Do not cross-use the ultrasound system between human and veterinary/animal use.

Contact Information

Contacting GE HealthCare Ultrasound

For additional information or assistance, please contact your local distributor or the appropriate support resource listed on the following pages:

Internet

<http://www.gehealthcare.com>

<https://cleaning.gehealthcare.com>

Clinical Questions

For information in the United States, Canada, Mexico and parts of the Caribbean, call the Customer Answer Center. TEL: (1) 800-682-5327 or (1) 262-524-5698

In other locations, contact your local Applications, Sales, or Service Representative.

Service Questions

For service in the United States, call GE HealthCare CARES.

TEL: (1) 800-437-1171

In other locations, contact your local Service Representative.

Information Requests

To request technical product information in the United States, call GE HealthCare.

TEL: (1) 800-643-6439

In other locations, contact your local Applications, Sales, or Service Representative.

Placing an Order

To order accessories, supplies, or service parts in the United States, call the GE HealthCare Technologies Contact Center.

TEL: (1) 800-558-5102

In other locations, contact your local Applications, Sales, or Service Representative.

Table 1-2 • Americas

ARGENTINA	GE HealthCare Argentina Nicolas de Vedia 3616 piso 5 Buenos Aires - 1307	TEL: (+54) 11-5298-2200
BRAZIL	GE HealthCare do Brasil Comércio e Serviços para Equipamentos Médicos - Hospitalares Ltda. Av. Magalhães de Castro, 4800, Andar 10 Conj. 101 e 102, Torre 3 Cidade Jardim CEP: 05676-120 - São Paulo/SP - Brasil	TEL: 3004 2525 (Capitals and Metropolitan Regions) 08000 165 799 (Other Locations)

CANADA	GE HealthCare Ultrasound 9900 Innovation Drive Wauwatosa, WI 53226	TEL:(1) 800-668-0732 Customer Answer Center TEL: (1) 262-524-5698
LATIN & SOUTH AMERICA	GE HealthCare Ultrasound 9900 Innovation Drive Wauwatosa, WI 53226	TEL:(1) 800-668-0732 Customer Answer Center TEL: (1) 262-524-5698
MEXICO	GE HealthCare Sistemas Medicos de Mexico S.A. de C.V Rio Lerma #302, 1° y 2° Pisos Colonia Cauhtemoc 06500-Mexico,D.F.	TEL:(1) 800-668-0732 Customer Answer Center TEL: (1) 262-524-5698
USA	GE HealthCare Ultrasound 9900 Innovation Drive Wauwatosa, WI 53226	TEL:(1) 800-668-0732 Customer Answer Center TEL: (1) 262-524-5698

Table 1-3 • Asia

ASIAPACIFIC JAPAN	GE HealthCare Asia Pacific 4-7-127, Asahigaoka , Hino- shi, Tokyo 191-8503, Japan	TEL:+81 42 585 5111
AUSTRALIA	32 Phillip Street Parramatta 2150 Sydney, NSW, Australia	TEL:1800 659 465
CHINA	GE HealthCare - Asia No.1, Yongchang North Road Beijing Economic & Technology Development Area Beijing 100176, China	TEL:(8610) 5806 8888 FAX:(8610) 6787 1162 Service: 4008128188 (24h)
INDIA	Wipro GE HealthCare Pvt Ltd No. 4, Kadugodi Industrial Area Sadaramangala, Whitefield Bangalore, 560067	TEL:+(91) 1-800-425-8025
KOREA	15F, 416 Hangang Dae ro, Chung-gu Seoul 04637, Korea	TEL:+82 2 6201 3114
NEW ZEALAND	Level7 Vero Centre 48 Shortland St, Auckland, 1010 New Zealand	TEL:0800 659 465
SINGAPORE	GE HealthCare ASEAN (Singapore) 11 North Buona Vista Drive #11-07 The Metropolis Tower 2 Singapore 138589	TEL:+65 6291 8528 FAX:+65 6291 7006

Table 1-4 • Africa

EGYPT	GE Medical Systems Egypt, LLC Plot 44 Tesseen El Shamaly Street Al Salam Axis First Sector City Centre, 5th settlement Cairo, Egypt	TEL: +20 2 25354200 FAX: +20 2 25370031
KENYA	GE HealthCare East Africa Services Limited General Mathenge Drive, Courtyard Building Westlands Nairobi 30 00100 KE	TEL: +254 719 093 044
NIGERIA	GE HealthCare International Operations (Nig) Ltd Bishop Aboyade Cole Street No. 927/928 Mansard Place, PO Box 54255 Victoria Island Lagos LA NG	TEL: +234 (01) 4607101 TEL: +234 (01) 4607102
KINGDOM OF SAUDI ARABIA	GE HealthCare Arabia Co. Ltd Platinum Centre, Building 1 Salahuddin Ayoubi Road Riyadh-12811 Kingdom of Saudi Arabia	TEL: +966 (11) 494 5779 FAX: +966 (11) 207 3946
SOUTH AFRICA	GE HealthCare South Africa (Pty) Ltd. 60 Glenhove Road Green on Glenhove Customer Innovation Centre Johannesburg GP 2196 ZA	TEL: +270100725000 FAX: +27 0862958385

Table 1-5 • Europe

AUSTRIA	GE HealthCare Austria GmbH & Co OG EURO PLAZA, Gebäude E Technologiestrasse 10 A-1120 Vienna	TEL: (+43) 1 97272 0 FAX: (+43) 1 97272 2222
BELGIUM & LUXEMBURG	GE HealthCare BVBA/SPRL Kouterveldstraat 20 1831 DIEGEM	TEL: (+32) 2 719 7204 FAX: (+32) 2 719 7205

CZECH REPUBLIC	GE Medical Systems Česká Republika, s.r.o. Bucharova 2641/14 158 00 Praha 5 Česká republika	TEL: (+420) 224 446 162 FAX: (+420) 224 446 161
DENMARK	GE HealthCare Park Allè 295 DK-2605 Brøndby, Denmark	TEL: (+45) 43 295 400 FAX: (+45) 43 295 399
ESTONIA & FINLAND	GE HealthCare Finland Oy Kuortaneenkatu 2, 000510 Helsinki P.O. Box 330, 00031 GE Finland	TEL: (+358) 10 39 48 220 FAX: (+358) 10 39 48 221
FRANCE	GE Medical Systems SCS Division Ultrasound 24 Avenue de l'Europe - CS20529 78457 Vélizy Villacoublay Cedex	TEL: (+33) 1 34 49 52 70 FAX: (+33) 13 44 95 202
GERMANY	GE HealthCare GmbH Beethovenstrasse 239 42655 Solingen	TEL: (+49) 212-28 02-0 FAX: (+49) 212-28 02-380
GREECE	GE HealthCare 8-10 Sorou Str. Marousi Athens 15125 Hellas	TEL: (+30) 210 8930600 FAX: (+30) 210 9625931
HUNGARY	GE HealthCare Hungary Zft. Bence utca 3 Budapest BU 1138 HU	TEL: (+36) 23 410 314 FAX: (+36) 23 410 390
IRELAND	NORTHERN IRELAND GE HealthCare Victoria Business Park 9, Westbank Road Belfast BT3 9JL. REPUBLIC OF IRELAND GE HealthCare 3050 Lake Drive Citywest Business Campus Dublin 24	TEL: (+44) 028 90229900 TEL: 1800 460 550 FAX: (+353) 1 686 5327
ITALY	GE Medical Systems Italia spa Via Galeno, 36, 20126 Milano	TEL: (+39) 02 2600 1111 FAX: (+39) 02 2600 1417
KAZAKHSTAN	«Дженерал Электрик Қазақстан» ЖШС Қазақстан, Алматы қаласы, Медеу ауданы, көшесі ЗЕНКОВ, үй 26/41, пошталық индексі 050010	T+7 727 3560020

LUXEMBORG	See Belgium	
NETHERLANDS	GE HealthCare De Wel 18 B, 3871 MV Hoevelaken PO Box 22, 3870 CA Hoevelaken	TEL: (+31) 33 254 1290 FAX: (+31) 33 254 1292
NORWAY	GE HealthCare Vingmed Ultrasound AS Sandakerveien 100C 0484 Oslo, Norway GE HealthCare Vingmed Ultrasound Strandpromenaden 45 P.O. Box 141, 3191 Horten	TEL: (+47) 23 18 50 50 FAX: (+47) 23 18 60 35 TEL: (+47) 33 02 11 16
POLAND	GE Medical Systems Polska Sp. z o.o., ul. Woloska 9 02-583 Warszawa, Poland	TEL: (+48) 22 330 83 00 FAX: (+48) 22 330 83 83
PORTUGAL	GE HealthCare Portuguesa SA Avenida do Forte 6 - 6A Edificio Ramazzotti 2790-072 Carnaxide	TEL: (+351) 21 425 1300 FAX: (+351) 21 425 1343
RUSSIA	GE HealthCare Presnenskaya nab. 10 Block C, 12 floor 123317 Moscow, Russia	TEL: (+7) 4957 396931 FAX: (+7) 4957 396932
SPAIN	GE HealthCare España C/ Gobelás 35-37 28023 Madrid	TEL: (+34) 91 663 2500 FAX: (+34) 91 663 2501
SWEDEN	GE HealthCare Sverige AB FE 314, 182 82 Stockholm Besöksadr: Venevagen 89 Danderyd, Sverige	TEL: (+46) 08 559 500 10 FAX: (+46) 08 559 500 15 Service Center (+46) 020-120 14 36
SWITZERLAND	GE Medical Systems (Schweiz) AG Europastrasse 31 8152 Glattbrugg	TEL: (+41) 1 809 92 92 FAX: (+41) 1 809 92 22
TURKEY	GE HealthCare Türkiye Istanbul Office Levent Ofis Esentepe Mah. Harman Sok. No:8 Sisli-Istanbul	TEL: +90 212 398 07 00 FAX: +90 212 284 67 00
UNITED ARAB EMIRATES (UAE)	GE HealthCare Dubai Internet City, Building No. 18 First Floor, Dubai - UAE	TEL: (+971) 4 429 6101 or 4 429 6161 FAX: (+971) 4 429 6201

UNITEDKINGDOM	GE Medical Systems Ltd Pollards Wood Nightingales Lane Chalfont St Giles Buckinghamshire HP8 4SP	TEL: (+44) 1494 544000 FAX: (+44) 1707 289742
For all other European countries not listed, please contact your local GE HealthCare distributor or the appropriate support resource listed on www.gehealthcare.com .		

Manufacturer

	GE Ultrasound Korea, Ltd. 9, Sunhwan-ro 214beon-gil, Jungwon-gu, Seongnam-si, Gyeonggi-do, 13204 Republic of Korea

Factory Sites

GE Ultrasound Korea, Ltd.

9, Sunhwan-ro 214beon-gil, Jungwon-gu, Seongnam-si,
Gyeonggi-do, 13204 Republic of Korea

Chapter 2

Safety

In this section

Owner Responsibility.	38
Safety Precautions.	39
Device Labels.	65

Owner Responsibility

Owner requirements

It is the responsibility of the owner to ensure that anyone operating the system reads and understands this section of the manual. However, there is no representation that the act of reading this manual renders the reader qualified to operate, inspect, test, align, calibrate, troubleshoot, repair or modify the system. The owner should make certain that only properly trained, fully-qualified service personnel undertake the installation, maintenance, troubleshooting, calibration and repair of the equipment.

The owner of the ultrasound unit should ensure that only properly trained, fully qualified personnel are authorized to operate the system. Before authorizing anyone to operate the system, it should be verified that the person has read, and fully understands, the operating instructions contained in this manual. It is advisable to maintain a list of authorized operators.

Should the system fail to operate correctly, or if the unit does not respond to the commands described in this manual, the operator should contact the nearest field GE HealthCare Ultrasound Service Office.

For information about specific requirements and regulations applicable to the use of electronic medical equipment, consult the local, state and federal agencies.

For USA only

Federal law restricts this device to use by, or on the orders of, a physician.

Notice against user modification

Never modify this product, including system components, software, cables, and so on. User modification may cause safety hazards and degradation in system performance. All modification must be done by a GE HealthCare qualified person.

Reporting

In the case of a serious incident occurring in relation to GE HealthCare ultrasound products, the incident should be reported to GE HealthCare and the competent Authority.

Safety Precautions

Precaution Levels

Various levels of safety precautions can be found on the equipment. Different levels of concern are identified by one of the following flag words and icons, which precede the precautionary statement.

**DANGER**

A danger symbol and icon indicates that a specific hazard is known to exist which, through inappropriate conditions or actions, will cause :

- Severe or fatal personal injury

**WARNING**

A warning symbol and icon indicates that a specific hazard is known to exist which, through inappropriate conditions or actions, may cause:

- Minor personal injury
- Substantial property damage (substantial property damage is defined as damage requiring service for the system to function)

**CAUTION**

A caution symbol and icon indicates that a potential hazard may exist which, through inappropriate conditions or actions, will or can cause:

- Property damage

NOTE

A note indicates precautions or recommendations that should be used in the operation of the ultrasound system, specifically:

- Maintaining an optimum system environment
- Using this Manual
- Notes to emphasize or clarify a point

Hazard Symbols

Icon Description

Potential hazards are indicated by the following icons:

Table 2-1 • Potential Hazards

Icon	Potential Hazard	Usage
	Biological Hazard Describes precautions necessary to prevent the risk of disease transmission or infections. • Patient/user infection due to contaminated equipment.	• Cleaning and care instructions • Sheath and glove guidelines
	Electrical Hazard Describes precautions necessary to prevent the risk of injury through electric hazards. • Electrical micro-shock to patient, e.g., ventricular.	• Probes • ECG, if applicable • Connections to back panel
	Moving Hazard Describes precautions necessary to prevent the risk of injury through moving or tipping hazard! • Console, accessories or optional storage devices that can fall on patient, user, or others. • Collision with persons or objects may result in injury while maneuvering or during system transport. • Injury to user from moving the console.	• Moving • Using brakes • Transporting
	Acoustic Output Hazard • Patient injury or tissue damage from ultrasound radiation.	• ALARA, the use of Power Output following the 'as low as reasonably achievable' principle
	Explosion Hazard Describes precautions necessary to prevent the risk of injury through explosion hazard. • Risk of explosion if used in the presence of flammable anesthetics.	• Flammable anesthetic
	Fire and Smoke Hazard • Patient/user injury or adverse reaction from fire or smoke. • Patient/user injury from explosion and fire.	• Replacing fuses • Outlet guidelines

Important Safety Considerations

The following topic headings (Patient Safety, and Equipment and Personnel Safety) are intended to make the equipment user aware of particular hazards associated with the use of this equipment and the extent to which injury can occur if precautions are not observed. Additional precautions may be provided throughout the manual.

**DANGER**

Improper use can result in serious injury. The use of the system outside the described conditions or intended use, and disregarding safety related information is considered abnormal use. The user must be thoroughly familiar with the instructions and potential hazards involving ultrasound examination before attempting to use the device. Training assistance is available from GE HealthCare if needed.

Disregarding information on safety is considered abnormal use.

Following are potential risks inherent to technology:

- Ultrasonic energy delivered to non-targeted tissue with the use of Ultrasound devices, and the interaction of sound energy with tissue at sufficiently high levels can produce biological effects.
- Monitoring the Mechanical Index (MI) can be a tool to help monitor the probability that cavitation could occur.

Patient Safety Related Hazards

**WARNING**

The concerns listed can affect the safety of patients undergoing a diagnostic ultrasound examination.

Patient Identification

Always include proper identification with all patient data and verify the accuracy of the patient's name and ID numbers when entering such data. Make sure correct patient ID is provided on all recorded data and hard copy prints. Identification errors could result in an incorrect diagnosis.

The ultrasound system is not meant to be long term storage for patient data or images. The customers are responsible for the data on the system and a regular backup is highly recommended.

It is advisable to back up system data prior to any service repairs to the hard drive. It is always possible during system failure and repair to lose patient data. GE HealthCare will not be held responsible for the loss of this data.

Diagnostic Information

The images and calculations provided by the system are intended for use by competent users, as a diagnostic tool. They are explicitly not to be regarded as the sole, irrefutable basis for clinical diagnosis. Users are encouraged to study the literature and reach their own professional conclusions regarding the clinical utility of the system.

The user should be aware of the product specifications and of the system accuracy and stability limitations. These limitations must be considered before making any decision based on quantitative values. If in doubt, the nearest GE HealthCare Ultrasound Service Office should be consulted.

Equipment malfunction or incorrect settings can result in measurement errors or failure to detect details within the image. The equipment user must become thoroughly familiar with the equipment operation in order to optimize its performance and recognize possible malfunctions. Applications training is available through the local GE HealthCare

representative. Added confidence in the equipment operation can be gained by establishing a quality assurance program.



WARNING

The system provides calculations (e.g estimated fetal weight) and charts based on published scientific literature. The selection of the appropriate chart and clinical interpretation of calculations and charts is the sole responsibility of the user. The authorized user should consider proper indications for the use of a calculation or chart as described in the scientific literature. The diagnosis, decision for further examination, and medical treatment must be performed by qualified personnel following good clinical practice.



CAUTION

Features that facilitate measurements such as VOCAL or SonoNT must be used with extreme care. The measurement results are a suggestion of the system, if in doubt verify with manual measurement methods.

The user is responsible for the diagnostic interpretation of the measurement results.

Mechanical hazards

The use of damaged probes or improper use and manipulation of intracavity probes can result in injury or increased risk of infection. Inspect probes often for sharp, pointed, or rough surface damage that could cause injury or tear protective barriers.

The use of damaged probes can result in injury or increased risk of infection. Inspect probes often for sharp, pointed, or rough surface damage that could cause injury or tear protective barriers.



ELECTRICAL HAZARD

A damaged probe can also increase the risk of electric shock if conductive solutions come in contact with internal live parts. Inspect probes often for cracks or openings in the housing and holes in and around the acoustic lens or other damage that could allow liquid entry. Become familiar with the probe's use and care precautions outlined in Probes and Biopsy.



DANGER

Ultrasound transducers are sensitive instruments which can easily be damaged by rough handling. Take extra care not to drop transducers and avoid contact with sharp or abrasive surfaces. A damaged housing, lens or cable can result in patient injury or serious impairment or operation.

ALARA



DANGER

Ultrasound can produce harmful effects in tissue and potentially result in patient injury. Always minimize exposure time and keep ultrasound levels low when there is no medical benefit. Use the principle of ALARA (As Low As Reasonably Achievable), increasing output only when needed to obtain diagnostic image quality. Observe the acoustic output display and be familiar with all controls affecting the output level. See the Bioeffects section of the Acoustic Output chapter in the Advanced Reference Manual for more information.

Training

It is recommended that all users receive proper training in applications before performing them in a clinical setting. Please contact the local GE HealthCare representative for training assistance.

ALARA training is provided in the Medical Ultrasound Safety booklet shipped in the eDOCs kit. The ALARA education program for the clinical end-user covers basic ultrasound principles, possible biological effects, the derivation and meaning of the indices, ALARA principles, and examples of specific applications of the ALARA principle.

Equipment and Personnel Safety

The concerns listed in the Equipment and Personnel Safety section can seriously affect the safety of equipment and personnel during a diagnostic ultrasound examination.

Equipment and Personal Safety Related Hazards



DANGER

This equipment contains dangerous voltages that are capable of serious injury or death.

If any defects are observed or malfunctions occur, stop operating the equipment and perform the proper action for the patient. Inform a qualified service person and contact a Service Representative for information.

There are no user serviceable components inside the console. Refer all servicing to qualified service personnel only.

Ensure that unauthorized personnel do not tamper with the unit.



ELECTRICAL HAZARD

To avoid injury:

- Do not remove protective covers. No user serviceable parts are inside. Refer servicing to qualified service personnel.
- To assure adequate grounding, connect the attachment plug to a reliable (hospital grade) grounding outlet (having equalization conductor ).
- Never use any adaptor or converter of a three-prong-to-two-prong type to connect with a mains power plug. The protective earth connection will loosen.
- Do not place liquids on or above the console. Spilled liquid may contact live parts and increase the risk of shock.
- In North America, a 220-240V installation requires the use of a center-tapped AC power source.



DANGER

The concerns listed below can seriously affect the safety of equipment and personnel during a diagnostic ultrasound examination.



EXPLOSION HAZARD

Risk of explosion if used in the presence of flammable anesthetics.

Never operate the equipment in the presence of flammable or explosive liquids, vapors or gases. Malfunctions in the unit, or sparks generated by fan motors, can electrically ignite these substances. Operators should be aware of the following points to prevent such explosion hazards.

- If flammable substances are detected in the environment, do not plug in or turn on the system.
- If flammable substances are detected after the system has been turned on, do not attempt to turn off the unit, or to unplug it.
- If flammable substances are detected, evacuate and ventilate the area before turning off the unit.



SMOKE AND FIRE HAZARD

The system must be supplied from an adequately rated electrical circuit. The capacity of the supply circuit must be as specified.



DANGER

Biological Hazard

For patient and personnel safety, be aware of biological hazards while performing invasive procedures. To avoid the risk of disease transmission:

- Use protective barriers (gloves and probe sheaths) whenever possible. Follow sterile procedures when appropriate.
- Thoroughly clean probes and reusable accessories after each patient examination and disinfect or sterilize as needed. Refer to Probes and Biopsy for probe use and care instructions.
- Follow all infection control policies established by your office, department or institution as they apply to personnel, equipment and accessories.



DANGER

Do not unpack the ultrasound system. This must be performed by qualified service personnel only. Improper unpacking could lead to injury.



DANGER

GE HealthCare recommends dedicated probes for use on humans only or animals only. Mark probes dedicated for animals with special labels.

Observe any country specific rules and regulations for handling equipment used on both animals and humans. Such national restrictions may prohibit transfer of probes used on animals to humans and vice-versa.

Failure to follow these instructions could lead to exposure to infectious agents.

**DANGER**

Reusable accessories should be cleaned and disinfected or sterilized as stated by the manufacturer, after each patient examination.

**WARNING**

- Be sure to verify the media after writing data, such as EZBackup, SaveAs or Export.
- Before deleting a patient or image from the patient screen, make sure you have saved the data by EZBackup/Backup or Export and verify that the media data transfer was successful. Not following the instructions could result in the loss of data requiring a rescan.

**WARNING**

When you move the Control Panel up/down with the monitor, place BOTH hands on the Control Panel. Touching other moving parts other than the Control Panel may cause personal injury.

**WARNING**

Contact with natural rubber latex may cause a severe anaphylactic reaction in persons sensitive to the natural latex protein. Sensitive users and patients must avoid contact with these items.

**WARNING**

The ultrasound system is not intended to be used as a data storage device; backup of the Patient and Image Database is your institution's responsibility. GE HealthCare is NOT responsible for any lost patient information or for lost images. Loss of image data may require a rescan.

**WARNING**

To minimize accidental loss of data, perform EZBackup and Backup on a regular basis.

1. First, perform EZBackup to save the images.
2. Next, perform Backup at Utility -> Backup/Restore. Enable the following checkboxes under Backup:
 - User defined configuration
 - Service

**CAUTION**

Only approved and recommended peripherals and accessories should be used.

All peripherals and accessories must be securely mounted to the ultrasound system. Failure to follow these instructions could lead to unexpected diagnostic performance.



CAUTION

Non-supported peripheral devices that use their own AC power source CANNOT be attached to the ultrasound system. DO NOT connect the peripheral device's power cord into the ultrasound system. Only peripheral devices purchased from GE HealthCare with the purpose of being used with the ultrasound system should be used.

Use a USB printer cable that is less than 3 meters in length.

Failure to follow these instructions could lead to unexpected diagnostic performance.



CAUTION

Do not use this equipment if a safety problem is known to exist. Have the unit repaired and performance verified by qualified service personnel before returning to use.



CAUTION

To avoid injury or system damage, NEVER place any object or liquid on the operator panel.

Material Safe Data

Rubber Part

Material: EPDM

Where Used: Probe holder/Gel holder/Keyboard bumper/Front and Rear Foot Pedal/System bumper at the four corners.

Monitor Related Hazards



WARNING

- DO NOT place a finger, hand or any object on the joint of the monitor or monitor arm to avoid injury when moving the monitor and monitor arm.
- To avoid result of injury or system damage, NEVER place any object or liquid on the monitor, whether in the home or flip down/transport position.
- DO NOT scratch or press on the panel with any sharp objects, such as a pencil or pen, as this may result in damage to the panel.
- To avoid injury or damage, make sure nothing is within the range of motion before moving the monitor and monitor arm. This includes both objects and people.
- Pay attention to the monitor arm position to avoid hitting it against anyone or anything.
- Before moving the system to another location, be sure to lock the monitor arm in the transport position.
- The monitor screen may have defective pixels. These pixels may appear as a slightly light or dark area on the screen. This is due to the characteristics of the panel itself, and not the product.
- The backlight of the monitor has a fixed life span. When the screen becomes dark or begins to flicker, contact a qualified Service Representative for information.

NOTE

Bright light could impact readability of screen.

Volume Navigation Related Hazards

**DANGER**

DO NOT use the Volume Navigation feature on any patient relying on life-sustaining electronic equipment, such as a pacemaker or defibrillator. Failure to follow this instruction could lead to interference with patient electronic device(s).

Specific Modes and Feature Related Hazards

**CAUTION**

- Be aware that diagnostic conclusions must not be drawn from a specific mode, such as Render Mode or LOGIQView Mode. Always check with other diagnostic procedures.
- The accuracy of measurements in specific modes such as Render Mode, LOGIQView, STIC or VOCAL is limited and can be lower than measurements in B-images.
- Do not diagnose based on 3D/4D Acquisition Mode. Always check and confirm diagnostic findings in B-Mode.

Classifications

Type of protection against electric shock

Class I Equipment (*1)

Degree of protection against electric shock

Type BF Applied part (*2) (for Probes marked with BF symbol)

Type CF Applied part (*3) (for ECG, PCG, AUX, eTRAX needle marked with CF symbol)

Continuous Operation

System is Ordinary Equipment (IPX0)

Footswitch is IPX8; Probes are IPX7

NOTE

Probe connector is not waterproof.

*1. Class I Equipment

EQUIPMENT in which protection against electric shock does not rely on BASIC INSULATION only, but includes an earth ground. This additional safety precaution prevents exposed metal parts from becoming LIVE in the event of an insulation failure.

*2. Type BF Applied Part

TYPE BF APPLIED PART providing a specified degree of protection against electric shock, with particular regard to allowable LEAKAGE CURRENT.

Table 2-2 • Type BF Equipment

Patient leakage current	Normal Mode	Single fault condition
	Less than 100 microA	Less than 500 microA

***3. Type CF Applied Part**

TYPE CF APPLIED PART providing a degree of protection higher than that for Type BF Applied Part against electric shock particularly regarding allowable LEAKAGE CURRENTS.

Table 2-3 • Type CF Equipment

Patient leakage current	Normal Mode	Single fault condition
	Less than 10 microA	Less than 50 microA

EMC (Electromagnetic Compatibility)

NOTE

This equipment generates, uses and can radiate radio frequency energy. The equipment may cause radio frequency interference to other medical and non-medical devices and radio communications. To provide reasonable protection against such interference, this product complies with emissions limits for a Group 1, Class A Medical Devices Directive as stated in EN 60601-1-2. However, there is no guarantee that interference will not occur in a particular installation.

NOTE

If this equipment is found to cause interference (which may be determined by turning the equipment on and off), the user (or qualified service personnel) should attempt to correct the problem by one or more of the following measure(s):

- Reorient or relocate the affected device(s)
- Increase the separation between the equipment and the affected device
- Power the equipment from a source different from that of the affected device
- Consult the point of purchase or service representative for further suggestions.

NOTE

The manufacturer is not responsible for any interference caused by using other than recommended interconnect cables or by unauthorized changes or modifications to this equipment. Unauthorized changes or modifications could void the users' authority to operate the equipment.

NOTE

To comply with the regulations on electromagnetic interference for a Class A FCC Device, all interconnect cables to peripheral devices must be shielded and properly grounded. Use of cables not properly shielded and grounded may result in the equipment causing radio frequency interference in violation of the FCC regulations.

**CAUTION**

Unexpected diagnostic system performance or failure from interference broadcasted by device due to electromagnetic incompatibility with cauterizing knife causing obvious deterioration of image quality.

NOTE

Do not use devices which intentionally transmit RF Signals (cellular phones, transceivers, or radio controlled products), other than those supplied by GE HealthCare, in the vicinity of the equipment, as it may cause performance outside the published specifications. Keep the power to these type devices turned off when near this equipment.

**CAUTION**

Portable RF communications equipment (including peripherals such as antenna cables and external antennas) at frequencies should be used no closer than 30cm (12 inches) to any part of the ultrasound system, including cables specified by GE HealthCare.

Otherwise, degradation of the performance of this equipment could result.

The medical staff in charge of this equipment is required to instruct technicians, patients, and other people who maybe around this equipment to fully comply with the above requirement.

EMC Performance

All types of electronic equipment may characteristically cause electromagnetic interference with other equipment, either transmitted through air or connecting cables. The term EMC (Electromagnetic Compatibility) indicates the capability of equipment to curb electromagnetic influence from other equipment and at the same time not affect other equipment with similar electromagnetic radiation from itself.

Proper installation following the service manual is required in order to achieve the full EMC performance of the product.

The product must be installed as stipulated in 4.2, Notice upon Installation of Product.

**CAUTION**

Use of this equipment adjacent to or stacked with other equipment should be avoided because it could result in improper operation. If such use is necessary, this equipment and the other equipment should be observed to verify that they are operating normally. Failure to follow these instruction could lead to unexpected diagnostic performance.

In case of issues related to EMC, please call your service personnel.

The manufacturer is not responsible for any interference caused by using other than recommended interconnect cables or by unauthorized changes or modifications to this equipment. Unauthorized changes or modifications could void the users' authority to operate the equipment.



WARNING

Do not use devices which intentionally transmit RF signals (cellular phones, transceivers, or radio controlled products), other than those supplied by GE HealthCare (wireless microphone, broadband over power lines, for example) unless intended for use with this system, in the vicinity of this equipment as it may cause performance outside the published specifications.

Keep power to these devices turned off when near this equipment.

Medical staff in charge of this equipment is required to instruct technicians, patients and other people who may be around this equipment to fully comply with the above regulation.

Portable and mobile radio communications equipment (e.g. two-way radio, cellular/cordless telephones, wireless computer networks) should be used no closer to any part of this system, including cables, than determined according to the following method:

Table 2-4 • Recommended separation distances between portable and mobile RF communications equipment and this system

This system is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of this system can help prevent electromagnetic Interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and this system as recommended below, according to the maximum output power of the communications equipment.			
Rated Maximum Output Power (P) of Transmitter Watts (W)	Separation distance according to transmitter frequency		
	150 kHz - 80 MHz $d = [3.5/3] \text{ square root of } P$	80 MHz - 800 MHz $d = [3.5/3] \text{ square root of } P$	800 MHz - 2.5 GHz $d = [7/3] \text{ square root of } P$
	Where: d= separation distance in meters, P = rated power of the transmitter.		
	The separation distance in meters should be		
0.01	0.12	0.12	0.23
0.1	0.37	0.37	0.74
1	1.17	1.17	2.33
10	3.69	3.69	7.38
100	11.7	11.7	23.3
For transmitters rated at a maximum output power not listed above, the separation distance can be estimated using the equation in the corresponding column, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer. NOTE: At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies. These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects, and people.			

Notice upon Installation of Product

Separation distance and effect from fixed radio communications equipment: field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast, and TV broadcast transmitter cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the ultrasound system is used exceeds the

applicable RF compliance level as stated in the immunity declaration, the ultrasound system should be observed to verify normal operation. If abnormal operation is observed, additional measures may be necessary, such as re-orienting or relocating the ultrasound system or using an RF shielded examination room may be necessary.

1. Use either power supply cords provided by GE HealthCare or ones designated by GE HealthCare. Products equipped with a power source plug should be plugged into the fixed power socket which has the protective grounding conductor. Never use any adaptor or converter to connect with a power source plug (e.g. three-prong-to-two-prong converter).
2. Locate the equipment as far away as possible from other electronic equipment.
3. Be sure to use only the cables provided by or designated by GE HealthCare. Connect these cables following the installation procedures (e.g. wire power cables separately from signal cables).
4. Lay out the main equipment and other peripherals following the installation procedures described in the Option Installation manuals.

General Notice



WARNING

Use of accessories, transducers and cables other than those specified or provided by GE of this equipment could result in increased electromagnetic emissions or decreased electromagnetic immunity of this equipment and result in improper operation.

1. Designation of Peripheral Equipment Connectable to This Product.
The equipment indicated in the Supplies/Accessories section can be hooked up to the product without compromising its EMC performance.
Avoid using equipment not designated in the list. Failure to comply with this instruction may result in poor EMC performance of the product.
2. Notice against User Modification.
The user should never modify this product. User modifications may cause degradation in EMC performance.
Modification of the product includes changes in:
Cables (length, material, wiring, etc.)
System installation/layout
System configuration/components
Securing system parts (cover open/close, cover screwing)
3. Operate the system with all covers closed. If a cover is opened for some reason, be sure to shut it before starting/resuming operation.
4. Operating the system with any cover open may affect EMC performance.

Peripheral Update for EC countries

The following is intended to provide the users in EC countries with updated information concerning the connection of the ultrasound system to image recording and other devices or communication networks.

Peripherals Used in the Patient Environment

The ultrasound system has been verified for overall safety, compatibility and compliance with the following image recording devices:

- Sony UP-D898DC Digital Printer
- Sony UP-D25 Digital Printer
- Sony UP-DR80MD Color Printer
- Drive Bay (for Tru3D and Volume Navigation)
- USB 2.0/3.0 Flash Drive
- Hard Disk Drive

NOTE

The ultrasound system supports the USB 3.0 standard (on the Operator Panel and Peripheral/Accessory Connector Panel) and the USB 2.0 standard on the Monitor.

The ultrasound system has also been verified for compatibility, and compliance for connection to a local area network (LAN) via the rear panel Ethernet connection, provided the LAN components are IEC/EN 60950 and IEC/EN 62368 compliant.

A Wireless LAN option is available. *Configuring the Wireless Network* on page 1088. Conforms to IEEE 802.11 ac/a/b/g/n Wi-Fi with Bluetooth 5.2 Standard.

The ultrasound system may also be used safely while connected to devices other than those recommended above if the devices and their specifications, installation, and interconnection with the system conform to the requirements of IEC/EN 60601-1.

Accessory equipment connected to the analog and digital interfaces must be certified according to the respective IEC standards (i.e., IEC60950 for data processing equipment and IEC60601-1 for medical equipment). Furthermore, all complete configurations shall comply with the valid version of the system standard IEC60601-1. Everyone who connects additional equipment to the signal input part or signal output part of the ultrasound system configures a medical system, and is therefore responsible to ensure that the system complies with the requirements of the valid version of IEC60601-1. If in doubt, consult the technical service department or your local GE HealthCare representative.

General precautions for installing an alternate on-board device would include:

1. The added device must have appropriate safety standard conformance and CE Marking.
2. The total power consumption of the added devices, which connect to the ultrasound system and are used simultaneously, must be less than or equal to the rated supply of the ultrasound system.
3. There must be adequate heat dissipation and ventilation to prevent overheating of the device.
4. There must be adequate mechanical mounting of the device and stability of the combination.
5. Risk and leakage current of the combination must comply with IEC/EN 60601-1.
6. Electromagnetic emissions and immunity of the combination must conform to IEC/EN 60601-1-2.

General precautions for installing an alternate off-board, remote device or a network would include:

1. The added device(s) must have appropriate safety standard conformance and CE Marking.
2. The added device(s) must be used for their intended purpose having a compatible interface.
3. Signal or mains isolation devices and additional protective earth may be needed to assure compliance with IEC/EN 60601-1.

Declaration of Emissions

This system is suitable for use in the following environment. The user must assure that it is used only in the electromagnetic environment as specified.

Table 2-5 • Declaration of Emissions

Guidance and manufacturer's declaration - electromagnetic emissions		
The system is intended for use in the electromagnetic environment specified below. The user of the system should assure that it is used in such an environment.		
Emission Type	Compliance	Electromagnetic Environment
RF Emissions CISPR 11	Group 1	This system uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF Emissions CISPR 11	Class A	This system is suitable for use in all establishments, other than domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes, provided the following warning is heeded: WARNING: The EMISSIONS characteristics of this equipment make it suitable for use in industrial areas and hospitals (CISPR 11 class A). If it is used in a residential environment (for which CISPR 11 class B is normally required) this equipment might not offer adequate protection to radio-frequency communication services. The user might need to take mitigation measures, such as relocating or reorienting the equipment.
Harmonic Emissions IEC 61000-3-2	Class A	
Voltage Fluctuations/ Flicker Emissions IEC 61000-3-3	Complies	

Declaration of Immunity

This system is suitable for use in the following environment. The user must ensure that the system is used according to the specified guidance and only in the electromagnetic environment listed.

Table 2-6 • Declaration of Immunity Compliance Statement Supporting

Guidance and manufacturer's declaration - electromagnetic immunity			
This system is intended for use in the electromagnetic environment specified below. The customer or the user of this system should assure that it is used in such an environment.			
Environments of INTENDED USE	Professional healthcare facility environment		
Immunity Type	IEC 60601-1-2 test level	IEC 60601-1-2 Compliance Level	Electromagnetic Environment Guidance
IEC 61000-4-2 Static discharge (ESD)	± 8 kV contact ± 15 kV air	± 8 kV air ± 15 kV contact	Floors should be wood, concrete, or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%.
IEC 61000-4-4 Electrical fast transient/burst	±2 kV for power supply lines 100 kHz rate ± 1 kV for input/output lines 100 kHz rate	± 2 kV for power supply lines 100 kHz rate ± 1 kV for input/output lines 100 kHz rate	Mains power quality should be that of a typical commercial and/or hospital environment.
IEC 61000-4-5 Surge Immunity	± 1 kV line-line ± 2 kV line-earth	± 1 kV line-line ± 2 kV line-earth	Mains power quality should be that of a typical commercial and/or hospital environment.
IEC 61000-4-11 Voltage dips and short interruptions and voltage variations on power supply input lines	0 % UT; 0.5cycle, Phase: 0,45,90,135,180,225,270,315° 0 % UT; 1cycle, Phase: 0° 70 % UT; 25/30cycle, Phase: 0° 0 % UT; 250/300 cycle Note: Apply IEC 60601-2-37 (202.6.2.7)	0 % UT; 0.5cycle, Phase: 0,45,90,135,180,225,270, 315° 0 % UT; 1cycle, Phase: 0° 70 % UT; 25/30cycle, Phase: 0° 0 % UT; 250/300 cycle Note: Apply IEC 60601-2-37 (202.6.2.7)	Mains power quality should be that of a typical commercial and/or hospital environment. If the user requires continued operation during power mains interruptions, it is recommended that the system be powered from a UPS or a battery option.
IEC 61000-4-8 Power frequency (50/60 Hz) magnetic field	30 A/m	30 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial and/or hospital environment.
NOTE: UT is the AC mains voltage prior to application of the test level.			

Guidance and manufacturer's declaration - electromagnetic immunity			
This system is intended for use in the electromagnetic environment specified below. The customer or the user of this system should assure that it is used in such an environment.			
Environments of INTENDED USE	Professional healthcare facility environment		
Immunity Type	IEC 60601-1-2 test level	IEC 60601-1-2 Compliance Level	Electromagnetic Environment Guidance
IEC 61000-4-6 Conducted RF	3 Vrms 150kHz to 80MHz 6 Vrms in ISM bands between 150 kHz to 80MHz	3 Vrms 150kHz to 80MHz 6 Vrms in ISM bands between 150 kHz to 80MHz	Portable and mobile RF communications equipment should be used no closer to any part of this system, including cables, than the recommended separation distance calculated from the equation appropriate for the frequency of the transmitter. Recommended Separation Distance (see <i>Table 2-4 • Recommended separation distances between portable and mobile RF communications equipment and this system</i> on page 50) $d = \left[\frac{3.5}{3} \right] \sqrt{P}$ (80 MHz to 800 MHz (see <i>Table 2-4 • Recommended separation distances between portable and mobile RF communications equipment and this system</i> on page 50) $d = \left[\frac{3.5}{3} \right] \sqrt{P}$ (800 MHz to 2.7 GHz (see <i>Table 2-4 • Recommended separation distances between portable and mobile RF communications equipment and this system</i> on page 50) $d = \left[\frac{7}{3} \right] \sqrt{P}$ where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in meters (m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey, a should be less than the compliance level in each frequency range b. Interference may occur in the vicinity of equipment marked with the symbol: 
IEC 61000-4-3 Radiated RF EM fields	3 V/m 80 MHz - 2.7 GHz 80% AM 1 kHz	3 V/m 80 MHz - 2.7 GHz 80% AM 1 kHz	

Guidance and manufacturer's declaration - electromagnetic immunity			
This system is intended for use in the electromagnetic environment specified below. The customer or the user of this system should assure that it is used in such an environment.			
Environments of INTENDED USE	Professional healthcare facility environment		
Immunity Type	IEC 60601-1-2 test level	IEC 60601-1-2 Compliance Level	Electromagnetic Environment Guidance
IEC 61000-4-3 Proximity fields from RF wireless communications equipment	9 V/m to 28 V/m spot frequencies 385 MHz: 27 V/m 450 MHz: 28 V/m 710 MHz: 9 V/m 745 MHz: 9 V/m 780 MHz: 9 V/m 810 MHz: 28 V/m 870 MHz: 28 V/m 930 MHz: 28 V/m 1720 MHz: 28 V/m 1845 MHz: 28 V/m 1970 MHz: 28 V/m 2450 MHz: 28 V/m 5240 MHz: 9 V/m 5500 MHz: 9 V/m 5785 MHz: 9 V/m PM 18 Hz or 217 Hz (50% duty cycle)	9 V/m to 28 V/m spot frequencies 385 MHz: 27 V/m 450 MHz: 28 V/m 710 MHz: 9 V/m 745 MHz: 9 V/m 780 MHz: 9 V/m 810 MHz: 28 V/m 870 MHz: 28 V/m 930 MHz: 28 V/m 1720 MHz: 28 V/m 1845 MHz: 28 V/m 1970 MHz: 28 V/m 2450 MHz: 28 V/m 5240 MHz: 9 V/m 5500 MHz: 9 V/m 5785 MHz: 9 V/m PM 18 Hz or 217 Hz (50% duty cycle)	Minimum separation distances for higher IMMUNITY TEST LEVELS shall be calculated using the following equation: $E = \frac{6}{d} \sqrt{P}$ Where P is the maximum power in W, d is the minimum separation distance in m, and E is the IMMUNITY TEST LEVEL in V/m.
a. Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast, and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which this system is used exceeds the applicable RF compliance level above, this system should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating this system. b. Over the frequency range 150 kHz to 80 MHz field strengths should be less than 3 V/m. c. As an alternative to FM modulation, 50 % pulse modulation at 18 Hz may be used because while it does not represent actual modulation, it would be worst case. NOTE: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects, and people.			

Table 2-7 • Declaration of Immunity Compliance Statement Supporting IEC 60601-1-2:2014+A1:2020

Guidance and manufacturer's declaration - electromagnetic immunity			
This system is intended for use in the electromagnetic environment specified below. The customer or the user of this system should assure that it is used in such an environment.			
Environments of INTENDED USE	Professional healthcare facility environment		
Immunity Type	IEC 60601-1-2 test level	IEC 60601-1-2 Compliance Level	Electromagnetic Environment Guidance
IEC 61000-4-39 Radiated fields in close proximity	65A/m 134.2kHz PM 2.1kHz 7.5A/m 13.56MHz PM 50kHz	65A/m 134.2kHz PM 2.1kHz 7.5A/m 13.56MHz PM 50kHz	Radiated fields in close proximity should be that of a typical commercial and/or hospital environment. Both 134.2 kHz and 13.56 MHz are used for RFID reader operation.

Table 2-8 • Spot Frequencies for IEC 61000-4-3 Proximity fields from RF wireless communications equipment

Test Frequency (MHz)	Band (MHz)	Service	Modulation	Maximum Power (W)	Distance (m)	Immunity Test Level (V/m)
385	380 - 390	TETRA 400	Pulse modulation 18 Hz	1.8	0.3	27
450	430 - 470	GMRS 460, FRS 460	FM $\pm$ 5 kHz deviation 1 kHz sine	2	0.3	28
710	704 - 787	LTE Band 13, 17	Pulse modulation 217 Hz	0.2	0.3	9
745						
780						
810	800 - 960	GSM 800/900 TETRA 800 iDEN 820 CDMA 850 LTE Band 5	Pulse modulation 18 Hz	2	0.3	28
870						
930						
1720	1700 - 1990	GSM 1800; CDMA 1900; GSM 1900; DECT; LTE Band 1, 3, 4, 25; UMTS	Pulse modulation 217 Hz	2	0.3	28
1845						
1970						

Test Frequency (MHz)	Band (MHz)	Service	Modulation	Maximum Power (W)	Distance (m)	Immunity Test Level (V/m)
2450	2400 - 2570	Bluetooth, WLAN, 802.11 b/g/n, RFID 2450, LTE Band 7	Pulse modulation 217 Hz	2	0.3	28
5240	5100 - 5800	WLAN 802.11 a/n	Pulse modulation 217 Hz	0.2	0.3	9
5500						
5785						

Essential performance

The essential performance of the ultrasound unit is:

- The ability to display B-mode image as input for diagnosis.
- The ability to display M-mode image as input for diagnosis.
- The ability to display Doppler-mode image as input for diagnosis.
- The ability to display Color Flow-mode image as input for diagnosis.
- The display of acoustic power indexes as an aid for safe use of ultrasound diagnostic (MI, TIS, TIB, TIC).

Patient Environmental Devices

Figure 2-1 • Patient Environmental Devices



1. USB port	7. Probe Storage
2. Front Panel	8. Power In/Out
3. Gel warmer	Power Line (AC~)
4. Peripheral Devices	Ground Line
DVD Multi Drive (CD, DVD, DVR), Black/White	Power Cable with Protective Earth
Printer, Patient I/O (ECG), VNav Connectors	9. Signals I/O Port (USB 3.0 Port, Ethernet
5. Network/Disk Indicators	Connector, HDMI Connector, S-Video
6. Probe Ports (4)	Option Connector)
Probe Ports with CW Option	

Figure 2-2 • Patient Environmental Devices

Acceptable Devices

The Patient Environmental devices shown on the previous page are specified to be suitable for use within the PATIENT ENVIRONMENT.



CAUTION

DO NOT connect any probes or accessories without approval by GE HealthCare within the PATIENT ENVIRONMENT.

Any device connected to the ultrasound system must conform to one or more of the requirements listed below:

1. IEC standard or equivalent standards appropriate to devices.
2. The devices shall be connected to PROTECTIVE EARTH (GROUND).

Unapproved Devices



DANGER

DO NOT use unapproved devices.

If devices are connected without the approval of GE HealthCare, the warranty will be INVALID.

Any device connected to the Ultrasound System must conform to one or more of the requirements listed below:

1. IEC standard or equivalent standards appropriate to devices.
2. The devices shall be connected to PROTECTIVE EARTH (GROUND).

Accessories, Options, Supplies



CAUTION

Unsafe operation or malfunction may result. Use only the accessories, options and supplies approved or recommended in these instructions for use.

Acoustic Output

When the “Auto Freeze Time (probe selection required)” preset is selected on the **Utility > System > System Imaging** screen, the system auto freezes after the time specified (10 or 30 minutes, 1 hour, or Never) of scanning if it detects no change in the image.

Located on the upper right section of the system display monitor, the acoustic output display provides the operator with real-time indication of acoustic levels being generated by the system. See the Acoustic Output chapter in the Advanced Reference Manual for more information.

Acoustic Output Display Specifications

The display consists of three parts: Thermal Index (TI), Mechanical Index (MI), and a relative Acoustic Output (AO) value.

The TI and MI are displayed at all times. The TI display starts at a value of 0.0, increments in steps of 0.1 and accuracy is $\pm 50\%$. For MI display values between 0 and 0.4, the display increments in steps of 0.01 and for values greater than 0.4, increments in steps of 0.1. The MI accuracy is $\pm 25\%$.

The AO percentage (AO%) value informs the user of where the system is operating within the range of available output. Accuracy of the AO% is $\pm 10\%$.

Thermal Index

Depending on the examination and type of tissue involved, the TI parameter will be one of three types:

- Soft Tissue Thermal Index (TIS). Used when imaging soft tissue only, it provides an estimate of potential temperature increase in soft tissue.
- Bone Thermal Index (TIB). Used when bone is near the focus of the image as in the third trimester OB examination, it provides an estimate of potential temperature increase in the bone or adjacent soft tissue.
- Cranial Bone Thermal Index (TIC). Used when bone is near the skin surface as in transcranial examination, it provides an estimate of potential temperature increase in the bone or adjacent soft tissue.

Mechanical Index

MI recognizes the importance of non-thermal processes, cavitation in particular. The Index is a relative indicator of the likelihood of mechanical bioeffect within the tissue.

Changing the Thermal Index Type

You can select the displayed TI type on Utility > Imaging > B-Mode. This preset is application dependent so each application could specify a different TI type.

Controls Affecting Acoustic Output

The potential for producing mechanical bioeffects (MI) or thermal bioeffects (TI) can be influenced by certain controls.

Direct. The Power Output control has the most significant effect on Acoustic Output.

Indirect. Indirect effects may occur when adjusting controls. Controls that can influence MI and TI are detailed under the Bioeffects portion of each control in the Optimizing the Image sections.

Always observe the Acoustic Output display for possible effects.

Best Practices While Scanning



HINTS

Raise the Acoustic Output only after attempting image optimization with controls that have no effect on Acoustic Output, such as Gain and TGC.

NOTE

Refer to the Optimizing the Image sections for a complete discussion of each control.



DANGER

Be sure to have read and understood control explanations for each mode used before attempting to adjust the Acoustic Output control or any control that can affect Acoustic Output. During a screening and diagnostic ultrasound examination, high frequency sound penetrates and interacts with tissue in and around the area of anatomy to be imaged. Only a small portion of this sound energy is reflected back to the transducer for use in constructing the image, while the remainder is dissipated within the tissue. The interaction of sound energy with tissue at sufficiently high levels can produce biological effects (aka bioeffects) of either a mechanical or thermal nature. Although the generation of bioeffect is intentional with therapeutic ultrasound, it is generally undesired in screening and diagnostic applications and may be harmful in some conditions.



ACOUSTIC OUTPUT HAZARD

Use the minimum necessary acoustic output to get the best diagnostic image or measurement during an examination. Begin the exam with the probe that provides an optimum focal depth and penetration.

Acoustic Output Default Levels

In order to assure that an exam does not start at a high output level, the ultrasound system initiates scanning at a reduced default output level. This reduced level is preset programmable and depends upon the exam category and probe selected. It takes effect when the system is powered on or **New Patient** is selected.

To modify acoustic output, adjust the Power Output level on the Touch Panel.




Figure 2-3 • Location of Power Output control

Acoustic Output on Pleural Preset

The U.S. Food and Drug Administration (FDA) recommends that MI levels should be lower than 1.4 when scanning lungs.

Select the Pleural preset when performing a lung exam. The displayed MI will not exceed 0.8 with Pleural preset selected. Differences in MI values are a result of probe to probe variation as well as the calculation accuracy of the displayed MI value.

RoHS Hazardous Substances

The following product pollution control information is provided according to SJ/T11364-2014 Marking for Control of Pollution caused by Electronic Information Products.



The symbol above indicates the product contains hazardous materials in excess of the limits established by the Chinese standard GB/T 26572 Requirements for Concentration Limits for Certain Hazardous Substances in Electronic Information Products. The number in the symbol is the Environment-friendly Use Period (EFUP), which indicates the period during which the hazardous substances or elements contained in electrical and electronic products will not leak or mutate under normal operating conditions so that the use of such electrical and electronic products will not result in any severe environmental pollution, any bodily injury or damage to any assets. The unit of the period is "Year."

In order to maintain the declared EFUP, the product shall be operated normally according to the instructions and environmental conditions as defined in the product manual, and periodic maintenance schedules specified in Product Maintenance Procedures shall be followed strictly. Consumables or certain parts may have their own label with an EFUP value less than the product. Periodic replacement of those consumables or parts to maintain the declared EFUP shall be done in accordance with the Product Maintenance Procedures.

This product must not be disposed of as unsorted municipal waste, and must be collected separately and handled properly after decommissioning.

Name and Concentration of Hazardous Substances

Table 2-9 • Table of hazardous substances' name and concentration

Component Name	Hazardous substances' name					
	Pb	Hg	Cd	Cr (VI)	PBB	PBDE
HDU Display	X	O	O	O	O	O
LCD Display	X	O	O	O	O	O
Circuit Boards	X	O	O	O	O	O
Touch Panel	X	O	O	O	O	O
Ultrasound Probes	X	O	O	O	O	O
Console Cabinet	X	O	O	O	O	O
Operator Panel	X	O	O	O	O	O
Console Frame (Base Casting, Castings, Card Rack)	X	O	O	O	O	O
System Covers	X	O	O	O	O	O
System Cables	X	O	O	O	O	O

This table is prepared according to SJ/T 11364.

O: Indicates that this hazardous substance contained in all of the homogeneous materials for this part is below the limit requirement in GB/T 26572.

X: Indicates that this hazardous substance contained in at least one of the homogeneous materials used for this part is above the limit requirement in GB/T 26572.

Data listed in the table represents best information available at the time of publication.

Applications of hazardous substances in this medical device are required to achieve its intended clinical uses, and/or to provide better protection to human beings and/or to environment, due to lack of reasonably (economically or technically) available substitutes

Note: Options may not be present on every system.

WEEE Passport

The WEEE (Waste Electrical and Electronic Equipment) Passport describes product recycling information. To access the WEEE passport for GE HealthCare products:

1. Go to the GE HealthCare Support Documentation Library at:
<https://www.gehealthcare.com/documentation>
2. Select the modality "Ultrasound (UL)."
3. Enter the document name or the keyword "WEEE."
4. Press "Search."
5. Select the desired WEEE passport.

Safe Product and Packaging Disposal

This product and package should be disposed of according to hospital disposal practices, and local environmental and waste disposal regulations. Components and accessories of the ultrasound system which have come into direct or indirect contact with the patient may be biohazardous, and should be disposed of according to facility guidelines for biohazardous material. The waste of electrical and electronic equipment must not be disposed as unsorted municipal waste and must be collected separately. Please contact an authorized representative of the manufacturer for information concerning the disposal/decommissioning of equipment.

Device Labels

Label Icon Description

The following table describes the purpose and location of safety labels and other important information provided on the equipment

NOTE


This machine should be used in compliance with law. Some jurisdictions restrict certain uses, such as gender determination (shown for country specific label).




Figure 2-4 • Rear Panel Label Location


Table 2-10 • Label Icons


No.	Label/Icon	Purpose/Meaning
1.		Monitor Label: GE created
		DO NOT place a finger, hand or any object on the joint of the monitor or monitor arm to avoid injury when moving the monitor and monitor arm.

No.	Label/Icon	Purpose/Meaning
1. and 3.		Follow instruction for use. IEC 60601-1:2005+A1:2012 Annex D.1 and ISO 7010-M002
1. and 3.	RoHS Label–China systems only (shown for country specific label) 	Indicates the presence of hazardous substance(s) above the maximum concentration value. Maximum concentration values for electronic information products, as set by the People's Republic of China Electronic Industry Standard SJ/T11364-2006, include the hazardous substances of lead, mercury, hexavalent chromium, cadmium, polybrominated biphenyl (PBB), and polybrominated diphenyl ether (PBDE). "20" indicates the number of years during which the hazardous substance(s) will not leak or mutate so that the use of this product will not result in any severe environmental pollution, bodily injury, or damage to any assets. China Electronic Industry Standard SJ/T11364-2014
1 and 3.		This WEEE symbol indicates that waste electrical and electronic equipment must not be disposed of as unsorted municipal waste and must be collected separately. Please contact an authorized representative of the manufacturer for information concerning the decommissioning of your equipment. Standard: EN 50419. WEEE Directive 2012/19/EU
1. and 3.		"General Warning Sign" Possible shock hazard. Do not remove covers or panels. No user serviceable parts are inside. Refer servicing to qualified service personnel. Standard: ISO 7010-W001.
2.		The system contains the following wireless module: INTEL AX210NGW. Australia - RCM certification mark Moldova - SM certification mark and number Malaysia - MCMC certification mark and number South Korea - KC certification mark and number Pakistan - PTA certification number South Africa - ICASA certification mark and number Serbia- RATEL certification mark Ukraine - NCCIR certification mark Brazil - ANATEL certification mark and number USA - FCC certification mark and number Canada - ISED certification number China - SRRC certification number Belarus - TR BY certification mark Singapore - IMDA certification mark Taiwan - NCC certification mark and number Taiwan - BSMI certification mark and number Japan - MIC certification mark and number UAE - TDRA certification mark and number
2.		Brazil - INMETRO Certification: TUV Rheinland Brazil

No.	Label/Icon	Purpose/Meaning
3.		Do not use the following devices near this equipment: cellular phone, radio receiver, mobile radio transmitter, radio controlled toy, broadband power lines, etc. Use of these devices near this equipment could cause this equipment to perform outside the published specifications. Keep power to these devices turned off when near this equipment. ISO 7010-P013
3.		DO NOT push the system. Use the handle to push/pull the system, e.g., DO NOT use the monitor. Failure to do so may cause serious injury or system damage. IEC 60601-1:2005+A1:2012 Annex D.2 and ISO 7010-P017
3.	Rx Only	United States only Prescription Device label 21 CFR 801.109 and Alternative to Certain Prescription Device Labeling Requirements Guidance to Industry 1/2/2000 U.S. Food & Drug Administration modified by General Electric for clarity that this is for the USA
3.		Type BF Applied Part (man in the box) symbol is in accordance with IEC 60417-5333.
3.		Symbol indicating that the Instructions for Use are supplied in electronic form. ISO 7000-3500
3.		Authorized European Representative address
3.		The CH-REP label indicates the authorized representative in Switzerland.
3.		Use two people to transport system on inclines. To avoid possible injury and equipment damage when transporting from one area of use to another: Be sure the pathway is clear. Limit movement to a slow careful walk. Use two or more persons to move the equipment on inclines or long distance. GE created
3.		Unlock/Lock the monitor arm GE created
3.		The MR Unsafe label is to warn that this Ultrasound System poses risks to the patient, medical staff or other persons within the MR environment and that the Ultrasound system should remain outside of the MR environment.
3.		The CE Mark of Conformity indicates this equipment conforms with the Council Directive 2017/745 EU. European Regulation 2017/745 EU

No.	Label/Icon	Purpose/Meaning
3.		This product is a medical device.
4.		Possible shock hazard. Do not remove covers or panels. No user serviceable parts are inside. Refer servicing to qualified service personnel. USB 3.0 Ports Ethernet Connection S-Video Connector HDMI Connector
		“Equipotentiality” indicates the terminal to be used for connecting equipotential conductors when interconnecting (grounding) with other equipment. Connection of additional protective earth conductors or potential equalization conductors is not necessary in most cases and is only recommended for situations involving multiple equipment in a high-risk patient environment to provide assurance that all equipment is at the same potential and operates within acceptable leakage current limits. An example of a high-risk patient would be a special procedure where the patient has an accessible conductive path to the heart such as exposed cardiac pacing leads. IEC60417-5021 IEC 60601-1:2005+A1:2012 Annex D.1 and IEC 60417-5021
		Mains OFF: indicates the power off position of the mains power breaker. IEC 60601-1:2005+A1:2012 Annex D.1 and IEC 60417-5008
		Mains ON: indicates the power on position of the mains power breaker. IEC 60601-1:2005+A1:2012 Annex D.1 and IEC 60417-5007

No.	Label/Icon	Purpose/Meaning
5.		Every system has a unique marking for identification, the Unique Device Identification (UDI) Label. The UDI label consists of a series of alpha-numeric characters and barcode which uniquely identify the ultrasound system as a medical device manufactured by General Electric. Scan or enter the UDI information into the patient health record as required by country-specific laws. This label also indicates the system weight.
		UDI Human Readable Label Text: Global Trade Item Number, GTIN, Manufacturing Date, Serial Number
		UDI Symbol and Data Matrix
		System Voltage (~100-240VAC) Frequency Power Rating
		Alternating Current symbol is in accordance with IEC 60417-5032
		Catalog/Model Number. Standard: ISO 7000-2493.
		Serial Number. Standard: ISO 7000-2498.
		Legal Manufacturer's name and address. Standard: ISO 7000-3082.
		Date of manufacture YYYY-MM. Standard: ISO 7000-2497.
6.		IN ACCORDANCE WITH ANSI/AAMI ES 60601-1:2005/A2:2021 CAN/CSA-C22.2 No. 60601-1 (Amendment2:2022)
		This system has Power Assistant option, GE created
		This system has Scan on Battery option, GE created
7.		Gender Caution (only for India, China, Korea)
8.		GE CARES Label (US/Canada systems only)

No.	Label/Icon	Purpose/Meaning
9.	Country-Specific Labels 	Example: INMETRO Certification: TUV Rheinland Brazil
		Rating label for Japan
		Rating label for Korea
ECG		eTRAX needle, PCG, AUX Type CF Applied Part (heart in the box) symbol is in accordance with IEC 60417-5335.
ECG		ECG Type CF Defib-Proof Applied Part (heart in the box with paddle) symbol is in accordance with IEC 60417-5336.
Monitor Arm		DO NOT place a finger, hand or any object on the joint of the monitor or monitor arm to avoid injury when moving the monitor and monitor arm. GE created
V-Nav		ISO 7010 - P007 Volume Navigation Pacemaker Warning
Op Panel		Power On/Off. Standard: IEC 60417-5010
Op Panel		To avoid injury, DO NOT place a finger, hand or any object below the operator panel when it is lowered.
Foot Pedal and Battery Pack		The product has been tested and approved in UL Laboratories, based on UL and CSA standards, through mutual approval activities.

No.	Label/Icon	Purpose/Meaning
Battery Pack		Battery Pack if contains Pb/Cd/Hg The separate collection symbol is affixed to a battery, or its packaging, to advise you that the battery must be recycled or disposed of in accordance with local or country laws. The letters below the separate collection symbol indicate whether certain elements (Pb=Lead, Cd=Cadmium, Hg=Mercury) are contained in the battery. To minimize potential effects on the environment and human health, it is important that all marked batteries that you remove from the product are properly recycled or disposed. For information on how the battery may be safely removed from the device, please consult the service manual or equipment instructions. Information on the potential effects on the environment and human health of the substances used in batteries is available at this url: http://www.gehealthcare.com/euen/weee-recycling/index.html . WEEE Directive 2012/19/EU

Probe Labeling

Each probe is labeled with the following information:

- Seller's name and manufacturer
- Operating frequency (not shown on all probes)
- GE HealthCare part number
- Probe serial number
- Month and year of manufacture
- Probe designation-provided on the probe grip and the top of the connector housing, so it is easily read when mounted on the system and is also automatically displayed on the screen when the probe is selected.
- UDI Symbol and Data Matrix
- DI Human Readable LabelText: Global Trade Item Number, GTIN

The following information appears on all probe labels, regardless of the connector type, except for "IPX7," "CE Mark," "MD symbol," and "XDclear™" which only appears on applicable probes.



Figure 2-5 • Probe Label

1. GE HealthCare Logo

9. CE Mark and Notified Body Number

2. Probe Model (Name)	10. REF: Catalog/model number
3. UDI Symbol and Data Matrix	11. Serial Number
4. UDI Human Readable LabelText: Global Trade Item Number, GTIN, (01), Manufacturing Date (11), Serial Number (21)	12. Manufacturer's site country of origin
5. Type BF/CF Applied Part	13. Legal Manufacturer's Name and Address
6. Caution: Consult the Manual.	14. Date of Manufacture, as YYYY-MM
7. WEEE Waste Symbol	15. Product Marketing Indicator information may appear here.
8. Chinese RoHS Hazardous Substance Symbol	16. IP Classification
	17. Symbol indicates the item is a medical device.


NOTE

Non-GE probes will also have a UDI symbol and equivalent information.

Probe Packaging Labels

The labels in *Table 2-11 • Probe Package Labels* on page 72 may be found on probe packaging.

Table 2-11 • Probe Package Labels

Label/Icon	Purpose/Meaning
	To indicate the acceptable upper and lower limit of atmospheric pressure for transport and storage.
	To identify the temperature limits, for example on transport packaging to indicate limits within which the package must be kept and handled. The temperature values may be shown adjacent to the symbol.
	To indicate the acceptable upper and lower limits of relative humidity for transport and storage.
	On transport packaging, to indicate that the content of the package is fragile and that the package must be handled with care.
	On transport packaging, to indicate that the package must be kept in dry conditions.
	To indicate that the marked item or its material is part of a recovery or recycling process.

Label/Icon	Purpose/Meaning
	"ATTENTION - Consult accompanying documents" is intended to alert the user to refer to the operator manual or other instructions when complete information cannot be provided on the label.

Chapter 3

Preparing the System for Use

In this section

- Site Requirements. 76
- Console Overview. 78
- Powering the System. 88
- System Language Configuration. 102
- Probe Connections. 107
- Operator Controls. 113
- Monitor. 123
- Monitor Display. 132
- System Positioning and Transporting. 139

Site Requirements

Only qualified physicians or sonographers should perform ultrasound scanning on human subjects for medical diagnostic reasons. Request training, if needed.

Do not attempt to set up the system alone. General Electric, Affiliate, or Distributor Field Engineers and Application Specialists will install and setup the system. See *Contact Information* on page 30 for more information.

The ultrasound system does not contain any operator serviceable internal components. Ensure that unauthorized personnel do not tamper with the unit.

Perform regular preventive maintenance. See *System Care and Maintenance* on page 702 for more information.

Maintain a clean environment. Turn off, and if possible, disconnect the system before cleaning the unit. See *Cleaning and Disinfecting the System* on page 703 for more information.

Never set liquids on the unit to ensure that liquid does not drip into the control panel or unit.

Before the system arrives

NOTICE

This medical equipment is approved, in terms of the prevention of radio wave interference, to be used in hospitals, clinics and other institutions which are environmentally qualified. The use of this equipment in an inappropriate environment may cause some electronic interference to radios and televisions around the equipment.

Ensure that the following is provided for the new system:

- A separate power outlet with a 15 amp circuit breaker for 120 VAC for 120 V area, 7.5 amp circuit breaker for 220-240 VAC for 220/240 V area or 15 amp circuit breaker for 100 VAC for Japan.
- Take precautions to ensure that the console is protected from electromagnetic interference.

Precautions include:

- Operate the console at least 5 meters (15 feet) away from motors, typewriters, elevators, and other sources of strong electromagnetic radiation (non-medical grade UPS must be at least 2 meters (6 feet) away from console).
- Operation in an enclosed area (wood, plaster or concrete walls, floors and ceilings) helps prevent electromagnetic interference.
- Special shielding may be required if the console is to be operated in the vicinity of radio broadcast equipment.

Environmental Requirements

The system should be operated, stored, or transported within the parameters outlined below. Either its operational environment must be constantly maintained or the unit must be turned off.

NOTE

You may get an overheating message with regard to fan speed. Ensure adequate system/room ventilation.

Table 3-1 • System Environmental Requirements

	Operational	Storage	Transport (<16hrs.)
Temperature	10° - 35°C 50° - 95°F	-10° - 50°C 14° - 122°F	-10° - 50°C 14° - 122°F
Humidity	30 - 80% non-condensing	30 - 80% non-condensing	30 - 80% non-condensing
Pressure	70 - 106kPa	70 - 106kPa	70 - 106kPa

Console Overview

Key parts of the ultrasound system are identified in this section.

Console Part Identification



CAUTION

For compatibility reasons, use only GE HealthCare approved probes, peripherals or accessories.

DO NOT connect any probes or accessories without approval by GE HealthCare. Failure to follow these instructions could lead to unexpected diagnostic performance.



Figure 3-1 • Ultrasound System

1. Monitor	11. Probe Ports (4)
2. Operator Panel with power on/off button, probe holders, gel warmer	12. Parking probe port (1)
3. Touch Panel with USB 3.0 Ports	13. CW Probe Port
4. Control panel swivel lever	14. Rear Handle
5. Control panel up/down lever	15. Power Cord Wrap Hook
6. Probe Cable Management Hooks (underneath Operator Panel)	16. Peripheral USB 3.0 Ports, Ethernet Connector, S-Video Connector, HDMI Port
7. Probe Holder and Gel warmer	17. Breaker, Ground and Power Cord
8. (Option) DVD Drive, Drawer, Blind cover, B/W printer, V-Nav controller, ECG connector	18. Air Filter (Inside the front cover)
9. Audio speaker	19. Wheel Brake (with each caster)
10. Storage bin (Option)	20. External battery (Option)
	21. Physical A/N keyboard (Option)

**DANGER**

DO NOT touch the patient and any of the connectors on the ultrasound unit simultaneously, including ultrasound probe connectors. DO NOT touch the conducting parts of the USB, Ethernet, Video or Audio cables when connecting equipment to the unit. Failure to follow these instruction could lead to electrical shock.

External drives (CD/DVD Multi Drive, USB Flash Drive, USB HDD)

The external DVD-R/CD-R drive (Optional) is located below the keyboard. Approved USB Hard Disk and USB Flash Drives may also be used for Save As, MPEGVue, and Backup/Restore in the powered system USB ports.

You can use these to perform software upgrades, image archiving, and service diagnostics.

You can use the following media for the multi-drive: DVD-R or CD-R.

USB Drives are an ESD-sensitive device. Only use USB 2.0/3.0 Drives recommended by GE HealthCare.

NOTE

When ejecting a CD-R/DVD-R, you are prompted if you wish to finalize the disk. If you do not finalize the disk, you can add files to the CD-R/DVD-R at a later time. However, you may not be able to view the files stored on this disk with a standard PC. Finalizing a CD-R/DVD-R, allows you to view these files using most standard PCs.

**CAUTION**

The system ONLY supports CD-R/DVD-R and DOES NOT support CD-RW/DVD+R/DVD-RW/DVD+RW or DVD-RAM media. DO NOT use or attempt to format CD-RWs, DVD+R, DVD-RAM, or DVD-RWs.

**CAUTION**

Make sure that the CD/DVD Multi Drive Disk tray is securely retracted during system operation. Mechanical damage may occur if other objects hit the tray.

**CAUTION**

When you put media into the DVD drive, please make sure the media is placed in the right position. If the media is not placed in the right position, the media may be damaged.

Storage areas

Storage areas are available and can be used to store gel, options, probe cables, accessories, etc.

Preparing the System for Use



- 1 Small tray
- 2 Probe Holder

Figure 3-2 • Storage area

Install the TVTR Probe holder (Option)

Push the TVTR Probe holder into the left-side probe holder until it clicks as below graphic.



Figure 3-3 • Install TVTR Probe holder



Figure 3-4 • TVTR Probe Holder

When you remove the TVTR Probe holder from the system, pull the claw and pull out the probe holder.



Figure 3-5 • Pull the claw to remove

Speakers

Audio is provided via internal speakers.

NOTE

You make volume adjustments on the Utility Touch Panel (Master Volume, Effects Volume).

- Audio Doppler operation
- Audio playback of recorded scan sessions
- Audio error notification.

Peripheral/Accessory Connection

Peripheral/Accessory Connector Panel

Ultrasound system peripherals CAN ONLY BE properly connected using the Peripheral/Accessory Connector Panel. Available connectors are two USB 3.0 ports, an HDMI Port, and S-Video and Ethernet connectors.



DANGER

The connection of equipment or transmission networks other than as specified in these instructions can result in electric shock hazard. Alternate connections will require verification of compatibility and conformity to IEC/EN 60601-1 by the installer. Equipment modifications and possible resulting malfunctions and electromagnetic interference are the responsibility of the owner.



CAUTION

For compatibility reasons, use only GE HealthCare approved probes, peripherals or accessories.

DO NOT connect any probes or accessories without approval by GE HealthCare. Failure to follow these instructions could lead to unexpected diagnostic performance.

Table 3-2 • Peripheral/Accessory Connector Panel

	1. USB 3.0 Port 2. RJ-45 Modular, 8-pin Ethernet 3. S-Video Option 4. HDMI Port

LAN Protection Cable (Option)

To prevent claw breakage of the LAN cable, follow the guidelines below.

If the LAN cable is pulled away from the console, the connection (see **(1)** in *Figure 3-6 • LAN Protection Cable* on page 82) will be unplugged first, preventing damage to the LAN connector in the unit.



Figure 3-6 • LAN Protection Cable



Figure 3-7 • LAN Cable with Protection Cable (Example)

Scan Converter Option (S-Video)

The Scan Converter Option enables the ability for the console to view or record scanning via an S-Video Cable.

NOTE

The S-video image displays a portion of the main display that includes the probe image in both single and dual probe display modes. To provide increased resolution and maintain the proper aspect ratio on the S-video display, the entire main display image is not projected on the S-video monitor. The S-video crop area was selected to optimize the probe image so when the console is not in an imaging mode the S-video will continue to display a cropped portion of the screen which may appear incorrect.

External Monitor

An external monitor can be connected to the ultrasound system via the HDMI port, under the following guidelines:

- The display resolution must be set to (1080i) 1920x1080 on the external monitor, matching the main system display. Other resolutions are not supported.
- An isolation transformer must be used to power the external monitor to prevent adverse effects on the ultrasound scanner.
- It is the customer's responsibility to ensure leakage current and grounding is tested and complies with electrical leakage standards.



Figure 3-8 • Connect HDMI Cable to back of ultrasound system and External Monitor

Front Panel Connections and Indicators

There are four sets of connections/indicators on the Front Panel: Hard Disk Drive and Network Status Indicators, CW Probe, Patient Cardiac/ECG Connections, and Volume Navigation (V Nav) Connections/Indicators

Network Status Indicators

There are three front-panel status indicators:

Table 3-3 • Local Drive and Network Status

	1. Local Drive Activity Status 2. Network Activity status 3. Network Activity status

Network and Spooler Status Icons

The following icons identify network and spooler statuses. These appear on the display monitor.

Table 3-4 • Network and Spooler Status Icons

Ethernet Active	Ethernet Error	Ethernet Active Spooler Active	Ethernet Active Spooler Error	Mobile

Mobile Error 	Mobile Bluetooth 	Mobile Wi-Fi 	Spooler Active 	Spooler Active Error 
Spooler Inactive 	Spooler Inactive Error 	Wi-Fi 1 Bar 	Wi-Fi 2 Bars 	Wi-Fi 3 Bars 
Wi-Fi 4 Bars 	Wi-Fi Alert 	Wi-Fi Spooler Active 	Wi-Fi Spooler Error 	Wi-Fi Error 

NOTE

The Mobile icons are used for the Smart Device Apps (Tablet and Mobile Phone).

ECG Connections**Table 3-5 • ECG connection**

	1. Auxiliary Connector 2. ECG Connector 3. PCG Connector

See *ECG* on page 907 for more information.

Table 3-6 • ECG Lead Placement

Lead	Patient Cable Marking		Position on Patient
	AHA	IEC	
I	RA (White)	R (Red)	Right Arm
II	LA (Black)	L (Yellow)	Left Arm
III	LL (Red)	F (Green)	Left Leg

Preparing the System for Use

Volume Navigation Connections/Indicator

Table 3-7 • Volume Navigation Connections

	1. Non-V Nav Inside Probe V Nav Receiver Connector (2 cables) 2. V Nav Needle Tip Tracker/Virtual Tracker Sensor Connector 3. V Nav Active Tracker Connector 4. V Nav Transmitter Indicator 5. V Nav Transmitter Connector

Cable Hook Direction

Cable hook direction can be changed depending on preference.

Tool required: Phillips screwdriver.

NOTE

Be careful not to hit your head against the system.

System without A/N Keyboard

1. Loosen screw indicated with orange circle.
2. Remove screw indicated with red circle and change direction of cable hook.



Figure 3-9 • System Without A/N Keyboard

3. Tighten screw in new position.



Figure 3-10 • Cable Hook Direction

System with A/N Keyboard

1. Loosen screw indicated with orange circle.
2. Remove screw indicated with red circle and change direction of cable hook.



Figure 3-11 • System with A/N Keyboard

3. Tighten screw in new position.



Figure 3-12 • Cable Hook Direction

Powering the System

The section describes how to connect power, start up, log on and shut down the system.

Connect System to Power Supply

**DANGER**

Under no circumstances should the AC power plug be altered, changed, or adapted to a configuration rated less than specified. Never use an extension cord or adapter plug.

To help assure grounding reliability, connect to a “hospital grade” or “hospital only” grounded power outlet.

**DANGER**

Use the appropriate power cord provided by or designated by GE HealthCare. Failure to follow these instructions could lead to exposure to electrical shock.

**WARNING**

Use caution to ensure that the power cable does not disconnect during system use.

If the system is accidentally unplugged, data may be lost.

**WARNING**

To avoid leakage current above safety limits as prescribed by IEC 60601-1 and to ensure continuity of protective earth, DO NOT connect ultrasound system and mains-operated accessories to a single or multiple socket extension cord or power strip.

To connect the system to the electrical supply:

1. Ensure that the wall outlet is of the appropriate type.
2. Ensure that the power switch is turned off.
3. Unwrap the power cable. Make sure to allow sufficient slack in the cable so that the plug is not pulled out of the wall if the system is moved slightly.
4. Attach the power plug to the system (see **b** in *Figure 3-13 • Circuit Breaker (a) and Power Plug (b)* on page 88).



Figure 3-13 • Circuit Breaker (a) and Power Plug (b)

5. Position the power cord latch over the power cord.

6. Push the power plug securely into the wall outlet.

NOTE

Do not use an extension cord or adapter plug except for GE HealthCare approved UPS.

7. Ensure the circuit breaker is on (see **a** in *Figure 3-13 • Circuit Breaker (a) and Power Plug (b)* on page 88). See *Circuit breaker* on page 89 for more information.



Figure 3-14 • Example Plug and Outlet Configurations

8. 100-120 VAC, 1200 VA
Plug and Outlet Configuration
9. 220-240 VAC, 1200 VA
Plug and Outlet Configuration

NOTE

Country-specific power cords are currently available for the United States, Japan, the United Kingdom, Europe, Denmark, Switzerland, Israel, India, Australia, Argentina, South Africa, China and Brazil.

Circuit breaker

The Circuit Breaker is located at the rear panel of the system. On supplies main power to all internal systems. Off removes main power from all internal systems. The circuit breaker automatically shuts off power to the system in case of a power overload.

If a power overload occurs:

1. Turn off all peripheral devices.
2. Reactivate the Circuit Breaker switch.

The Circuit Breaker switch should stay in the On position ("I"); DO NOT hold the switch in the On position. If the Circuit Breaker switch remains On, follow the Power On procedure.

NOTE

If the Circuit Breaker switch does not remain in the On position or trips again:

1. Disconnect the Power Cable.
2. Call for Service immediately.

DO NOT attempt to use the system.

Power On

Press the Power On/Off switch to turn the power on. The circuit breaker must also be in the on position. See *Circuit breaker* on page 89.

Turn on the system

1. Ensure that the unit is properly plugged into an AC outlet of sufficient capacity (120V/10A or 240V/5A).
2. Turn on the breaker at the back of the system (see **a** in *Figure 3-13 • Circuit Breaker (a) and Power Plug (b)* on page 88). At this point, the On/Off switch should be turned off.
3. Momentarily press the On/Off switch. The switch turns on a light.
4. The system should now go through its boot-up process with no further user intervention (approximately 1 to 2 minutes).



Figure 3-15 • Power On/Off Switch Location

Power Up Sequence

The system is initialized with Welcome screen.

During this time:

- The system boots up and the status is reflected on the monitor.
- Probes are initialized for immediate operation.

NOTE

If no probe is connected, the system goes into freeze mode.

- Peripheral devices are activated on power up.

After initialization is complete, controls on the Control Panel backlight and the default B-Mode screen is displayed on the monitor (if a probe is connected).

Password Protection

This section describes login, password protection and logoff procedures for the ultrasound system.

Login Screen

At login, you are notified that “You are accessing a diagnostic medical device that is provided by authorized usage only. Data stored on this device may be subject to various regulations including but not limited to regulations which govern disclosure and privacy of this data. By using this device you are acknowledging that you are authorized to do so and are trained in appropriate use and regulatory guidelines.”

NOTE

You can change the wording that appears on the Login screen. See *Logon Banner* on page 575 for more information.

1. **Operator:** Enter the Operator ID.
2. **Password:** Enter Operator’s password (optional).
3. **Logon or Cancel.**
 - **OK:** Proceed with the logon
 - **Emergency:** Data stored only for the duration of the current exam (EUSR).
4. **Change Password:** Change password, as specified by the Password Policy.

Initial Login

Upon first login to the ultrasound system:

1. The Administrator should log in, typing “ADM” as the Operator Login.

NOTE

No password is needed for the initial administrator log in.

The Touch Keyboard displays when a field is selected. Use the keyboard to enter the Operator ID and Password.



Figure 3-16 • First Admin Login

2. Upon ADM Login, specify the institution's Default Security Level for the ultrasound system: Lowest, Medium, High (Recommended), or Highest), then select **Apply Change**.

NOTE

You may choose **Skip For Now** up to 20 times total to postpone choosing the Security Level. After 20 skips the system requires a Security Level to be chosen.




Figure 3-17 • Default Security Level

Table 3-8 • Security Levels

Security Level	Complexity Rules
Skip For Now	Best for users not responsible for system security setup (manufacturing, vendors or new users browsing the system). Prompt will re-appear at each logon until security level is set for use. This skip can be chosen 20 total times before the system will require a security level to be chosen.
Lowest	Autologon enabled. No password complexity rules.
Medium	Autologon disabled. Minimum Password Requirements: • 8 characters. • Include at least two of the following: one upper case character, one lower case character, one number and/or one special character. • One or more changed character(s) in successive passwords. • No more than four repeated characters. • No more than four sequential characters. • Cannot contain username. • Cannot use the last 10 passwords.

Security Level	Complexity Rules
High (Recommended)	Autologon disabled. Minimum Password Requirements: • 10 characters. • One lower case character. • One upper case character. • One special character. • One number. • Two or more changed characters in successive passwords. • No more than four repeated characters. • No more than four sequential characters. • Cannot contain username. • Cannot use last 15 passwords. • May not be found in a dictionary of common passwords.
Highest	Autologon disabled. Minimum Password Requirements: • 14 characters. • One lower case character. • One upper case character. • One special character. • One number. • Four or more changed characters in successive passwords. • No more than two repeated characters. • No more than two sequential characters. • Cannot contain username. • Cannot use last 25 passwords. • May not be found in a dictionary of common passwords. • Require logon at system startup.

3. If a Security Level is chosen, the Confirm Change screen appears. If *Lowest* or *Medium* Security Level was chosen, you will be prompted to acknowledge that the security setting is lower than GE HealthCare recommends to secure the system. You will not be able to Confirm Change unless the “I Agree” box is selected. If *Medium*, *High* or *Highest* Security Level was chosen, you can choose to change your password immediately and/or to force all users to change their password at next login.




Figure 3-18 • Confirm Change Screen

4. You will be prompted to set up the ADM password, based on the selected Security Level.

NOTE

If the password you type doesn't meet the selected Security Level password complexity rules, the screen will prompt you (in red) to correct the password (see *Figure 3-19 • Change Password* on page 95).




Figure 3-19 • Change Password

5. After logging in, complete the system encryption setup. Navigate to the Disk Encryption Utility, **Utility > Admin > Disk Encryption** page. The system encrypts patient data by default (Encryption On. Disks are unlocked automatically). If you wish to change the default encryption setting, select the desired Encryption Policy, then press Accept.




Figure 3-20 • Initialize System Encryption

NOTE

If you (or the Field Service Engineer) reloads system software, you will need to press **Initialize System Encryption** to encrypt the system and reset the System Encryption password and preferences.

Table 3-9 • Encryption Policy Selections

Encryption Policy	Description
Encryption OFF	Patient Data will not be encrypted. Selecting “OFF” will unencrypt the drive. System drive and recovery partition will remain encrypted.
Encryption ON. Disks are unlocked automatically	System Default. Patient data is encrypted and unlocked at system boot-up. Recovery Key and Password tied to the hard drive.
Encryption ON. Require Pre-Boot PIN/Password before unlocking system drives	The system will not boot until the Pre-Boot PIN or Password has been entered. Unlike other manual key entry configurations, no system functionality is available without the PIN/Password. This encryption policy is intended for high security environments or customers with specific needs. NOTE If the Pre-Boot PIN is enabled and then disabled by choosing auto-unlock or another encryption option, the system must be rebooted and authenticated with the Pre-Boot PIN once more after the change to fully disable the functionality.
Encryption ON. Key is stored on USB/password is entered manually	The system will request the encryption password or recovery key at system startup. The system is not accessible until this password or a disk recovery key is provided. Requiring a password to access the patient archive may prevent emergency usage of the system.


NOTE

If you choose to dismiss this dialog, this dialog will be offered to you again at the next logon.


6. You must set the Encryption Password and record the Recovery Key in order to ensure access to your institution's patient data (required if replacing the system drive, ECB Board, or reformatting the C:\ Drive).

Recovery Keys are not backed up by the system; you must record / archive the Recovery Key in order to retrieve patient data.

1. Reset the Encryption Password by pressing **Change password**.
Reply **No** to this Question: Password is already set on a disk. Do you want to reuse it?
Press "No" to delete existing password.



You can now update the encryption password, then press **OK**.



- Record the Recovery Key by pressing **Show recovery key**, then printing it to a local printer or PACS. Or, save the Recovery Key to a USB Flash Drive by pressing **Save recovery keys**.



You can Show or Hide the Encryption key. Store the Recovery Key in a secure location, accessible to the ADM user as necessary.

Preparing the System for Use

Encryption Notes

The system is usable while it is being encrypted. Encryption can take between 20 minutes and several hours, depending upon system configuration; a status bar tracks encryption progress. It is a background task. You can scan while the system is being encrypted. You can also power the system down and back on while the disk is being encrypted; disk encryption will pick up where it left off in the encryption process.

Changing Your Password

The System Administrator manages system groups, users, and permissions. After you have been added as a valid user, the System Administrator will assign you a temporary password. When you log into the system for the first time, you will be prompted to change your password.

You can change your password at any time when first logging onto the system. To change your password:

1. Type your name in the Operator field.
2. Press the Change Password button. The Change Password pop-up appears.




Figure 3-21 • Password Change

3. Type the following:
 - **Password:** Type your current password.
 - **New Password:** Type your new password.
 - **Confirm Password:** Retype your new password.

NOTE

If you do not wish to set a password, or to change your password, select OK to continue.

Emergency User

If an Emergency User has been enabled by the Administrator via Utility > Admin > Logon, they need to login as "EUSR."



User Rights for the Emergency User are set by the Administrator via Utility > Admin > Groups. Emergency Users do not have access to your institution's patient data, as shown in the example below.

Figure 3-22 • Emergency User




Figure 3-23 • Patient Information Not Accessible

Logoff

To logoff, press the **Power On/Off** switch momentarily and a SYSTEM-EXIT window appears.



Figure 3-24 • System Exit Window

Power Off

For optimum system operation, we recommend that you restart the system at least once every 24-hour period. If you shut down the system at the end of the day, no other action is needed.



WARNING

To avoid losing patient data, ensure that you have properly ended the patient's exam and transferred all exam data / images / clips.

To power off the system:

1. Set the brake and use the operator panel movement controls to lock the control panel in place.



HINTS

If the Park Lock does not engage when you attempt to dock the Operator Panel in place, reset the Park Brake by selecting "Reset Control Panel Park Lock" in the Miscellaneous section on the **Utility > System > General** page.

2. When you shutdown the system, enter the scan screen and lightly press the **Power On/Off** switch at the front of the system once. The SYSTEM-EXIT window is displayed.

NOTE

DO NOT press and hold down the Power On/Off switch to shutdown the system. Instead, lightly press the Power On/Off switch and select Shutdown.

3. Using the trackball, select **Shutdown**.

The shutdown process may take up to two (2) minutes and is completed when the control panel illumination shuts down.

NOTE

If a system shutdown is initiated while the system is still processing an incoming or outgoing DICOM job, a dialog box appears, notifying the user to confirm shutdown, check the spooler status or cancel the shutdown (see *Figure 3-25 • Confirm Shutdown* on page 101).




Figure 3-25 • Confirm Shutdown

4. Disconnect the probes.

Clean or disinfect all probes as necessary. Store them in their shipping cases or another appropriate probe storage system to avoid damage.



WARNING

DO NOT turn off the circuit breaker before the Power On/Off switch LED is white. Data may be lost or system software damaged if the circuit breaker is turned off before the Power On LED is white.

NOTE

When Power Assistant is installed, Power Assistant battery replacement may be necessary if the circuit breaker is turned off for long periods of time (3 to 6 months).

Crash Recovery Instructions

In the event of a system crash, power cycle the system. Upon boot-up, all images and measurements, except for generic worksheets, are preserved in the system. When the system returns, the system alerts you that unsaved images are still in the system from the previous patient. Respond to the prompt to continue the current patient. Check that all images and measurements have been preserved in the system. Then resume the exam.

If the system fails to respond to your commands within a typical period of time, you need to manually reset the system. Simply hold down the power switch to initiate a normal power down sequence. After the system has completely shut down (power switch white), restart the system using the standard power-up sequence. All images and measurements, except for generic worksheets, are preserved in the system. When the system has fully powered up, the system alerts you that unsaved images are still in the system from the previous patient. Respond to the prompt to continue the current patient. Check that all images and measurements have been preserved in the system. If you do not have any images on the clipboard, the patient must be retrieved from the database. Then resume the exam.

System Language Configuration

This section describes how to set the system language and the keyboard language.

Select the System Language and Date/Time Format

The default operating system language and keyboard may be changed from the **Utility > System > General** page.



Figure 3-26 • Utility System General Screen

1. Select the desired system language from the **Language** dropdown menu (1). Then select **Save** at the bottom left of the screen. Do NOT restart yet.
2. Select **Regional Options** (2) to open the Regional Options Dialogue Box



Figure 3-27 • Regional Options Dialogue Box

3. Select the following Regional Options:
 - **Current OS Language** - Select the language from the dropdown menu (3) to match the system language selected in Step 1.
 - **Keyboard Preference** - Select the keyboard language preference from the dropdown menu (4).
 - **Selectable Keyboard** - Selectable Keyboards allows users to install/uninstall keyboards. The installed keyboards can be changed with Alt+Shift or Ctrl+Shift combination keys (5).
 - **Current Format** - Select the date and time format preference from the dropdown menu (6).
4. After making changes to the Regional Options, select **OK**; when prompted to restart, choose **OK** again to restart system.

NOTE

For the United Kingdom, it is recommended to use ENG Language (1), English (United States) OS Language (3), and modify "Current Format" time/date settings to English (United Kingdom) (4) (see *Figure 3-27 • Regional Options Dialogue Box* on page 103).

NOTE

For Greek, Russian, Chinese, Hebrew and Japanese, the Regional Options language must match the Current OS Language to prevent possible character corruption.

Change the Keyboard Language (Temporary)

To temporarily change the keyboard language, press the left Alt+Right-Shift or Alt+Left-Shift keys on the keyboard to cycle forward or backward through all available keyboard languages until you reach the desired language.

NOTE

When the system is restarted, the keyboard language will return to the keyboard language last set in Regional Options.

English (International) Keyboard

When the system language is set to English (United States), pressing the Ctrl+Shift on the keyboard toggles between the English (United States) keyboard and the English (International) keyboard.

Apostrophe/Quotation Marks

To type an apostrophe or quotation marks while using the English (International) keyboard, you must press the space key after the apostrophe or quotation mark to display the character.

Special Accented Characters on the English (International) Keyboard

To type special accented characters, from the English (United States) keyboard, press Ctrl+Shift to change to the English (International) keyboard, then press the AltGr+appropriate keyboard key listed below.

- [AltGr]+[Q] to type ä
- [AltGr]+[W] to type å
- [AltGr]+[E] to type é
- [AltGr]+[Y] to type ü
- [AltGr]+[I] to type ì
- [AltGr]+[P] to type ö
- [AltGr]+[A] to type á
- [AltGr] + [Function Key] + [,] to type ç
- [Shift] + [6] + [a] to type â
- [Shift] + [6] + [e] to type ê
- [Shift] + [6] + [i] to type î
- [Shift] + [6] + [o] to type ô
- [Shift] + [~] + [a] to type ã
- [Shift] + [~] + [n] to type ñ
- [Shift] + ["] + [a] to type ä
- [Shift] + ["] + [e] to type ë
- [Shift] + ["] + [i] to type ï
- [Shift] + ["] + [o] to type ö

Preset Restore User Interface Corruption

When restoring presets with the user interface in Russian, Greek, Japanese or Simplified Chinese, if the OS Region Language does not match the selected system software application language at the time of the preset restore, the user interface text may appear corrupted.

To prevent the corruption, before restoring presets, confirm that the OS Region language and system software application language match by following the procedure below, *Match the OS Language with System Software Application Language* on page 105, referring to *Figure 3-28 • Match OS Language with System Software Application Language (Preventative)* on page 105.

If the corruption has already occurred and the text on the screen is not readable, follow the same procedure (*Match the OS Language with System Software Application Language* on page 105) to correct the settings, referring to *Preset Restore User Interface Corruption* on page 104, which shows the location of the Regional Options and Language fields on the corrupted screens.

Match the OS Language with System Software Application Language

Change the default operating system language on the Utility > System > General page.

1. Select **Regional Options** to open the Regional Options dialogue box (1).
2. Select **Current OS Language** and select the OS language from the dropdown menu to match the system language (2).
3. Select **OK** (3). The system will restart.



Figure 3-28 • Match OS Language with System Software Application Language (Preventative)

Preparing the System for Use



Figure 3-29 • Match OS Language with System Software Application Language (Corrective)

Probe Connections

This section describes how to connect and select probes on the ultrasound system.

Concept



CAUTION

If the packaging has been:

- Damaged.
- Unintentionally opened before use.
- Exposed to environmental conditions outside of those specified.

Contact your GE HealthCare representative.

Connecting the Probe



DANGER

Fault conditions can result in electric shock hazard to user/patient. Do not touch the surface of probe connectors which are exposed when the probe is removed. Do not touch the patient when connecting or disconnecting a probe.



CAUTION

Inspect the probe before and after each use for damage or degradation to the housing, strain relief, lens, seal, cable and connector. **DO NOT** use a transducer which appears damaged until functional and safe performance is verified. A thorough inspection should be conducted during the cleaning process.

Probes can be connected at any time, regardless of whether the console is powered on or off. To ensure that the ports are not active, place the system in the image freeze condition.

To connect a probe:

1. Place the probe's carrying case on a stable surface and open the case.
2. Carefully remove the probe and unwrap the probe cord.
3. **DO NOT** allow the probe head to hang free. Impact to the probe head could result in irreparable damage. Use the integrated cable management hook to wrap the cord.

Preparing the System for Use



Figure 3-30 • 4 DLP Probe Port Connectors

4. Prior to inserting the probe, ensure that the connector locking handle is positioned to the left.
5. Align the connector with the probe port and carefully push into place.

NOTE

When connecting the probe, DO NOT turn the locking lever if resistance is felt. If this is the case, remove the probe connector and check for misaligned or damaged pins. Probes with damaged connector pins should be taken out of service immediately.

6. Turn the connector locking handle to the right to secure the probe connector.
7. Carefully position the probe cord so it is free to move and is not resting on the floor.

Probe Port Illuminator

The ultrasound system probe ports are illuminated to indicate the probe port locations in a dark room.

The illumination brightness can be adjusted on the Utility page.



Figure 3-31 • Probe port with illumination





Figure 3-32 • Touch panel control

Cable Handling

Take the following precautions with probe cables:

- Keep free from wheels. Use the cable hooks located below the operator panel.
- Do not bend the cable acutely
- Avoid crossing cables between probes.

Probe Management



- 1 Active probe port
- 2 Parking probe port (Extra probe can be stored by connecting to this port)
- 3 Pencil probe port

Figure 3-33 • Probe port

Selecting Probes

To activate the probe, select the appropriate probe from the probe indicators on the Touch Panel.

1. Select the probe that provides optimum focal depths and penetration for the patient size and exam.
2. Select the exam application.
3. Select the application preset.
4. Begin the scan session using the default Power Output setting for the probe and exam.



Figure 3-34 • Probe Application/Preset Touch Panel

The probe's default settings for the mode and selected exam are used automatically.

NOTE

Selecting a new probe unfreezes the image.

Deactivating the Probe

When deactivating the probe, the probe is automatically placed in freeze mode.

To deactivate a probe:

1. Ensure the selected probe is deselected or the ultrasound system is in freeze mode. If necessary, press the **Freeze** key.
2. Gently wipe the excess gel from the face of the probe.

3. Carefully slide the probe around the right side of the keyboard, toward the probe holder. Ensure that the probe is placed gently in the probe holder.

Disconnecting the Probe



CAUTION

DO NOT allow the probe head to hang free. Impact to the probe head could result in irreparable damage. Use the integrated cable management hook to wrap the cord.

Probes can be disconnected at any time. However, the probe should not be active when disconnecting the probe.

1. Ensure the probe is deactivated. Deactivate by selecting another probe or pressing Freeze.
2. Move the probe locking handle to the left.
3. Pull the probe connector straight out of the probe port carefully.
4. Carefully slide the probe and connector away from the probe port and around the right side of the keyboard.
5. Ensure the cable is free.
6. Be sure that the probe head is clean before placing the probe in its storage box or wall hanging unit.

Storing the Probe

It is recommended that all probes be stored in the provided carrying case or in the wall rack designed for probe storage.

When storing a probe in the carrying case:

1. Place the probe connector into the carrying case.
2. Carefully wind the cable into the carrying case.
3. Carefully place the probe head into the carrying case. DO NOT use excessive force or impact the probe head.

Storage/Transportation



DANGER

Placing a dirty or contaminated probe in a carrying case or shipping carton will contaminate the foam insert. Failure to follow proper cleaning guidelines could lead to patient exposure to contaminant.

Each probe should be supported in its own probe holder on the console. If a carrying case is provided with the probe, always use the carrying case to transport the probe from one site to another.

Preparing the System for Use

Secure the probe in its holder for moving short distances.

When transporting a probe a long distance, store it in its carrying case.

If possible, use a rigid container with a lid that secures the probe's connector in place so as not to damage the probe head or lens. Place a soft cloth in the bottom of the container to prevent movement during transport.

Operator Controls

This section describes the ultrasound system Control Panel, Keyboard and Touch Panel controls.

Control Panel Map

Controls are grouped together by function for ease of use.



1. Probe Holder and Cable Management
2. Touch Panel and Joystick controls
3. On Screen Keyboard (not display in this graphic)
4. User Defined Keys. See for detail.
5. Mode/Gain/XYZ (3D) Controls
6. Trackball, Trackball Keys, Pointer, Measure, Comment, Body Pattern, Clear, Zoom, Programmable Keys

7. L/R, Start/Stop, Freeze, Simultaneous
8. Steer/Width/Depth/Reverse
9. Auto, CF/PW Auto Positioning
10. P1, P2

Figure 3-35 • Control Panel

Control Panel Adjustment



CAUTION

To avoid injury or damage, make sure nothing is within the range of motion before moving the control panel. This includes both objects and people.

Ensure that the hands of the patient are away from the Control panel arm when moving the Control panel.

To Raise/Lower the Control Panel

1. Hold the front handle in two hands.
2. Push and hold down the Up/Down button next to the right front handle.
3. Raise or lower the control panel.
4. Release the Up/Down button at the desired height.



Figure 3-36 • Up/Down control button

To Swivel the Control Panel

1. Hold the front handle in two hands.
2. Push and hold down the swivel button next to the left front handle.
3. Move the control panel to the left or the right.
4. Release the swivel button at the desired position.



Figure 3-37 • Swivel control button

Keyboard

On Screen Keyboard

You can use "On Screen Keyboard" on touchscreen. Keyboard will show up when you press "Keyboard" User Defined key. And you can hide it with "Exit" button on Keyboard or Keyboard UD key.



Figure 3-38 • On Screen Keyboard

1. Language Indicator key; This key simulates [Alt] [Shift] key to change the input language.



Figure 3-39 • Keyboard User Defined key

Physical A/N keyboard (Option)

Physical A/N keyboard is under the control panel.



Figure 3-40 • Physical A/N keyboard

Push the keyboard to project forward.



Figure 3-41 • Push the keyboard

Functional keys

The standard alpha-numeric keyboard has some special functions.

Table 3-10 • Special Key Function

Keyboard key	Function
Esc	Exit current display screen.
F1	Help Access Online help/user manual.
F2	Arrow Annotation Arrow.

Keyboard key	Function
F3	Eject Eject media.
F4	Spooler Activates DICOM Job Spooler screen.
F5	Creates a Fast Key.
F6	Plays a Fast Key.
F7	Home/Set Home Move annotation cursor to home position; shift+key to set current annotation cursor position as the new home position.
F8	Text1/Text2 Switch between user text annotation overlays.
F9	Grab Last Activate the last selected data for edit.
F10	Word delete Erase word associated with comment cursor.

If you encounter a problem and cannot collect the logs immediately:

Table 3-11 • Key for Collecting the Log

Keyboard key	Function
Alt+1 or Alt+2	Place a marker in the log.
Alt+D	Collect the logs.

Once the logs are collected, the engineering team would be able to see the marker you added which will help engineering to troubleshoot the problem.

Touch Panel

The Touch Panel contains exam function and mode/function specific controls.

Exam Function Controls



1. Patient: Enters Patient screen
2. Scan: Enters scanning mode screen
3. Reports: Activates default report and Touch Panel of report choices.
4. End Exam: Activates Image Management and Touch Panel with end of exam options.
5. Utility: Activates system configuration menus. The Utility+ screen defaults to Utility screen, unless an app is installed on the system. If one more apps are installed, the Utility+ screen defaults to the Apps screen (App Launchpad).
6. Probe Indicator: Indicates and selects the probes.
7. TGC Control: Activates TGC function.
8. Indicates the number of pages for this mode. To move to the next page, touch the “dot” or swish your hand from right to left/left to right.
9. To view all/fewer of the controls for this Touch Panel, press this Research/Clinical button.

Figure 3-42 • Exam Function Controls

NOTE

Different menus are displayed depending on which Touch Panel Tab is selected.

At the bottom of the Touch Panel, there are six combination rotary dials/push buttons. The functionality of these rotaries changes, depending upon the currently-displayed menu. Press the button to switch between controls, or rotate the dial to adjust the value, or move the control left/right or up/down to adjust the value.

Mode/Function Specific Controls

In general, the key name is indicated at the top of the key. There are different types of Touch Panel keys as illustrated below:



1. Press to toggle control on/off.
2. Progress/Select keys are used for controls that have three or more choices.
3. Progression keys are used to assess the impact of the control on the image progressively.
4. 3-way functionality knobs (below the Touch Panel): Adjust controls by pressing (dot symbol), rotate (circled arrow symbol), move up/down (vertical line with arrows) or left/right (horizontal line with arrows).

Figure 3-43 • Mode/Function Specific Controls

Touch Panel or Operator Panel Reset

If an issue arises with Touch Panel or Operator Panel functionality, the functionality can be reset on either the Touch Panel or the Operator Panel.

Touch Panel Functionality

If the Touch Panel touch input becomes unresponsive or is not functioning correctly, it can be reset in two locations:

1. **Configure Touch Panel from the Main Display:** To reset touch screen functionality to the Touch Panel, at Utility > System > General screen, select **Configure Touch Panel** to launch a Windows utility with instructions on how to assign the touch panel as the touch screen (see #1 in Figure 3-44 • Main Display - Configure Touch Panel on page 120).




Figure 3-44 • Main Display - Configure Touch Panel

2. **On the Touch Panel:** At Utility > Touch Panel, swipe to the second Touch Panel page and select **Reset Touch Panel** to reset control on the Touch Panel (see #1 in Figure 3-45 • Touch Panel - Reset Touch Panel or Operator Panel on page 120).

NOTE

If the Touch Panel control does not work to select **Reset Touch Panel**, use the trackball to select with the mouse pointer.




Figure 3-45 • Touch Panel - Reset Touch Panel or Operator Panel

Operator Panel Functionality

If the Operator Panel is not functioning correctly (slow or no response to key input), it can be reset on the Touch Panel.

On the Touch Panel: At Utility > Touch Panel, swipe to the second Touch Panel page and select **Reset Operator Panel** to reset the Operator Panel and restore normal functionality (see #2 in *Figure 3-45 • Touch Panel - Reset Touch Panel or Operator Panel* on page 120).

Mode, Display and Print

This group of controls provides various functions relating to the display mode, display orientation, image recording/saving, freeze, gain and Cine scroll.

The Mode Controls select the desired display mode or combinations of display modes.

- During dual display modes the L and R keys activate the Left or Right displayed image. See *Split Screen* on page 242 for more information.
- Auto is used to:
 - initiate auto optimize
 - turn off auto optimize.
- Depth controls the image display depth.
- The Reverse key (via Depth key if preset) toggles the left/right orientation of the scan image.
- Print keys are used to activate/print the designated recording device.
- The Freeze key is used to stop the acquisition of ultrasound data and freeze the image in system memory. Pressing **Freeze** a second time continues live image data acquisition.
- To activate a specific mode, press the appropriate mode key.

Each mode has its own gain control via the larger gray knob surrounding the mode key.

Measurement and Annotation

This group of controls performs various functions related to making measurements, annotating and adjusting the image information.

- The Comment key enables the image text editor and displays the annotation library Touch Panel.
- The Clear key is generally used to erase functions, such as annotations/comments, body patterns and measurements. Pressing the Clear key again exits the selected function.
- The Body Pattern/Ellipse control has a dual purpose:
 - Press the Body Pattern/Ellipse control, it enables the Body Pattern Touch Panel and displays the default pattern on the screen. When body patterns are active, the knob rotates the probe position indicator.
 - Rotate the Body Pattern/Ellipse control, it activates the ellipse measurement function after the first distance measurement has been set and the second caliper is activated.

Press Set to fix the measurement after the ellipse adjustment is complete. The measurement is then displayed in the measurement result window.

- The Measure key is used in all types of basic measurements. When the Measure key is pressed, the measurement Touch Panel is displayed.

Preparing the System for Use

- The Set key, located on the Trackball on-screen controls, is used for various functions, but is generally used to fix or finish an operation (e.g. to fix a measurement caliper).
- The Trackball is used with almost every key function in this group. Trackball control depends on the last key function pressed.

Monitor

Moving the Wide Screen Monitor

Adjusting the Monitor Position


CAUTION

If any defects or damages are observed on the display screen or monitor itself, do not operate the equipment but inform a qualified service person.

Contact a Service Representative for information.

Grab the frame of the monitor with both hands when you adjust the position of the monitor and monitor arm.

Table 3-12 • How to grab

LCD Display	
Up/Down/Left/Right 	Flip down/Flip up 
HDU Display	
Up/Down/Left/Right 	Flip down/Flip up 



CAUTION

DO NOT apply force to the monitor surface by finger or other hard object.

The monitor surface may be damaged and permanently discolored.

- LCD Display



- HDU Display



Locking/Unlocking the Monitor

1. Turn the release knob clockwise to unlock the monitor. The monitor can be moved freely in all directions (see 1 in *Figure 3-46 • Unlock/Lock the monitor arm* on page 124).
2. Turn the release knob clockwise to raise the lock and move the monitor into the parked position, then turn the lock counterclockwise to lock. (see 2 in *Figure 3-46 • Unlock/Lock the monitor arm* on page 124).



1 Unlocked

2 Locked

Figure 3-46 • Unlock/Lock the monitor arm

Adjusting the Monitor

You make adjustments to all monitors via the second Utility Menu page. To view the monitor while making these adjustments

1. Press **Utility** and select the second page.
2. Press **Scan Screen**. The scan screen displays on the main display while the Utility Touch Panel is active so you can see the affect your adjustments are making to the monitor.

Table 3-13 • Monitor Adjustments

Monitor Adjustments	Parameters	Affect on the Monitor
Room Profile: Monitor luminance varies, depending on the room's profile (light, dark, semi-dark). Adjusting this control helps to adapt the monitor to different room lighting conditions.	Set the Room Profile to match the room:	Brightness/Contrast Equivalent:
	LCD display	
	Dark	50/85
	Semi-Dark	70/85
	Light	90/85
	HDU display	
	Dark	50/40
	Semi-Dark	50/65
	Light	50/100
	User Defined	If you select User Defined, adjust the Brightness/Contrast control at the bottom of the Touch Panel.
	Note: Brightness/Contrast setting is changeable when the Room Profile setting is available and set to "User Defined".	
Color Profile: Color Profile controls color temperature, or overall tint of the monitor.	0	6500K
	1	9000K (LCD display) / 9300K (HDU display)
	2	11000K
	3	13000K

Preparing the System for Use

Monitor Adjustments	Parameters	Affect on the Monitor
Gamma Setting: With Grayscale Standard Display Function (GSDF) enabled, the gamma button on the Touch Panel affects the image, emulating the appearance of that gamma. The system gamma is adjusted to match the GSDF gamma compensation to the new monitor gamma.	Set the Gamma to 2.2 or 2.4	Note: With GSDF disabled, the monitor still uses a gamma curve that may be selected on the Touch Panel (for backwards compatibility with sites that are happy with their PACS or may have a mix of older systems).
Return to Default settings.	Press “Reset Monitor”, then restart the system.	

Wide Screen Monitor Display Settings

Further clarification on three settings on the Utility > System > System Display configuration page is provided below:

- Image Size (Default, Large)
- Image Display Area (Default, Large, Extra Large)
- Use Wide Screen For... (On, Off, Auto)

Table 3-14 • Wide Screen Monitor Display Settings

Preset Setting	Description	Choices
Image Size:	The parameter “Image Size” changes the size of the ultrasound image without changing the screen image, which includes the title bar and the image parameter window. It also does not change the size of comments, measurement cursors and lines, or the measurement result box. With all other settings the same, increasing Image Size from Default to Large creates a larger ultrasound image both on the ultrasound system display and on PACS’ displays. It will not affect the appearance of screen image items like the title bar and the image parameter window.	• Default • Large

Preset Setting	Description	Choices
	 Image Size = Default (Left) and Image Size = Large (Right) Hint: If you select and save “Last Used” on this page, the system starts with last used Room Profile setting every time.	
Image Display Area	“Image Display Area” changes the size of the screen image. On the ultrasound system display, this parameter changes the size of the ultrasound image. It also changes the arrangement of the items like the title bar and image parameter window, but not their individual size. Comments, measurement cursors and lines, and the measurement result box also do not change. This parameter also changes the pixel count of the image. Many PACS software packages scale all images of the same aspect ratio to the same viewing size, regardless of their pixel count. Because the ultrasound image displayed is enlarged as the pixel count is increased, the PACS display size of the ultrasound image does not change. However, items like the title bar, image parameter window, comments, and measurements do not get larger as the pixel count is increased, so they appear smaller on PACS. If text appears too small on PACS, change the Image Display Area to Default (decreasing the screen image pixel count).	• Default • Large • Extra Large

Preparing the System for Use

Preset Setting	Description	Choices
	 Ultrasound System Image Display Area: Left = Default, Right = Large  PACS Image Display Area: Left = Default, Right = Large	
Use Wide Screen For...	“Use Wide Screen for...” determines when the system will use a wide screen aspect ratio (16:9) as opposed to a “standard” aspect ratio (4:3). If “Single Image” under this heading is set to On, then the 16:9 aspect ratio is always used. A third aspect ratio is used if Extra Large is selected under Image Display Area. In this case the Wide Screen display options are not used and a 16:10 aspect ratio is used for all images. The 16:10 aspect ratio will be only modestly different than the 16:9 aspect ratio. Note: PACS display the top and bottom of the image as black bars, just like watching a wide-format movie on a non-wide screen TV.	• On • Off • Auto

Preset Setting	Description	Choices
	 Ultrasound System Left&Right: Image Size=Large, Image Display Area=Default; Right: Wide Screen=On  PACS Left&Right: Image Size=Large, Image Display Area=Default; Right: Wide Screen=On	

Image Display Area Preset Setting Notes

The following table shows pixel counts (columns × rows) for saved images for all combinations of Image Display Area and Wide Screen on/off.

Table 3-15 • Wide Screen Single Image and CINE Clip Pixel Count

Image Display Area Setting	Single Image	Single Image	CINE Clip*	CINE Clip*	Stress Clip
	On	Off	On*	Off*	
Default	1456 x 819	1092 x 819	1346 x 748	982 x 748	1092 x 819
Large	1552 x 873	1164 x 873	1442 x 802	1054 x 802	1164 x 873
Extra Large	1552 x 970	1552 x 970	1442 x 899	1442 x 899	1552 x 970
*Under CINE Clip, “On” and “Off” mean Wide Screen Single Image On/Off, NOT CINE Clip On/Off.					

NOTE

Note the different pixel counts for single frames and cine clips. This is because saved cine clips cut off the title bar and image parameter window.

Use Wide Screen For... Preset Setting Notes

The following table shows the aspect ratio for saved images for all combinations of Image Display Area and Wide Screen on/off. For comparison, it also includes data for the image

Preparing the System for Use

display on 19" monitors. For CINE clips, the aspect ratios will not be exactly the same as for single frames, but they will be very close.

Table 3-16 • Wide Screen Aspect Ratio, Single Image

Image Display Area Setting	On	Off
Default	16:9	4:3
Large	16:9	4:3
Extra Large	16:10	16:10

Test Patterns



The ultrasound system monitors follow the DICOM Gray Scale Display Function (GSDF) standard* which uses a GSDF curve to evaluate the health and effectiveness of the monitor. The GSDF standard is what is commonly used for calibrating PACS' monitors. This setting may help to make the image appearance more uniform between the ultrasound system and PACS.

NOTE

The monitor is GSDF-compliant with all contrast settings, but not all brightness settings.

There is a GSDF on/off switch, "Enable DICOM grayscale display mode (GSDF)" via the Utility > System > System Display page that allows you to adjust the monitor you are using to the PACS.

Table 3-17 • Test Patterns

Test Pattern	Description	Screen Example
Gray Bars	Confirm that you can clearly see all of the different gradations of the grayscale in each gray cell.	
Color Bars	Confirm that you can clearly see all of the different colors in each colored cell.	
Resolution	Confirm that this image is clear and crisp, without jagged edges or lines.	

Test Pattern	Description	Screen Example
Brightness Contrast	Before making any adjustments, record the settings for contrast and brightness. In a dimly lit room, adjust the monitor to Brightness 50 & Contrast 40. Increase the contrast until the left most block in the second row from the bottom is just visible. All the remaining blocks in the last 2 rows of the image should now be visible. Reset the contrast and brightness to the recorded levels.	
LN1 to LN18	Luminance for different grayscale levels. Usually, the facility's Biomed Engineer tests each gray level map on the monitor from LN-1 to LN-18 with a light meter and then graphs these on a scale to ensure that the curve falls within the GSDF's standard. Generally, this is performed yearly.	LN-1  LN-18 
QC	Quality Control. Confirm that this image is clear and crisp, without jagged edges or lines.	
CT	Contrast. Confirm that you can you detect the circles within the square grayscale boxes.	
***Assessment of Display Performance for Medical Imaging Systems." Report of the American Association of Physicists in Medicine (AAPM) Task Group 18, Medical Physics Publishing, Madison, WI, AAPM On-Line Report No. 03, April 2005, Samei E, Badano A, Chakraborty D, Compton K, Cornelius C, Corrigan K, Flynn MJ, Hemminger B, Hanciandreu N, Johnson J, Moxley M, Pavlicek W, Roehrig H, Rutz L, Shepard J, Uzenoff R, Wang J, Willis C. *Also see https://www.aapm.org/pubs/reports/OR_03.pdf, section 4.3.		

Monitor Display



1. System Date and Time (Note: The date on the monitor may truncate the century when using the YYYY-MM-DD date format.)
2. Caps Lock
3. Network Connection Indicators (Wireless LAN, Wired, Mobile Bluetooth), Battery Status
4. DVR Status
5. InSite Indicator
6. Image Clipboard
7. Image Preview
8. Worksheet/Direct Report
9. Gray/Color Bar
10. Institution/Hospital Name, Date, Time
11. Operator Identification, Patient Name
12. Probe Orientation Marker
13. Measurement Calipers and System Message Bar
14. Color Window
15. Power Output Readout
16. Probe Identifier, Exam Preset
17. Live/Freeze Indicator
18. Imaging Parameters by Mode, Speed of Sound (SoS) Indicator (if applicable)
19. Focus Indicator and Color Flow Focal Zone Marker
20. Depth Scale
21. Image Management Icons and LOGIQ Apps Icon (not shown)
22. Body Pattern
23. CINE Gauge
24. Trackball Controls and Status
25. 6Tc-RS Probe temperature display
26. 6Tc-RS Probe angle display

Figure 3-47 • Monitor Display Tour

Live/Freeze Indicator

Live/Freeze Indicator shows active image scan status with Single/Dual/Quad format and has four choices (Live/Not Live/Update/B Pause).

Live/Freeze Indicator display (on/off) is configurable on Utility > System > System Display menu.

Table 3-18 • Live/Freeze Indicator

Live		Update	 
Not Live		B Pause	 

Monitor Display Layout

Monitor Display Layout is configurable on Utility > System > System Display menu.

See *System/System Display Preset Menu* on page 459 for more information.





Figure 3-48 • Default Image Format (4:3)

Preparing the System for Use

1. Title bar
2. Information window 1
3. Information window 2
4. Scan Area
5. Control Window
6. Preview window/User Label
7. Clipboard
8. Trackball mapping and Set keys
9. Status area

Information Window

Measurement Summary, Side Clipboard, Scan Assistant Guide, and My Desktop are displayed options for Info1 or Info2 window.



- 1 Measurement Summary
- 2 Side Clipboard
- 3 Scan Assistant Guide
- 4 My Desktop

Figure 3-49 • Information Window - Example

My Desktop

The user can import graphics (jpeg) and write a caption for that graphic to display on My Desktop in Utility > User Specific.

Table 3-19 • User Specific

Preset Parameter	Description
My Desktop Title	Free text
Picture 1 and 2	Browse to load jpeg and type the caption.
Background Color	Select Color: Select the background color. Default Color: Reset the background color to factory default.
Import	Import the graphic.

User Label

In Utility > Application > Settings > Label Area, you can enter User Label which displays in the Preview Window as a brief note.

For example, P1 - Stills only, P2 - Stills, Clips (3 sec.), P3 - Volume, P4 Screen Capture.



Figure 3-50 • Application Settings Preset Menu

Table 3-20 • Label Area

Preset Parameter	Description
Show labels	If selected, the system displays the User Label in the Preview Window at the left-bottom of the monitor.
Label 1 - 8	User labels have eight text lines. Each User Label is limited to 50 characters.




Change the Display Image Area

You can change the screen format layout by pressing the lower, right-hand corner of the Display Image Area icon on the display:


Table 3-21 • Display Image Area Selections

Display Image Area icons 	Click Image Display Area icon to toggle current display area and extra large area. Default Layout (4:3 Layout) Standard (16:9 Layout) Full Wide Large Image Layout (4:3 Layout) Standard (16:9 Layout) Full Wide Extra Large Layout (16:10 Layout) Partial Wide

Table 3-22 • Display Image Area Examples

Default Layout Standard	
Default Layout Full Wide	
Large Image Layout Standard	
Large Image Layout Full Wide	

Extra Large Layout Partial Wide

**NOTE**

Even If there are buttons on the rear cover of the monitor, DO NOT use those buttons for adjustment. The ultrasound system will override the parameter(s) and the monitor does not have capability to retain parameters.

Monitor Display Image Management Controls

You can manage images from the display via these on-display controls.

Table 3-23 • Image Management Icons

	Image Display Area Press Image Display Area to select Image Display Area - Default/Large/Extra large. See <i>Change the Display Image Area</i> on page 135 for more information.
	Worksheet Activate the Worksheet.
	Active Images Press Active Images to go to the Patient Active Images page.
	Compare Assistant Press to activate Compare Assistant.
	Activate Save As Menu - Note: Save As Menu icon is displayed only when an image is recalled or scan state is frozen/CINE Loop.
	Delete Recalled Image/Delete Last Image - Delete Recalled Image is displayed only when one image is recalled. / Delete last image icon is displayed only when the image is not recalled and there are non-stored images. You can use this to delete an image from the clipboard. Place the cursor on the clipboard image you want to delete, then press Set to select the image. Then place the cursor on the Delete icon and press Set.
	Next/Previous Image(s) and Clipboard Slide Show Press the left arrow to move to the previous image; press the right arrow to move to the next image. Clipboard Slide Show The Clipboard Slide Show plays all images on the clipboard and wraps around the ends. To activate, press and hold [Ctrl] + [Previous Arrow] or [Ctrl] + [Next Arrow]. Each image recalls for three seconds, or the length of the loop, whichever is longer. You can manually skip to a new image during the slide show by recalling it, as usual. To end the slide show manually, press [Ctrl] + [Previous]/[Next] again. Slide Show ends when you go to live scanning, or if the clipboard is not shown when it's time for the next image to load.

System Positioning and Transporting

Moving the System

When moving or transporting the system, follow the precautions below to ensure the maximum safety for personnel, the system, and other equipment.



CAUTION

This equipment is not to be used during transportation (e.g. ambulance cars, aircraft).



CAUTION

Handle carefully. A drop of more than 5 cm can cause mechanical damages.



MOVING HAZARD

Never move the system with locked wheels.

Before Moving the System

1. Press the **Power On/Off** switch to power off the system. See *Power Off* on page 100 for more information.
2. Unplug the power cord.
3. Wind the power cable around the cable hook below the rear handle.



DANGER

LOOSELY (not twisting tightly) wrap the Power Cord around the cable hooks below the rear handle. Occasionally wrap the Power Cord around the cable hooks in the opposite direction.

NEVER allow the Power Cord to drag on the floor.

NEVER roll over the Power Cord with the wheels.

Failure to follow these instructions could lead to exposure to electrical shock.

4. All cables from off-board peripheral devices (external Color Digital/Report printer, etc.) and the ethernet connection must be disconnected from the console.
5. Ensure that no loose items are left on the console.
6. Connect all probes to be used while off site. Ensure that probe cables are out of the way from the wheels and not protruding beyond the console. Use the probe management hooks located below the Operator Panel to further secure the probe cables.

NOTE

If more than four (4) probes are intended to be used, store the additional probes securely.

7. Store all other probes in their original cases or in soft cloth or foam to prevent damage. In addition, on-board storage bins are available as a system option.
8. Store sufficient gel and other essential accessories in the provided space.
9. Adjust the monitor and control panel to their lowest positions by using the up/down switch on the front of the operator panel. Make sure the operator panel is locked in place.



CAUTION

To prevent system damage while not in use AND/OR before moving the system, flip down the monitor and lock the monitor arm and operator panel firmly in place.



Figure 3-51 • Flip down the monitor and lock the monitor arm

10. Unlock the wheels.

When moving the system

1. Always use the rear handle grips to move the system.



CAUTION

- **DO NOT** attempt to move the console using any cables or fixtures, such as the probe connectors.
- **DO NOT** attempt to move the system with the monitor by pulling cables or belts placed around the monitor and/or monitor arm.
- **DO NOT** restrain the ultrasound system at the monitor or the monitor neck using a belt. Always secure the monitor at the body part of the console.

2. Take extra care when moving the system long distances and on inclines. Ask for help if necessary.

Avoid ramps that are steeper than ten degrees to avoid tipping over the system. Utilize additional care and personnel when moving on a steep incline (>5 degrees) or loading the system into a vehicle for transport.

NOTE

Wheel chair ramps are usually less than five degrees.

3. Use the foot brake (pedal), located on the bottom of the system in the front, when necessary.
4. Do not let the system strike walls or door frames.
5. Use extra care when crossing door or elevator thresholds.
6. Once the destination is reached, lock the wheels.

**WARNING**

To avoid possible injury and equipment damage:

- Be sure the pathway is clear.
- Limit movement to a slow careful walk.
- Use two or more persons to move the system on inclines or long distances.

Failure to follow instructions could lead to possible injury and/or equipment damage.

Refer to *Adjusting the Monitor Position* on page 123 for instructions on how to lower the monitor.

Wheels

Examine the wheels frequently for any obvious defects that could cause them to break or bind. Each wheel has an independent brake pedal. A left rear wheel also has a swivel lock.



1. Brake pedal

2. Swivel lock

Figure 3-52 • Wheel lock and Swivel lock

**CAUTION**

If you park the system on a slippery slope, you **MUST** use the brakes on the wheel.

Preparing the System for Use

Brake Pedal



Figure 3-53 • Front and right-rear caster

1. Step on Lower side pedal to activate Brake
2. Step on Upper side pedal to release Brake



Figure 3-54 • Left-rear caster

3. Step on to activate Brake
4. Raise up to release Brake

Swivel Lock



Figure 3-55 • Rear-left with Swivel lock

1. Step on Lower side pedal to activate Swivel Lock with clicking sound

2. Step on Lower side pedal to De-activate Swivel Lock

NOTE

Wheel needs to be straight when you activate Swivel Lock.



Figure 3-56 • Caster position for Swivel lock

Transporting the System

Use extra care when transporting the system using vehicles. In addition to the instructions used when moving the system (see *Before Moving the System* on page 139 for more information), also perform the following:

1. Before transporting, place the system in its special storage case.
2. Only use vehicles that are designed for transport of the ultrasound system.
3. Load and unload the system to a vehicle parked on a level surface.
4. Ensure that the transporting vehicle can handle the weight of the system plus the passengers.
5. Ensure that the load capacity of the lift is capable of handling the weight of the system and any other items on the lift at the same time.

A minimum lift load capacity of 85 kg [187 lbs] is recommended.

6. Ensure that the lift is in good working order.
7. Secure the system while it is on the lift so that it cannot roll. Use either wood chocks, restraining straps, or other similar types of constraints. Do not attempt to hold it in place by hand.

NOTE

Strap the system below its handle so that the system does not break loose.

8. Employ two to three persons to load and unload safely from a vehicle.

Preparing the System for Use

9. Load the unit aboard the vehicle carefully and over its center of gravity. Keep the unit still and upright.

NOTE

Do not lay the unit down on its side.

10. Ensure that the system is firmly secured while inside the vehicle. Any movement, coupled with the weight of the system, could cause it to break loose.
11. Secure system with straps or as directed otherwise to prevent motion during transport.
12. Prevent vibration damage by driving cautiously. Avoid unpaved roads, excessive speeds, and erratic stops or starts.

Acclimation Time

After being transported, the unit requires one hour for each 2.5 degree increment when its temperature is below 10 degree C or above 40 degree C before powering on.

Table 3-24 • System Acclimation Time Chart

Degree C	60	55	50	45	40	35	30	25	20	15	10
Degree F	140	131	122	113	104	95	86	77	68	59	50
Hours	8	6	4	2	0	0	0	0	0	0	0
Degree C	5	0	-5	-10	-15	-20	-25	-30	-35	-40	
Degree F	41	32	23	14	5	-4	-13	-22	-31	-40	
Hours	2	4	6	8	10	12	14	16	18	20	

Chapter 4

Preparing for an Exam

In this section

Beginning an Exam.	146
Starting an Examination	153
Retrieving and Editing Archived Information.	162
MyPreset.	166

Beginning an Exam

Begin an exam by entering new patient information.

The operator should enter as much information as possible, such as:

- Dataflow
- Exam category
- Patient ID
- Patient name
- Exam Information

The patient's name and ID number is retained with each patient's image and transferred with each image during archiving or hard copy printing.

Patient Screen

Press **Patient** on the Touch Panel to display the Patient Screen on the monitor.

- Enter Patient Data with the alphanumeric keyboard.
- To navigate through the Patient Entry menu, use the **Tab** key or **Trackball** and **Set** to move and fix the cursor.




Figure 4-1 • Patient Screen (Example: Category ABD)

Table 4-1 • Patient Screen

No.	Function	Description
1.	Image Management	• Patient – Provides a search and creation of patient. (currently selected) • Image History – Provides a list of images per exam for the currently selected patient. • Active Images – Provides preview of the currently selected exam and Compare Assistant. • Data Transfer – Provides an interface to handle patient data from a remote device.
2.	Function Selection	• New Patient – Used to clear patient entry screen in order to input a new patient's data into the database. • Register – Used to enter new patient information into the database prior to the exam. If you are using the auto-generate Patient ID feature, do not select Register. It is always a good practice to Register all patients. • Details – Select the Details box to activate/deactivate the exam details. The Exam Description pull-down selection is used as the DICOM identifier.
3.	Query This Patient	Allows for One Click Q/R for the current patient.
4.	Logoff	Used to logoff system.
5.	EZBackup	One-step method to backup patient images to an external media.
6.	Dataflow Selection	Select archive and other pre-defined services. If you use a DVD-R, select DICOM CD Read in Dataflow. If you place the cursor on the icon, the pop-up menu displays disk capacity.
7.	Save and Exit	Used to save all changes and exit the Patient Menu.
8.	Patient Information	• Patient ID Number • Alternate (Other, Second) Patient ID Name and Number. The system allows you to enter a second identification number for the same patient, which may be required in certain countries. This is only displayed if enabled on the Utility > Connection Manager > Advanced Settings > Patient Screen. • Patient Name–Last, First and Middle • DOB (Birthdate) • Age (automatically calculated when birthdate is input) • Sex
9.	Category Selection	Select from 8 exam application categories. When a category is selected, the measurement and category presets are displayed.

No.	Function	Description
10.	Exam Information	Shows the Current/Active Exam information. Information pertinent to the selected exam category appears in the window. All possible information needs to be entered. - Images-Displays the selected exam's images in the middle of screen. - Clear-Clears existing data. - Past Exam (only for OB) – Input past exam data (register the patient before using).
11.	Scan Assistant Program	The Scan Assistant Program is either selected automatically or manually, depending on the preset as configured on the Utility > System > General page.
12.	Patient View/Exam View	Display either patient or examination list. Refer to <i>Patient View/Exam View</i> on page 148 for details.

Patient View/Exam View



Figure 4-2 • Patient View and Exam View

Table 4-2 • Patient View/Exam View

No.	Function	Description
1.	Patient View	Lists the patients in the database. When you double-click a patient on the patient list using the Set key, the Review screen or New Exam entry screen displays depending upon how the “Double click on patient list to start” preset is set at Utility > Connectivity > Miscellaneous.

No.	Function	Description
2.	Exam View	Displays the list of all the exams for the Current Patient. The system can display the Detail Mode instead of Exam View when you select a patient on the patient list and press Review or Register (to Register, enter a patient ID and select). If the Detail Mode preset on Utility > Connectivity > Miscellaneous menu is selected, the Detail Mode displays.
3.	Search	Select a search criteria. Note: Criteria "Img. Archived" means that the exam was backed up to external media by EZBackup or Export.
4.	String	Enter appropriate information for search criteria. Note: If you select Locked (Y, N) or Archived (Y, N) for the Search key, enter Y (Yes) or N (No).
5.	Clear	Clears the entered string.
6.	Hide	Hide the patient view.
7.	Review	Highlight the patient and press Review. Exam view is displayed. Note: If the selected patient has a Current Exam or the selected exam is the Current Exam, the Review button changes to "Resume Exam" on Patient List.
8.	New Exam	Creates a new exam for the current patient.
9.	Resume Exam	Continues the exam for that patient if you select the last exam of the day.
10.	Delete	Delete one or more patient record from Patient View. Note: "Delete" is only displayed when you login as Administrator.
11.	Detail	The detail information for the selected category displays.
12.	Lock/Unlock	Locks the exam/patient. Prevents move and delete functions. If you select the patient, all exams are locked. If you select one exam, the selected exam is locked and the lock icon displays in the patient ID cell.
13.	Listing xx/xx	Displays the quantity of patients that match the search criteria in the search window or the quantity of patients in the database.
14.	Images	Display images of the selected exam in the middle of the monitor.
15.	Clear Exam	Clear exam (located on the Touch Panel).
16.	Delete Exam	Delete one or more exams from Exam View. Note: "Delete" is only displayed when you login as Administrator.
17.	Disk	Displays the disk name on which you saved the exam's image data. If "+" displays behind the disk name, the data is saved on two or more disks.
18.	Send To	Send the images to the DICOM Device.

No.	Function	Description
19.	Print	Print the search list to a standard printer. Highlight the patient and press left Set key. Select Print from the pop-up and press the right Set key. 

NOTE

Refer to *Change Patient ID of the Existing Patient (Edit & Copy)* on page 158 and *Anonymize the patient* on page 1046 for more information about Edit&Copy and Anonymize.

OB Exam

Exam Preparation

Prior to an ultrasound examination, the patient should be informed of the clinical indication, specific benefits, potential risks, and alternatives, if any. In addition, if the patient requests information about the exposure time and intensity, it should be provided. Patient access to educational materials regarding ultrasound is strongly encouraged to supplement the information communicated directly to the patient. Furthermore, these examinations should be conducted in a manner and take place in a setting which assures patient dignity and privacy.

- Prior material knowledge and approval of the presence of nonessential personnel with the number of such personnel kept to a minimum.
- An intent to share with the parents per the physician's judgement, either during the examination or shortly thereafter, the information derived.
- An offer of choice about viewing the fetus.
- An offer of choice about learning the sex of the fetus, if such information becomes available.

Ultrasound examinations performed solely to satisfy the family's desire to know the fetal sex, to view the fetus, or to obtain a picture of the fetus should be discouraged.

Acoustic Output Considerations



DANGER

The ultrasound system is a multi-use device which is capable of exceeding FDA Pre-enactment acoustic output (spatial peak-temporal average) intensity limits for fetal applications. The interaction of sound energy with tissue at sufficiently high levels and/or longer duration can produce biological effects (aka bioeffects) of either a mechanical or thermal nature.

**DANGER**

It is prudent to conduct an exam with the minimum amount and duration of acoustic output necessary to optimize the image's diagnostic value. The interaction of sound energy with tissue at sufficiently high levels can produce biological effects (aka bioeffects) of either a mechanical or thermal nature.

Concerns Surrounding Fetal Exposure

Always be aware of the acoustic output level by observing the Acoustic Output Display. In addition, become thoroughly familiar with the Acoustic Output Display and equipment controls affecting output.

Training

It is recommended that all users receive proper training in fetal Doppler applications before performing them in a clinical setting. Please contact a local sales representative for training assistance.

To Start an Obstetrics Exam

NOTE

Calculation formulas are listed in the Advanced Reference Manual.

To begin an Obstetrics exam, you have to enter the OB-specific information. Obstetric patient fields are listed in the following table.

Table 4-3 • Obstetric Fields

Field	Description
LMP	Last Menstrual Period; enter the date that the patient started her last menstrual period. You must enter 4 digits for the year. When you type the month and day, the system fills in the slash (/). The Date Format preset chosen in Utility > System > General determines the required format.
BBT	Basal Body Temperature.
EDD by LMP	Estimated Delivery Date by LMP; the system fills in the date after you enter the LMP.
GA by LMP	Gestational Age by LMP; the system fills in the age after you enter the LMP.
Gravida	Number of pregnancies.
Para	Number of births.
AB	Number of abortions.
Ectopic	Number of ectopic pregnancies.
Fetus #	Number of fetuses; default is 1. Can be 1-4.
Accession #	Exam number used with hospital information system (DICOM). This is a tracking number from the worklist.
Exam Description	Describe the type of exam.
Perf Physician	The physician who performs the exam. Choose from the list or type the name.

Field	Description
Ref. Physician	The physician who requested the exam. Choose from the list or type the name.
Operator	The person (not a physician) who performs the scan. Choose from the list.

NOTE

To fill in the following information, move the **Trackball** to highlight the Detail button and press **Set**.

Table 4-4 • Obstetric Fields: Detail

Field	Description
Indications	Why the patient needs the ultrasound exam.
Comments	Comments about the exam.

Starting an Examination

Warnings and Cautions

**WARNING**

Imaging functions may be lost without warning. Develop emergency procedures to prepare for such an occurrence. Failure to prepare for unexpected loss of functionality could lead to patient injury.

**WARNING**

Always ensure you have selected a dataflow. If No Archive is selected, no patient data is saved and a rescan may be required. A Ø appears next to Dataflow if No Archive is selected.

**WARNING**

To avoid patient identification errors, always verify the identification with the patient. Make sure the correct patient identification appears on all screens and hard copy prints.

**CAUTION**

Always use the minimum power required to obtain acceptable images in accordance with applicable guidelines and policies.

**CAUTION**

Always use the system on a flat surface in the patient environment.

**CAUTION**

Ensure that the hands of the patient are away from the system during the exam.

The position of the operator and the patient vary by scan region.

In most cases, the operator sits/stands straight in front of the operator console and the patient lies on the bed on the right (or left) side of the system.

Creating a New Patient Record

When starting a new patient's exam, ensure you do the following:

1. Press **Patient** on the Touch Panel.
2. Press **New Patient** on the Touch Panel or the monitor.
3. Fill in patient information.

NOTE

Press Tab or Enter on A/N keyboard to move the cursor to next field.

NOTE

Do not use the following characters when filling in patient information:

“ ‘ \ / : ; . , * < > | + = [] &

4. Choose the exam category.

Enter the required information for the selected exam category.



Figure 4-3 • Exam category

5. Verify the dataflow.

NOTE

DO NOT use the removable media dataflows on the New Patient menu.

NOTE

The system can display a warning dialog when the patient is registered to “No Archive”. If the “Warn register to No Archive” preset is selected in the **Utility > Connection Manager > Advanced Settings > Dataflow Notification**, a warning displays. A different dataflow for permanent storage of patient data should be selected.

6. Select **Register**.
7. Select the probe to start scanning (or select Save and Exit, Esc, Scan, or Freeze).

Perform an Exam

1. Select the probe, exam category, and application.




Figure 4-4 • OB Preset - example

- 1 Probe
- 2 Exam Category
- 3 Application Preset

NOTE

See *Imaging Preset Manager* on page 594 for information on creating/editing user defined presets.

- 2. Perform an exam.
- 3. Store the raw data to the clipboard.

To store the still image, press **Freeze** and run the cineloop using the **Trackball**. Select the frame and press **P1** (or the assigned Print key).

To store the cineloop, press Print key without freezing or press **Freeze** and run the cineloop using the **Trackball**. Select the start/end frame and run the selected loop. Press **P1** (or the assigned Print key).

Ending an Exam



WARNING

To ensure correct storage of measurements, verify that the measurement result window is updated before you send or save the image.

- 1. When you have completed the study, press **End Exam** on the Touch Panel.
- 2. The image management screen displays. Select the images (still frame or cineloop) you want to store or select **Select All** to store all images. Select **Permanent Store** to store the images permanently.

NOTE

Return to the patient screen automatically from the scan screen when you select OK from the "ID is not unique" warning message.

Scanning Without Entering Any Patient Data

To scan a patient without entering any patient data until the end of the exam:

1. Press **Scan**.
2. When you scan the patient and save images to clipboard without a patient, you will receive a warning that states, "A patient must be selected for permanent storage of image." Press OK.
3. Press **Patient** to display the Patient Search screen.
4. Enter the Patient ID.

NOTE

If you have images or measurements that are not attached to a Patient ID, the message "Unsaved images, measurements or fetus number will be linked to the current patient information, continue?" appears when the patient ID is registered.

5. Enter patient data and exam information as necessary.
6. Press **Active Images**.
7. Press **Permanent Store**.

Starting a New Exam on an Existing Patient

1. Press **Patient**.
2. Press **New Patient** on the Touch Panel or the monitor.
3. Enter Patient ID or Last name, etc. in the patient information to display the target patient in Patient View.

For example, if you type "v" in Patient ID, only display patients begins with patient ID "v". or select Search key and enter a string to search a patient.

4. Highlight the target patient and double click the right **Set** key. Selected patient information displays.
5. Choose the exam category for new exam.
6. Select **New Exam**. A new exam is created in Exam View.
7. Select the probe to start scanning (or select Exit, Esc, Scan, or Freeze).

Retrieve Patient/Exam Information from Worklist

NOTE

Before you retrieve data from the Worklist server, make sure that Network LAN is connected correctly. If the worklist is not displayed, reboot the system.

1. Press **Patient** and select **Data Transfer**. The Data Transfer screen displays.

- Click on a radio button of **Worklist**. The patient/exam list displays in the Transfer To section.



- Worklist
- Worklist radio button
- Transfer button
- Worklist server

Figure 4-5 • Data Transfer

- The Worklist used last time is displayed on the monitor display. Press **Refresh** to refresh the list or select another Worklist server from the Transfer From pull-down menu.
- Select the patient(s) or the exam(s) from the list.
- Press **Transfer**. The progress bar displays during the transfer.
- Enter the required items and start an exam.

Sending Multiple Exams and Patients with Single Click to PACS

In the Patient View window, select 1 or multiple patients -- you can select by whole exam or whole patient -- to send to PACS. Select multiple items by using the Shift+Trackball Set Key.



Figure 4-6 • Multiple Patient/Exam Send To

Change Patient Information except Patient ID and scan

1. Perform Step 1 to 6 in *Starting a New Exam on an Existing Patient* on page 156.
2. You can change Patient information except Patient ID.



Figure 4-7 • Change Patient Information

3. Select the probe to start scanning (or select Exit, Esc, Scan, or Freeze).

Change Patient ID of the Existing Patient (Edit & Copy)

1. Ensure that you are logged in as an Administrator on the system.
2. Change the dataflow to Local Archive.
3. Select the patient from the Patient View list and left click to bring up the “Edit & Copy” pop-up menu.

NOTE

If you do not see the Edit and Copy ID option on the pop up menu, make sure the top half of the patient screen is blank.



Figure 4-8 • Select an Exam to Copy and Edit

4. A confirmation dialog displays. Select **OK**.



Figure 4-9 • Copy Patient Pop-Up Confirmation

5. The “Edit and Copy Patient” dialog displays. All the fields inherit the values from the original patient’s exam, except for Patient ID.



Figure 4-10 • Edit and Copy Patient Dialogue

- **Generate Patient ID** generates a patient ID.
- **Clear All** clears all fields except for Patient ID and Other ID.
- **Cancel** button cancels the “Edit and Copy Patient” function.

6. Fill in the Patient information fields, then press OK.

NOTE

If the newly entered Patient ID is not unique in the database, the Patient ID turns to red and an error message displays on the status bar.



Figure 4-11 • Copy Status Bar

7. When the copy is done, the patient list is refreshed.

Make sure that the patient record with new Patient ID is created.

NOTE

If you want delete Patient record with old Patient ID, click **Delete** on the Touch Panel.

Tips



WARNING

Confirm the patient identification prior to deleting/entering/importing any patient data. The user is responsible for patient data, diagnostic information or any other patient-related information entered in the database. Failure to follow instructions could lead to loss of or incorrect patient data.

The “Edit and Copy Patient” function:

- Copies data of a registered patient from the local archive to a newly-created patient in the local archive. The newly- created patient would have a new identification: patient ID, other ID, patient name, and sex etc.
- Assigns a new identification: new UIDs to the copied exam data. The newly-copied patient would have the same medical data as the patient being copied but with a different identification.

NOTE

“Edit and Copy Patient” will only copy patient data, images inside the local archive; it will not allow patient data or images from outside of the local archive. This includes the following types of images: exported, MPEGvue, SaveAs, DICOM Store, or (DICOM) print.

NOTE

“Edit and Copy Patient” does not copy the patient’s report.

NOTE

“Edit and Copy Patient” does not deal with patient information already burned in image pixels.

NOTE

The image and title bar, including patient information, is copied as is.

NOTE

The “Edit & Copy” function does not display for a current patient.

NOTE

The “Edit & Copy” function does not display when multiple patients are selected.

Retrieving and Editing Archived Information

Searching for an Existing Patient

1. Press **Patient** to display the Patient Screen.
2. Select the search key (Patient ID, First Name, Last Name, etc.). Type the search string.

When default configured, the system automatically searches to see if the patient is already in the archive. The result of this search is displayed in the Patients list.

NOTE

When the number of patients on a hard disk is in the hundreds, it takes time to search for a patient or switch to another screen. In this case, do one of the following:

- Uncheck the “Auto search for patient” preset, found under Patient/Exam Menu Options in **Utility > Connection Manager > Advanced Settings > Search** page.
- Delete unnecessary patient data.

3. Highlight the patient in the Patients list.

Select the Exam View tab to display a list of examinations instead of the patient records.



- 1Select **Review** to review the exam history of this patient.
If the study was performed on the same day, **Resume Exam** will populate in place of **Review** to continue the exam.
- 2Select **New Exam** to create a new exam for this patient.
- 3Select **Delete** to delete this patient.
- 4Lock/Unlock. Use to lock/unlock the exam/patient.

Figure 4-12 • Patient Search Screen

NOTE

“Delete” is only displayed when you login as Administrator.

Pop-up Menu

If you select the patient and press the left **Set** key, the pop-up menu displays.



Figure 4-13 • Archived Patient

If the study was performed on the same day, **Resume Exam** will populate to continue the exam.



Figure 4-14 • Patient of the Day

NOTE

The preset “Double click on patient list to start”, located on the **Utility > Connection Manager > Advanced Settings > Dataflow Settings**, allows you to display either the Review or New Exam screen by double clicking the **Set** key on the patient name.

Changing Patient Information or an Exam

If patient information needs to be edited, pressing **Patient** enables the Patient Screen for modifying information.

If the patient is still active, you can go to the New Patient page to change any field as well as select a different category. The exam changes in the Exam View area.

If the exam category needs to be changed, pressing **New Exam** allows modification of the Patient Screen without erasing accumulated patient images, measurements, annotations, calculations and worksheets.

1. Display the Patient screen by pressing **Patient**.

Preparing for an Exam

2. Select patient from the Patient list. The system automatically searches to see if the patient is already in the database.
 - Select Search key (Patient Data: ID, First Name, Last Name, Birthdate, Sex or Exam Date).
 - Enter search string (for example, initial letter of Patient Name)

3. The appropriate patient is displayed.

If patient information needs to be edited or the exam category changed, use the New Exam feature. Pressing **New Exam** allows modification of the Patient Screen without erasing accumulated patient images, measurements, annotations, calculations and worksheets.

NOTE

Patient identification information cannot be modified.

4. To have the database shown in its entirety, **Backspace** on the Search string and all patient names appear.
5. Select Register to register the new exam.

A new exam is automatically created on that patient unless an exam already exists on that day for that patient.
6. To display the patient information on the title bar, press the **Esc** key, the **B-Mode** key or **Register**.

Select the Model and appropriate probe Touch Panel keys, if necessary.

Deleting the Existing Patient/Exam/Image



WARNING

Before deleting a patient or image from the Patient Screen, make sure you have already saved the data with EZBackup, Backup, or Export. Verify the media before deletion. Failure to follow instruction could lead to loss of patient data.

Deleting the Existing Patient

1. Search and select the patient in the patient list with the **Ctrl** or **Shift** keys.
2. Select **Delete**. The confirmation dialog box displays.

OR

Press the left **Set** key. A pop-up menu displays. Select **Delete**. The confirmation dialog box displays.



Figure 4-15 • Select the Patient in the Patient List

3. Select **Yes** to delete or **No** to cancel.

Delete Multiple Patients from the Patient List

1. Select the multiple patients to be deleted from the patient list.
2. Select **Delete**. The confirmation dialog box displays.

OR

Press the left **Set** key. A pop-up menu displays. Select **Delete**. The confirmation dialog box displays.

3. Select **Yes** to delete or **No** to cancel.

Deleting the Existing Exam

1. Search and select the patient in the patient list.
2. Select **Review**.
3. The patient exam screen displays. Select the exam to be deleted.
4. Select **Delete**. The confirmation dialog box displays.
5. Select **Yes** to delete or **No** to cancel.

Deleting the Existing Image

1. Search and select the patient in the patient list.
2. Select **Review**. The patient exam screen displays.
3. Select the exam which contains the image to be deleted.
4. Select **Active Images** to display the image list.
5. Select the image to delete and select **Delete**. The confirmation dialog box displays.
6. Select **Yes** to delete or **No** to cancel.

MyPreset

MyPreset allows configuration of specific presets, according to the probe.

Presets can be arranged and edited on the Touch Panel in Utility. See *Arranging MyPreset Tab* on page 167 for more information.

Activating MyPreset

1. Check **Default MyPreset** in Utility > System > System Imaging > Control to start MyPreset.
2. From the Touch Panel, select the active probe icon. The MyPreset exam tab displays on the Touch Panel.




Figure 4-16 • MyPreset Tab

NOTE

Select MyPreset (star icon) on the Touch Panel. Conventional exam tab displays.



Figure 4-17 • Conventional Exam Tab

Arranging MyPreset Tab

MyPreset configurations for each probe can be updated, if desired. The configuration must be updated for each probe separately.




Figure 4-18 • Imaging Preset Manager

1. Select MyPreset.
2. Select the desired Probe.
3. Select presets from Available Imaging Presets column to be displayed on the MyPreset tab. Use the up/down arrow to move the preset to the desired position.

Chapter 5

Optimizing the Image

In this section

Optimizing B-Mode.	170
Optimizing M-Mode.	191
Optimizing Color Flow.	195
Optimizing M Color Flow.	204
Optimizing Spectral Doppler.	205
Using 3D.	212

Optimizing B-Mode

Intended Use

B-Mode is intended to provide two-dimensional images and measurement capabilities concerning the anatomical structure of soft tissue.

Typical B-Mode Exam Protocol

A typical examination using B-Mode might proceed as follows:

1. Record exam-related patient information. Verify system setup (probes and presets).
2. Position the patient and the console for optimum operator and patient comfort. Perform the scan.
3. Complete the study by collecting all the data.

B-Mode Scanning Hints



HINTS

These B-Mode controls produce the following results:

Auto Optimize. Automatically improves the contrast resolution of the image by changing the gray scale to match the image data. Available in B-Mode and Doppler Mode.

Coded Harmonics. Improves image clarity and tissue contrast by reducing clutter and artifacts.

B-Flow. Provides a more intuitive representation of non-quantitative hemodynamics in vascular structures.

Frequency. Changes system parameters to best optimize for a particular patient type.

Gray Map. Affects the presentation of B-Mode information. Choose the gray map prior to making other adjustments. There is an interdependency between gray maps, gain, and dynamic range. If you change a map, revisit gain and dynamic range settings.

Dynamic Range. Changes the amount of gray scale information displayed. A higher dynamic range shows more gray scale information displayed, while a lower dynamic range displays less gray scale information onto the same display scale. If you increase the gain, you may want to decrease the Dynamic Range.


Frame Average. Smooths the image by averaging frames. Reduces noise in the image.

TGC. Adjust TGC to adjust Gain in specific areas.

Width. Sizes region of interest. Adjust the Width to the smallest reasonable size to maximize frame rate.

B-Mode Controls

Table 5-1 • B-Mode Controls

Control	Adjusts Acoustic Output	Description/Benefit
Depth	Yes	To increase/decrease, adjust Depth. Depth controls the distance over which the B-Mode images anatomy. To visualize deeper structures, increase the depth. If there is a large part of the display which is unused at the bottom, decrease the depth.
Gain	No	To decrease/increase, rotate Gain (Mode key). B-Mode Gain increases or decreases the amount of echo information displayed in an image. It may have the effect of brightening or darkening the image if sufficient echo information is generated. NOTE: Always optimize gain before increasing the Power Output.
Focus 	Yes	In B-Mode, 4D and Contrast (Reference and Contrast) mode, the Focus is applied uniformly across the entire field of view, as indicated by the Focus indicator overlaid on the Depth indicator (see example 1 on the left). In Contrast MVI, CF/PDI/MVI, the Focus position is focused in the Region of Interest, as indicated by the “Dot” indicator overlaid on the Depth indicator (see example 2 on the left). The focus position is assigned to a rotary on the Contrast MVI, CF/PDI/MVI tab.
Auto Optimize	No	See <i>Continuous Auto Tissue Optimization</i> on page 174.
ACE	No	Use ACE (Adaptive Contrast Enhancement) to emphasize echoes from real structures while reducing noise/haze. This results in enhanced signal-to-noise ratio. A sub-selection of ACE allows you to adjust how much ACE is applied, including a setting showing the coherence factor. Values are Light, Medium, or High filtering, or None (no filtering).
Mode Cursor	No	Displays the M/D-Mode cursor on the B-Mode image. To activate/deactivate the M/D-Mode cursor, press Mode Cursor (left Set key). Trackball to position the M/D-Mode cursor. Adjust the Angle and SV Length as necessary.
SRI-HD	No	See <i>SRI-HD</i> on page 175.

Control	Adjusts Acoustic Output	Description/Benefit
CrossXBeam	No	CrossXBeam Angle adjusts maximum steering angle of CrossXBeam. Higher values correspond to larger angles. See <i>CrossXBeam</i> on page 177.
Coded Harmonic Imaging (CHI)	Yes	To activate Coded Harmonic imaging, press CHI on the control panel. Coded Harmonics enhances near field resolution, diminishes low frequency high amplitude noise and improves imaging technically difficult patients. Coded Harmonics may be especially beneficial when imaging isoechoic lesions in shallow-depth anatomy in the breast, liver, and hard-to-visualize fetal anatomy.
Frequency	Yes	Adjust Frequency until the desired frequency is selected. Multi Frequency mode lets you downshift to the probe's next lower frequency or shift up to a higher frequency.
Steer	Yes	To slant the linear image to the left/right, select Steer. You can slant the B-Mode or Color Flow linear image left or right to get more information without moving the probe. The angle steer function only applies to linear probes. Linear probes can be steered left, center, or right up to a maximum of 15 degrees, depending on the probe.
Virtual Convex	Yes	To activate/deactivate Virtual Convex, select Virtual Convex. On Linear and Sector probes, Virtual Convex provides a larger field of view in the far field. Virtual Convex is defaulted as active with Sector probes.
Max Angle	Yes	With IC5-9-D and RIC9-5-D probes, the Virtual Convex control becomes Max Angle. Max Angle toggles the field of view to 179 degrees.
TGC	No	TGC amplifies returning signals to correct for the attenuation caused by tissues at increasing depths. TGC slide pots are spaced proportionately to the depth. The area each pot amplifies varies as well. A TGC curve may appear on the display (if preset), matching the controls that you have set (except during zoom). You can choose to deactivate the TGC curve on the image.
Width	Yes	To narrow/widen the sector width, rotate the Width control (located on the Depth control). You can widen or narrow the size of the sector angle to maximize the image's region of interest (ROI).
Tilt	Yes	Tilt is available on the Trackball and/or the Rotary control, allowing sector angle adjustment to obtain more information without moving the probe. Trackball Control: When shown as available on the Trackball controls on the display, use the Trackball key and the Trackball to tilt the angle to the left/right. Rotary Control: Rotate the Rotary control to tilt the angle to the left/right. Pressing the control resets Tilt to center. Tilt is not available on Linear probes. Tilt is not available while in CrossXBeam.

Control	Adjusts Acoustic Output	Description/Benefit
Dynamic Range	No	To increase/decrease, adjust Dynamic Range. Dynamic Range controls how echo intensities are converted to shades of gray, thereby increasing the adjustable range of contrast. Dynamic Range is useful for optimizing tissue texture for different anatomy. Dynamic Range should be adjusted so that the highest amplitude edges appear as white while lowest levels (such as blood) are just visible.
Reverse (if Preset)	No	To flip the image 180 degrees, press the Reverse key. Flips the image 180 degrees left/right. WARNING: When reading a reverse image, be careful to observe the probe orientation to avoid possible confusion over scan direction or left/right image reversal. Failure to follow instructions could result in misinterpretation.
Maps	No	To select a map, press the Gray Map on the Touch Panel. A map window displays. The image reflects the map as you go through the selections. The system supplies B, M, and Doppler Mode system maps. Maps are preset-specific and are arranged in order from the softest map on the top of the menu to the map with the most contrast on the bottom of the menu. The only exception is Map J, which is a very soft map.
Frame Average	No	Temporal filter that averages frames together, thereby using more information to make up one image. This has the effect of smoothing the image and reducing apparent noise.
Colorize	No	1. Select Colorize on the Touch Panel. 2. Trackball to cycle through available maps. 3. Press Set to select. Colorize is the colorization of a conventional B-Mode image or Doppler Spectrum to enhance the user's ability to discern B, M, and Doppler Mode intensity valuations. NOTE: You can colorize realtime or CINE images or Timeline CINE, but not DVR images. Colorizes the gray scale image to enhance the eye's discrimination capability. The gray bar displays while Colorize is activated. To deselect, select a gray map.
Rotation	No	Rotates the image 180 degrees. Beneficial in transvaginal and transrectal scanning. WARNING: When reading a rotated image, be careful to observe the probe orientation to avoid possible confusion over scan direction or left/right image reversal. Failure to follow instructions could result in misinterpretation.

Control	Adjusts Acoustic Output	Description/Benefit
Frame Rate	Yes	Optimizes frame rate or spatial resolution for the best possible image. High Frame Rate is useful in fetal heartbeat, adult cardiac applications, and clinical radiology applications which require significantly higher frame rates. Low Frame Rate (high resolution) is useful in situations where very small vessels are being imaged, e.g., thyroid, testicles.
Rejection	No	Selects a level below which echoes will not be amplified (an echo must have a certain minimum amplitude before it will be processed). Allows for the elimination from the display of low level echoes caused by noise.
Suppression	No	Suppresses the noise in the image.
Speed of Sound	No	See <i>Speed of Sound (SoS) Tissue Imaging</i> on page 188.

Continuous Auto Tissue Optimization

Continuous Auto Tissue Optimization (CATO) optimizes the image based on the B-Mode image data. Different preset levels (Low, Medium, High) allow users to choose a preference for contrast enhancement of the B-Mode image. Low provides the least amount of contrast enhancement, high provides most.

CATO is available in single or multi images, on live, frozen or CINE images (in B-Mode only), and while in zoom.

NOTE

The probe orientation mark color is blue with green lines above and below.



Figure 5-1 • Probe Orientation Marker

Auto Optimize Preset

The CATO preset can be set at Utility > Imaging > B-Mode > Auto Optimize Mode. Specify the CATO Level: Low, Medium, or High.



Figure 5-2 • CATO Preset

Continuous Tissue Optimization (CTO)

CTO optimizes the spatial continuous gain adjustment on the B-Mode image. CTO can be used alone or in combination with CATO.

Preset

Select Auto Optimize Mode at **Utility > Imaging > B > Auto Optimize** Mode. CTO only, CATO only or both may be selected to use a combination of CTO and CATO to optimize the B-Mode image.

Adjusting

To activate CTO, press the right **Auto** control.

To deactivate CTO, press the left **Auto** control.

To adjust the overall CTO Gain, use “CTO Gain” on the Touch Panel or B-Mode Gain. This can be set via **Utility > Imaging > B > CTO Gain**.

Availability

CTO is available on the 6S-D, BE9CS-D, C1-6-D, C1-6VN-D, C2-7-D, C2-7VN-D, C2-9-D, C2-9VN-D, C3-10-D, IC5-9-D, L2-9-D, L2-9VN-D, L3-9i-D, L3-12-D, L4-20t-D, L6-24-D, L8-18i-D, M5Sc-D, ML6-15-D, RAB6-D and RIC5-9-D probes, in certain applications.

SRI-HD

SRI-HD (High Detection Speckle Reduction Imaging) is an adaptive algorithm to reduce the unwanted effects of speckle in the ultrasound image. Image speckle usually appears as a grainy texture in otherwise uniform areas of tissue. Its appearance is related to image system characteristics, rather than tissue characteristics, so that changes in system settings, such as probe type, frequency, depth, and others, can change the appearance of the speckle. Too much speckle can impair image quality and make it difficult to see the desired detail in the image. Likewise, too much filtering of speckle can mask or obscure desired image detail. Extra care must be taken to select the optimal SRI-HD level.

SRI-HD

SRI-HD is available in B-Mode imaging and may be used with any transducer or clinical application when image speckle appears to interfere with the desired image detail.

Advanced SRI Type 1

Advanced SRI Type 1 is available in B-Mode for select applications, to produce images with clear borders and reduced speckle while maintaining natural image texture.

Advanced SRI Type 2

Advanced SRI Type 2 is available in B-Mode for OB/GYN applications (as a purchasable option), to produce smoother, cleaner images with enhanced border delineation.

SRI-HD Type 2 is available for the probes and applications listed in *Table 5-2 • Advanced SRI Type 2 OB/GYN Probes and Applications* on page 176.

Table 5-2 • Advanced SRI Type 2 OB/GYN Probes and Applications

Probe	OB1	OB23	FetalEcho	NT	GYN	Follicle
C1-6-D/C1-6VN-D	X	X	X	X	X	X
C2-9-D/C2-9VN-D	X	X	X	X	X	X
RAB6-D	X	X	X	X	X	X
L2-9-D/L2-9VN-D	X	X	X	X	X	X
L3-12-D	X	X	X	X		
IC5-9-D	X	X		X	X	X
RIC5-9-D	X	X		X	X	X

NOTE

When available per probe and application, the SRI Type can be chosen on the B-Mode tab, page 2.

Adjusting

Adjust SRI-HD levels on the Touch Panel. You can also set presets via Utility > Imaging > B-Mode.

NOTE

We recommend selecting the SRI-HD level by observing the enhanced image in side-by-side dual image comparison with the original, unprocessed image. Dual display mode is activated by pressing the L and R keys simultaneously.




Figure 5-3 • SRI-HD Adjustment

Values

Level values vary, depending upon the probe. 0 = Minimal Filtering. Increasing the level value increases the amount of SRI filtering and will result in smoother images.

Benefits

SRI-HD smooths the image when image speckle interferes with the desired image detail.

Tips and Notes

When selecting the SRI-HD level, observe the effect of SRI-HD in the desired region of interest and make a real-time comparison with the original image. The optimal level depends on the clinical situation. Observing the original and SRI-HD-processed images together helps to determine whether too much or too little SRI-HD has been applied.

Dual image mode for SRI-HD can also be activated on a stored CINE Loop. This allows you to always see the original, unprocessed or enhanced image by going into the Dual display mode and to change the SRI-HD settings when reviewing the CINE Loop.

SRI-HD is available on 3D/4D (sectional image and render image).

- You cannot change SRI-HD after the scan starts.
- The effects for the rendered image are less than the 2D-image.

CrossXBeam

CrossXBeam is the process of combining three or more frames from different steering angles into a single frame. CrossXBeam is available on Convex and Linear probes.

CrossXBeam combines multiple co-planar images from different view angles into a single image at real-time frame rates.

CrossXBeam Angle adjusts maximum steering angle of CrossXBeam. Higher values correspond to larger angles.

Optimizing the Image

Adjusting

To activate CrossXBeam, press the **CrossXBeam** key on the Touch Panel; an “X” appears on the display next to “CHI” or “B.”

You can also adjust the **CrossXBeam Angle** and **CrossXBeam Type** (Mean/Hybrid/Max/MotionCorrection) on the Touch Panel. **Max** detects maximum values; **Mean** detects averaged values; **Hybrid** detects a mix of both average and maximum values; **MotionCorrection** suppresses blurriness and/or artifacts caused by lateral motion.

Values

All linear and convex probes are supported. B-Mode CrossXBeam is available while in B-Mode, Color Flow, or PW Doppler Mode. Steering is optimized by probe.

Benefits

The combined single image has the benefits of:

- reduced speckle noise,
- reduced clutter, and
- continuity of specular reflectors.

Therefore, this technique can improve:

- contrast resolution with increased conspicuity of low contrast lesions,
- better detection of calcifications,
- biopsy needle visualization, and
- cystic boundary definition.

B-Flow

B-Flow is intended to provide a more intuitive representation of non-quantitative hemodynamics in vascular structures.

B-Flow is digitally-encoded Ultrasound, using digital codes to enhance weak signals from small particulate reflectors (blood flow) and suppress signals from strong reflectors (tissue). Flow and tissue are displayed simultaneously without threshold decision and overlay.

All B-Mode measurements are available with B-Flow active: depth, distance along a straight line, % stenosis, volume, trace, circumference, and enclosed area.

Presetting

Preset the default B Flow Mode via Utility > Imaging > General > BF button.

Activating

To activate/deactivate B-Flow, press **B-Flow** on the control panel. The preset default B Flow Mode appears. PW Doppler Mode is available while in B-Flow; however, M-Mode and Color Flow/PDI Modes are not available.

Using B-Flow

To optimize the image:

Adjust the frequency, display depth, and focal zone location based on the patient body type and anatomy of interest. Adjust Sensitivity/PRI and Background setting as needed (see below).

Adjust the remaining Imaging parameters and presets as needed; functionally is the same as B-Mode when in B-Flow mode.

Scanning Hints

B-Flow provides an intuitive view of blood flow, acute thrombosis, soft plaque, small vessel perfusion, and high grade stenosis. Compared to Color Doppler, it does not display bleeding, blooming, or aliasing artifacts.

To view high speed jets only for stenosis, use a PRI value of 10 or lower. Use the Background control to display an appropriate amount of tissue background including plaque.

To view slow flow, use a large PRI value while avoiding a PRI value that introduces a bar artifact.

Benefits

Compared to Color Doppler mode, B-Flow provides better spatial and temporal resolution, displays blood flow in the entire image, i.e. NO ROI, and is not angle dependent as it does not use the Doppler Principle. B-Flow is therefore a more realistic (intuitive) representation of flow information, allowing you to view both high and low velocity flow at the same time.

Affect on Other Controls

When you activate B-Flow, the system remembers the imaging parameters set while in B-Mode. When you optimize the frame rate via Line Density, you compromise the resolution and when you optimize the resolution, you compromise the frame rate. B-Flow is not available in 3DView; but is available in Easy 3D.

Bioeffects

Using B-Flow Frequency, Focus Position/Number, Sensitivity/PRI, Line Density, Visualization and Power Output may change the TI and/or MI. Observe the output display for possible effects.

Background

Background and Tissue controls work together to display the amount of background shown on the B-Flow image. Background provides a large change in the amount of tissue displayed; Tissue provides fine tuning to the amount of tissue displayed.

Adjusts the amount of tissue displayed on the B-Flow image. Works along with the Tissue control.

Value

0, 1, 2, 3

- 0 = Least amount of tissue displayed.
- 3 = Most tissue displayed.

NOTE

Not available for Frozen images or in CINE.

Accumulation

Accumulation enhances the flow in an image; ideal to capture dynamic flow in a still picture.

Optimizing the Image

Values

Off - Infinite. Infinite provides the same result as applying CINE Capture to a B-Flow CINE clip.

Benefit

Accumulation detects the maximum signal and holds it (accumulates it) for the level specified (Off - Infinite).

Capture

B-Flow Capture provides users the ability to create an accumulated image in B-Flow Live. Maximum intensity projection algorithm to detect the strongest signal between the frames and display the results.

Capture allows the user to represent the dynamic flow situation in the vasculature with a single still image. Adjustments can be made to modify the start and end frames used in the process. Select from the Touch Panel, a "C" will appear in the image parameter list in place of frame average.

NOTE

B-Flow Capture is assigned to the right Trackball key. You can operate B-Flow Capture from the Touch Panel or the right Trackball key. B-Flow Capture is available only while in Research mode.

NOTE

When turning on Live Capture, the cine buffer is cleared of prior data.

Sensitivity/PRI

Sensitivity/PRI (Pulse Repetition Interval) is proportional to the time interval between the pulses sent to develop the B-Flow image.

In general, a larger value is recommended for slow flow as slow flow detection requires more time separation between pulses so the system can detect the difference in flow profile. However, a larger value could cause bar artifacts on the image. Therefore, it is suggested to not increase the PRI value more than needed. A small value of PRI should be used when the interest is in fast flow only, e.g. viewing a jet in a stenosis case, where the jet is of interest.

NOTE

Sensitivity/PRI is Probe and Exam Category dependent.

Bioeffects

Using B-Flow Sensitivity/PRI may change the TI and/or MI. Observe the output display for possible effects.

Visualization

Define the display technique.

Values

- B-Flow. Displays only B-Flow image.
- Dual. Displays B-Mode and B-Flow images simultaneously using dual screen.

- Hybrid. Displays the B-Flow image over the B-Mode image using Hybrid Map.

Available Probes

NOTE

Available for C1-6-D, C1-6VN-D, C2-7-D, C2-7VN-D, C2-9-D, C2-9VN-D, C3-10-D, L2-9-D, L2-9VN-D, L3-12-D, L6-24-D, L8-18i-D, M5Sc-D and ML6-15-D probes.

LOGIQView

LOGIQView provides the ability to construct and view a static 2D image which is wider than the field of view of a given transducer. This feature allows viewing and measurements of anatomy that is larger than what would fit in a single image. Examples include scanning of vascular structures and connective tissues in the arms and legs.

LOGIQView constructs the extended image from individual image frames as the operator slides the transducer along the surface of the skin in the direction of the scan plane. The quality of the resulting image is somewhat user-dependent and requires some additional skill and practice to develop proper technique and become fully proficient.

LOGIQView is not available for the following:

- Multi Image
- Timeline Modes
- B-Flow Mode
- Color Flow Mode
- PDI Mode

Benefits

The user can look at a larger region of interest within one field of view that is wider than any given probe would normally provide.

Clinical Use

LOGIQView is intended for scanning areas too large to fit on one image.

Using LOGIQView

To perform an exam using LOGIQView:

1. Perform a detailed examination of the anatomy/pathology. Optimize parameters for tissue texture and visible window PRIOR TO activating LOGIQView.
2. Press the LOGIQView key on the Control Panel.
3. To start acquiring the image, press Start (Trackball key).



HINTS

When you scan, start with a strong, quick sweep in the direction of the acquisition and then complete the LOGIQView image with a slow steady sweep. LOGIQView acquires images via leading edge vectors (and does not acquire slices, as in CINE). The image is being stored as you perform the scan and you can watch the LOGIQView as it is being acquired.

Optimizing the Image

4. To restart the scan, press Start again. You can back up the probe, realign it, then go forward to redo a portion of the scan.
5. To complete the scan, press End or Freeze (or allow the scan to auto complete. The LOGIQView is then displayed, scaled to fit entirely on the screen.
6. Perform measurements and record images.
7. Select Frame Review to move through the LOGIQView one frame at a time. Use the Trackball to scroll.

NOTE

Measurement error is within 5% of the distance you measured for all linear probes.

Uniform Motion

The quality and usefulness of LOGIQView images is affected by transducer motion. Incorrect technique can contribute to image distortion.

Guidance and precautions for uniform motion:

- Continuous contact is required throughout the length of the extended image. DO NOT lift the transducer from the skin surface.
- Always keep the transducer perpendicular to the skin surface. DO NOT rock the transducer.
- Keep the motion within the same scan plane, if possible. DO NOT slide the transducer laterally.
- Lateral turning (change in direction to follow anatomical structure) can be accommodated with slower motion. DO NOT make abrupt changes in direction.
- The system accommodates a reasonable range of motion velocity. DO NOT make abrupt changes in speed of motion. Deeper scans generally require reduced speed.

Bioeffects

Activating LOGIQView has no affect upon Acoustic Output values.

BSteer+ (Option)

BSteer+ allows you to obtain improved biopsy needle visualization without changing B-Mode imaging parameters. B BSteer+ is available in B-Mode, Color Flow mode, and PDI mode. BSteer+ is available with all linear probes and the following convex probe: C1-6-D.

Benefits

BSteer+ provides better needle visualization compared to normal B-Mode. B Steer+ allows you to adjust beam angle and needle gain to achieve best possible needle visualization without affecting the B-Mode image. The enhanced needle is superimposed on top of the B-Mode image, giving the best possible image of both needle and tissue.

Activating

To activate BSteer+ press the BSteer+ button (1) in the B tab on the Touch Panel.

NOTE

To activate BSteer+ in Color Flow or PDI mode, first select the B tab on the Touch Panel.




Figure 5-4 • Button for activating BSteer+

The steer knob may also be used to activate BSteer+. To do this, check the option Steer knob activates BSteer+ in Utility > Imaging > General menu. Pushing the steer knob left or right will then activate BSteer+.

NOTE

When using the Steer knob for BSteer+, this functionality only applies only to the selected probe and preset in the Utility > Imaging menu.

NOTE

When using the Steer knob for BSteer+, normal steer functionality is not be available for the selected probe and preset.

Activating BSteer+ may decrease frame rate.

Using BSteer+

When BSteer+ is activated, the BSteer+ symbol (1), and a borderline appears on the image. The borderline indicates the boundary of the needle enhancement area. Beneath the line, the needle is not enhanced by BSteer+.



Figure 5-5 • BSteer+ Symbol (1) and Borderline (2)

NOTE

For convex probes, 2 borderlines may appear. The region between the lines is the needle enhancement area. This area is narrow relative to the width of the image for a convex probe because the beams are steered so that they are parallel. For convex probes, a limited part of the needle in the image will be enhanced (the tip as it first enters the issue, and the upper part of the shaft as it is inserted deeper).



Figure 5-6 • Two Borderlines May appear in BSteer+ for Convex Probes

The BSteer+ tab and the following controls appear on the Touch Panel.



- 1 Beam Direction
- 2 Beam Angle
- 3 Dynamic FOV (field-of-view)
- 4 Needle Resolution
- 5 Needle Gain
- 6 BSteer+ (for deactivating BSteer+)

Figure 5-7 • BSteer+ Tab and Controls on the Touch Panel

To optimize the biopsy image using BSteer+ functionality, adjust BSteer+ beam direction, angle and needle gain.

Adjust BSteer+ beam direction and angle to make needle and beam angle as perpendicular as possible.

BSteer+ beam direction and angle are selectable using the direction and angle buttons on the Touch Panel.

NOTE

BSteer+ angle can also be adjusted using the Set Keys above the trackball if **Use set keys to change BSteer+ Angle** is checked in Utility > System > User Configurable Key menu.

NOTE

When the BSteer+ tab is active on the Touch Panel, the Auto Keys can also be used to adjust BSteer+ angle. To use Auto Keys to activate and deactivate auto optimize while in BSteer+, select the B tab on the control panel.

Adjust Needle Gain to achieve desired needle brightness. Gain is adjustable by rotating the Needle Gain rotary or the Depth rotary (range from 0 to 100).

NOTE

If the Needle Gain is too high, noise may be amplified along with needle.

Turning on Dynamic FOV will make the needle signal stronger when the steering angle is not perfectly perpendicular to the needle.

NOTE

Turning on BSteer+ Dynamic FOV may decrease frame rate.

Select Needle Resolution to achieve desired spatial resolution for the needle. High will make the needle appear thinner, and Low will make the needle appear thicker.

BSteer+ values (including enable/disable, beam direction, angle and Needle Gain) are returned to factory default values or user preset values when the following change: Probe, Exam Category, Exam Calcs or New Patient.

Preset



Figure 5-8 • BSteer+ Preset

- BSteer+ Beam Direction - preset via Utility -> Imaging -> B Tab.
- BSteer+ Beam Angle - preset via Utility -> Imaging -> B Tab.
- BSteer+ Needle Gain - preset via Utility -> Imaging -> B Tab.
- BSteer+ Dynamic FOV - preset via Utility -> Imaging -> B Tab.
- BSteer+ Needle Resolution - preset via Utility -> Imaging -> B Tab.

Touch TGC

TGC amplifies returning signals to correct for the attenuation caused by tissues at increasing depths. TGC slide pots are spaced proportionately to the depth. A TGC curve may appear on the display (if preset), matching the controls that you have set (except during zoom).

NOTE

TGC adjusts the image automatically when using zoom.

Values

When you change the depth, TGC is rescaled across the new depth range. Each pot is proportionately scaled across the depth.

User-Defined Preset

There are 4 presets located on and selected via the Touch Panel. The right-most preset is locked and cannot be changed.

TGC Display On/Off -- preset via Utility > System > System Display.

Benefits

TGC balances the image so that the density of echoes is the same throughout the image.

Adjusting Touch TGC with B-Mode Image Overlay

You can adjust the TGC by overlaying the TGC slidepots over the B-Mode image on the Touch Panel (1), or via the Touch Panel TGC slide pots (2).




Figure 5-9 • Adjusting the TGC

Adjusting TGC with the B-Mode Image Overlay

To adjust the TGC interactively with the B-Mode image overlay:

1. Press the TGC control, located in the upper, right-hand corner of the B-Mode Touch Panel. TGC “slidepots” appear over the B-Mode image.



Figure 5-10 • TGC Control

Pressing the bottom, right rotary or bottom, right corner of the Touch Panel also enters Touch TGC and adjusts the image.

2. Adjust these slidepots by sliding each slidepot on the Touch Panel with your finger.



HINTS

To hold the B-Mode image while adjusting the TGC, select the Hold Image lock located in the upper left-hand corner of the Touch Panel. The Hold Image lock overrides the Utility screen Auto Hide delay timer and hides the TGC slidepots, allowing unobstructed view of the image on the Touch Panel.

When Hold Image is active (or the Auto Hide delay timer is set to Off on the Utility screen), the Full Image button appears. Select the Full Image button to toggle between active B-Mode and the full image area.

NOTE

Though the B-Mode image is locked, you can still adjust the TGC curve while the image lock is active.

3. To save this TGC setting as a preset, press and hold the box on the right until the box displays a blue outline. This TGC setting now appears in the “Custom” window.
4. To reset to the default setting, touch the TGC Presets preset at the bottom of the TGC display on the Touch Panel.

NOTE

Timeout for TGC display can be set in Utilities > System > General.

Adjusting TGC Using the Touch Panel

You can also adjust the TGC or select a TGC preset directly on the Touch Panel.

1. Adjust these slidepots by sliding each slidepot on the Touch Panel with your finger; or select TGC preset.
2. To save a TGC setting as a preset, press and hold an available TGC preset box on the bottom of the Touch Panel until the box displays a blue outline.

Speed of Sound (SoS) Tissue Imaging

Speed of Sound is available on all probes for the following applications: Abdomen, Abdomen 2, Renal, Bowel, and Breast. The Speed of Sound control is displayed only for these applications and are hidden in all other applications, even when research mode is enabled.

NOTE

Speed of Sound displays on the monitor display as “SoS” with the speed following, “SoS 1500” (when the speed of sound is not equal to 1540).

A control has been added on the Touch Panel to change the transmitted speed of sound for various breast tissue types:

To activate Speed of Sound for Breast, for example,

1. Select Probe > Small Parts > Breast > B-Mode > Speed of Sound.
2. Press Speed of Sound on the Touch Panel. Speed of Sound selections appear at the bottom of the Touch Panel in place of the Focus Position control.

3. Adjust the Speed of Sound control up/down to achieve the desired image. The system displays the Speed of Sound (SoS) on the Touch Panel in the upper, right-hand corner of the display as "SoS ####".

SoS settings are returned to the default for an SoS of 1540 (or when the SoS is not displayed on the display).

Minimizing Grating Lobe/Side Lobe Artifacts

Overview

Grating lobes/side lobes often create a spurious structure or clouding within an ultrasound image. Such artifacts are not uncommon in gray scale 2D imaging mode. They are caused by an undesired off-axis ultrasound beam contaminating the main beam.

The ultrasound system presumes that the reflections returning to the transducer always come from the steered direction. However, when these (weaker) off-axis beams encounter a strong specular reflector, the returning reflected energy will be received and "added" to the main beam.

Grating lobe/side lobe artifacts can be especially noticeable within anechoic regions (i.e. fluid, cysts, heart chamber, large blood vessels, etc.).

Corrective Adjustments

On the ultrasound system, if grating lobe/side lobe artifacts are suspected during the scan, adjustments can be made to reduce such artifacts.

1. Adjust scan window
2. Reduce CrossXBeam Angle
3. Decrease AO

For example, on C3-10-D probe, at default setting for NeoHead, grating lobe/side lobe artifacts are present in *Figure 5-11 • Image with Grating Lobe Artifact* on page 189.




Figure 5-11 • Image with Grating Lobe Artifact

Optimizing the Image

To correct, reduce the CrossXBeam Angle and decrease the AO from 100% to 90%. The grating lobe/side lobe artifacts are removed as shown in *Optimizing M-Mode* on page 191.




Figure 5-12 • Image with Grating Lobe Artifact Removed

Optimizing M-Mode

M-Mode is used to determine patterns of motion for objects within the ultrasound beam. The most common use is for viewing motion patterns of the heart.

The ultrasound system has three types of M-Mode:

- Conventional M-Mode: displays a distance/time plot of a cursor line in the axial plane of the 2D-image. Conventional M-Mode can be combined with Color Mode.
- Anatomical M-Mode (AMM)
- Curved Anatomical M-Mode (CAMM)

Intended Use

M-Mode is intended to provide a display format and measurement capability that represents tissue displacement (motion) occurring over time along a single vector.

M-Mode Display



Figure 5-13 • M-Mode Display

Typical Exam Protocol

A typical examination using M-Mode might proceed as follows:

1. Get a good B-Mode image. Survey the anatomy and place the area of interest near the center of the B-Mode image.
2. Press **Mode Cursor**.
3. Trackball to position the mode cursor over the area that you want to display in M-Mode.
4. Press M-Mode.
5. Adjust the Sweep Speed, TGC, Gain, Power Output, as needed.

Optimizing the Image

6. Press **Freeze** to stop the M trace.
7. Record the trace to disk or to the hard copy device.
8. Press **Freeze** to continue imaging.
9. To exit, press M-Mode.

M-Mode Controls

NOTE

You can set default value of each parameter by probe and application on the Utility > Imaging page. See *Imaging Presets* on page 473 for more information.

Table 5-3 • M-Mode Controls

Control	Adjusts Acoustic Output	Description/Benefit
Sweep Speed	No	To increase/decrease, select Sweep Speed. Changes the speed at which the timeline is swept. Available in M-Mode, Doppler Mode and M Color Flow Mode. You can speed up or slow down the timeline to see more or fewer occurrences over time.

Anatomical M-Mode (AMM) and Anatomical Color M-Mode (ACMM)

Anatomical M-Mode allows manipulation of the cursor at different angles and positions. The M-Mode display changes according to the position of the cursor.

Anatomical M-Mode displays a distance/time plot from a cursor line, which is independent from the axial plane. AMM is available in B, Color and TVI.

NOTE

To set up AMM, go to Utility > Imaging > AMM. Select the specific probe and parameters.

NOTE

Turn off CrossXBeam before activating AMM/CAMM.



Figure 5-14 • Anatomical M Mode

Activating

To activate Anatomical M-Mode, while in M-Mode, press the **Anatomical M** Touch Panel control.

NOTE

AMM is not available on Linear probes.

NOTE

Anatomical M-Mode can also be used with previously acquired digitally stored 2D images. More than one heart cycle should be stored if performing M-Mode in post processing.

To activate Anatomical Color M-Mode, after accessing Anatomical M-Mode, activate Color Flow (CF).

Adjusting

Use Touch Panel Trackball to position the M cursor over the required area of the B-Mode image.

1. Use the trackball (assigned function: Pos) to position the M cursor over the required area of the image.
2. Press the top Trackball key to allow free rotation of the solid arrow line throughout the 2D image (trackball assigned function: Angle).

NOTE

Rotate the Touch Panel control to angle the M cursor at a given angle.

Benefits

Color Flow Mode and Color M-Mode are Doppler Modes intended to add color-coded qualitative information concerning the relative velocity and direction of fluid motion within the B-Mode or M-Mode image.

Bioeffects

Changing the Packet Size, Scale, and ROI size may change the TI and/or MI. Observe the output display for possible effects.

Curved Anatomical M-Mode (Camm)

Curved Anatomical M-Mode (Camm) displays a distance/time plot from a free-drawn cursor line. Camm is available in B, CF and TVI.

1. Select **Curved AMM** on the Touch Panel.
2. Use the Trackball to position the start point of the time motion curve in the B-Mode image.
3. Press **Set** to fix the start point.
4. Use the Trackball to position the next point.

The time motion curve is drawn by the green line.

5. Press **Set** to fix the point.
6. Repeat step 4 and 5 to draw a complete time motion curve.

NOTE

The time motion curve can be edited by following the curve back to the desired point and redrawn as desired. Following the curve back to the starting point will delete the time motion curve.

7. Press **Set** twice to complete.

NOTE

Move the cursor to the desired anchor point and press **Set**. Move the point to the desired position and press **Set**.

8. The arrow cursor appears on the M-Mode image and the red bar appears on the time motion curve.

The red bar indicates the position of the time motion curve relative to the arrow cursor on the Camm image. They move relative to one another.

NOTE

Press **Set** to clear a cursor line.

NOTE

Curved Anatomical M-Mode can also be used with previously acquired digitally stored B-Mode images.

NOTE

Camm is not available on Linear probes.

Optimizing Color Flow

Intended Use



WARNING

Color Flow Doppler is intended for qualitative studies only.

Color Flow Mode is a Doppler Mode intended to add color-coded qualitative information concerning the relative velocity and direction of fluid motion within the B-Mode image.

Information regarding the flow velocity and direction is color-coded and rendered onto a B-Mode image. Flow that travels away from the transducer (negative Doppler shift) is depicted in blue, and flow that is traveling toward the transducer (positive Doppler shift) is depicted in red (colors are user-definable), with lighter shades of each color denoting higher velocities.

Typical Exam Protocol

A typical examination using Color Flow Mode,

1. Follow the same procedure as described under B-Mode to locate the anatomical area of interest.
2. After optimizing the B-Mode image, add Color Flow.

NOTE

Use all noise reduction controls with care. Excessive application may obscure low level diagnostic information.

3. Move the color flow area of interest as close to the center of the image as possible.
4. Optimize the color flow parameters so that a high frame rate can be achieved and appropriate flow velocities are visualized.
5. Press **Freeze** to hold the image in memory.
6. Record color flow images as necessary.
7. If more definitive information is needed about flow, utilize the procedures described under Doppler Mode.
8. To exit Color Flow, press **CF-Mode** or **B-Mode**.

NOTE

Most parameters are user presettable by probe and application in the preset menu (Utility > Imaging > CF).

Activating Color Flow

Color Flow is useful to see flow in a broad area. Color Flow allows visualization of flow in the CF ROI, whereas Doppler Mode provides spectral information in a smaller area.

Color Flow is also sometimes used as a stepping stone to Doppler. You use Color Flow to locate flow and vessels prior to activating Doppler.

Color Flow Mode Controls

Color Flow Mode and Color M-Mode are Doppler Modes intended to add color-coded qualitative information concerning the relative velocity and direction of fluid motion within the B-Mode or M-Mode image.

Table 5-4 • Color Flow Mode Controls

Control	Adjusts Acoustic Output	Description/Benefit
Flow Selection	No	In the Lower Extremity Vein (LEV) and Abdominal applications, you can quickly select the flow state via a shortcut on the Color Flow Mode Touch Panel menu.
Gain	No	Gain amplifies the overall strength of echoes processed in the Color Flow window or spectral Doppler timeline. Allows you to control the amount of color within a vessel or to fill in or clean out spectral information.
Scale (Velocity Scale)	Yes	To raise/lower the velocity scale, adjust Scale. Increases/decreases the Scale on the color bar. Imaging of higher velocity flow requires increased scale values to avoid aliasing.
Wall Filter	No	Filters out low flow velocity signals. It helps get rid of motion artifacts caused from breathing and other patient motion. Gets rid of excess, unnecessary low frequency signals caused by motion.
Size/Position of the color window	No	Adjust size and position of the color window. To adjust the size, press the top trackball key to select Size then move the Trackball left/right, up/down. To adjust the position, press the top trackball key to select Pos then move the Trackball to position the color window. Increase the color window to see a larger area; decrease the color window to improve frame rate and spatial resolution.
Invert (Color Invert)	No	To reverse the color flow, press Invert (Color Invert). Lets you view blood flow from a different perspective, e.g., red away (negative velocities) and blue toward (positive velocities). You can invert a real-time or frozen image. NOTE: Invert reverses the color map, NOT the color Scale.
Baseline	No	To adjust the baseline, adjust Baseline up/down, as necessary. Changes the Color Flow or Doppler spectrum baseline to accommodate higher velocity blood flow. Minimizes aliasing by displaying a greater range of forward flow with respect to reverse flow, or vice versa. Baseline adjusts the alias point. The default baseline is at the midpoint of the color display and at the midpoint of the color bar reference display.

Control	Adjusts Acoustic Output	Description/Benefit
Angle Steer	Yes	To slant the linear image to the left/right, adjust Angle Steer. You can slant the ROI of the Color Flow linear image left or right to get more information without moving the probe. The Angle Steer function only applies to linear probes. Provides a Doppler cursor angle suitable for linear probe orientation. Beneficial in Peripheral Vascular to image carotids.
Accumulation	No	Accumulation enhances the flow in an image. Accumulation detects the maximum signal and holds it for the level specified. NOTE: If Accumulation is turned off, then Frame Averaging is used.
Color Flow Line Density	Yes	Optimizes the Color Flow frame rate or spatial resolution for the best possible color image. Low line density is useful in fetal heartbeat, adult cardiac applications, and clinical Radiology applications which require significantly higher frame rates. High resolution is useful in situations where very small vessels are being imaged, e.g., thyroid, testicles.
Map	No	Allows you to select a specific color map. After you have made your selection, the color bar displays the resultant map. Shows the direction of the flow and highlights the higher velocity flows. Velocity Maps (V). Flow shown as blue away/red toward the probe. Velocity Variance Maps (VV). Provides a measure of turbulence (stenosis). Adds green to velocity maps.
Map Compress	No	When you increase the value, high velocity elements in the map are compressed so that the map darkens. When you decrease the value, low velocity elements in the map are compressed so that the map lightens. The effect is visible in the color bar.
Radiantflow	No	Radiantflow provides an easy, fast visualization of tiny vessels, displaying as a 3D effect.
Threshold	No	Threshold assigns the gray scale level at which color information stops. Limits color flow overlay to low level echoes inside vessel walls. Helps minimize color 'bleeding' outside vessel walls.
Frame Average	No	Averages color frames. Higher frame averaging keeps the color displayed longer for increased flow visualization while lower frame averaging provides greater flow dynamics.
Transparency Map	No	Brings out the tissue behind the color map. Helps demonstrate the tissues behind the color.
Spatial Filter	No	Smooths out the color, makes it look less pixelated.
Flash Suppression	No	Activates/deactivates Flash Suppression, a motion artifact elimination process. Beneficial to suppress flash.

Control	Adjusts Acoustic Output	Description/Benefit
Packet Size	Yes	Controls the number of samples gathered for a single color flow vector. Allows you to improve the color sensitivity and accuracy of color averaging (increase packet size) or frame rate (decrease packet size), as needed.
Power Doppler Imaging (PDI)	No	See <i>Power Doppler Imaging (PDI)</i> on page 200.

Radiantflow



CAUTION

Radiantflow is a display method which uses the amplitude of the Color flow signal. Note that this may alter the appearance of the Ultrasound information displayed. For diagnostic purposes, this must be taken into account or the Region of Interest must be checked without Radiantflow.

Radiantflow is a rendering technique for Color Flow and Power Doppler Imaging, available on all probes. Radiantflow provides an easy, fast visualization of tiny vessels, displaying as a 3D effect. Press Radiantflow on the Color Flow Touch Panel and select the amount of power signal gradient.

- Off. - Normal Color Flow/Power Doppler Imaging.
- Min. - Less gradient.
- Mid. - More gradient.
- Max. - Most gradient.



Figure 5-15 • Radiantflow Map

Flow Model Shortcuts

Flow Model Shortcuts values vary by application. You can configure these Shortcuts on the Utility > Imaging > CF. Below is an example of the Renal Flow Model Shortcuts and the following table lists all the following Flow Model Shortcuts by application.




Figure 5-16 • Application Flow Model Shortcuts

Table 5-5 • Flow Model Shortcuts

Application	Flow Model Shortcuts			
ABD, ABD Biopsy	Aorta	Slow Flow (Renal)	Fast Flow (Renal Art)	Penetration
Renal	Aorta	Twinkle	Fast Flow (Renal Art)	Penetration
OB1, OB23	Fetal Echo	Ovary		
PedAbd	Slow Flow (Renal)		Twinkle	Penetration
NeoHead	Slow Flow			Penetration
NeoAbd	Slow Flow (Renal)			Penetration
Cardiac (Adult, Pediatric)	Fast Frame Rate	Penetration	Slow Flow	
Scrotal	Slow Flow			
Thyroid	Slow Flow	Med Flow	Carotid	
Breast, MskGen	Slow Flow	Med Flow	Fast Flow	Penetration
MskSup	Slow Flow	Med Flow	Fast Flow	Rheuma
LEV	Slow Flow	Fast Flow		
UEV	Slow Flow			
Carotid	Vascular Surgery	Vertebral Arterial (Vert.Art.)		
UEA	Vascular Surgery	Slow Flow		
GYN	Ovary			

Editing Flow Model Shortcuts

1. Via **Utility > Imaging > CF**, highlight the Flow Model line you want to add/update.



Figure 5-17 • Setting up Flow Models

2. Type the Flow Model name in the Name field. The Flow Model is updated as you type.
3. Update Delta Gain (increases/decreases Displayed Gain by the dB selected here), Velocity, Line Density, CF/PDI Frequency for this Flow Model.
4. Set Default Flow Model.
5. Press Save, then Exit.

The updated Flow Model is located in the designed position on the Color Flow Touch Panel.

NOTE

Users can free type custom flow shortcuts into the Scan Assistant Creator to use them within Scan Assistant.

Power Doppler Imaging (PDI)

Power Doppler Imaging (PDI) is a color flow mapping technique used to map the strength of the Doppler signal coming from the flow rather than the frequency shift of the signal. Using this technique, the ultrasound system plots color flow based on the number of reflectors that are moving, regardless of their velocity. PDI does not map velocity, therefore it is not subject to aliasing. A sub-mode of PDI has been optimized for visualization of small slow flow vessels, called Micro Vascular Imaging (MVI).

Adjusting

Press **PDI**. The color flow window appears over the B-Mode image. Move the Trackball to move the CF window. To exit, press **PDI** or select a new mode.

NOTE

Most parameters are user presettable by probe and application in the preset menu (Utility > Imaging > PDI).

Values

On/Off.

Twelve power and four directional PDI maps are available.

Benefits

Since PDI does not display velocity, it does not alias.

Affect on Other Controls

When PDI is activated, the following controls are adjusted: Color Map is set to a power map. Line Density is adjusted. Threshold is set to 100%. Frame Averaging is adjusted. Packet Size is adjusted.

NOTE

These controls are reset to their previous values upon exiting PDI.



HINTS

When changing maps, higher gain settings may be needed.

Table 5-6 • Power Doppler Imaging Controls

Control	Adjusts Acoustic Output	Description/ Benefit
Map	No	Allows you to select a specific color map. After you have made your selection, the color bar displays the resultant map. Shows the direction of the flow and highlights the higher velocity flows. Power Maps. Assortment of Black/White, Pink, Blue, Yellow, Red, or B-Flow Maps. Directional Maps. Directional Power Doppler Maps from PDI. Note: If the image is aliasing while in Directional MVI Maps, increase the Scale. You can set the default map by probe via Utility > Imaging > MVI > Default Map.
Background	No	Displays the amount of background image shown on the MVI image. Use the Background control to display an appropriate amount of tissue background including plaque. 0=Least amount of image displayed 6=Most amount of image displayed

Presets

You can set the default PDI imaging mode via Utility > Imaging > General > Default PDI:

Optimizing the Image



Figure 5-18 • Set Default PDI

Micro Vascular Imaging (MVI)

Smaller vessels can be imaged using Micro Vascular Imaging (MVI), which is a sub-mode of PDI optimized for visualization of small, slow flow vessels. MVI is available under the Power Doppler Imaging (PDI) Touch Panel. It's available on the C1-6-D, C1-6VN-D, L8-18i-D, L2-9-D, L2-9VN-D, L3-12-D, ML6-15-D and L6-24-D probes. MVI is used to detect and visualize micro-vasculature by providing higher sensitivity and higher spatial resolution.

Activating MVI

Access PDI on the Operator Panel, then select MVI on the Touch Panel.




Figure 5-19 • Access MVI from the PDI Touch Panel

Directional Power Doppler

You can select the DPO 0, 1, 2, and 3 Directional Power Doppler maps while in PDI.

NOTE

If you store a PDI image and recall it, you can still switch to the Directional Power Doppler map and vice versa. However, an image stored as non-directional then switched to directional just adds direction to a non-directional map and vice versa.



HINTS

If the image is aliasing while in Directional Power Doppler, increase the Scale.

Optimizing M Color Flow

M Color Flow Mode

M Color Flow is used for cardiac applications. There is a cursor on the B-Mode image that determines the extent of the Color.

The Color Flow maps available in M-Mode are the same as in Color Flow Mode. The size and position of the Color Flow window in B-Mode determines the size and position of the Color Flow window in M-Mode.

All M-Mode measurements are available with M Color Flow active: depth, distance along a straight line, % stenosis, volume, trace, circumference, enclosed area, distance, time, slope, and heart rate.

Activating

To activate M Color Flow Mode, press **M** (M-Mode). Then press **CF** (Color Flow) - or - press **CF**, then press **M**.

To toggle between M Color Flow controls and Color Flow controls, press the appropriate Touch Panel Mode tab.

Benefits

Color Flow Mode and Color M-Mode are Doppler Modes intended to add color-coded qualitative information concerning the relative velocity and direction of fluid motion within the B-Mode or M-Mode image.

Bioeffects

Changing the Sweep Speed, Packet Size, Frame Rate/Resolution, Zoom, PRF, and ROI size may change the TI and/or MI. Observe the output display for possible effects.

Optimizing Spectral Doppler

Intended Use

Doppler is intended to provide measurement data concerning the velocity of moving tissues and fluids. PW Doppler lets you examine blood flow data selectively from a small region called the sample volume.

Typical Use - PW Doppler

In Pulsed Wave Doppler (PW) Mode, energy is transmitted from the ultrasound probe into the patient, as in B-Mode. However, the received echoes are processed to extract the difference in frequency between the transmitted and received signals. Differences in frequencies can be caused by moving objects in the path of the ultrasound signal, such as moving blood cells. The resultant signals are presented audibly through the system speakers and graphically on the system display. The X axis of the graph represents time while the Y axis represents the shift in frequency. The Y axis can also be calibrated to represent velocity in either a forward or reverse direction.

PW Doppler is typically used for displaying the speed, direction, and spectral content of blood flow at selected anatomical sites. PW Doppler operates in two different modes: conventional PW and High Pulse Repetition Frequency (HPRF).

PW Doppler can be combined with B-Mode for rapidly selecting the anatomical site for PW Doppler examination. The site where PW Doppler data is derived appears graphically on the B-Mode image (Sample Volume Gate). The sample volume gate can be moved anywhere within the B-Mode image.

Typical Exam Protocol

A typical examination using PW Doppler Mode might proceed as follows:

1. Locate the anatomy to be examined. Get a good B Mode image. Press CF to help locate the vessel you wish to examine.
2. Press Mode Cursor to display the sample volume cursor and gate.
or
Press PW. The PW Doppler spectrum appears and the system operates in combined B+Doppler Mode. Adjust Volume to adjust Doppler audio. The Doppler signal is heard through the speakers.
3. Position the sample volume cursor by moving the Trackball left and right. Position the sample volume gate by moving the Trackball up and down. Size the gate by clicking SV Length.
4. Optimize the PW Doppler spectrum, as necessary. Refer to the Doppler Optimization section of this chapter for more information.
5. Press **Update** to toggle between real time B-Mode with Doppler Mode (with audio).
6. Sample along the whole length of the vessel. Make sure that the probe is parallel to flow. Listen, then look, when positioning the sample volume cursor.
7. Press Freeze to hold the trace in memory and stop imaging. Activate CINE Timeline, as necessary. See *CINE* on page 231 for more information.

Optimizing the Image

8. Perform measurements and calculations, as necessary. Refer to the Measurements and Calculations chapter for more information.
9. Record results by pressing the appropriate print key, depending on the setup of your recording devices.
10. Press Freeze to resume imaging.
11. Repeat the above procedure until all relevant flow sites have been examined.

Spectral Doppler Display

Time zero (the start of the trace) appears on the left side of the graph. As time progresses, the trace moves to the right. The baseline of the graph (representing zero velocity, zero frequency shift, or no detected flow), appears as a solid line running horizontally across the display. By convention, movement toward the probe is positive and movement away from the probe is negative. Positive frequencies or velocities appear above the baseline. Negative frequencies or velocities appear below the baseline.

Typically, blood flow is not uniform but is composed of a mix of blood cells moving at different velocities and in different directions. Thus, the display is composed of a spectrum as gray scale values. Strong signals are displayed as bright while weak signals are displayed as varying shades of gray.

HPRF (High Pulse Repetition Frequency) is invoked when you are operating in PW Doppler Mode and conditions activate HPRF (when the velocity scale factor or sample volume gate depth exceeds certain limits). When HPRF is active, multiple sample volume gates appear along the Doppler mode cursor. Doppler information can be received from any of the multiple sample volume gates. The Doppler signals from all the gates are added together and displayed in one spectrum.

Information about the PW Doppler display is automatically written on the screen and updated when scanning parameters are changed.

This chapter includes:

- A discussion of PW Doppler.
- Activating Pulsed Wave Doppler.
- Optimizing the Doppler spectrum.

PW Doppler Mode Display




Figure 5-20 • PW Doppler Mode Display

Table 5-7 • Doppler Mode Display Explanations

Doppler Display	Description, Format, Values
Scale	Velocity Scale, displayed as PRF in kHz.
Wall Filter	Wall filter size, displayed as WF in Hz.
Doppler Gain*	Displays as GN in decibels (dB).
Sample Volume Depth	Displays (in Cm) when Doppler cursor is present.
Doppler Angle (AC ##)	Indicates angle in degrees between the Doppler mode cursor and the angle correction indicator. Displays when Doppler cursor is present. The Doppler Angle displays in red when the angle exceeds 60°. Velocities obtained when the angle is greater than 80° are displayed as asterisks (**).
Spectral Invert	INVERT appears when the spectral trace is inverted and the plus/minus signs (+/-) are reversed.
HPRF	HPRF mode is used when detected velocities exceed the processing capabilities of the currently selected PW Doppler scale or when the selected anatomical site is too deep for the selected PW Doppler scale.
Time Scale	Each selection represents a different sweep time.
Angle Correct	Indicates flow direction.
Sample Volume Gate	Indicates sample volume box. Each probe defaults to a specific range gate.
Doppler Velocity Scale	Flow direction has a positive and negative indicator, noted in centimeters per second (cm/sec). When the velocity scale is less than 10 cm/s, it is displayed to the first decimal point (4.6 rather than 5 cm/s). The Doppler velocity scale adjusts as you adjust the Scale.

Doppler Mode Controls

NOTE

You can set the default value of each parameter by probe and application on the Utility > Imaging page. See *Imaging Presets* on page 473 for more information.

Table 5-8 • Doppler Mode Controls

Control	Adjusts Acoustic Output	Description/Benefit
Doppler Sample Volume Gate Position (Trackball)	Yes	To move sample volume gate position, press the top trackball key to select Pos, move Trackball up or down until positioned inside the vessel. Moves the sample volume gate on the B-Mode's Doppler Mode cursor. The gate is positioned over a specific position within the vessel. Positions the sample volume gate to sample blood flow.
Doppler sample volume length (SV Length)	Yes	To increase/decrease the gate size, adjust SV Length on the Touch Panel. Hold down key to continuously size gate. Sizes the sample volume gate. A smaller gate produces accurate sampling results because it is more sensitive. You can also enlarge the gate for sampling large vessels or areas.
Angle Correct/Auto Angle	No	To adjust the angle relative to the probe face, adjust Angle Correct to the left/right. The velocity scale changes when you adjust angle correct. Press Angle Correct to access Auto Angle. Estimates the flow velocity in a direction at an angle to the Doppler vector by computing the angle between the Doppler vector and the flow to be measured. Optimizes the accuracy of the flow velocity. This is especially useful in vascular applications where you need to measure velocity. NOTE: When the Doppler Mode Cursor and angle correct indicator are aligned (the angle is 0), you cannot see the angle correct indicator.
Quick Angle	No	Quickly adjusts the angle by 60 degrees. Press Quick Angle to toggle between Off, Right and Left.
Steer and Fine Steer	Yes	To slant the linear image to the left/right, adjust Steer to the left or right. Press Steer to access Fine Steer. You can slant the ROI of the Color Flow linear image left or right to get more information without moving the probe. The angle steer function only applies to linear probes. Provides a Doppler cursor angle suitable for linear probe orientation. Beneficial in Vascular applications.

Control	Adjusts Acoustic Output	Description/Benefit
Audio Volume	No	Controls audio output. An audio representation of the flow within a vessel can be used to evaluate proper probe angle and position.
Cycles to Average	No	The average value over a number of cycles (from 1-5).
Display Format	No	Changes the horizontal/vertical layout between B-Mode and M-Mode, or timeline only.
Update	Yes	To activate, press Update to toggle between simultaneous and update. Doppler Mode does not restart each time the image is updated; however, a black bar may appear with a lightning bolt signaling a break in the timeline. Toggles between simultaneous and update presentation while viewing the timeline. Update increases the Spectral Doppler display quality.
Baseline	No	Adjusts the baseline to accommodate faster or slower blood flows to eliminate aliasing.
Compression	No	Compression controls how echo intensities are converted to shades of gray, thereby increasing the range of contrast you can adjust. Optimizes the image's texture and smoothness by increasing or decreasing the amount of gray scale.
Invert	No	Vertically inverts the spectral trace without affecting the baseline position. The plus (+) and minus (-) signs on the velocity scale reverse when the spectrum is inverted. If you change the probe angle to accommodate anatomy, blood flow still moves in the same direction, but the Doppler information will be reversed. It is easier in cases like this to invert the spectrum instead of reversing the probe orientation.
Scale (Velocity Scale)	Yes	To view signal detail, adjust Scale to enlarge the vertical spectral Doppler trace. Velocity range directly controls the pulse repetition frequency, which is responsible for the setting of the Nyquist limit (the ability to detect maximum velocity without aliasing). If the sample volume gate range exceeds single gate Scale capability, the system automatically switches to high PRF mode. Multiple gates appear, and HPRF is indicated on the display.
Trace Method (Spectral Trace)	No	To get a peak trace, click MAX. A green trace displays on the spectrum. To get a mean trace, click MEAN. A blue trace displays on the spectrum. Traces the average mean and peak velocities in realtime or frozen images. Lets you trace the cardiac cycle.
Trace Sensitivity	No	Adjust the trace to follow the waveform for signal strength. If the signal is very faint, increasing the Trace Sensitivity will allow the system to trace that signal strength.

Optimizing the Image

Control	Adjusts Acoustic Output	Description/Benefit
Trace Direction	No	Specifies trace direction. You can select where on the waveform to perform the trace, above, below, or both (above and below).
Cursor Moving	No	On Utility--> Imaging, specify No Action, Update 2D/CF-Long, Medium, or Short, or Update Doppler-Slow, Medium, or Fast. Cursor Moving lets you 'walk' Doppler through a vessel while the Doppler gate is moving. Updates are more frequent on Fast vs. Medium vs. Slow. If you set the preset to Update 2/D/CF, this causes the B Mode/Color Flow image to go live while you move the Doppler cursor.
Time resolution	No	Adjusts image appearance so that if you select a lower setting, the image appears smoother; if you select a higher setting, the image appears sharper.
Wall Filter	No	To increase/decrease, select Wall Filter on the Touch Panel. Insulates the Doppler signal from excessive noise caused from vessel movement. Gets rid of excess, unnecessary information. Cleans out low level noise above and below the baseline so you don't see or hear it on the spectrum.
Auto (Auto Optimize)	No	To activate, press Right Auto key. Press Left Auto key to turn off Auto. Auto in PW Doppler Mode (ASO: Auto Spectral Optimization) optimizes the spectral data. Auto adjusts the Velocity Scale/PRF (live imaging only), baseline shift, dynamic range, and invert (if preset). "Running Auto Spectral Optimization" appears at the bottom of the monitor upon activation. Upon deactivation, the spectrum is still optimized. See <i>Continuous Auto Tissue Optimization</i> on page 174 for more information.
Mode Cursor	No	Displays the Doppler Mode cursor on the B-Mode image. To activate/deactivate the Doppler Mode cursor, press Mode Cursor. Trackball to position sample volume graphic.
Modify Auto Calcs	No	Activates the menu to select which calculations are automatically calculated.
Auto Calcs	No	Activates the calculation automatically which you select in the Modify Auto Calculation when the system is in a state of freeze or live. Live: Auto calculation activates when the system in a state of live. Freeze: Auto calculation activates when you press Freeze. Off

Auto Doppler Assist

Auto Doppler Assist automatically positions and steers Color Flow ROI and PW Cursor. This to feature is available on all linear probes. Available in Color Flow, Color Flow + PW (Triplex and non-Triplex).

To use Auto Doppler Assist, position the cursor in the Color Flow ROI (the Color Flow ROI has to include the vessel of interest, or part of it). Press the appropriate Touch Panel Auto control (shown below) or press the Auto control to the left for Color Flow or to the right for Pulsed Wave. Manual map the “Auto” controls via **Utility > User Configurable Key**.



Figure 5-21 • Auto Doppler Assist

Upon pressing the Auto Doppler Assist button, the system will automatically:

- Pick the artery or vein (depending on the selected application).
- Center the Color Flow ROI on the vessel of interest.
- Steer with the direction of the vessel.
- Keeps the Sample Volume in the middle of the ROI.
- Steers the PW Cursor (if existing) to maintain the angle correct set by the operator under Auto Correct on the **Utility > Imaging > PW > Angle Correct**.

NOTE

If the angle is set to zero, then the system uses 60 degrees.

Using 3D



DANGER

DO NOT use the Volume Navigation feature on any patient relying on life-sustaining electronic equipment, such as a pacemaker or defibrillator. Failure to follow this instruction could lead to interference with patient electronic device(s).

Easy 3D is compatible with every 2D transducer using a freehand acquisition to generate a volume dataset.

3D Volume datasets are allowing the navigation in the 3D cube itself and providing access to the 3 different main planes – axial, sagittal and coronal.

There are three 3D Packages:

Table 5-9 • 3D Package Options

3D Type	Description	Sensor/No Sensor
Easy 3D	Designed for rendering B Mode and Color Flow Mode images, e.g., Baby Face scans.	No sensor
Advanced 3D	Designed for rendering B Mode and Color Flow Mode images, e.g., vessel trees.	No sensor
Tru3D	Designed for rendering B Mode and Color Flow Mode images, e.g., vessel trees.	Sensor

3D Acquisition

Acquiring a 3D Scan

To acquire a 3D scan:

1. Optimize the B-Mode image. Ensure even gel coverage.
2. Press the 3D/4D control panel key. Two screens appear.
3. Set appropriate values for Acq Mode and Scan Plane. Also, set the scan distance before scanning.
 - **Acquisition Mode**
Sensorless Parallel is for all acquisitions done with the linear probes and on regular shapes, where you can move the probe parallel on the skin.
Sensorless Sweep is for the sweep acquisition using the curved probe (i.e. intercostal liverscan or kidney).
 - **Scan Distance**
The Scan Distance is an indicator for the size of the Volume: Have you acquired longer distance than 6 cm, increase the Scan distance. Have you acquired a shorter distance than 6 cm, decrease the Scan distance. For sweep acquisition 6 means a transducer angulation of around 60°.

4. To start acquiring the image, press Start (Trackball key).
5. To perform a parallel scan, scan evenly. To perform a sweep (fan) scan, rock the probe once. Note the distance of the scan.
6. The 3D volume of interest (VOI) is dynamically assembled on the right side of the screen.

NOTE

If the image stops before you're done scanning, start acquiring the 3D volume of interest again.

7. To complete the 3D scan, press End (Trackball key).

NOTE

You can also press Freeze, but then you need to also press the 3D key to obtain the final render.

3D Notes

- Adjust the 3D dataset brightness with B-Mode or Color Flow Mode Gain.
- Use Colorize to change the color of the active dataset.
- Use Zoom to increase the zoom factor of the active dataset.
- Vertical lines may be seen in a resliced image. This usually happens when you scan too fast or if the scan distance is set to a high value.

Scan more slowly, adjust the frame rate for a faster rate or adjust the scan distance.

3D Acquisition Parameter Description




Figure 5-22 • 3D Acquisition

Table 5-10 • 3D Acquisition Description and Instructions for Use

3D Parameter	Description
App (Application) Presets	Selections: None, OB - Baby Face, Vascular, User 1, User 2, User 3 None. No application preset applied. OB - Baby Face. After having scanned in this mode, certain rendering parameters are set automatically. The gray surface mode is activated and the texture mode is switched off. The gray surface mode values for opacity and threshold are set automatically according to the datasets histogram. Vascular. Available only with Advanced 3D or Tru 3D package. After having scanned in this mode, certain rendering parameters are set automatically. The color image is rendered in the texture mode. The values for opacity and threshold of the texture mode are set automatically according to the datasets histogram. The B-Mode image is rendered in the gray surface mode. Opacity and threshold values are defined according to the histogram.
Delete Active User Preset	Select to delete a user preset (User 1, User 2, or User 3).
Acquisition Mode	Selections: Sensorless Parallel, Sensorless Sweep Sensorless Parallel. In this mode the probe must be moved during 3D data acquisition without angling it. You should scan the object you want to render in 2-4 seconds. The speed at which you scan should be constant. No sensor is mounted on the probe. Since the time for post-processing depends on the acquired number of frames, it is recommended that you check the frame rate. Low frame rates result in fewer acquired frames for the 3D dataset which results in intensive post-processing (interpolation). Therefore, low frame rate = long post-processing. Sensorless Sweep. In this mode the probe must be moved to a position where you can clearly see a middle cut of the object you want to scan and render. Tilt the probe to about 30 degrees until the object you want to scan disappears. Start the acquisition and tilt the probe over a distance of around 60 degrees until the object disappears again. The entire scan time should be around 2-4 seconds. During the sweep, the probe may not be moved parallel, just tilted. No sensor is mounted on the probe. Before starting an acquisition, take care that the transmitter is positioned correctly during data acquisition and that the transmitter cannot move.
Scan Plane	Selections: Front to Back, Side to Side Front to Back. After having scanned in this mode, the rendered dataset is shown in a frontal view. For acquiring a fetal face in sagittal cuts, use this mode. Side to Side. After having scanned in this mode, the rendered dataset is shown from a side view. For acquiring a fetal face in coronal cuts, use this mode.
Display Format 50/50 Only 2D	50/50. Display in Dual Image (2D and 3D). Only 2D. Display in Single Image.
3D	Starts the rendering process.

3D Parameter	Description
Scan Distance	Adjusts the distance covered during the scan. Depending on the real width of a scan acquired during a sensorless 3D acquisition, the volume of interest's width can be enlarged or reduced. You can adapt the form of a fetal face if the baby's head looks oval instead of round. The assumed default width of a parallel scan is 6 cm; or a fan scan 60 degrees.

NOTE

The selection of user presets is effective only while 3D mode is active. Exiting 3D mode and activating 3D mode again resets the 3D presets to the default setting, regardless if Patient or application changes.

NOTE

Changing the tab between [Easy] and [Adv3D] changes some parameters that are not common between those tabs.

NOTE

When a 3D image is recalled, no 3D presets are active and parameters are recalled from the image file.

NOTE


Default Scan Distance, Opacity and Threshold may not be consistent and may change per scan. After the User Preset is saved and recalled, Opacity and Threshold are consistent.

Manipulating the Volume of Interest




Imagine you are able to manipulate the 3D volume of interest (VOI) in your hand. The 3D VOI is a tangible anatomical object that you can see and manipulate easily using the Trackball and Set control panel keys.

Practice positioning the pointer at different places within the 3D VOI. Highlight different colors (white, red, yellow, or green). Press Set to select a VOI for manipulation. Use the hand to manipulate the 3D VOI.

Table 5-11 • Manipulating the Volume of Interest

Procedure	Example
Rotating the 3D VOI Left/Right or Forward/Backward You can rotate it left to right or right to left. You can rotate it forward/backward. Press right Set key when the white pointer finger is positioned on the white box. Move the closed white hand to manipulate the 3D VOI.	

Optimizing the Image

Procedure	Example
Moving Through the 3D VOI You can move through the 3D VOI using the red hand. Press Set when the red pointer finger is positioned on the red box. Move the closed red hand to move through the 3D VOI. Note: Any plane in the volume can be made active (highlighted with red box) by clicking on it.	
Viewing Specific Portions of the Anatomy You can pull back tissue to view specific portions of anatomy using the yellow hand. Press Set when the yellow pointer finger is positioned on the yellow box. Move the closed yellow hand to manipulate the 3D VOI. Note: This actually moves an edge. A yellow hand appears only when the pointer is on an edge of the VOI.	
Pulling Back a Corner of the VOI to View Specific Anatomy You can pull back a corner to view specific portions of anatomy using the green hand. Press Set when the green pointer finger is positioned on the green box. Move the closed green hand to manipulate the 3D VOI.	

Easy 3D




Figure 5-23 • Easy 3D

Table 5-12 • Easy 3D Description and Instructions for Use

3D Parameter	Description
Reset	Resets the 3D volume of interest back to its original orientation.
Utilities	Select Average Off, Average Light, Average Medium, or Average Strong. Use smoothed volume for rendering the 3D volume. Strong = Most Smoothing.
Auto Movie	Initializes the calculation and display of a 3D movie. A rotation of 30 degrees left and right around the actual image position (either the default position after acquisition or the position that was manually defined by manipulating the 3D volume of interest) is shown. For this 60 degree rotation, eleven images in steps of 6 degrees are calculated.
Save Preset	Save as a user preset (User 1, 2, or 3).
Undo	Undoes any manipulation you have done to your 3D dataset.
Scalpel	Structures, for example a part of the placenta hiding the view to a fetal face, can be cut out in a rendered image. All visible structures can be cut out. The option of 'erase inside' deletes all structures inside the marked region. The option of 'erase outside' deletes all structures outside the marked region. The region in the rendered image is marked with the right Set key. To define the contour of the region, press the right Set key for each vertex. To close the contour, double click the right Set key. As long as a contour is not closed, it can be traced back with the left Set key. The cut out process can be undone by the Undo Last function. As soon as the Apply button is pressed, a new dataset is generated.
Gray Surface	Activates the gray surface rendering mode. It leads to a transparent appearance of the object, generated by displaying only a surrounding shell of structures.

3D Parameter	Description
Texture	Activates the texture or photorealistic rendering mode. It creates a photorealistic appearance of the object. The shading depends on the orientation of the surface of the object. If both Texture and Gray Surface mode are switched on, the mixture percentage of both modes can be defined.
Render	Changes between the rendered image view and the view of a volume of interest. The volume of interest shows the acquired ultrasound images transformed into an isotropic rectangular coordinate system. The volume of interest can be manipulated as described above.
Orientation Marker	You can now specify/define, then add the following orientation markers while in 3D via the Orientation Marker key: • TRV Sup to Inf • TRV Inf to Sup • SAG Lt to Rt • SAG Rt to Lft • Defined • None
Threshold/Opacity	Threshold defines which gray values are used for rendering and which are considered noise. Opacity defines how strict Threshold is used for discrimination. A low opacity value creates a firmer appearance of the surface. A high opacity value leads to a transparent appearance of the rendered image.
Scan Distance	Adjusts the distance covered during the scan. Depending on the real width of a scan acquired during a sensorless 3D acquisition, the volume of interest's width can be enlarged or reduced. You can adapt the form of a fetal face if the baby's head looks oval instead of round. The assumed default width of a parallel scan is 6 cm; or a fan scan 60 degrees.
Colorize/Contrast	Colorizes the 3D render or adds contrast to the 3D rendered image.

Advanced 3D



Figure 5-24 • Advanced 3D

Table 5-13 • Advanced 3D Description and Instructions for Use

3D Parameter	Description
Tile	The display can be divided into 1, 2, 4, or 6 windows. Switching to a lower number of windows keeps the images from left to right.
3D Landscape	Shows a combination of 2D slices and a 3D rendered image. After a color acquisition you can combine the 2D B-Mode image slices with a 3D rendered color image. This mode allows stepping through the B-Mode images along a vessel structure. The 2D slice can be moved with the right Set key. The Trackball symbol has to be positioned inside the 2D plane.
Active Data	Manipulations of rendering parameters only have an effect on the data defined as Active Data. After having selected Active Data, a list of data is displayed, Gray Data or Inversion. Choose the data to be manipulated. Active Data is only available when you select both Inversion and Gray Data in Visible Data. NOTE: Inversion Mode is only available for Black-and-White mode.
Visible Data	After selecting Visible Data, a list of data is displayed, Gray Data or Inversion. Choose the data you want to display. For example, if only Inversion is chosen, the B-Mode image is switched off in the rendered image and only inversion mode is displayed.
Define Axis	For certain display and measurement modes (Angular Plane Mode, Angular Volume Measurement Mode), an axis in the volume of interest is required. To define the axis, set the start point by using the Trackball to position one end of the axis and pressing the right Set key, then positioning the other end of the axis and pressing the right Set key.

3D Parameter	Description
Group Planes	Selections: Off, Main, Parallel, Angular Off. A VOI or a rendered image is displayed. The Render button changes between the rendered image view and the view of the VOI. The VOI shows the acquired Ultrasound images transformed into an isotropic rectangular coordinate system. Main. Three orthogonal cuts (with colored frames) of the acquired VOI are displayed after pressing Main. The VOI shows the acquired ultrasound images transformed into an isotropic rectangular coordinate system. On the left top of the image a complete VOI is displayed. It shows the position of the three orthogonal planes in the VOI. A green point displayed in each plane defines the point of intersection of the three planes. This point can be set to different positions in the planes by double clicking the right Set key. A plane can be moved parallel in the VOI by pressing the right Set key on the position of the green point and moving the Trackball up and down inside the plane. Parallel. In this mode all displayed VOIs get the orientation of the last modified volume. Normally four VOIs are displayed. It is possible to display six VOIs by increasing the number of displayed volumes in the Tile area. Between the first and the last VOI, the selected planes are parallel and equidistant. A modification on the plane in one VOI results in a parallel modification of the planes in all other VOIs. Angular. Before starting the Long Axis Rotation Mode, make sure that a long axis has correctly been defined in the VOI (see Define Axis above). The function starts in the long axis display mode. In the upper, left-hand corner a short axis cut is shown which gives an overview of the orientation of the long axis planes. To move these planes, press and hold down the right Set key while moving the Trackball.
Type 1/2	Defines the rendering modes. Selections: Gray Surface, Texture, Maximum Intensity, Minimum Intensity, Average Intensity, and None. If both Type 1 and Type 2 rendering modes are switched on, the mixture of both modes can be defined. Gray Surface. Activates the gray surface rendering mode. It leads to a opaque appearance of the object, generated by displaying only a surrounding shell of anatomical structures. Adjust Threshold and Opacity as well. Texture. Activates the texture of photorealistic rendering mode. It creates a photorealistic appearance of the object. The shading depends on the orientation of the surface of the object. Adjust Threshold and Opacity as well. Maximum Intensity. Transparent appearance of the object. Generated by displaying the maximum gray values in the VOI. Minimum Intensity. The rendered image is generated by displaying the lowest gray values in the VOI that exceed the defined threshold. Dark anatomical structures, like cysts, can be shown in this mode. Average Intensity. Transparent appearance of the object. Generated by a summation of the gray values. None for Type 2. No second rendering mode is used in addition to the Type 1 rendering mode.

3D Parameter	Description
Render	Changes between the rendered image view and the view of a volume of interest. The volume of interest shows the acquired ultrasound images transformed into an isotropic rectangular coordinate system. The volume of interest can be manipulated as described above.
Reslice	Cube. The VOI shows the acquired ultrasound images transformed into an isotropic rectangular coordinate system. This mode allows you to work simultaneously with six cut planes. Virtual Rescan. The marked cut planes under Reslice Cube (red border) is displayed without any perspective distortions, e.g., parallel to the screen. This allows you to move through the volume one slice at a time in any direction. Cubic Plane. Only one cut plane view is shown in a perspective displayed VOI. The cut plane can be moved freely without any limitations.

Movie 3D




Figure 5-25 • Movie 3D

Table 5-14 • Movie 3D Descriptions and Instructions for Use

3D Parameter	Description
Manual Define Start/End	An animated rotation of the rendered image can be calculated and displayed by this function. Using this function, you first need to define the start and end position of the rotation. To define this, move the VOI to the start position, then press Define Start. Move the VOI to the end position and press Define End.
Movie 360 Degrees	The calculation and display of a complete rotation around the axis, defined by the Axis button, starts in steps of 15 degrees.

3D Parameter	Description
Auto Movie	Initializes the calculation and display of a 3D movie. A rotation of 30 degrees left and right around the actual image position (either the default position after acquisition or the position that was manually defined by manipulating the 3D volume of interest) is shown. For this 60 degree rotation, eleven images in steps of 6 degrees are calculated.
Axis	All rotations (Auto Move and Movie 360) are calculated as rotations around the specified axis (X, Y, or Z).
Movie Speed	You can adjust the speed of any 3D rotation.
Pause	Stops and restarts the rotation. As soon as Pause is pressed, the different rotation steps can be displayed by moving the Trackball.

Tru3D and Volume Measurement



Figure 5-26 • Tru3D

Descriptions and instructions for using Tru3D are the same as noted in the Advanced 3D section (see *Table 5-13 • Advanced 3D Description and Instructions for Use* on page 219).

Performing a Sensor Scan

To perform a sensor scan:

1. Attach the sensor cables to the front of the Ultrasound system.
2. Attach the probe bracket to the probe, unless using a VNav Inside (VN) probe.
Insert the receiver into the bracket on the probe.
3. Position the magnet near the patient, within 18 inches of the probe, next to the patient on the bed.
4. On the Touch Panel, select With Sensor.
5. Acquire the 3D scan.

Tru3D

Table 5-15 • Tru3D Descriptions and Instructions for Use

3D Parameter	Description
Acquisition Mode	With Sensor. Only available if you are using the Tru 3D package. In this mode the probe can be moved in different ways over the object you want to render. The probe can be moved parallel and angled during the same 3D acquisition. The sensor system registers any movement. It is recommended not to move in one direction and back again during one scan. It is also not recommended to turn the probe around its axis. The scanning speed does not have to be constant, but you should not change the velocity too much. In this mode the sensor has to be mounted on the probe. Before starting an acquisition, take care that the transmitter is positioned correctly during data acquisition and that the transmitter cannot move. The transmitter has to be placed in such an orientation that the probe with the receiver attached is always in the forward hemisphere of the transmitter during scanning. The patient MAY NOT move during the acquisition. The sensor device consists of an electromagnetic field transmitter and a field receiver. The field transmitter generates an alternating spheric electromagnetic field in a strength of up to five times the earth's magnetic field, depending on the distance between the transmitter and the receiver. For 3D data acquisition, the probe with the attached position sensor can be moved freely in an area of 70 cm around the transmitter. During data acquisition, the electromagnetic sensor device generates with a frequency of about 100 Hz a set of three translation and three angle values. These values describe the position of the ultrasound probe in space.

Magnetic Field Range from the Transmitter




Figure 5-27 • Magnetic Field Range from the Transmitter

From the center of the transmitter, the magnetic field ranges:

- 30cm above and below
- 28cm left and right
- 20 - 66cm forward

NOTE

Remove the metal product from the region of the magnetic field to increase accuracy of position detection.

Volume Measurement

The average measurement error for distance measurements is 5% and for volume measurements is 10%.

Workflow Example

NOTE

If you print out a Volume Measurement image with a B/W printer or store it onto the Clipboard, change the following presets before performing the Volume Measurement. On the **Utility > Connection Manager > Print Button > Configure Print Button**, choose Volume File Format* = 2 - Standard DICOM with Raw Data and in the Still Images section, choose Format = Secondary Capture Image.

1. Scan in 2D-mode. Select Auto Sweep. Acquire and store the image.
2. Recall Image. Activate **3D/4D**.

3. Select **Vol. Meas** Tab. Select **Angular Method**.



Figure 5-28 • 3D Volume Measurement

NOTE

Angular method is intended for spherical objects. Serial method is intended for rectangular objects.

4. Define the axis and press **Enter**.
5. Six cut planes are displayed. You measure the volume by marking the contour of the anatomy.
6. When all traces are completed, the system displays the volume in the Results Window.

Measurement examples are shown below:



Figure 5-29 • 2D Measurement (Example)




Figure 5-30 • Example of Segment Method




Figure 5-31 • Example of Angular Method

Table 5-16 • Volume Measurement Descriptions and Instructions for Use

3D Parameter	Description
2D	Type of 2D measurement: Distance, Angle, Circumference, Area Distance. Set the start and end distance using the right Set key. Angle. An angle is measured by marking two lines. Press the right Set key once marks the beginning of the first line. Press the right Set key again to mark the end of the first line and simultaneously the beginning of the second line (the intersection of two lines). Press the right Set key a third time to close the angle measurement. Circumference. The circumference of an area is measured by marking the contour of the anatomy by a polygon contour. Each point of the polygon is marked by pressing the right Set key. A double click of the right Set key closes the circumference. As long as the contour is not finished, each point can be deleted by a single press of the left Set key. Area. An area is measured by marking the contour of the anatomy by a polygon contour. The area of an area is measured by marking the contour of the anatomy by a polygon contour. Each point of the polygon is marked by pressing the right Set key. A freehand contour can be drawn by pressing and holding down the right Set key and moving the Trackball. Double click the right Set key to complete the area. As long as the contour is not finished, each point can be deleted by a single press of the left Set key.

3D Parameter	Description
Angular Method	This function allows you to mark any volume of interest in the dataset to measure its volume or to perform a segmentation of the object. The volume of a 3D object is determined by drawing the contour in several planes, which are rotated around an axis defined by the user. The contours are used to calculate the volume of the object. To determine the volume of an object based on the multiplanar Simpson rule, you need to define the rotation axis via Define Axis. When you press Angular Method, six cut planes are displayed. The Long Axis is marked in blue. You measure the volume by marking the contour of the anatomy. The contour can be marked in different ways: Polygon, Spline, Ellipse, Rectangle, or Rotate. Polygon. Each point of the polygon is marked with a single press of the right Set key. A freehand contour can be drawn by pressing and holding down the right Set key and moving the Trackball. Double click the right Set key to complete the area. As long as the contour is not finished, each point can be deleted by a single press of the left Set key. Curve. An area can be marked by single points positioned by pressing the right Set key. A double click on the right Set key closes the spline. The position of the points defining the contour can be changed by clicking on the point and moving it by pressing the right Set key and using the Trackball. As long as the contour is not finished, each point can be deleted by a single press of the left Set key.
Angular Method	Ellipse. When you select this mode, a circle is displayed. You can move the circle by pressing and holding down the right Set key while moving the circle with the Trackball. Press the right Set key to set the chosen position. To manipulate the shape of the circle, move the edges of the circle while pressing the right Set key. Rectangle. When you select this mode, a rectangle is displayed. You can move the rectangle by pressing and holding down the right Set key while moving the rectangle with the Trackball. Press the right Set key to set the chosen position. To manipulate the shape of the rectangle, move the edges of the circle while pressing the right Set key. Rotate. Using the rotate function, you can rotate an area around the Z axis. When you select this function, the Trackball symbol changes as soon as it is positioned to an edge of a region. The region can then be rotated by pressing and holding down the right Set key. A region can be selected and deleted by pressing the Clear key. To save a measurement, press Save Segment, Save to Report, or Cancel.

3D Parameter	Description
Serial Method	The Serial Method allows you to mark any volume of interest in the dataset to measure its volume or perform a segmentation of the object. A volume definition is done by defining areas at different depths. Before starting the volume measurement, you need to select a plane showing a cut in which the object can be clearly defined. When you press Serial Method, the display window shows two different views. The left side displays the active plane as a single plane. The VOI on the right side is displayed in cubic mode. In the right VOI select the cut plane position where the measurement process should start. In the left plane, mark the object of interest by designating one of the area definition modes (Curve, Ellipse, Rectangle, Polygon, Rotate). After you have completed defining the first area, the depth of the VOI should be changed on the right side. To change the depth, position the Trackball symbol inside the plane to be moved. press and hold down the right Set key while moving the Trackball backwards. By defining the contour of an object at different depths, its volume can be calculated by summing up the defined slices. To save a measurement, press Save Segment, Save to Report, or Cancel.
Define Axis	For certain display and measurement modes (Angular Plane Mode, Angular Volume Measurement Mode), an axis in the volume of interest is required. To define the axis, set the start point by using the Trackball to position one end of the axis and pressing the right Set key, then positioning the other end of the axis and pressing the right Set key.
Save Segment	After measuring a volume, you can use the defined volume for segmentation by pressing Save Segment. Segmentation means that a new dataset is created with voxel information based on the defined volume. A dataset containing only voxels inside the measured volume is created. The original dataset is saved additionally to the segmented data. The segmented data can be chosen in the Active Data or Visible Data list.
Save to Report	After measuring a volume, press Save to Report to register the measurement result in a database that is used for report generation.

Chapter 6

Scanning/Display Functions

In this section

Freezing an Image.	230
CINE.	231
Image Zoom.	239
Split Screen.	242
Annotating an Image.	247
Using the Fast Key.	262

Freezing an Image

Freezing a real-time image stops all movement and allows you to measure and print the image.

NOTE

While the image is frozen, all Power Output is suspended.

To freeze an image:

1. Press **Freeze**. The Freeze key backlight turns blue.

If you are in a mixed mode, both screen formats stop immediately. Deactivating Freeze restarts both modes and places a black bar on the trace to indicate the time discontinuity.

To reactivate the image:

2. Press **Freeze** again.

NOTE

Selecting a new probe unfreezes the image.

NOTE

Deactivating Freeze erases all measurements and calculations from the display (but not from the worksheet).

Use the Trackball to start CINE after pressing **Freeze**.

CINE

CINE images are constantly being stored by the system and are available for playback or manual review via CINE.

You can view CINE as a continuous loop via CINE Loop or manually review CINE images frame by frame via the Trackball.

Data in CINE is available until new data is acquired. CINE is stored on the system's memory and can be archived as well.

CINE is useful for focusing on images during a specific part of the heart cycle or to view short segments of a scan session.

Cine Gauge



Figure 6-1 • Cine gauge

- 1 Loop speed
- 2 Cine gauge
- 3 Current frame number/total frame number. The cine gauge indicates which frame you are viewing of the whole loop.
- 4 Current number of seconds/total number of seconds. The cine gauge indicates which frame you are viewing of the whole loop.
- 5 Start frame
- 6 End frame

Touch Panel

The following CINE Touch Panel appears:




Figure 6-2 • CINE Touch Panel

Table 6-1 • CINE Touch Panel Description

Parameter	Description
Select All	Select all frames of the cineloop.
Sync Mode	Phase synchronizes multiple cineloops.
Num Cycles	Control the number of heart cycles to be included in the cineloop.
First	Move to the first frame of cineloop.
Last	Move to the last frame of cineloop.
Run/Stop	Start/Stop the cineloop review.
Loop Speed	Adjust the cineloop playback speed.
Cycle Select	Select the heart cycle to review.
Start Frame	Rotate the rotary to select the start frame and push to set the frame.
End Frame	Rotate the rotary to select the end frame and push to set the frame.
Frame by Frame	Review the cine image frame by frame manually.
Capture	Searches through all images between the start frame and end frame and displays each peak or the highest velocity/tissue power.
Capture Recon.	Capture Recon. reconstructs small blood vessels in the cine mode.
Enhancement	Execute the enhancement to the Cine capture image. Select Off, Shade FW, Shade Rv, Enhance1, Enhance2 or Enhance3.
Motion Correction	Motion correction compensates each frame and cancels the motion.

Using CINE

NOTE

Preset the parameters as necessary.

1. Press **Freeze**.

2. Move the Trackball to activate Cine.
3. Use the trackball or **Frame by Frame** to scroll through the acquisition and find the sequence of interest.
4. Press **Start Frame** or **End Frame** to set the corresponding cineloop boundary to the current frame as necessary.
Rotate **Start Frame** and **End Frame** to trim or expand the cineloop boundaries.
5. Adjust **Cycle Select** to move from heart beat to heart beat and select the heart cycle of interest.
6. Adjust **Num cycles** to select the number of heart beats to play back.
7. Press **Run/Stop** to run the cineloop and then press the print key to store the cineloop.
Cine loops stored on the clipboard are indicated with a movie strip icon.
8. Press **Run/Stop** again to stop the cine loop.
9. Press **Freeze** to return to live scanning.

NOTE

If you don't need edit, only press Run/Stop to run the cineloop and print key to store.

NOTE

Cineloop storage can be configured to store heart cycles with additional time before and after the R-wave and to display a preview before storage. See *Print Controls* on page 489 for more information.

Adjust the CINE Loop Speed

Rotate Loop Speed to set the speed of the cineloop playback.

The speed factor (%) is displayed above the cine gauge.

To View a Cineloop Frame by Frame

In freeze, use the trackball or **Frame by Frame** to scroll through the cineloop frame by frame.

Synchronize CINE Loops

1. Recall stored cine loop to right side of dual screen.
2. Recall same cine loop to left side of dual screen.
3. Change visualization of left side image
4. Select **Sync mode** to start the synchronization.

NOTE

This is useful for Hybrid Contrast to display and check Contrast and Hybrid Contrast for example.

Recalling a cine loop

To recall a cine loop, double click on the cine loop on the clipboard.

NOTE

CINE Loops stored on the Clipboard are indicated with a movie strip icon.

CINE Mode Selection

To scroll the B-Mode cine loop only, toggle the top Trackball key and select **Scroll B**.

To scroll the Timeline cine loop only, toggle the top Trackball key and select **Scroll D**.

Velocity Scale with B-Mode Only

If you review the B-Mode cine loop while in Doppler Mode with the Timeline using Scroll B, the Velocity Scale displayed with the Timeline is for the time phase of the currently-displayed B-Mode image, NOT for the time phase of the acquired Doppler Spectrum.

Check the velocity value with the measurement function if you review the cine loop using Scroll B. Note that there may be a discrepancy between the velocity scale displayed and the velocity measured using the measurement function.

Mark CINE

Preset

1. Check **Enable Mark Cine Control** via Utility > Application > Print Controls.




Figure 6-3 • Mark Cine on Print Control

Table 6-2 • Mark Cine

Preset Parameter	Description
Enable Mark Cine Control	Lets you mark where you want the CINE Loop to start (prospective CINE).
Preview Loop Longer than... (s)	When selected, allows you to review cine loops before storage for loops longer than selected time frame (in seconds).

2. Press **Save**.

Using Mark CINE

1. Press **Mark CINE** to specify the starting point for loop storing or cine review.
When you press **Mark CINE**, the current image frame is noted as a start frame.
2. Press the appropriate print key while continuing to live scan to store the cine loop.

NOTE

The Mark CINE control on the Trackball key is available while live scanning in non-timeline modes (B-Mode, B-Flow and Color Flow Mode).

NOTE

Selecting Mark CINE when a Mark CINE already exists causes the new Mark CINE to replace the previous one.

NOTE

Changing modes or other actions that flush CINE memory causes the Mark CINE to be removed and the image data will not be saved.

NOTE

A Print button can be configured to store a Single Image during Mark CINE, without stopping the CINE loop.

Preview

Loop Preview can be enabled independently for Time-Based Store, ECG-Based Store, and Mark CINE. This is useful for setting preview preferences based on the application.

NOTE

The Contrast Time Span setting overrides the Time Span when in Contrast Mode.

Background Store

Live Clips are stored in the background to allow you to continue scanning. This works for both Raw Data and for DICOM Loops (with Direct Store On or Off). Image Ordering is preserved with Background Store.

The benefit of Background Store is that clips are stored with minimal interruption to live scanning.

NOTE

Background Store IS NOT supported with V Nav, 4D, or with previously-acquired CINE Loops.

NOTE

The system may stop acquisition while storing if CINE memory is at least 80% full. Monitor CINE memory while storing CINE loops to ensure continuous live scanning.

NOTE

DICOM loops take significantly longer to store. Storage time may approach or exceed loop time. Allow extra space in CINE memory when saving DICOM loops.

NOTE

The CINE gauge turns purple to indicate the section of CINE memory that is being stored in the background.

Image Storage Hints



HINTS

Setup Tips

- Print Button Setup is Application specific. When you access Application > Print Controls, the current Application is the default Application.
- To apply the same Print Control Settings for all Applications, select All Applications as the preset on the Print Controls Menu. Be sure to re-enter values if the field is green.
- Print button setup for the file format and destinations are still configured via the Connectivity Menus.



HINTS

Usage Tips

- If you select Mark CINE, the next time you press Print completes the Mark CINE Loop Store, independent of its configuration.
- The CINE gauge turns green when a Prospective CINE Clip is pending.
- You can cancel Prospective Store by pressing Freeze/Unfreeze or by changing Modes.

CINE Capture

Selecting **Capture** searches through all images between the start frame and end frame and displays each peak or the highest velocity/tissue power. Adjust the start frame and end frame points to limit the image frames used in the process.

1. Display the CINE loop which is in memory or recalled from archive.

NOTE

CINE Capture applies only to 2D images (B, B Flow, CF, PDI, Contrast, etc.).

NOTE

On 2D duplex modes (B/CF, B/PDI, etc.), CINE Capture is not applied to the background B-Mode image, even if the CF/PDI display is turned off.

2. Run the CINE loop.
3. Select **Capture** on the Touch Panel to display the captured image.

A character 'C' displays on the screen instead of the frame average level.

4. If necessary, save the captured image.
5. Press **Capture** again to turn CINE Capture off.

NOTE

CINE Capture can be used on exported files by using the Save As function. You can save the still image (jpeg) and CINE loop (avi) by using Save As.

NOTE

CINE Capture does not effect TIC Analysis.

Capture Recon.

Description

Capture Recon. reconstructs small blood vessels in the CINE mode.

It is available in B-Flow, Contrast, CF, PDI, and B-Mode without CrossXBeam.

Capture Recon. shows smaller blood flow information than conventional B-Flow image, which expected for the characterization of tumor, chronic liver disease, vascular abnormality, etc.

Procedure

1. Scan B-Flow mode in the ordinary way, except:
Sensitivity can be larger (20 (default) to 30 or more).
2. Acquire a short clip, then freeze.
 - Combination of Capture helps to imagine resulting image.
 - 3-dimensional sweep may be efficient to visualize many vessels.
3. Press **Capture Recon** on the CINE Touch Panel. Adjust threshold as necessary.
4. The order of the frames is rearranged by order of smaller artifact. The frame which has motion artifact is rejected. Move the Trackball or assigned rotary to review the image Frame by Frame.

NOTE

While the Capture Recon, is activated, the color of CINE gauge turns pink.

Enhancement

Enhancement executes the enhancement process to cine capture images.

1. Display the cine capture image using **Capture**.

NOTE

Enhancement applies only to CINE Capture images.

2. Select **Enhancement**. A character 'C' displays on the screen instead of the frame average level.
 - Shade Fw/Shade Rv
Select Fw (Forward) that brings start frame or Rv (Reverse) that brings last frame of cine to the front. The system has two thresholds for the Shade process. Comparison

will be made frame by frame after the thresholds have been met. If the two thresholds are satisfied, no comparison will be made with the rest of the frame.

Using Shade Fw/RV, lower intensity echoes in the near frame will be masked by higher intensity echoes in the far frame. ShadeFw/Rv makes it possible to show lower intensity echoes in the near frame despite the fact that there may be a higher echo projection in the far frame. Therefore, the anteroposterior position of the blood vessel is clearly displayed.

- Enhance1/Enhance2/Enhance3

The part corresponding to the data of the selected frame is enhanced and superimposed on the entire cine capture images. This allows you to visualize the spatial relationships with the B-Mode image and the flow appearance.




Figure 6-4 • Enhancement Image Example

3. If necessary, save the enhanced image.

NOTE

You can save the still image (JPEG) and CINE loop (WMV) by using Save As.

Motion Correction

Understanding of the vascular structure is important, for example, with liver cancer. Though, the captured image distorts or blurs due to the patient's breathing. Motion correction compensates each frame and cancels the motion.

NOTE

Motion Correction is activated only for B, CHI, Contrast and B-Flow.

1. Displays the cine capture image using **Capture**.
2. Select **Motion Correction**. The system generates the captured image with motion correction.

The user can store the cine clip with the state of motion correction kept in the raw data.

The user can store the still image as the single frame DICOM image without the raw data.

Image Zoom

There are three kinds of zoom on the system: Pan Zoom, HD Zoom and Magnification Zoom.

- HD Zoom only acquires the image data within the ROI, increasing the density of the image in the ROI. Image adjustments can only be performed during live scanning.
- Pan Zoom magnifies the display of the data within the ROI without making any changes to the ultrasound image data acquired. The entire image is acquired and the ROI can be adjusted (moved and resized).
- Magnification Zoom magnifies the entire image on the screen (non-ROI zoom).

Location of Zoom Control




Figure 6-5 • Zoom Control Knob

HD Zoom

In HD Zoom, the Ultrasound line density and/or sampling frequency increases, resulting in higher resolution. HD Zoom can be performed on live images.

1. To activate HD Zoom, push the **Zoom** control knob inward while scanning live.

NOTE

If already in Magnification Zoom, push the **Zoom** control knob inward to first turn off Magnification Zoom, then push again to activate HD Zoom.

Press the top trackball key to select **Size** for change ROI size or **Pos** for change ROI position.

Use the trackball to position the zoom area over the desired portion of the image.

2. To exit HD Zoom, push the **Zoom** control knob inward again.

NOTE

When HD Zoom is on and the image is live, the right trackball button toggles from HD Zoom to Pan Zoom.

Bioeffect

HD Zooming an image changes the frame rate which tends to change thermal indices. The position of the focal zones may also change which may cause the peak intensity to occur at a different location in the acoustic field. As a result, the MI (TI) may change.



ACOUSTIC OUTPUT HAZARD

Observe the output display for possible effects.

Pan Zoom

Pan Zoom can be performed on a live, frozen, CINE or recalled raw data image.

1. To activate Pan Zoom (when frozen or in CINE) push the **Zoom** knob inward.
Press the top trackball key to select **Size** to change ROI size or **Pos** to change ROI position. Use the trackball to change the ROI size or position the zoom area over the desired portion of the image.
2. To deactivate Pan Zoom push the **Zoom** control knob inward again.

NOTE

When Pan Zoom is on and the image is live, the right trackball button toggles from Pan Zoom to HD Zoom.



- 1 Zoom Image
- 2 Reference image is the small, un-zoomed image.
- 3 Zoom ROI: Zoom ROI indicates the region of the image to zoom.
- 4 Pos/Size: Use the top trackball key to change position and size of ROI.

Figure 6-6 • Zoomed Image Example

Magnification Zoom

Magnification Zoom magnifies the entire image on the screen (non-ROI zoom).

1. To activate Magnification Zoom, rotate the **Zoom** control knob clockwise.
2. There are three ways to deactivate Magnification Zoom:
 - Press Mag Zoom Rest on the trackball button.
 - Rotate the **Zoom** control value to 0.
 - Activate B-Mode (push the B-Mode button on the Front Panel).

Panning in Magnification Zoom

While in Magnification Zoom, you can pan the image across the screen.

1. Press the Top Trackball button to highlight the Position Indicator “Pos.”



Figure 6-7 • Position Indicator

2. Move the Trackball to pan the image across the screen.

Split Screen

The ultrasound system supports the following multiple image format:

- Dual (split the window area into 2 areas)
- Wide Dual (split the window area into 2 areas, but wider than the normal dual)
- Quad (split the window area into 4 small areas)
This is useful, for example, when measuring AFI of OB.
- Simultaneous (Dual) (split the module window into 2 areas, with both panes live and active)

NOTE

Recalled dual/quad images can be edited.

Dual Screen

1. Press **L** to activate a dual screen. The single image is placed on the left side.

NOTE

When you activate the dual screen by pressing L, the single image is placed on the left side; when you activate by pressing R, the single image is placed on the right side.

2. Press **R**. The left side image is frozen and the image displays in the right side.
3. Press **Freeze** to freeze the image of the right side.
4. Press **Freeze** again to unfreeze the active image which has the gray bar under the image.

To switch between active images, press L or R.

5. Press **B-mode** key to return to the single screen.

NOTE

To put a copy of the image on the opposite side when entering dual split screen, use the "When Entering Dual Image" preset found on Utility > Application > Settings preset page.

Quad Screen

1. Press and hold down **L** to activate a quad screen. The single image is placed on the upper left.

NOTE

When you activate the dual screen by pressing L, the single image is placed on the left side; when you activate by pressing R, the single image is placed on the right side.

2. Press **R**. The left side image is frozen and the image displays in the upper right.
3. Press **Freeze**.
4. Press **Freeze** again to unfreeze the image which has the gray bar under the image.

Press **L** or **R** to move the gray bar to the image of the left side or the right side.

5. Press **B-mode** key to return to the single screen.

Simultaneous Mode

While using CFM or PDI, press **L** and **R** keys simultaneously to display B and B+CFM, or B and B+PDI in real-time on the left and right side.

It is useful to observe the ROI in B-Mode.

NOTE

Simultaneous mode can also be used within B-Mode to view with (Right) and without CrossXBeam (Left) if CrossXBeam is turned on.

Dual Caliper

In split screen, you can draw a caliper, area, ellipse, or spline trace on both the left and right image at the same time. Whichever side of the screen that you annotate is called the “Original” graphic. The copy is called the “Shadow” graphic.

This feature is available in the following modes:

- B-Mode:B-Mode
- Color Flow Mode:Color Flow Mode
- B-Mode:Color Flow Mode
- Simultaneous Mode.
- Contrast
- Elastography
- Volume Navigation

NOTE

Dual Caliper IS NOT available in B-Mode: B/PW Mode or in B-Mode:B/M Mode, or with different probes.



Figure 6-8 • Original (Left), Shadow (Right)

NOTE

Only the Original graphic contains the graphic numbering. In this way you can always distinguish between the Original and the Shadow graphic.

NOTE

You can only edit the Original graphic; however, when you do edit the Original graphic, the Shadow graphic is also edited at the same time.

NOTE

If you delete either graphic, both are deleted.

NOTE

When a measurement is selected without Dual B-Mode images or with different probe images, a warning message is displayed on the status bar and the selected measurement is cancelled.

NOTE

If the first point of the Original graphic is out of the Shadow image area, then a warning message displays on the status bar and the Shadow graphic is not drawn.

NOTE

The Trackball move area is limited to the narrow area of both images.

NOTE

You cannot take a measurement across dual images.

NOTE

The 2D Dual measurement tool cannot be copied.

Dual Caliper for 2D Image

2D Dual Caliper / 2D Dual Area / 2D Dual Ellipse / 2D Dual Spline Trace / 2D Dual Circle are not available through the factory default. To enable these measurements, add a new measurement using “2D Dual Caliper”, “2D Dual Area”, “2D Dual Ellipse”, “2D Dual Spline Trace” or “2D Dual Circle” tool in the Utility > Measure > M&A preset menu.

1. Select Blank from Add measurement.



Figure 6-9 • Add Measurement

2. Select appropriate dual caliper tool from Tool drop-down menu.



Figure 6-10 • Drop-down menu

3. Type the measurement and parameter name as you like.
4. Add the created measurement to the Touch Panel.
5. Display the dual image and press **Measure**.
6. Select an added measurement from the Touch Panel to enable the appropriate measurement. A caliper displays.

NOTE

When the measurement is selected without dual B images or with different probe images, a warning message displays on the status bar and the selected measurement is cancelled.

7. To position the caliper at the start point, move the **Trackball**.

You can use both images as an original image.

NOTE

If the first point of the original graphic is out of the shadow image area, then the warning message displays on the status bar and the shadow graphic is not drawn.

8. To fix the start point, press **Set**. The caliper changes to an active caliper.

NOTE

Only original graphic has graphic numbering to distinguish between original image and shadow image.

NOTE

The trackball move area is limited to the narrow area of both images.

NOTE

Only the original graphic can be edited. When the original graphic is edited, the shadow graphic is also updated.

9. To complete the measurement, press **Set**. The system displays the measurement result in the Results Window.

Dual Caliper for V Nav

1. Check **Dual Caliper on VNav** in Utility > System > System Measure. Only available distance, area, trace and angle measurement for V-Nav.

NOTE

You cannot measure on the recalled image.

NOTE

Measure is available on the cine image shortly after freezing the image; however, you cannot measure on recalled images.

2. Select the measurement via the Trackball key for a live scan image. Select the measurement from the Measurement Type table for Volume data.

NOTE

Numbers are only displayed on the live scan image.

3. If you want to measure live scan and volume data separately, select **Off** for Dual Caliper using the Touch Panel key.

NOTE

After the Dual caliper measurement, graphics and measurement results are removed from the Volume dataset when you press **Measure**.

NOTE

If you use Zoom after the Dual caliper measurement, graphics and measurement results are removed.

Annotating an Image

The comment function provides the capability to type the comments of free text and/or insert the pre-defined comments from the comment library. It also provides the user with arrow markers to point to parts of the image.

Pressing the **Comment** key or any keys on the alphanumeric keyboard initiates the comment mode. This assigns the Trackball function to controlling the cursor and displays the comment library on the Touch Panel menu.



Figure 6-11 • Comment Key on the Front Panel

In comment mode, text can be added by using the comment library or by typing from the alphanumeric keyboard.

Comments can be erased by powering down, when you press **Clear** or **New Patient**, or when preset via Utility > Comments.

In addition, the display's home position can be changed (preferred comment area) for each display so that all subsequent comments begin in the same spot.

To return to the user specified position or factory default position, press the **F7 (Cursor Home)** key or the **Home** control on the Annotation Touch Panel (see *Figure 6-13 • Annotation Touch Panel* on page 250).

To establish a new cursor home position, place the cursor in the desired position and press **Shift+F7** or press and hold the Home control on the Annotation Touch Panel.

Comment Mode is activated by pressing the **Comment** key. Comment Mode can also be automatically activated by typing from the alphanumeric keyboard.

NOTE

In this case, the cursor begins at the same location where the comment mode was exited.

After activating the comment mode, a vertical bar type cursor appears on the screen. Use the **Trackball** to move the cursor.

The factory default color for comments is yellow. The color selection can be changed to any of the colors available on the system. The choices are white, yellow, bright red, orange, etc.

NOTE

The user cannot change the Font Family.

To indicate a specific comment or text group is selected, the color turns to blue. Once the comment is set or fixed, the color returns to yellow or to the user selected color.

On the Touch Panel annotations have been color-coded for easy replacement or to easily add annotations. See Annotation Groupings below.

To delete comments by character, press the **Backspace** key.

To delete all comments and arrow marks, press the **Clear** key twice immediately after entering the comment mode.

To exit the Comment/Library Comment function, press the next function you wish to do.

To move by words or by text group, press the **Tab** key.

Adding Comments to an Image

Comment Retention

Comments from the B-Mode images are retained and carried over when switching to multi-image format or duplex mode.

The position of the comments is adjusted so that it is at the same relative position with respect to the display window in the new format as it was in the single image format.

NOTE

Comments may not be retained when the image is switched to M-Mode image format depending on the preset.

Arrow Pointers

Arrow pointers can be used by activating the F2 (Arrow) key on the keyboard or by selecting the **Arrow** control on the Annotation Touch Panel (see *Figure 6-13 • Annotation Touch Panel* on page 250). When the pointer comes up, it is a blue color, indicating it is active and can be moved.

- Move the pointer using the **Trackball** to any place on the screen. The pointer head direction can be controlled by movement of the Trackball or **Arrow Rotate** control.
- To readjust the length and thickness of the pointer, use the **Arrow Resize** rotary control. The default for the pointer size can be preset.
- Press **Set** to fix the place of the pointer and direction of the pointer head. The blue color turns to yellow (or the default color if changed).
- To delete the arrow marks, press the **Clear** key right after pressing the F2 (Arrow) key or press and hold the **Arrow** control on the Annotation Touch Panel.

NOTE

To prevent the Trackball from changing the arrow angle, select the "Keep arrow angles" preset at Utility > Annotation > Comments.

Text Overlays

There are 2 layers of the text in comments, which can be selected by toggling the F8 (Text1/Text2) key on the keyboard or by selecting the **Comment Layer Switch** control on the Annotation Touch Panel (see *Figure 6-12 • Comment Layer Switch Control* on page 249). Text1 is the default choice.



Figure 6-12 • Comment Layer Switch Control

By using this function, users can perform a HIDE TEXT/SHOW TEXT, allowing the users to save or print an image without clearing the typed text.

You can specify to display text 1, text 2, or both. This allows you to have some comments that do not change during the exam while allowing you to change the other comment. Toggle the F8 key or **Comment Layer Switch** control to cycle through the three Text 1/Text 2 states:

1. Text 1 Only -- Only Text 1 comments displayed and editable.
2. Text 2 Only -- Only Text 2 comments displayed and editable.
3. Text 1 and Text 2 -- Both displayed; only Text 2 comments editable. Only Text 2 comments erased by Clear key. Word Delete only deletes Text 2 comment. Both Text 1 and Text 2 comments erased with new patient, new exam, or probe change.

To preset the Text Overlay Sequence, go to Utility > Comment > Comment and select either **Text 1 and Both** or **Text 1 and Text 2 and Both**.

The font color for the Text1 and Text2 overlays can be set separately. Go to Utility > Comment > Comment and specify the text color for Text 1 Color and Text 2 Color.

NOTE

If you check "Erase when image is unfrozen" in the Utility menu, only the editable text plane erases when you unfreeze the image.

Annotating an Image Using the Library

To reduce the amount of time spent annotating an image, store frequently-used comments in the Comment Library. As many as 6 libraries are available per study. One of the selected

libraries is designated as the default and its entries shall be displayed on the Touch Panel when the comment mode is activated for that study.

Press **Comment** and move the comment cursor location using the **Trackball**.

Select the desired comment from the Touch Panel.

Each Touch Panel key can also be configured to hold a small list of up to 3 comments. The first word in the list is displayed on the Touch Panel and the others can be accessed by toggling the key or pop-up menu. To show the presence of a small list stored under a particular key, a small indicator (>) is placed on the key.

To program your system with specific comments, see *Comments Libraries Presets* on page 476 for more information.

Annotating an Image by Comment Groupings

Comments can be grouped together for ease of annotating. For example, in the figure below there are blue, yellow, and green groups.



Figure 6-13 • Annotation Touch Panel

When you select comments with a different color, the selected comment is added to the existing comment. In the example below, Right was selected from the blue group of comments. Trans was selected from the yellow group, so it was added to the first comment. Bowel was selected from the green group, so it was also added.




Figure 6-14 • Added Comments (Different Color Groups)

When you select comments with the same color, the selected comment is overwritten with the new comment. In the example below, Bowel was selected from the green group of comments, but then Appendix was selected from the same green group, so it replaced Bowel.




Figure 6-15 • Replaced Annotation (Same Color Group)

You can configure these Comment color groupings via Utility > Comments > Libraries (up to 5 groups). Groups are organized by color. Position the cursor over the rectangular region to the left of the annotation to display the group drop-down selection menu, then select the color that matches your grouping.

NOTE

You can also group annotations by dots (: and .). Dots are for color blindness recognition in groupings.




Figure 6-16 • Tagging a Group Color to an Annotation

The configured comments are displayed on the Touch Panel.



Figure 6-17 • Color-Tagged Annotation Touch Panel

Rearranging Touch Panel Comments

You can easily move comments to a new location by selecting the annotation and moving it to the new, empty location.

NOTE

You cannot move a comment to a location with a comment. The new location must be empty.

Table 6-3 • Moving Touch Panel Annotations

Step	Example
1. Position the cursor over the annotation you want to move. In the example, we are going to move “Left” to a cell to the right.	
2. Hold and drag the “Left” annotation to the new cell; in this case the empty cell to the right.	

Step	Example
3. Release the cursor when you have the annotation positioned in the new cell location. Then Save and Exit the Utility Page.	

Annotating an Image with Typed Words

- Press **Comment** and type the comments where the cursor is currently located (the display's home position) and use the **Trackball** to further place the comment cursor in the desired location.
- Press Enter to move to the next line.

NOTE

Comments wrap to the next line when they are within one character of the right margin if Word Wrapping is selected in the Text Boundary preset. See *Comments Libraries/Comments Preset Menu* on page 478 for more information.

The word wrap starts one line below the start of that comment.

Comments appear on all prints, photos, DVR or CINE loops.




Figure 6-18 • Next Line Word Wrap

NOTE

The same word wrap principles apply for library scripts as typed comments.

Moving Text

You have the ability to move comments already on the screen and place them in different locations.

- Place the cursor on the desired text or text group and press **Set**.
- The selected text color turns to green.
- Use the **Trackball** to move the selected text and press **Set**.

Editing While Annotating

Backspace over any error(s) made. Blank spaces take the place of the letter(s) that were there. Continue typing the comment after backspacing over all incorrect letters.

To delete previous character(s):

- Press **Backspace** as many times as necessary to make the deletion.
- Once all texts within the selected text group are deleted, then the cursor will find another text group to delete to the upper left direction.
- If there is no more text to delete, the cursor will be located at home position.
- To delete all comments and arrow marks, press the **Clear** key twice immediately after entering the comment mode.

To move through the text a word at a time:

- Press **Tab** to move to the right by text group (Preset Keyboard Tab = Word)

NOTE

Press Shift + Tab to move to the left.

To activate the last text group typed or selected from the Library:

- Press F9 (Grab Word) key. The selected comment will be highlighted.
- To increase/decrease the area of the highlighted selection use HIGHLIGHT rotary.

NOTE

Once the text is highlighted, typing comments or choosing them from the library replaces the highlighted text.

NOTE

To select all text groups, Press Shift + F9 (Grab Word) key.

To cancel the last action:

- Press **Undo** key.

Body Patterns

An additional way to annotate the image display is with body patterns. Body patterns are a simple graphic of a portion of the anatomy that is frequently scanned. The body pattern and probe marker can serve as a reference for a patient and probe positioning when images are archived or scanned.

1. Press **Bodypattern/Ellipse**. Body patterns specific to the current application are displayed.



Figure 6-19 • Body Pattern/Ellipse Control

2. Touch the body pattern to insert. The selected body pattern with a probe marker is displayed on the scanning screen.
3. Using the trackball, adjust the position of the probe marker.
4. Rotate **Body Pattern/Ellipse** to set the probe marker orientation.
5. To move the body pattern:
 - Press **Move Pattern**.
 - Move the bodymark to a new location with the trackball.
 - Press **Set** to anchor the bodypattern to the new location.
6. Press **Set** on the keyboard or **Scan** on the Touch Panel to exit without erasing the body pattern.
7. To clear the body pattern, press the Body Pattern/Ellipse control to activate body patterns and then press the Clear key.

The body pattern packages may be customized to accommodate user preference. Up to 30 individual body patterns in the packages can be changed. See *Body Pattern Libraries/ Applications Preset Menu* on page 485 for more information.




Figure 6-20 • Touch Panel displays of Body Pattern - example

1Move Pattern: Each Touch panel key can also be configured to hold a small list of up to 3 comments. The first word in the list is displayed on the Touch Panel and the others can be accessed by toggling the key.

To program your system with specific comments, see *Creating a small list* on page 477 for more information.

2Save Position: Move the body pattern to the desired location and press the **Save Position**. Current position of Body Pattern is saved as a Home Position of current display format. Hold down **Save Position** to reset the home position to factory default.

3Save Probe Position: Rotate the arrow. The arrow can be rotated with the Trackball.

4Probe Type: The probe mark type is selectable by rotating the **Probe Type** control on the Touch Panel. There are different choices available with one being a blank selection.

To select the active side in dual B-Mode, use the **Active Side** rotary control at the bottom of the Touch Panel.

You can use the **Zoom** control to select the body pattern. If you want to assign the select function to the Zoom control, see *Body Pattern Libraries/Body Patterns Preset Menu* on page 484 for more information.

Notes for Body Pattern (Probe Mark)

- Probe Type is the type of probe mark displayed on the body pattern. It can be saved only for each body pattern on the Touch Panel while body pattern is activated, but not in the Utility preset menu. Therefore, Probe Type cannot be saved as an Application or System Preset.

To save the Probe Type:

1. Activate the **Body Pattern**.
2. Select a Body Pattern on the Touch Panel.
3. Select a type of probe mark with the **Probe Type** Touch Panel key.
4. Place the probe mark at the proper location.
5. Select the **Save Probe Position** Touch Panel key.

NOTE

“Save Probe Position” saves both the Probe Mark position and Probe Type.

- When a Body Pattern is selected and no Probe Mark has been saved on it, the latest used Probe Mark is carried over to the Body Pattern.
- Check the Body Pattern on the Touch Panel if the Probe Mark does not appear on the monitor.




Figure 6-21 • Body Pattern Touch Panel

- 1 When the Probe Mark is saved for the Body Pattern, the Probe Mark is displayed in yellow on the Touch Panel and should also be shown on the monitor.
- 2 When the Probe Mark is saved with "Probe Type - None", the Probe Mark is displayed in gray on the Touch Panel and is not shown on the monitor. Reselect an appropriate Probe Type and save as necessary.
- 3 When the Probe Mark is not saved, no Probe Mark is displayed either on the Touch Panel or the monitor. Select an appropriate Probe Type and save as necessary.



1. abdo 1


16. body 1

31. breast 1

46. carotid 1

2. abdo 2	17. body 2	32. breast 2	47. carotid 2
3. abdo 3	18. body 3	33. breast 3	48. carotid 3-Lt
4. abdo 4	19. body 4	34. breast 4-Lt	49. carotid 3-Rt
5. aorta	20. body 5	35. breast 4-Rt	50. carotid 4
6. arm 1	21. body 6	36. breast 5-Lt	51. carotid 4-Lt
7. arm 2	22. body 7-Lt	37. breast 5-Rt	52. carotid 4-Rt
8. arm 3	23. body 7-Rt	38. breast 6-Lt	53. carotid 5
9. arm 4	24. body 8-Lt	39. breast 6-Rt	54. chest 1
10. arm 5	25. body 8-Rt	40. breast 7-Lt	55. fetus 1
11. arm 6	26. body 9	41. breast 7-Rt	56. fetus 2
12. arm 7	27. body 10	42. breast 8-Lt	57. fetus 3
13. arm 8	28. body 11-Lt	43. breast 8-Rt	58. fetus 4
14. arm 9	29. body 11-Rt	44. breast 9-Lt	59. fetus 5
15. arm 10	30. body 12	45. breast 9-Rt	60. fetus 6

Figure 6-22 • Body Patterns Available



61. fetus 7	76. foot 3-Rt	91. hand 3-Lt	106. heart 1
62. fetus 8	77. foot 4-Lt	92. hand 3-Rt	107. heart 2
63. fetus 9	78. foot 4-Rt	93. hand 4-Lt	108. heart 3
64. fetus 10	79. foot 5-Lt	94. hand 4-Rt	109. heart 4
65. fetus 11	80. foot 5-Rt	95. hand 5-Lt	110. heart 5

66. fetus 12	81. foot 6-Lt	96. hand 5-Rt	111. heart 6
67. fetus 13	82. foot 6-Rt	97. hand 6-Lt	112. heart 7
68. fetus 14	83. foot 7-Lt	98. hand 6-Rt	113. heart 8
69. fetus 15	84. foot 7-Rt	99. hand 7-Lt	114. heart 9
70. fetus 16	85. foot 8-Lt	100. hand 7-Rt	115. heart 10
71. foot 1-Lt	86. foot 8-Rt	101. head 1	116. heart 11
72. foot 1-Rt	87. foot 9-Lt	102. head 2	117. legs 1
73. foot 2-Lt	88. foot 9-Rt	103. head 3	118. legs 2
74. foot 2-Rt	89. hand 1	104. head 4	119. legs 3
75. foot 3-Lt	90. hand 2	105. head 5	120. legs 4


Figure 6-23 • Body Patterns Available (continued)



121. legs 5	136. neck 3	151. organ 7	166. twin 2
122. legs 6	137. neck 4	152. organ 8	167. twin 3
123. legs 7	138. neck 5	153. organ 9	168. twin 4
124. legs 8	139. ob 1	154. pelvis 1	169. twin 5
125. legs 9	140. ob 2	155. pelvis 2	170. twin 6
126. legs 10	141. ob 3	156. post-breast-bilateral	171. twin 7
127. legs 11	142. ob 4	157. post-breast-Lt	172. twin 8
128. legs 12	143. ob 5	158. post-breast-Rt	173. uro 1
129. legs 13-a-Lt	144. ob 6	159. rheuma	174. uro 2

130. legs 13-a-Rt	145. organ 1	160. shoulder-back-Lt	175. uro 3
131. legs 13-Lt	146. organ 2	161. shoulder-back-Rt	176. uro 4
132. legs 13-Rt	147. organ 3	162. shoulder-front-Lt	177. uro 5
133. liver	148. organ 4	163. shoulder-front-Rt	178. uro 6
134. neck 1	149. organ 5	164. thyroid	179. uro 7
135. neck 2	150. organ 6	165. twin 1	180. uro 8

Figure 6-24 • Body Patterns Available (continued)



181. uro 9 female	196. ankle lat-Lt	211. foot-Rt	226. hip-Lt
182. uro 9 male	197. ankle lat-Rt	212. foot lat-Lt	227. hip-Rt
183. uro 10 female-Lt	198. ankle post-Lt	213. foot lat-Rt	228. hip lat-Lt
184. uro 10 female-Rt	199. ankle post-Rt	214. forearm-Lt	229. hip lat-Rt
185. uro 10 male-Lt	200. arm-Lt	215. forearm-Rt	230. hip post-Lt
186. uro 10 male-Rt	201. arm-Rt	216. forearm post-Lt	231. hip post-Rt
187. uro 11	202. arm post-Lt	217. forearm post-Rt	232. knee-Lt
188. uro 12	203. arm post-Rt	218. hand-Lt	233. knee-Rt
189. uro 13-Lt	204. elbow-Lt	219. hand-Rt	234. knee lat-Lt
190. uro 13-Rt	205. elbow-Rt	220. hand post-Lt	235. knee lat-Rt
191. uterus 1	206. elbow lat-Lt	221. hand post-Rt	236. knee post-Lt
192. uterus 2	207. elbow lat-Rt	222. head	237. knee post-Rt
193. uterus 3	208. elbow post-Lt	223. head lat-Lt	238. leg-Lt

194. ankle-Lt	209. elbow post-Rt	224. head lat-Rt	239. leg-Rt
195. ankle-Rt	210. foot-Lt	225. head post	240. leg post-Lt

Figure 6-25 • Body Patterns Available (continued)



241. leg post-Rt	248. thigh-Lt	255. wrist post-Rt	262. brain4
242. shoulder-Lt	249. thigh-Rt	256. new carotid 2	263. breastSA-Rt
243. shoulder-Rt	250. thigh post-Lt	257. carotid2-Lt	264. breastSA-Lt
244. shoulder lat-Lt	251. thigh post-Rt	258. carotid2-Rt	265. lung1
245. shoulder lat-Rt	252. wrist-Lt	259. brain1	266. organ10
246. shoulder post-Lt	253. wrist-Rt	260. brain2	267. organ11
247. shoulder post-Rt	254. wrist post-Lt	261. brain3	

Figure 6-26 • Body Patterns Available (continued)

Using the Fast Key

A keyboard Fast Key is available to record and run a sequence of often-run keystrokes.

NOTE

Ensure that you have a patient selected prior to running the Fast Key operation.

Create a Fast Key

1. Press the F5 key. The “Do you want to create the Fast Key?” dialog displays. Select OK to continue.
2. Select a key to assign a Fast Key to (a-z, 0-9).

If you select a Front Panel control, a Touch Panel key or any key besides a-z or 0-9, a warning dialog displays and the procedure is cancelled.

NOTE

Assign Fast Key Function to Key 0 - 9 in Utility > System > User Configurable Key before you create a Fast Key.

NOTE

There is no distinction between capital and small letters.

NOTE

The key code is the same in Russian and Greek (a-z, 0-9).

3. If the selected key is already assigned to a Fast Key, a warning dialog displays.
Select Yes to continue. The Fast Key file is overwritten.
Select No to cancel the Fast Key setup.
4. Input the key sequence to be assigned.

NOTE

It is impossible to save a power cycle sequence or any input from outside of the system.

NOTE

If a warning dialog displays due to the limitations of the number of key sequences, press F5 to finish and retry.

5. Press the F5 key to complete a Fast Key setup. The information dialog displays. Select OK.

Start a Fast Key

1. Press the F6 key to start a Fast Key. The message “Select the key which the Fast Key is assigned to” displays on the status bar.

NOTE

The F6 key is ignored if another dialog displays on the system.

NOTE

If you press F5 after F6, the F6 function cancels and the F5 function is enabled.

2. Press the key assigned to the Fast Key macro. The message “Fast Key playback is finished” displays on the status bar when the macro is finished.

To stop the Fast Key during the operation, press F6. The message “Fast Key playback is cancelled” displays on the status bar.

NOTE

Select the running speed in the Run Fast Key Speed preset on Utility > System > General.

Backup and Restore the Fast Key

You can backup/restore the Fast Key via Utility > System > Backup/Restore.

To Backup - Select User Defined Configuration in the Backup section.

To Restore - Select User Defined Configuration in the Restore section.

Chapter 7

General Measurements and Calculations

In this section

General Measurements and Calculations Introduction.	266
Generic Measurements.	278
Mode Measurements.	292
Worksheet.	306
Anatomical Survey.	315
Measurement and Calculation Setup.	318

General Measurements and Calculations Introduction

Measurements and calculations derived from ultrasound images are intended to supplement other clinical procedures available to the attending physician. The accuracy of measurements is not only determined by system accuracy, but also by the use of proper medical protocols by the user. When appropriate, be sure to note any protocols associated with a particular measurement or calculation. Formulas and databases used within the system software that are associated with specific investigators are so noted. Be sure to refer to the original article describing the investigator's recommended clinical procedures.



WARNING

The system provides calculations (e.g. estimated fetal weight) and charts based on published scientific literature. The selection of the appropriate chart and clinical interpretation of calculations and charts is the sole responsibility of the user. The authorized user should consider proper indications for the use of a calculation or chart as described in the scientific literature. The diagnosis, decision for further examination, and medical treatment must be performed by qualified personnel following good clinical practice.

Basic Operation

Measure and Assign

1. Press **Measure** on the control panel.
2. Select the measurement tool via the upper trackball key.
3. Perform the measurement. Follow the instructions displayed on the message area at the bottom of the screen.
4. To assign a label, select the measurement in the Measurement result window and press **Set**.



Figure 7-1 • Label menu

5. Select the required label from the menu. For example, if it is a distance measurement, the list includes all distance calculations for the current study.
or
Select **User Name** from the menu. The dialogue window displays.
Enter the appropriate name and select OK.



Figure 7-2 • Enter Measurement Name

List of General Measurements

The following types of general measurements are available when you press **Measure** but do not choose a specific calculation. The type of measurement depends on the current scan mode.

After pressing **Measure**, rotate between various measurement types with the upper Trackball keys.

B and CF Modes

- Dist (Caliper)
- Trace
- Spline
- Intensity
- Open Trace
- Open Spline

NOTE

You can preset the sequence of B and CF area measurements in the Measure Key Sequence (B/CF) preset in the Utility > Measure > Advanced screen. See the “M&A Advanced Preset” section for more information.

Doppler Mode

- Velocity
- Trace
- Slope
- Time

M-Mode

- Caliper
- Time
- Slope

Assign and Measure

1. Press **Measure** on the control panel.
2. Select the measurement on the Touch Panel.

General Measurements and Calculations

If you select the measurement folder, the sub menu tab is displayed. You can select and perform the measurement in the tab.

If the folder is configured with auto-sequence measurement, the next measurement in the study is pre-selected. To skip a pre-selected measurement, select another measurement.

3. Perform the measurement. Follow the instructions displayed on the message area at the bottom of the screen.

Measurement Controls



Figure 7-3 • Measurement Controls on the Control panel

Table 7-1 • Measurement controls

Control	Description
1. Cursor	Pointer Key. Select to display a pointer on the monitor.
2. Clear	During a measurement sequence, erases the measuring caliper and measurement data from the display. When not performing a measurement sequence, clears all calipers and measurements from the display. To remove all annotations/body markers/arrows, hold down the Clear button.
3. Annotation	Annotations can be added to measurements on the image.
4. Body Pattern	Body patterns can be added to measurements on the image.
5. Measure	Activates a measurement caliper and the calculation package associated with the currently selected preset.

Control	Description
6. Zoom Rotary	Zoom Rotary controls Pan Zoom, HD Zoom and Magnification Zoom. Pan Zoom magnifies the display of the data without making any changes to the ultrasound image data that is acquired. HD Zoom only acquires the image data within the ROI, and can only be performed during live scanning because of the acquisition adjustments that are done. When preset, you can adjust the Depth by moving the toggle up and down. If preset, you can scroll through the body patterns with the Zoom control. Magnification Zoom magnifies the entire image on the screen (non-ROI zoom).
7. Trackball	Moves the measurement calipers, selects the measurement on the Summary Window. Trackball also selects items on the Touch Panel with the Pointer and Set keys.
8. Trackball keys	The functionality of these keys changes (for example, Set, Change Measure, etc.) depending on the mode or action. Current functionality is displayed on the lower-right corner of the monitor.

Touch Panel

B/M/Doppler Mode Select Tab

The B, M, or Doppler mode select tab key allows the user to select measurements associated with a particular mode of active study.

The system provides a default mode selection.



Figure 7-4 • B-Mode - Example

Exam Calcs Tab

The Exam Calcs Tab is used for selecting calculations from other study in the selected Exam Category.




Figure 7-5 • Exam Calcs Tab - example

Selecting a Measurement in a Different Application

While scanning a patient, you may find that you want to measure an item that is not in the current application. In that case, Exam Category allows you to select other calculations without changing application.

1. Select **Exam Calcs** in the measurement mode.
2. Select **Exam Category**.
3. Select the exam category that has the calculation you want to make.
4. Select the study and the desired measurement.
5. After you complete the measurement, to return to the original application, repeat steps 1–4.

NOTE

This measurement DOES NOT appear on the original application worksheet.

Options Tab

The Options tab allows you to specify the following measurement and display options:



Figure 7-6 • Measurement and Display Options - Example

- 1 Select **Hide Display** to hide the Results Window and Summary Window. Hide Display on the Touch Panel key is lit.
To re-display them, select **Hide Display** again.
- 2 The system displays the following choices:
12 x 12
9 x 9
- 3 Select **Heart Rate Cycle** to change the number of cycles used in the calculation. The system displays a list of choices from 1 – 10.
- 4 You may want to change where the Results Window is positioned on the monitor display. To change the location of the Results Window, select **Move Res Win**. The system displays a list of choices.
- 5 Select Auto or Manual for Doppler Trace.
Auto – the system traces the Doppler waveform from the begin time to the end time.
Manual – trace the waveform manually.
- 6 After you press Set to complete the measurement, the dotted line remains on the display if the Show Caliper Line is selected.
If Show Caliper Line is not selected, the system erases the dotted line and only the measurement calipers with a number or symbol are displayed.
Note: The Show Caliper Line on the Options tab takes precedence over the Cursor Line Display preset, found on the System > System Measure screen.

Push/Rotary Button Area

At the bottom of the Touch Panel, there are six (6) combination rotary dials/push buttons. The functionality of these rotaries changes, depending upon the mode, exam category, study, etc.

Side Rt/Lt

The system has measurements for the patient's right and left side. To change side, push or rotate the Side rotary button.

Delete

Delete active caliper.

Cursor Select

When there are several measurements on the display, to rotate through and activate previously fixed calipers, push or rotate the Cursor Select rotary button.

Exam Category/Study/Measurement

For each patient, the system organizes information by exam category, study, and measurement. The definitions of these terms are as follows:

- **Exam Category** – categories include the following:
 - Abdomen
 - Obstetrics
 - Gynecology
 - Cardiology
 - Vascular
 - Urology
 - Small Parts
 - Pediatrics
- **Study/Preset** – after you choose an exam category, the system allows you to select a study. For example, when you choose the Obstetrics exam category, you can choose one of the following studies:
 - Generic
 - OB-1
 - OB-2/3
 - OB-General
 - Fetal Heart
 - OB/GYN Vessel
- **Measurement** – the measurements and calculations needed to analyze an item of anatomy. For example, a femur length is a measurement. A measurement can include several pieces of measurement data. For example, to calculate the ovarian volume, you need to measure width, length, and height.

General Instructions

General Guidelines

Any measurement can be repeated by selecting that measurement again from the Touch Panel.

The system retains all measurements, but the worksheet retains only the last six measurements of each type.

Measurement and Calculation Results

As you take measurements, each measurement is given a sequential number on the display and in the Results Window. The system can display nine measurements on the screen at one time.

Once the Results Window has nine measurements, if you make any further measurements, the system erases the first measurement and adds the new measurement ("first in, first out").

Measurement graphics are kept while in cine scroll. The measurement graphic is redisplayed on the frame where it is taken, if preset "Keep Graphics with Cine Scroll" on the Advanced M&A page.

Selecting a Calculation

When you take measurements, you can select the calculation before you take the measurement or after you take it. For example, in Obstetrics, if you select the calculation before you take the measurement, the estimated fetal age is displayed as you take the measurement. If you select the calculation after you take the measurement, the estimated fetal age is displayed after you complete the measurement.

NOTE

After you take a measurement, if you select a calculation and the measurement is not applicable for the calculation, then the system assumes you want to start the calculation. The system then uses the calculation for the next measurement.

Erasing Measurements

These actions erase measurements from the system's memory:



HINTS

- If you adjust the Trackball, unfreeze the image, or press Clear, the system erases all completed measurements and calculations on the display. Measurements and calculations, however, remain on the worksheets.
- If you select New Patient, the system erases all measurements and calculations on the display and clears the worksheets.
- If you make a new measurement that exceeds the maximum number of allowable measurements, the system erases the first (oldest) measurement and adds the new measurement.
- If the second caliper is active, to erase the second caliper and activate the first caliper, press Clear.

These actions you can take while performing measurements.



HINTS

- Before making measurements, to stop the acquisition of image data, press **Freeze**.
- For measurements such as distance, to make fine adjustments before completing the measurements, press the top Trackball key to toggle between active calipers.
- Before completing the measurement sequence, to erase the active measuring caliper and the current data measured, press **Clear**.
- After the sequence is complete, to erase all data that has been measured to this point, but not data entered on worksheet pages, press **Clear**.
- When there are several measurements on the display, to rotate through and activate previously fixed calipers, adjust the **Cursor Select** knob. After a cursor is activated, you can change the measurement.
NOTE: If you want to change a trace measurement, you must erase it and trace again.
- To repeat any measurement, select that measurement again from the Touch Panel.

Calculation formulas are available in the Advanced Reference Manual.

Measurement Cursor

While you are making a measurement, the measurement cursor is either active (open plus sign) or fixed (closed plus sign). An active cursor is green and a fixed cursor is yellow.

The system allows you to identify measurements by number or by unique symbol. The symbols are used in sequence as listed. The first symbol is used for the first measurement, the second symbol for the second measurement, and so on.



Figure 7-7 • Fixed Caliper Symbols

Measurement graphics are kept while in cine scroll. The measurement graphic is redisplayed on the frame where it is taken, if preset on the Advanced M&A page.

Cursor Preset

You can preset the measurement cursor in Utility > System > System Measure.



Figure 7-8 • System/System Measure Preset Menu

Table 7-2 • Cursor

Preset Parameter	Description
Cursor Type	Select whether to mark measurements with numbers or symbols.
Cursor Size	Specify 12x12 or 9x9.
Cursor Line Display	If selected, after you press Set to complete a measurement, the cursor line is displayed. If not selected, after you press Set to complete a measurement, only the cursor number or symbol is displayed.
Cursor Ellipse Cross Line Display	Check box to display the cross line in Ellipse.
D Manual Trace Cross Line Display	Check box to display the cross line with the caliper.
Cursor Position	Select 1st Cursor, 2nd Cursor, or Image Center.
Color When Set (reboot)	Select white, yellow, bright red, or orange.
Cursor is Displayed when Trackball is Moved	The active cursor does not display until you move the Trackball. This assumes the following presets are set: Repeat Measurement, Repeat, Default Measurement, and Cursor.

Copy, Move and Paste Measurement Tools

You can copy, move and paste the measurement graphic.

NOTE

This function is supported with trace, area trace, spline trace, volume trace, ellipse, 3-point ellipse, circle and intensity. The Double tools and Dual tools are not supported.

Copy and Paste

1. Measure the trace.
2. If present, clear the active caliper using the **Clear** key. Press the **Arrow** key to display the green arrow cursor on the screen. Move the cursor to the + mark of the measurement graphic. The selected graphic color changes from yellow to green.
3. Press **Set**. The pop-up menu displays. Select **Copy**.



Figure 7-9 • Copy and Move Menu

4. Press **Set** on the outer side of the measurement graphic. The pop-up menu displays.
5. Select **Paste**. The copied graphic displays on top of the original graphic in green. Move it to the desired position using the **Trackball** and press **Set** to fix the location.

NOTE

If the copied graphic is bigger than the pasted area, "Paste" fails and "The copied graphic cannot be pasted to this area" message displays on the status bar.

Copy and Move

1. Measure the trace.
2. If present, clear the active caliper using the **Clear** key. Press the **Arrow** key to display the green arrow cursor on the screen. Move the cursor to the + mark of the measurement graphic. The selected graphic color changes from yellow to green.
3. Press **Set**. The pop-up menu displays. Select **Copy&Move**.
4. The copied graphic displays on top of the original graphic in green. Move it to the desired position using the **Trackball** and press **Set** to fix the location.

Move

1. Measure the trace.
2. If present, clear the active caliper using the **Clear** key. Press the **Arrow** key to display the green arrow cursor on the screen. Move the cursor to the + mark of the measurement graphic. The selected graphic color changes from yellow to green.
3. Press **Set**. The pop-up menu displays. Select **Move**.
4. Move the selected graphic to the desired position using the **Trackball** and press **Set** to fix the location.

Performing Measurements on Saved Images

You can perform measurements on recalled images. Select the image, then perform the measurement. If the image was not saved as a raw DICOM image, you need to calibrate the image prior to performing the measurement.

To calibrate the image:

1. Recall the image.
2. Press **Measure**. The Measurement Calibration Touch Panel appears.
3. Select the mode you need to be in to perform the measurement.

4. Press the appropriate mode key on the Touch Panel (**2D calib** for B-Mode, **MM calib** for M-Mode, or **Dop. calib** for Doppler mode). The specified mode calibration pop-up appears.
5. The system prompts you, depending on the mode.

B-Mode:

1. Place the first point of the caliper on the ruler. Press **Set**.
2. Position the cursor at the 5 cm point on the ruler. Press **Set**.
3. Type "5" into the 2D-Mode Calibration pop-up window. Press OK.

M-Mode or Doppler Mode:

1. Place the cross on zero depth and minimum or zero time.
2. Place the cross on maximum depth and time.
3. Type the time (in seconds) and velocity (cm/sec) in the M-Mode/Doppler Mode calibration pop-up window.

Generic Measurements

Each exam category has a Generic study. The Generic studies provide you quick access to measurements such as volume, angle, A/B ratio, and % stenosis. The particular measurements available in each Generic study vary, depending on the exam category and the mode. This section describes generic measurements, organized by mode.

1. On the Control Panel, press **Measure**.
2. On the Touch Panel, select **Exam Calcs**.
3. On the Touch Panel, select the **Generic** folder.

Calculation formulas are available in the Advanced Reference Manual.

B-Mode Measurements

In B-Mode, the Generic study includes the following measurements:

- % Stenosis
- Volume
- Angle
- Velocity
- Stiffness
- V Ratio
- Stiff. Ratio
- A/B Ratio



Figure 7-10 • B-Mode Generic Study

NOTE

Prior to performing B-Mode measurements, scan the patient and then press **Freeze**.

% Stenosis

You can calculate % Stenosis by diameter or by area, depending on the mode.

NOTE

The ultrasound system automatically activates the % Stenosis with the default selection. If another method is preferred, select it from the Touch Panel.

Diameter

NOTE

When you use diameter to calculate the %stenosis, always take the measurement from a cross-sectional view of the vessel.

To calculate percent stenosis by diameter:

1. From the Generic Touch Panel, select **% Stenosis**.
2. Select **%sten(Diam)**.
The system displays an active caliper.
3. Make a distance measurement of the inner area of the blood vessel.
The system displays an active caliper for the second distance measurement.
4. Make a distance measurement of the outer area of the blood vessel.
The system displays each distance measurement and the % Stenosis in the Results Window.

For details on how to make a distance measurement, see *Distance Measurement* on page 292.

NOTE

For the diameter calculation, do NOT take a distance measurement from a longitudinal view. This may lead to an inaccurate assessment of % stenosis.

Area

To calculate percent stenosis by area:

1. From the Generic Touch Panel, select **% Stenosis**.
2. Select **%sten(Area)**.
The system displays a caliper.
3. Make a trace measurement of the inner area of the blood vessel.

NOTE

To erase an open trace, move the **Trackball**.

4. Press **Set**.
The system displays a second caliper.
5. Make a trace measurement of the outer area of the blood vessel.

General Measurements and Calculations

The system displays the two area measurements and percent stenosis in the Results Window.

Ellipse + Area

To calculate percent stenosis by ellipse and area:

1. From the Generic Touch Panel, select **% Sten[E+A]** folder.
2. Ellipse is selected by default.

The system displays a caliper.

NOTE

You can select the trace at this time.

3. Make an ellipse measurement of the inner area of the blood vessel.
4. Press **Set**.

The system displays a caliper.

5. Make a trace measurement of the outer area of the blood vessel.

The system displays the two area measurements and percent stenosis in the Results Window.

NOTE

% Stenosis (E+A) is not available through the factory default. To enable %Stenosis (E+A), add "%Steno(E+A)" to the Measure & Study list on the Utility > Measure > M&A screen.

Volume

The volume calculation can be made from any of the following measurements:

- One distance
- Two distances
- Three distances
- One ellipse
- One distance and one ellipse

For details on how to make a distance measurement, see *Distance Measurement* on page 292.

For details on how to make an ellipse measurement, see *Circumference and Area (Ellipse) Measurement* on page 293.

NOTE

IMPORTANT!! If you want to make a volume calculation using one or two distances, you must select **Volume** BEFORE you make the measurements.

NOTE

If you select Fix Caliper by Print Key on the Utility > System > System Measure, the print key does not function like the Set key, but instead ends the measurement sequence and initiates the volume calculation based on the number of measurements taken so far.

To make a volume calculation using one or two distances:

1. Select **Volume**.
2. Make one or two distance measurements.
3. Select **Volume**.
The system displays the distances and the volume in the Results Window.

NOTE

Use the **Clear** key to erase the green caliper.

To make a volume calculation using three distances:

1. Make three distance measurements.

NOTE

Three distances can be done in the dual format mode (side by side images). One measurement is usually made in the sagittal plane and two measurements in the axial plane. To use the dual format mode, press the **L** or **R** key on front panel.

2. Select **Volume**.
The system displays the distances and the volume in the Results Window.

To make a volume calculation using one ellipse:

1. Make one ellipse measurement.
2. Select **Volume**.
The system displays the ellipse measurement and the volume in the Results Window.

To make a volume calculation using one ellipse and one distance:

1. Make one distance measurement and one ellipse measurement.
2. Select **Volume**.
The system displays the distance and ellipse measurement and the volume in the Results Window.

**HINTS**

- Volumes are most accurate when measurements are taken in the sagittal and axial scan planes.
- To display sagittal and axial plane images simultaneously, use the side-by-side dual format option.

NOTE

If you change the parameters or category during the volume measurement, please follow the procedure below before you restart the measurement.

1. Check the number of each measurement in the summary window.
2. If the numbers are not all the same, it shows that you have the calculation which is not completed. Open the Worksheet and clear that calculation.

Table 7-3 • Volume Calculations

Calc Name	Input Measurements
Volume (spherical)	One distance
Volume (prolate spheroidal)	Two distances, $d1 > d2$
Volume (spheroidal)	Three distances
Volume (prolate spheroidal)	One ellipse: (d1 major axis, d2 minor axis)
Volume (spheroidal)	One distance d1, and one ellipse (d2 major axis, d3 minor axis)




Figure 7-11 • Volume Calculation Examples

Calculation formulas are available in the Advanced Reference Manual.

Post-Assignment for General Volume

You can input a unique name for the general volume measurement. You can group the general volume measurements for each application.

1. Complete the volume measurement.
2. Move the caliper to the measurement result box (with green frame) and select **Set**.
3. The volume name menu appears. Select Name Volume.



Figure 7-12 • Volume Name Menu

4. The dialog box displays. Enter a new name or choose the existing name.



Figure 7-13 • Volume Name Dialog box

NOTE

The factory default volume name cannot be changed (for example, Renal Volume).

Angle

This function measures the angle between two intersecting planes.

1. From the Generic Touch Panel, select **Angle**.
The system displays an active caliper.
2. To position the caliper at the start point, move the **Trackball**.
3. To fix the position of the first caliper, press **Set**.
The system displays a second active caliper.
4. To position the second caliper at the apex of the angle, move the **Trackball**.
5. To fix the position of the second caliper, press **Set**.
The system displays a third active caliper.
6. To position the third caliper, move the **Trackball**.
7. To complete the angle measurement, press **Set**.
The system displays the angle in the Results Window.

NOTE

To rotate through and activate previously fixed calipers, adjust the **Cursor Select** control.

A/B Ratio

In B-Mode, you can calculate A/B ratio by diameter or by area.

NOTE

The ultrasound system automatically activates the A/B Ratio with the default selection. If another method is preferred, select it from the Touch Panel.

Diameter

1. From the Generic Touch Panel, select **A/B Ratio**.
2. Select **ratio(Diam)**.
The system displays an active caliper.

General Measurements and Calculations

3. Make a distance measurement of the first diameter.
The system displays an active caliper for the second distance measurement.
4. Make a distance measurement of the second diameter.
The system displays each distance measurement and the A/B ratio in the Results Window.

NOTE

The first distance is the A diameter. The second distance is the B diameter.

For details on how to make a distance measurement, see *Distance Measurement* on page 292.

Area

To calculate A/B ratio by area:

1. From the Generic Touch Panel, select **A/B Ratio**.
2. Select **ratio(Area)**.
The system displays a caliper.
3. Make a trace measurement of the A area.

NOTE

To erase an open trace, move the **Trackball**.

The system displays a second caliper.

4. Make a trace measurement of the B area.
The system displays the two area measurements and the A/B ratio in the Results Window.

For details on how to make a trace measurement, see *Circumference and Area (Trace) Measurement* on page 294.

Dual Caliper

If “Dual Caliper on VNav and Simultaneous” is checked in Utility> System> System Measure menu, then when you enter simultaneous view, there is a “Dual Caliper: On/Off” setting on the Touch Panel. If the setting is on, dual caliper is enabled in simultaneous view.



Figure 7-14 • Dual Caliper

Velocity and Stiffness Ratios

If set to display via configuration settings, these two measurements appear. Refer to Elastography Option documentation.

M-Mode Measurements

In M-Mode, the Generic study includes the following measurements:

- % Stenosis
- A/B Ratio
- HR (Heart Rate)

% Stenosis

See *% Stenosis* on page 279 for more information.

A/B Ratio

In M-Mode you can measure A/B ratio by diameter, time, or velocity.

NOTE

The ultrasound system automatically activates the A/B Ratio with the default selection. If another method is preferred, select it from the Touch Panel.

Diameter

See *Diameter* on page 283 for more information.

Time

To calculate A/B ratio by time:

1. Select **A/B**.

2. Select **ratio(Time)**.

The system displays an active caliper.

3. To position the caliper at the A point, move the **Trackball**.

4. To fix the measure point, press **Set**.

The system displays a second active caliper.

5. To position the second caliper at the B point, move the **Trackball**.

6. To complete the measurement, press **Set**.

The system displays the two time measurements and A/B ratio in the Results Window.

Velocity

To calculate AB ratio by velocity:

1. Select **A/B**.

2. Select **ratio(Velocity)**.

The system displays an active caliper with vertical and horizontal dotted lines.

3. To position the caliper at the A velocity, move the **Trackball**.

4. To fix the measure point, press **Set**.

The system displays a second active caliper.

5. To position the second caliper at the B velocity, move the **Trackball**.

6. To complete the measurement, press **Set**.

The system displays the two velocity measurements and the A/B ratio in the Results Window.

Heart Rate

To calculate the heart rate from M-Mode:

1. Obtain an image and press Measure. Select **HR**.

The system displays an active caliper.

2. To position the caliper at a recognizable point in the first cycle, move the **Trackball**.

3. To fix the first caliper, press **Set**.

The system displays a second active caliper.

4. To position the caliper at the identical point in the next cycle (depending on preset), you need to move the **Trackball**.

NOTE

In the message bar at the bottom of the display, the system indicates the number of cycles you should measure.




Figure 7-15 • Two Heart Beat Reference (example in Doppler mode)

5. To complete the measurement and transfer the calculation to the worksheet, press **Set**.

NOTE

For information about how to specify the number of heart beats that the system will use, see *Options Tab* on page 270.

Doppler Mode Measurements

In Doppler Mode, the Generic study includes the following measurements:

- PI (Pulsatility Index)
- RI (Resistive Index)
- PS/ED Ratio or ED/PS Ratio
- A/B Ratio
- HR (Heart Rate)

Prior to performing Doppler Mode measurements:

1. In the B-Mode part of the display, scan the anatomy you want to measure.
2. Go to the Doppler Mode part of the display.
3. Press **Freeze**.

Control Assignment

Cancel Transfer

NOTE

Only for Vascular, Abdomen, OB and GYN.

After the Auto Vascular calculation results are assigned to a particular vessel, the user can cancel the assignment and assigned parameters are removed from Worksheet and Report page.

When Cancel Transfer occurs, a message appears on the screen to indicate the value was erased from Worksheet and Report page.

Vessel Location

If the vessel has a location, you can select one of the following:

- Proximal (**Prox**)
- Middle (**Mid**)
- Distal (**Dist**)

NOTE

If you do not wish to assign a vessel location, press the lit location, then no location is assigned. Choose the folder you want the value assigned to.

To select one of the locations, adjust the Touch Panel control.

Pulsatility Index (PI)

For auto trace:

1. Select **PI**.
The system displays a caliper and a vertical dotted line.
2. Position the caliper at the beginning of the waveform.
3. To fix the start point, press **Set**.
The system displays a second active caliper.
4. Position the caliper at the end of the waveform.
5. To complete the measurement, press **Set**.
The system displays peak systole, minimum diastole, end diastole, TAMAX, and PI in the Results Window.

For manual trace:

1. Select **PI**.
The system displays a caliper and a vertical dotted line.
2. Position the caliper at the beginning of the waveform.
3. To fix the start point, press **Set**.
The system displays a second active caliper.
4. Manually trace the entire waveform.
5. To complete the measurement, press **Set**.
The system displays peak systole, minimum diastole, end diastole, TAMAX, and PI in the Results Window.

Resistive Index (RI)

1. From the Doppler Generic Touch Panel, select **RI**.
The system displays an active caliper with vertical and horizontal dotted lines.
2. To position the caliper at the peak systolic velocity, move the **Trackball**.
3. To fix the measure point, press **Set**.
The system displays a second active caliper.
4. To position the second caliper at the end diastolic velocity, move the **Trackball**.
5. To complete the measurement, press **Set**.

The system displays PS, ED, and RI in the Results Window.

PS/ED or ED/PS Ratio

To calculate the Peak Systole/End Diastole ratio or End Diastole/Peak Systole ratio:

1. Select **PS/ED** or **ED/PS**.
The system displays an active caliper with vertical and horizontal dotted lines.
2. To position the caliper at peak systole (PS) or end diastole (ED), move the **Trackball**.
3. To fix the measure point, press **Set**.
The system displays a second active caliper.
4. To position the second caliper at end diastole (ED) or peak systole (PS), move the **Trackball**.
5. To complete the measurement, press **Set**.
The system displays the peak systole, end diastole, and PS/ED or ED/PS ratio in the Results Window.

Heart Rate

To measure heart rate, see *Heart Rate* on page 286.

A/B Ratio

In Doppler Mode you can measure A/B ratio by velocity, time, or acceleration.

NOTE

The ultrasound system automatically activates the A/B Ratio with the default selection. If another method is preferred, select it from the Touch Panel.

Velocity

See *Velocity* on page 286 for more information.

Time

See *Time* on page 285 for more information.

Acceleration

To measure A/B ratio by acceleration:

1. Select **A/B**.
2. Select **ratio(Acc)**.
The system displays an active caliper.
3. Make a distance measurement of the A acceleration point.
 1. To position the active caliper at the start point, move the **Trackball**.
 2. To fix the start point, press **Set**.
The system fixes the first caliper and displays a second active caliper.
 3. To position the second active caliper at the end point, move the **Trackball**.

A dotted line connects the measurement points.

4. To complete the measurement, press **Set**.
The system displays the distance value in the Results Window and displays an active caliper for the second distance measurement.

4. To make a distance measurement of the B acceleration point, repeat steps a–d.

The system displays the two acceleration measurements and the A/B ratio in the Results Window.

Acceleration

1. Select **Accel**.

The system displays an active caliper with vertical and horizontal dotted lines.

2. To position the caliper at peak systole, move the **Trackball**.

3. To fix the measure point, press **Set**.

The system displays a second active caliper.

4. To position the second caliper at end diastole, move the **Trackball**.

5. To complete the measurement, press **Set**.

The system displays the peak systole, end diastole, acceleration time, and acceleration in the Results Window.

Acceleration Time (AT)

1. Select **AT**. The system displays an active caliper and a vertical dotted line.

2. To position the caliper at the start point, move the **Trackball**.

3. To fix the first caliper, press **Set**.

The system displays a second active caliper.

4. To position the caliper at the end point, move the **Trackball**.

5. To complete the measurement, press **Set**.

The system displays the acceleration time in the Results Window.

Peak Systole (PS), End Diastole (ED), or Minimum Diastole (MD)

To calculate the peak systole, end diastole, or minimum diastole:

1. Select **PS**, **ED**, or **MD**.

The system displays an active caliper with vertical and horizontal dotted lines.

2. To position the caliper at the measurement point, move the **Trackball**.

3. To complete the measurement, press **Set**.

The system displays the peak systole, end diastole, or minimum diastole in the Results Window.

Helpful Hints



HINTS

- Prior to making measurements, use the CINE function, if necessary, to display the best image.
- As you take measurements, each measurement is given a sequential number on the display and in the Results Window. Nine measurements can be displayed in the Results Window at one time.
- Once the Results Window has nine measurements, if you make any further measurements, the system erases the top (first) measurement and adds the new measurement last ("first in, first out").
- While you are taking a measurement, the value in the Results Window updates until you complete the measurement.

Mode Measurements

B-Mode Measurements

The following measurements can be made in B-Mode.

- Distance
- Circumference
- Circumference and Area
 - Ellipse Method
 - Trace Method
 - Spline Method
 - Intensity (Echo level) Method

NOTE

Prior to performing B-Mode measurements, scan the patient and then press **Freeze**.



WARNING

DO NOT perform a depth measurement using 4D probes.

Distance Measurement

To make a distance measurement:

1. Press **Measure** once; an active caliper displays.
2. To position the active caliper at the start point, move the **Trackball**.
3. To fix the start point, press **Set**.
The system fixes the first caliper and displays a second active caliper.
4. To position the second active caliper at the end point, move the **Trackball**.
A dotted line connects the measurement points, if preset accordingly.
5. To complete the measurement, press **Set**.

The system displays the distance value in the Results Window.



HINTS

- **Before** you complete a measurement:
 - To toggle between active calipers, press the top Trackball key.
 - To erase the second caliper and the current data measured and start the measurement again, press **Clear** once.
- **After** you complete the measurement:
 - To rotate through and activate previously fixed calipers, adjust **Cursor Select**.
 - To erase all data that has been measured to this point, but not data entered onto worksheets, press **Clear**.

Circumference Measurement (Open Trace)

To trace the circumference of a portion of the anatomy and calculate its length:

NOTE

Set OpenTrace to the Touch Panel in Utility > Measure before perform the measurement. See *Display/Hide a Folder or a Measurement in the Touch Panel* on page 321 for more information.

1. Press **Measure**.
2. Select **Open Trace** from the Touch Panel.
3. Position the caliper at the start point.
4. To fix the trace start point, press **Set**. The caliper changes to an active caliper.
5. Move the **Trackball** to trace the measurement area. A dotted line shows the traced area.
6. To complete the measurement, press **Set**. The system displays the circumference in the Results Window.

Circumference and Area (Ellipse) Measurement

You can use an ellipse to measure circumference and area. To measure with an ellipse:

1. Press **Measure** once; an active caliper displays.
2. To position the active caliper, move the **Trackball**.
3. To fix the start point, press **Set**. The system fixes the first caliper and displays a second active caliper.
4. To position the second caliper, move the **Trackball**.
5. Adjust the **Ellipse** control; an ellipse with an initial circle shape displays.
6. To position the ellipse and to size the measured axes (move the calipers), move the **Trackball**.
7. To increase the size, adjust the **Ellipse** control in a clockwise direction. To decrease the size, adjust the **Ellipse** control in a counterclockwise direction.

8. To toggle between active calipers, press the top **Trackball key**.
9. To complete the measurement, press **Set**. The system displays the circumference and area in the Results Window.



HINTS

Before completing the ellipse measurement:

- To erase the ellipse and the current data measured, press **Clear** once. The original caliper is displayed to restart the measurement.
- To exit the measurement function without completing the measurement, press **Clear** a second time.

Circumference and Area (Trace) Measurement

To trace the circumference of a portion of the anatomy and calculate its area:

1. Press **Measure**.
2. Press the top **Trackball key** to select Trace; a caliper displays.
3. To position the caliper at the start point, move the **Trackball**.
4. To fix the trace start point, press **Set**. The caliper changes to an active caliper.
5. To trace the measurement area, move the **Trackball** around the anatomy. A dotted line shows the traced area.
6. To complete the measurement, press **Set**. The system displays the circumference and the area in the Results Window.



HINTS

Before completing the trace measurement:

- To erase the line (bit by bit) back from its current point, move the **Trackball** or adjust the **Ellipse** control counterclockwise.
- To erase the dotted line but not the caliper, press **Clear** once.
- To clear the caliper and the current data measured, press **Clear** twice.

Circumference and Area (Spline Trace) Measurement

To trace the circumference of a portion of the anatomy and calculate its area:

NOTE

Spline trace is not available through the factory default. The system defaults to trace. To enable spline trace, modify the Measure Key Sequence preset found in Utility > Measure > Advanced preset menu.

1. Press **Measure**.
2. Press the top **Trackball key** to select Spline Trace; a caliper displays.
3. To position the first caliper at the start point, move the **Trackball**.

4. To fix the trace start point, press **Set**. The first caliper turns yellow. The second caliper appears at the same position as the first caliper and is green.

NOTE

When pressing the **Clear** key once, the second caliper disappears and the first caliper is activated.

If **Clear** is pressed again, the first caliper disappears and the Spline trace is cancelled.

5. To position the second caliper, move the **Trackball** and press **Set**. The third caliper appears at the same position.

NOTE

The **Clear** key functionality is the same as noted in the previous step.

The spline trace requires at least three points to draw the trace. Continue setting the points of the trace until the desired points are set.

6. Press **Set** again after the last caliper is fixed to finalize the spline trace. All points are removed from the line and the spline trace turns yellow.

NOTE

Pressing **Set** twice finishes the trace measurement.

If **Clear** is pressed twice when more than 3 points exist on the trace, all points are removed and the first caliper again displays.

Edit the Spline Trace

1. Select **Cursor Select**. The spline trace changes to green and all points appear on the trace as yellow.

A pick-caliper appears on the center of the image and the message "Edit spline trace" displays at the bottom of the screen.

NOTE

The pick-caliper is used to select and move the trace points.




Figure 7-16 • Edit spline trace

Select **Cursor Select** again. The trace is deactivated (changes to yellow) and all points, including the pick-caliper, are removed.

If the previous/next fixed caliper exists on the image, it is activated.

NOTE

Pressing **Clear** at this time removes all points and the trace graphic.

2. Move the pick-caliper to the desired point and press **Set**. The point is activated and turns green.

3. Move the point to the desired position and press **Set**. The point is fixed and turns yellow. The pick-caliper appears on the center of the image.

NOTE

The spline trace is updated at run time.

NOTE

To remove a point, press **Clear** while moving the point. The trace turns green and the remaining points continue to be shown as yellow. If there are less than three points, the spline trace is removed.

4. Press **Set** again. All points are removed from the trace and the trace is shown as yellow.

Intensity (Echo Level) Measurement

To make an echo level measurement:

1. Press **Measure**.
2. Press the top Trackball key to select Intensity. A caliper displays.
3. To position the caliper at the start point, move the **Trackball**.
4. To fix the trace start point, press **Set**. The caliper changes to an active caliper.
5. To trace the measurement area, move the **Trackball** around the anatomy. A dotted line shows the traced area.
6. To complete the measurement, press **Set**. The system displays the echo level, as EL ___ dB, in the Results Window.

NOTE

The echo level measurement is only available on a frozen image, not on a B-paused image.

NOTE

Echo Level is not available through the factory default. To enable echo level, modify the Measure Key Sequence preset, found in the Utility > Measure > Advanced preset.

Doppler Mode Measurements

Five basic measurements can be made in Doppler Mode.

- Velocity
- TAMAX and TAMEAN (Manual or Auto Trace)
- Two Velocities with the Time Interval and Acceleration between them
- Time Interval
- Volume Flow

Prior to performing Doppler Mode Measurements:

1. In the B-Mode part of the display, scan the anatomy you want to measure.
2. Go to the Doppler Mode part of the display.

3. Press **Freeze**.

Velocity

To measure velocity:

1. Press **Measure**; an active caliper with a vertical dotted line displays.
2. To position the caliper at the desired measurement point, move the **Trackball**.
3. To complete the measurement, press **Set**. The system displays the velocity measurement in the Results Window.

Slope (Velocity, Time Interval and Acceleration)

To measure two velocity values, the time interval (ms), and acceleration (m/s^2):

1. Press **Measure**. Press the top Trackball key to select Slope; an active caliper with vertical and horizontal dotted lines displays.
2. To position the caliper at the start point, move the **Trackball**.
3. To fix the start point, press **Set**. The system fixes the first caliper and displays a second active caliper.
4. To position the second caliper at the end point, move the **Trackball**.
5. To complete the measurement, press **Set**. The system displays the two peak end point velocities, the time interval, and the acceleration in the Results Window.

Time interval

To measure a horizontal time interval:

1. Press **Measure**. Press the top Trackball key to select Time; an active caliper with vertical and horizontal dotted lines displays.
2. To position the active caliper at the start point, move the **Trackball**.
3. To fix the start point, press **Set**. The system fixes the first caliper and displays a second active caliper.
4. To position the second caliper at the end point, move the **Trackball**.
5. To complete the measurement, press **Set**. The system displays the time interval between the two calipers in the Results Window.

TAMAX and TAMEAN

Manual Trace

The value measured depends upon the Vol Flow Method preset. The two selections available are: Peak (TAMAX) and Mean (TAMEAN).

To do a manual trace of TAMAX or TAMEAN:

1. Press **Measure**. Press the top Trackball key to select Trace; a caliper displays. Select Manual on the Touch Panel.
2. To position the caliper at the trace start point, move the **Trackball**.
3. To fix the start point, press **Set**.

4. To trace the velocity spectrum boundary, move the **Trackball**.

NOTE

To edit the trace line, move the **Trackball**.

5. To complete the measurement, press **Set**. The system displays the measurement values in the Results Window.

Auto Trace

The value measured depends upon the Vol Flow Method preset. The two selections available are: Peak (TAMAX) and Mean (TAMEAN).

To auto trace TAMAX:

1. Press **Measure**. Press the top Trackball key to select Trace; an active caliper with a vertical dotted line displays. Select Auto on the Touch Panel.
2. To position the caliper at the trace start point in the Doppler spectrum, move the **Trackball**.
3. To fix the start point, press **Set**.
4. To position the vertical caliper at the end point, move the **Trackball**.
5. To complete the measurement, press **Set**. The system automatically fixes both calipers and traces the maximum value between the two points. The system displays this value in the Results Window.

NOTE

When you set the Auto Trace for Both (above and below), the system picks up the maximum power of the signal, NOT the maximum velocity. If the maximum velocity is not the maximum power, the system may not trace accurately. If you want to use maximum velocity, select either Above or Below.

Edit Trace

Auto Trace can be edited after taking an Auto Trace measurement.

1. After taking an Auto Trace measurement, select the measurement result on the result window. The Edit Trace (Edit Peak or Edit Mean) menu window appears.

NOTE

If the system cannot take the trace data correctly from the image, Edit Trace does not work.

2. Select Edit Trace. The first caliper (manual trace caliper) appears on the center of the image. Use the **Trackball** to move the caliper on the trace line to the start point.

NOTE

To cancel Edit Trace at this time, press **Clear**, **Scan**, or **Freeze**.

3. Press **Set** to fix the first caliper. The second caliper appears. Edit the trace manually using the second caliper.

The Ellipse control is used to edit the trace.

NOTE

When pressing the **Clear** key once at this time, the second caliper disappears and the first caliper appears in the center of the image.

NOTE

If you press **Scan** or **Freeze** at this time, the caliper is automatically fixed and the result window updates.

4. Press **Set** to fix the second caliper. The trace and the result window update. The trace data (TAMAX and TAMEAN) are updated, though the other points (e.g. PS, ED) are not updated by trace. The points can be edited with **Cursor Select**.

NOTE

While in Edit Trace, Cursor Select is disabled.

5. Repeat Edit Trace as needed.

Doppler Auto Calc Average Cycle

When using Auto Calc, a selection is available to average a number of cycles automatically. There is also a preset selection in the Utility Imaging PW page for this feature. When using average cycle:

- Selected cardiac cycle lines display on the image. Point calipers are not displayed.
- When changing the number of cycles from 1 to >1, all the data is reacquired from the image, recalculated and updated.
- When multiple cycles are selected in AutoCalc, the average values calculate and display automatically.
- When selecting Peak Value (PV), average cycle is not available.

NOTE

You cannot edit the lines while in Average Cycle. Cursor Select is not available at that time.

NOTE

Average Cycle data is acquired from the display image area only, for both live and frozen. The average cycle data fails if the setting for the number of cycles is larger than the number of image cycles.



Figure 7-17 • PW Touch Panel

Volume Flow - Manual Calc

You perform a manual Volume Flow measurement using the TAMAX plus a Volume Flow coefficient compensation.

1. To perform the Volume Flow measurement using the TAMAX plus a Volume Flow coefficient compensation, in Utility > Measure > Advanced, select the following:
 - Trace = Manual
 - Vol Flow Method = TAMAX [a Volume Flow coefficient MUST be selected for use with TAMAX.]
 - Vol Flow Compensation with TAMAX = [select value from 0.5 to 1.0]

2. Set Auto Calcs to Off via Doppler Mode > Modify Auto Calcs > Off.
3. Select a folder in Doppler Mode > select a calculation folder > select Show All.
4. Select **Volume Flow**. You'll notice that TAMAX is automatically selected.

NOTE

Ensure the caliper is placed in the spectral window when selecting the Volume Flow measurement.

5. Trace the TAMAX. The system prompts to "Mark the first point on the spectral doppler." Press **Set**.
6. The system prompts to "Trace the velocity spectrum boundary." Press **Set**.

NOTE

Use the Trackball to back up while tracing the TAMAX.

7. Trace the vessel diameter. The system prompts to "Mark first point of vessel diameter for volume flow calculation." Press **Set**.
8. The system prompts to "Mark last point of vessel diameter for volume flow calculation." Press **Set**.
9. The Volume Flow is calculated in ml/min.

Volume Flow - Auto Calc

An automatic Volume Flow measurement can be performed using TAMEAN or using the TAMAX and a Volume Flow coefficient.

1. To perform the Volume Flow measurement using the TAMEAN, in Utility > Measure > Advanced, select the following:
 - Trace = Auto
 - Vol Flow Method = TAMEAN
 OR, to perform the Volume Flow measurement using the TAMAX plus a Volume Flow coefficient compensation, select the following:
 - Trace = Auto
 - Vol Flow Method = TAMAX [if TAMAX is used, a Volume Flow coefficient must also be selected.]
 - Vol Flow Compensation with TAMAX = [select value from 0.5 to 1.0]
2. Set Auto Calcs to Live via Doppler Mode > Modify Auto Calcs > Live.
3. Perform the scan.
4. Select **Volume Flow** via Doppler Mode > Modify Auto Calcs > VOLUME FLOW. The system prompts you through the measurement.
5. Take vessel diameter for volume flow calculation. Set the first cursor.
6. Mark last point of vessel diameter for volume flow calculation. Press **Set**.
7. The calculation automatically completes the Volume Flow measurements as ml/min.

NOTE

If the TAMAX coefficient is changed, the Volume Flow is automatically adjusted when in Auto Calcs (but not in Manual Calcs).

Flow Volume (FV)

Flow Volume estimates the volume of blood that flows through a vessel per unit time. It is derived from a vessel's cross-sectional diameter obtained from the B-Mode portion of the image and the mean velocity of flow in the vessel obtained from the Doppler portion of the image. It is measured in milliliters. When the FV measurement is made, FVO is automatically calculated.

To measure flow volume:

1. Select **FV** from Doppler Touch Panel.
2. Place the dotted horizontal line caliper at each of the time base on the Doppler spectrum.
 - If Trace Auto is selected, the waveform is automatically traced.
 - If Trace Auto is not selected, manually trace the desired portion of the waveform.
 The caliper moves to the B-Mode area.
3. Use the Ellipse or Trace method to measure the circumference and area of the vessel.

The flow volume (FV) is calculated and displayed in milliliters. The flow volume output (FVO) is also calculated and displayed in milliliters/minute.

Flow Volume Output (FVO)

This measurement is used to measure the flow volume output in a vessel on the Doppler spectrum. It is measured in milliliters/minute. When the FVO measurement is made, FV is automatically calculated.

Auto vs. Manual Calculations

The same calculations can be performed using either manual or auto calcs.

Manual Calcs

To perform manual calcs:

1. To turn Auto Calcs off and perform manual measurements, choose **Auto Calcs > OFF** on the PW tab of the Touch Panel.
2. After obtaining a waveform, press **Measure**. Choose the appropriate vessel folder or calculation. The system walks you through the measurement.

NOTE

To program which calculations are done under manual calcs when using measurement folders for measuring specific vessels, press the Utility key. Select Measure > Doppler and program your manual calcs (Auto Calcs OFF). Each vessel must be programmed individually and saved after each change.

Auto Calcs

To perform auto calcs:

1. Ensure that the auto calcs function is on by choosing **Auto Calcs > Frozen** or **Live** on the Doppler tab of the Touch Panel.
 - Live: Auto calculation activates when the system is in real-time.
 - Frozen: Auto calculation activates when you press Freeze.
 - Off
2. After obtaining a waveform, press **Measure**. Choose the appropriate vessel folder, side and location. The measurements that are pre-programmed are performed automatically and entered in the worksheet.

To modify auto calcs:

1. Select **Modify Auto Calcs** on the Touch Panel.
2. Choose the measurements to be performed with this preset.
3. To save these measurements:
 - If this is a temporary change, press Return.
 - If this is a permanent change, select **Save as default**.

The measurements are saved and can be performed with the auto calcs function.

Edit Auto Calcs

Auto Calcs can be edited after taking an Auto Trace measurement.

1. After taking an Auto Calc with a trace, select the measurement result on the result window. The Edit Trace menu window appears.

NOTE

If the system cannot take the trace data correctly from the image, Edit Trace does not work.

2. Select Edit Trace. The first caliper (manual trace caliper) appears on the center of the image. Use the **Trackball** to move the caliper on the trace line to the start point.

NOTE

To cancel Edit Trace at this time, press **Clear**, **Scan**, or **Freeze**.

3. Press **Set** to fix the first caliper. The second caliper appears. Edit the trace manually using the second caliper.

The Ellipse control is used to edit the trace.

NOTE

When pressing the **Clear** key once at this time, the second caliper disappears and the first caliper appears in the center of the image.

NOTE

If you press **Scan** or **Freeze** at this time, the caliper is automatically fixed and the result window updates.

4. Press **Set** to fix the second caliper. The trace and the result window are updated. The data is retaken from the trace and updated.

NOTE

While in Edit Trace, Cursor Select is disabled.

The trace data (TAMAX and TAMEAN) is updated, but the other selections (e.g. PS, ED) are not updated by trace. The points can be edited using **Cursor Select** if needed.

5. Repeat Edit Trace as needed.

Modify Auto Calcs

When you select this key, the Modify Calculation menu is displayed as below. In this menu, you select parameters to display in the Auto Vascular Calculation window. Only parameters that can be used by the calculation are displayed.

Select **Save as Default** to save the selected parameters as the default calculations for this application.

Select **Return** to return to the previous Touch Panel screen.

If you select **PV**, all selected parameters are turned off. When you deselect **PV**, the system returns to the previously selected calculation.




Figure 7-18 • Modify Auto Calculation Menu

M-Mode Measurements

Basic measurements that can be taken in the M-Mode portion of the display are:

- Tissue Depth (Distance)
- Time Interval
- Time Interval and Velocity



WARNING

DO NOT perform a depth measurement using 4D probes.

Prior to performing M-Mode measurements:

1. In the B-Mode part of the display, scan the anatomy you want to measure.
2. Go to the M-Mode part of the display.
3. Press **Freeze**.

Tissue Depth

Tissue depth measurement in M-Mode functions the same as distance measurement in B-Mode. It measures the vertical distance between calipers.

1. Press **Measure** once; an active caliper with a vertical and horizontal dotted line displays.
2. To position the active caliper at the most anterior point you want to measure, move the **Trackball**.
3. To fix the start point, press **Set**.

The system fixes the first caliper and displays a second active caliper.

4. To position the second caliper at the most posterior point you want to measure, move the **Trackball**.
5. To complete the measurement, press **Set**.
The system displays the vertical distance between the two points in the Results Window.

Time Interval

To measure a horizontal time interval and velocity:

1. Press **Measure**. Press the top Trackball key to select Time; an active caliper with vertical and horizontal dotted lines displays.
2. To position the caliper at the start point, move the **Trackball**.
3. To fix the first caliper, press **Set**. The system fixes the first caliper and displays a second active caliper.
4. To position the second caliper at the end point, move the **Trackball**.
5. To complete the measurement, press **Set**. The system displays the time interval between the two calipers in the Results Window.

Slope (Time Interval and Velocity)

To measure time and velocity between two points:

1. Press **Measure**. Press the top Trackball key to select Slope; an active caliper with vertical and horizontal dotted lines displays.
2. To position the active caliper at the start point, move the **Trackball**.
3. To fix the start point, press **Set**.
The system fixes the first caliper and displays a second active caliper.
4. To position the second caliper at the end point, move the **Trackball**.
5. To complete the measurement, press **Set**.

The system displays time(s) and slope between the two points in the Results Window.

Worksheet

The worksheet function enables the user to review, edit, delete or print data independently of a report. All measurements and calculations taken during the examination can be viewed at any time using the worksheet.

As measurements are completed, measurement data is included in the appropriate worksheets.

NOTE

Worksheets are not saved if the system crashes.

To View a Worksheet

To view a worksheet, select **Worksheet** on the Touch Panel.
OR

Select **Worksheet** on the measurement summary window.

The system displays the worksheet for the current study.

The OB Worksheet has three sections of information.



Figure 7-19 • OB Worksheet

- 1 Patient data
- 2 Measurement information
- 3 Calculation information

To return to scanning, do one of the following:

- Select **Worksheet**.

- Press **Esc**.
- Select the Exit button.

To view a different worksheet, select the worksheet key for the desired exam.

To view worksheet data for a particular mode, select the key for that mode. To view a worksheet with data for more than one mode, select **Expand**. When Expand is selected, it defaults to view all measurements, noted by mode, on the worksheet.

If a worksheet has more data on a second page, to view the next page, adjust the **Page Change** control.

To Edit a Worksheet



HINTS

Some fields on the worksheet are view only. To see which fields can be changed or selected, move the **Trackball**. As the cursor moves over a field that can be changed or selected, the field is highlighted.

Change Data

1. Select **Worksheet** from any page of the Vascular Calculation Touch Panel.
2. Position the cursor at the field to be changed by moving the **Trackball**.
The cell is highlighted. Press **Set**. The field backlights.
3. Type the new data in the field and move the cursor to another place. Press **Set**. The new data, displayed in blue with an asterisk, is appended to the updated value and resultant value to indicate that it was manually entered.

The average measurements, calculations and ratios are automatically updated to reflect the edited values.

NOTE

If the user moves the cursor to the edited value and presses the Set key once, the value returns to the original value before the edit was made.

Exclude Data

When the user selects a particular value on the Worksheet and selects **Exclude Value**, this value is excluded from result line and resultant value is re-calculated without this value and also calculation values using this value is 'blank.'

1. To highlight the field to delete or exclude, position the cursor with the **Trackball**.
2. Perform one of the following:
 - To exclude the field, select **Exclude Value**.
The data in the field is not visible and is not included in worksheet calculations.
 - To include a previously excluded value, select **Exclude Value**.

Delete Data

1. Select **Worksheet** from any page of the Vascular Calculation Touch Panel Menu.
2. To highlight the field to delete or exclude, position the cursor with the **Trackball**.
3. Select **Delete Value** from the Touch Panel.

For example:

1. If the user measured RI 4 times, the latest 3 sets of RI measurements are displayed in the worksheet.

Table 7-4 • Example of Latest Measurements in Worksheet

Result Number	#2	#3	#4
PS	0.500	0.600	0.700
ED	0.100	0.200	0.300
RI	0.800	0.667	0.571

2. Then the user deleted PS value of #3 from the worksheet.
3. Then, if the user deletes the PS value in column #3 from the worksheet, the whole set of measurements from column #3 is deleted from the worksheet and measurements from column #1 are shifted and displayed, as below.

Table 7-5 • Example of Latest Measurements in Worksheet

Result Number	#1	#2	#4
PS	0.500	0.600	0.700
ED	0.100	0.200	0.300
RI	0.800	0.667	0.571

Examiner's Comment

To type a comment on a worksheet:

1. Select **Examiner's Comments**. The Examiner's Comments window opens.
2. Type comments about the exam.
3. To close the Examiner's Comments window, select **Examiner's Comments**.



Figure 7-20 • Examiner's comments field

Volume Measurement Value Off

Select the method type **Off**. The value field becomes blank.

To Select a Method

1. Move the cursor over the value in the method column and press **Set**.
2. The pull-down menu displays. Move the cursor to select the method and press **Set**. The selected method is displayed in the column.



- 1 Avg.: Average of the measurements taken
- 2 Max.: Maximum measurement
- 3 Min.: Minimum measurement
- 4 Last: Last measurement that was taken

Figure 7-21 • Pop-up menu of methods - example

Delete All Worksheet Values

To delete all worksheet values on a worksheet:

1. When the Worksheet is displayed on the monitor, select **Clear**; the following warning message appears:



Figure 7-22 • Delete All Warning Message

2. Select OK to delete all.
Select Cancel to cancel the deletion.

Recording Worksheet

The worksheet can be saved as would any ultrasound image. Once it is displayed on the screen, it can be recorded on the DVR, printed on the B/W printer, stored on media with the Image Archive option, or placed on regular paper with a line printer.

OB Worksheet

Patient Data

The Patient data section, at the top of the worksheet, lists information from the Patient Data Entry screen.

The following fields can be selected:

- FetusNo – If this is a multi-gestational patient, select the fetus in this field. **Fetus** selection can be adjusted to change the fetus.
- CUA/AUA – Select the ultrasound age calculation method
 - Composite Ultrasound Age (CUA) – Regression calculation
 - Average Ultrasound Age (AUA) – An arithmetic average

Select the method in this field or adjust the **Select CUA/AUA** control.

NOTE

CUA/AUA is only available when USA OB Type is selected in the Utility > System > System Measure menu.

Information can be entered in the following fields:

- FetusPos – Enter information about the fetus position.
- PLAC – Enter information about the placenta.

Calculation Information

This section of the worksheet provides calculation choices and lists calculation results.

- EFW – Lists the parameters used to calculate EFW. This is followed by the calculation result.

To change which parameters are used:

- Select this field or press **Select EFW**.
- Select the desired parameters.

- EFW GP – Lists the source used to calculate EFW–GP (growth percentile). This is followed by the growth percentile.

To change the source:

- Select this field or press **Select GP**.
- Select the desired source.

The remaining calculation information shows ratios for several measurements, and the Cephalic Index (CI).

The worksheet shows if any of the ratios are out of range (OOR). Out of range indicates one of the following:

- The measurement is out of the normal range based on the gestational age that is calculated from the LMP. The system determines OOR from the ultrasound age compared to the gestational age. The gestational age is calculated from the last menstrual period or the estimated delivery date.
- The measurement is outside of the range for the data used in the calculation. That means that the measurement is either less than or more than the range of measurements used to determine fetal age based on the measurement.

For more information about how to use the worksheet, see *Worksheet* on page 306.

Vascular Worksheet

Intravessel Ratio

To calculate the Intravessel ratio on the Vessel Worksheet page:

1. Select **Intrav. Ratio** from the Touch Panel. The Intravessel Ratio pop-up window displays in the header section of the worksheet.




Figure 7-23 • Intravessel Pop-up Window

2. Select the first velocity. The value displays in the window.



Figure 7-24 • Intravessel Ratio One

3. Select the second velocity. The second value and Result value display in the window.



Figure 7-25 • Intravessel Ratio Two

- To save the Intravessel ratio to the Vessel Summary, move the cursor to **Save** and press **Set**.
- To clear values, move the cursor to **Clear** and press **Set**.
- To cancel and exit Intravessel ratio, move the cursor to **Cancel** and press **Set**.

NOTE

Intravessel Ratio is only displayed and saved in the Vessel Summary as Intra-Ratio.

Vessel Summary

The Vessel Summary is designed to automatically display measurements made at specific anatomical sites. Calculated ratios are automatically summarized and displayed.

The Vessel Summary can be displayed at any time during the exam by selecting **Vessel Summary** from the Vascular Worksheet Touch Panel.

	Right			Left		
	PS	ED	AC	PS	ED	AC
Prox CCA	8.1 cm/s	1.8 cm/s	0.04g	11.0 cm/s	1.0 cm/s	0.04g
Mid CCA	22.9 cm/s	11.7 cm/s	0.04g	22.4 cm/s	9.3 cm/s	0.04g
Dist CCA	26.3 cm/s	8.9 cm/s	0.04g	29.4 cm/s	11.5 cm/s	0.04g
Vessel Ratio	6.57					

Figure 7-26 • Vessel Summary Example

1. The first row, indicating Right or Left, is not displayed when the side is not defined in the vessel. In the third column on the second line, select the calculations. Move the cursor to the third column, and the pop-up menu is displayed. The selected parameter is displayed in every third column.




Figure 7-27 • Pop-up menu

2. Vessel Name with location information.
3. Check Box. Use to select the vessel velocity for calculating the vessel ratio (ex. ICA/CCA). Only one location (position) in a vessel can be selected.
4. Result value column. This value cannot be changed or excluded from this page.
5. Calculation name and result. ICA/CCA: The ICA/CCA ratio selects the highest systolic ICA and CCA velocities when calculating this ratio, and displays the velocities.

Carotid Study

In the configuration page for ICA/CCA ratio, specify which portion of the CCA vessel (Prox, mid, distal) is chosen. The selections on the Vessel summary can be overridden.

The ICA/CCA ratio can be configured for either systole or diastole.

General Measurements and Calculations

The vertebral vessel also has systole and diastole selections. In the summary page, there is a box to select flow reversal for vertebral flows. The choices are Ante (Antegrade), Retr (Retrograde), and Abs (Absent).

To select the method:

Move cursor to the box and press **Set**. After the pop-up menu (Blank, Ante, Retr, Abs) displays, select from a menu of choices. The selected choice displays in the column.

The box is independent of Left and Right.

Renal Artery Study

For renal arteries, RENAL/AORTIC ratio (RAR) can be calculated based on peak systolic velocities.

The two renal summary pages can be combined, with a heading to separate the different measurements (main renal, intra renal). Scroll between the measurements. The most commonly used measurement, the main renal artery, is the default.

Lower Extremity Artery Study

For the lower extremity artery, an intra vessel ratio (assessing pre vs. stenotic velocities) is required. You can specify which (ratio is stenotic/pre).

The intra-vessel ratio needs to be available for all vascular measurements. This appears on the worksheet only if used.

Recording Worksheet

The worksheet can be saved like any ultrasound image. Once it displays on the screen, it can be recorded on the DVR, printed on the B/W printer, stored on media with the Image Archive option, or placed on regular paper with a line printer.

Print All Pages

When there are multiple pages, a "Print All Pages" button displays on the Touch Panel. The default is ON. If Print All Pages is set to ON, select the Print button; all pages are stored or printed, depending upon the print button configuration.

Anatomical Survey

The Anatomical Survey page provides a checklist that indicates which anatomy was imaged, and its appearance.



Figure 7-28 • OB Anatomical Survey



Figure 7-29 • Abdominal Anatomical Survey



Figure 7-30 • GYN Anatomical Survey

Editing

1. To activate the Anatomical Survey, select **Anatomy** on the Worksheet Touch Panel.
2. Fill the required field.

Table 7-6 • Abdomen Anatomical Survey

Field	Description
ANATOMY	Enter the following information for each part of the anatomy imaged: 1. Imaged?: Check the box for the part of the anatomy imaged. 2. Appearance: If the Imaged? box was checked, indicate whether the appearance was normal or abnormal. Additional anatomy can be added to this worksheet. Up to 9 additional items can be added. Move the Trackball to the blank field. Enter the anatomy name.
COMMENTS	Free text

Table 7-7 • OB Anatomical Survey

Field	Description
Fetus Pos	Indicate the fetal position within the uterus.
PLAC	Identify the location of the placenta.
ANATOMY	Enter the following information for each part of the anatomy imaged: 1. Imaged?: Check the box for the part of the anatomy imaged. 2. Appearance: If the Imaged? box was checked, indicate whether the appearance was normal or abnormal. Additional anatomy can be added to this worksheet. Up to 9 additional items can be added. Move the Trackball to the blank field. Enter the anatomy name.
BIOPHYSICAL	The score is _ of 10 possible total points, depending upon the number of parameters entered. Enter the following information to assess the fetus's biophysical well-being.
Movement	Type 0, 1 or 2
Tone	Type 0, 1 or 2
Breathing	Type 0, 1 or 2
Fluid	Type 0, 1 or 2
Reactive NST (Reactive non-stress test)	Type 0, 1 or 2
COMMENTS	Free text

Table 7-8 • GYN Anatomical Survey

Field	Description
ANATOMY	Enter the following information for each part of the anatomy imaged: 1. Imaged?: Check the box for the part of the anatomy imaged. 2. Appearance: If the Imaged? box was checked, indicate whether the appearance was normal or abnormal.
	Additional anatomy can be added to this worksheet. Up to 9 additional items can be added. Move the Trackball to the blank field. Enter the anatomy name.
COMMENTS	Free text

3. Select **Exit** to return to the Scan screen.
Select **Worksheet** to return to the Worksheet.

NOTE

The patient specific contents input on the Anatomical Survey page are returned to the factory default settings after starting a new patient.

Measurement and Calculation Setup

Measurements and studies are organized for typical work flows. This set up can be modified to specify which studies are in each exam category and which measurements and calculations are in each study. The measurements available on the Touch Panel can be changed. The ultrasound system allows quick and easy set up.

This section describes how to:

- Change a study to include different measurements
- Add a new study or measurement
- Remove a study from an exam category
- Change measurement parameters
- Create a measurement formula to correctly handle unit conversions
- Edit user-defined calculations
- Define application-specific measurement parameters
- Specify the default manual calc measurements for a selected study or folder

Starting Study and Measurement Setup

Changes to studies and measurements can be made in the Measurement & Analysis screen. To open the screen:

1. On the Touch Panel, select **Utility**.
2. On the Touch Panel, select **Measure**.
3. The system displays the Measurement & Analysis screen on the monitor display.



Figure 7-31 • Measurement & Analysis Screen

- **Selection Menu:** Select exam category, study, or measurement folder/measurement.
- **Measurement Menu:** Add and delete studies (folders) and measurements; select mode.
- **Folder or Measurement:** Define studies and measurements. This section changes between Folder and Measurement, depending on what is selected in the Selection menu.

Select the Exam Category

To select the exam category to work with:

1. Move the cursor to the exam category at the top of the Selection menu.
2. Press **Set**.

The system displays a list of exam categories.

3. Move the cursor to the desired exam category.
4. Press **Set**.




Figure 7-32 • Select the Exam Category

Select the Study

1. Move the cursor to the current study in the Selection menu.



Figure 7-33 • Select the Study

2. Press **Set**. All study displays.



Figure 7-34 • Select the Study

3. Move the cursor to the desired study you want to open.
4. Press **Set**.

Select a Measurement Folder or a Measurement

1. Move the cursor to the desired measurement folder or the desired measurement in the Selection menu.
2. Press **Set**.
 - When a measurement folder is selected, the folder information displays.



Figure 7-35 • Folder - Example

- If a measurement is selected, the Measurement information is displayed.




Figure 7-36 • Measurement - Example

Select the Measurement Mode

In the Measurement menu, select the measurement mode:

- 2D (B-Mode)
- MM (M-Mode)
- Dop (Doppler Mode)
- Plot (Plot Mode—The measurement on the plot graph of the TIC/QAnalysis)




Figure 7-37 • Measurement Menu

Display/Hide a Folder or a Measurement in the Touch Panel

The Folder section of the Measurement & Analysis screen has two lists of folders and measurements. This is where items to go in a study or folder are specified.

- **Available Folders and Measurements.** The list contains all possible folders and measurements for the selected study or folder.
- **Measure & Study.** The list defines where the folder or measurement is located on the Touch Panel.

To display an item in the Touch Panel:

1. In the Measure & Study list, move the **Trackball** to highlight which folder to put the item in, and press **Set**.

2. Move the **Trackball** to highlight an item in the Available folders and measurements list, and press **Set**. The selected item is assigned to the Touch Panel.



Figure 7-38 • Select the Position and an Item



Figure 7-39 • Measure & Study list: New Item Added

The selected item is now displayed in the Touch Panel and the Summary Window.

Touch Panel Positions

Each Touch Panel has 25 positions, five across and five down. The items in the first row across are numbered 1–5, in the second row 6–10, and so on. Positions 1–5 are system programmed and cannot be edited.



Figure 7-40 • Measure & Study Display with Touch Panel Positions, Page 1

Page 2 of the Touch Panel has 25 positions, five across and five down. The items in the first row across are numbered 26-30, in the second row 31-35, and so on. Positions 26-30 are system programmed and cannot be edited.



Figure 7-41 • Measure & Study Display with Touch Panel Positions, Page 2

Move, Remove or Hide Touch Panel Items



Figure 7-42 • Move (1), Remove (2) and Hide (3) Icons

Move Touch Panel Items

To move items on the Touch Panel, change the item position in the Measure & Study list.

1. Highlight an item in the Measure & Study list.
2. Move the cursor to the Up or Down Arrow Icon and press **Set** (see 1 in *Move Touch Panel Items* on page 324).

The item is displayed at the selected position on the Touch Panel.

Remove Touch Panel Items

To remove items from the Touch Panel, remove the item from the Measure & Study list.

1. Move the **Trackball** to highlight the item in the Measure & Study list and press **Set**.
2. Select the Trash Icon to the right of the list (see 2 in *Move Touch Panel Items* on page 324).

The system removes the item from the Measure & Study list and from the Touch Panel. The item is still listed in the Available folders and measurements list.

Hide Touch Panel Items

Modifiers for Stenosis, as well as Prox, Mid, Distal and Origin location identifiers, can be hidden on the Touch Panel using the Hide command in the Measure & Study list.

1. Move the **Trackball** to highlight the item in the Measure & Study list and press **Set**.
2. Select the Hide Icon to the right of the list (see 3 in *Move Touch Panel Items* on page 324).

The system hides the modifier from the Measure & Study list and from the Touch Panel. Once a modifier is hidden, another measurement can be assigned to the Touch Panel location.

To reestablish the hidden modifier, highlight the hidden modifier in the Measure & Study list, press **Set** and select the Hide Icon.

NOTE

If a modifier is hidden in either 2D or Doppler mode, it will be hidden in both modes.

Setting up an Automatic Measurement Flow

In some cases, related measurements are put in a measurement folder. This allows measurements to be organized logically. It also allows the system to be specified to automatically start each measurement in a folder, one after the other. This is the automatic sequence feature. To use this feature:

1. In the Selection menu, select the folder that contains the desired measurements.
2. In the Folder section, select Auto sequence.



Figure 7-43 • Measurement & Analysis Screen: Auto Sequence

Change the Tool Used to Make a Measurement

Measurement changes can be made. For example, Head Circumference can be measured with an ellipse, a trace, or two distances. The desired measurement tool can be selected.

1. Select the measurement to change in Selection menu.
2. In the Measurement section, select the desired tool from the Tool list.

NOTE

If the Tool field is gray, it cannot be changed.




Figure 7-44 • Change the Tool

NOTE

The diagram to the right of the Tool list shows the measurement type. In the example in *Figure 7-44 • Change the Tool* on page 326, ellipse is selected and the diagram shows an ellipse.

Adding Folders and Measurements

Adding a Folder

When adding a folder, it can be a study, or a measurement folder that includes related measurements.



Figure 7-45 • Measurement & Analysis: Add Folder

1. In the Selection menu, select the study or folder to add the folder to.
2. Select the mode in the Measurement menu.
3. In the Measurement menu section, select Add folder.
 - If Blank is selected, the system adds a folder with a name such as USERDEFS1. It is listed in the Selection menu.
 - For Generic study, only Blank can be used.
 - To use an existing folder, select Insert, and then select a folder from the list.




Figure 7-46 • Add folder Window

4. Select the user-defined folder in the Selection menu.
- NOTE**
An existing folder name cannot be changed.

5. Move the **Trackball** to the Name field and press **Set** twice.
Type the name.

NOTE
DO NOT use "single quotes" for a parameter name, a measurement name, a folder name or an author name.

6. Add a measurement to the folder.

Adding a Measurement



DANGER

Remember that you are responsible for confirming the correctness and accuracy of the user input formula that you enter or is entered into the system on your behalf. Failure to confirm the correctness and accuracy of user defined calculations could lead to patient injury.

You can add a measurement to the folder.

NOTE

DO NOT use “single quotes” for a parameter name, a measurement name, a folder name or an author name.

1. In the Selection menu, select the study or folder where to add the measurement to.
2. In the Measurement menu section, select Add measurement.

The system displays the Add Measurement window.



Figure 7-47 • Add Measurement Window

3. Do one of the following:
 - To create this measurement from a copy of an existing measurement, select Use copy of, and then select a measurement from the list. The list includes all measurements defined for the current exam category and selected mode.
- NOTE**
This only applies to OB and Cardiac.
- To use an existing formula, select Insert, and then select a measurement from the list. The list includes all measurements defined for the current exam category and selected mode. This formula cannot be edited.
 - To create a blank new measurement, select Blank.
4. Select OK.
 - If a blank measurement was created, the system adds a measurement with a name such as USERDEFM3.
 - If a measurement was created from a copy of an existing measurement, the system lists the measurement and its parameters in the Measurement section.

- When creating a new measurement, the measurement name is automatically highlighted. Type a name for the new measurement. The name of a measurement created from a copy can be changed.

NOTE

2D Dual Caliper, 2D Dual Area, 2D Dual Ellipse, and 2D Dual Spline Trace are not available through the factory default. To enable these measurements, add a new measurement using “2D Dual Caliper,” “2D Dual Area,” “2D Dual Ellipse,” or “2D Dual Spline Trace” tool.

Changing Measurement Parameters

**DANGER**

Remember that you are responsible for confirming the correctness and accuracy of the user input formula that you enter or is entered into the system on your behalf. Failure to confirm the correctness and accuracy of user defined calculations could lead to patient injury.

You can add a measurement to the folder.

To change a measurement parameter:

- In the Selection menu, select the measurement.
- To change the name of the Parameter, move the **Trackball** to the parameter name and press **Set** twice. Type a name for the parameter.

Adding Measurement Parameters

To add a measurement parameter:

- In the Selection menu, select the measurement.
- To change the tool used to make a measurement:

In the Measurement section of the Measurement & Analysis screen, select the desired tool from the Tool list. Select the arrow to display the drop-down list.

NOTE

If the Tool field is gray, it cannot be changed.

- If necessary, check Fetus (OB only), Location (Loc), or Side:
 - Fetus: If this is an OB measurement, check this box. (Default ON).
 - Location: If this measurement includes a Prox, Mid, or Dist location, check this box.
 - Side: If this measurement includes a Left or Right side, check this box.
- In the Measurement section, move the **Trackball** to an empty line at the bottom of the Parameter list. Press **Set**.
The system adds a parameter with a name of (Name).
- To change the name of the Parameter, move the Trackball to the (Name) and press **Set** twice. Type a name for the parameter.
- Move the **Trackball** to the Tool result field, and double click the **Set** key.
The Edit Formula window is displayed.



Figure 7-48 • Edit Formula

7. To create a formula:
 1. In the Value Type field, select a value.
 2. Do one of the following:
 - Type a formula in the Formula field.
 - Select formula components from the Operators, Parameters, and Functions drop-down lists. When a component is selected, the system displays it in the Formula field.
8. To test the formula, select Check.

If there are no problems, the system displays "Syntax OK!"

If there are any problems with the formula, the system displays an Error message in place of the Formula field label.
9. When the formula is correct, select OK to save it.

The Edit Formula window closes. The formula is displayed in the Tool result field.

Formula Unit Conversion

When a formula is created, the system changes the calculation result into an output unit as defined in the following table.

Table 7-9 • Formula Unit Conversion

	Unit	Conversion (coefficient value)
Time	s	x1
	ms	x1,000
	min	x0.0167
	h	x0.00027778
Ratio	%	x100
Frequency	bpm or BPM	x1.0
Angle	rad	x1.0
	deg	x57.2958
	grad	x63.6620

	Unit	Conversion (coefficient value)
Distance	cm	x100
	m	x1
	dm	x10
	mm	x1,000
	inch	x39.37
	feet	x3.281
	pixels	x1
Velocity	m/s	x1
	dm/s	x10
	cm/s	x100
	mm/s	x1,000
	inch/s	x39.37
Acceleration	m/s ²	x1
	dm/s ²	x10
	cm/s ²	x100
	mm/s ²	x1,000
	inch/s ²	x39.37
Area	m ² or m ^{^2}	x1
	dm ²	x100
	cm ² or cm ^{^2}	x10,000
	mm ² or mm ^{^2}	x1,000,000
	inch ²	x1550
Volume	m ³	x1
	dm ³	x1,000
	cm ³	x1,000,000
	l	x1,000
	dl	x10,000
	cl	x100,000
	ml	x1,000,000
	gallon	x264,178
	quart	x1056.71

	Unit	Conversion (coefficient value)
Volume Flow	m ³ /s	x1
	dm ³ /s	x1,000
	cm ³ /s	x1,000,000
	mm ³ /s	x1,000,000,000
	l/s	x1,000
	dl/s	x10,000
	cl/s	x100,000
	ml/s	x1,000,000
	m ³ /min	x60
	dm ³ /min	x60,000
	cm ³ /min	x60,000,000
	mm ³ /min	x60,000,000,000
	l/min or L/min	x60,000
	dl/min	x600,000
	cl/min	x6,000,000
	ml/min	x60,000,000
	ml/m ²	x1,000,000
Pressure	mmHg	x1
	Pa	x133.322
	kPa	x0.133322
	bar	x0.00133322
Pressure/Time	mmHg/s	x1
Mass	kg	x1
	g	x1,000
	ounce	x35.273962
	pound	x2.2046226
Others	l/minm ²	x60000.0
	g/m ²	x1000.0
	cm/m ²	x100.0
	cm ² /m ²	x10000.0
	ml/kg/min	x60000000.0

For example, when a Volume formula is created:

$$\text{Vol [ml or cm}^3\text{]} = 0.523598 \times \{D1\} \times \{D2\} \times \{D3\}$$

(D1, D2, and D3 indicate a measurement result.)

In this case, the measurement (D1, D2, and D3) is a distance measurement, so the measured data is a meter [m] unit according to the above table.

To change into a milliliter, the system multiplies each measurement value by 100. As a result, it multiplies a formula by 1,000,000.

The standard unit of volume is a cube meter, so the system multiplies the result by 1,000,000.

The system multiplies the calculation result by the coefficient and converts it. To get a correct result, when defining the formula, the coefficient itself must be converted, such as the coefficient of 10^4 .

For example, to define the following formula:

$$efg[g] = 10^{(1.5662 - 0.0108 \cdot \{P1\} + 0.0468 \cdot \{P2\} + 0.171 \cdot \{D1\} + 0.00034 \cdot \{P1\} \cdot \{P1\} - 0.003685 \cdot \{P2\} \cdot \{D1\})}$$

D1[cm]: Distance

P1[cm]: Perimeter

P2[cm]: Perimeter

The system defines the standard value of each measurement as a meter [m]. If the unit of each measurement value of this formula is defined as centimeter [cm], the formula must be defined as follows:

$$efw[g] = 10^{(1.5662 - 0.0108 \cdot \{P1\} \cdot 100 + 0.0468 \cdot \{P2\} \cdot 100 + 0.171 \cdot \{D1\} \cdot 100 + 0.00034 \cdot \{P1\} \cdot \{P1\} \cdot 100 \cdot 100 - 0.003685 \cdot \{P2\} \cdot \{D1\} \cdot 100 \cdot 100)}$$

(This converts each measurement value to a centimeter [cm], since the system standard unit is a meter [m].)

The output unit of this formula is a gram. Since the standard unit of the system is defined as a kilogram [kg], the system multiplies the output by 1,000.

Because the output of this formula is defined as a gram, it is necessary to define the formula as follows.

$$efw[g] = 10^{(1.5662 - 0.0108 \cdot \{P1\} \cdot 100 + 0.0468 \cdot \{P2\} \cdot 100 + 0.171 \cdot \{D1\} \cdot 100 + 0.00034 \cdot \{P1\} \cdot \{P1\} \cdot 100 \cdot 100 - 0.003685 \cdot \{P2\} \cdot \{D1\} \cdot 100 \cdot 100)} / 1,000$$

As shown, an exact calculation result can be obtained.

Editing Calculations

To modify user-defined calculations:

1. Select Add Measurement from the Measurement menu. The system displays the Add Measurement window.
2. Select Blank and OK.
3. Type the appropriate name and select "Calculation" from the Tool pull-down menu.




Figure 7-49 • Measurement Window

4. Type the parameter name.

5. Double click on the = Calculated symbol under Tool Result. The Edit Formula window displays.
6. Select OK.
7. In the Measurement menu section, select Edit Calc. The Modify User CALC window displays.



Figure 7-50 • Edit Calc

8. In the User Defined list, select the calculation to be modified, then select OK.



Figure 7-51 • Modify User CALC Window

9. The Measure tab for user-defined calculations displays. Double click on the equals sign symbol under Tool Result for the desired parameter.
10. Edit the formula as needed and select OK.

Deleting a Folder or Measurement

NOTE


You can only delete user-defined folders or measurements. You cannot delete default system folders or measurements.

1. Select the folder or measurement in the Selection menu.
2. In the Measurement menu section, select the Trash icon next to Delete measure and study.

M & A Advanced Preset

The Advanced tab allows application-specific values for certain parameters to be specified.

1. On the monitor display, select the Advanced tab.



M&A Category: Display and select current exam category.
Parameter: Lists application specific parameters.
Value: Select the value for a parameter.

Figure 7-52 • M&A Advanced Preset Menu

2. Select an exam category.
The Parameters list displays parameters for the selected category.
3. Select a value for a parameter.

NOTE

The parameters that appear are category dependent.

Table 7-10 • M&A Advanced

Preset Parameter	Description
Absolute Value	Displays the absolute value of the Doppler Velocity measurement (On or Off)
Default Cursor Position	Image Center/Summary Window Select the display position of the cursor when the measurement key is pressed.
Display Angle Correct Value	On or Off

General Measurements and Calculations

Preset Parameter	Description
Display Sample Volume Depth (TCD)	On or Off
Display Summary Window	On or Off
Doppler Default Method	Avg, Max, Min or Last
Heart Rate Cycle	1, 2, 3, 4, 5, 6, 7, 8, 9, or 10 NOTE: For Cardiac, you can select only "1".
Keep Graphics with Cine Scroll	If "On" is selected, the measurement graphics remain while in CINE scroll. The measurement graphic redisplay on the frame where the measurement was taken in B-Mode.
Start M&A on a Freeze (B)	Off: Select measurement manually on Freeze
Start M&A on a Freeze (M)	On: Measurement menu appears automatically on Freeze.
Start M&A on a Freeze (D)	Caliper: Measurement menu and caliper appear automatically on Freeze.
Measure Key Sequence (B/CF)	2 Sequences: Dist, Trace; Dist, Spline 2 Sequences: Dist, Open Trace; Dist, Open Spline 3 Sequences: Dist, Trace, Spline; Dist, Spline, Trace; Dist, Spline, Intensity; Dist, Trace, Intensity; Dist, Trace, Open Trace; Dist, Spline, Open Trace 4 Sequences: Dist, Trace, Spline, Intensity; Dist, Spline, Trace, Intensity; Dist, Spline, Trace, Open Trace; Dist, Trace, Open Trace, Spline
Show Result Parameters	Preview or After Set cursor: Preview: Displays while taking the measurement. After Set Cursor: Displays after completing the measurement.
Default Location	Off, Prox, Mid or Dist
Default Side	Left, Right or Off
PI Calc Method	MD or ED
RI Calc Bidirectional Flow	On or Off
RI Calc Method	MD or ED
Shear Measure Size	Sets the Default Diameter size of the shear wave measurement circle.
Shear Measure Fixed Size	On or Off
Shear Calculation Method	Specify Mean or Median Mean averages all of the shear wave points within the measurement circle. Median sorts, then selects the middle point of all points within the measurement circle.
Shear Units Determine Folder	If shear wave units have been preset via the Display Units on the Utility > System > System Imaging page and On is specified, the unit specified pre-selects the measurement folder. If m/s is specified as the unit, then the Velocity folder is used; if kPa is specified as the unit, then the Stiffness folder is used.

Preset Parameter	Description
Show Location Select Button	Both on B and Doppler, Doppler only or No Display NOTE1: Only Abdominal, Vascular, Obstetrics and Gynecology have this preset. NOTE2: For Obstetrics and Gynecology, only Doppler only or No Display can be selected.
Show BM Folder Name on Worksheet	On or Off
Show Measure Name on Worksheet	On or Off
Show Point Velocity	On or Off
Show Tissue Depth	On or Off
Keep Result Window	Auto, On or Off
Trace	Auto or manual
Length Unit	mm, cm, Default
Velocity Unit	mm/s, cm/s, m/s, Default
Acceleration Unit	mm/s ² , cm/s ² , m/s ² , Default
Area Unit	mm ² , cm ² , Default
Volume Unit	cm ³ , ml, l, Default
Volume Flow Unit	cm ³ , ml, l, Default
Time Unit	ms, s, Default
Show Area Value While Tracing	On or Off
Vol Flow Method	TAMEAN or TAMAX
Vol Flow Compensation with TAMAX	If TAMAX is selected as the volume flow method, the coefficient to use MUST be specified. Select from 0.5 to 1.0.
Worksheet Default Display	Mode/Expand (Abdominal, Small Parts, Obstetrics, Gynecology, Urology and Pediatrics) or Worksheet Summary (Vascular)
Primary Worksheet	For Abdominal and Small parts. Abdominal: Worksheet or Summary, Small parts: Worksheet or SWE summary
Doppler AutoCalc Velocity Unit	Velocity, Hz, Both or Auto
Default CCA location for ICA/CCA ratio	Prox/Mid/Dist/Off Select the default location of CCA which is used for the ICA/CCA ratio.
MCA/ICA Ratio	TAMAX or PS
Default ICA location for MCA/ICA ratio	Prox/Mid/Dist/Off Select the default location of CCA which is used for the MCA/ICA ratio.
WMS Freeze Loop at ES	On or Off
WMS Segment Model	16 segments or 18 segments
WMS Initial Scoring	Undefined or Normal
WMS Scoring Legend	ASE, European or Asian
Hip Orientation	Cranial-left or Caudal-left
Show area value while tracing	On or Off

Preset Parameter	Description
Measure Assistant BPD Method	Outer to Inner, Outer to Outer
Restrict Breast Contour Caliper Edit	On or Off
AFI/AutoEF auto processing	Off, delay 1s, delay 2s, delay 3s or delay 4s
AFI/AutoEF ROI method	Auto ROI or 3 Points
AFI/AutoEF YOYO	Play or Stop
AFI Default Color Palette	Red-Blue, Green-Yellow-Red
AFI segment model	17 segments, 18 segments
AFI PSS/PSI Mode	PSS only, PSS&PSI
Auto sequence - Trigger for next measurement	Measure or Freeze
Link contents	Select to link parameters by cardiac cycle.
Default value for RAP	Blank, 3, 5, 7, 8, 10,15 (system default is 3)
UGAP Print and Unfreeze	On or Off
Length unit	Default, mm or cm
Velocity unit	Default, mm/s cm/s, or m/s
Acceleration unit	Default, mm/s ² , cm/s ² , or m/s ²
Area unit	Default, mm ² , or cm ²
Volume unit	Default, cm ³ /s, ml/s, l/s, cm ³ /min, ml/min, or l/min
Volume Flow unit	Default, cm ³ /s, ml/s, l/s, cm ³ /min, ml/min, or l/min
Time unit	Default, ms, or s
Primary Worksheet	Worksheet or Summary

Doppler Tab - Modify Calculation

The Doppler tab allows the parameters for manual calculations to be preset.

1. On the monitor display, select the Doppler tab.
2. The following example describes how to configure the Carotid Doppler calculations.
Select Vascular next to M&A Categories. The Vascular measurement category is displayed.
3. Select Carotid. The available calculations are displayed in Modify Calcs.



Figure 7-53 • M&A Doppler Preset Menu

4. Check the desired calculations to be performed.

Application Measurement Preset

The Application Measurement presets allow different calculation packages to be available under different application presets.

The presets allow configuration of the Measurement Categories and Measurement Exam Calcs. These presets are found on the Utility > Application > Measurements screen.




Figure 7-54 • Application Measurements Menu

Chapter 8

Application M&A

In this section

General Information.	342
Abdomen.	343
Small Parts.	344
OB.	346
GYN.	371
Cardiac.	373
Vascular.	377
Urology.	393
Pediatrics.	397

General Information

Measurements and calculations derived from ultrasound images are intended to supplement other clinical procedures available to the attending physician. The accuracy of measurements is not only determined by the system accuracy, but also by use of proper medical protocols by the user. When appropriate, be sure to note any protocols associated with a particular measurement or calculation. Formulas and databases used within the system software that are associated with specific investigators are so noted. Be sure to refer to the original article describing the investigator's recommended clinical procedures.

General Guidelines

New Patient information must be entered before beginning an exam. See *Beginning an Exam* on page 146 for more information.

Any measurement can be repeated by selecting that measurement again from the Touch Panel.

The system retains all measurements, but the worksheet retains only the last six measurements of each type.

Abdomen

Abdominal measurements offer a few different types of measurement studies. Select the desired study.

- Generic – Common to all applications. See *Generic Measurements* on page 278 for more information.
- Abdomen
- Renal
- Bypass Graft
- Aorta Iliac
- Vascular Renal
- Mesenteric
- Abdomen Vein




Figure 8-1 • Abdomen Exam Category Touch Panel

Small Parts

B-Mode Measurements

The Small Parts exam category includes the following two folders:

- Generic—Common to all applications. See *Generic Measurements* on page 278.
- Small Parts, which includes the breast, thyroid and scrotal measurement packages.




Figure 8-2 • Small Parts Exam Category Touch Panel

Thyroid

Thyroid Left/Right

Each of these is a standard distance measurement. Length and height are typically measured in the sagittal plane. Width is measured in the transverse/axial plane.

To measure thyroid length, width, or height:

1. On the **Exam Calcs**, select **Small Parts**.
2. Select **Thyroid**.
3. Select **Lt or Rt Thyroid**. Change the orientation (side), if necessary.
4. Select **Thyroid L**, **Thyroid W**, or **Thyroid H**.

An active caliper displays.

5. Perform a standard distance measurement.

Isthmus AP

To measure the anterior/posterior isthmus tissue, perform a distance measurement.

Scrotal

Scrotal Left/Right

Each of these is a standard distance measurement. Length and height are typically measured in the sagittal plane. Width is measured in the transverse/axial plane.

To measure scrotal length, width, or height:

1. On the **Exam Calcs**, select **Small Parts**.
2. Select **Scrotal**.
3. Select **Lt or Rt Testicle**. Change the orientation (side), if necessary.
4. Select **Testicle L**, **Testicle W**, or **Testicle H**.

An active caliper displays.

5. Perform a standard distance measurement.

Epididymis

To measure the epididymis structure, perform a distance measurement.

OB

Out of Range – If the system indicates that a measurement is out of range (OOR), it means one of the following:

- The measurement is out of the normal range based on the gestational age that is calculated from the LMP. The system determines OOR from the ultrasound age compared to the gestational age. The gestational age is calculated from the last menstrual period or the estimated delivery date.
- The measurement is outside of the range for the data used in the calculation. That means that the measurement is either less than or more than the range of measurements used to determine fetal age based on the measurement.

NOTE

Calculation formulas are listed in the Advanced Reference Manual.

NOTE

Nuchal Translucency is not available through the factory default. To enable Nuchal Translucency, add NT to the measurement folder in Utility > Measure > M&A > Add measurement (Insert).

OB Type change

The ultrasound system includes measurements for the following studies: USA, Europe, Tokyo, Osaka and ASUM.

Select OB Type in Utility > System > System Measure.



Figure 8-3 • OB Type selection




Figure 8-4 • OB Type: Tokyo - example

NOTE

ASUM studies include the following measurements:

- ASUM: AC, BPD, and CRL
- ASUM 2001: AC, BPD, CRL, FL, HC, HL, and OFD

To Start an Obstetrics Exam

NOTE

Calculation formulas are listed in the Advanced Reference Manual.

To begin an Obstetrics exam, enter patient data, or, if the patient data from a previous exam is saved in the system, find the patient information.

1. On the control panel, press Patient.
The Patient Data Entry screen is displayed.
2. On the Patient Data Entry screen, select New Patient.
3. To choose an Obstetrics exam, move the **Trackball** to highlight Obstetrics, then press **Set**.
The obstetric fields are listed in the Exam Information section of the Patient Data Entry screen.
4. Do one of the following:
 - If the patient data is already stored in the system, search for the data. Use the search fields in the bottom section of the Patient Data Entry screen. For information about how to search for patient data, see *Changing Patient Information or an Exam* on page 163.
When the correct patient data is listed in the search list, move the **Trackball** to highlight the patient name and press **Set**. The system displays the patient data.

NOTE

To change patient data, use the **Trackball** to move the cursor to the field and press **Set**. Press **Backspace** to delete the data, and then type the correct data.

- If the patient data is not stored in the system, enter the data. To enter data in a field, move the **Trackball** to highlight the field and then press **Set**. Use the **Tab** key to move between fields. Obstetric patient fields are listed in the following table.

NOTE

For information about entering general patient data such as Patient ID and name, see *Beginning an Exam* on page 146.

Table 8-1 • Obstetric Fields

Field	Description
LMP	Last Menstrual Period; enter the date that the patient started her last menstrual period. 4 digits for the year must be entered. When the month and day are typed, the system fills in the /. The Date Format preset chosen in Utility > System > General determines the required format.
BBT	Basal Body Temperature.
EDD by LMP	Estimated Delivery Date by LMP; the system fills in the date after the LMP is entered.
GA by LMP	Gestational Age by LMP; the system fills in the age after the LMP is entered.
Gravida	Number of pregnancies.
Para	Number of births.
AB	Number of abortions.
Ectopic	Number of ectopic pregnancies.
Fetus #	Number of fetuses; default is 1. Can be 1-4.
Accession #	Exam number used with hospital information system (DICOM). This is a tracking number from the worklist.
Exam Description	Describe the type of exam.
Perf Physician	The physician who performs the exam. Choose from the list or type the name.
Ref. Physician	The physician who requested the exam. Choose from the list or type the name.
Operator	The person (not a physician) who performs the scan. Choose from the list.

NOTE

To fill in the following information, move the **Trackball** to highlight the Detail button and press **Set**.

Table 8-2 • Obstetric Fields: Detail

Field	Description
Indications	Why the patient needs the ultrasound exam.

Field	Description
Comments	Comments about the exam.

After the patient information is completed, the scan can begin.

- To change from the Patient Data Entry screen to the Scan screen, do one of the following:
 - On the keyboard, press **Esc**.
 - On the Touch Panel, select **Scan**.
 - On the Control Panel, select **Patient** or **Freeze**.
 - On the Control Panel, press the **B-Mode** key.
The system displays the Scan screen.
- To choose the appropriate probe, select the probe icon on the Touch Panel.
- On the Control Panel, press **Measure**.
The default Obstetrics study is displayed on the Touch Panel.




Figure 8-5 • OB-General Study

To Choose a Study

- To change the study on the exam category, select **Probe**.
Choose the Obstetrics exam category from the following studies:
 - Generic
 - OB-1
 - OB-2/3
 - OB-General
 - Fetal Heart
 - OB/GYN Vessel
- To select a study, select the appropriate study on the Touch Panel.




Figure 8-6 • OB Study Touch Panel

NOTE

The folders on the Touch Panel may be different if the system has been customized.

OB Measurements Performed Over Multiple Planes

Gestational Sac

To calculate the gestational sac, make three distance measurements in two scan planes. To display two scan planes, press the **L** or **R** key. Get an image in each scan plane and press **Freeze**.

1. Select **GS**; an active caliper displays.
 - a. To position the active caliper at the start point, move the **Trackball**.
 - b. To fix the start point, press **Set**.
The system fixes the first caliper and displays a second active caliper.
 - c. To position the second active caliper at the end point, move the **Trackball**.
A dotted line connects the measurement points.
 - d. To complete the measurement, press **Set**.
The system displays the distance value in the Results Window and displays an active caliper.
2. To make the second and third distance measurement, repeat steps a–d.
After completing the third distance measurement, the system displays the gestational sac measurement in the Results Window.

To calculate the gestational sac by a one distance measurement:

1. Select **GS**; an active caliper displays.
 - a. To position the active caliper at the start point, move the **Trackball**.

- b. To fix the start point, press **Set**.
The system fixes the first caliper and displays a second active caliper.
- c. To position the second active caliper at the end point, move the **Trackball**.
A dotted line connects the measurement points.
- d. To complete the measurement, press **Set**.
After completing the measurement, the system displays the gestational sac measurement in the Results Window.

Amniotic Fluid Index (AFI)

To calculate the amniotic fluid index, make measurements of the four quadrants of the uterine cavity. The system adds these four measurements together to calculate the Amniotic Fluid Index.

NOTE

The four quadrants can be measured with distance (caliper) or circumference (circle) measurements. Press the appropriate AFI quadrant Touch Panel key to toggle between caliper and circle.

1. Select **AFI**.
The first distance measurement, AFI-Q1, is already selected.
2. Make a standard distance measurement for the first quadrant:
 1. To position the active caliper at the start point, move the **Trackball**.
 2. To fix the start point, press **Set**.
The system fixes the first caliper and displays a second active caliper.
 3. To position the second active caliper at the end point, move the **Trackball**.
A dotted line connects the measurement points.
 4. To complete the measurement, press **Set**.
The system displays the distance value in the Results Window.
3. When the measurement of the first quadrant is completed, unfreeze and move to the second quadrant.
4. After obtaining the image, press **Freeze** and then **Measure**.
The system prompts to continue with the AFI measurements. Make sure that the next quadrant has been selected.
5. Perform a standard distance measurement for the second, third, and fourth quadrants (see step 2).
When all four quadrants have been measured, the system calculates the AFI total and displays it in the Results Window.



HINTS

- If the image is unfrozen after doing an AFI measurement, the system does not delete the previous measurements. Unfreeze and change scan planes as necessary.
- To specify that an unassigned distance measurement be used for an AFI measurement:
 - Select **AFI**.
 - Press the top **Trackball** key.
 - Move the **Trackball** to highlight the unassigned distance measurement in the Results Window.
 - Select the AFI measurement on the Touch Panel.
- If the fluid in a pocket is zero, set the second caliper on top of the first one to give it a zero value.
- An AFI quadrant that is zero (0) can be measured by pressing **Set** twice.

OB Calculation

SonoNT (Nuchal Translucency)

NOTE

This measurement can be adjusted and customized in the system setup.

To measure the contour detection of the NT border:

1. Select NT. The measurement cursor appears.
2. Select the fetal position ("Face Up" or "Face Down").
3. Position and fix the first point (P1) of the rectangular ROI.
4. Position and enter the second point (P2) of the rectangular ROI. The NT border detection is performed. If a valid result is found, the borders are shown in red and the NT distance is displayed with two crosses.
5. If the measurement is correct according to the guidelines, accept and confirm the result to store in the report. If the system cannot detect a result, a warning message appears.

NOTE

To edit the measurement, move the trackball and/or press Change to readjust the start and end point before accepting the measurement.

NOTE

It is possible to select the calculation method by pressing Method: (i-i: inner-inner or i-m: inner-middle).




Figure 8-7 • SonoNT

- 1 SonoNT Measurement
- 2 Detected NT Border

SonoIT (Intracranial Translucency)

SonoIT (Sonography based Intracranial Translucency) is a system supported measurement for Intracranial Translucency. Starting from the routinely used midsagittal view of the fetal face, obtained for assessment of the Nuchal Translucency and nasal bone, the ultrasound system uses a semi-automated mode to measure the anterior-posterior diameter of the fourth ventricle recognizable as intracranial translucency.

The workflow is identical to SonoNT.

OB Graphs

OB Graphs allow assessment of fetal growth compared to a normal growth curve. When a patient has completed two or more ultrasound exams, the graphs can also be used to look at fetal trending. For multi-gestational patients all fetuses can be plotted and the growth can be compared on the graphs.

The ultrasound system provides the following two basic types of graphs:

- **Fetal Growth Curve graphs** – show one measurement per graph. These graphs show the normal growth curve, positive and negative standard deviations or applicable percentiles, and ultrasound age of the fetus using the current measurement. For multi-gestational pregnancies, all fetuses can be viewed. If previous exam data is available, the graph can show fetal trending.

- **Fetal Growth Bar graph** – shows the ultrasound age and the gestational age based on patient data. Plots all measurements on one graph.

To View OB Graphs

To view OB graphs:

1. Press **Measure**.
2. Select **Graph**.

The system displays the OB Graph keys.



Figure 8-8 • OB Graph Keys on Touch Panel

Fetal Growth Curve Graph




Figure 8-9 • Fetal Growth Curve Graph (Single)

The horizontal axis shows the fetal age in weeks. The system determines this age from the data on the Patient Data Entry screen. The vertical axis shows one of the following:

- For measurements, mm or cm
- For ratios, percent
- For fetal weight, grams

The Fetal Growth Curve Graph shows the following information for the selected measurement:

- The normal growth curve
- The standard deviations or relevant percentiles
- The gestational age of the fetus, using patient data (vertical dotted line)
- Using the current ultrasound measurement data, where the fetus is on the growth curve

To Select the Measurement

To select which measurement you want to display on the Fetal Growth Curve Graph, do one of the following:

- To select a specific measurement:
 1. On the graph display, move the **Trackball** to the Measurement Type field and press **Set**.
The system displays a list of measurements.
 2. Move the **Trackball** to select the desired measurement and press **Set**.
The system displays the Fetal Growth Curve Graph for the selected measurement.
- To scroll through all Fetal Growth Curve Graphs, adjust the **Graph Change** control.

To Select the Age to Use

To plot the fetus age, the system allows you to use the gestational age (GA) from the LMP, or to use the composite ultrasound age (CUA). To select, adjust the **Select GA** control. The information in the left column changes between CUA and GA(EDD), and the data may change.

To View a Single or Four Graphs

A single Fetal Growth Curve Graph can be viewed, or four graphs can be viewed at the same time. To select each view, press **Single** or **Quad** to view 4 graphs at once.




Figure 8-10 • Fetal Growth Curve Graph: Quad View

The measurement values are displayed at the bottom of the graph.

To Change Measurements in Quad View

When four graphs are viewed simultaneously, the four graphs can be selected. To change each graph in quad view:

1. On the graph display, use the **Trackball** to move the cursor to the small box that is upper left of each graph, then press **Set**.

The system displays a list of measurements.

2. Move the **Trackball** to select the desired measurement and press **Set**.

The system displays the Fetal Growth Curve Graph for the selected measurement.

To scroll through all Fetal Growth Curve Graphs, adjust the **Graph Change** control.

The order of a quad graph view can be saved by selecting **Save**.

To Edit Patient Data

When working with graphs, the following patient data can be changed or entered:

- GA(LMP) – This field is computed using the LMP date on the Patient Data Entry screen. To change this field:

NOTE

This field can only be changed on the Fetal Growth Curve Graph in single view.

1. Move the **Trackball** to the field, which is left of the graph. To select the field, press **Set**.
The system displays a window with the GA weeks and days.
2. To select each field, move the **Trackball** to the field and press **Set**.
3. Type the correct weeks or days.
4. Select OK.

The system makes the following changes:

- GA (LMP) is now GA (GA) and shows the age you entered.
- In the Patient Data section, the GA changes.
- In the Patient Data section, The EDD (LMP) changes to EDD(GA) and shows an updated date, using the GA you entered.

The LMP is erased.

- FetusPos – Type information about the fetus position.
- PLAC – Type information about the placenta.

To Return from a Graph to the Scan Display

After viewing graphs, to return to the scan display, do one of the following:

- On the graph display, select Exit.
- On the Touch Panel, select **Graph**.

Fetal Trending

When there is ultrasound data for more than one exam for a patient, the data can be used to look at fetal trending on the Fetal Growth Curve Graphs.

1. Select **Graph Display** and select the desired Fetal Growth Curve Graph.
2. Select **Plot Both**.

The system automatically finds the data from previous ultrasound exams, and displays it on the graph with the present data.




Figure 8-11 • Fetal Trending on Fetal Growth Curve Graph

The legend at the bottom of the graph shows the symbols and colors that indicate Past and Present data.

To Manually Enter Past Exam Data

If you have data from a previous ultrasound exam that you want to use for fetal trending, but it is not in the system, you can manually enter the data.

1. After you have registered the patient for this exam, on the Patient Data Entry screen, in the Exam Information (Obstetrics) section, select Past Exam.

The system displays the Input Past Exam screen. See *To Edit Patient Data* on page 356.

2. Enter the data from previous exams.
3. To enter data on page 2, select Next.
4. After you enter the previous exam data, select Exit to Save.

The system saves the previous exam data. When you view the Fetal Growth Curve Graphs, select **Plot Both** to view fetal trending. The system automatically uses the data you entered.

Exam Date	Wt kg	BP cm	HC cm	AC cm	FI cm
mm/dd/yyyy	History	History	History	History	History

Figure 8-12 • Input Past Exam screen, page 1

Fetal Growth Bar Graph

The fetal growth bar graph shows current exam measurements and the normal growth range based on the gestational age. It shows all measurements on one graph.

To view the Fetal Growth Bar Graph:

1. Press **Measure**.
2. Select **Graph**.
3. Select **Bar**.



Figure 8-13 • Fetal Growth Bar Graph

- The horizontal axis shows the gestational weeks.
- The red vertical line shows the gestational age using the patient data.
- The blue dotted vertical line shows the ultrasound age using the current measurements.
- The yellow x shows the ultrasound age for each measurement.
- The green rectangle shows the normal age range for the measurement.

Fetal trending cannot be performed and multiple gestation data cannot be viewed on the bar graph.

OB-Multigestational

Multiple Fetus

The ultrasound system allows measurement and reporting of the development of multiple fetuses. The system can report a maximum of four fetuses.

To Enter the Number of Fetuses

If more than one fetus is imaged during the exam, enter the number of fetuses in the Patient Data Entry Menu.



Figure 8-14 • Fetus Number

When an OB exam is started, the system automatically fills in the Fetus # field with 1. To change the number:

1. Move the cursor to the fetus number and press **Set** twice.
The number is highlighted.
2. Type the correct number and press **Set**.
The system displays a message to confirm changing the fetus number.
3. Select **Yes**.

To Identify Each Fetus

For measurements, calculations, and worksheet displays, the system labels each fetus A, B, C, or D. Each fetus is identified by a letter and the total number of fetuses. For example, fetus A/3 is fetus A from a total of 3.

When scanning, information can be entered about the fetus position and placenta location. The information can be entered in the Patient Data section of the worksheets and the graphs. Up to 23 characters can be entered in the FetusPos and PLAC fields.



Figure 8-15 • Patient Data section of the OB Worksheet

To Select a Fetus

During measurements and calculations, to change between fetuses, do one of the following:

- Adjust the **Fetus** selection.
- Move the **Trackball** to the Summary Window and select the fetus.

The fetus selection can be changed at any time during the exam.

NOTE

After changing to the next fetus, any measurements made are recorded and reported to that fetus. If any active measurement or calculation is not completed when changing the fetus, the system cancels the measurement or calculation.

To View Multiple Fetuses Data on Graphs

Gestational data for multiple fetuses can be displayed on fetal growth curve graphs. After making measurements for each fetus, select **Graph**.

1. To view the graph for each fetus, do one of the following:
 - Adjust the **Fetus** selection.
 - In the Patient Data section, move the **Trackball** to highlight the FetusNo field. In the list of fetuses, move the **Trackball** to select the desired fetus and press **Set**.
2. To display data for multiple fetuses on the same graph, select **Fetus Compare**.

The legend at the bottom of the graph shows the symbols and colors that represent each fetus.

To Compare Multiple Fetus Data on a Worksheet

Measurements of multiple fetuses can be listed and compared on the worksheet.

Select **Worksheet**, then select **Fetus Compare**.

When selecting **Fetus Compare**, the system lists the measurement results for each fetus on the Worksheet.

To Show Fetal Trending for Multiple Fetuses

When there is data for more than one exam, fetal trending for multiple fetuses can be viewed and compared on one graph.

To view fetal trending for multiple fetuses:

1. Select **Graph**.
2. Select **Fetus Compare**.
3. Select **Plot Both**.

NOTE

Fetal trending for multiple fetuses can only be viewed in single graph display.

The symbol key for fetal trending and multiple fetuses is shown at the bottom of the graph.

OB Table Editor

User programmable OB tables can be added to the system.

OB Table Settings Menu

OB Tables can be added in the Measurement & Analysis menu. To open the menu:

1. On the Touch Panel, select **Utility**, then select **M&A**.
2. Check the Exam Category on the far left of the monitor screen. Make sure that Obstetrics is selected.

If it is not selected, select Obstetrics and continue selecting the folders until the appropriate area is selected as to where this new OB Table will be entered. For example, select Obstetrics, then select OB-2/3. If there are further folders within OB-2/3, select that appropriate folder.

3. On the monitor display, select the OB Table tab.

The system displays the OB Table settings menu.




Figure 8-16 • OB Table settings

4. The OB Table settings menu lists OB Table parameters. Specify the following parameter values as appropriate:
 - **Study:** Shows the study to which this measurement table belongs.
 - **New/Edit:** To create a new OB table, select New Table. To edit an existing user-programmable OB table, select Edit Table.

NOTE

System OB Tables cannot be edited.

- **OB Table Template:** To create a new OB table, select the Template (1 - 7) to use as the basis of the user programmable OB Table. *OB Table Templates* on page 363
To edit an existing user OB table, select the desired OB table to edit.
- **Tool Type:** Select the type of measurement: Distance or Circumference.
- **Measure Name:** Type the name of measurement that will display on the Touch Panel.
- **Author Name:** Type the author's name.
- **Table Type:** If necessary, select the Table Type: Fetal Age or Fetal Growth.
- **Measure Type:** Select a measurement type that can be used to calculate EFW, for example BPD.

NOTE

Measure Type is used only when calculating EFW.

NOTE

The following items are display only: Table Format, Table Unit, SD/GP Range, and Graph Range. The system determines these values automatically, based on the type of OB table being created.

5. After specifying all parameter values, move the **Trackball** to Edit Table and press **Set**.

The system displays the Edit Menu.

NOTE

If any of the OB table parameters are not correct, the Edit Menu is not displayed.

OB Table Templates

Tool Type:

- Distance: 2D Caliper
- Circumference: 2D Ellipse, 2D Trace, 2D Caliper

Table 8-3 • Template 1 (based on Hadlock)

Template 1: SD Range Table				
Fetal Age Table	Table Format	MEAS	MEAN	SD
	Table Unit	mm	Week	Week
	Table Range	1SD		
	Graph Range	1SD		
Measurement Result	Value [cm]			
	GA [#w#d]			
	Min [#w#d]			
	Max [#w#d]			
Fetal Growth Table	Table Format	AGE	MEAN	SD
	Table Unit	Week	mm	Week
	Others are same as above			

Table 8-4 • Template 2 (based on Tokyo)

Template 2: SD Range Table				
Fetal Age Table	Table Format	MEAS	MEAN	SD
	Table Unit	mm	Day	Day
	Table Range	1SD		
	Graph Range	1SD		
Measurement Result	Value [cm]			
	GA [#w#d]			
	SD: day(+/-)			
	EDD (Date)			
	GA-Min [#w#d]			
	GA-Max [#w#d]			
Fetal Growth Table	Table Format	AGE	MEAN	SD
	Table Unit	Day	mm	Day
	Others are same as above			

Table 8-5 • Template 3 (based on Osaka)

Template 3: SD Table				
Fetal Age Table	Table Format	MEAS	MEAN	SD
	Table Unit	mm	Day	mm
	Table Range	1SD		
	Graph Range	1SD		
Measurement Result	Value [cm]			
	GA [#w#d]			
	SD: sd=(mv-pv)/sd			
	EDD (Date)			
	GA-Min [#w#d]			
	GA-Max [#w#d]			
Fetal Growth Table	Table Format	AGE	MEAN	SD
	Table Unit	Day	mm	mm
	Others are same as above			

Table 8-6 • Template 4 (based on several European tables)

Template 4: 5%-95% Table					
Fetal Age Table	Table Format	MEAS	MIN	MEAN	MAX
	Table Unit	mm	WeekDay	WeekDay	WeekDay
	Table Range	5%:95%			
	Graph Range	5%:95%			

Template 4: 5%-95% Table					
Measurement Result	Value [cm]				
	GA [#w#d]				
	GP [%] GP is calculated by Fetal Growth Table. If you did not edit Growth Table, GP is not calculated by the system,				
	EDD (Date)				
	GA-Min [#w#d]				
	GA-Max [#w#d]				
Fetal Growth Table	Table Format	AGE	MIN	MEAN	MAX
	Table Unit	WeekDay	mm	mm	mm
	Table Range	5%:95%			
	Graph Range	5%:95%			

Table 8-7 • Template 5 (based on several European tables)

Template 5: 5% - 95% Table				
Fetal Age Table	Table Format	MEAS	MEAN	SD
	Table Unit	mm	WeekDay	mm
	Table Range	1SD		
	Graph Range	5%:95%		
Measurement Result	Value [cm]			
	GA [#w#d]			
	GP [%] GP is calculated by Fetal Growth Table. If you did not edit Growth Table, GP is not calculated by the system,			
	EDD (Date)			
	GA-Min [#w#d]			
	GA-Max [#w#d]			
Fetal Growth Table	Table Format	AGE	MEAN	MAX
	Table Unit	WeekDay	mm	mm
	Table Range	1SD		
	Graph Range	5%:95%		

Table 8-8 • Template 6 (based on several European tables)

Template 6: 5%-95% Table					
Fetal Age Table	Table Format	MEAS	MIN	MEAN	MAX
	Table Unit	mm	WeekDay	WeekDay	WeekDay
	Table Range	10%:90%			
	Graph Range	10%:90%			

Template 6: 5%-95% Table					
Measurement Result	Value [cm]				
	GA [#w#d]				
	GP [%] GP is calculated by Fetal Growth Table. If you did not edit Growth Table, GP is not calculated by the system,				
	EDD (Date)				
	GA-Min [#w#d]				
	GA-Max [#w#d]				
Fetal Growth Table	Table Format	AGE	MIN	MEAN	MAX
	Table Unit	WeekDay	mm	mm	mm
	Table Range	10%:90%			
	Graph Range	10%:90%			

Table 8-9 • Template 7 (Based on several European tables)

Template 7: 10% - 90% Table				
Fetal Age Table	Table Format	MEAS	MEAN	SD
	Table Unit	mm	WeekDay	mm
	Table Range	1SD		
	Graph Range	10%:90%		
Measurement Result	Value [cm]			
	GA [#w#d]			
	GP [%] GP is calculated by Fetal Growth Table. If you did not edit Growth Table, GP is not calculated by the system,			
	EDD (Date)			
	GA-Min [#w#d]			
	GA-Max [#w#d]			
Fetal Growth Table	Table Format	AGE	MEAN	MAX
	Table Unit	WeekDay	mm	mm
	Table Range	1SD		
	Graph Range	10%:90%		

OB Table Edit Menu

The data entered in the OB Table Edit Menu depends on whether the table type is Fetal Age or Fetal Growth.

Fetal Age Table

When creating or editing a Fetal Age table, the OB Table Edit Menu is as follows:

OB Table Edit Menu: Fetal Age Table

Parameters: Min: 100, Max: 1000, Interval: 100

	MEAS	MEAN	SD
	mm	Week	Week
1	100.00		
2	200.00		
3	300.00		
4	400.00		
5	500.00		
6	600.00		
7	700.00		
8	800.00		
9	900.00		
10	1000.00		

Buttons: Cancel, Exit to Save

Figure 8-17 • OB Table Edit Menu: Fetal Age Table

1. Input value to Min, Max and Interval of the Parameters field.
The system automatically fills in the MEAS column.
2. Input value to the columns of MEAN and SD.

NOTE

To move between the fields in the table, use the up, down, left, and right arrow keys.

NOTE

A minimum of two rows of data must be entered. Lines with blank cells are not saved.

To save the Table Data, move the **Trackball** to Exit to Save and press **Set**. To cancel this table, move the **Trackball** to Cancel and press **Set**.

Fetal Growth Table

When creating or editing a Fetal Growth table, the OB Table Edit Menu is as follows:




Figure 8-18 • OB Table Edit Menu: Fetal Growth Table

1. Input value to the required columns.

NOTE

To move between the fields in the table, use the up, down, left, and right arrow keys.

NOTE

A minimum of two rows of data must be entered. Lines with blank cells are not saved.

2. To save the Table Data, move the **Trackball** to Exit to Save and press **Set**. To cancel this table, move the **Trackball** to Cancel and press **Set**.

After the OB table is completed, it is now available for the selected study. The measurement must be assigned to a Touch Panel to be used. See *Measurement and Calculation Setup* on page 318 for more information.

EFW for OB User Table/Formula Editor

EFW Table Editor

An EFW Formula can be edited with the OB Table Editor.

1. Select Utility > Measure > OB Table.
2. Select the appropriate parameters and press **Edit Table**.
 - New/Edit: Select "New Table"
 - OB Table Template: Select appropriate template.
 - Tool Type: Select "Weight"
 - Measure Name: Enter measurement name.
 - Author Name: Enter author's name.

- Table Type: Select “Fetal Age”




Figure 8-19 • OB Table Tab Screen

3. Edit the table data and press **Exit To Save**.




Figure 8-20 • OB Table Editor Screen

EFW Formula Editor

1. Select the M&A tab and select **Edit Calc**. The Modify User CALC window displays.

Select the user table previously added from the User Defined pull-down menu and press **OK**.

2. Select the “Calculated” button for the EFW parameter.
3. The EDIT FORMULA window displays. Edit the formula and select **OK**.

NOTE

When editing a formula:

- To calculate EFW by centimeter, add “ *100 ” to the {parameter}.
- If EFW is calculated in grams, add “ /1000 ” to the formula.

For example:

$10^{(1.56\{AC[Hadlock]\} * 100 + 0.08 * \{FL[Hadlock]\} * 100) / 1000}$

GYN

The Gynecology exam category includes the following three studies:

- Generic. This study is common to all exam categories. See *Generic Measurements* on page 278 for more information.
- General Gynecology. This study includes uterine, ovarian, ovarian follicle, and endometrium measurements.
- OB/GYN Vessel. This study includes the following vessels: uterine, ovarian, umbilical, middle cerebral artery, aorta, placenta, and descending aorta.

NOTE

The calculation formulas are listed in the Advanced Reference Manual.

B-Mode Measurements

In B-Mode, measurements are made in the General Gynecology study. These measurements include:

- Ovarian follicle
- Endometrium thickness
- Ovarian length, width, and height
- Uterine length, width, and height
- Cervix
- Fibroid

Follicle Measurements

Left and right ovary follicle measurements can be made from one, two, or three distances.

To select the left or right, adjust the **Side** selection.

Endometrium Thickness

To measure the endometrium thickness, make one distance measurement.

1. Select **Endometrium**; an active caliper displays.
2. To position the active caliper at the start point, move the **Trackball**.
3. To fix the start point, press **Set**.
The system fixes the first caliper and displays a second active caliper.
4. To position the second active caliper at the end point, move the **Trackball**.
A dotted line connects the measurement points.
5. To complete the measurement, press **Set**.
The system displays the endometrium thickness in the Results Window.

Ovary Length, Width and Height

Length, width, and height of the left and right ovaries can be measured. Each measurement is a typical distance measurement made in the appropriate scan plane.

Typically, length and height are measured on the sagittal plane while the width is measured on the axial/transverse plane.

To measure ovarian length, width, or height:

1. Scan the patient's right or left ovary in the appropriate plane.
2. To select left or right, adjust the **Side** selection.
3. Select the **OV** folder, then select **OV L**, **OV W**, or **OV H**.
4. Perform a standard distance measurement.

Uterus Length, Width and Height

Each of these is a standard distance measurement. Typically, length and height are measured on the sagittal plane while the width is measured on the axial/transverse plane.

To measure uterus length, width, or height:

1. Scan the patient in the appropriate scan plane.
2. Select the **UT** folder, then select **UT L**, **UT W**, or **UT H**.

An active caliper displays.

3. Perform a standard distance measurement.

Cervix Measurements

Cervix measurements can be made from one distance or spline trace.

Cardiac

Cardiology measurements offer two types of measurement studies: Generic and Cardiac.

- **Generic:** Common to all applications. See *Generic Measurements* on page 278 for more information.
- **Cardiac:** This study includes all cardiac measurements.

Naming Format for Cardiac Measurements

When making a measurement, select the abbreviation for the measurement on the Touch Panel. Most abbreviations are made using acronyms. The following table lists acronyms used for naming cardiac measurements.

Table 8-10 • Cardiology Abbreviations

Acronym	Name
% STIVS	% Interventricular Shortening
A	Area
Acc	Acceleration
AccT	Flow Acceleration Time
ALS	Aortic Leaflet Separation
Ann	Annulus
Ao	Aorta
AR	Aortic Regurgitation
Asc	Ascending
ASD	Atrial Septal Defect
AV	Aortic Valve
AV Cusp	Aortic Valve Cusp Separation
AVA	Aortic Valve Area
AV-A	Aortic Valve Area by Continuity Equation
BSA	Body Surface Area
CI	Cardiac Index
CO	Cardiac Output
d	Diastolic
D	Diameter
Dec	Deceleration
DecT	Deceleration Time
Desc	Descending
Dur	Duration
EdV	End Diastolic Volume
EF	Ejection Fraction
EPSS	E-Point-to-Septum Separation

Acronym	Name
EsV	End Systolic Volume
ET	Ejection Time
FS	Fractional Shortening
FV	Flow Volume
FVI	Flow Velocity Integral
HR	Heart Rate
IVRT	IsoVolumetric Relaxation Time
IVS	Interventricular Septum
L	Length
LA	Left Atrium
LAA	Left Atrium Area
LAD	Left Atrium Diameter
LPA	Left Pulmonary Artery
LV	Left Ventricle
LVA	Left Ventricular Area
LVID	Left Ventricle Internal Diameter
LVL	Left Ventricle Length
LVM	Left Ventricular Mass
LVPW	Left Ventricle Posterior Wall
ML	Medial to Lateral
MPA	Main Pulmonary Artery
MR	Mitral Regurgitation
MV	Mitral Valve
MVcf	Mean Velocity Circumferential Fiber Shortening
MVO	Mitral Valve Orifice
OT	Outflow Tract
P	Papillary Muscles
PA	Pulmonary Artery
PAP	Pulmonary Artery Pressure
PDA	Patent Ductus Arteriosis
PEP	Pre-Ejection Period
PFO	Patent Foramen Ovale
PG	Pressure Gradient
PHT	Pressure Half Time
PI	Pulmonary Insufficiency
PISA	Proximal Isovelocity Surface Area
PR	Pulmonic Regurgitation
PV	Pulmonic Valve
PV-A	Pulmonic Valve Area by Continuity Equation
PVein	Pulmonary Vein

Acronym	Name
PW	Posterior Wall
Qp	Pulmonic Flow or CO
Qs	Systemic Flow or CO
RA	Right Atrium
RAA	Right Atrium Area
Rad	Radius
RAD	Right Atrium Diameter
RPA	Right Pulmonary Artery
RV	Right Ventricle
RVA	Right Ventricle Area
RVAW	Right Ventricle Anterior Wall
RVD	Right Ventricle Diameter
RVID	Right Ventricle Internal Diameter
RVL	Right Ventricle Length
RVOT	Right Ventricle Outflow Tract
s	Systolic
SI	Stroke Index
ST	Shortening
SV	Stroke Volume
SVI	Stroke Volume Index
T	Time
TA	Tricuspid Annulus
TAML	Tricuspid Annulus Medial to Lateral
TR	Tricuspid Regurgitation
TV	Tricuspid Valve
TVA	Tricuspid Valve Area
Vcf	Velocity Circumferential Fiber Shortening
Vel	Velocity
VET	Valve Ejection Time
Vmax	Maximum Velocity
Vmean	Mean Velocity
VSD	Ventricular Septal Defect
VTI	Velocity Time Integral

In this manual, the abbreviation for each measurement is listed in parenthesis after the measurement, as follows:

- Aortic Root Diameter (**Ao Diam**)
- Left Ventricle Posterior Wall Thickness, Diastolic (**LVPWd**)

For example, to measure the Aortic Root Diameter, select **Ao Diam** on the Touch Panel.

Cardiac Doppler Measurements

E/E' Ratio

The ratio of early transmitral velocity to early diastolic velocity of the mitral annulus (E/E') is measured in Doppler Mode and TVD mode.

1. Measure MV E/A Velocity to get "E".
2. Measure E'.

The system calculates the E/E' ratio automatically.

Vascular

Vascular measurements offer several different types of measurement studies:



Figure 8-21 • Vascular Exam Category Touch Panel

- Generic – Common to all applications. See *Generic Measurements* on page 278 for more information.
- Carotid
- LEA (Lower Extremity Artery)
- LEV (Lower Extremity Vein)
- TCD (Trans Cranial Doppler)
- UEA (Upper Extremity Artery)
- UEV (Upper Extremity Vein)
- Renal
- Bypass Graft
- UEV (Upper Extremity Vein) Map
- LEV (Lower Extremity Vein) Map

A vascular study is a group of particular vessels. Vessel exam calcs can be customized in the configuration menu. See *Measurement and Calculation Setup* on page 318 for more information.

When using Auto Vascular calculation, use the vessel keys on the Touch Panel to post-assign vascular calculations. When not using Auto Vascular calculation, the vessel key is used for manual measurement.

IMT Measurement

The average of the intima media thickness can be measured for use as the index of arterial sclerosis.

IMT can be measured both on the posterior and the anterior walls of the vessel.

NOTE

Due to the physical properties of ultrasound imaging, the posterior IMT measurement is generally more accurate than the anterior IMT measurement.

IMT Measurement - Auto

Auto IMT automatically measures the thickness of the Intima Media on the far and near vessel walls. Near Wall IMT is the distance between the trailing edges of the adventitia and intima; the Far Wall IMT is the distance between the leading edges of the adventitia and intima.

Set up the parameters you want to record on the worksheet on the Utility > Measure > M&A page while are in the Carotid application. Select CCA/ICA/BIF > IMT Far/Near > Parameter (Average, Max, Min, Standard Deviation, Points, or Distance).




Figure 8-22 • Configuring Auto IMT

In the Vascular Carotid application, the Auto IMT measurement is available.




Figure 8-23 • Auto IMT Touch Panel

The following controls are available.

Table 8-11 • Auto IMT Touch Panel Description

Parameter	Description
Worksheet	Select to view the Worksheet
IMT Far	Select to begin the Far Field IMT measurement.
IMT Near	Select to begin the Near Field IMT measurement.
AP	Anterior Posterior
Trans	Transverse
Length/Offset Rotary	Push to save Length/Offset as a preset. -40/+40 Length. At zero, you can freely adjust the length, but only vertically. Press key to save value as default. Offset distance, -20 (Left) / +20 (Right)
Overall / IMT Trace Fit / Intima	Adjusts (remeasures) the IMT automatically measured by the system.
Rt / Lt Side	Select Left / Right Side.
Cursor Select	Allows you to update cursor placement.

To measure the IMT:

1. In the Carotid application, press **Freeze**, press **Measure**.
2. Position the cursor, then select **IMT Far**.
3. Use the **Trackball** to set the length.

Or

Use the **Length / Offset** control on the Touch Panel to set the length and offset distance. The Offset key controls how far away from the vertical line the measurement starts. Length is the length of the tool itself. If set to zero, you can adjust it anywhere on the image.

4. Press **Set**.

Either adjust the trace prior to pressing the Print key, or press the Print key to store the image which also saves the measurement to the Worksheet.

To adjust the trace, use the **Overall IMT Trace Fit Intima** control on the Touch Panel. The Trace fit (up/down) adjusts the inter luminal line whereas the overall (rotate) adjusts both IMT lines.



Figure 8-24 • Example of Auto IMT Far Measurement

5. Position the cursor, then select IMT Near.
6. Use the **Trackball** to set the length.
Or
Use the **Length / Offset** control on the Touch Panel to set the length and offset distance.
7. Press **Set**. “Store image to accept IMT measurement” displays in the message area. If the traces fit both layers of the wall, approve the measurement by pressing the **Print** key to store the image.

To adjust the trace prior to pressing the Print key, use the IMT Trace Fit control on the Touch Panel. The measurement is saved to the Worksheet.

NOTE

Since the IMT measurements are semi-automatic, the operator has to approve the detection by visual inspection before storing the results in the worksheet and report.

IMT Measurement - Manual

1. Before measuring the IMT, add the IMT measurement to the Carotid folder via the Measurement & Analysis screen (by selecting one of the three types of IMT measurements under Add Measurement in the M&A screen).
 - IMT: Three vertical lines are parallel. Place the start point on the line and place the end point anywhere.
 - IMT2: Each vertical line can be rotated with the **Ellipse** control. The start and end points must be placed on the line.
 - 5mm Scale: The horizontal line can be rotated with the **Ellipse** control. A maximum of 20 distance values which produce one average value can be taken. The number of distance values is specified when adding the measurement in the M&A screen.

IMT and IMT2 have three kinds of measurements:

 - IMT --+/IMT2 --+: Measure from right to left.
 - IMT -+-/IMT2 -+ -: First measure at the center, then right and left are last.

- **IMT+--/IMT2 --+:** Measure from left to right.
2. Acquire a longitudinal scan of the carotid artery and optimize the image. Then press **Freeze**.
 3. Scroll to an end-diastolic frame where the intima layer is clearly visible.
 4. Press **Measure**, then select **IMT1**, **IMT2** or **5mm scale**. An active caliper displays.




Figure 8-25 • IMT caliper (Example)

5. Use the **Trackball** to move the caliper and the **Ellipse** control to adjust the angle. Press **Set** to fix the caliper.

NOTE

The interval of the vertical line for IMT1 and IMT2 is 1cm and for the 5mm scale is 5mm.

6. Measure the thickness of three points for IMT1 and IMT2.

OR

Measure the thickness of the specified number of points for the 5mm scale.

NOTE

The caliper moves to the next point automatically.

7. After completing the measurement, the system calculates the average automatically.



Figure 8-26 • IMT Measurement

IMT Measurement - Manual C(10)

1. Before you measure the IMT C10, create the IMT C10 measurement in the Carotid folder via the Measurement & Analysis screen.
 - IMT C10: Two vertical lines are parallel. One is the base line. Each vertical line can be rotated with the Ellipse control. Place the start point on the line and place the end point anywhere.
 - IMT2 C10: Two vertical lines are parallel. One is the base line. Each vertical line can be rotated with the Ellipse control. You must place the start and end points on the line.

IMT C10 and IMT2 C10 have two kinds of measurements:

- IMT C10 -+/IMT2 C10 -+: Base vertical line shows right side.
 - IMT C10 +-/IMT2 C10 +/-: Base vertical line shows left side.
2. Acquire a longitudinal scan of the carotid artery and optimize the image. Then press **Freeze**.
 3. Scroll to an end-diastolic frame where the intima layer is clearly visible.
 4. Press **Measure**, then select **IMT C10**. An active caliper displays.




Figure 8-27 • IMT C10 Active Caliper

5. Use the trackball to move the caliper and the Ellipse control to adjust the angle. Press **Set** to fix the caliper.

NOTE

The interval of the vertical line for IMT1 and IMT2 is 1cm.

6. Measure the thickness for IMT C10.
7. After you complete the measurement, the system shows base vertical line.

NOTE

The base vertical line length is 1cm.




Figure 8-28 • Base Vertical Line

Plaque Score Tool

1. Before measuring, add Plaque Score to the Carotid folder at the Utility > Measure > Measurement & Analysis screen (by selecting 2D Plaque Score under Add Measurement in the M&A screen).
Label the parameters you want measured:
 - Distance value (up to 20)
 - Sum (greater or equal to 1.1 mm)
 - Count (greater or equal to 1.1 mm)
 - Maximum value of each area
 - Average value
 - Average value of each area
2. Scan the carotid artery and press **Freeze**. Display dual images to measure across split screen images.
3. Press **Measure** and select **Plaque Score Tool**. An active caliper, one horizontal line and five vertical lines display. The interval of the vertical lines are 1.5 cm.




Figure 8-29 • Plaque score example

4. Use the **Trackball** to move the caliper and the **Ellipse** control to adjust the angle. Press **Set** to fix the caliper.
5. Measure the thickness (up to twenty times) on any place as necessary.




Figure 8-30 • Plaque Score Example 2

6. The system displays the measurement result.




Figure 8-31 • Plaque Score Example 3

NOTE

Only calculates the value larger than 1.1mm.

7. Double click the **Set** key to finish the measurement.

Auto Vascular Calculation

Auto Vascular Calculation enables the ultrasound system to detect and identify a cardiac cycle. It allows you to assign measurements and calculations during live timeline imaging, while the image is frozen, or in CINE. Peak values are detected for venous flow.

During cardiac cycle detection, the system identifies the cardiac cycle using calipers, vertical bars, and/or highlighting of timeline data. Use of identifiers is based on measurements and calculations selected by an operator for the current application. The system may place calipers at early systolic peak, peak systole, minimum diastole and end diastole. Vertical bars may also be placed to indicate the beginning and end of the cardiac cycle. The peak and/or mean trace may be highlighted. You can edit the cardiac cycle identified by the system or select a different cardiac cycle.

Calculations to be displayed in the M&A Result window can be selected during live scanning or on a frozen image. These calculations are displayed at the top of M&A Result Window located adjacent to the image. These calculations are presettable by application, which means they can be set up the default calculations to be displayed for each application.

Activating Auto Vascular Calculation

To activate Auto Vascular Calculation, select **Auto Calc** from Live (calculations displayed on the real-time image), or Freeze (calculations displayed on the frozen image).

To deactivate Auto Vascular calculation, select Off.




Figure 8-32 • Auto Calculation Touch Panel key

Set up Auto Vascular Calculation Parameters

- **Select Auto Trace**
To set continuous auto trace of the max or mean velocities:
 - Select Max or Mean using the **Trace Method** Touch Panel pull-down menu.
- **Select Trace Direction**
Trace Direction allows peak timeline data above, below, or composite (above and below) the baseline.
 - Select Positive, Negative or Both to set the peak timeline data.
- **Modify Calculation**
 1. Select the **Modify Calcs** Touch Panel key.
The Modify Calculation menu is displayed.
 2. Select which measurements and calculations are to be displayed in the Auto Vascular calculation window.

Select from the following parameters: PS, ED, MD, HR, TAMAX, PI, RI, Accel, PS/ED, ED/PS, AT, Volume Flow, PV.

Auto Vascular Calculation Exam

1. Preset the system.
2. Perform the scan and press **Freeze**.
3. Activate Auto Vascular Calculation.

The system performs a calculation automatically.




Figure 8-33 • Auto Vascular Calculation

The Auto Vascular calculation is assigned to particular vessel measurements.

1. Press Measure to display the Measurement menu.
2. Select the location of the vessel (Prox, Mid, or Dist) and Side (Right or Left).
3. Select the desired vessel name from the Touch Panel.

Selected vessel measurements are automatically assigned with the Auto Vascular calculation. The results are then displayed in the Results Window.



Figure 8-34 • Assigned Vessel

NOTE

To cancel the assignment, use the **Cancel Transfer** Touch Panel key. See *Cancel Transfer* on page 287 for more information..

During the course of an exam, the cardiac cycle may be indicated between two yellow bars; the peak trace and the mean trace may appear in green; calculation indicators appear on the spectral trace as a caliper identifier (these vary, depending on the selected calculation in the Results Window).

The right-most, most complete cycle is typically chosen to be the selected cardiac cycle. To select a different cardiac cycle:

- Move through CINE memory with the Trackball until the desired cardiac cycle is selected by the system.

NOTE

Several good cycles are required in front of the new cardiac cycle for this to be successful. Oftentimes, this is problematic near a freeze bar.

- Use the **Cycle Select** control to cycle to a different cardiac cycle.

To move the systole or diastole position:

- Use the **Cursor Select** control to move the start systole position or the end diastole position.

Manual Vascular Calculation

Calculations can be performed manually when Auto Doppler Calculation is not activated.

1. Press **Measure**.

If necessary, you can select another Exam Calc and then select parameters from Modify Calculation.

2. Select the location of the vessel (Prox, Mid, or Dist) and Side (Right or Left).
3. Select the desired vessel folder.

The Measurement menu is displayed.




Figure 8-35 • Measurement Menu Example

4. Make the required measurements according to the system, or select your preferred measurements.

To Select Vascular Measurements

The system is set up to show the measurements usually made for each vessel. To make a measurement that is not shown for the selected vessel:

1. Select the folder for the vessel you want to measure.
2. Select **Show All**.

The system displays all possible vessel measurements.

3. Select the desired measurement.

NOTE

Prior to performing vascular measurements, scan the patient and then press **Freeze**.

Intravessel Ratio

Information on Intravessel Ratio calculations is available in *Intravessel Ratio* on page 311.

Bypass Graft Anastomosis Graph

If there aren't any known Grafts associated with the current exam, Grafts can be created by describing the Graft anastomosis locations and anastomosis modifiers.

NOTE

New Grafts can be created by selecting one from the "Choose to Add" table, and then modifying.

To graph Bypass Graft anastomosis:

1. Select Bypass Graft on the Touch Panel. The Add/Edit Grafts pop-up appears.

NOTE

This pop-up appears if there are not any known Grafts.



Figure 8-36 • Add/Edit Graft

2. Use the Trackball to assign graft locations on the Add/Edit Grafts pop-up. Press Continue.

NOTE

Grafts can be selected by choosing it in the table of Grafts that exist in the current exam.

NOTE

Grafts can be modified by choosing new anastomosis locations and anastomosis modifiers.

NOTE

The Stent pop-up is nearly identical to the Graft pop-up, with the option to specify the Stent's laterality.

- Graft assignments now appear on the Touch Panel: This is saved on the system and can be edited. Specify the Topographic Modifier (location), Stenosis Modifier, and Anatomic Modifier for any measurement you would like to take.




Figure 8-37 • Graft table update

NOTE

For each measurement within a blood vessel, you can choose a Topographic Modifier (i.e., location) and a Stenosis Modifier. For Grafts and Stents, you can also choose an Anatomic Modifier (i.e., eye, ankle, or liver).

- To edit the Bypass Graft Anastomosis, select the “Edit 1/1” toggle on the Touch Panel to specify which of the Grafts you are measuring within.



HINTS

When creating Grafts or Stents, it is always recommended to perform the accompanying measurements. The measurements folder can be found in the lower, right-hand portion of the Touch Panel.

- On the Doppler Mode M&A Touch Panel, you have the option to select the Stenosis Modifiers, Pre-Steno, At-Steno, or Post-Steno.



Figure 8-38 • Select the Stenosis Modifier

Urology

Urology measurements offer three different types of measurement studies:

- Generic—Common to all applications. See *Generic Measurements* on page 278 for more information.
- Urology
- Pelvic Floor. See *Pelvic Floor Measurements* on page 395 for more information.

NOTE

Bladder(0.7) Vol, Bladder Vol, Post Void Vol, Prostate Vol, Renal Vol, Renal (0.8) Vol and Volume can be displayed on the Touch Panel if preset at the Utility > Measure screen.

Bladder Volume

This calculation uses a standard distance measurement. Length is typically measured in the sagittal plane. Width and height are measured in the axial plane.




Figure 8-39 • Bladder Volume Touch Panel

Renal Volume

This calculation uses a standard distance measurement. Length is typically measured in the sagittal plane. Width and height are measured in the axial plane.

To select the left or right, adjust the **Side** selection.

To measure Renal Volume:

Prostate Volume

This calculation uses a standard distance measurement. Length is typically measured in the sagittal plane. Width and height are measured in the axial plane.




Figure 8-40 • Prostate Volume Touch Panel

PSA Measurement

Entering the value of PSA (Prostatic Specific Antigen) and PPSA Coefficient at the Urology Patient screen, automatically calculates PSAD and PPSA.

The values are displayed on the Worksheet and Report (if set appropriately on the Report Designer page).



Figure 8-41 • Urology Patient Screen




Figure 8-42 • Measurement result window

PSAD: Prostatic Specific Antigen (PSA) Density – defined as: $PSAD = PSA/Volume$

PPSA: Predicted Prostate Specific Antigen – defined as: $PPSA = Volume \times PPSA \text{ Coefficient}$

Pelvic Floor Measurements

Pelvic floor measurements can be performed in the Pelvic Floor study. The measurements are located in the Exam Calc folder in the Urology preset.




Figure 8-43 • Pelvic Floor Study Touch Panel

BN (Bladder Neck) Rest

Obtain an image with the patient at rest (relaxed).

1. Create a straight line (zero or baseline) to line up with the inferior/posterior of symphysis pubis bone.
2. Once the baseline is positioned, a caliper appears. Position the caliper at the anterior margin of the bladder neck. A positive number displays since the caliper is placed below the baseline.
3. A distance is calculated in millimeters.

BN (Bladder Neck) Stress

Obtain an image after the patient performs the Valsalva maneuver.

1. Create a straight line (zero or baseline) to line up with the inferior/posterior of symphysis pubis bone.
2. Once the baseline is positioned, a caliper appears. Position the caliper at the anterior margin of the bladder neck.

If the bladder neck is below the baseline, the Bladder Neck Stress is a positive number. If the bladder neck is above the baseline (closer to the transducer face), the number is negative.

BN (Bladder Neck) Descent

The Bladder Neck Descent is a calculation that should be calculated after measuring the Bladder Neck Rest and Bladder Neck Stress.

$BND = \text{Bladder Neck Rest} - \text{Bladder Neck Stress}$

NOTE

If the Bladder Neck Stress is a negative number, it becomes positive and is added to the bladder neck rest measurement.

DWT (Detrusor Wall Thickening)

Three distance measurements of the bladder wall dome are calculated into a mean dimension and displayed in millimeters.

UT (Uterine) Descent Max

1. Create a straight line (zero or baseline) to line up with the inferior/posterior margin of symphysis pubis bone.
2. Measure using a 2-caliper dimension to the inferior position of the uterus in a stress image and display in millimeters

Rect Amp Des Max (Rectal Ampulla Descent Max)

1. Create a straight line (zero or baseline) to line up with the inferior/posterior margin of symphysis pubis bone.
2. Measure using a 2-caliper dimension to the inferior position of the rectal ampulla in a stress image and displayed in millimeters

Rectocele (Depth and Width)

Two 2-caliper diameter measurements to measure depth and width of the rectocele. Displayed in millimeters.

Lev Hiat Stress (Levator Hiatus Stress)

Two 2-caliper diameter measurements and calculate an area displayed as cm squared.

Residual Urine

Two 2-caliper diameter measurements calculate as:

$(x) \times (y) \times 5.9 \text{ minus } 14.9 \text{ equals Residual Volume displayed in ml.}$

Pediatrics

Pediatrics measurements offer two different types of measurement studies:

- Generic. The Generic Calculations study is common to all applications. See *Generic Measurements* on page 278 for more information.
- Pediatric Hip (PedHip).




Figure 8-44 • Touch Panel - Pediatrics Exam Calcs

Pediatrics Hip




Figure 8-45 • B-Mode Measurement Touch Panel – PedHip

Hip Dysplasia Measurement


The HIP calculation assists in assessing the development of the infant hip. In this calculation, three straight lines are superimposed on the image and aligned with the anatomical features. The two angles are computed, displayed, and can be used by the physician in making a diagnosis.

The three lines¹ are:

- 1 - The baseline connects the osseous acetabulum convexity to the point where the joint capsule and the perichondrium unite with the iliac bone.
- 2 - The inclination line connects the osseous convexity to labrum acetabulare.
- 3 - The Acetabulum roof line connects the lower edge of the osilium to the osseous convexity.

The α (Alpha) angle is the supplement of the angle between 1 and 3. It characterizes the osseous convexity. The β (Beta) angle is the angle between lines 1 and 2. It characterizes the bone supplementing additional roofing by the cartilaginous convexity.

Figure 8-46 • Hip Dysplasia



Anatomical Landmarks

- a Ilium
- b Iliac Bone
- c Labrum
- d Bony Roof
- e Cartilaginous acetabular roof
- f Femoral Head

¹Source: R GRAF, Journal of Pediatric Orthopedics, 4: 735-740(1984)

To make a Hip Dysplasia measurement:

1. From the Touch Panel, select either the **right** or **left side** (orientation) and then select **Beta Alpha HIP**.
A horizontal dotted line displays.
2. To place the baseline, move the **Trackball**. Position the crosshairs edge at the osseous convexity of the ilium.
3. To rotate or change inclination, adjust the **Ellipse** control or **Hip Rotate**.
4. To fix the baseline, press **Set**.
The system displays a second dotted line at an angle.
5. To place the line along the inclination line of the osseous convexity to labrum acetabulare, move the **Trackball**.
6. To rotate or change inclination, adjust the **Ellipse** control or **Hip Rotate**.
7. To fix the second measurement line, press **Set**.
The system displays a third dotted line at an angle.
8. To place the caliper along the acetabular roof line, move the **Trackball**.

9. To rotate or change inclination, adjust the **Ellipse** control or **Hip Rotate**.
10. To fix the third measurement line and complete measurement, press **Set**.

The system displays the hip measurements (α and β) in the Results Window.

Alpha HIP

The Alpha HIP measurement measures the angle between the iliac baseline and the bony roof line. To make an Alpha HIP measurement:

1. From the Touch Panel, select either the **right** or **left side** (orientation) and then select **Alpha HIP**.
A horizontal dotted line displays.
2. To place the baseline, move the **Trackball**. Position the crosshairs edge at the osseous convexity of the ilium.
3. To rotate or change inclination, adjust the **Ellipse** control or **Hip Rotate**.
4. To fix the baseline, press **Set**.
The system displays a second dotted line at an angle.
5. To place the caliper along the acetabular roof line, move the **Trackball**.
6. To rotate or change inclination, adjust the **Ellipse** control or **Hip Rotate**.
7. To fix the second measurement line, press **Set**.

The system displays the alpha hip measurement (α) in the Results Window.

d:D Ratio Measurement

The d:D Ratio measurement measures the percentage of the femoral head coverage under the bony roof. To make this measurement:

1. From the Touch Panel, select either the **right** or **left side** (orientation) and then select **d:D Ratio**.
A horizontal dotted line displays.
2. Use the **Trackball** to place the baseline along the ilium. Position the crosshairs edge at the osseous convexity of the ilium.
3. Use the **Ellipse** control to adjust or change inclination or **Hip Rotate**.
4. Press **Set** to fix the baseline.
5. The system displays a circle representing the femoral head. Use the **Trackball** to position the circle.
6. Use the **Ellipse** control to size the femoral head circumference.
7. Press **Set** to fix the femoral head circumference.

The system displays the d:D ratio for the femoral head in the Results Window.

Chapter 9

Recording Images

In this section

Set Up to Record Images.	402
Storing Images and Cineloops.	405
Review Images in Archive.	409
Clipboard.	418
SaveAs.	420
EZBackup.	425
Data Transfer.	430
External Drives.	438
Printing Options.	444
Portable Exam.	452

Set Up to Record Images

A typical workflow for connectivity might be as follows (this setup varies by each user setup):

1. Select the dataflow, worklist for example.
2. Start a new exam. Select the patient.
3. Perform the patient scan.
4. Store images as multi-frame CINE Loops and Raw DICOM data via the **P1** key.
5. Store secondary capture for DICOM print via **P2** key.
6. Store images to the color printer or B/W printer via **P3** key.
7. Check the DICOM Job Spooler via F4 to verify delivery.
8. End the exam.
9. Permanently store images via the Patient menu to permanent storage.

During an examination, the operator stores data, images and cineloops for immediate purposes. The ultrasound system includes an integrated patient archiving system for data and image storage.

The ultrasound system enables also storing of data and images to external databases (removable media).

Dataflow combines archive, data, DICOM, and onboard records into one coherent workflow. Destination devices are configured and assigned to the print keys. You select the appropriate dataflow (Portable, etc.) according to your requirements. You manage the patient database (local, shared, or via a worklist broker).

- DO NOT use the internal hard drive for long-term image storage. Daily backup is recommended. External storage media is recommended for image archive.

NOTE

DICOM images are stored to external media storage devices separately from patient data, which also needs to be backed up to a dedicated database-formatted external storage media.

- If working off-line with a dataflow pointing to a DICOM server, the images stored during the examination may have to be manually resent in the DICOM spooler when reconnecting the unit. Resend all jobs that failed or are on hold.

In addition, stored images and cineloops can be saved to a removable media in the standard formats JPEG, MPEG, WMV, and DICOM.

- You need to set up a process for locating images stored to external storage media for easy recall
- GE HealthCare **IS NOT** responsible for lost data if you do not follow suggested back-up procedures. GE HealthCare **WILL NOT** aid in the recovery of lost data.

Refer to the Customizing Your System chapter for instructions on setting up your system's connectivity.

Image Management Guide

Save As to View on any PC: Use to save images in a computer-friendly format to view on any PC.

EZBackup Images to Archive: Use to save images from ultrasound system to removable media for long-term archive. (This frees hard disk space, rather than deleting images.)

Export/Import Data/Images Between Systems: Use to copy both patient data and images for specified patient(s) from one system to another.

Media Requirements

The system ONLY supports medical grade CD-R, DVD-R, USB Hard Disk Drive, and USB Flash Drive.

The system DOES NOT support CD-RW / DVD+R / DVD-RW / DVD+RW / or DVD-RAM.



CAUTION

DO NOT use or attempt to format CD-RW, DVD+R, DVD-RW, DVD+RW or DVD-RAM media.

Media Handling Tips

Always finalize the media to allow it to be read them on any other PC. DO NOT finalize the media more images are to be added.

To eject the media:

1. Press **F3**. The Eject device menu is displayed.

NOTE

Always press **F3**. **DO NOT** press the eject button on the drive.

2. Select the relevant media.
3. Select *USB Drive* from the pull-down menu to disconnect the USB drive.

When the *Success* dialog appears, remove the USB drive from the USB port.

NOTE

If the unsuccessful dialogue is displayed, retry after a while.

NOTE

Verify is NOT available on Flash Drive or Hard Disk Drive media.

Adding Devices

To add a destination device (printer, worklist server, etc.) to this system, or to verify a DICOM device, see *Add Dicom Device* on page 513.

Adding a Dataflow

To add a new dataflow to this system, see *Dataflow Page* on page 545.

Adding Devices to a Print Button

To add devices/dataflows to a print button, see *Print Button Page* on page 542.

Formatting Removable Media

To format removable media, see *Removable Media* on page 508.

Storing Images and Cineloops

Images and cineloops that are stored during a current examination are displayed as thumbnails on the clipboard.

When an image is stored, all the additional information that is displayed is saved with it (i.e. probe and application selected, image setting, annotations or measurements).

See Dataflow in Connectivity for detailed settings related storing image/Cine.

Image archive is set by the dataflow selected (See Dataflow in Connectivity for more information.)

To print/store an image, **P1** is most commonly used for the primary destination and internal hard drive.

Storing an Image

To store an image:

1. While scanning, press **Freeze**.
2. Scroll through the cine Loop and select the desired image.
3. Press the appropriate Print key.

The selected image is stored (per your preset instructions) and a thumbnail is displayed on the clipboard.

NOTE

The ultrasound system numbers the images which are saved in the Local Archive (Instance Number). But Instance Number may change or get duplicated when adding/deleting images to the exam. So for identification, the recommendation is to use Content Date/Content Time on the DICOM server instead of Instance Number.

Storing a CINE Loop

A CINE loop is a sequence of images recorded over a certain time frame. The stored CINE Loops are displayed chronologically on the clipboard.

CINE loops can be stored at any time during scanning. You can choose to preview the cine loop before storage and save the cine loop directly, as described below.

The system can be configured to perform either

- Prospective clip: The system begins storing CINE from when you press the Print button, based on the Time Span setting.
- Retrospective clip: The system stores CINE predetermined time before you press the Print button, based on the Time Span setting.

Refer to *Retrospective CINE/Prospective CINE* on page 407 about the setting.

NOTE

The ultrasound system numbers the images which are saved in the Local Archive (Instance Number). But Instance Number may change or get duplicated when adding/deleting images to the exam. So for identification, the recommendation is to use Content Date/Content Time on the DICOM server instead of Instance Number.

Previewing and Storing a CINE Loop

1. While scanning, press **Freeze**.
2. Move the Trackball to activate Cine.
3. Use the trackball or **Frame by Frame** to scroll through the acquisition and find the sequence of interest.
4. Press **Start Frame** or **End Frame** to set the corresponding cineloop boundary to the current frame as necessary.

Rotate **Start Frame** and **End Frame** to trim or expand the cineloop boundaries.
5. Press **Run/Stop** to run the cineloop and then press the print key to store the cineloop.

Cine loops stored on the clipboard are indicated with a movie strip icon.
6. Press **Run/Stop** again to stop the cine loop.
7. Press **Freeze** to return to live scanning.

Depending on whether the system has been configured to enable or disable "Preview Loop before store" (see *Print Controls* on page 489), the following procedures enable the CINE loop to be stored directly.

Storing a CINE Loop Without Preview

If "Preview Loop before store" is disabled:

1. While scanning, press the appropriate print key.
2. The last valid CINE loop is stored in the archive and a movie clip thumbnail is displayed on the clipboard.
3. Scanning resumes immediately.

Storing a CINE Loop With Preview

If "Preview Loop before store" is enabled:

1. While scanning, press the appropriate print key.
2. The last valid cine loop is previewed.
3. Adjust the cine loop, as necessary.
4. Press the appropriate print key.

The movie clip thumbnail is displayed on the clipboard.

Preview

Loop Preview can now be enabled independently for Time-Based Store, ECG-Based Store, and Mark CINE. This is useful for setting preview preferences based on the application.

NOTE

The Contrast Time Span setting overrides the Time Span when in Contrast Mode.

Retrospective CINE/Prospective CINE



Figure 9-1 • Print Control

Retrospective CINE

When selecting to store a CINE loop retrospectively, the system stores CINE for a specified time before the Print Button is pressed.

1. Set Live Store P1 - P4 to "Retrospective Clip" via Utility > Application > Print controls.
2. Specify the recording time in Time-Based Store or ECG-Based Store.
3. Press **Save**.

NOTE

A Print button can be configured to store a Single Image during Retrospective CINE, without stopping the CINE loop.

Prospective CINE

When selecting to store a CINE loop prospectively, the system begins storing CINE from the moment the Print button is pressed.

1. Set Live Store P1 - P4 to "Prospective Clip."
2. Specify the recording time in Time-Based Store or ECG-Based Store.
3. Press **Save**.

NOTE

A Print button can be configured to store a Single Image during Prospective Cine, without stopping the Cine loop.

NOTE

The CINE gauge turns green when a Prospective CINE Clip is pending.

NOTE

Prospective Store can be cancelled by pressing Freeze/Unfreeze or by changing Modes.

Review Images in Archive

There are two ways to access archived images:

- Review the images from a selected examination.
- Select images from the Active Image screen displaying all the images sorted by examination for the actual patient record.

Review the Patient Exam/Image

To review the patient exam:

1. Move the cursor to the patient in the Patient View and double-click. The Exam View displays.
or
Move the cursor to the patient and select Exam View tab or **Review**. The Exam View Displays.
2. Move the cursor to the desired exam and double-click. The Active Images screen displays.
3. Move the cursor to the image and double click or press **Review**. The review screen displays.
4. Select the image from clipboard.

NOTE

See *Clipboard* on page 418 for more information.

Active Images

Active Images displays the images in the exam.

NOTE

CINE loops are not played interactively as you view the active images on the Patient screen.



Figure 9-2 • Active Images Screen

1. Select the exam including the image to review.
2. Select **Active Images**.
3. Select the image and press **Review** or double click on the image. The image is displayed.

If 2 - 4 images are selected and **Review** is selected, the archived images are displayed in the split screen.

NOTE

If the size of an image is larger than 2GB, the image does not display on the Active Image screen. Take care when you scan a long CINE Loop, such as in Contrast.

Table 9-1 • Active Images

Parameters	Description
Delete Images Delete Selected Images/ Delete All Temp. Images	To delete selected images, select the image in the active screen, then select "Delete" on the monitor display or "Delete Selected Images" on the Touch Panel. To delete all images, select the image, then select "Delete All Temp. Images" on the Touch Panel.
Permanent Store	Select the images which to save to the Local hard disk drive.
Standard Print	To print an image: 1. Select the image to print from the Active Images screen. One (1) image or 2x3 images per sheet can be printed. 2. Press Standard Print. NOTE: If the printer is not assigned to the button, a message displays to Check Printer Button Configuration. NOTE: There is no warning to notify that the printer is not functioning. Check the printer. The printer must be configured to the Standard Print button via Utility > Connectivity > Button.

Parameters	Description
SaveAs Images	Refer to ' <i>SaveAs</i> ' <i>Images</i> on page 423 for more detail. Multiple images to save with SaveAs can be selected in the Active Image screen. NOTE: It is suggested to save the images page by page with 'SaveAs' Images in Active Images. Saving multiple images or raw data can take a long time.
Select Images/Select All Unselect All	To select one or multiple image, place the cursor on the image and press Set. To select all images, press Select all on the Touch Panel. To deselect images, press Deselect all on the Touch Panel.
Send To	See <i>Sending Multiple Exams and Patients with Single Click to PACS</i> on page 157 for more information. NOTE: "Send To" button is not displayed in Active Images menu and the Touch Panel if the patient is not selected.

Analyzing Images

To analyze the archived images, select the image, then select **Review**. The archived images is displayed with the date and time of archival.

To compare the analyzed image to a live image, press **L/R**. Now both the archived and live images appear on the monitor display. Unfreeze the live image area.

Image Tags

Image Tag buttons included on the Active Images Touch Panel screen can be used to tag or untag selected images. There are three Image Tag options (see *Tagging Recalled Images* on page 412):

- Red mark
- Green mark
- Yellow mark




Figure 9-3 • Active Images Touch Panel Screen

Tagging Recalled Images

When recalling an image, the “Tags” tab displays (see *Tag Icons on Image Clipboards* on page 412).



Figure 9-4 • Touch Panel Screen

Tag Icons on Image Clipboards

All image clipboards display the image tag icons on the image thumbnails (see *Patient View/Exam View Tags Column* on page 412).



Figure 9-5 • Active Images Screen Clipboard

Patient View/Exam View Tags Column

Image Tags icons are also displayed in a “Tags” column in the Patient View and Exam View screens (see *Patient Search by Tags* on page 413). The “Tags” column can also be sorted by tags by clicking the “Tags” column heading.



Figure 9-6 • Patient View Tags Column

Patient Search by Tags

A patient search can be performed using Tags as a Search key (see *Image Tag Presets* on page 413).




Figure 9-7 • Tag Based Search

Image Tag Presets

Image Tag preset configurations are available on the Utility > System > System Display screen (see *Figure 9-8 • Image Tag Presets* on page 414). A custom label can be entered for each tag icon, which will be displayed as the button label.

Tags can be set for a specific Print button and the tags are applied automatically on the images saved with that Print button. The ultrasound system provides the option to show/hide the Tagging option for recalled images.




Figure 9-8 • Image Tag Presets

Image Tag options are as listed below.

Table 9-2 • Image Tags Presets

Preset Parameter	Description
Image Tag Labels	Edit the text to create custom labels for each image tag.
Image Tagging on Print buttons	Assign Image Tagging on Print buttons.
Image Tag Options	Enable Image Tags when recalling images.

Tagged Image Export

Images can be exported based on selected tags from the Patient View and Exam View screens. (Tagged Image Export is not supported from Active Image screen.)




Figure 9-9 • Export Tagged Images

Image Reorder

To reorder the sequence of the images displayed on the Active Images screen:

1. Select Image Reorder on the left side of the screen.
2. Select the image(s) to move with the left and right trackball buttons.



Figure 9-10 • Select Image(s)

Recording Images

3. There are three ways to freeze the selection and toggle from Choose to Move mode:
 - Press the top trackball button
 - Double-click the left or right trackball button
 - Drag and drop by holding the left or right trackball button
4. Move the placement cursor with the trackball to the desired position to move the images to.



Figure 9-11 • Move/Paste Image(s)

5. Press the top trackball button to paste the image(s) in the new location. After pasting, the system switches from Move mode to Choose mode.
6. Select Save Reorder to save the new image order, Reset Reorder to return to the original order, or Exit Reorder to return to the Active Images screen.

Image History

Image History displays the images of each exam in chronological order of the patient.

1. Select the patient.
2. Select **Image History**.
3. Select the appropriate button showing the old exam by date and storage location. 'Active Exam' displays on the button.



Figure 9-12 • Image History Screen

4. Move the cursor to the required image for review.
5. Select **Review**. The selected image (maximum of 4 images) is displayed.

If the image data is saved on a disk and the disk is not inserted when displaying the Image History page, a triangle icon displays instead of a thumbnail.

Place the cursor on the icon. The disk name displays under the preview window. Insert the appropriate disk.

- Name of the disk displays under the preview window
- Raw data B-Mode image appears in gray.
- Raw data Color image appears in color
- Image which does not have raw data (screen capture image) appears with a question mark.

Viewing Two Different Studies From the Same Patient

To view images side-by-side from two different studies on the same patient:

1. Select the patient.
2. Go to the Image History page.
3. Select the first image.
4. Select the next image from the other exam.
5. Press Review.

Clipboard

The clipboard displays thumbnail images of the acquired data for the current exam. Images from other exams are not displayed on the current patient's clipboard.

NOTE

If there are unsaved images in the clipboard and the exam is changed, the dialogue “You have unstored images. They will be saved to your current exam” displays. The unsaved images are saved to the current exam.

All of the images can be viewed in the Active Images screen or in the Image History screen, available from the display or from the Patient menu.




Figure 9-13 • Clipboard example

Clipboard Icon



Figure 9-14 • Side Clipboard Icon

1. Number of Images in Exam

The number of images in an exam is tracked on the bottom of these Monitor Display Controls.

2. Thumbnail Size

Place the cursor on one of the Thumbnail size box icons and press Set.

3. Prev/Next Image

4. Save As
5. Delete recalled image / Delete last image

Saving the Image/CINE to the Clipboard

The active image/CINE is stored and placed on the clipboard when you press the print key (this assumes a print key has been set up to do this). The clipboard contains preview images with enough resolution to clearly indicate the contents of the image. CINE Loops are indicated by a movie clip icon.

The clipboard fills from left to right, starting in the left-hand corner. Once the top row is full, the second row starts to fill. Once both rows are full, the next image stored starts to fill a 'third' row (the first row disappears from the clipboard display, with the second row now becoming the first row, and the third row now becoming the second row).

Previewing Clipboard Images

1. Select the Pointer key to obtain a cursor arrow.
2. Move the **Trackball** to position the pointer over the clipboard image to recall.
3. An enlarged preview of the image is displayed on the left-hand side of the monitor.

Recalling Images from the Clipboard

To recall images from the clipboard:

1. Select the Pointer key to obtain a cursor arrow.
2. Move the **Trackball** to position the pointer over the clipboard image to recall.
3. Press **Set** to recall the image.
4. Press the left/right arrow of Menu icon to move to the previous/next image.

To Delete an Image from the Clipboard

1. If in image is live, press **Freeze**.
2. Select the **Pointer** key to obtain a cursor arrow.
3. Place the cursor on the clipboard image to delete, then press **Set** to select the image.
4. Place the cursor on the Delete icon and press **Set**.

A warning message is displayed asking the user to confirm the action to perform.

5. Select **Yes**.

SaveAs

Images and cineloops can be saved to a removable media storage to View on a Windows PC in the following standard formats:

- Still images: JPEG, DICOM and RawDICOM (Raw data + DICOM)
- Cineloops: WMV, DICOM and RawDICOM (Raw data + DICOM)

Using SaveAs

To save images to media:

1. Insert the media into the drive or connect the USB drive to the system.

NOTE

If the media is not formatted, it will be formatted when you select Save As.

2. On the scan screen, press the left **Set** key. The arrow cursor displays.
3. Place the cursor on the image or CINE Loop in the clipboard to be saved and press **Set**. The image displays on the screen.
4. Select **SaveAs** in the lower, right-hand corner of the screen. The Save As menu appears.

NOTE

If the image is saved as an .WMV file, run the CINE Loop before you select SaveAs.

NOTE

A 2D cineloop image can't be saved as a .jpeg file.

5. Select the media from the **Save in Archive** pull-down menu.
6. Folder name: Create the folder for the saved file.
 - Default is blank (The folder is not created)
 - Max 32 characters

NOTE

The folder name can't be edited when the folder is opened.

7. File Name: The name of the file is automatically completed, but a desired file name can be typed as well.
 - Max 64 characters

NOTE

DO NOT use the following special characters when saving images: !, @, #, \$, %, ^, &, *, (,), |, :, ;, <, >, ?, /, ~, [,], {, }, and Yen sign.

8. Store: Select Image only or Secondary capture.

- Image only: Saves only the ultrasound image area
- Secondary capture: Saves the ultrasound image area, title bar, and scan information area. Not available for DICOM or RawDICOM images.

NOTE

If “WMV” is selected for the Save as type, Secondary capture is disabled.

9. Compression: Specify Compression.

- None
- Rle
- Jpeg
- Jpeg2000
- Lossless-Jpeg

NOTE

If “WMV” is selected for the Save as type, Compression is disabled.

10. Quality: Specify image quality (between 10-100). A high quality setting gives a lower compression.

NOTE

If “WMV” is selected for the Save as type, Quality is disabled.

11. Save as type: Select one of the following.

- RawDICOM: saves the still image or CINE Loop in both GE HealthCare raw format and DICOM format.
- DICOM: saves the still image or CINE Loop in pure DICOM format.
- WMV: Saves the CINE Loop in WMV(Windows Media Video) format.

NOTE

Store “Image Only” is available if you select WMV for Type.

NOTE

WMV type is only available with CINE loop image.

- JPEG: Saves the still image in jpeg format.
- JPEG2000: Saves the still image in jpeg2000 format.

NOTE

The Save button is disabled when you select “AllFiles.” Select each Save as type when saving the data.

To see all data saved onto the local drive, select “AllFiles(.*)”. All the data names display in the window.

12. For images transferring to USB, press **Save**.

The images are saved directly to the USB drives storage whenever you press Save.

If saving to a CD/DVD, select “For Transfer to CD/DVD,” the image is saved to the local drive buffer.

- If free space of the destination is not enough to save all selected images, then warning dialog appears.
- If the same file name exists in the destination, the warning dialog displays.

OK: Overwrite file and continue to save selected images.

Cancel: Cancel.



WARNING

Buffer images for transfer to CD/DVD. Images are not cleared out of the image buffer after you start a new exam or new patient. It is the responsibility of the user to clear the image buffer before storing new patient data so that the CD/DVD does not contain images from multiple patients.

13. Repeat steps 4-12 for each additional image to be stored to the CD/DVD.
14. After adding all of the images/loops to save and ready to write to the media, transfer all the images at the same time. Press Save As > Transfer To CD/DVD.

A progress bar displays, notifying that the “Media transfer is in progress.”

If the total transfer size is bigger than CD/DVD free space size, then only the files that can be copied to CD/DVD are transferred. After the copy is finished, a warning dialog displays showing the total required file size and transferred file size. Press **OK**, insert a new CD/DVD, and press **Transfer to CD/DVD** again.

15. To not save the image(s) to CD/DVD, select “**Delete Files for Transfer**” All images are deleted.
16. Press **F3** to eject the media. Select CD/DVD Recordable or USB drive. Select **Yes and Verify files** for CD/DVD. This compares the expected number of files with the actual number of files on the media. The files are also checked to ensure that they are readable.

NOTE

The Report Save As feature is somewhat different. As soon a report is selected to be saved, the report is saved.

NOTE

If a 3D image is saved as an WMV file, an annotation text “COMP” appears at the top of the saved image which represents the compressed image.

NOTE

Time line image can be saved as multi frames image with SaveAs.

Table 9-3 • Save As Formats

	.wmv format	MPEGvue in Data Transfer
B, B+CF	Multi frames	Multi frames
B+Doppler	Multi frames	Single frame
B+M	Multi frames	Single frame

	.wmv format	MPEGvue in Data Transfer
3D	N/A	Single frame

Verify the saved image works correctly on the Windows PC. If the image does not work, please save it again on the ultrasound system.

'SaveAs' Images

Multiple images can be selected to save at one time by selecting 'SaveAs' in the Active Image screen.

Features are almost the same as the SaveAs feature. See *SaveAs* on page 420 for more information.

NOTE

It is suggested to save images page by page with 'SaveAs' Images in Active Images. Saving several images or raw data at once can take a longer time.

NOTE

If the image has a filmstrip icon, this indicates a CINE Loop, which gets saved as a .wmv file; single images are saved as a jpeg file.

NOTE

'SaveAs' Images function doesn't support images which are query/retrieved.

1. In the Active Images screen, place the cursor on the image or CINE Loop to be saved and press **Set**. Multiple images with multiple pages can be selected.
2. Press **'SaveAs' Images** on the monitor display or the Touch Panel. The SaveAs menu appears.
3. Ensure that Jpeg&WMV is selected, then press Save.

NOTE

If saving to USB, the images are transferred as soon as you press Save; if saving via Transfer to CD/DVD, images must be saved to the hard drive buffer, select 'SaveAs Images' again and then Transfer to CD/DVD. See *Notes on Transfer to CD/DVD* on page 423 for more information.

Notes on Transfer to CD/DVD

Transfer to CD/DVD transfers all saved files on the hard drive buffer to CD/DVD.

To Transfer saved files to CD/DVD,

1. After all images/loops to be saved have been added and ready to write to the CD/DVD, transfer all the images at the same time. Press Save As > Transfer To CD/DVD.

A progress bar displays "Media transfer is in progress."

Recording Images

If the total transfer size is bigger than CD/DVD free space size, then only the files that can be copied to CD/DVD are transferred. After the copy is finished, a warning dialog displays the total required file size and transferred file size. Press **OK**, change the CD/DVD and press **Transfer to CD/DVD** again.

2. To not save the image to CD/DVD, select “**Delete Files for Transfer.**” All images are deleted.
3. Press **F3** to eject.

NOTE

Verify the saved image works correctly on the Windows PC. If the image/loop does not work, please repeat the Save operation again on the ultrasound system.

Storing Images with More Resolution

To store images with more resolution than is available with the JPEG selection, select Save As and select WMV as the Save As Type. Single images can be saved as WMV files.

Table 9-4 • Store Options

Image Type	Store as Image Only	Store as Secondary Capture
CINE Loop	Provides a loop of just the image (no title bar and scan information).	Provides a single image of the video area. DO NOT USE, UNCLEAR WHICH IMAGE FROM THE LOOP IS SAVED.
Still Image	Provides a single image (no title bar and scan information).	Provides a single image of the video area.

EZBackup

EZBackup allows management of hard disk space while maintaining the patient database on the scanner, as well as backing up the patient database and images.

- EZBackup: Copy the data from the local drive to the removable media.



HINTS

Ensure that data management protocol is established for the office/institution. Media backup media **MUST** be managed by keeping a log and by creating a media filing system.

Generally speaking, back up the system when 10 GB of images are available to back up.

Assign a person in charge of performing the backups. Backups will vary by the volume of work. Track how long it takes the office/institution to get to 10 GB, and set the back-up parameters accordingly.

The office/institution should determine a backup strategy; for instance, weekly backup. It should be an easy strategy to perform and to remember and follow consistently.

It's also useful to keep more recent information on the hard drive, where it's easier to recall.

When a backup is cancelled the system completes backing up the current media and then cancels the operation.

Only cancel the backup in case of an emergency. The system completes backing up the current media and then cancels the operation.



WARNING

The ultrasound system is not intended to be used as a storage device; backup of the Patient and Image Database is the responsibility of the end user. GE HealthCare is **NOT** responsible for any lost patient information or for lost images. Loss of image data may require a rescan.



WARNING

The system crash can cause the SSD corruption. The SSD is not considered a permanent storage device. Backup data on a regular basis. Loss of image data may require a rescan.



WARNING

When EZBackup requires more space for backup, a message appears when the first disk is full. If you select "Cancel" to stop the backup procedure and later try EZBackup again, all the data may not be backed up.

Select "Full Backup" on the first EZBackup wizard screen if the last time you were performing EZBackup you selected "Cancel." Failure to follow these instructions could result in loss of patient data.

**WARNING**

DO NOT turn off the power while EZBackup is running. The data may be lost. It may take several hours for EZBackup to finish, depending on the amount of data being backed up.

The following may give the impression of a lockup, but EZBackup is continuing in the background.

- The progress bar does not move.
- The screen may become white.
- The hourglass icon keeps turning.

**WARNING**

EZBackup does not back up the exams which were previously backed up once by EZBackup or Export. If you want to back up those exams by EZBackup again, select “Full Backup Option” on the EZBackup wizard. Failure to follow these instructions could result in loss of patient data.

**CAUTION**

NEVER restore the patient archive from media made previous to the last backup.

**CAUTION**

“Archived” information is saved to each exam during EZBackup. When you perform EZBackup, the system backs up the exams except for the archived exam.

Use a USB Hard Drive for EZBackup.

NOTE

EZBackup saves data as RAW data. If you import data to the system, you can modify the image data.

NOTE

To display exported Raw DICOM images on a PC, a dedicated viewer is required.

NOTE

When backing up reports using EZBackup, use the USB HDD.

NOTE

EZBackup cannot span a single image across two (2) or more media. Therefore, if EZBackup encounters an image that is greater than the capacity of the media, it skips the oversized image.

NOTE

EZBackup does not store images to media in sequential order. Instead it maximizes the most amount of images per media.

NOTE

If the system locks up during the media auto format process, shutdown the system by holding down the power button and boot it up again. After the system is up, replace the media to a new one and execute EZBackup again. To avoid data loss, do not reuse the failed media for any other function.

NOTE

If exporting a previously backed-up exam, the message “Can’t Find Source file” displays. The image data had already been removed from the hard disk drive with EZBackup.

Perform EZBackup

When performing the EZBackup procedure, a USB HDD is connected, if applicable; the system backs up the images (or reports) and creates a reference between the patient database and the media volume.

NOTE

EZBackup can take up to 20 minutes (or longer, depending upon the size of the backup). Perform EZBackup at the same time daily, when no patients are scheduled.

1. Prepare unformatted media or the USB HDD before starting EZBackup.

NOTE

BEFORE starting EZBackup, select “Unlock All” in Utility > Admin > Logon.

2. Specify the EZBackup setup on the Utility > System > Backup/Restore page.
3. To start the EZBackup procedure, go to the Patient menu and select EZBackup. The EZBackup Wizard starts.

NOTE

If using a USB HDD, some wizards and pop-up messages DO NOT appear.

4. Verify the information on the first page of the EZBackup Wizard, then press Next.

Full backup options display on the first page of the EZBackup wizard. To backup all of the exams in the range (even if the exam was previously backed up), check this option.

If this option is unchecked, the system only backs up exams which have not yet been backed up.

EZBackup does not back up the exams which were previously backed up once by EZBackup or Export.

NOTE

If exam which is already backed up is updated, the exam is also backed up.

5. Verify the information on the EZBackup Wizard, Page 2. The backup may span multiple media. This page specifies how many media you need to perform this backup. After the media is available (allow for one extra media, just in case), the backup is ready to begin. Press Next.

Free Space/Total Size: Specifies the size of the data selected to store/and the total size of the USB Hard Drive storage media. If the storage capacity of the USB HDD is insufficient, the message “Selected Location does not have enough free space” displays.

NOTE

The message “Please insert a blank media...” appears if Next is pressed without inserting the backup media. Insert the media and continue.

6. A pop-up message appears providing the media label. Label the media, then insert the media. Press OK.

Ensure the media is labelled with not only the volume name indicated on the Insert Media Message, but with the name of the ultrasound system on which this backup procedure was performed.

Update the EZBackup log with this information the volume information and the location of the media.

After the backup has been completed, file the media.

Table 9-5 • Typical EZBackup Log

Date	Scanner ID Name	Backup Images Y/N	Reminder Dialog Interval Days	Media Label (and Scanner ID)
-	-	-	-	-

7. The status menu appears. When the backup has been completed, press Next.

NOTE

When inserting additional media for the same backup, a message appears providing you with the media label. Label the media, then insert the next media and press OK.

8. When the backup is complete, the completed wizard page appears. Press Finish.
9. Perform a patient archive after each EZBackup.

It is recommended to attach the patient list to the EZBackup media. Insert the media and select DICOM CD Read from the Patient Screen for dataflow (if using a USB drive, select DICOM USB Drive Read). Select any patient and press the left Set key to print the patient list on the digital printer.

NOTE

Use Import to restore EZBackup images.

To Review EZBacked Up Images

Backed up media can be reviewed via the Patient Menu, Import, and the DICOM Read dataflow.

To review an EZBacked Up image:

1. Select the patient on the Patient Menu (on the same system where the backup was performed).
2. Insert the media volume indicated on the Patient Menu.
3. View the exam from the media.

NOTE

A an additional media volume may need to be inserted prior to or after the recommended media.

NOTE

If the patient is split over multiple media, images on the previous or next media are displayed as triangles.

NOTE

To view the whole patient on the system, use Import, from as many media as needed for that patient. However, do not to import studies over existing studies; duplicate or missing images may result. Delete the existing exam first.

Data Transfer

The user can select and access the Exam Transfer services from the Exam Data Transfer screen.

- Import
- Export
- Worklist
- Unified Background Export

NOTE

Ensure that all patients are exported or backed up BEFORE deleting them.

Export/Import

To move exams from one ultrasound system to another system or to back up/retrieve exam information, exam information must be imported/exported.

NOTE

Both database information and images are exported. No data is deleted from the local archive when exporting data.

NOTE

Importing patient records may take more than ten (10) minutes. Please allow sufficient time to import patients.

NOTE

The media used MUST be verified BEFORE performing Export/Import. Verify the media once each session. If problems are encountered, eject the media and then re-insert the media; then try the Export/Import again.

NOTE

If exporting a previously backed-up exam, the message “Can't Find Source file” displays. The image data has already been removed from the hard disk drive with EZBackup.

NOTE

It is STRONGLY recommended to verify files on Eject when using Export.

NOTE

When dealing with a very large dataset (~500GB), the DICOMDIR file may be corrupted by the system during the DICOM Export operation. This happens in extremely rare situations, only with very large datasets, and the patient/exam dataset is not damaged or corrupted; the user may not see the patient/exam listing. In this case, regenerate the DICOMDIR using 3rd party tools. Please contact GE HealthCare service for any assistance.

Exporting Data

See *Unified Background Export* on page 432 for information on exporting data.

Importing Data

To import an exam(s) to another ultrasound system:

1. On the ultrasound system to import to, insert the media.
2. Press **Patient** and select **Data Transfer**.
3. The Data Transfer screen displays. Press Import.
4. Select the media from the Transfer From pull-down menu.
5. The Transfer From search field shows the patients available for import from the removable media just loaded onto the system.
6. Select the patient(s) or the exam(s) from the list to be imported.
7. Press Transfer. The progress bar displays during the transfer.
8. Wait for the patient information to be copied to the system. Informational messages appear while the import is taking place.
9. Press F3 to eject the media.

NOTE

Use Import to restore EZBacked up and images.

NOTE

Data can be retrieved from the media to the local drive, played back, or exam information can be processed on the system as Raw Data.

DICOM Import

Graphics that have DICOM DIR can be imported from CD/DVD/USB.



CAUTION

If the following message displays, there is a possibility that Import may not work properly ("Data detected is not LOGIQ Fortis. Measurements and RawData will not transfer.").

Worklist (Search and Retrieve the Patient/Exam Information)

NOTE

Before retrieving data from the Worklist server, ensure the default IP address is input in the Default Gateway field in **Utility > Connection Manager > Network > Wireless**.

NOTE

Select the patient prior to sending images to a PACS device.

1. Press **Patient** and select **Data Transfer**. The Data Transfer screen displays.
2. Select Worklist. The patient/exam list in the Local Archive displays in the Transfer To section.

NOTE

Only “Local Archive - Int.HD” is enabled for Transfer To.

3. The previous Worklist used is displayed on the monitor display. Press Refresh to refresh the list or select another Worklist server from the Transfer From pull-down menu.

NOTE

The worklist server is configured on the Utility screen. Multiple servers can be configured.

The auto-refresh worklist can be enabled/disabled on the Utility screen. The system automatically refreshes the list when the exam data transfer accesses the Worklist server or changes the Worklist server.

4. Select the patient(s) or the exam(s) from the list.
5. Press Transfer. The progress bar displays during the transfer.

Unified Background Export

Export from Patient Screen

1. Export Patients/Exams from the Exam List on the Patient screen by selecting **Export To** (see *Figure 9-15 • Patient Exam List* on page 432).



Figure 9-15 • Patient Exam List

2. Selected exams can be exported as DICOM or MediaFileType formats (WMFV, JPEG, etc). (DICOM is the default format.) If **Export Only Tagged Images** is selected the Tags will be displayed to select from (see *Figure 9-16 • Export Dialogue with Tags* on page 433).



Figure 9-16 • Export Dialogue with Tags

After selecting “Export” a message box appears with Anonymize Patient information. Select “OK” on the popup to proceed and select “OK” for the desired Anonymization to submit the exams to spooler for background export (see Figure 9-17 • Anonymization on page 433).




Figure 9-17 • Anonymization

NOTE

If the export Patient ID already exists, a prompt will appear to confirm that existing exam information may be overwritten.

NOTE

To create a Local Copy, check **Anonymize**, and then **Create Local Copy**.

- After submitting the job to the spooler, a message box displays the background job submission status.




Figure 9-18 • Background Export Submission Status

Export from Active Images Screen

1. Export images from the Active Images screen by selecting the images and selecting **Export To** from the menu (see Figure 9-19 • Active Image Export on page 434). Images can be exported as DICOM or as the MediaFileType formats listed in Table 9-6 • Export To Media File Types on page 434. (MediaFileType is the default Export As format.)

Table 9-6 • Export To Media File Types

Still Image	Clip
BMP	AVI
JPEG	MP4
PNG	WMV



Figure 9-19 • Active Image Export

2. After submitting the job to the spooler, a message box displays the background job submission status (see Export from Active Images Screen on page 434).

Spooler Screen

Launch the spooler by pressing the F4 key on the keyboard or by clicking the status bar media icon. The status bar media icon updates based upon the status of the media spooler. The DICOM spooler is accessed on the **Network Jobs** tab and the Media spooler on the **Media Jobs** tab.



Figure 9-20 • Spooler Screen

During background export to optical media, if the media is unformatted it is formatted automatically, as a background process, during which the spooler status column displays Formatting.

After the background export to optical media, the exported content is burned to the media, as a background process, during which the spooler status column displays Writing.

For failed spooler jobs, the error description and recommendations are displayed. **The Retry** and **Transfer To...** selections are enabled based upon the nature of the error.

Spooler Status Icons

The Media spooler status updates with the media icons, displayed in the status bar.

Table 9-7 • Media Spooler Status

Icon	Spooler Status
	Export to USB media is active.
	Failed export jobs to USB in the spooler.
	Export to optical media is active.
	There are failed export jobs to optical media in the spooler.

If media is ejected while an active job is saving to that media, a warning message displays.

On system Shutdown, if any export is active a shutdown confirmation dialogue displays. The shutdown can be cancelled, the spooler status can be checked, or the shutdown can be continued. If the shutdown is continued, ongoing and pending spooler jobs are suspended.

NOTE

The suspended jobs will not be auto resumed when the system is powered on again.

DICOM Viewer

This feature enables the export of a self-contained DICOM viewer to removable media, along with the transfer of selected images.

NOTE

The exported DICOM Viewer software is a GE HealthCare developed DICOM Viewer. It is standalone software to display ultrasound images exported from GE HealthCare LOGIQ Ultrasound systems.

NOTE

This viewer is NOT intended to be a diagnostic tool. It is only meant to be used for reference.

Installing the DICOM Viewer

DICOM Viewer allows the user to transfer examinations to DICOM removable media together with the DICOM Viewer.

To install:

1. Place the media in any Windows PC. Open the media and click the DicomViewer.exe file to install the DICOM Viewer.

NOTE

Minimal Windows PC requirements are Windows 10/Windows 11.

2. The License Agreement appears. Select **I Agree**.



Figure 9-21 • License Agreement

3. The DICOM Viewer is displayed.



Figure 9-22 • DICOM Viewer

Configuring the DICOM Viewer

The user can configure the embedding of the DICOM viewer on the media on the Export To screen.

1. Export exams from the Exam List on the Patient screen by selecting **Export To**.
2. Check **Add DICOM Viewer to Media**.
The DICOM Viewer will be exported when transferring examinations to the corresponding DICOM removable media.



Figure 9-23 • Add DICOM Viewer to Media



Figure 9-24 • Create Local Copy

External Drives

Intended Use

Removable media can be used for the following purposes:

- Backup of patient database and system configuration presets (see *System Backup and Restore Preset Menu* on page 467).
- Export to copy a set of patient records to a third party DICOM review station.
- Copy system configuration presets from one system to another using the Backup/Restore feature (see *Preset Synchronization Using Media* on page 603).
- Use SaveAs to save images as JPEG, WMV, DICOM and RawDICOM for review on a Windows PC.
- Privacy and Security Encryption (see *Privacy and Security* on page 688).

Supported Removable Media

The following removable media is supported:

- USB Flash Drive
- USB external hard disk drive
- CD-R/DVD

If a CD/DVD is used, it is recommended to use an Archive Grade or Medical Grade CD/DVD.

Whichever media is used, it is highly recommended to make a backup of the media, which is the responsibility of the customer.



HINTS

Storing the media disc in an original media case or caddy will prevent it from becoming dirty or damaged.

DVD Drive

The media below can be used in the DVD drive to perform software upgrades, image archiving and service diagnostics:

- CD-R (maximum write speed: 48x)
- DVD-R (maximum write speed: 16x)

CD and DVD Handling

To avoid data loss, never touch the recordable surface of a disk. Handle the disk only by the outer edge. Do not place it face down on a hard surface. Fingerprints or scratches will make the disk unusable. Before usage, verify that the disk surface has no visible scratches. If there are any scratches, do NOT use the disk.

Inserting Media

Insert the media by pressing the eject button on the DVD drive.



CAUTION

When putting media into the DVD drive, please make sure the media is placed in the right position. If the media is not placed in the right position, the media may be damaged.



CAUTION

Make sure the disc tray is securely placed in each device during system operation. Mechanical damage may occur if other objects hit the tray.

Ejecting Media

1. To eject the media, always press F3. DO NOT use the eject button on the drive.
2. The Eject device menu is displayed. Select **CD/DVD Recordable**.
3. When ejecting media, a prompt appears asking whether or not to finalize the disc. If the disc is not finalized, files can be added to the media at a later time. However, the files stored on this disk may not be viewable with a standard PC. Finalizing media allows viewing of files using most standard PCs.

To verify that the data was successfully transferred to the media, Press F3, then select "Finalize" > "Yes and Verify Files." If the media was corrupted during an operation, the message "An error occurred on the last disk. Please discard it and start over." displays. In this case, please repeat the operation with new media.

NOTE

When pressing **F3** with a blank media inside the DVD drive, the finalization menu displays; selecting "Yes" or "Yes and Verify Files" ejects the media without finalizing the media.

USB Hard Disk Drive and USB Flash Drive

Cautions and Warnings



WARNING

Before removing the USB drive from the USB port, press Eject (F3) and select USB Drive from the pull-down menu. Disconnect the USB drive after the success dialogue is displayed. If the unsuccessful dialogue is displayed, retry after a while. Failure to follow these instructions could result in loss of patient data.



WARNING

If a problem occurs while exporting to the USB-HDD, such as a crash, the export may not have completed. Try again with a smaller number of patients.



WARNING

DO NOT use "Select All" when you export the patient data to the USB-HDD.

NOTE

Connect USB Flash Drives and Hard Disk Drives to the USB ports located on the Touch Panel (USB 3.0). Connect peripheral devices to the USB ports located at the rear of the system. (USB3.0).

NOTE

Do not insert USB Memory devices (hard drives or flash drives) that contain multiple partitions into the scanner. Use single partitioned USB Drives.

NOTE

Some USB memory device manufacturers allow for executable partitions or ship pre-formatted new USB memory devices with multiple partitions pre-configured. BEFORE inserting any memory device into the scanner, insert it into a PC or MAC to verify that there is only a single partition. If multiple partitions exist, contact the USB manufacturer for the steps in reformatting the memory to a single partition.

USB Ports



CAUTION

Non-supported peripheral devices that use their own AC power source CANNOT be attached to the ultrasound system. DO NOT connect the peripheral device's power cord into the ultrasound system. Only peripheral devices purchased from GE HealthCare with the purpose of being used with the ultrasound system should be used.

Use a USB printer cable that is less than 3 meters in length.

Failure to follow these instructions could lead to unexpected diagnostic performance.

System USB Ports

Side of Monitor

The two Monitor USB 2.0 Ports SHOULD ONLY BE USED for Bus-powered USB Hard Disk Drives and USB Flash Drives. The following configurations can be used:

- One or two USB Flash Drives
- One Flash Drive and One Bus-powered Hard Disk Drive
- One Bus-powered Hard Disk Drive



CAUTION

DO NOT plug in TWO Bus-powered Hard Disk Drives at the same time.

Operator Panel and Rear of System

The two USB 3.0 ports on the Touch Panel and at the back of the system SHOULD ONLY BE USED for the following devices:

- Color or Report Printer

NOTE

When connecting an external printer to the ultrasound system via the USB port on the back of the system, ensure that the power supplied to the printer is fed from the same power feed as the ultrasound system. This assures compliance to leakage currents.

- Footswitch
- PC Printer with isolation USB

- USB Storage (USB-HDD and USB Flash Device)

Ejecting a USB Flash Drive/USB HDD

1. To eject a removable media, always press F3.
2. The Eject device menu is displayed. Select the relevant media.
3. Select USB Drive from the pull-down menu to disconnect the USB Drive. Disconnect the USB drive after the Success dialogue is displayed.

Remove the USB Drive from the USB port.

NOTE

If the Unsuccessful dialogue is displayed, retry after a short period of time.

NOTE

Verify is NOT available on Flash Drives or Hard Disk Drive media.

SaveAs

NOTE

See SaveAs on page 420 for more information.

To save images to the USB Flash Drive or USB HDD:

1. Insert the USB Drive into the USB port.
2. Select the image(s) to be saved.
3. Select **Save As** menu in the lower, right-hand corner of the screen. Select the USB Drive as the archive media.
4. Specify: Image only or Secondary Capture, type of compression, quality, and image save format (Raw DICOM, DICOM, Jpeg, or WMV).
5. Press **Save**. When the images have been saved, press Eject (F3).

NOTE

If the SaveAs function is performed to the USB drive (:\\Export) by RawDICOM format and review the data on your PC, the title of the data appears as “:\\GEMS_IMG\\2006_Oct\\08(date)\\xxxxx(PatientID)”.

Direct SaveAs

The image can be saved directly to the USB Drive just by pressing a **Print** key.

1. Insert the USB Drive into the USB port.
2. Select **Save As** from the pull-down menu in Utility > Connectivity > Service. Press **Add**.
3. Select **Save As** in the list. Rename it in the Name field if needed.
4. Select USB Drive in the Destination field.
5. Verify the service.
6. Press **Save**.

Recording Images

7. Assign Save As to the appropriate print key in Button tab.
8. Display the image on the monitor and press the print key.

Export/Import

NOTE

Before exporting exams to the USB HDD, check **UtilityConnection ManagerMyComputer DeviceHD Export**.

To export/import exams using the USB Flash Drive or USB HDD:

1. Insert the USB Drive into the USB port.
2. On the Patient menu, select Data Transfer, then Export/Import. Specify USB Drive in the transfer To: pull-down menu. Select the patient/exam you want to transfer. Press **Transfer**.
3. When Export/Import has completed, press **F3**.

EZBackup (USB HDD Only)

1. Select “USB Drive” at Utility > Backup/Restore > EZBackup > Media.
2. Follow the instructions for EZBackup. (See *EZBackup* on page 425 for more information.)
3. When EZBackup has completed, press F3.

USB Quick Save

USB Quick Save easily sends images to a USB flash drive storage.

The images are stored either in .jpg or .WMV format.

USB Quick Save Setup

In Utility > Connectivity > Service, select USB Quick Save.



Figure 9-25 • Service USB Quick Save

Assigning USB Quick Save to Print Keys

The USB Quick Save service can also be assigned to the Print keys via the Utility > Connection Manager > Print Button preset menu. See *Print Button Page* on page 542 for more information.



Figure 9-26 • Assigning USB Quick Save to Print Keys

Printing Options

Setting up Digital Peripherals

Set up digital peripherals from the **Utility > System > Peripherals** menu.

The following printer can be connected to the power outlet supplied by the system.

- BW Printer: UP-D898DC

For the UP-DR80MD and UP-D25MD, insert the USB cable into the USB port of the rear panel and plug it into a wall socket.

NOTE

Printing using a standard printing service overrides the orientation and N-up feature of the printer preferences. Printer preferences are set up in the printer folder (via **Utility > System > Peripherals**. Select Properties under Standard Printer Properties).

Digital Printer Setup

There are two steps to do when setting up a digital printer: 1.) Follow the procedure below for each printer, then 2.) Set up specific properties for each printer (specific instructions are provided for each printer following this section).



CAUTION

- Before powering on the ultrasound system, connect the printer via the USB cable and turn the printer power on.
- DO NOT remove the cable while the ultrasound system is powered on.

Follow this procedure for each printer:

1. Select Utility > Connectivity > Service. Add the Standard Print service. Select the printer from the Printer pull-down Properties menu. For the UP-D898DC printer, select "Portrait" as orientation.
2. Type the printer name in the Name field. This name is used on the Button screen. After selecting the printer from the Printer pull-down Properties menu again, it turns white. Press Save.
3. Select Button. Select the appropriate print key (Print1, Print2...) from the Physical Print Buttons section. Select the printer from the MyComputer column and press >> to move it to the Printflow View column. Press Save.

Sony UP-D25MD Instructions

Follow these steps to set up the paper size of Sony UP-D25MD printer.

1. Press Utility-->System-->Peripherals. Select the UP-D25MD from the pull-down menu under Standard Printer Properties. Click **Properties**.



Figure 9-27 • Standard Printer Properties

2. Select Properties.



Figure 9-28 • Properties

3. Select Printing Preferences. Select Paper Size. Press **Apply**. Press **OK**. Press **Save**, then **Exit**.



Figure 9-29 • Printing Preferences



CAUTION

- Before powering on the Ultrasound system, connect the UP-DR80MD and turn the printer power on.
There is no warning to let you know that the printer is not functioning. Check the printer.
- DO NOT remove the cable while the Ultrasound system is powered on.
- DO NOT plug in a DR80MD while the Ultrasound system is turned on.

UP-D898DC Printer Settings

Preferred printer settings for the UP-D898DC are shown below; and instructions follow.

Table 9-8 • Recommended Settings

Recommendations	Use These Settings
Recommended Setting	Paper Size: 1920x1280 Windows Orientation: Landscape Utility > Connectivity > Service > Standard Print > Orientation: Landscape
Avoid using WIDE Setting	Utility > System > System Display > Use the settings below: Utility > System > System Display > Image Display > Image Display Area: Default or Large Utility > System > System Display > Use Wide Screen For...: turn OFF all parameters
Use Large Print Setting	Paper Size: 1920x1280 Windows Orientation: Landscape Utility > Connectivity > Service > Standard Print > Orientation: Portrait

Setting up the UP-D898DC Printer Settings

1. Select Portrait for Orientation in Utility > Connectivity > Service > Standard Print and press Save.
2. Select Utility > System > Peripherals.
3. Select the printer to adjust (UP-D898DC) from the pull-down menu under Standard Printer Properties. Click Properties.
4. Select Properties from the Printer pull-down menu.
5. Click Printing Preferences at the bottom of Properties Window.
6. Select the Layout tab and select:
 - Paper: 1920x1280
 - Orientation: Landscape
 - Interpolation Method: Bilinear
7. Select the Density Adjust tab and select:
 - Gamma: TONE2
 - Sharpness = 0; Dark = 0; Light = 0; Sharpness = 2
8. To save the adjusted printer settings, click Apply and then OK.
9. Close the 'Printers' window with the close button.

10. Exit System Setup with Save&Exit.
11. Assign the Printer to the remote keys.

Network Printer

A network printer can be added to the system.

NOTE

It is solely the customer's responsibility to determine printer and network settings and configuration to enable and ensure the specific network printer usage and functionality.

Add Network Printer

To add a network printer, the user must be logged on to the system as an administrator.

1. On the Utility > System > Peripherals page, Under "Network Printer," select **Add Network Printer**.




Figure 9-30 • Network Printer

The "Add Printer" Dialog Box appears.

2. In the Add Printer dialog box, enter a Printer Name, the IP Address of the desired printer, and select the Printer Driver from the following dropdown options:
 - Microsoft enhanced Point and Print compatibility driver
 - EPSON Universal Print Driver

Select **Add Printer**.



Figure 9-31 • Add Printer

Adding a printer may take several minutes. A dialog box confirming installation displays when the printer has installed. Select **OK**.

NOTE

Printer installation does not guarantee the printer will function correctly. The printer may install even if the IP address is incorrect or inaccessible.

3. After installing the printer, go to Utility > Connectivity to add a Standard Print Service and a Print Button.

Print a test exam image or report and verify the full image or all information has printed. Verify the resolution, aspect ratio, clarity and color accuracy of the prints.

Remove Network Printer

To remove a network printer, the user must be logged on to the system as a Network Administrator.

1. On the Utility > System > Peripherals page, Under “Network Printer,” select **Remove Network Printer** (Figure 9-30 • Network Printer on page 447).
2. Select the printer to remove from the drop-down list and select “Remove Printer.”



Figure 9-32 • Remove Printer

A dialog box confirming printer removal displays when the printer has been uninstalled. Select “OK.”

External Paper Printer

You can connect an external paper printer via the USB connection.



CAUTION

ONLY plug in devices to the USB ports located at the rear of the system WHILE the ultrasound system is NOT powered up. If a device is plugged in while the ultrasound system is powered on, the system may become unusable.



CAUTION

DO NOT place an external paper printer inside the patient environment. This assures compliance to leakage current.




Figure 9-33 • Patient Environment

External Paper Printer Setup

NOTE

The printer driver is customized for the ultrasound system at the factory; you do not need to change the settings.

1. Connect the printer to the USB port.
2. Select **Utility**--> **Connectivity** > **Service**. Add the Standard Print service.



Figure 9-34 • Connectivity -> Service Screen

3. Select the printer from the Printer pull-down Properties menu.

NOTE

After selecting the printer, the field turns white.

Recording Images

4. Set the following parameters in Properties: Rows, Columns, Orientation, and Right Margin.
 - Rows=3
 - Columns=2
 - Orientation=Portrait
 - Right Margin (mm)=10
5. Type the printer name in the Name field.

NOTE
This name is used on the Button screen.
6. Press Save, then select the Button tab.
7. Select the appropriate print key (Print1, Print2...) from the Physical Print Buttons section.
8. Select the printer from the MyComputer column and press ">>" to move it to the Printflow View column.



Figure 9-35 • Connectivity -> Button Screen

9. If you want to assign this printer to the Standard Print Button on the Active Image Screen, select this printer at the Active Image Printer section.
10. Press Save.

NOTE

Adjust the rows, columns, and margins as necessary to obtain the image size and quality that you find acceptable.

Setting up the Printer to Print Reports and Patient List Print

To set up the external printer to print reports and Patient List Print,

1. Press **Utility > System > Peripherals** and select Printers under Setup.
2. Select the printer from Default Printer pull-down menu.



Figure 9-36 • Report Printer Setup

3. Press **Save**.
4. Press Print on the Report screen to print the report.

Portable Exam

Perform a Portable Exam (Using the Worklist)

To perform a portable exam (using the worklist),

1. Go to the worklist and get the patient(s) you will need for this portable exam(s).
2. Ensure that images will be saved to the local hard drive.
3. Press the Power On/Off switch.
4. Proceed to perform the exam(s). Press Power On/Off. Select the patient, do the exam. Store images. Images are held in the spooler.
5. If there are additional patients to scan, repeat steps 4 and 5. After you have completed all portable exams, press Power On/Off.
6. When you have returned from performing the portable exam(s), reconnect to the network. Press Power On/Off.
7. Press F4. Send all of the images in the spooler to the printer/storage device.

Chapter 10

Customizing Your System

In this section

Presets.	454
System Presets.	456
Imaging Presets.	473
Comments Libraries Presets.	476
Body Patterns Presets.	483
Application Presets.	487
User Specific.	492
Test Patterns.	493
3D/4D.	494
Configuring Connectivity.	497
Measure.	561
Reports.	562
System Administration.	563
Imaging Preset Manager.	594
Backup and Restore	599
Search.	604

Describes how to create system, user, and exam presets.

Presets

Preset Menus provide the following functionality:

Utility, Page 1 Configuration Menus




Figure 10-1 • Utility Menu, Page 1

- System presets. View and update general system configuration settings, measurement and analysis settings, and video settings; backup and restore data and configuration files.
- Imaging presets. View and update exam and imaging parameters.
- Comment library presets. Set up comment libraries by application.
- Body Pattern library presets. Set up body pattern libraries by application.
- Application and User Defined presets. Configure application- and user-specific settings.
- User Specific. Configure My Desktop.
- 3D/4D. Real-time 4D and Static 3D scanning.
- Connectivity Setup. Define connection and communication setup, DICOM, Wireless Networking, Tricify, and includes exam dataflow information.
- Measurement and Analysis presets. Customize exam studies, create measurements, set up manual sequencing, and create OB Tables. Refer to the “General Measurement and Calculations” chapter for more information.
- Reports Presets. Allows you to edit the report template, diagnosis codes, and report comments. Refer to the Advanced Features chapter for more information.
- Administration presets. Perform system administrator activities such as setting up user IDs and logon formats.
- Image Preset Manager. Activates the Image Preset Manager to create, edit, import, and export user imaging presets.
- Scan Assistant. Create, import/export, and manage Scan Assistant programs.
- Search. You can search for a parameter on the Utility pages (Measure, Reports, and Service pages cannot be searched.)

Utility, Page 2 Configuration Menus



Figure 10-2 • Utility Menu, Page 2

- Electronic Instructions for Use (eIFU) Icon. Activates the Online User Manual.
- Scan Screen. Activate the Scan Screen so that you can adjust the monitor's brightness/contrast via the Room Profile control and adjust the monitor's RGB level/color temperature/gamma via the Color Profile control.
 - Room Profile. Change Brightness and Contrast of main display when set to User Defined.
 - Color Profile. Control RGB level, color temperature, gamma, etc.
 - Gamma. Set the Gamma at 2.2 or 2.4.
 - Color Space (OLED only). Select ITU 709 Standard for High Definition TV, Native 1 for most vivid OLED colors, or Native 2 for a setting between ITU 709 and Native 1.
 - Reset Monitor. Resets the monitor to system default settings.
- Test Patterns. Test the Monitor screen's performance.
- Service. Activate the Service Browser.

To access these functions, select the **Utility** tab on the Touch Panel, then select the appropriate Touch Panel key.

In addition, you can adjust the following via the rotaries located beneath the Touch Panel.

- Master Volume. Select to adjust the system volume, e.g., Doppler volume.
- Effects Volume. Select to adjust the system notifications, e.g., Touch Panel pushes, Print sounds, etc.
- Touch Panel Light. Select to adjust the brightness of the Touch Panel.
- Kbd/Touch Lock. Select to lock the Keyboard and Touch Panel controls in order to clean the system.
- Brightness/Contrast. Select to adjust the monitor brightness/contrast.
 - User-adjustable when Room Profile is set to "User Defined."
- OPIO backlight. Controls the level of brightness of the orange and green backlights on the hard keys.

System Presets

System presets allows you to view or change the following parameters:

- **General** – Location, Date/Time, Patient Info, Key Usage, and Utility configuration
- **System Display** - Presets related to the monitor display format.
- **System Imaging** – CINE Loop Store, Cardiac, Biopsy Guides, and Image Control and Display configuration
- **System Measure** – Measurement, Cursor, and Results Window configuration
- **Backup/Restore** – Backup, Media, EZBackup, Detailed Restore of User Defined
- **Peripherals** – DVR, Print and Store Options, and Setup configuration
- **User Configurable Key** -- BT (Breakthrough) Key, Keyboard Key
- **About** – Release information and part numbers for: Application Software (including version), Online Help, Service Platform, Security Package, Base Image and Base Image Update.
- **Licenses** – Licenses for software used on the ultrasound system.

Changing system parameters

To change system parameters:

1. On the Touch Panel, select **Utility**.
2. On the Touch Panel, select **System**.
The System screen is displayed.
3. On the monitor display, move the **Trackball** to select the tab that has the information you want to change.
4. Select values for the parameters you want to change.
5. To save the changes, select the Save button. Select Exit to return to scanning. In some cases, you may need to reboot the system for the change to take effect.

System/General Preset Menu

The System/General screen allows you to specify hospital name and system date and time.

Table 10-1 • Location

Preset Parameter	Description
Hospital	Type the institution's name.
Department	Type the institution's department name.
Machine Description (1&2)	Type the machine name.
Preset Region (restart needed)	Select region (Global, Europe).

Preset Parameter	Description
Language (restart needed)	Select the appropriate language from the drop-down list. Note: If you select Japanese (JPN), only the warning and status messages are displayed in Japanese. You can not type in Japanese.
Online Help Language	Select the language you would prefer to use to read the online user manual.
Units	Select metric or US units of measurement.
Regional Options (restart needed)	Select to set up the keyboard.

Table 10-2 • Date/Time

Preset Parameter	Description
Time Format	Select the time format: 12 Hr. AM/PM or 24 Hr.
Date Format	Select the date format: dd/mm/yyyy, mm/dd/yyyy, or yyyy/mm/dd.
Default Century	Select the default century for the system to use.
Date/Time (restart needed)	Select to display the Date/Time Properties window, to specify the system date, time, time zone, and to auto adjust for daylight savings time.

Table 10-3 • General User Interface

Preset Parameter	Description
Color Level (restart needed)	Select the darkness of the utility pages and any message windows: Brightest, Bright, Standard (Dark Text), Standard (Light Text), Dark, Darkest.

Table 10-4 • Title Bar

Preset Parameter	Description
Hide Patient Data	Always - Patient information is removed from the scanning screen Title bar (and while storing images). On Store - Patient information is removed only when storing the image. Never - Patient information is always displayed. NOTE: Upon recall of images with measurements, Dual image and/or V Nav, the DICOM image is recalled. In this case, there is no patient data burned into the DICOM image. If you want the data to appear on the image, set to Never.
Hide DateTime	Always - Date and time is removed from the scanning screen Title bar (and while storing images). On Store - Date and time is removed only when storing the image. Never - Date and time is always displayed. NOTE: Upon recall of images with measurements, Dual image and/or V Nav, the DICOM image is recalled. In this case, the date and time are not burned into the DICOM image. If you want the date and time to appear on the image, set to Never.
Font Size (restart needed)	Select to display patient information in the title bar using a small or large font size. You need to reboot the system for this change to take effect.

Table 10-5 • Worksheet

Preset Parameter	Description
Use Bold Worksheet Fonts (Restart Needed)	Select to bold fonts in Worksheets. The setting will take effect when the system is restarted.

Table 10-6 • Trackball

Preset Parameter	Description
Speed	Set how fast you want the Trackball to move while performing actions such as tracing the anatomy. Speed is used for Menu and Scan Mode. 0=Slow; 20=Very Fast
Acceleration	Set how fast you want the Trackball to move across the display. Acceleration is used for Menu and Scan Mode. 0, 1, and 2 with 0 being the slowest acceleration.

Table 10-7 • Key Usage

Preset Parameter	Description
Run Fast Key speed	Select the maximum value of the key interval when running Fast Key.

Table 10-8 • Utility

Preset Parameter	Description
Prompt for Save on Exit	If selected, the system prompts you to save data when you select exit without saving.
Utility Font Size	Select the font size you want to use to view the Utility menus: Small, Medium, or Large.

Table 10-9 • Scan Assistant

Preset Parameter	Description
Always Use Doppler Cursor	Use the Doppler Cursor when you activate Scan Assistant.

Table 10-10 • Start Assistant

Preset Parameter	Description
Use Start Assistant	On: Use Exam Description (default) - Saves exam mapping automatically from worklist On: Use Scan Assistant only - Exam description is ignored and selection of probe and preset is based on Scan Assistant protocol only. Exam Category and Scan Assistant program are not automatically selected. Start Assistant Editor does not show the Exam description column. Manual entry of an Exam Description for table entry is prevented. Off
Create/Edit	Select to open Start Assistant Mapping Editor.

Table 10-11 • V Nav

Preset Parameter	Description
Max.non US images per exam on Image History	Specify the number of non-Ultrasound images per exam to appear on the Image History page.

Table 10-12 • Touch TGC

Preset Parameter	Description
Delay Time for Auto Hide (sec)	Off, 3, 6, 8, 12, 16
Custom Settings	Select Application or Category
Include TGC Curve when Saving Presets	Check to select.

Table 10-13 • Miscellaneous

Preset Parameter	Description
Reset Control Panel Park Lock	Press to reset operator panel park lock.

Table 10-14 • Scan on Battery

Preset Parameter	Description
Auto Switch to Power Saving Mode	15 minutes, 30 minutes, 1 hour or Never
Freeze When AC is Unplugged	Select to freeze when AC power is unplugged from system.

Table 10-15 • Touch Keyboard

Preset Parameter	Description
Auto Popup Keyboard on Patient Screen	Display Touch Keyboard automatically when Patient screen is selected.
Auto Hide Delay Time for Touch Keyboard (sec)	Off, 6, 8, 12, 16, 20, 24, 30
Display Keyboard with Comment Button	Display Touch Keyboard automatically when Comment mode is activated.

System/System Display Preset Menu

The System/System Display screen allows you to specify parameters for the Monitor Display.

Table 10-16 • Image Display

Preset Parameter	Description
Image Display Area	Select Image Display Area Size: Default, Large, Extra Large

Preset Parameter	Description
Image size (probe selection required)	Select Default or Large.
Magnification Zoom Available	Select to enable Magnification Zoom on images (enabled by default).

Table 10-17 • Side Panel Content

Preset Parameter	Description
Side Clipboard	Display On/Off.
My Desktop	
Measurement Summary	

Table 10-18 • Clipboard

Preset Parameter	Description
Preview Image	Display On/Off
Show Zoom Reference Image	
Bottom Clipboard	Display On (Always display)/ Off (Never display)/ Auto (Display whenever there is no side clipboard)
Bottom Clipboard Auto Dimming	On/Off/Auto When the windows pointer is moved over the clipboard area, the pointer is undimmed.
Side Clipboard Auto Dimming	

Table 10-19 • Use Wide Screen for...

Preset Parameter	Description
Dual Screen	Automatically switch to Wide Screen when in Dual Screen.
DualView (Simultaneous)	Automatically switch to Wide Screen when in Simultaneous DualView Screen.
Contrast DualView	Automatically switch to Wide Screen when in Contrast DualView Screen.
LOGIQView	Automatically switch to Wide Screen when in LOGIQView.
Volume Nav	Automatically switch to Wide Screen when in Volume Navigation.
QAnalysis	Automatically switch to Wide Screen when in QAnalysis. Side by Side Timeline automatically switches to wide screen when in Timeline mode.
Display Format Horizontal Timeline	Side by Side Timeline automatically switches to wide screen when in Timeline mode.
Single Image	On/Off/Auto Auto turns on wide screen if 2D image exceeds width of non wide screen image area.

Table 10-20 • Display

Preset Parameter	Description
Horizontal Scale	Select to display width markers.
TGC Display	Select to display TGC curve.
PW Velocity Units in cm/s	Select to change scale on timeline from centimeters per second to meters per second.
Shear Elasto Display Units	Select m/s or kPa.
Shear Stiffness and Velocity Measurement	Specify whether to display Shear Wave Stiffness and Velocity measurements.
Shear Elasto Color Map	Select Red as Hard or Blue as Hard.
Strain Elasto Color Map	Select Red as Hard or Blue as Hard.
UGAP Display Units	Select dB/cm/Mhz or dB/m
Attenuation Coefficient and Rate Measurement	Select to display Attenuation Coefficient and Rate Measurement
Image Parameter Size (restart needed)	Choose Small, Medium, Large, or Extra Large. Must reboot the system.
Highlight Image Parameter Changes	Select if you want the display to indicate which controls you adjusted by highlighting the new value on the display.
Use Thicker V Nav Graphics	Select to display thicker V Nav graphics
Overlay Color (single visible dataset)	Select Red, Green, Blue or BW (Black/White).
PET Color in V Nav (if grayscale-only dataset)	Select Hot Body, Hot Metal, Warm Metal, BW (Black/White).
Hide Multiple Dataset Menu When Storing	Select to hide the Multiple Dataset Menu when storing.
Live/Freeze Indicator	Display On/Off.
Enable DICOM grayscale display mode (GSDF)	Enable On/Off. Adjust Gamma curve on DICOM GSDF. Note: With GSDF disabled, the monitor still uses a gamma curve that may be selected on the Touch Panel (for backwards compatibility with sites that are happy with their PACS or may have a mix of older systems). With GSDF enabled, the gamma button on the touch panel affects the image, emulating the appearance of that gamma, but does not affect the monitor.
Room Profile	Dark, Semi Dark, Light, User Defined, Last Used

System/System Imaging Preset Menu

The System/System Imaging screen allows you to specify parameters for key usage and image control and display.

Table 10-21 • Biopsy Guides

Preset Parameter	Description
Show Center Line	Displays center biopsy guideline.
Show Outer Lines	Displays outer biopsy guidelines.
Enable 0.5cm markers	Activates biopsy depth markers every 0.5cm.
Show Biopsy Mark on CFM Simultaneous Mode	Displays the Biopsy Guideline on the image while in Simultaneous Mode.
Show Biopsy Mark on Dual View Mode	Displays the Biopsy Guideline on the image while in Dual View Mode.
Show Biopsy Circle	Specify whether to display the Biopsy Circle with the Biopsy guideline.

Table 10-22 • Compare Assistant

Preset Parameter	Description
Comparison Image Side	Select Left or Right.
Comparison Image Date	Select All Dates, Different Date or None.
Copying Settings	Select Automatic: Imaging and Annotations, Automatic: Imaging Only, Automatic: Annotations Only, Manual: Imaging and Annotations, Manual: Imaging Only, Manual: Annotations Only or Off.

For details on the Compare Assistant presets, see *Setting Up Compare Assistant* on page 1014.

Table 10-23 • Image Label Layout

Preset Parameter	Description
Clipboard	Select 1-Line Label, 2-Line Label or No Label.
Active Images	Select 1-Line Label, 2-Line Label or No Label.
Image History	Select 1-Line Label, 2-Line Label or No Label.

Table 10-24 • Image Label Color

Preset Parameter	Description
Clipboard	Select the color of the Clipboard Image Label.
Active Images	Select the color of the Active Image Label.
Image History	Select the color of the Image History Image Label.

Table 10-25 • Contrast Clock Highlight

Preset Parameter	Description
Interval (sec)	Select the interval notification by contrast clock highlight.
Highlight Time (sec)	Select the duration of contrast clock highlight.
Contrast Timer Sound	Select the sound of contrast clock highlight.

Table 10-26 • Controls

Preset Parameter	Description
Auto Invert on Linear Steer	When selected, automatically inverts the color scale or spectral timeline when the Steer function is used.
Auto Invert on ASO	Automatically inverts the spectrum with ASO.
Link Color/Doppler Invert	When selected, the Doppler timeline scale inverts along with the color ROI.
Pushing Depth Rotary Performs Image Reverse	When selected, you can reverse the image when you push down on the Depth rotary.
Toggling Zoom Rotary Performs Depth	When selected, you can adjust the Depth by moving the toggle up and down.
Audio Volume	Adjusts the Doppler audio volume via a drop-down menu (for example, 0=softer; 20=louder).
Auto Freeze Time (probe selection required)	Automatically freezes the system after 10, 30 minutes, 1 hour of inactivity, or never.
Countdown Time For Contrast (sec)	Specify time for the Contrast Clock to countdown during a contrast study, 0 (off), 3, and 5 seconds.
Reverse Depth Control	Changes key direction for the Depth control.
Reverse Steer Controls	Changes key direction for the Steer controls.
Turn Off CrossXBeam for LOGIQView (non-linear probes)	Deactivates CrossXBeam when you activate LOGIQView.
3D Postprocessing when reloading	When selected, the system re-processes the recalled 3D CINE Loop.
Tru 3D/Easy 3D Resolution	Set Easy 3D/Tru 3D Resolution: Default, High, or Very High.
Doppler Scroll Priority	Set to 2D, Doppler, or Last Live Mode.
Start Doppler in Update	Select to allow the B/CF image to continue live while the PW image is frozen in triplex.
Assign PW Sample Volume control to rotary	Select to assign PW Sample Volume control to rotary.
CF Knob Changes Shear Gain	Enables Shear Wave Elastography gain control via the CF knob on the console. Check box to enable Shear gain control with CF knob (Off by default).
Default Rotation when changing mode	Sets the default rotation when changing mode.
Default MyPreset	Check the box to start MyPreset. Uncheck the box to start ConventionalExam tab (Default).

Table 10-27 • V Nav 3D Marker

Preset Parameter	Description
Inner Alpha	0-100
Margin Alpha	0-100
Color	Yellow, Orange, Red, Blue, Purple, Pink, or White
Margin Color	Yellow, Orange, Red, Blue, Purple, Pink, or White

Preset Parameter	Description
Diameter (mm)	1-100
Margin Dist. (mm)	0-15 in 0.5 mm increments
Short Axis	1-100
Long Axis	1-100
Reposition	Check to reposition.

Table 10-28 • EZ Settings

Preset Parameter	Description
EZ Touch Panel Page	Check to enable EZ Touch Panel on first page of B and flow mode tabs. Uncheck disable EZ Touch Panel (default).
MyPreset Shortcuts	Select “By Probe (Default)” or “By Category”.
Maintain icon usage	Always: The Symbol (Icon) button appears regardless of whether the EZ Touch Panel is On (Check) or off (uncheck). With EZ touch panel: If the EZ Touch Panel is On (Check), the Symbol (Icon) button appears. If the EZ Touch Panel is Off (Unchecked), the Text display button appears. Never: Regardless of whether the EZ Touch Panel is On (Check) or Off (uncheck), the Text display button appears.
B mode button	No function, Colorize, SRI HD, Reverse
Color mode button	No function, Map, Radiant Flow, Biopsy Guideline.
PDI mode button	No function, Map, Radiant Flow, Biopsy Guideline.
BFlow mode button	No function, Visualization, Background
MVI mode button	No function, Map, Radiant Flow, Biopsy Guideline.
PW mode button	No function, Simultaneous, Colorize, Wall Filter, Modify Auto Calcs, Trace Sensitivity, Quick Angle
CW mode button	No function, Simultaneous, Modify Auto Calcs, Trace Sensitivity, Trace Method, Map

System/System Measure Preset Menu

The System/System Measure screen allows you to specify measurement parameters such as the type of default OB measurements and calculations. You can also define cursor and Results Window default functionality.

Table 10-29 • Measurement

Preset Parameter	Description
Repeat Measurement	Select Repeat, No, DefaultMeasure Repeat = After you take a measurement, the system automatically starts the same measurement again. No = After you take a measurement, you have to touch a Touch Panel key or Trackball key to start another measurement. DefaultMeasure = After you take a measurement, the system automatically starts a default measurement based on the current scanning mode (B-Mode = basic length measurement, M-Mode = basic length measurement, Doppler Mode = velocity measurement except after a volume flow calculation).
OB Type	Select which OB measurements and calculations studies to use: USA, Europe, Tokyo, Osaka, or ASUM.
EFW GP	Select the source used to calculate EFW-GP Estimated Fetal Weight-Growth Percentile): Hadlock, Williams, Brenner, Kramer (f), Kramer (m), WHO.
CUA/AUA for Hadlock	Select to use CUA (Composite Ultrasound Age) or AUA (Average Ultrasound Age) as the default
Hadlock Table Type	Select Hadlock 82 or Hadlock 84 tables
EFW Formula (Europe)	Select the source used to calculate EFW (Europe) (Estimated Fetal Weight), Hadlock, Hansmann, Merz, Rich/Berk, Shep/Wars
EFW Formula (Tokyo)	Select the source used to calculate EFW (Tokyo) (Estimated Fetal Weight): Tokyo, Tokyo S-1, Tokyo S-2, Tokyo S-3.
Add 1 week to EDD	Select to add additional week to estimated date of delivery
OB Graph Display	Select Single or Quad for displaying OB Graphs.
OB Graph Single Display	Select Last Meas or EFW Single OB Graph displayed by default.
Fix Caliper by Print key	Select to use the Print key like the Set key. NOTE: If you select this during a generic volume measurement, the print key does not function like the Set key, but instead ends the measurement sequence and initiates the volume calculation based on the number of measurements taken so far.
LV Study using straight line	Sets straight line as the default for 2D LV studies.
Side selections of Rt, Lt and Off	Select to use "Rt, Lt and Off" for Side Selection. When not selected, displays only "Rt and Lt".
Dual Caliper on V Nav and Simultaneous.	Select to enable Dual Caliper on Volume Navigation and Simultaneous.
Map Cycle Select to Trackball Key	Map "AutoCalc cycle select" to Left/Right Set key.
Use WeekDay format for OB GA in DICOM SR	Select to use WeekDay format.

Preset Parameter	Description
Volume Method	Select "Single" to calculate one volume or "Multiple" to calculate multiple volumes when performing volume measurement. Single:  Multiple: 
Display AutoContor	Check to display trace line when AutoContor is fixed.

Table 10-30 • Worksheet (USA/ASUM)

Preset Parameter	Description
Show Individual Growth Percentiles	Check to display individual growth percentiles on the Worksheet.
OB Range Type	Selections: Min-Max, Standard Deviation.

Table 10-31 • Cursor

Preset Parameter	Description
Cursor Type	Select whether to mark measurements with numbers or symbols.
Cursor Size	Specify 12x12 or 9x9.
Cursor Line Display	If selected, after you press Set to complete a measurement, the cursor line is displayed. If not selected, after you press Set to complete a measurement, only the cursor number or symbol is displayed.
Cursor Ellipse Cross Line Display	Check box to display the cross line in Ellipse.
D Manual Trace Cross Line Display	Check box to display the cross line with the caliper.
Cursor Position	Select 1st Cursor, 2nd Cursor, or Image Center.
Color When Set (restart needed)	Select white, yellow, bright red, or orange.
Cursor is Displayed when Trackball is moved	The active cursor does not display until you move the Trackball. This assumes the following presets are set: Repeat Measurement, Repeat, Default Measurement, and Cursor.

Table 10-32 • Results Window

Preset Parameter	Description
Result Window Mode Depend	Select this if you want the measurement result window to be repositioned, depending on the mode.
Result Window Position X[0-800]	You can set the coordinates for the measurement result window when you do not have the result window set to be mode dependent. This is the X coordinate (left/right)

Preset Parameter	Description
Result Window Position Y[0-600]	You can set the coordinates for the measurement result window when you do not have the result window set to be mode dependent. This is the Y coordinate (up/down)
Result Window Location-2D	Select the Result Window location on the Monitor Display: Left-Bottom, Left-Top, Right-Bottom, Right-Top, Extreme Right-Top, or Extreme Right-Bottom.
Result Window Location-TimeLine	Select the Result Window location: Left-Bottom, Left-Top, Right-Bottom, Right-Top, Extreme Right-Top, or Extreme Right-Bottom.
Result Window Format	Select Wide or Narrow.
Font Color (restart needed)	Select White, Off White, Yellow, Bright Red or Orange (reboots the system)
Font Size (restart needed)	Select mini, small, medium, large, or extra large (reboots the system)

System Backup and Restore Preset Menu

Table 10-33 • Backup

Preset Parameter	Description
User Defined Configuration	Select to back up the user-defined configuration settings.
Service	Select to back up Service (InSite and Network) settings.
For Report templates, use Utility-> Report-> Export	
Backup	Select to begin the backup.

Table 10-34 • Backup To/Restore From

Preset Parameter	Description
Location	Select media type to use for Backup, Restore and Detailed Restore. CD/DVD, USB Drive F, Local Backup or Cloud.

Table 10-35 • EZBackup

Preset Parameter	Description
Reminder Dialog Interval days	Specify the number of days after the last backup that you want the system to prompt you to perform an EZBackup procedure (only for moving images).
Enable Reminder Dialog	Select to activate the EZBackup reminder pop-up dialog.
Media	Select media type.

Table 10-36 • Restore

Preset Parameter	Description
User Defined Configuration	Select to restore the user-defined configuration settings.
Service	Select to restore service InSite and Network settings.

Preset Parameter	Description
Restore	Select to begin the restore process for the selected configuration files from media storage backup.

The detailed section of this menu allows you to restore one area at a time from the user defined configuration. This allows you to selectively restore what you want to restore across multiple machines. Check the box(es) you want to restore, insert the appropriate media, and press Restore.

Table 10-37 • Detailed Restore of User Defined

Preset Parameter	Description
Imaging Presets	Select to restore imaging presets.
Connectivity Configuration	Select to restore connectivity configurations.
Measurement Configuration	Select to restore measurement configurations.
Comment/Body Pattern Libraries	Select to restore comment and body pattern configurations.
Protocol Templates	Select to restore protocol (Scan Assistant) templates.
Report Templates (Same Software Version Only)	Select to restore Report templates.
3D/4D	Select to restore 3D/4D settings.
Fast Key	Select to restore Fast Key.
Utility >Application Presets	Select to restore Utility > Application presets.
Custom Scan Assistant Programs	Select to restore Scan Assistant programs.
All Others	Select to restore all other configurations not listed in the Detailed Restore section. This includes parameters defined on the System preset menus.
Detailed Restore	Select to begin the restore process for the selected configuration files from media backup.

Table 10-38 • Local Backup or Local and Cloud Backup

Preset Parameter	Description
Backup Automatically	Select to automatically backup user-defined configuration settings to the system hard drive after a new setting or configuration is saved on the Utilities pages.
Upload for Fleet	Upload preset to Cloud for Fleet share. (Device Mgmt must be activated for Cloud Backup and Upload for Fleet button to appear.)

System/Peripherals Preset Menu

The System/Peripherals screen allows you to specify video and system setup parameters.

Table 10-39 • DVR

Preset Parameter	Description
Media (restart needed)	Select recording media: DVD or USB storage.
DVD Format (restart needed)	Select the video format: NTSC or PAL.
Picture Quality (reboot required)	SP, HQ, EP, LP. Extended Play or Long Play.
DVD Chapter Record Interval (sec.)	Select the Interval of automatic chaptering for DVD recording from 15, 30, 60 and 120 seconds.
USB Playback Skip Interval (sec.)	Select the Interval of time skipping for USB playback from 15, 30, 60 and 120 seconds.

Table 10-40 • S-Video

Preset Parameter	Description
S-Video Output Format (restart needed):	Select NTSC or PAL.

Table 10-41 • Standard Printer Properties

Preset Parameter	Description
Properties	Select to add an additional standard printer via the USB serial port and to configure digital printers. This activates the Windows Add Printer wizard. NOTE: Most printer drivers are available via Windows; however, newer printers may require you to load the manufacturer-supplied print driver (must be on CD/DVD). Refer to the Basic Service Manual for more information.

Table 10-42 • Standard Printer Restore

Preset Parameter	Description
Restore Defaults	Select to restore the selected printer to default settings.

Table 10-43 • Network Printer

Preset Parameter	Description
Add Network Printer	Select to add network printer.
Remove Network Printer	Select to remove network printer.

Table 10-44 • Default Printer

Preset Parameter	Description
Default Printer	Select to choose default printer.

Table 10-45 • Setup

Preset Parameter	Description
Print Full Screen	Select for the standard printer to print the full screen.
Enable Video Invert	Select for the standard printer to print black on white rather than white on black.

System/User Configurable Key

The User Configurable Key screen allows you to reconfigure the User Configurable Keys and configure Trackball Key functionality.

*Figure 10-3 • Programmable Operator Panel Controls***Table 10-46 • User Configurable Keys**

Preset Parameter	Description
1 User Defined Key	Select desired function to each User Defined key.
2 - User Defined Trackball Set Key	Specify which controls you want to be used on the Trackball: B/B-Flow/Contrast: Auto, Frequency, or Auto Preset Color Flow/TVI/PDI: Auto, Steer, PRF, or Auto Preset PW, CW, TVD: Auto, Baseline, or PRF Use Set keys to change BSteer Angle (On or Off)
3 - Replaceable Key caps	Replaceable key caps are provided should you choose to reconfigure the User Configurable Keys.
Keyboard Key 0 - 9 (Not pictured)	Check "Enable" in Keyboard key menu. Select desired function to each Keyboard key.



Figure 10-4 • User Configurable Key Preset Menu

To reconfigure User Defined key (UD), numeric Keyboard Keys, or Trackball Keys,

1. Press **Utility** > **System** > **User Configurable Key**.
To configure the User Defined Keys, select the function for each key (refer to *Figure 10-3 • Programmable Operator Panel Controls* on page 470). There are 10 User Defined Keys.
2. To configure the numeric Keyboard Keys, select the function for each Keyboard key (refer to *Figure 10-4 • User Configurable Key Preset Menu* on page 471). There are ten (10) Keyboard keys that are configurable.
3. To configure the Trackball Keys, select the function for each user-defined Trackball Set Keys by mode.
4. Press **Save**.

Replace User Defined Key caps

To move and replace a User Defined key cap,

1. Insert the flat blade of a screwdriver into the hole on the top side of the key cap and lift it up to remove it.



2. Repeat for each key, to match how you have configured the User Defined keys.

3. Replace the key caps by positioning it and pushing down on the key cap until it clicks into place.

System/About Preset Menu

The System/About screen lists information about the system software.

Table 10-47 • Installed Software Packages

Preset Parameter	Description
Application Software	The version, revision and part number of the current software on this system.
Online Help	The current version and part number of the Online Help on this system.
Application Service Patch	The current version and part number of the Service Platform on this system.
Security Package	The current version and part number of the Security Package on this system.
Base Image	The current version and part number of the Base Image on this system.
Base Image Update	The current version and part number of the Base Image Update on this system.

Table 10-48 • Copyright

Preset Parameter	Description
Copyright	Lists the copyright for the system.

Table 10-49 • Patents

Preset Parameter	Description
Patents	Lists system patent information.

Licenses

The Licenses screen lists information about the software licensed for use on the system.

Table 10-50 • Software

Preset Parameter	Description
License Titles	Scroll to select the License; the license is displayed in the License window.

Imaging Presets

Imaging screens allow you to specify parameters by Preset and Probe by Mode. For information about the specific parameters, refer to Chapter 5 Optimizing the Image.

- B-Mode (B)
- Color Flow Mode (CF)
- Power Doppler Imaging (PDI)
- Micro Vascular Imaging (MVI)
- Elastography (ELASTO)
- UGAP
- M-Mode (M)
- Anatomical M-Mode (AMM)
- Pulse Wave Mode (PW)
- Continuous Wave Mode (CW)
- Harmonics (HAR)
- B-Flow (BF)
- Contrast Reference (Ref)
- Contrast (CON)
- Tissue Velocity Imaging (TVI)
- Tissue Velocity Doppler (TVD)
- General



Figure 10-5 • Imaging Preset Example

1. Preset/Application-dependent setup parameters.
2. Probe-dependent setup parameters.

Changing imaging presets

To change imaging presets:

1. On the Touch Panel, select **Utility**.
2. On the Touch Panel, select **Imaging**.
The system displays the Imaging screens.
3. In the row across the top of the screen, select the mode.

The system displays two sets of parameters and settings. The left column lists all settings for the exam (for example, Abdomen). The right column(s) list settings that apply only to the exam and probe combination.

4. In the Preset list, select the exam.
5. In the Probe list, select the probe.
6. To change a parameter, do one of the following:
 - Select the value from a list
 - Select one value from a choice of two or more buttons
 - Select or clear a check box
7. After changing the parameters, to save the changes, select the Save button.

NOTE

When you Save changes to imaging parameters, the system saves changes to all modes, not just the mode currently displayed.

If you have problems with imaging, you can return parameters back to the original settings. Select the exam, probe, and mode, and then select Reload Factory Defaults. The system returns the selected parameters to the original settings.

For information about the specific parameters, refer to Chapter 5 Optimizing the Image.

General

You can specify a default probe per application and a default application per probe, EGC Display or Sync Mode.

Default probe per application

1. To specify a default probe per application, select **Utility > Imaging > General**.
2. Check the parameter if you want to start it automatically.
3. Select the default probe from the pull-down menu.

Default mode and application per probe

1. To specify a default application per probe, select **Utility > Imaging > General**.
2. Under Probe, specify the desired mode and application from the pull-down menu.

Checkmark the following fields when you want the system to activate a certain display. Values vary by probe.

- Simultaneous
- Automatically Retain Field of View

If selected, the system automatically activates the Retain Field of View control when a field of view setting is changed. If not selected, you can activate the Retain Field of View manually from the Probe Touch Panel.

- Application Default Mode
 - B-Mode

- Harmonic
- Enable B Steer +
- Default PDI:
 - PDI
 - MVI
- Default Elasto
 - Shear
 - Strain
- ECG
 - ECG Display
 - Sync Mode
- PDI/TVI button
 - PDI
 - TVI
- BF/CHI button
 - BF
 - CHI

Comments Libraries Presets

Comment screens allow you to specify comment text and pointer options, to define comment libraries, and assign comment libraries to applications.

Comments Libraries/Libraries Preset Menu

On the comments Libraries tab, you can change and create comment libraries. A comment library is a list of comments that are associated with a specific application. The comments are listed in the library in the order in which they display on the Touch Panel. For each library, you can define two Touch Panel displays of comments (Page1 and Page2), with 30 comments on each Touch Panel. Home and Arrow controls can also be added to the Touch Panel displays.

You can configure these comment color groupings via **Utility > Comments > Libraries** (up to 5 groups).



Figure 10-6 • Comment Libraries Preset Menu

Table 10-51 • Libraries

Preset Parameter	Description
Library	The name of the comment library.
Reload Factory Defaults	Select to reload factory defaults.
Small List	Fields where you define a small list.
User Defined Library	The name of a new comment library that you want to create/delete.
Copy from Existing	You can add to or delete from the selection of comments.

Defining Comments

1. In the Library field, select the library you want.

The system displays all comments for the library. You can have two Touch Panel displays of comments for each library. The comments are listed in the order that they are shown on the Touch Panel when you use comments.

2. To change or add an comment, select the comment or blank location and press **Set**, then do one of the following:
 - Type the comment.
 - Select the comment in the Copy from Existing list, and press **Set**.
3. To save the changes, select the **Save** button.

Creating a new comments library

1. In the User Defined Library field, type a name for the library, then select **Create**.
The system creates a new library.
2. Enter comments as described in step 2 above.
3. To save the changes, select the **Save** button.

Deleting a user defined library

1. Select the library name which you want to delete from the pull-down menu.
2. Press **Delete**.
3. Press **Save** to save the changes.

Creating a small list

A small list is a list of up to three comments attached to one comment location on the Touch Panel. You can use a small list to group similar comments, such as those indicating a probe location. For example, you can specify that a small list include the following comments: Long, Transverse, and Coronal. To make comments easier to use, you can define the small list in the same location in each comment library.

To define a small list:

1. Move the **Trackball** to the comment field on Page1 or Page2 where you want to create a small list, and press **Set**.
2. Move the **Trackball** to the first field in the Small List section, and press **Set**.
3. To enter comments in the fields in the Small List section, select the field and press **Set**, then do one of the following:
 - Type the comment
 - Select the comment in the Copy from Existing list, and press **Set** twice.
You can enter up to three comments. When you enter an comment in the first field of the Small List section, the selected comment field on Page1 or Page2 changes to SMALL LIST.
4. To save the changes, select the **Save** button.

NOTE

The small list can be displayed as a pop-up window or as a toggle field. The Small List Operation field on the General tab allows you to specify how it is displayed.

Comments Libraries/Comments Preset Menu

On the Comments tab you can specify text and pointer options.

Table 10-52 • Text

Preset Parameter	Description
Text Font Size	Specify the font size. The font size increases as the number increases.
Text color (Text1 and Text2)	Select the color for comment Text1 and Text2.
Arrow Color	Select the color for comment Arrow.
Text Boundary	Select Group Move or Word Wrapping.
Small List Operation	Select whether you want small list options to display in a Pop-up window, by a Toggle function, or in a Pop-up with replace.
Enable Type Over Mode	Select to type over existing comments. Position the cursor over the text to be changed, then start typing.
Reset Small List	Select to indicate that small lists should not be reset to the first item.
Automatically Set Text	Sets the comment as you are typing it.
Replace Mode	Select to replace comment.

Table 10-53 • Arrow

Preset Parameter	Description
Arrow Length	Select the default pointer length.
Arrow Size	Select the default pointer size.
Keep Arrow Angle	Keep the angle of arrow pointer head until next change.
Rotation Angle	Set Rotation Angle of the arrow: 8 directions, 16 Directions, or Fine.

Table 10-54 • General

Preset Parameter	Description
Retain while entering or leaving timeline mode	If selected, the system keeps the comment(s) on the monitor display when you enter or leave timeline mode.
TextOverlay in Multiple Image	When selected, and you select the F8 key to hide or show comments, if you are in multiple image, the system hides the text in both images. When cleared, the system only hides the text for the active image.
TextOverlay Sequence	You can specify to display Text1, Text2, or both. This allows you to have some comments that do not change during the exam while allowing you to change other comments. Toggle the F8 key to cycle through the 3 Text1/Text12 states.

Preset Parameter	Description
Erase when the probe or application is changed	Deletes annotations when you change the application or probe.
Clear Non Active Image Comments	Select if you want comments to be removed from the non-active multi image.
Map Arrow to top trackball key	Check to map Arrow to top trackball key.

After you change comment options, select **Save** to save the changes.

Comments Libraries/Applications Preset Menu

The Comments Libraries/Applications tab is a link to the Applications preset menu. The Applications preset screen allows you to specify which libraries belong to an application. You also specify which is the default library that displays when you use comments.

The Applications/Comments screen can be accessed through either the Comments Libraries or Applications Touch Panel key.

Specifying which libraries belong to an application

1. On the Applications tab, in the Application field, select the application.
2. In the Library Group Tabs fields, select the libraries for this application. You can select up to six libraries.
3. In the Default Library Group field, select the default library you want the system to display when you use comments.

NOTE

When you use comments, the default library is displayed. To use other libraries for the application, press the tab for the library.

4. To save the changes, select the Save button.

Table 10-55 • Applications

Preset Parameter	Description
Preset	The name of the application preset.
Tabs	A list of libraries for the application. You can select up to six libraries.
Default Tab	The default library that the system displays when you use comments.

Using comments from a library

To use comments, press the **Comment** key on the Control Panel. Comments are then displayed on the Touch Panel.

To select a comment library, press the appropriate tab (e.g. the tabs are OB23 and OB23_1).



Figure 10-7 • OB 2/3 Comments Touch Panel

Comments Libraries/Mapping Preset Menu

The system uses the annotation/body pattern information associated with the image to automatically assign the segment/position qualifier of the breast lesion.

On the Mapping tab, you add/delete/reset the user-defined mapping for the qualifier.




Figure 10-8 • Comment Mapping

Table 10-56 • Mapping preset

Preset Parameter	Description
Annotation to Location Conversion	Add/Delete/Reset the use-defined mapping annotation for the qualifier for Breast. - Add: Add the use-defined mapping annotation. - Delete: Delete selected user-defined mapping. - Reset: Reload factory default mappings to the selected qualifier.
Use Auto Positioning	Configure a mapping of annotation and body pattern to location (Position and Segment) mappings. Note: This is effective only against the following body pattern: Breast4 Lt/Rt, Breast5 Lt/Rt, Breast6 Lt/Rt, Breast7 Lt/Rt, Breast8 Lt/Rt

After you change Mapping options, select Save to save the changes.

For example,

1. Select Qualifier "12 O'Clock" from the pull-down menu.
2. Type the mapping annotation "12" and press Add.
3. Check "Use Auto Positioning".




Figure 10-9 • Comment Mapping

4. In **Utility > Body Pattern > Libraries**, select SMLP from pull-down menu.
5. Select an empty cell of the Library table.
6. Select Breast4_Rt in Copy from Existing to add.



Figure 10-10 • Comment Mapping - SMLP

7. Press **Save and Exit**.
8. Scan the patient.
9. Type "Right Breast 12 Zone 1" as Comment on the image.
10. Activate Breast measurement and select Lesion folder from the Touch Panel for add Lesion1.

Position Rotary is set 12 O'Clock, Segment Rotary is set A, and "12 O'Clock A" displays as measurement name automatically.




Figure 10-11 • Comment Mapping - Measurement

NOTE

The system assigns "-" to position qualifier if the value is not found in the annotation or body pattern information associated with the image.

11. Select the frame for Lesion 2.
12. Select Breast4_Rt as Body pattern on the image and locate the probe mark in the appropriate position.
13. Activate Breast measurement and select Lesion folder from the Touch Panel.
14. Position Rotary and Segment Rotary is set automatically associated with the position and segment of the probe mark.

Body Patterns Presets

Body patterns screens allow you to specify body pattern options, to define body pattern libraries, and assign body pattern libraries.

Body Pattern Libraries/Libraries Preset Menu

On the Body Patterns Libraries tab, you can change and create body pattern libraries. A body pattern library is a list of body patterns that are associated with a specific application. The body patterns are listed in the library in the order in which they display on the Touch Panel. For each library, you can define two Touch Panel displays of body patterns (Page1 and Page2), with 15 body patterns on each Touch Panel.



Figure 10-12 • Body Patterns Libraries Preset Menu

Table 10-57 • Body Patterns Libraries

Preset Parameter	Description
Library	The name of the body pattern application library.
Reload Factory Defaults	Select to reload factory defaults.
Body Pattern Image	Displays the image of the currently selected body pattern.
User Defined Libraries-Create	The name of a new body pattern application library that you want to create.
User Defined Libraries-Delete	Allows the selection of the user defined library to be deleted.
Copy from Existing	A list of body patterns you can use to create an application library.

Defining body patterns

1. In the Library field, select the application library you want.

The system displays all body patterns for the library. You can have two Touch Panel displays of body patterns for each library. The body patterns are listed in the order that they are shown on the Touch Panel.

2. To change or add a body pattern, select the body pattern or blank location and press **Set**, then do one of the following:

- Type the body pattern name.
- Select the body pattern in the Copy from Existing list, and press **Set**.

When you select a body pattern name in a Touch Panel location or in the Copy from Existing list, the system displays the pattern in the lower left corner of the screen.

3. To save the changes, select the **Save** button.

Creating a new body pattern library

1. In the User Defined Libraries field, type a name for the library, then select Create.

The system creates a new library.

2. Enter body patterns as described in step 2 above.
3. To save the changes, select the **Save** button.

Body Pattern Libraries/Body Patterns Preset Menu

Table 10-58 • Body Patterns

Preset Parameter	Description
Erase When the probe or application is changed	If checked, when you change probes or applications, the system erases the body pattern.
Erase When the image is unfrozen	If checked, when you unfreeze the image, the system erases the body pattern.
Copy to active side in multiple image	If checked, when you use dual B-Mode, the system copies the body pattern to the active side of the dual image.
Body pattern background	Select whether you want the body pattern background to be Transparent or Opaque.
Use Zoom Rotary knob to select Body pattern	If selected, you can scroll through the body patterns with the Zoom control.
Body Pattern knob Recall On	When the image is recalled, Body pattern knob works as below: • Body Pattern on: Activate Body Pattern by press or move Up/Down/Left/Right the Body Pattern knob. • U/D: Body Pattern On, L/R: Prev/Next image: Move Body Pattern knob Up and Down to Body Pattern On, Move left recalls previous image, move right recalls next image. • Prev/Next image: Move up and left recall previous image, move down and right recall next image. • None: Activate Body Pattern by press the Body Pattern knob.

Preset Parameter	Description
Body Pattern knob Recall Off	When the image is not recalled, Body pattern knob works as below: - Body Pattern on: Activate Body Pattern by press or move Up/Down/Left/Right the Body Pattern knob. - U/D: Body Pattern On, L/R: Prev/Next image: Move Body Pattern knob Up and Down to Body Pattern On, Move left recalls previous image, move right recalls next image. - Prev/Next image: Move up and left recall previous image, move down and right recall next image. - None: Activate Body Pattern by press the Body Pattern knob.
Preset Parameter: Body Pattern knob: Scan Assistant On	Select None or Scan Assistant Control for Body Pattern knob. Description: When Scan Assistant is activated, - Scan Assistant Control: Navigate Scan Assistant protocol by move Up/Down the Body Pattern knob. Pause/Resume Scan Assistant by move Left/Right the Body Pattern Knob. - None: Activate Body Pattern by press the Body Pattern knob.

After you change body pattern options, select Save to save the changes.

Body Pattern Libraries/Applications Preset Menu

The Body Patterns Library/Applications tab is a link to the Applications preset menu. The Body Patterns Applications tab allows you to select body pattern application libraries. You also specify which is the default library that displays when you use body patterns.

The Applications/Body Patterns screen can be accessed through either the Body Pattern Libraries or Applications Touch Panel keys.



Figure 10-13 • Body Patterns Applications Preset Menu

Table 10-59 • Applications

Preset Parameter	Description
Preset	Defines the Body Pattern option.
Tabs	A list of body pattern applications.
Default Tab	The default library that the system displays when you use body patterns.

Selecting body pattern application libraries

1. On the Applications tab, in the Application field, select the body pattern.

2. In the Library Group Tabs fields, select the application libraries for Body Patterns. You can select up to six libraries.
3. In the Default Library Group field, select the default application library you want the system to display when you use body patterns.

NOTE

When you use body patterns, the default library is displayed. To use other application libraries, press the tab for the library.

4. To save the changes, select the **Save** button.

Using body pattern application libraries

See the following Body Patterns Small Parts Touch Panel.



Figure 10-14 • Body Patterns Small Parts Touch Panel

To select a body pattern library, select the tabs (for example, ABD or OB).

To select body patterns, use the ***Ellipse/Body Pattern*** control on the Control Panel.

Application Presets

Application Presets allow you to configure the application-specific settings (presets).

Settings

Table 10-60 • Preset

Preset Parameter	Description
Preset	Select the application that you want to specify the presets. Along with the various applications available on the system, there are four user-defined application presets that can be set.

Table 10-61 • Image Control and Display

Preset Parameter	Description
Show kHz scale	When selected, displays the kHz scale on the left side of the Doppler spectrum.
Show Doppler Rate	When selected, displays the Doppler rate (mm/s) below the Doppler spectrum.
Anatomical Angle Correction	Select to keep the angle constant with regard to the anatomy.
Join Dual Image for Linear	Select to place linear probe dual images directly next to each other.
Hide Mode Cursor Key	Select to unmap (hide) the Mode Cursor key, which normally appears on the left Trackball key during live scanning in B-Mode or Color Flow Modes.
Horizontal Display for Biplane	Display of the two planes of the Biplane probe in a top and bottom layout.

Table 10-62 • Auto Zoom Linear Probe Images at Shallow Depth...

Preset Parameter	Description
Check Desired Setting	Single Screen Dual Screen and DualView Virtual Convex BSteer+

Table 10-63 • When Entering Dual Image...

Preset Parameter	Description
Duplicate Frozen Image to Opposite Side	When entering Dual Image, duplicate the frozen image to the opposite side.
Duplicate Live Image to Opposite Side	When entering Dual Image, duplicate the live image to the opposite side.

Table 10-64 • Patient Info

Preset Parameter	Description
Titlebar Line 1	Select the patient information to display on the scanning screen Title bar.
Titlebar Line 2	Select the patient information to display on the scanning screen Title bar.
Titlebar Line 3	Select the patient information to display on the scanning screen Title bar.

Table 10-65 • Comments

Preset Parameter	Description
Active function at Freeze	Select None, Body Pattern, or Comments. If Body Pattern or Comment is selected, the Body Pattern or Comment is activated automatically when freezing the system.
Erase when the image is unfrozen	Select to erase image when system Freeze is deactivated.

Table 10-66 • Footswitch

Preset Parameter	Description
Left, Middle, Right	Specify from the following for each footswitch pedal: No Function, Record/Pause, Freeze, Next Heartcycle, Previous Heartcycle, Print 1,2,3,4, Update, Next Step (Scan Assistant), Previous Step (Scan Assistant), Scan Assistant Pause/Resume, or Mark Cine.

Table 10-67 • Protocol

Preset Parameter	Description
Show Protocol Tab	Check to display the Protocol Tab on the Touch Panel.
Template	Select the default stress echo template: Bicycle Normal, Bicycle Sporty, Contrast Pharmacological, Pharmacological 4x4, Pharmacological 8x5, Exercise 2x4, Exercise 2x4 B, Pharmacological US 4x4

Table 10-68 • ECG

Preset Parameter	Description
Show ECG Tab	Check to display the ECG Tab on the Touch Panel.
ECG Lead	Specify the number of ECG Leads (1, 2, or 3).
Resp Lead	Specify the number of Respiration Leads (1 or 2).

Table 10-69 • ELASTO

Preset Parameter	Description
Show Quality Bar	Check to display a Quality Bar for Elastography. The more bars, the better the quality. As the quality increases, the bars go from red, to yellow, to green.
Show Quality Graph (restart needed)	Select to display a Quality Graph for Elastography. The higher the level, the higher the data quality for the frames.

Table 10-70 • User Label

Preset Parameter	Description
Show Label	If selected, the system displays the User label in the Preview Window at the bottom left portion of the monitor.
Label 1-8	User Labels have eight (8) text lines. Each User Label is limited to 50 characters.

Print Controls

Table 10-71 • Live Store

Preset Parameter	Description
P1, P2, P3, P4, PrintScreen	Select Retrospective Clip (CINE prior to pressing Print), Prospective Clip (CINE after pressing Print), Single Image or None to store from a live image.

Press the “Connectivity/Button/Physical Print Buttons” hyperlink to go directly to the Connectivity Print Button Setup page.

Table 10-72 • Time-Base Store

Preset Parameter	Description
Time span (s)	Select the number of seconds of CINE Loop storage. The default is 3 seconds.
3D/4D Time span (s)	Select the number of seconds of CINE Loop storage while in 3D/4D.
Contrast Time span (s)	Select the number of seconds of CINE Loop storage while in Contrast.
Preview clip before store	When selected, allows you to review cine loops before storage.
Segment prospective loops longer than time span	Specify length of Clip (15, 30, or 45 seconds, Off, or Max).

Table 10-73 • ECG-Based Store

Preset Parameter	Description
Time before heart cycle [ms]	Sets the storage time span before R-wave of the first heart cycle.
Time after heart cycle [ms]	Sets the storage time span after R-wave of the last heart cycle.
Number of heart cycles	Select the number of heart cycles to store. (Must be de-selected for single frame.)
Preview cine clip before store	When selected, allows you to review Cine loops before storage.

Table 10-74 • Mark Cine

Preset Parameter	Description
Enable Mark Cine Control	Lets you mark where you want the Cine Loop to start (prospective CINE).

Preset Parameter	Description
Preview Loop Longer than... (s)	When selected, allows you to review Cine loops before storage for loops longer than selected timeframe (in seconds).

Imaging Controls

You can select which controls you want to be available via the Touch Panel during a clinical scan. When you select **Preset > Application and Control Mode > Clinical**, deselect the controls you DO NOT want to appear while scanning in this clinical application.

NOTE

If you select Research, all controls appear.



Figure 10-15 • Imaging Controls, Research

Clinical vs Research Touch Panel Example. To view all controls press the View All/View Less control (circled).



Figure 10-16 • Clinical Control Mode




Figure 10-17 • Research Control Mode

Comments and Body Patterns

Comments and Body Patterns were described earlier in this chapter.

Measurements

You can set the exam category measurement and calculation package you want to appear when you select the exam category Preset.

User Specific

Refer to “Monitor Display” in Chapter 3 for more information.

Test Patterns

For more information, see Test Patterns in the Wide Monitor section of Chapter 3.

3D/4D

3D/4D presets allow you to set up application-specific settings (presets) for each 4D image acquisition type. You can define different application-specific settings for each probe. Refer to Chapter 5 for more information.

4D Presets

To set up 4D presets:

1. On the Touch Panel, select **Utility**.
2. On the Touch Panel, select **3D/4D**.
The system displays the 4D Presets screen.
3. To select a probe, click on the plus sign (+) that appears next to the desired probe.
4. To select the application, click on the plus sign (+) that appears next to the desired application.
5. To select the acquisition type, click on the plus sign (+) that appears next to the desired application.
6. Double-click the desired application under the acquisition type. The Display tab is selected.

Display Presets Tab

Table 10-75 • Image Display

Preset Parameter	Description
Tile	Determines the number of display windows. Values include: 1 (Single), 2 (Dual), and 4 (Quad).
Visualization	Determines the method of display for working with images. Selections available: Sectional, Render, VOCAL, VCI static and TUI.
3D Orientation (degrees)	Determines the orientation of the ROI on the monitor display. Values include: 0, 90, 180, 270.
Zoom Factor	Determines the magnification factor of the zoom. Values include: 0.3 through 4.0, in .01 increments.
Orientation Help	Activate Orientation Help.
Gain	Set the desired Gain.

Table 10-76 • Pre-mode ROI

Preset Parameter	Description
ROI Center (cm)	Determines the vertical center of the region of interest. Values vary by probe.
ROI Span (cm)	Determines the height of the region of interest. Values vary by probe.

Preset Parameter	Description
Tilt (degrees)	Determines the degree of tilt from the vertical center location of the ROI. Values vary by probe.
Width (degrees)	Determines the width of the ROI. Values vary by probe.
Volume Angle	Set the range of the volume sweep. Values vary by probe. Listed in degrees for curved probes, cm for linear probes.

Table 10-77 • Quality Setting

Preset Parameter	Description
Setting	Set quality setting -- balances speed with line density. Selections are Low, Mid1, Mid2, Hi1, Hi2, Max. High combines the highest density with the slowest speed. Low combines the lowest density with the highest speed.
CF Setting	Set quality setting -- balances speed with line density. Selections are Low, Mid1, Mid2, Hi1, Hi2. High combines the highest density with the slowest speed. Low combines the lowest density with the highest speed.

Render Tab

Table 10-78 • Render and Render for VCI Static

Preset Parameter	Description
Render Mode 1:Mode2 (Gray Inversion, and VCI Static)	Set render mode values. Surface Smooth, Surface Texture, Transp Max, Transp X-Ray, TransMin (Render 1), or HDlive Texture Surface Smooth, Light, Gradient Light, Transp Max, Transp X-Ray, Transp Min (Render 2), or HDlive Smooth.
Mix/Mix Inversion/eMixVCI Static (% Render Mode 2)	Set mix of Render 1 / Render 2 Mode, 0-100.
Lower Threshold	Set a lower threshold below which weaker echoes are removed, 0-255.
Transparency/ Transparency (Inversion)	Set the transparency of the image, 10/20-250. The higher the number, the more transparent the gray scale information.
Render Direction	Set the direction in which the ROI is viewed.
SonoRenderLive	Check to use SonoRenderLive.
SonoRenderLive Sensitivity	Set the SonoRenderLive sensitivity from 1-100.
[VCI Static] Slice Thickness	Set slice thickness, 2-20.

Color / PDI Render tab

Table 10-79 • Color / PDI Render

Preset Parameter	Description
Render Mode 1 (Color) / Render Mode 2 (Color)	Determine the render mode, selected from Render Mode 1 and Render Mode 2.
Mix (Color)	Set the percentage of Render Mode 1 to be mixed with Render Mode 2.
Lower Threshold (Color)	Set the lower threshold below which weaker echoes are removed.

Preset Parameter	Description
Transparency (Color)	Determine the transparency of the image. The higher the number, the more transparent the gray scale information. Values: 20 to 255.
Render Gray : Color)	Determine the render mode, selected from Render Mode 1 and Render Mode 2.
Mix (Gray Color)	Set the percentage of Render Mode 1 to be mixed with Render Mode 2.

VOCAL tab

Table 10-80 • VOCAL

Preset Parameter	Description
Vocal Method	Set Sphere, Manual, Contour Detect, or Semi-Auto Contour Detect.
Vocal Semi-Auto-Detect type	Set Hypo, Cystic, or Hyper/Iso.
Vocal Rotation step	Set 6, 9, 15, or 30.

TUI tab

Table 10-81 • TUI

Preset Parameter	Description
Display Format	Set 1x1, 1x2, 2x2, or 3x3.
Total Slices	Set 3, 5, 7, 9, 11, 13, 15, 17, or 19.
Slice Distance (mm)	Set 0.5-40 (0.1 step increments).

OmniView tab

Table 10-82 • OmniView

Preset Parameter	Description
Show OmniView View Direction Marker	Select to view the direction marker in OmniView.

Advanced tab

Table 10-83 • Advanced

Preset Parameter	Description
Upper Threshold	Sets the higher threshold above which weaker echoes are removed.
Volume Calibration Shift	See a Field Service Engineer for information on this parameter.

Configuring Connectivity

You use Connectivity functionality to set up the connection and communication protocols for the ultrasound system. The following page gives an overview of each of the Connectivity functions. Each function is described in detail in the following pages.

Connection Manager

Connection Manager allows configuration of all connectivity functions on the Ultrasound system. Select Connectivity on the Utility Touch Panel page to launch Connection Manager.

If Connectivity was not previously set up on the Ultrasound system, the Scanner page appears when Connectivity is selected.

It is recommended to initially set up each area of Connectivity in the sequence they are listed across the top of the Connection Manager page, from left to right.

Navigate through Connection Manager from left to right by selecting the icons at the top of the screen or by selecting Next or Previous at the bottom of the screen. The recommended setup sequence is:

- *Scanner Page* on page 497
- *Network Page* on page 499
- *MyComputer Device Page* on page 505
- *Dicom Page* on page 512
- *Special Devices Page* on page 522
- *Print Button Page* on page 542
- *Dataflow Page* on page 545
- *Advanced Settings Page* on page 546

If Connection Manager has previously been set up, the Summary page appears when Connectivity is selected.

Navigate to the desired Connection Manager configuration screens by selecting the icons at the top of the screen or by selecting Next or Previous at the bottom of the screen to advance to through the screens.

Scanner Page

Use the Scanner page to configure or modify details about the DICOM properties used for the Ultrasound system.




Figure 10-18 • Scanner Page

This table shows all the preset parameters available on Scanner with descriptions.

Table 10-84 • Scanner Page

Preset Parameter	Description
Computer Name	Enter a name for the Ultrasound system. This may be the same as the station name.
Scanner AE Title	Assign an AE Title to the Ultrasound system. (AE stands for Application Entity. DICOM services use this to identify the Ultrasound system.) AE Title is case-sensitive. This title may contain the Computer Name. Maximum number of characters in AE Title is 16 characters. It is NOT recommended to use the factory default. This is not prohibited, but more than one system with the same AE Title can cause confusion
Scanner Port	If needed, edit the default port number of 104. Restricted port numbers are listed in the user manual.
Allow Trusted (Configured) Devices Only	To prevent unsolicited inbound DICOM conversations, enable Allow Trusted (Configured) Devices Only. When enabled, the Ultrasound system only responds to TCP/IP ping or DICOM Echo from a defined DICOM source IP address. On = Enabled/Off - Disabled
Use DICOM TLS Encrypted Connection	To allow DICOM connectivity to a secure network connection using encryption and certificates using the TLS (transport layer security) protocol, enable Use DICOM TLS Encrypted Connection. Internal communication is still DICOM-structured. On = Allowed/Off = Not allowed

Preset Parameter	Description
Ping ICMPV4	To allow the Ultrasound system to respond to a network ping or traceroute command from another device on the network. Disabled = Ultrasound system will not respond to incoming IP ping requests from any remote device LocalSubnet = Ultrasound system responds to all incoming IP ping requests from any remote device that belongs to same local subnet. Any = Ultrasound system responds to all incoming IP ping requests from any remote device.
Discard Scanner Changes	Select to remove any changes made to Scanner.

Configure Scanner

1. Navigate to **Connectivity > Connection Manager > Scanner**.
2. In **Computer Name**, enter a name for the Ultrasound system.
3. In **Scanner AE Title**, enter an AE Title for the Ultrasound system.
4. In **Scanner Port**, enter a port number for the Ultrasound system. The default port number is 104.
5. To allow DICOM connectivity to a secure network connection using encryption and certificates using the TLS (transport layer security) protocol, enable **Use DICOM TLS Encrypted Connection**.
6. To prevent unsolicited inbound DICOM conversations, enable **Allow Trusted (Configured) Devices Only**.
7. Under **Ping ICMPV4**, select to allow the Ultrasound system to respond to a network ping or traceroute command from another device on the network. Select Disabled if you do not want the Ultrasound system to respond to incoming IP ping requests from any remote device.

Network Page

Use the Network page to configure or modify network properties for the Ultrasound system.



Figure 10-19 • Network Page

Refer to the following sections for more information:

- *General* on page 500
- *Wired* on page 501
- *Wireless* on page 502
- *Proxy* on page 504

General

Use Scanner Network Properties - General to configure the protocol setting.




Figure 10-20 • Scanner Network Properties - General

This table shows all the preset parameters available on Scanner Network Properties - General with descriptions.

Table 10-85 • Scanner Network Properties - General

Preset Parameter	Description
Protocol	Select either the IPV4 or IPV6 static IP setting.

Add Scanner Network Properties - General

1. Navigate to **Connectivity > Connection Manager > Network**.
2. Select **General**.
3. Under **Protocol**, select either the IPV4 or IPV6 static IP setting.
4. Select **Done**.

Wired

Use Scanner Network Properties - Wired to configure the local area network settings.




Figure 10-21 • Scanner Network Properties - Wired

This table shows all the preset parameters available on Scanner Network Properties - Wired with descriptions.

Table 10-86 • Scanner Network Properties - Wired

Preset Parameter	Description
Automatic IP (DHCP)	Enable to automatically configure your wired network settings. On = Enabled/Off - Disabled
Network Speed	Select the network speed (Auto Detect, 10Mbps/Half/Full Duplex, or 100 Mbps/Half/Full Duplex, and 1000Mbps/Auto-negotiate).
Network adapter details	Displays the actual network configuration the Ultrasound system is currently using and has recognized.

Add Scanner Network Properties - Wired

1. Navigate to **Connectivity > Connection Manager > Network**.
2. Select **Wired**.
3. To automatically configure your wired network settings, enable **Automatic IP (DHCP)**.
4. Under **Network Speed**, select the network speed.
5. Select **Done**.

Wireless

Use Scanner Network Properties - Wireless to configure the wireless network settings.




Figure 10-22 • Scanner Network Properties - Wireless

This table shows all the preset parameters available on Scanner Network Properties - Wireless with descriptions.

Table 10-87 • Scanner Network Properties - Wireless

Preset Parameter	Description
Wireless Network Name	Name used for the wireless network.
IP Address	IP Address of the Ultrasound system. IP stands for Internet Protocol. Every device on the network has a unique IP address.
Default Gateway	Default gateway address (optional).
Subnet Mask	IP address filter that eliminates communication/messages from network devices of no interest to your system.
Primary DNS Server	IP address for the primary DNS server (optional - at least one valid DNS address is needed for Insite remote service connectivity).
Secondary DNS Server	IP address for the secondary DNS server (optional). Do not configure Secondary DNS server only; if only one DNS IP address is being used, enter it in the Primary DNS Server field.
MAC Address	Unique network card address.
Speed (Mbps)	Actual network speed in Megabits per second.
Connected (min)	Number of minutes the Ultrasound system has been connected to the network.

Preset Parameter	Description
Connection Status	Current network status. - Operational: Network adapter has been disabled, for example because of an address conflict. - Unreachable: Network adapter that is not connected. - Disconnected: For LAN adapters: network cable disconnected. For WLAN adapters: no carrier. - Connecting: Network adapter that is in the process of connecting. - Connected: Network adapter that is connected to a remote peer.
Configure wireless network	Select to display the Wireless Network Configuration page.

Add Scanner Network Properties - Wireless

1. Navigate to **Connectivity > Connection Manager > Network**.
2. Select **Wireless**.
3. Select **Configure Wireless Network**.
4. On the Wireless Network Configuration dialog box, select the wireless connection and select **Connect**.
5. When prompted, enter the new wireless network properties.
6. Select **Done**.

Proxy

Use Scanner Network Properties - Proxy to configure a proxy server.




Figure 10-23 • Scanner Network Properties - Proxy

This table shows all the preset parameters available on Scanner Network Properties - Proxy with descriptions.

Table 10-88 • Scanner Network Properties - Proxy

Preset Parameter	Description
Use Proxy	Enable to use a proxy server. On = Enabled/Off = Disabled
Proxy Server	When Use Proxy is enabled, enter a proxy server address of the facility.
Proxy Port	When Use Proxy is enabled, enter a port number for the proxy server.
Proxy Credentials	Enable to use credentials for a proxy server. On = Enabled/Off = Disabled
Proxy User Name	When Proxy Credentials is enabled, enter an account user name for the proxy server.
Proxy Password	When Proxy Credentials is enabled, enter an account password for the proxy server.
Discard Proxy Changes	Select to remove any changes made to Scanner Network Properties - Proxy.

Add Scanner Network Properties - Proxy

1. Navigate to **Connectivity > Connection Manager > Network**.
2. Select **Proxy**.
3. To use a proxy server, enable **Use Proxy**.
4. When Use Proxy is enabled, in **Proxy Server**, enter a proxy server address of the facility.
5. When Use Proxy is enabled, in **Proxy Port**, enter a port number for the proxy server.
6. To use proxy credentials, enable **Use Proxy Credentials**.
7. When Use Proxy Credentials is enabled, in **Proxy User Name**, enter an account user name for the proxy server.
8. When Use Proxy Credentials is enabled, in **Proxy Password**, enter an account password for the proxy server.
9. Select **Done**.

MyComputer Device Page

Use the MyComputer Device page to configure or modify details about the Ultrasound system.




Figure 10-24 • MyComputer Device Page

Refer to the following sections for more information:

- *EzBackup* on page 506
- *Removable Media* on page 508
- *DICOM Export* on page 509
- *USB Quick Save* on page 510
- *HD Export* on page 511

EzBackup

Use EzBackup to manage hard disk space while maintaining the patient database on the Ultrasound system and back up the patient database and images by copying the data from the local drive to removable media.




Figure 10-25 • Edit EzBackup

This table shows all the preset parameters available on Edit EzBackup with descriptions.

Table 10-89 • Edit EzBackup

Preset Parameter	Description
Name	Logical name of the service.
Convert Clips	When enabled, the raw cine clips get converted to DICOM multi-frames during EzBackup operation. On = Enabled/Off = Disabled
Convert Volumes	When enabled, the raw volumes get converted to Enhanced US Volume (DICOM 3D) format during EzBackup operation. On = Enabled/Off = Disabled
Apply Display Precision to Measurements	When enabled, the measurement values are rounded to configured number of decimal places. On = Enabled/Off = Disabled
Dicom Coding Scheme	Measurements are encoded with configured coding scheme in DICOM SR objects.
Use Button Settings	When enabled, uses button settings. On = Enabled. The Ultrasound system will use the default image storage and compression presets./Off = Disabled. Properties menu expands and image storage and compression presets can be defined.

Add EzBackup

1. Navigate to **Connectivity > Connection Manager > MyComputer Device**.
2. Select **EzBackup > Configure**.
3. In **Name**, enter a logical name of the service.

4. To have raw cine clips converted to DICOM multi-frames during EzBackup operation, enable **Convert Clips**.
5. To have raw volumes get converted to Enhanced US Volume (DICOM 3D) format during EzBackup operation, enable **Convert Volumes**.
6. To round measurement values to configured number of decimal places, enable **Apply Display Precision to Measurements**.
7. Under **Dicom Coding Scheme**, select how measurements are encoded with configured coding scheme in DICOM SR objects.
8. To use button settings, enable **Use Button Settings**.
9. Select **Done**.

Removable Media

Use Removable Media to do the following:

- Verify the DICOM directory on removable media.
- Verify the free space of the media.
- Verify that the media is finalized or unfinalized.
- Verify that the media is formatted or unformatted.
- Format removable media (rewritable CD/DVD or USB device).



Figure 10-26 • Edit Removable Media

This table shows all the preset parameters available on Edit Removable Media with descriptions.

Table 10-90 • Edit Removable Media

Preset Parameter	Description
Removable Media	Select the removable media to format or verify.
Label	Type a label for a new removable media (free text).
SetLabel	When text has been entered under Label, select to set the text.
Format	Select to format the media. New media should always be formatted.
Media Properties	Shows properties for the configured media.

Preset Parameter	Description
Done	Select to save the removable media settings.

Add Removable Media

1. Navigate to **Connectivity > Connection Manager > MyComputer Device**.
2. Select **Removable Media > Configure**.
3. Under **Removable Media**, select the removable media to format or verify.
4. Under **Label**, enter a label for a new removable media (free text).
5. If you entered a label, select **SetLabel**.
6. Select **Format** to format the removable media.
7. Select **Done**.

DICOM Export

Use DICOM Export to export DICOM content in either .mp4 and .avi formats.




Figure 10-27 • Edit DICOM Export

This table shows all the preset parameters available on Edit DICOM Export with descriptions.

Table 10-91 • Edit DICOM Export

Preset Parameter	Description
Name	Logical name of the service.
Convert Clips	When enabled, the raw cine clips get converted to DICOM multi-frames during DICOM Export operation. On = Enabled/Off = Disabled

Preset Parameter	Description
Convert Volumes	When enabled, the raw volumes get converted to Enhanced US Volume (DICOM 3D) format during DICOM Export operation. On = Enabled/Off = Disabled
Apply Display Precision to Measurements	When enabled, the measurement values are rounded to configured number of decimal places. On = Enabled/Off = Disabled
Dicom Coding Scheme	Measurements are encoded with configured coding scheme in DICOM SR objects.
Use Button Settings	When enabled, uses button settings. On = Enabled. The Ultrasound system will use the default image storage and compression presets./Off = Disabled. Properties menu expands and image storage and compression presets can be defined.

Add DICOM Export

1. Navigate to **Connectivity > Connection Manager > MyComputer Device**.
2. Select **DICOM Export > Configure**.
3. In **Name**, enter a logical name of the service.
4. To have raw cine clips converted to DICOM multi-frames during EzBackup operation, enable **Convert Clips**.
5. To have raw volumes get converted to Enhanced US Volume (DICOM 3D) format during EzBackup operation, enable **Convert Volumes**.
6. To round measurement values to configured number of decimal places, enable **Apply Display Precision to Measurements**.
7. Under **Dicom Coding Scheme**, select how measurements are encoded with configured coding scheme in DICOM SR objects.
8. To use button settings, enable **Use Button Settings**.
9. Select **Done**.

USB Quick Save

Use USB Quick Save to save images or video clips to a USB flash drive with a print button. The images are stored in either .jpg or .wmv formats.




Figure 10-28 • Edit USB Quick Save

This table shows all the preset parameters available on Edit USB Quick Save with descriptions.

Table 10-92 • Edit USB Quick Save

Preset Parameter	Description
Name	Logical name of the service.
Destination	Destination device for the service.

Add USB Quick Save

1. Navigate to **Connectivity > Connection Manager > MyComputer Device**.
2. Select **USB Quick Save > Configure**.
3. In **Name**, enter a logical name of the service.
4. Under **Destination**, enter a destination device for the service.
5. Select **Done**.

HD Export

Use HD Export to save images or video clips to a USB flash drive or HD media with a print button. The images are stored in DICOM format.




Figure 10-29 • Edit HD Export

This table shows all the preset parameters available on Edit HD Export with descriptions.

Table 10-93 • Edit HD Export

Preset Parameter	Description
Name	Logical name of the service.
Destination	Destination device for the service.

Add HD Export

1. Navigate to **Connectivity > Connection Manager > MyComputer Device**.
2. Select **HD Export > Configure**.
3. In **Name**, enter a logical name of the service.
4. Under **Destination**, enter a destination device for the service.
5. Select **Done**.

Dicom Page

Use the Dicom page to add a Dicom device and then add one or more Dicom services to that device.




Figure 10-30 • Dicom Page

This table shows all the preset parameters available on Dicom with descriptions.

Table 10-94 • Dicom Page

Preset Parameter	Description
System Information	Information settings for the Ultrasound system.
Enable Dicom Verbose Logging	When enabled, DICOM traffic messages are logged to protected logs. The default time limit is set for 5 minutes. On = Enabled/ Off = Disabled
Devices	List of created devices.
Show special Dicom devices	When enabled, displays special dicom devices. By default, special dicom devices are hidden. When special devices (for example, Koios and Tricefy) are activated, the Ultrasound system automatically creates special dicom devices on the DICOM page.
Add Dicom Device	Select to add a new Dicom device.
Services	List of created services.
Add Dicom Service	Select to add a new Dicom service.

Refer to the following sections for more information:

- *Add Dicom Device* on page 513
- *Add Dicom Service* on page 515

Add Dicom Device

Use Add Dicom Device to add a destination device (printer, worklist server, etc.) to the Ultrasound system.




Figure 10-31 • Add Dicom Device (with Dicom Web Device On)

This table shows all the preset parameters available on Add Dicom Device with descriptions.

Table 10-95 • Add Dicom Device

Preset Parameter	Description
Name	Name for the DICOM device.
Dicom Web Device	Determines whether the device is a web device. On = Enabled/ Off = Disabled
Basic Authentication	When Dicom Web Device is enabled, determines whether basic authentication is used for the device. On = Enabled/ Off = Disabled
Username	Username used when Basic Authentication is enabled.
Password	Password used when Basic Authentication is enabled.
IP Address	When Dicom Web Device is Off, IP address for the device. IP stands for Internet Protocol. Every device on the network has a unique IP address.
Use DICOM TLS Encrypted Connection	When enabled, the Ultrasound system communicates securely with all the configured services for that device with secure communication by encrypting the DICOM objects during transfer through TLS1.2 protocol. On = Enabled/ Off = Disabled
Certificate Name	When Use DICOM TLS Encrypted Connection is enabled, name of imported certificate.
Ping Status	When Dicom Web Device is Off, indicates whether a ping to the device is allowed.
Ping	When Dicom Web Device is Off, select to confirm that the device is connected.

Preset Parameter	Description
Add	Select to add the DICOM device.
Discard and Close	Select to discard any changes you have made and to close the dialog box.

Add a Dicom device

1. Navigate to **Connectivity > Connection Manager > Dicom**.
2. To display special Dicom devices, enable **Show special Dicom devices**.
3. Select **Add Dicom Device**.

NOTE

To edit an existing device, select the **Edit** for the device.

4. In **Name**, enter a name for the Dicom device.
5. To add a device (not a web device), set **Dicom Web Device** to Off.
6. In **IP Address**, enter an IP address for the device.
7. If **Use DICOM TLS Encrypted Connection** is On, select the pencil icon to display the TLS Encrypted Configuration dialog box to add a certificate.
8. Select **Add**.

Add a Dicom Web device

1. Navigate to **Connectivity > Connection Manager > Dicom**.
2. To display special Dicom devices, enable **Show special Dicom devices**.
3. Select **Add Dicom Device**.

NOTE

To edit an existing device, select **Edit** for the device.

4. In **Name**, enter a name for the Dicom device.
5. To add a web device, set **Dicom Web Device** to On.
6. If Basic Authentication is On, enter a **Username** and **Password**.
7. If **Use DICOM TLS Encrypted Connection** is On, select the pencil icon to display the TLS Encrypted Configuration dialog box to add a certificate.
8. Select **Add**.

Add Dicom Service

For each added device, set up the service(s) that the device supports (you must be an administrator to update these screens).

Use Add Dicom Service to set the properties for the service. The name and properties in this section change, depending on what service is currently selected.




Figure 10-32 • Add Dicom Service

Refer to the following sections for more information:

- *Add Dicom Service - Service* on page 516
- *Add Dicom Service - Connection* on page 517
- *Add Dicom Service - Content* on page 518
- *Add Dicom Service - Compression* on page 520
- *Add Dicom Service - Summary* on page 520

Add Dicom Service - Service

Use Add Dicom Service - Service to set the properties for the service. The name and properties in this section change, depending on what service is currently selected.




Figure 10-33 • Add Dicom Service - Service

This table shows all the preset parameters available on Add Dicom Service - Service with descriptions.

Table 10-96 • Add Dicom Service - Service

Preset Parameter	Description
Name	Descriptive name for the service.
URL	URL for the service
Anonymize images Before	When enabled, anonymizes patient data. On = Enabled/ Off = Disabled

Add Dicom Service - Connection

Use Add Dicom Service - Connection to set the properties for the service. The name and properties in this section change, depending on what service is currently selected.




Figure 10-34 • Add Dicom Service - Connection

This table shows all the preset parameters available on Add Dicom Service - Connection with descriptions.

Table 10-97 • Add Dicom Service - Connection

Preset Parameter	Description
Maximum Retries	Maximum number of times to try establishing a connection to the service.
Retry Interval (sec)	Specify how often (in seconds) the Ultrasound system should try to establish a connection to the service.
Transmission Timeout (sec)	Specify the amount of time (in seconds) after which the Ultrasound system will stop trying to establish a connection to the service.

Add Dicom Service - Content

Use Add Dicom Service - Content to set the properties for the service. The name and properties in this section change, depending on what service is currently selected.




Figure 10-35 • Add Dicom Service - Content

This table shows all the preset parameters available on Add Dicom Service - Content with descriptions.

Table 10-98 • Add Dicom Service - Content

Preset Parameter	Description
Color Support	Select Mixed or Color. Best performance when set to Color.
Allow Raw Data	Select to save data in both TruAccess (raw data) and DICOM format. Clear to save in DICOM format only.
Allow Clips	Select to allow cine loop storage. Deselect to send only Stills to PACS.
Allow Enhanced DICOM Objects	Select to allow enhanced DICOM objects.
Key Image Notes	Image deletion notification. ONLY available for the Direct Store Workflow and ONLY generated when there are images deleted during the exam. Selecting this lets the reader at the PACS system know which images have been deleted. An indicator is placed on deleted images with a reason, "Rejected for Quality Reasons," for example.
Enable Structured Reporting	Select for Structured Reporting.
Dicom Coding Scheme	
Apply Display Precision to Measurements	When Enable Structured Reporting is selected, Apply Display Precision to Measurements is available; select to apply the same precision for a Structured Report that is used on the scanner display. (Apply Display Precision to Measurements is always available for a DICOM SR Storage service.)

Add Dicom Service - Compression

Use Add Dicom Service - Compression to set the properties for the service. The name and properties in this section change, depending on what service is currently selected.




Figure 10-36 • Add Dicom Service - Compression

This table shows all the preset parameters available on Add Dicom Service - Compression with descriptions.

Table 10-99 • Add Dicom Service - Compression

Preset Parameter	Description
Use Button Settings	If “Use Button Settings” is checked, the system will use the default system image storage and compression presets. If “Use Button Settings” is unchecked, the Properties menu expands and the user can define the image storage and compression presets as desired.

Add Dicom Service - Summary

Use Add Dicom Service - Summary to set the properties for the service. The name and properties in this section change, depending on what service is currently selected.




Figure 10-37 • Add Dicom Service - Summary

This table shows all the preset parameters available on Add Dicom Service - Summary with descriptions.

Table 10-100 • Add Dicom Service - Summary

Preset Parameter	Description
Service	Summary of the service settings for the service. See <i>Add Dicom Service - Service</i> on page 516.
Connection	Summary of the connection settings for the service. See <i>Add Dicom Service - Connection</i> on page 517.
Content	Summary of the content settings for the service. See <i>Add Dicom Service - Content</i> on page 518.
Compression	Summary of the compression settings for the service. See <i>Add Dicom Service - Compression</i> on page 520.
Verify Connection	Checks the connection status with the server and the compatibility with the PACS device.
Verify Timeout	Amount of time after which the system will stop trying to establish a connection to the service.

Add a Dicom service

1. Navigate to **Connectivity > Connection Manager > Dicom**.
2. Do one of the following:
 - a. If the service is already displayed in Services, select **Edit** for the service.
 - b. If the service is not displayed in Services, select **Add Dicom Service**.

3. Select **Service** and then configure the fields on the page.
4. Select **Connection** and then configure the fields on the page.
5. Select **Content** and then configure the fields on the page.
6. Select **Compression** and then configure the fields on the page.
7. Select **Summary** to review the settings.
8. Select **Done**.

Special Devices Page

Use the Special Devices page to configure or modify details about the special devices used with the Ultrasound system.




Figure 10-38 • Special Devices Page

Refer to the following sections for more information:

- *Tricify* on page 523
- *KOIOS* on page 524
- *Device Mgmt* on page 531
- *Insite* on page 533
- *Syslog* on page 537
- *Data Streaming* on page 539
- *Vscan Air* on page 540

Tricefy

Tricefy is a cloud-based image viewer and a platform to archive, collaborate, and share. The corresponding DICOM destinations can be used through the Print keys. An internet connection is necessary for uploading data to Tricefy. DICOM connectivity to the optional Tricefy Cloud PACS system occurs over a proprietary protocol protected by TLS/HTTPS encryption, although the internal communication is still DICOM-structured.

As soon as the Tricefy option is enabled, relevant Tricefy items are displayed.




Figure 10-39 • Tricefy Page

Table 10-101 • Configure Special Devices - Tricefy on page 523 lists the preset parameters available on the Configure Special Devices - Tricefy page.

Table 10-101 • Configure Special Devices - Tricefy

Preset Parameter	Description
Account Email	Email address for the Tricefy account.
Account Name	Name for the Tricefy account.
Customer Name	Name for the customer on the Tricefy account.
Uplink ID	Uplink ID for the Tricefy account.
Account Status	Status of the Tricefy account.
Activate	Select to activate the Tricefy option.
Discard Tricefy Changes	Select to remove any changes made to the Tricefy settings.

Configure Tricefy

1. Navigate to **Connectivity > Connection Manager > Special Devices**.
2. Select **Tricefy**.
3. Configure the fields on the page.
4. Select **Activate**.

5. Select **Done**.

KOIOS

Koios DS is a Breast and Thyroid Analysis Option. Koios DS is integrated with the LOGIQ Fortis via DICOM and is configured similar to a DICOM Service. The user can accept/dismiss analysis results. If accepted, these results are included in the DICOM Structured Report.




Figure 10-40 • KOIOS Page

This table shows all the preset parameters available on Configure Special Devices - KOIOS with descriptions.

Table 10-102 • Configure Special Devices - KOIOS

Preset Parameter	Description
IP Address	IP address for the Koios DS server.
Account Status	Status of the Koios DS account. Active or inactive.
Activate	Select to receive notification that you have successfully connected to the Koios DS server and the required device, service, and printflows will be automatically created.

Refer to the following sections for more information:

- *KOIOS DS Breast* on page 524
- *KOIOS DS Breast - Results Dialog* on page 525
- *KOIOS DS Breast - Assessment Menu* on page 526
- *KOIOS DS Breast - Summary* on page 527
- *KOIOS DS Thyroid - Results Dialog* on page 528
- *Accept Key* on page 529
- *Review Key* on page 530

KOIOS DS Breast




Figure 10-41 • KOIOS Page - KOIOS DS Breast

This table shows all the preset parameters available on KOIOS DS Breast with descriptions.

Table 10-103 • KOIOS Page - KOIOS DS Breast

Preset Parameter	Description
Lesion Classification System	Select BI-RADS or U1-U5.
Include Result in DICOM SR	Select to save Koios Breast results to Structured Reporting.

KOIOS DS Breast - Results Dialog




Figure 10-42 • KOIOS Page - KOIOS DS Breast Results Dialog

This table shows all the preset parameters available on KOIOS DS Breast Results Dialog with descriptions.

Table 10-104 • KOIOS Page - KOIOS DS Breast Results Dialog

Preset Parameter	Description
Show Classification Name	Select to show classification name.
Show Banner in Color	Select to show banner in color.

KOIOS DS Breast - Assessment Menu




Figure 10-43 • KOIOS Page - Koios DS Breast - Assessment Menu

This table shows all the preset parameters available on KOIOS DS Breast - Assessment Menu with descriptions.

Table 10-105 • KOIOS Page - KOIOS DS Breast - Assessment Menu

Preset Parameter	Description
Show Classification Name	Select to show classification name.
Show Banner in Color	Select to show banner in color.

KOIOS DS Breast - Summary




Figure 10-44 • KOIOS DS Breast - Summary

This table shows all the preset parameters available on KOIOS DS Breast - Summary with descriptions.

Table 10-106 • KOIOS Page - KOIOS DS Breast - Summary

Preset Parameter	Description
Show Classification Name	Select to show classification name.
Show Banner in Color	Select to show banner in color.

KOIOS DS Thyroid - Results Dialog




Figure 10-45 • KOIOS DS Thyroid - Results Dialog

This table shows all the preset parameters available on KOIOS DS Thyroid - Results Dialog with descriptions.

Table 10-107 • KOIOS Page - KOIOS DS Thyroid - Results Dialog

Preset Parameter	Description
Show Classification Name	Select to show classification name.
Show Banner in Color	Select to show banner in color.
Show Recommendation	Select to show recommendation.

Accept Key




Figure 10-46 • KOIOS Page - Accept Key

This table shows all the preset parameters available on Accept Key with descriptions.

Table 10-108 • KOIOS Page - Accept Key

Preset Parameter	Description
Display Key	Select to display this key in the Results dialog.
Save Image for Off Scanner Koios Analysis	Select to save the image sent to Koios in the exam for off scanner analysis.
Save Descriptors to Exam	Select to save Descriptors from Koios Analysis to the exam.
Save Koios Analysis to Exam	Select to save Koios Analysis result to the exam.

Review Key




Figure 10-47 • KOIOS Page - Review Key

This table shows all the preset parameters available on Review Key with descriptions.

Table 10-109 • KOIOS Page - Review Key

Preset Parameter	Description
Display Key	Select to display this key in the Results dialog.
Save Image for Off Scanner Koios Analysis	Select to save the image sent to Koios in the exam for off scanner analysis.
Save Descriptors to Exam	Select to save Descriptors from Koios Analysis to the exam.
Save Koios Analysis to Exam	Select to save Koios Analysis result to the exam.

Configure KOIOS

1. Navigate to **Connectivity > Connection Manager > Special Devices**.
2. Select **KOIOS**.
3. Configure the fields on the page.
4. Select **Activate**.
5. Select **Done**.

Device Mgmt

Device Mgmt is a remote device management tool that enables bi-directional management capabilities on the device. Device Mgmt allows Cloud management of system preset configurations to a fleet of systems on network, as well as one to one system preset configuration Cloud backup and restore.

NOTE

For Cloud operation please refer to Device Mgmt online user manual after sign-up at <http://AVURI.gehealthcare.com/signup>



Figure 10-48 • Device Mgmt Page

This table shows all the preset parameters available on Configure Special Devices - Device Mgmt with descriptions.

Table 10-110 • Configure Special Devices - Device Mgmt

Preset Parameter	Description
Account Email	Email address for the Device Management account.
Account Status	Indicates whether Device Management is active or not. Valid values are Inactive or Active. This field will be Inactive after activation has failed. Otherwise, the field will be empty. If you never activated or activated and then deactivated, this field will be empty.
Server Information	When enabled, displays the server information for the Device Management account. On = Enabled/Off = Disabled
Registration Key	Registration key for the Device Management account. Product specific key strings, pre-populated by the device (can be overwritten if necessary).
Server URL	URL of registration server strings, pre-populated by the device (can be overwritten if necessary).
Use Proxy	Enable to use a proxy server. On = Enabled/Off = Disabled
Use System Proxy	When selected, uses the proxy settings configured under the Network page > Proxy tab.
Proxy Server	Name of the proxy server IP provided by the customer. This field is optional unless required by the customer facility infrastructure.

Preset Parameter	Description
Proxy Port	Number of the proxy server port provided by the customer. This field is optional unless required by the customer facility infrastructure.
Proxy Credentials	Enable to use credentials for a proxy server. On = Enabled/Off = Disabled
Proxy User Name	Name of the proxy user provided by the customer. This field is optional unless required by the customer facility infrastructure.
Proxy Password	Password for the proxy user name provided by the customer. This field is optional unless required by the customer facility infrastructure.
Activate	Select to activate/deactivate Device Management.
Service Certificate	Select to download the certificate for Device Management from the backoffice server.
Remove Registration	Select to remove registration for Device Management from the backoffice server.

Configure Device Mgmt

1. Navigate to **Connectivity > Connection Manager > Special Devices**.
2. Select **Device Mgmt**.
3. Configure the fields on the page.
4. Select **Activate**.
5. Select **Done**.

Insite




Figure 10-49 • Configure Special Devices - Insite

This table shows all the preset parameters available on Configure Special Devices - Insite with descriptions.

Table 10-111 • Configure Special Devices - Insite

Preset Parameter	Description
Serial Number	Serial number of the agent (read-only). If the agent is not registered with a serial number, this field is populated with the serial number of the Ultrasound system. The serial number of the agent is tied to the serial number of the Ultrasound system.
Agent Model Number	GE part number for the Ultrasound system. The same number as listed on the rating plate.
CRM No	Customer Relationship Management (CRM) number. System identifier assigned to the customer unit by the service region. This number must match the System ID loaded in Siebel CRM and the GE CARES sticker.
Display Name	Displayed name of the agent.
Agent Version	When Agent Connected is Yes, version of the RsVP Agent that is installed on the machine which facilitates the connectivity to the backoffice.

Autoconnect





Figure 10-50 • Insite Page - Autoconnect

This table shows all the preset parameters available on Autoconnect with descriptions.

Table 10-112 • Insite Page - Autoconnect

Preset Parameter	Description
Autoconnect Status	Indicates whether Autoconnect to Insite is enabled or disabled.
Enable/Disable	Select to enable or disable Autoconnect to Insite.

Advanced Configuration

**Figure 10-51 • Insite Page - Advanced Configuration**

This table shows all the preset parameters available on Advanced Configuration with descriptions.

Table 10-113 • Insite Page - Advanced Configuration

Preset Parameter	Description
Server	Name of the enterprise server. This field should be normally set to Production for USCAN customers. Valid values are: • Production – Use to configure an enterprise server for USCAN. • Production-EU – Use to configure an enterprise server for Europe. • CURRENT – Use to configure an enterprise server. CURRENT will be available after configuring a server with the Others selection. • Others – Use to configure a staging server. Select Others, add the server url (stginsite.healthcare.ge.com) and port (443). Do not use the IP address of the enterprise server using the Others selection as Windows 10 has removed the ability to bypass a certificate error that occurs when using the IP address. This error can block connectivity.
Enterprise Host	Name of the enterprise host. This field should be normally set to insite.gehealthcare.com.

Preset Parameter	Description
Enterprise Port	Number of the enterprise port. This field should be normally set to 443.
Software Download URL	Address where software will be downloaded from to the Ultrasound system.
Add Enterprise Configuration	Use as an alternative option to add an enterprise server. Once you use Add Enterprise Configuration to add an enterprise server, that enterprise server will be listed under Server. Add Enterprise Configuration can be used instead of Others.

Proxy Configuration




Figure 10-52 • Insite Page - Proxy Configuration

This table shows all the preset parameters available on Proxy Configuration with descriptions.

Table 10-114 • Insite Page - Proxy Configuration

Preset Parameter	Description
Use Proxy	Enable to use a proxy server. On = Enabled/Off = Disabled
Use System Proxy	When enabled, uses the proxy settings configured under the Network page > Proxy tab. On = Enabled/Off = Disabled
Proxy Server	When Use Proxy is On, proxy server address of the facility.
Proxy Port	When Use Proxy is On, port number of the proxy server.

Preset Parameter	Description
Proxy Credentials	When Use System Proxy is Off, enable to use credentials for a proxy server. On = Enabled/Off = Disabled
Proxy User Name	When Proxy Credentials is On, account user name of the proxy server.
Proxy Password	When Proxy Credentials is On, account password of the proxy server

Configure Insite

1. Navigate to **Connectivity > Connection Manager > Special Devices**.
2. Select **Insite**.
3. Configure the fields on the page.
4. Select **Submit**.
5. Select **Done**.

Syslog




Figure 10-53 • Configure Special Devices - Syslog

This table shows all the preset parameters available on Configure Special Devices - Syslog with descriptions.

Table 10-115 • Configure Special Devices - Syslog

Preset Parameter	Description
Server IP Address/Hostname	IP address of the server.

Preset Parameter	Description
Mode of Transfer	Mode of transferring the data to the server (UDP, TCP, or TLS).
UDP Listener Port	User Datagram Protocol Listener Port. This is the port which we transfer logging data to use the syslog data protocol. It can be done by UDP or TCP, depending on the configuration of the customer logging server.
Ping Status	Indicates whether the connection has been confirmed or not.
Ping	Select to confirm that the server is connected.

WindowsEventLog




Figure 10-54 • Syslog - WindowsEventLog

This table shows all the preset parameters available on WindowsEventLog with descriptions.

Table 10-116 • Syslog - WindowsEventLog

Preset Parameter	Description
Transfer Application Event Log	When enabled, allows the transfer of application event logs. On = Enabled/Off = Disabled
Transfer Security Event Log	When enabled, allows the transfer of security event logs. On = Enabled/Off = Disabled
Source of application event log to transfer	Where to draw data from for the audit report.
Add Source	Select to specify where to draw data from for the audit report.

Transfer




Figure 10-55 • Syslog - Transfer

This table shows all the preset parameters available on Transfer with descriptions.

Table 10-117 • Syslog - Transfer

Preset Parameter	Description
Send Time Period (Sec)	How often the system will attempt to contact the logging server to send logs.
Number of days of Historic data to send	If logging server connectivity has not been available, how many days of history should be sent when it is restored?
EventLog Batch Send Size	When sending historic data, how much to send in each transaction – this can help manage network traffic and server load, not typically necessary to change this.

Configure Syslog

1. Navigate to **Connectivity > Connection Manager > Special Devices**.
2. Select **Syslog**.
3. Configure the fields on the page.
4. Select **Ping**.
5. Select **Done**.

Data Streaming

Use Data Streaming to stream live/recall/CINE ultrasound image data over the network connection to enabled devices. The data stream will contain grayscale, color map, geometry, view settings (flip/rotate/reverse), probe and system information, VNav position information

and CT/MR registration information with ultrasound data. No patient information is transferred with the streamed data.

It is recommended to use a 1 Gbps network connection for Data Streaming. The required bandwidth often lies in the 100-300 Mbps range. Usage of a 100 Mbps network leads to dropped frames and the risk of latency buildup.

The Data Streaming option key needs to be installed to enable the Data Streaming page.




Figure 10-56 • Configure Special Devices - Data Streaming

This table shows all the preset parameters available on Configure Special Devices - Data Streaming with descriptions.

Table 10-118 • Configure Special Devices - Data Streaming

Preset Parameter	Description
System Information	Shows system information properties for the configured Ultrasound system.
Enable Streaming	When enabled, data streaming is allowed.
Close Stream on Patient/Exam Change	When enabled, data streaming will close when the patient/exam is changed.
PortNo	Port number used for data streaming.
Revoke Selected	Select to remove selected client certificates.
Revoke All Expired	Select to remove all expired client certificates.

Configure Data Streaming

1. Navigate to **Connectivity > Connection Manager > Special Devices**.
2. Select **Data Streaming**.
3. Configure the fields on the page.
4. Select **Done**.

Vscan Air

Use Vscan Air to pair the Vscan Air CL probe with the Ultrasound system.

The Vscan Air CL probe is a battery-operated, wireless, general-purpose diagnostic handheld ultrasound imaging system.

The Vscan Air option key needs to be installed to enable the Vscan Air page.




Figure 10-57 • Configure Special Devices - Vscan Air (Not Paired with the Probe)



Figure 10-58 • Configure Special Devices - Vscan Air (Paired with the Probe)

This table shows all the preset parameters available on Configure Special Devices - Vscan Air with descriptions.

Table 10-119 • Configure Special Devices - Vscan Air

Preset Parameter	Description
Connection Status	Indicates the pairing status (with the serial number) of the Vscan Air probe.

Preset Parameter	Description
Wi-Fi Environment	Indicates whether the current Wi-Fi environment (signal strength, channel usage) is appropriate for Vscan Air usage and displays errors if there are issues with Wi-Fi hardware.
Country Code	Setting for the Vscan Air probe to meet country regulations (for example, Wi-Fi frequencies).
Use probe button as	Maps the Vscan Air probe button to different functions (for example, Freeze/Print/Toggle).
Probe Serial Number	Serial number of the paired Vscan Air probe.
Probe Model	Model of the paired probe.
Probe Firmware Version	Firmware version on the paired Vscan Air probe.
Battery Level	Battery level for the paired Vscan Air probe.
Probe Temperature	Current temperature of the paired Vscan Air probe.
Activation Status	Indicates whether the Vscan Air probe is activated or not.
Acoustic Output Table (R1)	Link to the acoustic power output table for the paired Vscan Air probe.
Supporting Documents	Link to manuals and tips for the Vscan Air probe (if applicable).
Firmware Compatibility	Lists all the compatible Vscan Air probe firmware versions that can be paired with the Ultrasound system. If the firmware on the paired probe is not compatible with the Ultrasound system, firmware downgrade menu will be activated to allow user to downgrade probe firmware to a compatible probe firmware version.

Configure Vscan Air CL

1. Navigate to **Connectivity > Connection Manager > Special Devices**.
2. Select **Vscan Air**.
3. Configure the fields on the page.
4. Select **Done**.

Print Button Page

Use the Print Button page to configure or modify details about the print buttons used for the Ultrasound system.




Figure 10-59 • Print Button Page

Configure Print Button




Figure 10-60 • Configure Print Button




Figure 10-61 • Configure Print Button - Advanced

This table shows all the preset parameters available on Configure Print Button with descriptions.

Table 10-120 • Configure Print Button

Preset Parameter	Description
Format	Format: Ultrasound Image, Secondary Capture (Image, Video, Screen)
Still Image Compression	Compression: None, Rle, Jpeg, Jpeg2000
Add Multiframe Data	Clips: Add Multiframe Data: Checkbox
Clip Compression	Compression: None, Rle, Jpeg, Jpeg2000
Clip Quality	Quality: Lossless, 99, 98, 97, ... 50
Volume File Format	1-Standard DICOM (Default), 2-Standard DICOM with Raw Data; 3-Enhanced DICOM, 2&3 (2 files)
Compare Assistant	Compare Assistant: Comparison view (Default), New image, Both Comparison & New (2 files)
VNav Data	VNav Data: V Nav View (Default), Ultrasound Only, or VNav & Ultrasound (2 files)
Scan Assistant Advance	Scan Assistant Advance: On, Off, Use program (system uses setting from the Scan Assistant program which allows a user to configure two print keys identically except that one advances Scan Assistant and the other does not.) On =advances to the next step when that print key is pressed independently of the program setting. Off = does not advance to the next step when the print key is pressed independently of the program setting.

Configure a print button

1. Navigate to **Connectivity > Connection Manager > Print Button**.
2. Select the print button.
3. Configure the fields on the page.
4. Select **Done**.

Dataflow Page

Use the Dataflow page to configure or modify details about the data flow properties used for the Ultrasound system.

A dataflow is a set of pre-configured services. When you select a dataflow, the ultrasound system automatically works according to the services associated with the dataflow. The Dataflow tab allows you to select and review information about dataflows. You can also create, change, and remove dataflows.

Set up dataflows for the services.

NOTE: You must be logged on as Administrator to use the Dataflow tab.




Figure 10-62 • Dataflow Page

This table shows all the preset parameters available on Dataflow with descriptions.

Table 10-121 • Dataflow Page

Preset Parameter	Description
Direct Store	Select to store data directly to archive (no buffer storage).
Hidden	Select so that this dataflow does not appear as a Dataflow on the Patient menu.

Preset Parameter	Description
Set As Default	Select to use this dataflow as the default dataflow when the Ultrasound system starts.
Add New Dataflow	Select to add a new dataflow.

Add a dataflow

1. Navigate to **Connectivity > Connection Manager > Dataflow**.
2. Select **Add New Dataflow**.
3. Configure the fields on the page.
4. Select **Add**.

Advanced Settings Page

Use the Advanced Settings page to configure or modify details about the advanced setting properties used for the Ultrasound system.

Refer to the following sections for more information:

- *Advanced Settings - Patient Screen page* on page 546
- *Advanced Settings - Search Settings page* on page 548
- *Advanced Settings - Worklist Settings page* on page 549
- *Advanced Settings - Dataflow Settings page* on page 550
- *Advanced Settings - Dataflow Notification page* on page 551
- *Advanced Settings - Transfer Settings page* on page 552
- *Advanced Settings - Print Button Settings page* on page 554
- *Advanced Settings - Image Numbering page* on page 555
- *Advanced Settings - Imaging Insights page* on page 556
- *Advanced Settings - Measurement page* on page 557
- *Advanced Settings - Spooler Settings page* on page 558
- *Advanced Settings - LOGIQ Apps page* on page 559

Advanced Settings - Patient Screen page




Figure 10-63 • Advanced Settings - Patient Screen

This table shows all the preset parameters available on Advanced Settings - Patient Screen with descriptions.

Table 10-122 • Advanced Settings - Patient Screen

Preset Parameter	Description
Automatic generation of patient ID	In the Search/Create Patient window: When selected, the Patient ID is not required when entering a new patient in the archive. The system automatically generates an ID number. When cleared, the Patient ID is required when entering a new patient in the archive. On = Enabled/Off - Disabled
Show BBT	Show BBT field on the OB patient screen to input the basal body temperature. On = Enabled/Off - Disabled
Use birthdate	In the Patient information window, enter either the patient age or the birth date: When selected, enter birth date, then the age is calculated. When cleared, enter age (birth date field not available). On = Enabled/Off - Disabled
Detail Mode	Select to display Detail Mode, rather than Exam View, when you select the patient name in the patient list on the Patient menu. You can also type comments while in Detail Mode. On = Enabled/Off - Disabled
Quick New Patient Entry	Select to store new patient automatically by pressing the Patient key. On = Enabled/Off - Disabled

Preset Parameter	Description
Enable Other ID	Not selected is the Default. If selected, allow entering Other ID, such as Citizen Service Number, Burger Service Number (BSN), National Health System (NHS) number, along with patient ID information on the Patient Screen. On = Enabled/Off - Disabled
Validation Format	If the Enable Other ID preset is selected, the system validates the format of "Other ID" when an ID is entered. Choose: NHS Number *** ** ***** , Letters and Numbers, Numbers, or Any (no restriction).
Columns in examination listing	Create new columns, remove columns, and select the information to display in a column. Use the arrows (<< or >>) to reposition column headings.

Configure Patient Screen

1. Navigate to **Connectivity > Connection Manager > Advanced Settings**.
2. Select **Patient Screen**.
3. Configure the fields on the page.
4. Select **Next** to proceed to Search Settings.

Advanced Settings - Search Settings page




Figure 10-64 • Advanced Settings - Search Settings

This table shows all the preset parameters available on Advanced Settings - Search Settings with descriptions.


Table 10-123 • Advanced Settings - Search Settings

Preset Parameter	Description
Auto search for Patient	In the Search/Create Patient window: When selected, the system automatically searches through the selected patient archive, while the user enters patient information. When cleared, the automatic search tool is turned off. If you are trying to keep the past patient data confidential, DO NOT use this feature. On = Enabled/Off - Disabled
Keep Search String	Search string is kept rather than cleared. On = Enabled/Off - Disabled
Remember cursor position on Transfer Screen	Enable to set a default cursor location on the Data Transfer screen. On = Enabled/Off - Disabled

Configure Search Settings

1. Navigate to **Connectivity > Connection Manager > Advanced Settings**.
2. Select **Search Settings**.
3. Configure the fields on the page.
4. Select **Next** to proceed to Worklist Settings.

Advanced Settings - Worklist Settings page

**Figure 10-65 • Advanced Settings - Worklist Settings**

This table shows all the preset parameters available on Advanced Settings - Worklist with descriptions.

Table 10-124 • Advanced Settings - Worklist Settings

Preset Parameter	Description
Worklist Auto Query	Automatically queries the worklist server. On = Enabled/Off - Disabled
Validate Incoming Worklists	Confirm incoming Worklists are valid. On = Allowed/Off = Not allowed

Configure Worklist Settings

1. Navigate to **Connectivity > Connection Manager > Advanced Settings**.
2. Select **Worklist Settings**.
3. Configure the fields on the page.
4. Select **Next** to proceed to Dataflow Settings.

Advanced Settings - Dataflow Settings page




Figure 10-66 • Advanced Settings - Dataflow Settings

This table shows all the preset parameters available on Advanced Settings - Dataflow with descriptions.


Table 10-125 • Advanced Settings - Dataflow Settings

Preset Parameter	Description
Automatic Disable Patient Data	Select to automatically disable patient data. If selected, locks the patient name, date of birth and gender (like Patient ID). The Factory Default for this preset is unchecked. On = Enabled/Off - Disabled
Auto Archiving Patient Data	Archives patient data automatically. On = Enabled/Off - Disabled
After [End Current Patient], go to:	Select Worklist screen or Patient screen.
Double click on patient list to start	Select Review or New Exam to display each time you double click on the patient name in the patient list on the Patient menu.

Configure Dataflow Settings

1. Navigate to **Connectivity > Connection Manager > Advanced Settings**.
2. Select **Dataflow Settings**.
3. Configure the fields on the page.
4. Select **Next** to proceed to Dataflow Notification.

Advanced Settings - Dataflow Notification page

**Figure 10-67 • Advanced Settings - Dataflow Notification**

This table shows all the preset parameters available on Advanced Settings - Dataflow Notification with descriptions.

Table 10-126 • Advanced Settings - Dataflow Notification

Preset Parameter	Description
Warn image store without patient	Select to receive a warning when you press the Print key without an active patient. On = Enabled/Off - Disabled
Warn image store to Read Only dataflow	The system posts a warning message if you attempt to store images to a read-only Dataflow. On = Enabled/Off - Disabled
Warn register to No Archive	Select to receive a warning when you register a patient to the "No Archive" data flow. Select a different data flow for permanent storage of patient data. On = Enabled/Off - Disabled
Request acknowledge of End Exam action	When selected, the user is asked to confirm action when ending an examination. On = Enabled/Off - Disabled
Warn video titles exist in the internal storage	The system posts a warning if the video titles exist on the internal DVR flash memory. On = Enabled/Off - Disabled

Configure Dataflow Notification

1. Navigate to **Connectivity > Connection Manager > Advanced Settings**.
2. Select **Dataflow Notification**.
3. Configure the fields on the page.
4. Select **Next** to proceed to Transfer Settings.

Advanced Settings - Transfer Settings page




Figure 10-68 • Advanced Settings - Transfer Settings

This table shows all the preset parameters available on Advanced Settings - Transfer Settings with descriptions.

Table 10-127 • Advanced Settings - Transfer Settings

Preset Parameter	Description
Send Images via Wireless	When enabled and connected to the network through a wireless LAN, images will be sent to the DICOM device over the wireless LAN. If disabled, images spooled in the Spooler will be sent when the Ultrasound system is connected to the wired network. On = Enabled/Off - Disabled
Receive Images Wireless	When enabled and connected to the network through a wireless LAN, images will be received from the DICOM device over the wireless LAN. If disabled, images will be received when the Ultrasound system is connected to the wired network. On = Enabled/Off - Disabled
Send To: Default Service	Select default Send To service from the dropdown list.
Process Dicom Job Without Network	When enabled, preset that allows communication between the Ultrasound system and the Koios Server VM (installed in the Ultrasound system) even when there is no network. On = Enabled/Off - Disabled

Configure Transfer Settings

1. Navigate to **Connectivity > Connection Manager > Advanced Settings**.
2. Select **Transfer Settings**.
3. Configure the fields on the page.

4. Select **Next** to proceed to Print Button Settings.

Advanced Settings - Print Button Settings page




Figure 10-69 • Advanced Settings - Print Button Settings

This table shows all the preset parameters available on Advanced Settings - Print Button Settings with descriptions.

Table 10-128 • Advanced Settings - Print Button Settings

Preset Parameter	Description
Allow press and hold print key to replace an image	Select to enable pressing and holding print key to replace an image. On = Enabled/Off - Disabled
Enable Smart Capture Area	Check box to select. On = Enabled/Off - Disabled
Add Titlebar information to Multiframe loops	Adds a title bar to the DICOM image. On = Enabled/Off - Disabled
Store Dicom MultiFrame When Collecting RF Data	Select to store DICOM multiFrame images when collecting RF Data. On = Enabled/Off - Disabled
Store 2D Loop with Timeline Data	Check box to select. On = Enabled/Off - Disabled
Add Scan Parameter information to Multiframe Loops	Adds scan parameter(s) to the DICOM image. On = Enabled/Off - Disabled
Store Dicom MultiFrame in QAnalysis pack	Select to store DICOM MultiFrame images in Quantitative Analysis mode. Selected by default. On = Enabled/Off - Disabled

Preset Parameter	Description
Patient List Print - Font Size	Select font size.
P[1-4] Key Sound	Select None, Click, Chimes, Ding, Ding-Dong, or Whoosh.

Configure Print Button Settings

1. Navigate to **Connectivity > Connection Manager > Advanced Settings**.
2. Select **Print Button Settings**.
3. Configure the fields on the page.
4. Select **Next** to proceed to Image Numbering.

Advanced Settings - Image Numbering page




Figure 10-70 • Advanced Settings - Image Numbering

This table shows all the preset parameters available on Advanced Settings - Image Numbering with descriptions.

Table 10-129 • Advanced Settings - Image Numbering

Preset Parameter	Description
Use Scan Assistant image ordering/numbering	When enabled, uses Scan Assistant image reordering/renumbering. On = Enabled/Off - Disabled
Image Numbering Scheme	Preset that defines how the images are numbered during the examination/scanning process.

Preset Parameter	Description
Show image number on Active Images screen	When enabled, shows the image number on the Active Images screen. On = Enabled/Off - Disabled
Image Reordering Scheme	Preset that defines how the images are numbered during image reorder operation.

Configure Image Numbering

1. Navigate to **Connectivity > Connection Manager > Advanced Settings**.
2. Select **Image Numbering**.
3. Configure the fields on the page.
4. Select **Next** to proceed to Image Insights.

Advanced Settings - Imaging Insights page




Figure 10-71 • Advanced Settings - Imaging Insights

This table shows all the preset parameters available on Advanced Settings - Imaging Insights with descriptions.

Table 10-130 • Advanced Settings - Imaging Insights

Preset Parameter	Description
Store incremental analytic data in each image	Select to store incremental data in each image. On = Enabled/Off - Disabled
Send consolidated analytic data to server	Select to send consolidated analytic data to server. On = Enabled/Off - Disabled

Preset Parameter	Description
Image Analytics Server	Select server from dropdown list.
Trigger interval outside an exam (hours)	Off, 1, 2, 3, 5

Configure Imaging Insights

1. Navigate to **Connectivity > Connection Manager > Advanced Settings**.
2. Select **Imaging Insights**.
3. Configure the fields on the page.
4. Select **Next** to proceed to Measurement.

Advanced Settings - Measurement page




Figure 10-72 • Advanced Settings - Measurement

This table shows all the preset parameters available on Advanced Settings - Measurement with descriptions.

Table 10-131 • Advanced Settings - Measurement

Preset Parameter	Description
Link Measurement to Images	Links measurements to images when sent to PACS. On = Enabled/Off - Disabled
Verify all Measurements have Image References	Verifies that all measurements have image references when sent to PACS. On = Enabled/Off - Disabled

Configure Measurement

1. Navigate to **Connectivity > Connection Manager > Advanced Settings**.
2. Select **Measurement**.
3. Configure the fields on the page.
4. Select **Next** to proceed to Spooler Settings.

Advanced Settings - Spooler Settings page




Figure 10-73 • Advanced Settings - Spooler Settings

This table shows all the preset parameters available on Advanced Settings - Spooler Settings with descriptions.

Table 10-132 • Advanced Settings - Spooler Settings

Preset Parameter	Description
Media Spooler Jobs Display Time (mins)	Number of minutes completed jobs will display in the media spooler.
Network Spooler Jobs Display Time (mins)	Number of minutes completed jobs will display in the network spooler.

Configure Spooler Settings

1. Navigate to **Connectivity > Connection Manager > Advanced Settings**.
2. Select **Spooler Settings**.
3. Configure the fields on the page.
4. Select **Next** to proceed to LOGIQ Apps.

Advanced Settings - LOGIQ Apps page




Figure 10-74 • Advanced Settings - LOGIQ Apps

This table shows all the preset parameters available on Advanced Settings - LOGIQ Apps with descriptions.

Table 10-133 • Advanced Settings - LOGIQ Apps

Preset Parameter	Description
Unpair Bluetooth Devices	Select to unpair the LOGIQ Apps device.

Configure LOGIQ Apps

1. Navigate to **Connectivity > Connection Manager > Advanced Settings**.
2. Select **LOGIQ Apps**.
3. Configure the field on the page.
4. Select **Next** to proceed to Summary.

Connection Manager Summary Page

The Summary page lists the status of the wired or wireless connectivity and proxies configured on the left and Dicom Devices and Services on the right (see *Figure 10-75 • Summary Page* on page 560.)



Figure 10-75 • Summary Page

Device status is noted with different colors:

- Green - Device and all services are available
- Red - Device is offline, or all the services are offline
- Orange - Device is available. However, one more of its services are offline
- Gray - Device has no service or service status is being updated

Navigate to configure connectivity functions by clicking the icons at the top of the screen to access the desired configuration screen, or by selecting **Next** or **Previous** at the bottom of the screen.

Measure

Please refer to Chapter 7, General Measurements and Calculations for more information on setting up Measurement and Analysis Presets.

Reports

Refer to Chapter 13 for more information.

System Administration

The Admin screen has the following sections:

- **System Administration** – lists all the options implemented in the system.
- **Users** – allows you to define user IDs, specify operator's registration, operator's rights, registration of staff related to an examination (for example, referral doctors and sonographers) and password update requirements.
- **Logon** – defines logon procedures, allows the System Administrator to set password policies, and LDAP Configuration.
- **Groups** – The System Administrator can set up user groups on this page.
- **System Password** – Password for the Application Windows Logon Account.
- **Disk Encryption** – The System Administrator can encrypt the disk for highest system security, if required.
- **Audit Report** – Generates an audit report.
- **Vulnerability Scan Mode** – Part of the Advanced Security Option. Vulnerability Scan Mode configures the system so that an external scanner can analyze the LOGIQ Fortis.

NOTE

Only the System Administrator can access the pages.

Administrator Tasks

Here are the tasks typically performed by the LOGIQ Fortis System Administrator. Instructions for each of these tasks can be found in this chapter.

- Privacy and Security Configuration options
- Define connection options for LDAP Directory Server
- Control encryption of the Patient Data drive
- Create Users/Groups
- Change rights for Users/Groups
- Change Encryption status for the system
- Change/Add Encryption Passwords
- Save/Print Recovery keys for encrypted drives
- Configure Password Policies
- Configure Session Policies
- Vulnerability Scan Mode
- Generate Audit Reports
- Configure Remote Logging Servers

Privacy and Security

Privacy protects both personal and private interests and information of persons. Security protects both system and information from risks to confidentiality, integrity, and availability. The system Privacy and Security capabilities are discussed in Chapter 12; Privacy and

Security configuration is discussed below, followed by a description of each system administrative page.

Creating Password Policies and User Groups

The foundation of setting up effective privacy and security is controlling user groups, users, and their system permissions.

Enabling Password Policies

Check the “Enable Password Policies” box on the **Utility > Admin > LOGON** page.

In addition, set the password policy for each selection. These policies should be set BEFORE creating new users. Save the changes and reboot the system. This applies rules for the passwords.



HINTS

Ensure the ADM password is known BEFORE rebooting the system.

Figure 10-76 • Enable Password Policies

Groups

You define Group access rights via the Utility-> Admin-> Groups page. When each user logs in, they will have access to the system according to the rights assigned to their group(s). Default system groups are preset with pre-determined access rights. To view these access rights, highlight a group name in the Group List column on the left, then look in the Groups Rights column to see the permissions by group.



HINTS

Admin has full access rights.




Figure 10-77 • Groups Page

Adding users

You define Group access rights via the **Utility > Admin > Utility** page.



Figure 10-78 • Page to Add Users

Creating a user

1. Press **Add**.
2. Type the user ID.

NOTE

ENSURE that you DO NOT include the following characters in a user's ID: slash (/), dash (-), asterisk (*), question mark (?), an underscore (_), ampersand (&), or blank spaces.

3. Type the user's information in the Identity section, ensuring that you are using the defined policies. User IDs are enforced as all uppercase letters for local users.

NOTE

If a password is created, the user will be required to enter the password when logging on, even if Password Policies is not enabled. Remember, passwords are case-sensitive.

4. Select the user's group(s). Multiple groups can be selected, if needed.
5. If the user needs full configuration and advanced operations access, select **Admin**.
6. Press **Save**.

NOTE

DO NOT add users with the same User ID. The system allows you to do this; however, the first user is erased and only the second user remains.

When adding a new user, press Add first. Then edit the ID from the default of "NewUser" and edit the other fields. DO NOT press Add again unless you actually want to create another user. Press Save after adding one or more users. The user listed as NewUser on the list will be updated with the edited ID when you re-enter this screen.

User Login

After the system administrator adds a new user, the new user should login with their new user name and password. If designated by the system administrator, they may be required to change their password.

The user will be prompted for their password when logging into the system or when selecting their user name from the Patient menu as well.

After logging in, the user will have access according to the rights available to them within their assigned group(s).

Changing a user configuration

The system administrator can update a user. The system administrator can also specify whether the user's account is "Active," "Blocked," or requires a password change. If needed, select the check box, "User must change password." Then the user will be prompted to change their password the next time they log into the LOGIQ Fortis.

1. Move the **Trackball** to a user ID in the User List.
2. Make the desired changes.

Deleting a user

The system administrator can delete a user.

1. Move the **Trackball** to a user ID in the User List.

2. Select **Remove**. This marks the account as inactive.

The user is removed from the User List.

3. Select Remove again.

NOTE

Accounts are not removed immediately because User data is retained for auditing purposes. This can be useful because their name will still appear in audit reports. Also, they can be reactivated.

If you permanently remove the user, by selecting "Remove" on an inactive user, this traceability will be lost.

User Accounts and Password Policies Frequently Asked Questions

Here are answers to some frequently-asked questions:

Q I lost or forgot my password.

A System Admin can change your password.

Q I entered the wrong password multiple times and now the system says I am locked out.

A Wait until you are unlocked, enter the correct password. If you have forgotten your password, the System Admin can access the system and user details. **Utility > Admin > Users**.

Q I am locked out after multiple attempts to remember my password, does the System Admin have to wait until my account is unlocked before accessing the system?

A The System Admin can logon with their own logon information before the users blocked time is over. The box "Block user account" in **Utility > Admin > User**, should be unchecked to allow the user to logon.

Q I have created a list of users on one u/s system, I have three more in the department, can I do a backup to disk and restore the user list onto the other systems?

A Yes, you can copy both User and Password.

Q I have a sonographer who left the facility, how do I delete this user?

A The System Admin can deactivate or delete a user. **Utility > Admin > Logon**. Select the user and press "remove"

Q I have a sonographer who is on medical leave for a few weeks. I do not want to remove their user logon but I want to make sure they cannot access the system while not working.

A The System Admin can block a user account. **Utility > Admin > Logon**, select the user and check "Block user account," or deselect "active user account."

Q We have information that a weekend users password may have been compromised, I need to ask the user to change their password the next time they logon.

A The System Admin can require the user to change their password. **Utility > Admin > Logon**, select the user and check "Require password change"

Q I have created a new user for a new sonographer, how do I assign the correct group?

A When a new user is added to the list, the groups list is located on the right column of the screen. Assign the user to a group or multiple groups with the appropriate access rights.

Q Can I change the access rights for a pre-defined group?

A Currently this is not supported. You can select multiple groups from the factory default list for any user that needs additional access but not full admin rights; or create your own groups with desired rights.

Q The default “ADM” user account does not have a password, can I create a password for this account?

A Yes, you can create a password for the default ADM. Be sure to write this down for anyone who will need access to the system, such as GE Service.

Q I was the last user on the system, the screen has gone black, do I need to logon again?

A Yes, once the Screen Lock time has been reached touch any button or the trackball on the operator panel to display the logon screen. Logon with your current credentials.

Q Can I use the Auto Logon feature?

A Yes, the Auto logon if checked will logon with the last user if the “Use password policies” is unchecked and the user has blank password. If the user is not assigned a password they will be logged on with no other entry required. If the user is assigned a password the logon window will pop-up requiring the password to be entered.

Locking the Screen and Logon

You define Automatically Lock Screen via the **Utility > Admin > Logon** page by checking the Enable Session Timeout (Lock Screen) box. Then go to the bottom of the column to designate after how many minutes the screen should be locked (Session Lock Screen Timeout (min)).



Figure 10-79 • Enable Lock Screen

The screen will lock after the designated time. During this time, the system will be completely black. To reactivate the system, the user will need to logon again.

NOTE

If “use password policies” is unchecked and no password is set to the user, and when Auto Logon is checked, the system will start by using the ID of the last operator. If Admin is used as the operator, assure that all characters are removed from the password field.

To lock the screen,

- Type [Alt+L].

System Admin

The System Admin screen has information about any options implemented for the system.



Figure 10-80 • Administrative System Admin Preset Menu

Table 10-134 • Product

Preset Parameter	Description
Product	The name of the product.
HW Number	The hardware number of the product.

Table 10-135 • Software Option Key

Preset Parameter	Description
Enter New Option Key	Type the key for the option you wish to add and press Add. Note: The Option Key may contain alphanumeric characters (2-9, A-H, J-N, P-Z) and special characters ("?", "%", "&").
Installed Option Keys	Lists the key for the installed options.
Add	Press to add a software option key.
Remove	To remove a software option key, select the key in the SW Option Key list, and then select Remove.
Import	Used to import option strings stored on media (USB or DVD) or in the system (OptionKeys.txt may be stored at d:\service). When you press Import, a dialog box comes up that allows you to select from the following locations: USB drive, DVD drive, and d:\service.

Table 10-136 • Service

Preset Parameter	Description
Enable Automatic Request for Service	Check this box to enable the system to send system-generated requests for service, without your intervention.
Transfer Machine Logs to GE HealthCare	Allows transfer of monitoring errors and status logs to GE HealthCare back office for data analytics.
Transfer Operator Information to GE HealthCare	Allows transfer of operator information to GE HealthCare back office for usage analytics.

Preset Parameter	Description
Enable Automatic Probe Diagnostics	Enable Automatic Probe Diagnostics to run and save data to log file for transfer to GE HealthCare backoffice for analysis.
Transfer Images (captured without patient information) to GE HealthCare	Allows transfer of images for analysis.
Include Annotation with Image transfer to GE HealthCare	Usage Analysis: Allows transfer of operator information to GE HealthCare back office for usage analytics.

If your site decides to deactivate the InSite ExC Agent, then Remote Connectivity is no longer available on the LOGIQ Fortis. This also means that Remote Service can no longer connect to the LOGIQ Fortis via Disruptive Mode to diagnose system issues. and the “Service” section will be removed from this Utility page. In addition, you will not be able to initiate a Request for Service or Clinical Support Request via the “GE HealthCare InSite ExC” icon control located at the bottom of the Monitor Display.



Figure 10-81 • Remote Connectivity Deactivated; Service Fields Removed

Table 10-137 • Protecting Health Information (PHI)

Preset Parameter	Description
Prevent Writing to Removable Media CD/DVD/USB (requires reboot)	Check this box to prevent users from copying/saving information to removable media.

Table 10-138 • USB External Media

Preset Parameter	Description
USB External Media disabled (requires reboot) By Checking this box you will disconnect all external USB Mass Storage devices	Select to disable any USB media from connecting to the ultrasound system. Please remember to enable this feature to reload software or to install the eIFU USB Media.

Table 10-139 • Rights

Preset Parameter	Description
Require Admin Operator Rights to Save Imaging Settings	Check this box to require the User to have Administrative rights in order to save image settings.

Table 10-140 • Option Status

Preset Parameter	Description
Options	A list of the option name and status.
Status	Lists each option's effectivity.

Users

The Users screen allows you to define user IDs. It also allows you to specify operators registration, operator's rights setting, and registration of staff related to an examination (for example, referring and interpreting physicians).

**Figure 10-82 • Users Preset Menu****Table 10-141 • User List**

Preset Parameter	Description
User List	Lists the user ID for all system users.
Identity	Type the operator's user ID, Password, Prefix, Last Name, First Name, Middle Name, Suffix, Phone Number.
User Controls	The System Administrator can specify whether a user's account is active, blocked, or requires a password update.

Preset Parameter	Description
Group Membership	Select the user's group: Operator (sonographers, doctors, or any person using the ultrasound system); Ref.Phys. (referring physician can be associated to the patient examination in the extended Patient information window); Perf.Phys. – physician performing the exam can be associated to the patient examination in the extended Patient information window. Note, other groups may exist as set up by the System Administrator.

Logon

The Logon section defines log on procedures.



Figure 10-83 • Administrative Logon Preset Menu

Table 10-142 • Logon

Preset Parameter	Description
Auto Logon	Specifies logon procedures: - Use Auto Logon - When selected, the system is started automatically, using the last user logon. - When blank, the user must select a user ID and enter a password when logging on. NOTE Auto LOGON only works when password policies are disabled and if there is no password assigned to the user. - Allow Auto Logon for Admin users: The system is started automatically when logging in as an ADM user. - Allow Emergency User: The system will allow the EUSR user for emergency use.

Preset Parameter	Description
Common Network Login	Specifies the user ID and password used to access the network. - User – User ID for network access - Password – Password for network access
Connectivity Maintenance	Reset to factory default.
LDAP Configuration	Refer to the next section for LDAP instructions.
Security Baseline	Choose Security Baseline. Select a baseline set of security policies for password and session management. For example: Lowest (default): Autologon available. No password complexity rules. Medium: Autologon unavailable. Passwords must meet the following criteria: - Minimum password length of 8. - Minimum of 2 character sets. - Password should not contain username. - Password should not be any of the last 10 passwords. High: Autologon unavailable. Passwords must meet the following criteria: - Minimum password length of 10. - Minimum of 4 character sets. - Minimum of 1 lower case characters. - Minimum of 1 upper case characters. - Minimum of 1 special characters. - Minimum of 1 digits. - Password should not contain username. - Password should not be any of the last 15 passwords. Highest: Autologon unavailable. Passwords must meet the following criteria: - Minimum password length of 14. - Minimum of 4 character sets. - Minimum of 1 lower case characters. - Minimum of 1 upper case characters. - Minimum of 1 special characters. - Minimum of 1 digits. - Password should not contain username. - Password should not be any of the last 25 passwords. Reset Security Baseline: Clears the stored value for customer-selected security baseline. This forces the dialog which allows you to choose a baseline to be presented at the next Admin logon.

Table 10-143 • Policies

Preset Parameter	Description
Enable Password Policies	Specify whether to enable establishing policies for acceptable passwords, This specifies Password requirements such as # of letters, numbers, symbols, etc.

Preset Parameter	Description
Enable Session Timeout (Lock Screen)	2-step logout: 1. System will display a lock screen after a set amount of time (configurable below under "session Lock Screen Timeout"). 2. System attempts to logout the user if "Automatic Logoff" is enabled and the timeout for it is reached or another user logs in instead.
Enable Session Timeout (Automatic Logoff)	This logs out a user automatically after a set amount of time.
Require Logon At Startup	Select to require logon with password at startup.

Table 10-144 • User Name Policies

Preset Parameter	Description
Display Login User List	Check this box to display a list of Users.

Table 10-145 • Password Policies

Preset Parameter	Description
Password cannot contain username	Password policy stating that the password cannot contain the user's name.
Minimum Password Length	Password policy for password minimum length.
Minimum Number of Character Sets	Password policy for minimum number of characters types (Upper-Case, Lower-Case, Digits, Symbols).
Minimum Number of Upper Case Letters	Password policy for minimum number of upper-case letters allowed (A, B, C, etc.)
Minimum Number of Lower Case Letters	Password policy for minimum allowed for number of lower-case letters (a, b, c, etc.)
Minimum Number of Digits	Password policy for minimum number of numbers (1, 2, 3, etc).
Minimum Number of Symbols (~#\$% etc.)	Password policy for minimum number of symbols allowed (#, @, etc.).
Minimum Password Age (hours)	Password policy for the minimum age for a password, in hours.
Maximum Password Age (days)	Password policy for the maximum age for a password, in days.
Minimum Changes Between Passwords	Password policy for the minimum number of changes between passwords.
Maximum Number Of Repeated Characters	Password policy for the maximum number of repeated password characters.
Maximum Number Of Sequential Characters	Password policy for the maximum number of sequential password characters.
Password Reuse History Count	Password policy -- users can't reuse old password -- # of old passwords it stores so you can't reuse it
Failed Logins Before Account Blocked	Password policy for the number of failed attempts to login that are allowed.

Preset Parameter	Description
Do Not Allow Common Passwords	Select to disallow commonly used passwords.
Account Block Time (min)	There's a policy which will block an account for a certain amount of time after a certain number of failed logins -- you can set how long that time is
Session Lock Screen Timeout (min)	Password policy for the time, in minutes, before the system will lock the screen.
Session Auto Logoff Timeout (min)	Password policy for the time, in minutes, before the system will automatically log a user off the system.

Logon Banner

The Power On Logon Banner displays the text at power-on and requires user confirmation to complete booting the system. This is user-configurable.

Updating the Logon Banner Page

You can change the text that displays when the user logs onto the LOGIQ Fortis via the **Utility > Admin > Login** screen.

There are three login screens that can be used:

- GE HealthCare Default Login Screen
- Department of Defense (DoD) Default Login Screen
- User-customized Login Screen




Figure 10-84 • GE HealthCare Default Login Window

To configure a customer-specific Login Banner page:

1. Access the **Utility > Admin > Login** screen.



Update the Logon Banner portion of this screen:



2. Type the Title into Logon Title Text.
3. Type the text into Logon Disclaimer Text.
4. Press **Save**.

To set the GE HealthCare Default as the Login window, press **Load GE HealthCare Default**. Then press **OK > Save**.



To set the DoD Default as the Login window, press **Load DoD Default**. Then press **OK > Save**.

Figure 10-85 • Default GE HealthCare Logon Banner



Figure 10-86 • Department of Defense Logon Banner

Enabling the Power On Logon Banner

To ensure the Power On Logon Banner displays and requires user confirmation to complete booting the system,



Figure 10-87 • Power On Logon Banner

1. Access the **Utility > Admin > Login** screen.
2. Checkmark **Enable Power On Logon Banner**.
3. Press **Save and Exit**.

Adding a Logon Logo

The system default logo on the LOGIQ Fortis is the GE HealthCare Logo. To add your own Logo,

1. Insert the USB Flash Drive with the Logo into a USB port on the LOGIQ Fortis.
Specifications: The 'Logo' image should be a bitmap image (.BMP) format ONLY, with a size of 128x128.
2. Navigate to the **Utility > Admin > Login** screen.
3. Select **Browse Logo**. A pop-up window appears for you to navigate to the USB Flash Drive and Logo you want to add.
4. Select the Logo. Select **OK**. Then **Save**.

Adding a "Lock Screen" Image

You can add a custom lock-screen image.

Specifications: The 'Lock Screen' image should be a bitmap image (.BMP) format ONLY, with a size of 1920x1080.

NOTE

Please select a screen that does not have high contrast or bright colors.

To add a 'Lock Screen' image,

1. Insert the USB Flash Drive with the image into a USB port on the ultrasound system.
2. Navigate to the **Utility > Admin > Login** screen.
3. Select **Lock Screen Image**. A pop-up window appears for you to navigate to the USB Flash Drive and image you want to add.
4. Select the image. Select **OK**. Then **Save**.

Groups

System rights are set by the Administrator. If you do not have rights to a feature/function, please contact the Administrator. If access is denied, a message in red will appear in the status line (such as "You do not have the required permission to perform..." or a dialog will be displayed with more information. In some cases, buttons or sections of the screen may simply be disabled if you do not have appropriate rights.




Figure 10-88 • Groups Configuration

Table 10-146 • Group List

Preset Parameter	Description
Add	This creates a new group.
Default Groups	BioMed, DiagPhys, Emergency User, Lead Sonographer, Operator, Physician, RefDoc, Sonographer, and SysAdmin.
SysAdmin	All Group Rights
BioMed	All Group Rights, except Admin. Can edit Imaging/Connectivity Utility pages. Can change the system time and date. Can access the Service Desktop. Cannot edit Admin Utility pages.
Lead Sonographer	Login, Create/Update/Delete/Transfer/Export Patient Data, Basic/Imaging Configuration, Authorize Remote Service Access, and Capture Logs. Can edit Imaging Utility pages. Cannot edit Connectivity or Admin Utility pages and cannot change the system time and date.
Sonographer	Login, Create/Update/Transfer/Export Patient Data, Basic Configuration, Authorize Remote Service Access, and Capture Logs
Physician	Login, Create/Update/Delete/Transfer/Export Patient Data, Basic/Imaging Configuration, Authorize Remote Service Access, and Capture Logs
Emergency User	Login, Create/Transfer Patient Data. Only access to the Utility System and Search Tabs.

Table 10-147 • Name and Description

Preset Parameter	Description
Id	The Id, from the Group List.

Preset Parameter	Description
Description	A description of the Group List Id.

Table 10-148 • Group Rights

Preset Parameter	Description
Rights assigned to each Group, by default	
Login	Allows users in this group to logon to the system.
Admin	Allows users in this group to have system administrator's rights. Can add/remove users and enable/disable encryption, etc.
Create Patient Data	Allows users in this group to create patient and exam data. This right is needed to register a patient and start an exam. Access to the following controls is prohibited without this right: New Patient, Register, and Save and Exit.
Update Patient Data	Allows users in this group to change patient and exam data. This right is needed to view patient and exam information stored in the patient database. Access to the following is prohibited without this right: Patient List, Exam List, Data Transfer screen, ability to delete images from the clipboard or active images screen, ability to delete reports from the Reports screen, ability to change to the USB/CD Read Only workflow, ability to see all the jobs (except one's own) in the Spooler.
Delete Patient Data	Allows users in this group to delete patient and exam data. This right is needed to use the Delete button on the Patient List and Exam List. Access to the following is prohibited without this right: Patient delete via Patient Registration and Patient List.
Transfer Patient Data	Allows users to transfer patient and exam data over DICOM. This right is needed for DICOM transfers over the network configured for Print keys and with Workflows. Access to the following is prohibited without this right: DICOM Image Storage, DICOM SR Storage, DICOM MPPS, DICOM Storage Commitment, and DICOM Print; Print button and Workflow; Koios DS in TCS; and Sent To from Patient Registration and Active Image Patient screens.

Preset Parameter	Description
Export Patient Data	Allows users to export patient and exam data to media. This right is needed to Export, Save As, USB quick Save to media and for print functionalities. Access to the following is prohibited without this right: Executing a Save As, USB Quick Store, or Video Capture via a configured Print key, EZBackup, Printing or Saving a Report As, Print the Patient List; in TCS Standard Print Button, and Save As Images controls, Data Transfer screen functionality to perform an Export or MPEGvue export.
Import Patient Data	Allows users to import patient and exam data from media. This right is needed to import from media, Worklist download, or to use the Query/Retrieve feature.
Basic Configuration	Allows access to and modification of basic Utility pages, which everyone has access to, except members of the Emergency User's Group.
Imaging Configuration	Allows modification of imaging preset Utility pages. Note, If the "Require Admin Operator Rights to Save Imaging Settings" is checked on the Utility page, then the "Imaging Configuration" right can only be used by a SysAdmin.
Advanced Configuration	Allows access to and modification of advanced configuration pages.
Device Mgmt Configuration	Allows users in this group to activate and configure Device Mgmt Cloud configuration management tool.
Authorize Remote Service Access	Allows users to authorize service engineers to connect to the system remotely and perform service tasks.
Capture Logs	Allows users in this group to capture and export a system log to monitor system performance.
Capture Logs with PHI	Allows users in this group to export log files with PHI included.
Access Service Desktop	Allows users in this group to access the Service Desktop.
Software Management	Allows users in this group to access Software Download.
ReceiveStreaming	Allows users in this group to use Data Streaming.

To set up a new group,

1. Select the Group.
2. Assign the Rights you wish this Group to have.

LDAP Configuration

To enable LDAP authentication, check the "Enable LDAP authentication" box at the top of the LDAP Configuration page.




Figure 10-89 • LDAP Configuration

Table 10-149 • Connection Configuration

Preset Parameter	Description
Enable LDAP authentication	The administrator can select to enable system credentials using LDAP
Connection Configuration:	
Directory Server	URL for the Directory Server
Port	Port for the Directory Server
SSL	SSL (Secure Sockets Layer) -- Select to enable LDAPs (LDAP over SSL).
Lookup...	Button that opens up a Query pop-up for groups, domain, LDAP server, etc.
Domain	Domain name for LDAP server (if needed).
Connect...	Select to connect to server as...
DN for users	Distinguished Name for users. The LDAP API references an LDAP object by its distinguished name (DN). A DN is a sequence of relative distinguished names (RDN) connected by commas.
...	Select to open a list of DNs supported by the server.
User caching:	
Enable caching	The administrator can select to allow the user to cache their password credentials.
Remember user ____ days	This field specifies the number of days to remember the user for caching purposes.
Cleanup cache	The user can cache credentials. When they select this, it clears out the cache of credentials.
Field mapping:	

Preset Parameter	Description
Field mapping	The mapping between attributes on the Ultrasound system and LDAP attributes. This allows you (for instance) to select which of several phone numbers that may be stored in your AD server are mapped to the Phone Number field of a user on the system or to refrain from mapping any AD element to that field.
Group mapping:	
Group mapping	Maps an LDAP group to a local group on the machine.
Selection buttons on bottom of screen:	
Reset field mapping	Resets field mapping to the factory defaults
Advanced config...	Entries in this menu should only be changed by experienced network administrators.
Certificates...	Server Certificate
Reload settings	Reloads LDAP settings.
Return to Admin	Press to go back to the Utility--> Admin configuration screens.

Lookup LDAP Servers

To look up LDAP Servers, select “Lookup” from the LDAP Configuration menu. Select the Domain from the pull-down menu, then select the LDAP Server from the list and press OK.

Logging on to the LDAP Server


To logon to the LDAP server, type your User Name and Password, then press **OK**.

Cleaning Up the Cache of Credentials

To clean up (empty) the cache of credentials, type your User Name and Password, then press **OK**.

Certificate Management

The Certificate Manager displays the system’s Intermediate and Trusted Root Certification Authorities, as noted on the Certificate Manager’s Tabs (Intermediate Certification Authorities and Trusted Root Certification Authorities). You can import or remove certificates.



Select the appropriate Certification Authority, then press “Import” to import Certificates or press “Remove” to remove Certificates.

Figure 10-90 • Certificate manager




Figure 10-91 • Certificate Import Dialog

Advanced LDAP Configuration

Advanced LDAP Configuration Settings can be set on the Advanced LDAP Configuration menu.




Figure 10-92 • Advanced LDAP Configuration

Table 10-150 • Advanced LDAP Configuration

Preset Parameter	Description
Default naming context attribute	Attribute name for the naming context.
User search filter	Search for the user using a keyword
Account name attribute	Attribute name for LDAP Attribute to be used as Account/User Name
Group search filter	Search for the group using a keyword
Group name attribute	Attribute name for the LDAP group.
Member Of attribute	Attribute name for the "Member Of".
Search timeout(s)	Limits the time to perform a search
Search page size (objects)	Limits the number of search results
Two step authentication	Under LDAP there's a option for a secondary authentication so you give it a user name and password to login to the LDAP server. (Advanced Configuration)
Enabled	Check this box to enable two-step authentication.
User distinguished name	The user's distinguished name.
User password	The user's password.
Load default settings	The user can select to load system default settings.


System Password

The System Password is the Windows Password used by the ultrasound system to automatically log into Windows. Users of the system will never need to use this password. The only reason to change the password will be if the user, for security reasons, prefers to define their own password instead of using the factory-created default password. As with all passwords, this should be treated with care and archived appropriately so that it can be provided to service personnel if necessary.

If you have changed the System Password, you must have the Current Password to change it.

To change the System Password,

1. Type your current password in the System Password field



NOTE

There is no need to enter the “current password” to change the System Password if the System Password has not been previously changed from the factory default. The same is true for the Database Password.

2. Type the new password in the New Password field.
3. Type the new password again in the Confirm Password field.
4. Press **Save** to save the new Password; then press **Exit**.

NOTE

‘Windows System password’ is the password for the underlying Windows OS user running this application. Do not change this password unless you are the system administrator of the device. This password is not needed for users of the system. It will only be needed by GEHC service in special situations. If you change the password, be sure to keep the new password secured and available if needed by GEHC service.

Disk Encryption

Disk Encryption is designed to protect data privacy and assist your organization with HIPAA/HITECH compliance. Safeguards include:

- All patient data on the system’s patient archive drive can be encrypted to provide protection in the event of a stolen device or hard drive.

**WARNING**

If you've selected "Encryption ON. Key is stored on USB / password is entered manually" encryption and both password and recovery key are lost, you will not be able to access archived patient data (images and measurements included) nor store new patient data on this system.

The only way to recover the system to allow storing patient data is to reset the entire disk, which deletes all the archived patient data on the disk.

It is strongly recommended that all the patient data be stored in PACS or backed up to media prior to encrypting the disk.

In addition, it is recommended that the recovery key be stored on a USB storage device, printed, and kept in a secure location, ensuring that it will not be lost.

**CAUTION**

Make sure the system power cable is plugged into external power. DO NOT attempt to perform the initial encryption function on battery power.

The system encrypts patient data by default (Encryption On. Disks are unlocked automatically).

1. If you wish to change the default encryption setting, select the desired Encryption Policy, then press Accept.




Figure 10-93 • Initialize System Encryption

NOTE

If you (or the Field Service Engineer) reloads system software, you will need to press "Initialize System Encryption" to encrypt the system and reset System Encryption password and preferences.

Table 10-151 • Encryption Policy Selections

Encryption Policy	Description
Encryption OFF	Patient Data will not be encrypted. Selecting “OFF” will unencrypt the drive. System drive and recovery partition will remain encrypted.
Encryption ON. Disks are unlocked automatically	System Default. Patient data is encrypted and unlocked at system boot-up. Recovery Key and Password tied to the hard drive.
Encryption On. Require Pre-Boot PIN/Password before unlocking system drives	The system will not boot until the Pre-Boot PIN or Password has been entered. Unlike other manual key entry configurations, no system functionality is available without the PIN/Password. This encryption policy is intended for high security environments or customers with specific needs. NOTE If the Pre-Boot PIN is enabled and then disabled by choosing auto-unlock or another encryption option, the system must be rebooted and authenticated with the Pre-Boot PIN once more after the change to fully disable the functionality.
Encryption ON. Key is stored on USB / password is entered manually	The system will request the encryption password or recovery key at system startup. The system is not accessible until this password or a disk recovery key is provided. Requiring a password to access the patient archive may prevent emergency usage of the system.


NOTE

If you choose to dismiss this dialog, this dialog will be offered to you again at the next login.


- You must set the Encryption Password and record the Recovery Key in order to ensure access to your institution's patient data (required if replacing the system drive, ECB Board, or reformatting the C:\ Drive).

To reset the Encryption Password, “Change password.”

Press NO if this question pop-up appears: "Password is already set on a disk. Do you want to reuse it? Press Yes to reuse existing password. Press No to delete existing password. You will be prompted to set new."



You can now update the encryption password, then press OK.



- Recovery Keys are not backed up by the system; you must record / archive the Recovery Key in order to retrieve patient data.



CAUTION

Make sure to keep the password, recovery key, and any backup of these in a secure place, not accessible for any unintended audience.



You can Show or Hide the Encryption key. Store the Recovery Key in a secure location, accessible to the ADM user as necessary.

Insert the USB flash drive into a USB port to save the recovery key.

Use the USB flash drive as the repository only for the recovery key. DO NOT use it for data archiving or DVR recording.

Save recovery key to a USB Flash Drive by pressing Save recovery keys.

View the Recovery Key by pressing Show recovery key, then print it to a local printer or PACS.

Press the **Print key** to print the recovery key on the local printer.

Press **Show recovery key** to display the recovery key on the screen.

Press **Hide recovery key** to hide the recovery key.

Change password

Press **Change Password** to change the password as necessary.

Change the recovery key

If you want to change the recovery key, select **Change recovery key** to generate a new key.

NOTE

Generating a new key causes the previous key to expire.

Disk Encryption Frequently Asked Questions

Here are answers to some frequently-asked questions:

Q What type of encryption technology is used on my ultrasound system?

A The system uses Microsoft Bitlocker FDE configured to use FIPS-Compliant encryption protocols.

Q The system was accidentally turned off or lost power during the disk encryption process.

A Restart the system, and then go to **Utility > Admin > Disk Encryption**, select “On”, ...resume, the disk encryption will continue

Q The USB recovery key was accidentally formatted. How do I access the system?

A If a password was created or the recovery key was written down, enter the information in the pop up window. Then go to **Utility > Admin > DiskEncryption** and Select Change recovery key. Be sure to print and save the new recovery key and store the USB in a secure location.

Q Is it possible to stop encrypting the data?

A **Utility > Admin > Disk encryption**, select the “Off” radio button. The process to unencrypt may take up to 15 minutes.

Q Is there a step I need to take every day to encrypt new data?

A No, the encryption will happen in the background without the need for any additional manual steps.

Q The encryption was done it took nearly 90 minutes, is this going to happen every time the data is encrypted?

A No, after the initial encryption process is completed, new data will be encrypted automatically.

Q How does this protect the patient data?

A The data is unreadable by anyone who attempts to use the system without entering the password or recovery key.

Q I am using the system portable with Power Assistant, do I have to enter the password or recovery key after each exam?

A No, the password or recovery key will only need to be entered at each full boot up.

Q I do not have Power Assistant but I take the system portable occasionally do I need to enter the recovery key or password each time?

A Yes, the recovery key or password must be entered each time the system is turned on. If the USB with Recovery key is connected with each start up you will not need to manually enter the password or recovery key.

Q The USB recovery key has been lost, how do I unlock the system?

A Manually enter the password or recovery key, then create a new USB key from the Disk Encryption page.

Q The USB recovery key and the password has been lost how do I unlock the disk?

A Contact your service, the system will need a base image and application reload. Create a new encryption key. *Note the patient data on the system is not recoverable from this scenario. Always store patient data to a PACS or backup to external media.

Q If it is not possible to unlock the data can I still scan and what cannot be used?

A Yes: Unless the “Require Pre-Boot PIN/Password” Encryption policy option is selected, Live scan, measurement, “Save As” can be used to store on other media (USB, printer, etc).

You can not access patient information, cannot access system Archive, cannot create an exam, cannot store to the hard drive, and DICOM Transfer is not possible.

Audit Report and System Log Server Configuration



Figure 10-94 • Audit Report

Table 10-152 • Audit Report

Preset Parameter	Description
Report Filter: Specify the following criteria.	
Search Type	Select the type of search.
Date Range	Specify the “From” and “To” date range.
Status	Specify the status.
Anonymize	Indicate whether to anonymize the report (remote patient data).
Generate Report	Press to generate the report.
Report Information: Generated report information.	
Report File Path	The location where to place the generated report.
Report Status	Report status.

The Audit Report System Log Server Configuration page allows you to customize where to draw information from for the Audit Report.

You are also able to connect the ultrasound system to a centralized logging database at your institution to monitor for pattern analysis. This way, if your system is compromised, the data can be analyzed for incident response.

This centralizes all the system’s data in the customer’s data center -- It’s like a DICOM connection page, but connects logs to a central location.




Figure 10-95 • System Log Server Configuration

Table 10-153 • System Log Server Configuration

Preset Parameter	Description
Connection	
IP Address of the Server	The Server's IP address
TCP Listener Port	Transfer Control Protocol Listener Port. This is the port which we transfer logging data to use the syslog data protocol. It can be done by UDP or TCP, depending on the configuration of the customer logging server.
UDP Listener Port	User Datagram Protocol Listener Port. This is the port which we transfer logging data to use the syslog data protocol. It can be done by UDP or TCP, depending on the configuration of the customer logging server.
Mode of Transfer	Mode of transferring the data to the server (UDP, TCP, or TLS).
Windows Event Log	
Type of Logs to transfer	- Application Logs - Security Logs
Source of application event log to transfer: (Add * for no filter)	- Add Source: The user needs to specify where to draw data from for the audit report. - Delete Source: Delete the event log source from the application.
Transfer	
Send Time Period (Sec):	How often the system will attempt to contact the logging server to send logs.

Preset Parameter	Description
Number of days of Historic data to send:	If logging server connectivity has not been available, how many days of history should be sent when it is restored?
EventLog Batch Send Size	When sending historic data, how much to send in each transaction – this can help manage network traffic and server load, not typically necessary to change this.

Imaging Preset Manager

The Imaging Preset Manager allows you to:

- Create and Edit User Presets
 - Update User Presets
 - Rename User Presets
 - Delete User Presets
- Arrange presets on the Touch Panel
- Share User Presets across LOGIQ Fortis systems
 - Export User Presets
 - Import User Presets
- Configure MyPreset for probe
 - Update MyPreset Config for each probe if desired. Need for each probe separately. See *Arranging MyPreset Tab* on page 167 for more information.

Creating a User-Defined Application Preset

To create a User-Defined Application Preset,

1. From the Touch Panel, select the **Probe** icon at the top of the Touch Panel.
2. Select the **Application** you want to use as a basis for the new Application Preset.
You are now ready to create your own user preset.
3. Press **Save**. A pop-up menu appears: The Create New Application menu appears.




Figure 10-96 • Create New User Pop-up Menu

4. Select **Create New**. The Create New Application menu appears.

NOTE

The new user application is based upon the current exam and application, plus any modifications you have made, including the comment library and M&A calcs.

The name of the new application cannot include spaces or symbols. However, the name can include numbers and letters.

There is no limit to the number of user-defined application presets you can create for each exam category; you do not need to map all created presets to the Touch Panel.

After you select to create the new preset, the Imaging Preset Manager screen appears. The preset you just created now appears in the Available Imaging Presets column. You'll notice that it has the name you assigned it ("GE HealthCare").

Arranging Presets on the Touch Panel

On this screen, you specify where you want the new user (and existing) presets to appear on the exam's Touch Panel screen.



Figure 10-97 • Steps to Add a User Preset

You can move the location of where the application appears on the Touch Panel via the Imaging Preset Manager (accessed from the Utility Touch Panel).

To reposition an application on the Touch Panel grid,

1. Select the Application to copy. Right click and select Create New.
2. Provide the new name.
3. The Imaging Preset Manager page appears (number not shown above).
4. Position the new application to appear at the desired Touch Panel location.
5. The new application appears on the Touch Panel.

Updating User Presets

You can edit, reset to factory default, or delete any user preset you create, as long as you have selected it in the “Available Imaging Presets” column on the left.

To view/edit the parameters for the user defined preset

1. Press **Utility > Imaging Preset Manager**. Select the user preset you want to view/edit from Available Imaging Presets Column.
2. Press **Edit**. The Imaging page appears.
3. Edit the presets as necessary and press **Save**.
or
4. Adjust the image while in the user preset you want to edit.
5. Press the **Probe**.
6. Press **Save > Overwrite [Preset Name]**.

Renaming a User Preset

To rename a user preset,

1. Press **Utility > Imaging Preset Manager**. Select the user preset you want to rename.
2. Press **Rename**. The Rename Preset pop-up menu appears.
3. Type the new name and press **Rename**.

Deleting a User Preset

To delete a user preset,

1. Press **Utility > Imaging Preset Manager**. Select the user preset you want to delete.
2. Press **Delete**. The Delete Preset pop-up menu appears.
3. Confirm that you want to delete this user preset and press **OK**.

Sharing User Presets between LOGIQ Fortis Systems

You can share the user presets you have created between LOGIQ Fortis systems by exporting/importing the preset(s) you want to share.

To move a user preset from one LOGIQ Fortis to another LOGIQ Fortis system (same software level), first export the user preset(s) you wish to share.

Exporting User Presets

To export a user preset (or presets),

1. Activate the **Imaging Preset Manager** from the Utility Touch Panel.
2. Insert the media.
3. Press **Export** (on the bottom).
4. An Export Presets pop-up menu appears that indicates:
 - destination location (USB Flash Drive/Hard Disk Drive drive location).
 - preset directory where the preset should be saved (Preset Export).
 - available presets on the scanner.

Select the name for the Preset Directory from the Preset Directory pull-down.
5. Select the User Defined Presets under Available presets on Scanner and press Export.
6. Upon a successful Export, an informational message will pop-up indicating that “1 preset successfully exported.” Press **OK**. Then press **Exit** to close the Export Presets pop-up menu.
7. Press **F3** to Eject the media. Take the media to the other LOGIQ Fortis and follow the Importing User Presets instructions below.

Importing User Presets

To import a User Preset,

1. Activate the **Imaging Preset Manager** from the Utility Touch Panel.
2. Insert the media (Flash Drive, USB Hard Drive).
3. Press Import. The Import Presets pop-up appears and displays the Source Directory and Available Imaging Presets.
4. Select the “User Defined Presets” under Available Imaging Presets and press Import.

If these presets are already on this LOGIQ Fortis, you will be asked whether you want to:

 - Overwrite this preset (Yes, Yes to All, No, or No to All).
 - Rename this preset (Type the new name and press Rename).
 - Cancel
5. Upon a successful Import, an informational message will pop-up indicating that “1 preset successfully imported.” Press Ok. Then press Exit to close the Import Presets pop-up menu.
6. Press F3 to Eject the media.

Retain Field of View

Selecting Retain Field of View ensures that the Imaging Parameters shown in the table below stay constant over Probe and Preset changes.

Table 10-154 • Retain Field of View

Mode	Probe	Retain Field of View Imaging Parameters
B-Mode, Harmonics, Contrast and B-Flow	Convex and Sector	Depth, Tilt, Zoom, Width
	Linear	Virtual Convex, Zoom, Depth, Steer
Color Flow Mode	Convex and Sector	ROI Size/Position
	Linear	ROI Size/Position, CF Virtual Convex, CF Steer
Doppler Mode		Doppler Cursor Position

Backup and Restore

The Backup/Restore function enables the user to copy and restore system presets, settings and service configurations, and enables the user to configure several units with identical configurations (providing the units have the same software version).

Depending on the system, you can use either a CD-R, DVD-R, USB Flash Drive, or USB Hard Disk for system backup/restore. For the sake of simplicity, we have used the CD-R in the following examples.

NOTE

The system ONLY supports CD-R / DVD-R and DOES NOT support CD-RW / DVD+R.

To minimize accidental loss of data, backup system presets, settings and service configurations, DAILY to formatted media and/or to the local hard drive (manually or automatically). Presets and service configurations can be restored to the local hard drive using the restore procedure.

NOTE

To perform backup and restore procedures, you must login with administrator privileges.

Backup

Media Backup

1. Insert media into the drive or USB device into a USB port.
2. On the Touch Panel, press **Utility**.
3. On the Utility Touch Panel, press **System**.
4. On the monitor display, select **Backup/Restore**.

The Backup/Restore screen is displayed.




Figure 10-98 • System/Backup/Restore Preset Menu

5. In the Backup field, select User Defined Configuration and/or Service to copy system presets, settings and service configurations.
6. Select CD/DVD or USB Drive F to save data in the Backup To/Restore From field.
7. Select **Backup**.

The system performs the backup. As it proceeds, status information is displayed on the Backup/Restore screen.

8. At the end of the process, the Backup completed message is displayed on the monitor.

Press **Eject** (F3) to eject media/disconnect USB device.

9. Make sure to physically label the media. An identification of the system should also be noted on the media and a backup log should be kept.

Store the media in a safe place.

Local Backup

Select Backup Automatically under Local Backup to automatically backup User Defined configurations to the system hard drive when the configuration is changed and saved.

Select Manual Backup to manually backup User Defined configurations to the system hard drive.

Backup Automatically

1. Check **Backup Automatically** under Local and Cloud Backup.
2. Select **Local Backup** under Backup To/Restore From.

3. Press **Save**.

User Defined Configuration files are automatically backed up to Local Backup, when configuration is changed and then saved.

Manual Backup

1. Select **Local Backup** under Backup To/Restore From.
2. Press **Backup** under Backup.

Restore from Media or Local Backup



WARNING

The restore procedure overwrites the existing database on the local hard drive. Make sure to insert the correct media. You cannot restore system presets, settings and service configurations between systems with different software versions. To minimize accidental loss of data, perform backup of the patient archives stored on the local hard drive periodically.

1. On the Touch Panel, press **Utility**.
2. On the Utility Touch Panel, press **System**.
3. On the monitor display, select **Backup/Restore**.
The Backup/Restore screen displays. (See *Figure 10-98 • System/Backup/Restore Preset Menu* on page 600.)
4. In the Restore list, select User Defined Configuration and/or Service to restore system presets, settings and/or service configurations.
5. If restoring from Media: Ensure the appropriate source device is selected in the Media field and select **Restore from Media**.
If restoring from Local Backup: Select **Restore from Local Backup**.
The system performs the restore. As it proceeds, status information displays on the Backup/Restore screen.
6. The LOGIQ Fortis restarts automatically when Restore is done.

Preset Restore User Interface Corruption

When restoring presets with the user interface in Russian, Greek, Japanese or Simplified Chinese, if the OS Region Language does not match the selected system software application language at the time of the preset restore, the user interface text may appear corrupted.

To prevent the corruption, before restoring presets, confirm that the OS Region language and system software application language match by following the procedure below, *Match the OS Language with System Software Application Language* on page 105, referring to *Figure 3-28 • Match OS Language with System Software Application Language (Preventative)* on page 105.

If the corruption has already occurred and the text on the screen is not readable, follow the same procedure (*Match the OS Language with System Software Application Language* on page 105) to correct the settings, referring to *Preset Restore User Interface Corruption* on page 104, which shows the location of the Regional Options and Language fields on the corrupted screens.

Match the OS Language with System Software Application Language

Change the default operating system language on the Utility > System > General page.

1. Select **Regional Options** to open the Regional Options dialogue box (1).
2. Select **Current OS Language** and select the OS language from the dropdown menu to match the system language (2).
3. Select **OK** (3). The system will restart.



Figure 10-99 • Match OS Language with System Software Application Language (Preventative)



Figure 10-100 • Match OS Language with System Software Application Language (Corrective)

Preset Synchronization Using Media

The procedure for preset synchronization of several scanners using media is as follows:

1. Make a backup of the user-defined configurations on a removable media from a fully configured LOGIQ Fortis system.
2. Restore user-defined configurations from the removable media to another LOGIQ Fortis system (you can restore all the user-defined presets or select specific presets to restore via Detailed Restore).

NOTE

For information on Device Mgmt cloud-based system preset configuration management see *Device Mgmt* on page 1096.

Search

Opens up a search window to find a parameter on the utility pages.

To search for a utility parameter,

1. Press **Search** from the Utility Touch Panel or from another Utility page.
2. Type in the search string. For instance, if you're searching for Zoom you would just type 'zoom'.
3. A list of possible matches appears to the right. Select the correct match.

NOTE

You cannot perform a search on the Measure, Reports, Imaging Preset Manager, Scan Assistant, or Service Utility pages.

Chapter 11

Probes and Biopsy

In this section

Probe Overview.	606
Probe Safety.	618
Care and Maintenance.	623
Biopsy Special Concerns.	653
Performing a Biopsy.	655
Surgery/Intra-operative Use.	676

This chapter consists of the information of each probe and describes some special concerns, biopsy kits and accessories as well as basic procedures for attaching a biopsy guide to the different types of probes.

Probe Overview

Ergonomics

Probes have been ergonomically designed to:

- Handle and manipulate with ease
- Connect to the system with one hand
- Be lightweight and balanced
- Have rounded edges and smooth surfaces.
- Stand up to typical wear by cleaning and disinfectant agents, contact with approved gel, etc.

Cables have been designed to:

- Connect to system with appropriate cable length

Supported Probes

The supports the following types of probes:

- Matrix Array probes
- Convex Array probes
- Linear Array probes
- Micro Convex Array probes
- Sector Phased Array probes
- Split Crystal
- Volume Probes (4D)



CAUTION

Probes for transvaginal and transrectal applications require special handling. Transvaginal/transrectal examinations and probe insertions should be performed only by personnel with adequate training. Refer to the user documentation enclosed with these probes.

Probe Type, Common Capabilities and Features

Table 11-1 • Probe Type, Common Capabilities and Features

Probe Type	Probe Names	Common Capabilities and Features		
Curved Array	BE9CS-D, C1-6-D, C1-6VN-D*, C2-7-D, C2-7VN-D*, C2-9-D, C2-9VN-D*, C3-10-D*, IC5-9-D*, Vscan Air CL	CrossXBeam	Easy 3D / Advanced 3D / Tru3D	The following are commonly used probe- independent features: LOGIQView, Scan Assistant, Compare Assistant, Speed of Sound, Color Quantification, and SRI-HD
Linear Array	L2-9-D, L2-9VN-D, L3-9i-D, L3-12-D, L4-20t-D, L6-24-D, L8-18i-D*, ML6-15-D*, Vscan Air CL	B Steer+, Virtual Convex, CrossXBeam		
Sector Phase Array	M5Sc-D*, 6S-D, 6Tc-RS	Virtual Convex		
Mechanical 4D	RIC5-9-D, RAB6-D	CrossXBeam	Static 3D, 4D, OmniView, STIC, VCI, TUI, VOCAL	
Split Crystal	P2D, P6D, P8D	CW		
	*Tru3D Capability			

Probe Description

Table 11-2 • Probe Applications and Features

Probe	Clinical Applications	Capabilities and Features	Illustration
BE9CS-D	Urology	- Easy3D/Advanced3D - PDI - M-Mode - LOGIQView - Contrast - CrossXBeam - Strain Elastography - Biopsy	

Probe	Clinical Applications	Capabilities and Features	Illustration
C1-6-D	- Abdomen (incl. Pleural) - OB/GYN - Pediatric - Peripheral Vascular (No transcranial) - General musculoskeletal	- Easy3D/Advanced3D - PDI - M-Mode - Anatomical M-Mode - LOGIQView - Contrast - B-Flow/Hybrid B-Flow - CrossXBeam - Shear Wave and Strain Elastography - UGAP - MVI/Contrast MVI - Biopsy	
C1-6VN-D		- Easy3D/Advanced3D - Tru3D - Volume Navigation - PDI - M-Mode - Anatomical M-Mode - LOGIQView - Contrast - B-Flow/Hybrid B-Flow - CrossXBeam - Shear Wave and Strain Elastography - UGAP - MVI/Contrast MVI - Biopsy	

Probe	Clinical Applications	Capabilities and Features	Illustration
C2-7-D	- Abdomen - Pediatric	- Easy3D/Advanced3D - PDI - M-Mode - Anatomical M-Mode - LOGIQView - CrossXBeam - Contrast - B-Flow - Biopsy	
C2-7VN-D		- Easy3D/Advanced3D - PDI - M-Mode - Anatomical M-Mode - LOGIQView - CrossXBeam - Contrast - V-Nav - Tru3D - B-Flow - Biopsy	
C2-9-D	- Abdomen - OB/GYN - Pediatric - Peripheral Vascular (No transcranial) - Neonatal - Neonatal transcranial - General musculoskeletal	- Easy3D/Advanced3D - PDI - M-Mode - Anatomical M-Mode - LOGIQView - CrossXBeam - Contrast - B-Flow/Hybrid B-Flow - Strain Elastography - UGAP - Biopsy	
C2-9VN-D		- Easy3D/Advanced3D - PDI - M-Mode - Anatomical M-Mode - LOGIQView - CrossXBeam - Contrast - B-Flow/Hybrid B-Flow - V-Nav - Tru3D - Strain Elastography - UGAP - Biopsy	

Probes and Biopsy

Probe	Clinical Applications	Capabilities and Features	Illustration
C3-10-D	• Neonatal transcranial • Neonatal Abdomen • Pediatric • Peripheral Vascular (No transcranial) • Small Parts • Abdomen	• Easy3D/Advanced3D • PDI • M-Mode • Anatomical M-Mode • LOGIQView • CrossXBeam • Contrast • B-Flow • V-Nav • Tru3D • Volume Navigation	
IC5-9-D	• OB/GYN • Urology	• Easy3D/Advanced3D • PDI • M-Mode • Anatomical M-Mode • LOGIQView • Contrast • Shear Wave and Strain Elastography • V-Nav • Tru3D • Biopsy • Volume Navigation	

Probe	Clinical Applications	Capabilities and Features	Illustration
L2-9-D	- Peripheral Vascular (No transcranial) - Abdomen (incl. Pleural) - OB/GYN - Small Parts - Pediatrics - Neonatal - Neonatal transcranial - General musculoskeletal	- Easy3D/Advanced3D - PDI - M-Mode - LOGIQView - Virtual Convex - Contrast - CrossXBeam - B-Flow/Hybrid B-Flow - MVI/Contrast MVI - Shear Wave and Strain Elastography - Biopsy	
L2-9VN-D	Superficial musculoskeletal	- Easy3D/Advanced3D - PDI - M-Mode - LOGIQView - Virtual Convex - Contrast - CrossXBeam - B-Flow/Hybrid B-Flow - MVI/Contrast MVI - Shear Wave and Strain Elastography - V-Nav - Tru3D - Biopsy	
L8-18i-D	- Breast - Intraoperative: Peripheral Vascular - General Musculoskeletal and Superficial Musculoskeletal - Neonatal Abdomen - Neonatal Transcranial - Peripheral Vascular (No transcranial) - Small Parts	- Easy3D/Advanced3D - PDI - M-Mode - LOGIQView - Virtual Convex - Contrast - CrossXBeam - B-Flow/Hybrid B-Flow - MVI - Shear Wave and Strain Elastography - V-Nav - Tru3D - Volume Navigation	

Probes and Biopsy

Probe	Clinical Applications	Capabilities and Features	Illustration
L3-9i-D	Intraoperative: Abdominal	- M-Mode - CF/PDI - PW - BSteer+ - Virtual Convex - CrossXBeam - Contrast - Strain Elastography	
L6-24-D	- General Musculoskeletal and Superficial Musculoskeletal - Small Parts - Abdomen - Neonatal Abdomen - Neonatal Transcranial - Peripheral Vascular (No transcranial)	- LOGIQView - Virtual Convex - CrossXBeam - B-Flow/Hybrid B-Flow - MVI - CF/PDI - PW - Strain and Shear Wave Elastography - BSteer+ - Strain Elastography - Shear Wave Elastography - Available on L6-24-D probes with the probe name underlined on the probe connector label ("<u>L6-24-D</u>")	
L3-12-D	- Vascular - Abdomen (incl. Pleural) - OB - Small Parts - General musculoskeletal - Superficial musculoskeletal - Neonatal - Neonatal transcranial - Pediatrics	- Easy3D/Avanced3D - PDI - M-Mode - Anatomical M-Mode - LOGIQView - Virtual Convex - Contrast - CrossXBeam - B-Flow/Hybrid B-Flow - MVI/Contrast MVI - Strain and Shear Wave Elastography - Biopsy	

Probe	Clinical Applications	Capabilities and Features	Illustration
L4-20t-D	- General musculoskeletal - Superficial musculoskeletal - Peripheral Vascular (No transcranial)	- M-Mode - MVI - CF/PDI - PW - Virtual Convex - LOGIQView - CrossXBeam - BSteer+ - Biopsy	
ML6-15-D	- Abdomen - Small Parts - Peripheral Vascular (No transcranial) - Pediatrics - Neonatal - Neonatal transcranial - General musculoskeletal - Superficial musculoskeletal	- Easy3D/Advanced3D - PDI - M-Mode - LOGIQView - Virtual Convex - Contrast - CrossXBeam - B-Flow/Hybrid B-Flow - MVI/Contrast MVI - Shear Wave and Strain Elastography - V-Nav - Tru3D - Biopsy - Volume Navigation	
M5Sc-D	- Adult cardiac - Pediatric cardiac - Adult cephalic - Abdomen (incl. Pleural)	- Easy3D/Advanced3D - PDI - M-Mode - Anatomical M-Mode - Curved Anatomical M-Mode - Color M - LOGIQView - Virtual Convex - Contrast - Anatomical M-Mode - B-Flow - CW - TVI/TVD - Volume Navigation - Tru3D - Biopsy	

Probe	Clinical Applications	Capabilities and Features	Illustration
6S-D	- Pediatric cardiac - Pediatric abdomen (incl. Pleural)	- Easy3D/Advanced3D - PDI - M-Mode - Anatomical M-Mode - Curved Anatomical M-Mode - Color M - PDI - LOGIQView - Virtual Convex - Anatomical M-Mode - CW - TVI/TVD	
RAB6-D	- OB/GYN - Abdomen - Pediatric - Neonatal	- PDI - M-Mode - Anatomical M-Mode - LOGIQView - Contrast - CrossXBeam - Realtime 4D - Static3D - Biopsy	
RIC5-9-D	- OB/GYN - Urology	- PDI - M-Mode - Anatomical M-Mode - LOGIQView - Contrast - CrossXBeam - Realtime 4D - Static3D - BetaView - Biopsy	
6Tc-RS	Adult cardiac	- CW - Virtual Convex - PDI - M-Mode - Anatomical M-Mode - Curved Anatomical M-Mode - Color M - TVI/TVD	

Probe	Clinical Applications	Capabilities and Features	Illustration
6TC-RS Adapter			
P2D	• Adult cardiac • Pediatric cardiac • Peripheral vascular • Adult cephalic	• CW	
P6D	• Adult cardiac • Pediatric cardiac • Peripheral vascular • Adult cephalic	• CW	
P8D	• Adult cardiac • Pediatric cardiac • Peripheral vascular • Adult cephalic	• CW	

Beta View

Beta View enables you to steer the probe head in elevation direction without moving the probe. This feature is available on the 4D RIC5-9-D probe during live scanning and is especially useful during an endovaginal or neonatal head exams.

The Beta View Touch Panel control appears on the B-Mode Touch Panel when you select either of these two probes. You can adjust the Beta View control Right/Left or Up/Down. When you press down on the Beta View control, the probe head re-centers itself.

Beta View is not available while the image is Frozen, during an image Recall, during a Biopsy procedure, or during Volume Navigation. When these controls are selected, the Beta View Touch Panel control is hidden and unavailable.

L4-20t-D Probe Buttons

Certain probes are equipped with buttons on the probe handle that allow assigning functions without accessing the Touch Panel or front panel.

To turn on the probe buttons and assign functions to them, select "L4-20t" in the Utility > Imaging menu. **Select Enable Probe Button** and select the desired User Defined Probe Button selections (see *Figure 11-1 • Assigning Different Functions to the Probe Buttons* on page 616).



Figure 11-1 • Assigning Different Functions to the Probe Buttons



CAUTION

Button 1 may not be accessible with out-of-plane needle guides.



HINTS

Holding a button for few seconds shows probe buttons helper.



Figure 11-2 • Probe Buttons Helper



HINTS

Holding two buttons (for example, button 1 and 2 / button 3 and 4) for a few seconds enables the probe buttons function. To disable the probe buttons, hold down two buttons. The system shows the current probe buttons status on the status bar.

Probe Description

See *Probe Description* on page 607 for the L4-20t-D Operating Modes and Probe Applications and Features.

Probe Immersion Level

See *Figure 11-14 • Probe Immersion Levels, 1 = Fluid Level* on page 639 for the L4-20t-D probe immersion level for cleaning/disinfecting.

Probe orientation

Each probe is provided with an orientation marking. This mark is used to identify the end of the probe corresponding to the side of the image having the orientation mark on the display.





Figure 11-3 • Orientation Marking on Probe (Example)



1. Orientation Mark

Figure 11-4 • Probe orientation marker on the display

Probe Naming Conventions

Table 11-3 • Probe Naming Convention

Real Time 4D	Type	Application	Frequency	Connector Type
"R"	C=Convex L=Linear M=Matrix S=Sector	AB=Abdominal IC=Intracavitary	"1-5"	D=DLP RS= RS Connector with RS-DLP Adapter LC=Long Cord

Probe Safety

Care and Handling

The following recommendations help to reduce preventable probe damage.



WARNING

Failure to follow the precautions listed in the Probe Care Recommendations table can result in serious injury and/or equipment damage.



WARNING

3D/4D probes should not be used for continuous scanning for more than 25 minutes. Continuous scanning is defined as scanning without freezing the image. 3D/4D probes can be used for more than 25 minutes during an exam in cases where imaging is periodically frozen.

NOTE

If 3D/4D probes are operated in continuous 4D mode for an unusually extended period of time, the surface temperature of the handle might get warm and exceed the limit specified in IEC60601-1. The temperature of the applied part will stay within the limits according to IEC60601-2-37.

Table 11-4 • Probe Care Recommendations

Do:	Don't Do:
Handle all probes with extreme care.	DO NOT drop or knock the probe or probe lens. Impacting the probe lens face can cause fractures of the crystal elements leading to failure.
Ensure that connected probes are placed in the probe holder yoke when not in use. Be sure to utilize the endocavitary probe holder and the probe inserts for the 3D and small aperture probes that were provided at delivery.	DO NOT leave probes in places where they may be knocked over or dropped.
Use wall-mounted probe holders with lens facing upward.	Ultrasonic cleaning IS NOT approved for GE HealthCare probes.
Visually inspect probes and cables for damage prior to connecting them to the ultrasound system. If a probe appears to be damaged, discontinue use and notify your GE HealthCare Customer Service Representative. Possible damage may include, but is not limited to: • Bent or broken probe pins • Cable cuts or splitting • Surface cracks • Exposed wires or shielding • Fluid leaks	DO NOT drop probes into holders or into disinfectant containers with the probe's lens face down. Even a short drop can damage a probe.

Do:	Don't Do:
Disconnect probes from the system prior to cleaning or disinfecting the probe.	DO NOT let probe cables dangle loosely from the ultrasound system where they might be caught in the wheels/casters while moving.
Ensure you follow the chemical manufacturer recommendations regarding the use and handling of the chemical.	DO NOT immerse probes deeper than permissible levels. NEVER immerse the connector or adapter into any liquid.
Always use bite guards with Transesophageal (TEE) probes.	DO NOT apply excessive bending or pulling force to the probe cable.
Clean and disinfect all probes following the procedures contained in this manual.	DO NOT kink, tightly coil, or apply excessive force on the probe cable or TEE shaft. Insulation failure may result.

Handling precautions



DANGER

Ultrasound probes are highly sensitive medical instruments that can easily be damaged by improper handling. Use care when handling and protect from damage when not in use. DO NOT use a damaged or defective probe. Failure to follow these precautions can result in an increase probability of disease progression, injury and equipment damage.



CAUTION

Transesophageal, endocavity and intraoperative probes require a special handling. Refer to the user documentation enclosed with these probes.

Transesophageal probe safety



WARNING

Never use excessive force when manipulating the transesophageal probe. The detailed operator manual enclosed with the transesophageal probe must be read carefully. Failure to follow these instructions could lead to patient injury and/or equipment damage.



CAUTION

TEE probes should always be used with an appropriate Bite Guard in place.

Electrical Shock Hazard



ELECTRICAL HAZARD

The probe is driven with electrical energy that can injure the patient or user if live internal parts are contacted by conductive solution:

- DO NOT immerse the probe into any liquid beyond the level indicated by the immersion level diagram. Refer to the immersion illustration in the Probe Cleaning Process section. Never immerse the probe connector or probe adaptors into any liquid.
- DO NOT drop the probes or subject them to other types of mechanical shock or impact. Degraded performance or damage such as cracks or chips in the housing may result.
- Prior to each use, visually inspect the probe lens and case area for cracks, cuts, tears, and other signs of physical damage. DO NOT use a probe which appears to be damaged until you verify functional and safe performance. You must perform a more thorough inspection, including the cable, strain relief, and connector, each time you clean the probe.
- Before inserting the connector into the probe port, inspect the probe connector pins. If a pin is bent, do not use the probe until it has been inspected and repaired/replaced by a GE HealthCare Service Representative.
- DO NOT kink, tightly coil, or apply excessive force on the probe cable. Insulation failure may result.
- Electrical leakage checks should be performed on a routine basis by GE HealthCare Service or qualified hospital personnel. Refer to the service manual for leakage check procedures.

Special handling instructions

Using protective sheaths



WARNING

Protective barriers may be required to minimize disease transmission. Probe sheaths are available for use with all clinical situations where infection is a concern. Use of legally marketed, sterile probe sheaths is mandatory for intra-cavitary and intra-operative procedures. Failure to follow these instructions could lead to exposure to infectious agents.



WARNING

Devices containing latex may cause severe allergic reactions in latex sensitive individuals. Refer to FDA's March 29, 1991 Medical Alert on latex products.



WARNING

DO NOT use an expired probe sheath. Before using probe sheaths, verify whether the term of validity has expired. Failure to follow these instructions could lead to exposure to infectious agents.

**CAUTION**

Do not use pre-lubricated condoms as a sheath. In some cases, they may damage the probe. Lubricants in these condoms may not be compatible with probe construction.

Instructions. Custom made sheaths are available for each probe. Each probe sheath kit consists of a flexible sheath used to cover the probe and cable and elastic bands used to secure the sheath.

Sterile probe sheaths are supplied as part of biopsy kits for those probes intended for use in biopsy procedures. In addition to the sheath and elastic bands, there are associated accessories for performing a biopsy procedure which are included in the kit. Refer to the biopsy instructions for the specific probes in the Discussion section of this chapter for further information.

Reordering. To reorder sheaths, please contact your local distributor or the appropriate support resource.

Endocavitary Probe Handling Precautions

If the disinfectant solution comes out of the endocavitary probe, please follow the cautions below.

**WARNING**

Sterile/sanitary sheaths are to be used on the probe during its actual use with patients. Wearing gloves protects the patient and operator. Failure to follow these instructions could lead to exposure to infectious agents.

**WARNING**

Disinfectant Exposure to Patient (e.g., Cidex)—Contact with a disinfectant to the patient's skin or mucous membrane may cause an inflammation. If this happens, refer to the disinfectant's instruction manual.

Disinfectant Exposure from Probe Handle to Patient (e.g., Cidex)—DO NOT allow the disinfectant to contact the patient. Only immerse the probe to its specified level. Ensure that no solution has entered the probe's handle before scanning the patient. If disinfectant comes into contact with the patient, refer to the disinfectant's instruction manual.

Disinfectant Exposure from Probe Connector to Patient (e.g., Cidex)—DO NOT allow the disinfectant to contact the patient. Only immerse the probe to its specified level. Ensure that no solution has entered the probe's connector before scanning the patient. If disinfectant comes into contact with the patient, refer to the disinfectant's instruction manual.

Endocavitary Probe Point of Contact—Refer to the disinfectant's instruction manual.

Failure to follow these instructions could lead to inflammation of skin or mucous membrane.

NOTE

Sporadically, silicone grease can leak in small amounts from the probes' cable bushing. This leakage is not a failure and is not harmful to the human body. Silicone grease does not contain any hazardous substances and is only use to seal the cable bushing. In the case of a leakage, wipe the grease with a cloth.

Probe handling and infection control

**WARNING**

ALWAYS clean and disinfect the probe according to the probe specific instructions, including probe compatible chemicals between patients to the level appropriate for the type of examination and use FDA-cleared probe sheaths where appropriate. Failure to follow these instructions could lead to exposure to infectious agents.

**WARNING**

Adequate cleaning and disinfection are necessary to prevent disease transmission. It is the responsibility of the equipment user to verify and maintain the effectiveness of the infection control procedures in use. Always use sterile, legally marketed probe sheaths for intra-cavitary and intra-operative procedures.

**WARNING**

To minimize the risk of infection from blood-borne pathogens, you must handle the probe and all disposables which have contacted blood, other potentially infectious materials, mucous membranes, and non-intact skin in accordance with infection control procedures. You must wear protective gloves when handling potentially infectious material. Use a face shield and gown if there is a risk of splashing or splatter.

This information is intended to increase user awareness of the risks of disease transmission associated with using this equipment and provide guidance in making decisions directly affecting the safety of the patient as well as the equipment user.

Diagnostic ultrasound systems utilize ultrasound energy that must be coupled to the patient by direct physical contact. Depending on the type of examination, this contact occurs with a variety of tissues ranging from intact skin in a routine exam to recirculating blood in a surgical procedure. The level of risk of infection varies greatly with the type of contact.


One of the most effective ways to prevent transmission between patients is with single use or disposable devices. However, ultrasound transducers are complex and expensive devices that must be reused between patients. It is very important, therefore, to minimize the risk of disease transmission by using barriers and through proper processing between patients.

Care and Maintenance

Inspecting probes

**WARNING**

If any damage is found, do not use the probe until it has been inspected and repaired/replaced by a GE HealthCare Service Representative. Failure to follow these precautions can result in injury and equipment damage.



- 1 Casing
- 2 Strain Relief
- 3 Seal
- 4 Lens

Figure 11-5 • Probe parts

Perform After Each Use

Inspect the probe's lens, cable, casing, and connector. Look for any damage that would allow liquid to enter the probe.

NOTE

Keep a log of all probe maintenance, along with a picture of any probe malfunction.

Environmental Requirements

**CAUTION**

Note that the environmental conditions of the ultrasound system and the probe may differ.

**CAUTION**

Ensure that the probe face temperature does not exceed the normal operation temperature range.



CAUTION

Check the room temperature before you use the 4D probe.

Probes should be operated, stored, or transported within the parameters outlined below.

NOTE

See provided TEE probe user manual for the environmental requirements of the TEE probe.

Table 11-5 • 2D Probe Environmental Requirements

	Operational	Storage	Transport
Temperature	10 to 35 degrees C	0 to 50 degrees C	0 to 50 degrees C
Humidity	30 - 85 % non-condensing	30 - 85% non-condensing	30 - 85% non-condensing
Pressure	700 - 1060hPa	700 - 1060hPa	700 - 1060hPa

Table 11-6 • L2-9-D and L2-9VN-D Probe Environmental Requirements

	Operational	Storage	Transport
Temperature	5 to 40 degrees C	-40 to 55 degrees C	-40 to 55 degrees C
Humidity	10 - 90 % non-condensing	5 - 95% non-condensing	5 - 95% non-condensing
Pressure	700 - 1060hPa	700 - 1060hPa	700 - 1060hPa

Table 11-7 • 4D Probe Environmental Requirements

	Operational	Storage	Transport
Temperature	18 to 30 degrees C 64 to 86 degrees F	-10 to 50 degrees C 14 to 122 degrees F	-10 to 50 degrees C -40 to 122 degrees F
Humidity	Max 75% non-condensing	Max 90% non-condensing	Max 90% non-condensing
Pressure	700 - 1060hPa	700 - 1060hPa	700 - 1060hPa

Probe Processing

To prevent transmission of disease, adequate cleaning and disinfection are necessary before first use, between patient use, and every time the probe is stored or transported in a non-clean container like shipping case.

All probes must be thoroughly cleaned prior to disinfection.

The required level of disinfection is based on the type of tissue the probe will contact during use. These recommendations suit the Spaulding classification for intended use, and federal guidelines from FDA¹ and CDC² for proper processing of reusable medical instruments. Refer to *Table 11-8 • Care Methods by type of tissue* on page 625.

Table 11-8 • Care Methods by type of tissue

Type of tissue	Care Method
Intact skin only	Cleaning followed by Intermediate-Level Disinfection (ILD) (wipe or spray)
Mucosal or Non-intact skin	Cleaning followed by High-Level Disinfection (HLD) (soaking or wipe or use of a trophon® EPR or trophon®2)

NOTE

According to FDA Guidance and CDC Guideline, ILD have HBV, HIV, and tuberculocidal claims. HLD refer to FDA-cleared high-level disinfectants.

NOTE

There are no adequate means to disinfect a probe that has been contaminated by prions, such as Creutzfeldt Jacob's disease. In this case, the contaminated probe **MUST BE** discarded in accordance with local biologic waste hazard procedures.

The following picture indicates the appropriate care method of the different portions of the probe.

Table 11-9 • Care Methods by Probe Area

Probe	Care Method	Picture
Connector (1)	Cleaning only	
Cable (2)	Processed using cleaning/ disinfectant wipes	
Probe (3)	Cleaning followed by appropriate level of disinfection	

The processing instructions provided in this document have been validated per ISO 17664: 2021 to properly prepare GE HealthCare Ultrasound probes for re-use, with the chemicals mentioned in *Table 11-10 • Products with validated efficacy performed by GE HealthCare* on page 625. It remains the responsibility of the processor to ensure that the processing is performed as specified in this document. This may require verification and routine monitoring of the process.

Table 11-10 • Products with validated efficacy performed by GE HealthCare

Product Type	Trade Name	Manufacturer	Minimum Contact Time	Active Ingredient
Cleaning (Wipe)	Oxivir® Tb	Diversey	N/A	Hydrogen Peroxide

¹ FDA Guidance, Reprocessing Medical Devices in Health Care Settings: Validation Methods and Labeling, March 17 2015

² CDC – Guideline for Disinfection and Sterilization in Healthcare Facilities, 2008

Product Type	Trade Name	Manufacturer	Minimum Contact Time	Active Ingredient
Cleaning (Soak)	Enzol®	Advanced Sterilization Products® (J&J)	1-Minute Soak	Proteolytic Enzymes
	(Cidezyme®)			
	MetriZyme™	Metrex™		
	Prolystica® 2X Concentrate Presoak & Cleaner	Steris		
Intermediate-level Disinfectant (wipe or liquid)	Oxivir® Tb	Diversey	10-Minute Exposure	Hydrogen Peroxide
Intermediate-level Disinfectant (wipe)*	Sani-Cloth AF3	PDI	3-Minute Exposure	Quaternary ammonium
High-level Disinfectant (Soak)	Cidex® OPA (FDA-cleared)	Advanced Sterilization Products (J&J)	12-Minute Soak	Ortho-phthalaldehyde
	McKesson OPA/28	McKesson	10-minute soak	
High-level Disinfectant (wipe)**	Tristel Trio	Tristel	1-minute Exposure	Chlorine dioxide
	Tristel Duo ULT			
* For 12S series and 6S series, Sani-Cloth AF3 is validated				
** Tristel Trio and Tristel Duo ULT are validated only for RIC5-9A-RS, RIC5-9-D, RIC5-9W-RS and RIC6-12-D probes				

To download the latest revision of the probes processing manual, visit the Customer Documentation Portal and type reference 5661328 in the search bar.

Table 11-11 • Customer Portal

Customer Documentation Portal
https://www.gehealthcare.com/documentation

Probe Care Cards and Website

The Care Card, supplied with every probe, contains a list of additional processing products that are compatible with probe materials, but GE HealthCare has not validated their efficacy. When using these products, refer to their manufacturer's instructions for use.

A complete list of compatible products is also available on the Ultrasound Probes website. This list updates periodically with new products and probes. (Refer to *Table 11-12 • Documentation Website* on page 627).

Table 11-12 • Documentation Website

Ultrasound Probes Website
https://www.gehealthcare.com/transducers

Probe Safety Symbols

Pictograms which may be included on probe care cards.

Symbols

Symbol	Meaning	Standard
	Signifies a general warning	ISO 7010-W001
	Consult accompanying documents. This symbol advises the user to consult the accompanying documents.	ISO 7010-M002
	Describes precautions necessary to prevent the risk of injury through electric hazards. Warning, electricity	ISO 7010-W012
	Describes precautions necessary to prevent the risk of disease transmission or infections.	ISO 7010-W009
	Manufacturer name and address	ISO 7000-3082
	Indicates that the product is in compliance with all relevant European Directives and under surveillance by Notified Body XXXX	N/A- by certification body
	Symbol indicates useful information	N/A- by GE HealthCare
	Use care when handling ultrasound probes and protect the probe head from damage.	N/A- by GE HealthCare

Symbol	Meaning	Standard
	Do not immerse the probe into any liquid beyond the level specified for that probe. Refer to the Instructions for Use of the ultrasound system.	N/A- by GE HealthCare
	This is to illustrate compatible cleaners or disinfectants available in wipes format (to be used according to instruction from the manufacturers of these products)	N/A- by GE HealthCare
	This is to illustrate compatible cleaners or disinfectants available in liquid format (to be used according to instruction from the manufacturers of these products)	N/A- by GE HealthCare
	This is to illustrate compatible ultrasound coupling gels	N/A- by GE HealthCare
	This is to illustrate compatible cleaners or disinfectants available in spray format (to be used according to instruction from the manufacturers of these products)	N/A- by GE HealthCare
	This is to illustrate compatible cleaners or disinfectants available in powder (to be used according to instruction from the manufacturers of these products)	N/A- by GE HealthCare
	This is to illustrate compatible automated reprocessors (to be used according to instruction from the manufacturers of these products)	N/A- by GE HealthCare

Probe Pre-Treatment at the Point of Use

The pre-treatment step is for removal of the protective sheath, if used, the gel and gross contamination.

After each use, remove protective sheath, if used, from the probe and remove the coupling gel by wiping from the cable strain relief to the acoustic lens end (i.e. from cleanest to dirtiest area), with a soft, low-/non-linting cloth or cleaning wipe.



CAUTION

USE non-abrasive cloth or wipe, such as Kimwipes™ Delicate Task Wipers or equivalent.

DO NOT use a twisting motion when wiping the probe. To extend the life of the probe acoustic lens, pat dry only.

Preparation before Cleaning

Some probes need to be disassembled before cleaning and disinfection. Either follow the disassembly instructions mentioned in your ultrasound system user manual or pictures below.

E7C8L-RS



P2D/P2D-RSe



P6D/P6D-RS



Probe Manual Cleaning

Thorough cleaning is a mandatory first step to allow adequate subsequent disinfection or sterilization. **Choose the most convenient method, either the wipe or enzymatic soak.**



WARNING

DO NOT clean the probe in an automated washer-disinfector, due to the possible damage of the connector/console interface.

Use necessary precautions (e.g. gloves, face screen and gown), as directed by your facility.

Cleaning with Wipes

1. Hold the probe by the handle near the cable strain relief. DO NOT suspend or hold the probe by the cable as this may damage the probe.
2. Dispense a cleaning wipe from the wipe canister.
3. Gently wipe the probe with a cleaning wipe from the cable strain relief to the acoustic lens (i.e. from cleanest to dirtiest area). Gently wipe the probe's acoustic lens.

NOTE

Pay special attention to acoustic lens, edges, and crevices, removing all gel, product, and body fluids.

4. Turn the probe and continue wiping until the entire surface of the probe has been wiped. As the wipe becomes visibly soiled, discard the wipe into clinical trash and dispense fresh wipes as needed.

As needed for additional focused cleaning to crevices, wrap a clean wipe around a soft nylon bristle brush or other suitable instrument to access crevices, such as biopsy notches.

5. Visually inspect the probe for any remaining soil and, if necessary, repeat steps 3 and 4 until the probe is visibly clean.

NOTE

Clean the probe holder of the ultrasound system before returning the probe back to the system (refer to the probe holder cleaning instruction in the ultrasound system user manual for details).

6. Thoroughly dry the probe using a clean, low-/non-linting, soft cloth or wipe. Pat dry acoustic lens.

Cleaning with Enzymatic Detergent

1. Ensure the probe has been disconnected from the console.

Use necessary precautions (e.g. gloves, face screen and gown), as directed by your facility.

Fill a sink or basin with warm utility water (30 - 40°C) to a level allowing immersion of the probe up to the immersion line.

Example of a probe immersion diagram is shown in *Figure 11-6 • Examples of Probe Immersion levels* on page 631 below. Refer to the specific immersion diagram in your ultrasound system user manual.



1 Fluid level

Figure 11-6 • Examples of Probe Immersion levels

2. Prepare the cleaning solution in accordance with the detergent manufacturer's instructions.
3. Immerse the probe in the prepared cleaning solution, for the minimum contact time mentioned in *Table 11-10 • Products with validated efficacy performed by GE HealthCare* on page 625, and up to the immersion line. Ensure no air bubbles are trapped on the probe's surface.



CAUTION

DO NOT submerge probe beyond the immersion line shown in the System user manual.

NOTE

Over-exposing ultrasound probes to cleaning solution may damage the ultrasound probe.

NOTE

For IC5-9-D, E8C, and E8C-RS probes with serial numbers listed below or greater, refer to *Figure 11-7 • Probe Immersion Level for IC5-9-D, E8C, and E8C-RS with serial numbers listed above* on page 631 for the soaking level:

- IC5-9-D: 780333WX1 or greater (example 780334WX1, 780335WX1)
- E8C: 783101WX9 or greater (example: 783102WX9, 783103WX9)
- E8C-RS: 782439WX4 or greater (example 782440WX4, 782441WX4)

The prefix number, i.e. 780333 for 780333WX1, indicates serial number sequence.



1 Serial number location

Figure 11-7 • Probe Immersion Level for IC5-9-D, E8C, and E8C-RS with serial numbers listed above

4. While immersed in the cleaning solution, brush with a clean, soft, nylon bristle brush. Maintain the immersion level while brushing. This step is critical to ensure cleaning efficacy.



CAUTION

Do not use the brush on the acoustic lens.




Figure 11-8 • Cleaning the probe using a brush

5. Visually inspect the probe for soil. Repeat Steps 3 and 4 until all visible soil has been removed from the surface of the probe.
6. Rinse the probe under running warm utility water (30 - 40°C) for not less than 2 minutes, while scrubbing the surface of the probe with a clean, soft, nylon bristle brush from the base of the cable strain relief to the acoustic lens.



CAUTION

DO NOT use the brush on the acoustic lens.

NOTE

Discard solutions and rinse waters in accordance with local regulations.

7. Visually inspect the device in a well-lit area to ensure all surfaces are free from residual cleaning solution. Repeat Step 6 if visible cleaning solution is observed.
8. Thoroughly dry the probe using a clean, low-/non-linting, soft cloth or wipe. Pat dry acoustic lens.

**CAUTION**

USE non-abrasive cloth or wipe, such as Kimwipes™ Delicate Task Wipers or equivalent.

DO NOT use a twisting motion when wiping the probe. To extend the life of the probe acoustic lens, pat dry only.

NOTE

Clean the probe holder of the ultrasound system before returning the probe back to the system (refer to the probe holder cleaning instruction in the ultrasound system user manual for details).

Cable and Connector Cleaning/Disinfection

The connector can be cleaned with a wipe dampened with alcohol. Use caution when cleaning the connector, wring wipe to remove excess of liquid before wiping the connector. Prevent introduction of foreign objects in the system connector assembly. Do not apply excessive force on any component of the system connector.

**WARNING**

Exposure to excessive moisture will result in damage to the probe and possibly the ultrasound console. DO NOT wet the connector/console interface surface or labels (Refer to red circles in picture below). DO NOT clean the probe in an automated washer-disinfector.



The cable should be processed using cleaning/disinfectant wipes. If the cable has been in contact with risk factors, such as blood and/or mucous, cleaning should be followed by disinfection.

1. Dispense a cleaning/disinfectant wipe from the wipe canister
2. Wipe the cable with a cleaning/disinfectant wipe from the handle strain relief to the connector strain relief (refer to *Figure 11-9 • Cleaning the Probe Cable* on page 634). As the wipe becomes visibly soiled, discard the wipe into clinical trash and dispense fresh wipes as needed




Figure 11-9 • Cleaning the Probe Cable

NOTE

Some detergents and disinfectants might cause discoloration to the probe's cable.

3. Visually inspect the cable for any remaining soil and, if necessary, repeat cleaning until the cable is visibly clean.

If disinfection is needed, dispense a new cleaning/disinfectant wipe and continue wiping the cable. Use as many wipes as needed, to ensure all surfaces remain wet for the minimum required contact time mentioned in *Table 11-10 • Products with validated efficacy performed by GE HealthCare* on page 625. Discard the wipes into clinical trash. Let the cable air dry until visibly dry

4. Saturate a soft, low-/non-linting cloth with Critical Water (remove excess water, wipe should be damp but not dripping) and thoroughly wipe all surfaces of the cable to remove chemical residues. Discard the cloth into clinical trash.

NOTE

Critical Water is water that is treated (usually by a multistep treatment process that could include a carbon bed, softening, DI, and RO or distillation) to ensure that the microorganisms and the inorganic and organic material are removed from the water to an appropriate level (Refer to AAMI TIR34/ST108).

Use of this type of water will reduce the recontamination of probes during processing.

5. Let the cable air dry until visibly dry.

Probe Intermediate-Level Disinfection (ILD)

For Intermediate-Level Disinfection, choose either the spray or wipe method.

NOTE

Probes that contact only intact skin may be disinfected in this manner. All probes that contact non-intact skin, mucosal membranes (e.g., endocavitary) require High-Level Disinfection.

Probe ILD - Disinfectant Spray or Wipe

NOTE

Disinfectant exists either in pre-impregnated wipe or in spray. The spray should be sprayed onto a low-/non-linting cloth and then used in same way as a pre-impregnated wipe. In steps 1 to 4 of this section, “wipe” will then stand for a pre-impregnated wipe as well as for a low-/non-linting cloth saturated with disinfectant.

Do not spray disinfectant onto the probe directly.

Use necessary precautions (e.g. gloves, face screen and gown), as directed by your facility.

1. Dispense a new wipe
2. Holding the probe near the strain relief, wipe the acoustic lens and handle areas.

Slightly rotate the probe after each wiping pass and continue wiping until all areas of the probe and handle have been wetted.

Wring the wipe above recessed areas and ridges for dripping liquid directly onto the less accessible surfaces.

3. Using fresh wipes, repeat step 2 as many times as needed to ensure all surfaces remain wet for the minimum required contact time listed in *Table 11-10 • Products with validated efficacy performed by GE HealthCare* on page 625.




Figure 11-10 • Disinfecting the probe with slight rotation

4. Saturate a soft, low-/non-linting wipe with Critical Water (remove excess water, wipe should be damp but not dripping) and thoroughly wipe all surfaces of the probe to remove chemical residue. Discard the wipe into clinical trash.

NOTE

Critical Water is water that is treated (usually by a multistep treatment process that could include a carbon bed, softening, DI, and RO or distillation) to ensure that the microorganisms and the inorganic and organic material are removed from the water to an appropriate level (Refer to AAMI TIR34/ST108).

Use of this type of water will reduce the recontamination of probes during processing.

5. Thoroughly dry all surfaces of the probe using a soft, low-/non-linting wipe or cloth, changing wipes/cloths when necessary to ensure the probe is completely dry. Pat dry acoustic lens. Visually inspect the probe to ensure all surfaces are dry. Repeat drying steps if any moisture is visible.

6. If the probe is not immediately reused, store the probe in a manner that will protect and keep the probe from being recontaminated. Refer to the Probe Transportation and Storage section for additional information.

NOTE

Ensure that probe holder of the ultrasound system has been disinfected before returning the probe back to the system (refer to the probe holder disinfection instruction in the ultrasound system user manual for details).

Probe High-Level Disinfection (HLD)

High-Level Disinfection is required for devices that contact intact mucosal membranes or non-intact skin. High Level Disinfection can be performed using either disinfecting wipes (only for some probes), a disinfectant soaking method or an automated system such as trophon® EPR and trophon®2.



DANGER

If the probe is damaged, remove it from patient use. Clean and disinfect the probe before contacting your GE HealthCare Service Representative for inspection and repair/replacement.



WARNING

DO NOT disinfect the probe in an automated washer-disinfector, due to the possible damage of the connector/console interface.



CAUTION

After each use, inspect the acoustic lens, cable, and housing of the probe. Look for any damage that would allow liquid to enter the probe.

NOTE

Handles of semi-critical probes that are not submerged during High-Level Disinfection require at minimum Intermediate-Level Disinfection to avoid cross contamination.

NOTE

All probes must be thoroughly cleaned and dried prior to High-Level Disinfection.

Probe HLD – Soak

1. Ensure the probe has been disconnected from the console. Use necessary precautions (e.g. gloves, face screen and gown), as directed by your facility and fill a sink or basin with High-Level Disinfectant prepared in accordance with the disinfectant manufacturer's instructions to a level allowing immersion of the probe up to immersion line shown in the system user manual.

Example of probe immersion diagram is shown in *Figure 11-11 • Examples of Probe Immersion levels* on page 637 below. Refer to the specific immersion diagram in your ultrasound system user manual.



1 Fluid level

Figure 11-11 • Examples of Probe Immersion levels



CAUTION

Ensure no liquid comes into contact with the probe connector pins or labels.

2. Immerse probe in the disinfectant up to the immersion line shown in the system user manual and ensure no air bubbles are trapped on probe's surface. Ensure the probe remains in the disinfectant for at least the minimum contact time listed in *Table 11-10 • Products with validated efficacy performed by GE HealthCare on page 625.*



CAUTION

The acoustic lens/cap should not be resting against the tank/basin surface. Carefully place the probe in the basin, taking care not to damage the transducer acoustic lens. The probe may be suspended using a special support.



Figure 11-12 • Carefully suspend probe to avoid damage to the acoustic lens

For IC5-9-D, E8C, and E8C-RS probes with serial numbers listed below or greater, refer to *Figure 11-13 • Probe Immersion Level for IC5-9-D, E8C, and E8C-RS with serial numbers listed above* on page 638 for the soaking level:

- IC5-9-D: 780333WX1 or greater (example 780334WX1, 780335WX1)
- E8C: 783101WX9 or greater (example: 783102WX9, 783103WX9)
- E8C-RS: 782439WX4 or greater (example 782440WX4, 782441WX4)

The prefix number, i.e. 780333 for 780333WX1, indicates serial number sequence.



- 1 Serial number location

Figure 11-13 • Probe Immersion Level for IC5-9-D, E8C, and E8C-RS with serial numbers listed above

NOTE

Over-exposing ultrasound probes to high-level disinfectants may damage the ultrasound probe.

3. Thoroughly rinse the probe by immersing it in a large volume of Critical Water for a minimum of 1 (one) minute. Remove the probe and discard the rinse water.

Repeat rinsing two additional times, for a total of 3 (three) rinses.

Do not reuse the water. Always use fresh volumes of water for each rinse.

NOTE

Critical Water is water that is treated (usually by a multistep treatment process that could include a carbon bed, softening, DI, and RO or distillation) to ensure that the microorganisms and the inorganic and organic material are removed from the water to an appropriate level (Refer to AAMI TIR34/ST108).

Use of this type of water will reduce the recontamination of probes during processing.



WARNING

Failure to properly rinse probes with water following disinfection may cause skin irritation.

NOTE

Discard solutions and rinse waters in accordance with local regulations.

4. Thoroughly dry all surfaces of the probe using a soft, low-/non-linting wipe or cloth, changing wipes/cloths when necessary to ensure the probe is completely dry.



CAUTION

USE non-abrasive cloth or wipe, such as Kimwipes™ Delicate Task Wipers or equivalent.

DO NOT use a twisting motion when wiping the probe. To extend the life of the probe acoustic lens, pat dry only.

Visually inspect the probe to ensure all surfaces are dry, including crevices. Repeat drying steps if any moisture is visible.

5. If the probe is not immediately reused, store the probe in a manner that will protect and keep the probe from being recontaminated. Refer to the Probe Transportation and Storage section for additional information.

NOTE

Ensure that probe holder of the ultrasound system has been disinfected before returning the probe back to the system (refer to the probe holder disinfection instruction in ultrasound system user manual for details).


Probe Immersion Levels

Figure 11-14 • Probe Immersion Levels, 1 = Fluid Level

Probe HLD – Tristel Trio and Tristel Duo ULT**NOTE**

These instructions can only be used for disinfection of RIC5-9A-RS, RIC5-9-D, RIC5-9W-RS and RIC6-12-D.

1. Use necessary precautions (e.g. gloves, face screen and gown), as directed by your facility.
2. Activate a new disinfection wipe as per manufacturer's instructions (Tristel Sporicidal Wipe & Tristel Activator Foam or Tristel Dry Wipe & Tristel Duo ULT)
3. Holding the probe near the strain relief, wipe the acoustic lens and handle areas.

4. Slightly rotate the probe after each wiping pass and continue wiping until all areas of the probe and handle have been wetted.
5. Wring the wipe above recessed areas and ridges for dripping liquid directly onto the less accessible surfaces.
6. Ensure all surfaces remain wet for minimum 1 (one) minute.
7. To remove chemical residue, use the Tristel Rinse Wipe or saturate a soft, low-/non-linting wipe with Critical Water (remove excess water, wipe should be damp but not dripping) and thoroughly wipe all surfaces of the probe. Discard the wipe.

NOTE

Critical Water is water that is treated (usually by a multistep treatment process that could include a carbon bed, softening, DI, and RO or distillation) to ensure that the microorganisms and the inorganic and organic material are removed from the water to an appropriate level (Refer to AAMI TIR34/ST108).

Use of this type of water will reduce the recontamination of probes during processing.

8. Thoroughly dry all surfaces of the probe using a soft, low-/non-linting wipe or cloth, changing wipes/cloths when necessary to ensure the probe is completely dry. Pat dry acoustic lens. Visually inspect the probe to ensure all surfaces are dry. Repeat drying steps if any moisture is visible.
9. If the probe is not immediately reused, store the probe in a manner that will protect and keep the probe from being recontaminated.

NOTE

Ensure that the probe holder of the ultrasound system has been disinfected before returning the probe back to the system.

Probe HLD - trophon® EPR and trophon®2

When performing High-Level Disinfection of GE HealthCare ultrasound probes with the trophon® EPR and trophon®2, it is not necessary to disconnect the probe from the ultrasound system. The probe must be inactive (not selected) during the disinfection cycle.

1. This automated disinfection replaces the manual HLD, but manual cleaning still needs to be performed prior to automated HLD. Refer to manual cleaning instructions mentioned in this document.
2. Follow the trophon instructions for probe placement and operation of the trophon system. Incorrect positioning of the probe may lead to probe/cable damage and High-Level Disinfection not being achieved.



CAUTION

Damage to the cable may occur if improperly hung. Damage to the probe may occur if the probe is placed in contact with the trophon chamber wall. Curved probes must be correctly positioned in the chamber using the Curved Probe Positioner (CPP) supplied with the trophon® EPR system, or with the Integrated Probe Positioner (IPP) of trophon®2 system.

3. Once the trophon High-Level Disinfection cycle is complete, use necessary precautions (e.g. gloves), as directed by your facility and promptly remove the probe from the trophon machine. DO NOT store the probe in the trophon chamber.
4. Hold the probe's handle near the strain relief cable. DO NOT suspend or hold the probe by the cable, as this may damage the probe.

5. Wipe the probe from the acoustic lens to strain relief with a clean, low-/non-linting, soft cloth or wipe to remove any residual hydrogen peroxide from the probe surface.



CAUTION

USE non-abrasive cloth or wipe, such as Kimwipes™ Delicate Task Wipers or equivalent.

DO NOT use a twisting motion when wiping the probe. To extend the life of the probe lens, pat dry only.

6. If the probe is not immediately reused, store the probe in a manner that will protect and keep the probe from being recontaminated. Refer to the Probe Transportation and Storage section for additional information.

Reassembly of probes

When probes were disassembled for cleaning and disinfection, reassemble them before storage or reuse. Either follow the reassembly instructions mentioned in your ultrasound system user manual or pictures below.

E7C8L-RS



P2D/P2D-RSe



Covering the Transducer using a Protective Sheath



CAUTION

Probe sheaths are disposable and must not be reused.



CAUTION

Protective barriers may be required to minimize disease transmission, but their use does not replace cleaning and disinfection. Probe sheaths are available for use with all clinical situations where infection is a concern. Use of legally marketed, sterile probe sheaths is recommended for intra-cavitary, intra-operative, and biopsy procedures.

NOTE

Endoscopic, rectal, and transvaginal probes should be used with a single-use sterile sheath. (Marketing Clearance of Diagnostic Ultrasound Systems and Transducers, FDA June 27, 2019)". Although a non-sterile transducer cover for all semi-critical procedures (e.g., endocavitary) is the minimal recommendation, a sterile cover is preferred. (Guidelines for Infection Prevention and Control in Sonography: Reprocessing the Ultrasound Transducer, SDMS, 2019).

1. Place an appropriate amount of gel (sterile preferred) inside the protective sheath and/or on the transducer face.

NOTE

Failure to use imaging gel may result in poor image quality.

2. Insert probe into sheath, making sure to use proper clean technique. Pull cover tightly over transducer face to remove wrinkles and air bubbles, taking care to avoid puncturing the sheath.




Figure 11-15 • Applying the Sheath

NOTE

No gel was applied to the probe in these photos.

3. Secure the sheath in place.
4. Inspect the sheath to ensure there are no holes or tears. If the sheath becomes compromised, stop the procedure and replace immediately.
5. After usage, discard the sheath into clinical trash.

Probe Cleaning and Disinfecting Notes



DANGER

Ultrasound transducers can easily be damaged by improper handling and by contact with certain chemicals. Failure to follow these precautions can result in serious injury and equipment damage.

- Do not immerse the probe into any liquid beyond the level specified for that probe. Never immerse the transducer connector or probe adapters into any liquid.
- Avoid mechanical shock or impact to the transducer and do not apply excessive bending or pulling force to the cable.
- Transducer damage can result from contact with inappropriate coupling or cleaning agents:
 - Do not soak or saturate transducers with solutions containing alcohol, bleach, ammonium chloride compounds or hydrogen peroxide.
 - Avoid contact with solutions or coupling gels containing mineral oil or lanolin
 - Avoid temperatures above 60°C .
- Inspect the probe prior to use for damage or degeneration to the housing, strain relief, lens and seal. Do not use a damaged or defective probe.
- Use only the products that are listed on the Transducer website or on the Probe Care Card enclosed with the probe. In addition, refer to the local/national regulations.



WARNING

Do not steam, heat autoclave on general surface probes.



CAUTION

CREUTZFELDT-JAKOB DISEASE

This device is not indicated for neurological use. Neurological contact on patients with this disease **MUST BE** avoided. If a device/probe becomes contaminated, there is no adequate means to disinfect it. In this case, the contaminated device/probe **MUST BE** discarded in accordance with local biologic waste hazard procedures.



CAUTION

DO NOT expose the system/probe connector to any moisture or liquids.



CAUTION

Take extra care when handling the lens face of the Ultrasound transducer. The lens face is especially sensitive and can easily be damaged by rough handling. **NEVER** use excessive force when cleaning the lens face.

**CAUTION**

In order for liquid chemical disinfectants to be effective, all visible residue must be removed during the cleaning process. Thoroughly clean the probe, as described earlier before attempting disinfection.

You **MUST** disconnect the probe from the ultrasound system prior to cleaning/disinfecting the probe. Failure to do so could damage the system.

DO NOT soak probes in liquid chemical disinfectant for longer than is stated by the disinfectant instructions for use. Extended soaking may cause probe damage and early failure of the enclosure, resulting in possible electric shock hazard.

**CAUTION**

Avoid cross-contamination, follow all infection control policies established by your office, department or hospital as they apply to personnel and equipment.

**CAUTION**

- Do not use paper products or products that are abrasive when cleaning the probe. They damage the soft lens of the probe.
- Before storing the probes, ensure that they are thoroughly dry. If it is necessary to dry the probe after cleaning, blot the probe with a soft cloth.

**CAUTION**

Probes must be cleaned and disinfected before they are replaced or disposed of.

NOTE

Cleaning products should be as close to neutral PH as possible. Any gel, cleaning, or disinfectant products containing concentrations, surfactants, methanol, ethanol, benzyl or methyl alcohol, mineral oil, lubricant oil, oil-based lotions, acetone, ammonia, anhydrous ammonia, iodine, iodine compounds, acids with 5PH or greater may damage or discolor your probe. Ultrasonic cleaning is not approved for GE HealthCare probes.

NOTE

DO NOT re-use cloths or wipes. Soap, detergents, or enzymatic cleaners should be used in accordance with the manufacturer's instructions. GE HealthCare is not responsible for damage incurred during the cleaning process for products which no material compatibility evaluation has been conducted.

Probe Disinfectants

Choosing a Disinfectant

NOTE

For latest chemicals tested for compatibility, check the GE HealthCare Probe Website at the link listed in *Table 11-13 • Probe Web Link* on page 646.

[Table 11-13 • Probe Web Link](http://www.gehealthcare.com/transducers)

Ultrasound Probe Web Site
http://www.gehealthcare.com/transducers

When choosing a disinfectant, determine the required level of disinfection. If the possibility of cross-contamination or exposure to unhealthy or non-intact skin exists, then high level disinfection should be performed. Good hand hygiene practice is highly recommended to help further reduce the risk of cross-contamination.



CAUTION

Disinfectant wipes and topical spray products are not FDA cleared high level disinfectants and do not provide adequate protection should the probe become cross contaminated or in contact with unhealthy or non-intact skin. Failure to appropriately disinfectant could lead to exposure to infectious agent(s).



CAUTION

Review the probe care card that is packed with each probe.

NOTE

For additional information about cleaning and disinfection, refer to the recommendations of the Association for Professionals in Infection Control (APIC), the U.S. Food and Drug Administration (FDA), and the U.S. Centers for Disease Control (CDC). For country-specific disinfection regulations, check with your local regulatory infection control authorities.

NOTE

GE HealthCare publishes a list of material-compatible disinfectants (see below and also refer to the GE HealthCare website at http://www3.gehealthcare.com/en/Products/Categories/Ultrasound/Ultrasound_Probes). DO NOT use non-GE-approved disinfectants or products that have not been evaluated by GE HealthCare for material compatibility. Damages linked to the use of disapproved chemicals are not covered under product warranty or service contract.

Coupling Gels



CAUTION

Do not use unrecommended gels (lubricants). They may damage the probe and void the warranty.

Table 11-14 • Probe Gels

Gel	11-1101	11-1102	11-1103	11-1104	11-1105	11-1106	11-1107	11-1108	11-1109	11-1110	11-1111	11-1112	11-1113	11-1114	11-1115	11-1116	11-1117	11-1118	11-1119
Aquasonic 100	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
Clear Image	X	O	X	O	X	X	X	X	X	X	O	X	X	X	O	O	O	O	O
EcoGel 200 Ultrasound Gel	X	O	X	O	X	X	X	X	X	X	O	X	X	X	X	X	O	O	O
EcoVue Ultrasound Gel	X	O	X	O	X	X	O	X	X	X	O	X	X	X	X	X	O	O	O
Haiyin	X	O	X	O	X	O	X	X	X	X	O	X	O	X	O	O	O	O	O
Kendall Life Trace Ultrasound Gel	X	O	X	O	X	X	X	X	X	X	O	X	X	X	X	X	O	O	O
Konix Ultrasound Gel	X	O	X	O	X	X	X	X	X	X	O	X	X	X	X	X	O	O	O
MediChoice Standard Ultrasound Gel	X	O	X	O	X	X	X	X	X	X	O	X	X	X	X	X	O	O	O
Medline Ultrasound Gel/ Ultrasound Transmission Gel	X	O	X	O	X	X	X	X	X	X	O	X	X	X	X	X	O	O	O
Natural Image	O	X	O	X	O	O	O	O	O	O	X	O	O	O	O	O	O	O	X
Scan	X	O	X	O	X	X	X	X	X	X	O	X	X	X	X	X	O	O	O
Sonogel	X	O	X	O	X	X	X	X	X	X	O	X	X	X	X	X	O	O	O
Wavelength Multi-Purpose Ultrasound Gel	X	O	X	O	X	X	X	X	X	X	O	X	X	X	X	X	O	O	O

X Recommended
O Unrecommended

Applying

In order to assure optimal transmission of energy between the patient and probe, a conductive gel or couplant must be applied liberally to the patient where scanning will be performed.



CAUTION

Do not allow gel contact with eyes. If there is gel contact with the eye, flush eye thoroughly with water.

NOTE

Gel is not applied to the 6Tc-RS Transesophageal probe prior to use.

Precautions

Coupling gels should not contain the following ingredients as they are known to cause probe damage:

- Methanol, ethanol, isopropanol, or any other alcohol-based product
- Mineral oil
- Iodine
- Lotions
- Lanolin
- Aloe Vera
- Olive Oil
- Methyl or Ethyl Parabens (para hydroxybenzoic acid)
- Dimethylsilicone
- Polyether glycol based
- Petroleum

Gel Warmer

The gel warmer allows three temperature set points: low, medium and high. The set points are the temperatures the warmer is designed to meet. The high and low limits are the minimum and maximum values the warmer allows.

Table 11-15 • Gel Warmer Temperature Set Points

Range	Set Point	Low Limit	High Limit
Low	28 °C	23 °C	31 °C
Medium	33 °C	28 °C	36 °C
High	38 °C	35 °C	40 °C

Sterile Ultrasound Procedures

ONLY ultrasound gel that is labeled as sterile, is sterile.

Ensure you always use sterile ultrasound gel for those procedures that require sterile ultrasound gel.

Once a container of sterile ultrasound gel is opened, it is no longer sterile and contamination during subsequent use is possible.

V Nav Cleaning Requirements

Cleaning and Disinfecting V Nav Probe Brackets

The probe bracket SHOULD NOT be autoclaved or gas sterilized. The bracket can be sterilized with Cidex. Details are in the CIVCO reference guide that is included with the bracket kit.

Cleaning and Disinfecting Cables, and Transmitter

Periodically clean the equipment (transmitter, sensor, and cables) by wiping them down with a cloth dampened in a cleaning solution such as mild soap and water, isopropyl alcohol, or a similar acceptable cleaning solution. If the tracker's components come in contact with biological fluid or tissue, be sure to follow your organization's procedures for proper cleaning and disinfection. The transmitters and sensors are not designed to withstand autoclaving or gamma radiation. Sensors are ETO-compatible. DO NOT immerse the transmitter, sensor, or cables in liquids. Components are not waterproof.

Disinfecting and Sterilizing General Purpose Sensor

High-level disinfect general purpose sensor using CIDEX OPA® ortho-Phthalaldehyde Solution (Johnson & Johnson) or equivalent .55% ortho-phthalaldehyde-based solution. Follow manufacturer's instructions and recommendations for concentration, time of contact and postprocess procedure.

High-level disinfect or sterilize general purpose sensor using CIDEX® Activated Dialdehyde Solution (Johnson & Johnson) or equivalent 2% glutaraldehyde-based solution, CIDEX Plus® (Johnson & Johnson) or equivalent 3.4% glutaraldehyde-based solution, or a hydrogen peroxide-based solution. Follow manufacturer's instructions and recommendations for concentration, time of contact and post-process procedure.

DO NOT gas sterilize or autoclave general purpose sensor.

Covers

Covers can be used to go over the probe, bracket, sensors, and transmitter.

Planned Maintenance

The following maintenance schedule is suggested for the system and probes to ensure optimum operation and safety.



DANGER

Improper handling can lead to early probe failure and electric shock hazards.

Failure to do so will void probe warranty.

DO follow the specific cleaning and disinfection procedures provided in this chapter and the disinfectant manufacturer's instructions.

Table 11-16 • Planned Maintenance Program

Do the Following	Daily	After Each Use	As Necessary
Inspect the Probes	X	X	X
Clean the Probes		X	X
Disinfect Probes		X	X

Automatic Probe Diagnostics

Automatic Probe Diagnostics uses automated lens echo acquisition and advanced processing to provide real-time probe health information.

Enable Automatic Probe Diagnostics by selecting "Enable Automatic Probe Diagnostics" on the Utility>Admin page.



Figure 11-16 • Enable Automatic Probe Diagnostics

Automatic Probe Diagnostics will execute once at each system startup (on only the first supported probe connected from the left to right) and generate a dump file in the "D:\log\diags\VITA" directory. The dump file will be transferred to the Back Office together with other files under D:\log directory every day.

Automatic Probe Diagnostics supports the following probes at this time: C1-6-D, IC5-9-D, M5Sc-D, ML6-15 and RIC5-9-D.

Probe Check (not available in all countries)

The **Probe Check** utility appears on the B-Mode Touch Panel Page 2. (The Probe Check utility does not appear while Scan Assistant is running.) A manual probe check can be run with this utility for all probes except: P2D, P6D, P8D and 6Tc-RS probes.




Figure 11-17 • Probe Check

Run Probe Check

Ensure the probe lens surface is clean and free of dirt, water or coupling gel before beginning the probe check.

1. Select the **Probe Check** button on the Touch Panel.
2. A pop-up appears specifying to ensure the probe lens surface is clean. If the lens is clean, select **Yes** on the pop-up (see *Figure 11-18 • Probe Check Pop-up* on page 651). The probe check takes approximately 5 to 10 seconds.




Figure 11-18 • Probe Check Pop-up

3. When the probe check is complete, a pop-up specifies if the probe “Passed” (1) or “Failed” (2) the check (see *Figure 11-19 • Probe Check Pass or Fail Pop-ups* on page 651). If the probe passed the check, continue using the probe. If the probe failed the check, follow the instructions on the pop-up.



Figure 11-19 • Probe Check Pass or Fail Pop-ups

NOTE

While Probe Check can correctly identify most damaged probes, there are some less common types of probe defects that Probe Check is not able to detect. If a probe passes the probe check but the image quality of the probe is unacceptable, please contact your service partner.

Automatic Probe Check

Automatic Probe Check runs during probe selection at a time interval specified at Utility > Admin page > Enable Automatic Probe Diagnostics (see *Figure 11-20 • Automatic Probe Check* on page 652). The time interval options are: Every Time, Once a day, Once a week, Once a month and Never.

Automatic Probe Check runs for all probes except the P2D, P6D, P8D and 6Tc-RS probes. Automatic Probe Check increases the probe selection time by about five seconds. If the probe fails the probe check, a pop-up displays (see *Figure 11-19 • Probe Check Pass or Fail Pop-ups* on page 651). If the probe fails the check, follow the instructions on the pop-up.




Figure 11-20 • Automatic Probe Check

Returning/Shipping Probes and Repair Parts

US Department of Transportation and GE HealthCare policy requires that equipment returned for service MUST be clean and free of blood and other infectious substances.

When you return a probe or part for service (Field Engineer or customer), you need to clean and disinfect the probe or part prior to packing and shipping the equipment.

Ensure that you follow probe cleaning and disinfection instructions provided in the User Manual.

This ensures that employees in the transportation industry as well as the people who receive the package are protected from any risk.

Biopsy Special Concerns

Precautions Concerning the Use of Biopsy Procedures


WARNING

Do not freeze the image during a biopsy procedure. The image must be live to avoid a positioning error.

Biopsy guide zones are intended to assist the user in determining optimal probe placement and approximate the needle path. However, actual needle movement is likely to deviate from the guideline. Always monitor the relative positions of the biopsy needle and the target mass during the procedure, otherwise it could result in repeated biopsies or patient injury.


WARNING

NEVER reuse the TR5° disposable biopsy guide attachment, disposable sterile Ultra-Pro II needle guide kits or Verza Needle guide kits. Failure to follow the manufacturer's instructions could lead to potential exposure to infectious disease.


WARNING

The use of biopsy devices with accessories that have not been evaluated for use with this equipment may not be compatible and could result in injury. Failure to follow these instruction could result in repeated biopsies or patient injury.


WARNING

The invasive nature of biopsy procedures requires proper preparation and technique to control infection and disease transmission. Equipment must be cleaned as appropriate for the procedure prior to use.

- Follow the probe cleaning and disinfection procedures and precautions to properly prepare the probe.
 - Follow the manufacturer's instructions for the cleaning of biopsy devices and accessories.
 - Use protective barriers such as gloves and probe sheaths.
 - After use, follow proper procedures for decontamination, cleaning, and waste disposal.
- Failure to follow these instructions could lead to exposure to infectious agents.


CAUTION

Improper cleaning methods and the use of certain cleaning and disinfecting agents can cause damage to the plastic components that will degrade imaging performance or increase the risk of electric shock.

See *Probe Safety* on page 618 for more information.


WARNING

Consult the biopsy needle manufacturer's instructions for acceptable reprocessing of biopsy needles. Failure to follow the manufacturer's instructions could lead to exposure to infectious agents.



CAUTION

- Use of legally marketed, sterile probe sheaths is recommended for biopsy procedures.
- Before using a biopsy guide, read and follow the biopsy guide independent manuals given by the manufacturer.
- Every time before using a biopsy guide, ensure its correct position and optimal fit on the probe.
- Always use a straight needle for each biopsy procedure.
- Before performing a biopsy, ensure that the selected and displayed biopsy line corresponds to the biopsy needle guide mounted to the ultrasound probe (left/right).
- The biopsy needle and the biopsy needle guide (and the bore inside) must be sterile.

Biopsy Guide Sterilization

Sterilization with autoclave is possible for the reusable stainless steel Biopsy Guides for the following probes:

- IC5-9-D
- RIC5-9-D
- RAB6-D

See *Probe Biopsy Reprocessing* on page 673 for more information.

Performing a Biopsy

Displaying the Guidezone

Activate the Biopsy Kit by selecting it from the B-Mode Touch Panel.



Figure 11-21 • B-Mode Touch Panel Menu

The available biopsy options appear when Biopsy Kit is selected. There are fixed and adjustable angle biopsy kits and plastic/disposable and reusable biopsy guides available with the ultrasound system depending on the probe. Select the desired biopsy kit.



CAUTION

Some angles (for example, if outside of the Field of View) may not be supported on all probes.

NOTE

You can display the biopsy guideline on the CFM image in simultaneous mode. Enabling Color Flow allows for visualization of the vascular structure around the area to be biopsied. Select the Show Biopsy Mark on CFM simultaneous Mode preset in the Utility -> System -> System Image -> Biopsy Guide screen.

NOTE

Be sure to match the angle setting on the bracket to the Biopsy Kit setting on the system.




Figure 11-22 • Biopsy Guidezones for the M5Sc Probe

1. Biopsy Kit Name and Biopsy Needle Angle
2. 1 cm increments
3. 5 cm increments

The biopsy guidezone represents the path of the needle. The dots which make up the guidezone represent the depth readout, where:

- Yellow represents 1 cm increments.
- Red represents 5 cm increments.

The display should be carefully monitored during a biopsy for any needle deviation from the center line or guidezone.

Before scanning, verify the needle can be visualized within the imaging plane. Use appropriate needle length to reach target area. Adjust the guide settings on the system and confirm they pass through the target, then match the setting on the system to the pin settings on the guide.

NOTE

Biopsy Needle Angle is defined with respect to the horizontal axis. This is equivalent to $(90 \text{ deg}) - (\text{angle specified by CIVCO})$. Displayed angle may be slightly different from CIVCO's specifications due to part variability.

The Biopsy Guidezone adjusts along with image adjustments, such as image inversion/rotations, zoom and depth changes.

The needle may vary from the center line or guidezone for various reasons:

- Needle barrel to needle clearance or strength.
- Bracket manufacturing tolerance.
- Needle deflection due to tissue resistance.
- Needle size chosen. Thinner needles may deflect more.



WARNING

Failure to match the guidezone displayed to the guide may cause the needle to track a path outside the zone.

It is extremely important that when using the adjustable angle biopsy guides, the angle displayed on the screen matches the angle set on the guide, otherwise the needle will not follow the displayed guidezone which could result in repeated biopsies or patient injury.

Guide circle on the biopsy line

You can use a guide circle on the biopsy line.

Mode: B/CF/PDI/Elastography/Contrast/Volume Navigation

Display format: Single/Dual



Figure 11-23 • Guide circle

- Guide circle
- Biopsy line
- The system displays current size of Target diameter (inside circle) and margin distance in the status area.

Preset for guide circle

Check **Show Biopsy Circle** in **Utility > System > System Imaging** to display a guide circle control on the monitor display and the trackball key.

Biopsy circle control



Figure 11-24 • Biopsy circle control

1. Target diameter
2. Margin distance
3. Guide circle display On/Off
4. Circle position (move Trackball Up/Down)

Preparing the Biopsy Guide Attachment

Convex, Sector and Linear probes have optional biopsy guide attachments for each probe. The guide consists of a non-disposable bracket to attach to the probe, disposable needle clip to attach to the bracket, sheath, gel (sterile gel if necessary) and disposable needle barrels.

The disposable needle barrels are available for a variety of needle sizes.



CAUTION

DO NOT attempt to use the biopsy bracket and needle guide until the manufacturer's instructions, provided with the biopsy bracket and needle guide in the kit, have been read and thoroughly understood. Failure to follow these instructions could result in repeated biopsies or patient injury.

The bracket is packaged non-sterile and is reusable. To avoid possible patient contamination, ensure bracket is properly cleaned, sterilized or disinfected before each use.

Disposable components are packaged sterile and are single-use only. Do not use if integrity of packing is violated or if expiration date has passed.

Fixed Needle Biopsy Guide Assembly



WARNING

DO NOT attempt to use the biopsy bracket and needle guide until the manufacturer's instructions, provided with the biopsy bracket and needle guide in the kit, have been read and thoroughly understood. Failure to follow these instructions could result in repeated biopsies or patient injury.

1. Identify the appropriate biopsy guide bracket by matching the label on the bracket with the probe to be used.
2. Orient the bracket so that the needle clip attachment will be on the same side as the probe orientation mark (ridge).



1. Probe Orientation Mark
2. Bracket

Figure 11-25 • Probe/Bracket Alignment

3. Attach the biopsy bracket to the probe by sliding the bracket over the end of the probe until it clicks or locks in place.
4. Place an adequate amount of coupling gel on the face of the probe.
5. Using a sterile technique, place the proper sanitary sheath over the probe and biopsy bracket. Use the rubber bands supplied to hold the sheath in place.





Figure 11-26 • Applying Sanitary Sheath

6. Snap the fixed or adjustable needle clip onto the biopsy guide bracket.



1. Sheath

Figure 11-27 • Fixed Needle Clip Attachment

7. Push the locking mechanism towards the bracket to secure the lock. Make sure the needle guide is firmly attached to the bracket.



Figure 11-28 • Locking the Needle Clip

NOTE

If using an in-plane needle guide, the needle clip appears as shown here. Be sure to choose an in-plane guide that matches the gauge of the needle being used. The in-plane guide does not support any on-screen graphics. This is because the guide allows variable angles. For in-plane needle guides, steps 8 and 9 are not applicable.



Figure 11-29 • Example of In-Plane Needle Guide

8. Place the needle barrel into the needle clip with the desired gauge facing the needle clip and snap into place.



Figure 11-30 • Needle Barrel Selection

9. Place the needle barrel into the needle clip with the desired gauge facing the needle clip and snap into place.



Figure 11-31 • Needle Barrel Installation



WARNING

Ensure that all guide parts are seated properly prior to performing a biopsy. Failure to follow these instructions could result in repeated biopsies or patient injury.

Multi Angle Biopsy Guide Assembly



WARNING

DO NOT attempt to use the biopsy bracket and needle guide until the manufacturer's instructions, provided with the biopsy bracket and needle guide in the kit, have been read and thoroughly understood. Failure to follow these instructions could result in repeated biopsies or patient injury.

1. Scan the patient and identify the target for biopsy. Move the probe to locate the target to the center of the image. Enable the system biopsy guidezone and try guidezone angles MBX1 to MBX3 to decide the best angle setting for needle path.




Figure 11-32 • Example

2. Pull up on the pin (a) to freely move the needle guide attachment. Align the pin with the selected position of the needle guide attachment.
Push the pin down (b) into the desired slot to secure the angle position of the needle guide attachment.

Probes and Biopsy



Figure 11-33 • Pull up and push down the pin

3. Fit a convex piece of the biopsy bracket (a) into the concave position of the probe (b).



Figure 11-34 • Probe/Bracket Alignment

Hold the side (a) and tuck down the needle guide side (b) until it clicks or locks in place.



Figure 11-35 • Probe/Multi-angle Bracket Alignment 2

4. Place an adequate amount of coupling gel on the face of the probe.
5. Place the proper sanitary sheath tightly over the probe and biopsy bracket. Use the rubber bands supplied to hold the sheath in place.




Figure 11-36 • Applying Sanitary Sheath

6. Snap the needle guide onto the biopsy guide bracket.




Figure 11-37 • Snap the needle guide

7. Push the locking mechanism towards the bracket to secure the lock (a). Make sure the needle guide is firmly attached to the bracket.



Figure 11-38 • Lock the Needle guide

8. Choose the desired gauge (size) needle barrel. Twist it back and forth to remove it from the plastic tree.



Figure 11-39 • Needle Barrel

9. Place the needle barrel into the needle clip with the desired gauge facing the needle clip and snap into place.



Figure 11-40 • Needle Barrel Installation

Remove the biopsy guide

Hold the other side and push out the needle clip attachment side (a).



Figure 11-41 • Remove the biopsy guide



CAUTION

Prevent damage to the probe lens with finger nails.

Releasing the needle

According to the following procedure, you remove the needle from a probe and an assembly without moving the needle.




Figure 11-42 • Release the needle from assembly

1. Push the knob portion of a sleeve in the direction of the arrow.
2. The needle is released from the assembly.
3. Push the probe and the assembly in the direction of the larger arrow to remove the needle.

Endocavitary Probe Biopsy Guide Assembly - Representative Example



WARNING

DO NOT use the needle with the catheter (soft tube). There is a possibility of breaking the catheter in the body.




WARNING

Before inserting the needle, scan the patient to determine the correct puncture depth and site. Only the sterile/sanitary sheath and rubber band are on the probe during the pre-needle placement scanning. Failure to follow these instructions could result in repeated biopsies or patient injury.

Preparation

To prepare the endocavitary probe for use:

1. Remove the probe from the box and carefully examine it for any damage.
2. If the biopsy guide is to be attached, use the filling removal tool to clean out the attachment area on the probe head.



Probes and Biopsy

1. Probe Head
2. Attachment
3. Filling Removal Tool

Figure 11-43 • Attachment Filling Removal

3. Clean, then disinfect the probe.

NOTE

Ensure that protective gloves are worn.

Installing the sheath

To install the sheath:

1. Remove the sheath from its package. Do not unroll the sheath.

NOTE

Remember to rinse all sanitary probe sheaths of powder before placing on the probe. Powder can degrade the displayed image.

2. Place an adequate amount of ultrasound gel inside the sheath tip (the gel is between the sheath inner surface and the probe aperture).

NOTE

Ensure that only acoustic coupling gel is used for this purpose.

3. Place the sheath tip over the probe aperture and then pull the sheath end toward the probe handle.
4. Inspect the sheath for nicks, cuts or tears.



1. Secure the Sheath with a rubber band.
2. The probe sheath should extend past the end of the probe to the probe's cable.

Figure 11-44 • Endocavitary Probe with Sheath

5. Rub a finger over the tip of the probe to ensure all air bubbles have been removed.

Endocavitary Probe Biopsy Guide Preparation

1. If a biopsy is to be performed, snap the metal or plastic biopsy guide on to the probe over the sheath.

**WARNING**

Ensure that all guide parts are seated properly prior to performing a biopsy. Failure to follow these instructions could result in repeated biopsies or patient injury.

2. Fix with a screw

NOTE

For the RIC5-9-D and IC5-9-D probes, use the TR5 guidelines for the plastic (disposable-only) biopsy guides; use the RU guidelines with the stainless steel reusable biopsy guides.

3. Place an adequate amount of ultrasound gel on the gel-filled sheath tip's outer surface.
4. Ensure the guide is properly seated and secure by pushing forward on the needle insertion end of the guide until the attachment node is firmly in place in it's hole.

4D Biopsy Guide Assembly - Representative Example

4D Probe

1. Place the needle guide onto the probe.
2. Push the needle guide forward until the bracket (b) catches in the support on the housing of the probe (a).



Figure 11-45 • Support on the housing



Figure 11-46 • Biopsy Needle Guide

3. Fix the biopsy guide by locking the frame on the opposite side.



Figure 11-47 • Mounting the Biopsy Needle Guide to the 4D Probe

NOTE

Needle guide sterilization with autoclave possible.

4D Endocavitary Probe

1. Place an adequate amount of ultrasound gel inside the sheath tip (the gel is between the sheath inner surface and the probe aperture).

NOTE

Ensure that only acoustic coupling gel is used for this purpose.

2. Place the sheath tip over the probe aperture and then pull the sheath end toward the probe handle.
3. Inspect the sheath for nicks, cuts or tears.
4. Rub a finger over the tip of the probe to ensure all air bubbles have been removed.
5. Position the small swelling of the needle guide on the notch at the probe tip. Snap the needle guide.



Figure 11-48 • Installation (without probe sheath)

NOTE

Material: Stainless Steel

NOTE

Needle guide sterilization with autoclave possible.

Verza Biopsy Needle Guide

The Versa Biopsy Needle Guide is available for the C1-6-D, C1-6VN-D, L2-9-D and L2-9VN-D probes.

Activate the Biopsy Kit by selecting it from the B-Mode Touch Panel.



Figure 11-49 • B-Mode Touch Panel Menu

The available biopsy options appear when Biopsy Kit is selected. Select the desired biopsy kit.

**CAUTION**


Be sure to match the 'pin' setting on the bracket to the pin setting on the system.

Verza Needle Guide Attachment Procedure


NOTE

The following procedure is shown with a C1-6-D probe.

Table 11-17 • Verza Biopsy Guide Attachment Procedure

Step	Instructions	Illustration
1.	Attach the Biopsy Guide Bracket to the probe. Match the indentations on the Biopsy Guide Bracket and the probe.	
2.	Tightly secure the Biopsy Guide Bracket to the probe. NOTE Be sure you are using the correct bracket for the probe. The bracket and probe are clearly marked.	
3.	Attach the probe sheath. 1. Apply gel. 2. Cover the probe. 3. Attach the rubber band.	

Step	Instructions	Illustration	
4.	Tightly secure the Needle Guide to the Biopsy Guide Bracket through the sheath.		
5.	Unlock angle lock to adjust the needle guide angle; then re-lock.	Locked	Unlocked
	Adjust Needle Guide Angle		
		Re-lock	

Step	Instructions	Illustration
6.	Insert correct Needle Gauge Holder. 	
7.	Push the Quick Tab Release, insert the biopsy needle into the needle guide. Close the Quick Tab Release.	

4D Probe Biopsy Needle Path Selection

To select the needle path and verify that the path of the needle is accurately indicated within the guidezone on the system monitor, perform the following before use:

1. Properly install the bracket and biopsy guide.
2. Scan in a container filler with water (47° C).
3. Select **Biopsy kit**. The available biopsy options from the Touch Panel. Select the biopsy guidezone where the needle echo passes through the center of the guidezone. Use the selected biopsy guidezone when performing the biopsy.

Biopsy Needle Path Verification

To verify that the path of the needle is accurately indicated within the guidezone on the system monitor, perform the following:

- Properly install the bracket and biopsy guide.
- Scan in a container filled with water (47° C).
- Display the biopsy guidezone on the monitor.

- Ensure that the needle echo falls within the guidezone markers.

The Biopsy Procedure



WARNING

Biopsy procedures must only be performed on live images. Failure to follow these instructions could result in repeated biopsies or patient injury.



WARNING

Ensure that all guide parts are seated properly prior to performing a biopsy. Failure to follow these instructions could result in repeated biopsies or patient injury.

1. Place coupling gel on the scanning surface of the probe/sheath/biopsy guide assembly.
2. Activate the biopsy guidezone on the system through the B-Mode Touch Panel. When using multi-angle guides, ensure that the proper guidezone angle is displayed.
3. Scan to locate the target. Center the target in the electronic guidezone path.

NOTE

Enabling color flow would allow for visualization of the vascular structure around the area to be biopsied.

4. Place the needle in the guide between the needle barrel and needle clip. Direct it into the area of interest for specimen retrieval.

Post Biopsy

When the biopsy is complete, remove the needle barrel, needle clip and probe sheath. Properly dispose of these items in accordance with current facility guidelines.

Clean and disinfect the probe. See *Probe Processing* on page 624 for more information.

The biopsy bracket can be cleaned and disinfected in a recommended disinfecting agent and reused.



WARNING

When the biopsy needle guide kit is opened, all parts must be discarded after the procedure whether they have been used or not. Failure to follow these instructions could result in repeated biopsies or patient injury.

Probe Biopsy Reprocessing

Manual Cleaning

NOTE

Efficacy of this manual cleaning process has been validated using Cidezyme/ENZOL Enzymatic Detergent.

1. Remove the biopsy guide and protective sheath(s) from the probe.
2. Whenever possible the biopsy guide should be rinsed immediately after use. If the biopsy guide cannot be cleaned immediately after use, cover the biopsy guide with a towel dampened with purified water. Devices may remain in this condition for a maximum of 4 hours.
3. Remove all visible soil. Flush the biopsy guide using utility water (30 - 40°C) for not less than 2 minutes.
4. Prepare the enzymatic detergent according to the manufacturer's recommendations, using utility water.
5. Submerge the biopsy guide in the prepared solution and soak for no less than 2 minutes.
6. After the 2-minute soak, while the biopsy guide is submerged in the cleaning solution, vigorously scrub the device with a soft nylon bristle brush.
7. Use a round nylon cleaning brush to clean the biopsy lumen. Use a syringe to flush cleaning solution through the lumen. Scrub the device for a minimum of 2 minutes.
8. Remove the device from the cleaning solution and rinse thoroughly under running utility water (30 - 40°C) taking care to remove any visible detergent. Rinse the device for a minimum of 1 minute.
9. Visually inspect the device for any residual soil or detergent. Repeat steps 6 through 8 until the device is visibly clean.



CAUTION

Do not clean any portion of the attachment with methanol, ethanol, isopropanol, or any other alcohol base detergent. Such substances can cause irreparable damage to the attachment.

NOTE

Discard solutions and rinse waters in accordance with local regulations

High-Level Disinfection

NOTE

High-Level disinfection efficacy of this manual process has been validated using Cidex OPA.

1. Fill a sink or basin with high-level disinfectant prepared in accordance with the disinfectant manufacturer's instructions to a level allowing immersion of the biopsy guide.
2. Immerse the devices in the disinfectant solution and agitate to ensure all air bubbles are removed from the surface of the device.
3. Allow the devices to soak in the disinfectant solutions for at least the minimum contact time listed in the disinfectant manufacturer's instructions for use.
4. Thoroughly rinse the device by immersing in a large volume of critical water for a minimum of 1 minute.

NOTE

Critical Water is water that is treated (usually by a multistep treatment process that could include a carbon bed, softening, DI, and RO or distillation) to ensure that the microorganisms and the inorganic and organic material are removed from the water to an appropriate level (Refer to AAMI TIR34/ST108).

Use of this type of water will reduce the recontamination of biopsy guide during processing.

- Repeat Step 4 two additional times, for a total of 3 (three) rinses using fresh volumes of water for each rinse.

NOTE

Discard solutions and rinse waters in accordance with local regulations

- Thoroughly dry the biopsy guide using a low-/non-linting wipe. Visually inspect the biopsy guide to ensure all surfaces are clean and dry. Repeat drying steps if any moisture is visible.

Autoclave Sterilization

NOTE

Sterilization efficacy testing was performed using worst-case parameters for time, temperature and load density. Parameters listed in the tables are the minimum required to ensure a Sterility Assurance Level (SAL) of 10^{-6} or better.

- Place the cleaned, disinfected, and disassembled biopsy guide in an approved autoclave pouch.
- Autoclave using the following parameters:

Table 11-18 • Autoclave parameters

Parameter	Cycle Type 1	Cycle Type 2
Sterilizer	Pre-vacuum	Pre-vacuum
Preconditioning Pulses	3	3
Temperature (Minimum)	132 degrees C	134 degrees C
Exposure Time (Minimum)	4 Minutes	3 Minutes
Drying Time (Minimum)	15 Minutes	15 Minutes
Package Configuration	Tyvek Pouch (14 x 25 cm)	Tyvek Pouch (14 x 25 cm)

Surgery/Intra-operative Use

Preparing for Surgery/Intra-operative Procedures

Preparing the transducer for intra-operative use follows the same sterile procedure as for biopsy use except that no biopsy attachments are used. See *Preparing the Biopsy Guide Attachment* on page 658 for more information. Sterile gel is applied to the transducer face and a sterile sheath completely covers the transducer and cable which has first undergone a thorough cleaning and high-level disinfection.

The invasive nature of biopsy procedures requires proper preparation and technique to control infection and disease transmission. Equipment must be cleaned as appropriate for the procedure prior to use.



CAUTION

For surgery/intra-operative procedures, a sterile environment is required. Therefore, both the operator and probe sheath needs to be sterile.

To ensure a sterile environment during the procedure, it is recommended that this be a two-person job.

1. Perform a high level disinfection of the probe.
2. The scanner (surgeon, sonographer, etc.) should be sterile and gloved.
3. Place an adequate amount of sterile coupling gel on the face of the probe.
4. Place the proper sterile sheath over the probe and cord.




Figure 11-50 • Applying Sterile Sheath

1. Depending on the type of procedure, use either sterile water or sterile gel on the sheath cover.

NOTE

Follow your institutions guidelines on post surgery/intra-operative procedures for probe cleaning and disinfection.

Chapter 12

User Maintenance

In this section

System Data.	678
Privacy and Security.	688
Anti-Virus Software Note.	691
Loading Windows Patches.	693
Software Download.	696
Base Image and Software Load.	700
Copyrighted Material.	701
System Care and Maintenance.	702
Quality Assurance.	715
Image Quality Check.	726
Assistance.	727

This chapter supplies system data, assistance information, and system care and maintenance instructions.

System Data

Features/Specifications

Table 12-1 • Physical Attributes

Dimensions and Weight (for Transport) - Height: 1250 - 1800 mm, 49 - 71 in (HDU display), 1210 - 1760 mm, 48 - 69 in (LCD display) - Width: 530mm, 20.9 inch (Caster), 545mm, 21.5 inch (LCD display), 565mm, 22.2 inch (HDU display) - Depth: 885mm, 34.8 inch - Weight: 85 kg (187.4 lb.) User interface - Operating keyboard adjustable in height and rotation - Ergonomic hard key layout - Interactive Back-Lighting - Integrated recording keys for remote control of up to 4 peripheral or DICOM devices - Integrated Gel Warmer (Option) Electrical Power - Voltage: 100 - 240 Vac - Frequency: 50/60 Hz - Power consumption maximum of 0.9 KVA with peripherals. Touch Screen 12.1 in High Resolution, Color, Touch, Display screen - Interactive, dynamic software menu - Brightness adjustment - User-configurable layout	Console Design 4 Active Probe Ports, 1 Inactive Probe Storage Port - Integrated SSD (1TB) - Integrated DVD-R Multi Drive - On-board storage for peripherals: Thermal printer. - Integrated speakers - Integrated locking mechanism that provides rolling lock and caster swivel lock - Integrated cable management - Front and Rear handles - Easily-removable air filters HDU Display 23.8" Wide screen High-Resolution HDU Display - Monitor translation: 350mm, 13.7 in. horizontal; 120mm, 4.7 in. vertical; 90 degree swivel - Fold-down and Lock Mechanism for transport - Resolution 1920 x 1080 - Anti Glare - Viewing Angle 89/89/89/89 degrees LCD Display 23.8" Wide screen High-Resolution LCD Display - Monitor translation: 350mm, 13.7 in. horizontal; 120mm, 4.7 in. vertical; 90 degree swivel - Fold-down and Lock Mechanism for transport - Resolution 1920 x 1080 - Anti Glare - Viewing Angle 89/89/89/89 degrees

Table 12-2 • System Overview

<u>Applications</u>	<u>Display Mode</u>
- Abdominal - Obstetrical - Gynecological - Breast - Small parts - Peripheral Vascular - Transcranial (Adult and neonatal) - Pediatric and neonatal - Musculoskeletal (general and superficial) - Urological - Cardiac (adult and pediatric) - Interventional - Pleural	- Live and Stored Display Format: Full and Split Screen - both with thumbnails for Still and CINE - Review Image Format: 4x4 and Thumbnails for Still and CINE - Time line Display [Independent Dual B or CrossXBeam/PW Display; CW; Top/Bottom selectable display format; Side/side selectable format - Virtual Convex
<u>Operating modes</u>	<u>Simultaneous Capability</u>
- B-Mode - M-Mode - Color Flow Mode (CFM) and Microvascular Imaging (MVI) - Power Doppler Imaging (PDI) - B-Flow - Extended Field of View (LOGIQ View) - PW Doppler - CW Doppler (Option) - Volume Mode (3D/4D)	- B or CrossXBeam/PW, B or CrossXBeam/CW(Optional), B or CrossXBeam/CFM or PDI, B/M, B/CrossXBeam, B-Flow/PW, Realtime Triplex Mode [B/CrossXBeam + CFM/PDI + PW] - Colorized Image - Colorized B - Colorized M - Colorized D - Multi Image Split/Quad Screen [Live and/or Frozen, B or CrossXBeam+B or CrossXBeam/CFM or PDI, PW/M, Independent Cine playback] - Zoom: Write/Read/Pan
3D Static 4D Real Time (Option)	<u>Selectable alternating modes</u>
- Anatomical M-Mode - Contrast Imaging (Option) - Strain Elastography - Shear Wave Elastography (Option) - UGAP (Option)	B or CrossXBeam/PW; B or CrossXBeam + CFM (PDI)/PW; B/CW(option)
<u>Scanning methods</u>	<u>Multi Image (split/quad screen)</u>
- Electric sector - Electric convex - Electric linear - mechanical volume sweep	Live and/or Frozen, B or CrossXBeam+B or CrossXBeam/CFM or PDI, PW/M, Independent Cine playback
<u>Transducer Types</u>	<u>Peripheral Options</u>
- Sector Phased Array - Curved Linear Array - Microcurved linear array - Linear Array - Matrix Array (Linear) - Volume probe (4D) - Split Crystal - TEE probe	- Integrated Options - Digital B&W thermal printer - DVD vide recorder - DVD Drive - Digital color thermal printer - Foot switch, with programmable functionality - Console protective cover - LOGIQ smart device apps - Photo Assistant - Remote Control - CRF200-U Card Reader (Japan only)

Table 12-3 • System Overview

<u>Standard Features</u>	<u>Options</u>
• Advanced user interface with high resolution 12.1 inch touch screen • Automatic Optimization • CrossXBeam • Speckle Reduction Imaging (SRI-HD, Advanced SRI Type 1) • Fine Angle Steer • Coded Harmonic Imaging • Virtual Convex • Patient information database • Image Archive on integrated CD/DVD and Hard Drive • Advanced 3D • Raw Data Analysis • Real-time Automatic Doppler Calculations • OB Calculations • Fetal Trending • Multigestational Calculations • Hip Dysplasia Calculations • Gynecological Calculations • Vascular Calculations • Urological Calculations • Renal Calculations • Cardiac Calculations • InSite Capability • On-board electronic documentation • Auto Doppler Assist • Privacy and Security, including user and rights management • LOGIQView • Breast Productivity Package • Thyroid Productivity Package • External USB Printer connection • Network Printer Support • HDMI output available for compatible devices • DICOM • App Launchpad	• Tricefy • B-Flow • Auto IMT • Compare Assistant • Scan Assistant • OB Measure Assistant • Color Quantification • Strain Elastography • Elastography quantification • Advanced Privacy & Security (Vulnerability Scan) • Power Assistant and Scan on Battery • Storage Bins • Shear wave elastography • Volume navigation • UGAP • Hepatic assistant • Coded Contrast Imaging • Stress echo • Cardiac Strain (Automatic Function Imaging) • On-board reporting • TVI • Wireless LAN • CW • DVR • Tablet tools • Advanced probes • Breast Assistant, Powered by Koios DS • Thyroid Assistant, Powered by Koios DS • SonoNT SonoIT • Advanced SRI-HD Type 2 • Auto Preset Assistant • Auto Abdominal Color Assistant • Auto Renal Measure Assistant • CEUS LI-RADS • Raw Data Streaming • DVD Drive

Table 12-4 • System Parameters

<u>Controls Available on Freeze or Recall</u>	<u>Scanning Parameters</u>
• Max Pan Zoom to 8 • B/M/CrossXBeam Mode (Gray Map Optimization; TGC, Colorized B and M; Frame Average [Loops only]; Dynamic Range) • PW-Mode (Gray Map; Post Gain; Baseline Shift; Sweep Speed; Invert Spectral Waveform; Compression; Rejection, Colorized Spectrum; Display Format; Angle Correct; Quick Angle Correct, Auto Angle Correct, Doppler Audio) • Color Flow (Overall Gain [Loops and Stills]; Color Map; Transparency Map; Frame Averaging [Loops only]; Flash Suppression, CFM Display Threshold; Spectral Invert for Color/Doppler) • Automatic Optimization • SRI-HD • CrossXBeam (display non-compounded and compounded image simultaneously in split screen) • 3D Reconstruction from a stored CINE loop • 4D (Gray Map, Colorize; Post Gain; Change display between single or rendered) • Anatomical M-Mode • Anatomical M-Mode on CINE Loop	• Digital Beamformer • Imaging Depth: 0-50 cm [Minimum: 0-2 cm (Zoom); Maximum: 0-50 cm] • Continuous Dynamic Receive Focus/Continuous Dynamic Receive Aperture • Adjustable Dynamic Range • Adjustable Field of View [FOV] • Image Reverse: Right/ Left • Image Rotation: 0, 90, 180, 270 degrees
<u>Controls Available While “Live”</u> • HD Zoom • B/M/CrossXBeam-Mode (Gain; TGC; Dynamic Range; Acoustic Output; Frame Rate Control, Sweep Speed for M-Mode; # of Angles for CrossXBeam) • PW-Mode (Gain; Dynamic Range; Acoustic Output; Transmission Frequency; PRF; Wall Filter; Spectral Averaging; Sample Volume Gate for PW-Mode Length and Depth; Velocity Scale) • Color Flow (CFM Gain; CFM Velocity Range; Acoustic Output; Wall Echo Filter; Frame Rate Control; CFM Spatial Filter; CFM Frame Averaging; CFM Line Resolution; Frequency/ Velocity Baseline Shift)	<u>Image Storage</u> • Storage Format: DICOM (compressed/ uncompressed, single/multiframe, enhanced [3D/4D], with/without Raw Data), Export JPEG, and WMV (MPEG4) • Storage Devices: USB Memory Stick (64MB to 64GB, for exporting individual images/ clips); CD-R Storage (700MB); DVD-R Storage (4.7GB); Hard Drive Image Storage (~730GB) • On-Board database of patient information from past exams • Compare previous exam images with current exam • Reload of archived data sets CINE Memory/Image Memory

Table 12-5 • System Parameters

<u>Connectivity</u> • Ethernet network connection • Wireless LAN 802.11ac/a/b/g/n (option) • DICOM 3.0 (option) with Verify, Print, Store, Modality Worklist, Storage Commitment, Modality Performed Procedure Step [MPPS], Media Exchange, Off network/mobile storage queue, Query/Retrieve) • Public SR Template • Structured Reporting - compatible with Vascular, OB, Cardiac, and Breast standard • InSite ExC capability	<u>CINE Memory/Image Memory</u> • 1 GB of CINE Memory • Selectable CINE Sequence for CINE Review • Prospective CINE Mark • Measurements/Calculations & Annotations on CINE Playback • Scrolling timeline memory • Dual Image CINE Display • Quad Image CINE Display • CINE Gauge and CINE Image Number Display • CINE Review Loop • CINE Review Speed

Table 12-6 • Measurements and Calculations

<u>B-Mode</u> • Depth and Distance • Circumference and Area (Ellipse/Trace) • Volume (Ellipsoid) • Angle between 2 Lines • % Stenosis (Area or Diameter) • Dual B-Mode capability <u>M-Mode</u> • M Depth and Distance • Time • Slope • Heart Rate <u>Doppler Measurements/Calculations</u> • Velocity • Time • Peak Systole • End Diastole • Acceleration • Acceleration Time • Ratios A/B Ratio (Velocities/Frequency Ratio) Peak Systole/End Diastole (PS/ED Ratio) End Diastole/Peak Systole (ED/PS Ratio) • Heart Rate • TAMAX (Time Averaged Maximum Velocity) • Volume Flow [TAMEAN and Vessel Area] • PI (Pulsatility Index) • RI (Resistivity Index) Shear Wave Measurements/Calculations	<u>Shear Wave Measurements/Calculations</u> • Stiffness • Velocity <u>Vascular Measurements/Calculations</u> • Carotid, Vertebral, Subclavian Measurements, Auto IMT • Summary Worksheet • Summary Report <u>Small Part Measurement Analysis</u> • Koios DS Breast Lesion Decision Support • Koios DS Thyroid Lesion Decision Support

Table 12-7 • Measurements and Calculations

<u>Obstetrics Measurements/Calculations</u>	<u>Gynecology Measurements/Calculations</u>
• Gestational Age Calculation • Calculations and Ratios • EFW Calculation • Growth Percentiles • Multi-Gestational Calculation • Fetal Qualitative Description (Anatomical Survey) • Fetal Environmental Description (Biophysical profile) • Programmable OB Tables • Fetal Graphical Trending • Measurements / Calculations by: ASUM, ASUM 2001, Alexander, Bahlmann, Baschat, Berkowitz, Bertagnoli, Brenner, Campbell, CFEF, Chervenak, Chitty, Doubilet, Ebbing, Eik-Nes, Ericksen, Goldstein, Hadlock, Hansmann, Hellman, Hill, Hohler, Jeanty, JSUM, Kurmanavicius, Kurtz, Mari, Mayden, Mercer, Merz, Moore, Nelson, Osaka University, Paris, Pexsters, Rempen, Robinson, Shepard, Shepard/Warsoff, Sonek, Tokyo University, Tokyo/Shinozuka, WHO, Yarkoni • Over 20 selectable OB Calcs • Summary Worksheet • Summary Report	• Ovarian, Uterine, Pelvic Floor, Endometrium, Follicular Measurements, Cervix • Summary Worksheet • Summary Report • Qualitative Description (Anatomical Survey)
	<u>Urology Measurements/Calculation</u>
	• Bladder, Prostate, Renal, Generic, Post-Void Bladder Volume Measurements • Summary Worksheet • Summary Report
	<u>Cardiology Measurements/Calculations</u>
	• Cardiology Measurements and Calculations • Summary Worksheet • Summary Report

Table 12-8 • Biopsy Guides

• Single-Angle, disposable with a reusable bracket • Multi-Angle, disposable with a reusable bracket	• Single-Angle, disposable with a disposable bracket

Table 12-9 • Inputs and Outputs Signal

• HDMI • S-Video	• Ethernet • Multiple USB 2.0/3.0 ports

Clinical Measurement Accuracy

Basic Measurements

The following information is intended to provide guidance to the user in determining the amount of variation or measurement error that should be considered when performing clinical measurements with this equipment. Error can be contributed by equipment limitations and improper user technique. Be sure to follow all measurement instructions and develop uniform measurement techniques among all users to minimize the potential operator error. Also, in order to detect possible equipment malfunctions that could affect measurement accuracy, a quality assurance (QA) plan should be established for the equipment that includes routine accuracy checks with tissue mimicking phantoms.

Please be advised that all distance and Doppler related measurements through tissue are dependent upon the propagation velocity of sound within the tissue. The propagation velocity usually varies with the type of tissue, but an average velocity for soft tissue is assumed. This equipment is designed for, and the accuracy statements listed on are based on, an assumed average velocity of 1540 m/s. The percent accuracy when stated applies to the measurement obtained (not the full scale range). Where the accuracy is stated as a percent with a fixed value, the expected inaccuracy is the greater of the two.

Table 12-10 • System Measurements and Accuracies

Measurement	Units	Useful Range	Accuracy	Limitations or Conditions
Depth	mm	Full Screen	±20%	
Distance:				
Axial (Equal to 1540m/s)	mm	Full Screen	±3% or ±1mm, whichever is greater	
Lateral (Equal to 1540m/s)	mm	Full Screen	±5% or ±1mm, whichever is greater	Linear Probes
Lateral (Equal to 1540m/s)	mm	Full Screen	±6.5% or ±3.5mm, whichever is greater	Convex Probes
Lateral (Equal to 1540m/s)	mm	Full Screen	±5% or ±1mm, whichever is greater	Sector Probes
Axial/Lateral (Not Equal to 1540m/s)	mm	Full Screen	±7.5% or ±5.0mm, whichever is greater	All probes Not Equal To 1540m/s
Circumference:				
Trace (Equal to 1540m/s)	mm	Full Screen	±5% or ±1mm, whichever is greater	
Ellipse (Equal to 1540m/s)	mm	Full Screen	±5% or ±1mm, whichever is greater	
Trace (Not Equal to 1540m/s)	mm	Full Screen	±7.5% or ±5mm, whichever is greater	All probes Not Equal To 1540m/s
Ellipse (Not Equal to 1540m/s)	mm	Full Screen	±7.5% or ±5mm, whichever is greater	All probes Not Equal To 1540m/s
Area:				

Measurement	Units	Useful Range	Accuracy	Limitations or Conditions
Trace (Equal to 1540m/s)	mm ²	Full Screen	±10% or ±5mm ² , whichever is greater	
Ellipse (Equal to 1540m/s)	mm ²	Full Screen	±10%, or ±5mm ² , whichever is greater	
Trace (Not Equal to 1540m/s)	mm ²	Full Screen	±20% or ±20mm ² , whichever is greater	All probes Not Equal To 1540m/s
Ellipse (Not Equal to 1540m/s)	mm ²	Full Screen	±20% or ±20mm ² , whichever is greater	All probes Not Equal To 1540m/s
Time	s	Timeline Display	±5%, not to exceed 10ms	M or Doppler Mode
Slope	mm/s	Timeline Display	±5%, not to exceed 1mm/s	M-Mode Only
Doppler Sample Volume Depth (SVD)	mm	Full Screen	2mm (0.2cm) in any direction	When SVD is at least half the depth of the image
Velocity	cm/s	Full Range	10%	PW and CW Doppler Mode
Doppler Angle Correction	degrees	From 0-59° From 60-90°	±1% ±2%	
Shear Wave Elastography Velocity/Stiffness Measurements				
Velocity	m/s	Shear wave ROI	±10%, or 0.5 m/s, for the 9L-D and ±15%, or 0.5 m/s, for the C1-6-D, C1-6VN-D, L2-9-D and L2-9VN-D, whichever is greater over the range of depths for which measurements can be made, limited by shear wave penetration.	SD less than or equal to 5% for multiple repeated measurements over the range of depths for which measurements can be made, limited by shear wave penetration.
Stiffness	kPa	Shear wave ROI	±20%, or 0.75 kPa, for the 9L-D and 30%, or 0.75 kPa, for the C1-6-D, C1-6VN-D, L2-9-D and L2-9VN-D, whichever is greater over the range of depths for which measurements can be made, limited by shear wave penetration.	SD less than or equal to 10% for multiple repeated measurements over the range of depths for which measurements can be made, limited by shear wave penetration.

Measurement	Units	Useful Range	Accuracy	Limitations or Conditions
The formula for Stiffness: $E = 3 \cdot \rho \cdot c^2$ Where E = Young's modulus of tissue ρ = density of tissue (assumed to be 1 g/cc) c = shear wave speed in m/s Note: The conversion from shear wave speed (m/s) to Young's modulus is done under the assumption that the underlying material in which the shear wave propagates is linear, isotropic, incompressible, and homogenous.				

Clinical Calculation Accuracy

Estimate the overall inaccuracy of a combined measurement and calculation by including the stated inaccuracy from the basic measurement accuracy statements.

Calculation formulas and databases are provided as a tool to assist the user, but should not be considered an undisputed database, in making a clinical diagnosis. The user is encouraged to research the literature and judge the equipment capabilities on an ongoing basis in order to assess its utility as a clinical tool.

V Nav Magnetic Field Data For the Patient



DANGER

DO NOT use the Volume Navigation feature on any patient relying on life-sustaining electronic equipment, such as a pacemaker or defibrillator. Failure to follow this instruction could lead to interference with patient electronic device(s).



WARNING

The Tracker's magnetic fields may possibly interfere with nearby electrical systems, e.g., an EKG. It is the user's responsibility to identify nearby devices to ensure their performance is not degraded when you are simultaneously using Volume Navigation.

Electromagnetic Field Strength is measured in units called Tesla, or for smaller fields, milliTesla (mT). According to this article from the World Health Organization (WHO), <https://www.who.int/peh-emf/publications/facts/fs299/en/>, people with pacemakers should avoid fields that exceed 0.5 mT.

The electromagnetic field created by the transmitter in our Volume Navigation offering is 0.5 mT once you are 7.5cm away from the transmitter face and goes down quickly after that (0.1 mT at 18cm away, e.g.). Nevertheless, we recommend not using V Nav on patients relying on a pacemaker or defibrillator.

For the Operator

The World Health Organization website addresses many items regarding exposure to electro-magnetic (EM) fields at <http://www.who.int/peh-emf/about/WhatIsEMF/en/index.html> (see extract below).

Extracted from World Health Organization:

"Many people are surprised when they become aware of the variety of magnetic field levels found near various appliances. The field strength does not depend on how large, complex, powerful or noisy the device is. Furthermore, even between apparently similar devices, the strength of the magnetic field may vary a lot. For example, while some hair dryers are surrounded by a very strong field, others hardly produce any magnetic field at all. These

differences in magnetic field strength are related to product design. The following table shows typical values for a number of electrical devices commonly found in homes and workplaces. The measurements were taken in Germany and all of the appliances operate on electricity at a frequency of 50 Hz. It should be noted that the actual exposure levels vary considerably depending on the model of appliance and distance from it.”

In the context of this, the VNav transmitter has the following EM field strengths at various distances from the transmitter face:

- 928 μ T at 5.1cm (2 inches) distance
- 287 μ T at 10.2cm (4 inches) distance
- 64 μ T at 20.3cm (8 inches) distance
- 11 μ T at 40.6cm (16 inches) distance
- 2.4 μ T at 71cm (28 inches) distance

NOTE

Note that the field is only active when the green light on the position sensing electronics unit is on. This is the same unit that the transmitter plugs into.

Typical Magnetic Field Strength Data

Typical magnetic field strength commonly found in your home are shown below:

Table 12-11 • Typical magnetic field strength of household appliances at various distances

Electric Appliance	3 cm Distance (μ T)	30 cm Distance (μ T)	1 m Distance (μ T)
Hair Dryer	6 – 2000	0.01 – 7	0.01 – 0.03
Vacuum Cleaner	200 – 800	2 – 20	0.13 – 2
Portable Radio	16 - 56	1	< 0.01
Microwave Oven	73 - 200	4 - 8	0.25 - 0.6
Fluorescent Light	40 – 400	0.5 - 2	0.02 - 0.25
Electric Oven	1 - 50	0.15 - 0.5	0.01 - 0.04

Privacy and Security

Privacy & Security considerations, capabilities and configuration of the LOGIQ Fortis Ultrasound System are described in the LOGIQ Fortis Privacy and Security Manual, available in the GE HealthCare Product Security Portal at the GE HealthCare Documentation Library website at:

<https://www.gehealthcare.com/support/documentation>

How to Contact GE HealthCare

For privacy and security concerns regarding GE HealthCare products, visit: <https://www.gehealthcare.com/security>

Privacy & Security Environment

The GE HealthCare LOGIQ Fortis Ultrasound System has been designed for an intended use with the following expectations of Privacy & Security protections included in the environment where this product will be used:

- The system should be connected to a secured LAN or VLAN, configured and segmented according to appropriate networking best practices, not open to unintended users or generally connected to a WAN.
- The LOGIQ Fortis Ultrasound System should be physically secured in such a way that it is not accessible to unintended users.
- Default application users and passwords should be replaced with customized users and passwords.
- External media containing images, patient data, reports and logs should be secured. When no longer used, the data should be securely erased and/or the media should be securely deleted.
- The monitors of the LOGIQ Fortis Ultrasound System should be placed in a way limiting the visibility of the screen content to the user only.

The LOGIQ Fortis is designed as a self-contained system, with system-level access reserved only for qualified GE HealthCare Service personnel. The LOGIQ Fortis is designed to be secure-by-default, with all security controls in place by default with no user intervention required. If customers experience issues or concerns related to cybersecurity, they should contact their GE HealthCare Service representative as soon as possible. If GE HealthCare makes the determination that a cybersecurity incident has occurred, GE HealthCare will take the appropriate steps to return the device to proper working order accordingly.

Identity Provisioning

The provisioning of user accounts includes the steps of account creation, maintenance, and suspension of the account when it is no longer needed. A user account is created for the use by a specific individual. It is associated with access rights.

Management of user accounts

User accounts are created, maintained and suspended by users with administrator privileges. When received from factory the system has three predefined user accounts:

- “ADM” : an administrative user account
- “USR” : a normal user account template without administrative privileges
- “EUSR” : an emergency user account template with limited privileges

When receiving the Ultrasound System, it is recommended to do the following steps to ensure control of the user accounts on the system:

- Change the password of the “ADM” account.
- Change the password of the “USR” user account.
- Create groups for local or LDAP users and set their rights/privileges appropriately for your workflows and operating environment.
- Create user accounts for each individual user of the system:
 - Give each user the needed group memberships / privileges.
 - Make sure to give administrative privileges only to users intended to perform administrative tasks on the system, like configuring dataflows, managing users on the system, inspecting audit logs etc. As administrative privileges may give the user access to privacy and security related configurations on the system, there should be a limited number of users with these privileges.
 - It is recommended to create individual users for each user of the system. This is particularly relevant for the audit logging, to associate actions performed on the system with individual persons.

Maintenance of user accounts

It is recommended to establish administrative routines for removing user accounts no longer being used.

User information stored on the system

The following information can be entered for a user defined on the system:

- User ID / User name (Required)
- Display ID (Optional)
- Last name (Required)
- First name (Optional)
- Title (Optional)
- Phone number (Optional)
- E-mail address (Optional)
- Address (Optional)

The user password is encrypted and not accessible in the system.

User name and password restrictions

The restrictions for usernames and passwords are:

- User names can be 1 - 32 characters long.
- Password can be 0 - 256 characters long.

It is recommended to enable and configure username and password policies as described below.

Identity provisioning by use of Backup/Restore

Users defined on a system can be copied from one system to another by use of the system's built-in Backup/Restore functionality.

Network Infrastructure

The infrastructure of the network where the LOGIQ Fortis Ultrasound System is connected must be configured to allow traffic. For information on inbound and outbound firewall configuration, refer to the LOGIQ Privacy and Security Manual, available in the GE HealthCare Product Security Portal at:

<https://www.gehealthcare.com/security>

Audit Logs

The LOGIQ Fortis systems provide an audit log feature which enables users to export the audit logs and detect any recorded events. Audit logs typically capture events by recording who performed an activity, what activity was performed, and how the system responded. In addition, LOGIQ Fortis supports the capability for administrators to import audit logs into customer-owned Security Incident and Event Monitoring (SIEM) solutions for additional monitoring functionality.

The types of events logged locally are as follows:

- User authentication and failed login attempts
- Configuration changes
- Creating, deleting, transfer, and access to patient records
- Role-based access control groups and changes
- DICOM events (Store, Commit, etc.)

In addition, SIEM integration adds Operating System-level event logs, which include anti-malware, code integrity, and system integrity protection logs in addition to the log types above. Monitoring these logs is the responsibility of the healthcare institution.

Anti-Virus Software Note

Security

Since the LOGIQ systems are integrated into your IT-network, GE HealthCare wants to make sure that you are aware of the proactive measures we are taking to secure the system. Below are measures we have implemented to secure the LOGIQ systems.

- Use of Windows* Embedded Standard 10, a componentized version of Windows 10 specifically made for embedded systems. Only the components required are used for the LOGIQ scanners, thereby reducing the OS attack surface. Please note that Windows 10 IoT Enterprise is NOT the same operating system as Windows 10.
- Disabled the user's ability to access the internet and Windows desktop.
- Disabled, or made inaccessible, functionality that is typically used as malware vectors for spreading viruses (e.g. email services, web browsers).
- Disabled AutoRun functionality on removable media.
- Closed network entry points that are not in use by the LOGIQ scanner software by strict firewall configuration and by disabling Services. The only Internet connection needed is on outbound port to GE HealthCare's remote service platform (Insite™ ExC), which is only opened on request by the user and through a secure HTTPS connection (port 443), and inbound and outbound connections through port 104 for DICOM connectivity.
- System provides user access controls under customer management to control scanner access.
- Uses secure integration and communication between systems (Scanners, Workstations and Servers).
- Monitor public security bulletins from software vendors and news services, analyze for applicability to the LOGIQ scanner, and include third party software security patches as necessary within GE HealthCare software.
- Release GEHC Ultrasound validated software or use other measures as necessary to resolve or mitigate product vulnerabilities.
- Assess potential vulnerabilities of our systems using up-to-date commercially available vulnerability scanning tools. Identified vulnerabilities are mitigated as appropriate based on risk assessment of the product.

GE HealthCare believes that this Defense in Depth strategy using the combination of the security measures above and the security standards of Microsoft Windows 10 IoT Enterprise will provide security against malware, especially for a system used in a professional hospital grade networking environment that itself should provide a high level of security measures.

Finally, a few points as to why GE HealthCare (as well as all other manufacturers of PC-based medical Ultrasound devices) do not use Anti-Virus software: Commercial Anti-Virus software is commonly used on general-purpose computers to detect the presence of malicious software (e.g. virus, Trojan horse, worm). Anti-Virus software is useful on general-purpose computers as they typically cannot be sufficiently hardened against the attack vectors used by malicious software.

The LOGIQ ultrasound systems however are single purpose (dedicated) devices that have controlled intended use, and thus are well hardened. For the LOGIQ ultrasound systems, the potential patient safety and security risks introduced by using commercial Anti-virus software would outweigh the security benefits. Such risks include:

- Real-time anti-virus scanning can affect ultrasound system performance.
- The effectiveness of Anti-Virus software depends on regular updates of the virus definitions files. This would typically require internet connectivity for the ultrasound system.
- The Anti-Virus software itself is a popular attack vector.
- Disruptive nature of the support of the Anti-Virus software throughout the life cycle of the medical device. The operating system of a medical ultrasound system is part of a medical device that requires a special and controlled release process. Any update of the Anti-Virus software would require a change of the system software.

Due to the cited risks, the use of commercial anti-virus software is not part of the LOGIQ systems product security strategy.

*Microsoft and Windows are trademarks of Microsoft Corporation.

Loading Windows Patches

Loading a Windows Patch

As part the product lifecycle management, GE HealthCare regularly analyzes and integrates software updates from our third party vendors into our products. These are typically released as part of regular updates or software releases.

To load a Windows patch onto the ultrasound system:

1. Power down the ultrasound system and insert the Windows Patch USB Flash Drive into a rear USB port.

NOTE

Ensure that the system is USB Device Enabled (check setting on System Admin Utility page).

2. Power on the ultrasound system. Windows Patch files will be loaded onto the ultrasound system automatically, following several screen prompts:

Select Install SW... on the Start Application screen.

Select OK on the first StartLoader screen.

Select the package and then select Install on the second StartLoader screen; software patch installation will begin.

The Patch package is installed at the root folder or under the SWLoad directory in the root folder of the USB.

Multiple screens appear during the software re-imaging installation process. DO NOT interrupt this process AND follow instructions as they appear on the display.

Once the installation is complete, a message displays saying that the “Software Installation completed successfully. System will reboot soon.” The ultrasound system restarts.

3. When the system starts up after the software installation has finished, the following dialog displays: the “New Software Verification” Checklist.

Performing Software Patch Verification

The Software Patch Verification dialog is critical. You **MUST** perform software verification after downloading and installing a software update.

If you are able to successfully perform each function, then Pass all of the items and type your name into the Signature box and press **OK**. Press the Question Mark if you have questions about how to perform this task.



CAUTION

Perform a check for all the features listed. You **MUST** ensure that the entire system functions normally, as expected, in each of the categories listed on the New Software Verification checklist.

These verification results are tracked for regulatory purposes, sent back to GE HealthCare for tracking, and approved with your signature.

As you verify that each feature works correctly, select “Passed.” If all features work correctly and “Passed” is filled in for all features, then the signature field is enabled at the bottom of the New Software Verification Checklist.


Type your signature (minimum of three characters) and press **OK**. The system is now ready for use.



CAUTION

If any of the features **DID NOT** function as expected, you need to select “Failed” next to the feature that failed. Type your signature (minimum of three characters) and press **Reload**.

The version of the installed Patch appears on the **Utility > System > About** page.



If one of the steps fails, then Fail the item(s).



Type your Signature, then press **Reload**. The previous version of software will be reloaded onto the system automatically.

Load the latest system patch which should be stored with the software located under the ultrasound system covers.

Patch Installation Notes

If by accident you try to load a patch that is not compatible with the software on the ultrasound system, an error message will notify you of this incompatibility.

If there is any issue with the media, an error message will indicate “The package cannot be installed. The package is not compatible or has been tampered. Please contact GE HealthCare Service.”

The system may reboot multiple times during patch/software update.

If the software reload fails, an error message will indicate “Software Reload operation failed. Please contact GE HealthCare Service.”

Reloading Software

When reloading software, the latest released system patches should be loaded in the system software.

Software Download

Overview

The LOGIQ Fortis is designed to download software updates when they are available. Software Download monitors, notifies, delivers and installs available system software updates. This feature requires active InSite ExC connectivity.

When system software is available for download, a Software Download icon appears in the system status bar at the bottom of the monitor screen.

Table 12-12 • Software Download Icons

	Software Download Available Icon
	Software Download In Progress Icon
	Software Download Paused Icon
	Software Download Complete/Ready to Install Icon
	Software Download Information Icon

Software Download Available

When the “Software Download Available” icon appears in the status bar, select the icon to bring up the following Software Download and Install menu. Choose the desired option.




Figure 12-1 • Software Download and Install Menu

Start Software Download and Install

1. Select **Start** to begin the software download. The software download status displays in the menu and the Software Download in Progress icon appears in the system status bar at the bottom of the monitor.



Figure 12-2 • Software Download in Progress

While the software is downloading, you can choose to pause the download, hide the Software Download and Install menu or stop the download.

- Select **Pause** to temporarily pause the download. The Software Download Paused icon appears in the system status bar. Select the icon again to bring up the Software Download menu and press **Resume** to resume the download from the previous download point.
 - Select **Hide** to hide the Software Download menu (the download continues in the background). The Software Download in Progress icon appears in the system status bar. Select the icon again to bring up the Software Download menu.
 - Select **Stop Download** to discontinue the download process. The download progress is not saved, and the Software Download Available icon must be selected again to bring up the Software Download menu to restart the software download process.
2. When the download is complete the Software Download and Install menu informs you the software is ready to install.



Figure 12-3 • Software Ready to Install Menu

Select **Start** to begin the installation. The installation begins and the software Installation in Progress status is displayed.

3. New Software Verification - When the software package installation is complete, the system shuts down and restarts. When the system restarts, the New Software Verification checklist dialog displays. You **MUST** perform the new software verification.


The image shows a 'New Software Verification' dialog box with a dark background. At the top, it says 'New software is installed. Functional checks are required to verify that the product works as intended. Please check the following:' followed by a '<<' button. Below this is a list of features: 3D Mode, M Mode, CF Mode, PW / CW Doppler Mode, Probes, Patient Archive, Presets, and Peripherals. Each feature has three buttons: 'Passed' (with a blue circle icon), 'Failed' (with a red circle icon), and a question mark icon. At the bottom, there is a 'Signature' field with a blue bar next to it.

Figure 12-4 • New Software Verification Dialog



CAUTION

Perform a check for all the features listed. You **MUST** ensure that the entire system functions normally, as expected, in each of the categories listed on the New Software Verification checklist.

As you verify that each feature works correctly, select “Passed.” If all features work correctly and “Passed” is filled in for all features, then the signature field is enabled at the bottom of the New Software Verification Checklist. Type your signature (minimum of three characters) and press **OK**. The system is now ready for use.

If any of the features DID NOT function as expected, you need to select “Failed” next to the feature that failed. Type your signature (minimum of three characters) and press **Reload**.

Contact your GE HealthCare Service Representative for assistance.

More Info

Select **More Info** from the Software Download and Install menu to see details about the available software download.



Figure 12-5 • More Info

Remind Me Later

Select **Remind Me Later** to close the Software Download and Install Menu and download the software at a later time. The Software Download Available icon remains in the system status bar at the bottom of the monitor screen.

Decline

Select **Decline** to decline the software download package. A popup displays.



Figure 12-6 • Decline Software Download Package Popup

You can then select OK or Cancel.

- Select OK to decline the package. You must then enter a signature consisting of three or more characters.
- OR
- Select Cancel to go back to the Software Download and Install menu.

Base Image and Software Load

Base Image and Software Load Procedure

For the Base Image and Software Load procedure, refer to the Basic Service Manual.

Copyrighted Material

Viewing Copyrighted Third Party Software License Information

The LOGIQ Fortis contains copyrighted material. For more information, you can view Copyrighted Third Party Open Source Software Licenses via the **Utility > System > Licenses** page.

- Select **Utility > System > Licenses**
- Scroll to select the license you wish to view in the License Title section. Press **Set**.



Figure 12-7 • Third Party Software License Information

- The selected software license appears in the License window. Use the Set+Trackball to scroll through the license.

System Care and Maintenance

Overview

Refer to Section 10 of the LOGIQ Fortis Service Manual for any additional maintenance guidance.

The user must ensure that safety inspections are performed at least every 12 months according to the requirements of the patient safety standard IEC 60601-1. Refer to the Service manual, Chapter 10.

Only trained persons are allowed to perform the safety inspections mentioned above.

Technical descriptions are available on request.

To ensure that the unit constantly operates at maximum efficiency we recommend that the following procedures be observed as part of the customer's internal routine maintenance program.

Contact the local Service Representative for parts or periodic maintenance inspections.

Expected Service Life Description

The expected service life for the LOGIQ Fortis system and probes is identified in this table:

Table 12-13 • Expected Service Life

Equipment / Accessory	Expected Service Life
LOGIQ Fortis system	The expected service life for the LOGIQ Fortis is at least seven (7) years from the manufacturing date under the provision of regular maintenance by authorized service personnel.
Probes	The expected service life for the probes meets or exceeds five (5) years from the date the probe is placed in service, under the provision that the customer follows the care instructions provided on the Probe Care Card / Accompanying LOGIQ Fortis Instructions for Use.

Maintenance Schedule

Follow this Maintenance Schedule to maintain optimum system function and patient care:

Table 12-14 • Maintenance Schedule

Monthly	Weekly	Daily	After Each Patient
Inspect the following on a monthly basis: - Connectors on cables for any mechanical defects. - Entire length of electrical and power cables for cuts or abrasions. - Equipment for loose or missing hardware. - Control panel and keyboard for defects. - Casters for proper movement and locking operation.	Clean the following on a weekly basis: - Console - System Cabinet - Removable Trackball/Trackball - Air Filters (weekly, or as needed) - Footswitch - B/W Printer - CD/DVD-R Multi Drive - Monitor and Monitor Frame - Operator Controls and Touch Panel	Clean and Disinfect the following areas where Cross Contamination can occur: - Operator Panel and Touch Panel - Monitor Frame - Front and Rear Handles	Clean and Disinfect the following after each patient: - Probe - Biopsy Bracket, as applicable Additionally, Clean and Disinfect any area on the system that has visible contamination from the previous exam. Note: Biopsy Accessories must be cleaned and disinfected or disposed of after each patient. Refer to the Probes Chapter, for probe cleaning and disinfecting instructions.

**DANGER**

To avoid electrical shock hazard, do not remove panels or covers from console. This servicing must be performed by qualified service personnel. Failure to do so could cause serious injury.

If any defects are observed or malfunctions occur, do not operate the equipment but inform a qualified service person. Contact a Service Representative for information.

Cleaning and Disinfecting the System

**WARNING**

All cleaners and disinfectants NOT on this list are unapproved by GE HealthCare. Failure to follow guidelines could result in damage to the device.

**CAUTION**

When reprocessing the operator control panel, make sure not to spill or spray any liquid on the controls, into the system cabinet, or in the probe connection receptacle.

**CAUTION**

To avoid liquids entering the product, DO NOT spray any liquid directly onto the surfaces. ALWAYS use a cloth or wipe.

**CAUTION**

Avoid using ALCOHOL (ISOPROPANOL) 70% inside the trackball. ALCOHOL (ISOPROPANOL) 70% may also compromise the durability of the paint used on the console controls.

**CAUTION**

Oxivir TB Wipes may compromise the durability of the paint used on the console handles and controls. Avoid the use of any Hydrogen Peroxide (H₂O₂) based disinfectants on the HDU monitor, as it may result in damage to screen.

These cleaners/disinfectants can be used anywhere on the console (Operator Panel, Monitor, Probe Holders, etc.), except for the probes. Refer to *Probe Disinfectants* on page 646 for probe disinfectant information and web links.

Always consult the cleaner or disinfectant manufacturer's instructions for proper use of their product. Wear appropriate PPE as indicated by the manufacturer.

Appropriate cleaners/disinfectants for the console that have been validated for compatibility are shown below:

Table 12-15 • Appropriate Cleaners/Disinfectants

Product	Manufacturer	Notes	Chemistry	Locale	Cleaner/ Disinfectant
Mild Soap					Cleaner
Acryl-Des Wipes	Schülke & Mayr GmbH	Same as Mikrozid Sensitive Wipes also by Schulke	Quat	Europe	Cleaner/ Disinfectant
Alcohol (Isopropanol) 70%	Generic		Alcohol	Global	Disinfectant
Cleanisept Wipes	Dr. Schumacher GMBH	Same as GE branded Septiwipe	Quat	Europe	Cleaner/ Disinfectant
Clinell Clorox Wipes	GAMA Healthcare	Same as Clorox Wipes sold in USA	Bleach	Europe	Cleaner/ Disinfectant
Clinell Universal Sanitizing Wipes	GAMA Healthcare		Quat+ Polyhexanide (PHMB)	Europe	Cleaner/ Disinfectant
Clorox Healthcare Bleach Germicidal Wipes	Clorox Professional	Same formula as Clinell Clorox Wipes	Bleach	US Canada Europe	Cleaner/ Disinfectant
Easy Screen Cleaning Wipes	Professional Disposables Inc. (PDI)	This product is 70% IPA	Alcohol	US	Cleaner
Mikrobac Tissues	BODE Chemie GmbH		Quat	Europe	Cleaner/ Disinfectant

Product	Manufacturer	Notes	Chemistry	Locale	Cleaner/ Disinfectant
Mikrozyd Sensitive Wipes	Schülke & Mayr GmbH	Same formula as Acryl-Des wipes	Quat	Europe	Cleaner/ Disinfectant
Oxivir TB Wipes	Sealed Air		H2O2	US Canada	Cleaner/ Disinfectant
Protex Ultra Wipes	Parker Laboratories	Same formula as SONO Wipes	Quat	US	Cleaner/ Disinfectant
Sani-Cloth HB Germicidal Disposable Wipe	Professional Disposables Inc. (PDI)		Quat	US	Cleaner/ Disinfectant
Sani-Cloth Plus Germicidal Disposable Cloth	Professional Disposables Inc. (PDI)	Same Formula as Asepti Wipes II	Quat + Alcohol	US	Cleaner/ Disinfectant
Sani-Cloth Prime Germicidal Disposable Cloth	Professional Disposables Inc. (PDI)		Quat + Alcohol	US	Cleaner/ Disinfectant
Septiwipes	EDM Medical Imaging	Same formula as Cleanisept Wipes	Quat	Europe	Cleaner/ Disinfectant
Sodium Hypochlorite 5.25% (Bleach) diluted 10:1	Generic	Clorox Healthcare Bleach Germicidal Wipes	Bleach	Worldwide	Cleaner/ Disinfectant
SONO ULTRASOUND WIPES	Advanced Ultrasound Solutions, Inc.		Quat	US	Cleaner/ Disinfectant
Super Sani-Cloth Germicidal Disposable Wipes	Professional Disposables Inc. (PDI)		Quat + Alcohol	US	Cleaner/ Disinfectant
Tristel Distel	Tristel		Quat+PHMB	Europe	Cleaner/ Disinfectant
Trophon Companion Cleaning Wipes	Nanosonics	Same formula as SONO Wipes	Quat	US	Cleaner/ Disinfectant
Virox Accel TB wipes	Virox Technologies Inc. (owned by Sealed Air)	Same formula as Oxivir TB	H2O2	CAN	Cleaner/ Disinfectant

Console Cleaning the Console

The LOGIQ Fortis console includes the System enclosure, Monitor, Monitor Frame, Operator Panel, Touch Panel, and Probe Holders. For Probes Reprocessing, see *Probe Processing* on page 624.

Always clean visible soil from surfaces first before disinfecting the Console.

Follow the cleaning/disinfecting frequency suggested in *Maintenance Schedule* on page 702.

To clean the system,

1. Moisten a soft, non-abrasive folded cloth with a mild, general purpose, non-abrasive soap and water solution or approved cleaning/disinfecting agent.

NOTE

The cloth/wipe should be damp, not dripping wet and running. Moisture should not drip into the crevices anywhere on the console.

NOTE

Refer to *Cleaning and Disinfecting the System* on page 703 for a list of acceptable solutions to be used on the Console.

2. Use a gentle wiping action to clean any surface on the console.



HINTS

A scrubbing action with the wipe may be necessary to help remove stubborn soil from the surfaces. However be careful with this action over crevices and gaps in the surface to prevent liquid from being scraped off the wipe and entering the product.

3. Wipe off excess cleaning agents.

NOTE

Do not spray any liquid directly into the unit.

NOTE

DO NOT scratch or press on the panel with any sharp objects, such as pencils or pens, as this may result in damage to the panel.

Disinfecting the Console

For disinfectants to be effective, the surface must first be clean. Refer to *Cleaning the Console* on page 706.

ALWAYS follow the manufacturer's instructions concerning the use of the disinfectant and follow the contact time to be sure the disinfectant's kill claims are accomplished.

Follow the cleaning/disinfecting frequency suggested in *Maintenance Schedule* on page 702.

Disinfect the desired surfaces of the console. To prevent cross contamination, surfaces that are often touched during exams should be disinfected after every patient. To disinfect the system,

1. Moisten a sterile cloth with a liquid disinfectant or remove pre-moistened disinfectant wipe from the container.

**CAUTION**

If a cleaner/disinfectant wipe was used to clean off visible soil per the above section, a second, fresh cleaner/disinfectant product should be used for the disinfectant step.

2. Wet the surfaces by gently applying the cloth or wipe. Avoid high pressure or squeezing the wipe to avoid having the liquid enter the gaps and crevices of the Console. Scrubbing is not necessary in the disinfecting step; evenly applying the liquid is the goal.
3. Let the surface remain wet for the appropriate contact time.
4. If the surface does not remain wet for the full contact time, apply an additional application of the disinfectant, as necessary, to extend the time.
5. After the contact time has expired, remove excess liquid with a dry sterile cloth.
6. To avoid disinfectant buildup, or to remove disinfectant residue which may cause skin irritation, perform a rinse step with a sterile damp cloth.

Probe Cleaning Notes

When cleaning/disinfecting probes, take care NOT to damage the Console with a solution that is approved for the probe but that IS NOT APPROVED for the Console. Refer to *Probe Disinfectants* on page 646 for a list of acceptable solutions to be used on the Console. Refer to the Probes Chapter, for probe cleaning and disinfecting instructions.

**CAUTION**

NEVER use any cleaner or disinfectant containing alcohol.

When cleaning/disinfecting probes using a spray cleaner/disinfectant, DO NOT spray the probe while the probe is set in its probe holder on the Ultrasound system (1). Overspray can damage the TGC controls.

If you use a spray cleaner, spray AWAY from the Ultrasound system (2).

If you are cleaning/disinfecting probes while they are on the Ultrasound system, use a wipe cleaner/disinfectant instead (3).



Figure 12-8 • DO NOT Spray a Probe While in its Holder

Monitor and Monitor Frame

Monitor

To clean the Monitor:

1. Moisten a soft, non-abrasive folded cloth with a mild, general purpose, non-abrasive soap and water solution.

NOTE

The cloth should be damp, not dripping wet.

2. Wipe down the top, front, back, and both sides of the monitor.
3. Wipe off excess cleaning agents.

NOTE

Never use thinner, benzene, alcohol (ethanol, methanol, or isopropyl alcohol), abrasive cleaners, or other strong solvents, as these may cause damage to the monitor.

Monitor Frame

To clean the monitor frame:

1. Moisten a soft, non-abrasive folded cloth with a mild, general purpose, non-abrasive soap and water solution.

NOTE

Moisten a soft, non-abrasive folded cloth with a mild, general purpose, non-abrasive soap and water solution.

2. Wipe down the top, front, back, and both sides of the monitor frame.
3. Wipe off excess cleaning agents.

Other acceptable cleaning agents are:

- Ammonia
- Bleach (10 to 1 ratio of water to 5% home bleach)
- Hydrogen Peroxide / Hydrogen Peroxide Wipes

NOTE

DO NOT scratch or press on the panel with any sharp objects, such as pencils or pens, as this may result in damage to the panel.

Operator Controls and Touch Panel

ONLY USE the following cleaners on the Operator Panel:

- Palmolive Dishwashing Liquid (manufactured by Colgate-Palmolive, www.colgatepalmolive.com)
- Sani-Cloth® HB, green cap (manufactured by Professional Disposables International, www.wearepdi.com). NOTE: May cause yellowing to the operator panel and probe cabling.
- Super Sani-Cloth®, purple cap (manufactured by Professional Disposables International, www.wearepdi.com)
- Sani-Cloth® Prime, plum cap (manufactured by Professional Disposables International, www.wearepdi.com)
- Sono® Ultrasound Wipes (manufactured by Advanced Ultrasound Solutions, Inc, www.UltrasoundWipes.com)
- Bleach (10 to 1 ratio of water to 5% home bleach)
- Hydrogen Peroxide / Hydrogen Peroxide Wipes

DO NOT USE:

- Cleaners or Disinfectants NOT listed above

Cleaning the Trackball

To clean the Trackball,

1. Twist and remove the Trackball prior to cleaning both the Trackball and Trackball housing (1 - 3).
2. Clean the Trackball and Trackball Housing with dry, soft cloth (4 - 6).
3. After cleaning the Trackball, replace and twist the Trackball into the Trackball housing (7-8).



Other Maintenance

Cleaning the air filter

Clean the system's air filters to ensure that a clogged filter does not cause the system to overheat and reduce system performance and reliability. It is recommended the filters be cleaned every two weeks, but the requirements will vary due to your system use.



CAUTION

Be sure to lock the wheels before cleaning the air filters to avoid injury by any unexpected movement of the system.

DO NOT operate the unit without the air filters in place.

Allow the air filters to dry thoroughly before re-installing them on the unit.

Cleaning

1. Pull the front cover of cabinet.



Figure 12-9 • Remove the front cover

2. Pull out the air filter with the frame.



Figure 12-10 • Remove the air filter

3. Dust the filter with a vacuum cleaner and/or wash it with a mild soapy solution.
If washed, rinse and dry the filter before re-installation.
4. Put back the air filter with both hands and the front cover.



Figure 12-11 • Put back the air filter
LOGIQ Fortis™ - User Manual
5919773-1EN Rev. 2

Cleaning Storage Tray and TVTR holder Storage Tray

1. Loosen a screw ((1) in *TVTR Probe Holder* on page 713) to remove the tray from the console.
2. Wash the tray with mild detergent.
3. Thoroughly dry the tray using a clean, low-lint, soft cloth or wipe.
4. Attach the tray to the system. Screw tightly.



Figure 12-12 • Screw of the storage tray

TVTR Probe Holder

1. Pull the claw to remove the probe holder from the console.



Figure 12-13 • Pull the claw to remove

2. Wash the probe holder with mild detergent.
3. Thoroughly dry the probe holder using a clean, low-lint, soft cloth or wipe.
4. Put the probe holder to the original position.

Replacing key caps

Contact a local Service Representative when a key cap needs to be replaced.

For Feature Keys, the user can rearrange these keys if desired. Please see *System/User Configurable Key* on page 470 for more information on how to remove and replace these Feature keycaps.

Footswitch

To clean the footswitch:

1. Disconnect the footswitch from the ultrasound system.
2. Moisten a soft, non-abrasive folded cloth with a mild, general purpose, non-abrasive soap and water solution.

NOTE

The cloth should be damp, not dripping wet.

3. Wipe the external surfaces of the unit then dry with a soft, clean, cloth.

CD/DVD Discs

To clean medical grade CD/DVD discs:

- If the disc is dirty, you can clean it with a commercially available dust removal spray. Be careful not to scratch the disc when wiping it off.

NOTE

CD type cleaner cannot be used.



HINTS

Keeping your media disc in an original media case or caddy all the time will prevent it from becoming dirty or damaged.

Quality Assurance

Introduction

A good Quality Assurance Evaluation program consists of periodic systematic actions that provide the user with adequate confidence that their diagnostic ultrasound system will produce consistently high quality images and quantitative information.

Therefore, it is in the best interests of every ultrasound user to routinely monitor equipment performance.

The frequency of Quality Assurance evaluations should be based on user's specific needs and clinical practice.

Periodic monitoring is essential in order to detect the performance changes that occur through normal aging of system components. Routine equipment evaluations may also reduce the duration of exams, number of repeat exams, and maintenance time required.

NOTE

For information on setting up the Image Quality Check, refer to *Image Quality Check* on page 726.

For details on system and peripheral routine preventive maintenance instructions, *System Care and Maintenance* on page 702

Typical Tests to Perform

Quality assurance measurements provide results relating to system performance. Typically these are:

- Axial Measurement Accuracy
- Lateral Measurement Accuracy
- Axial and Lateral Resolution
- Penetration
- Functional & Contrast Resolution
- Gray Scale Photography.

With these tests, a performance baseline can be set at installation with the phantom in your department. Future test results can be compared to the baseline in order to maintain a record of system performance trends.

The phantom shown is shown as a representative example of a phantom. You can select from any number of phantoms available on the market.

Frequency of tests

Quality assurance tests are used to determine whether a scanner is providing the same level of performance from day to day.

The frequency of testing varies with the amount of system usage and modes to be tested. It is recommended that the user perform quality assurance tests at least every three months

or every 400 patient studies. Tests should also be performed when a question about system performance exists.

A mobile system may require more frequent tests.

Image quality should also be tested immediately after the following events:

- Service calls
- System upgrades/modifications
- Dropped probe, power surge, etc.

Phantoms

Quality Assurance Evaluations may be done with phantoms and test objects that are applicable to the parameters being evaluated or to the user's clinical practice.

Typical phantoms are composed of material that acoustically mimic human tissue. Pins, anechoic and echogenic targets are physically positioned to provide information for a variety of tests.

The RMI 403GS phantom is shown in the illustration below as a representative example of a phantom.




Figure 12-14 • Representative Phantom Example

1. Penetration
2. Axial Distance Measurement
3. Functional Resolution
4. Lateral Resolution
5. Lateral Distance Measurement
6. Axial Resolution
7. Contrast Resolution and Gray Scale Photography
8. Gray Scale Plane Targets

Baselines

An absolute necessity for a quality assurance program is establishing baselines for each test or check. Baselines are established after the system has been verified to be working properly at installation or after a repair. If a probe or major assembly is replaced, new baselines should be generated.

Baselines can be made by adjusting system parameters to prescribed levels or to the best possible image. The key factor to remember is reproducibility. The same conditions must be reproduced for each periodic check.

All system parameters not displayed on the monitor should be recorded for the permanent record.

Periodic Checks

Periodic checks should be performed in accordance with your facility's quality assurance requirements. For the data to be valid, periodic checks should mimic the baseline setup parameters.

The resulting image, when scanning the phantom exactly as before, should be recorded and compared to the baseline. When a matching image is obtained, it can be assumed that the system performance has not degraded from the baseline.

If a significant difference between the baseline and periodic check is noted, double check the system setup and repeat the test. If the difference between the baseline and periodic check persists, contact a local Service Representative.

Failing to reproduce the control settings as in the baselines will introduce errors in the data and potentially invalidate the results.

Results

Lack of standardization among test instruments, the wide range of acceptance criteria, and incomplete knowledge regarding the significance of certain performance parameters prohibit the establishment of absolute performance criteria for these tests.

Quality Assurance Evaluation results should be compared to previously-recorded results.

Performance trends can then be detected. Unacceptable performance or diminishing trends should be identified for maintenance or repair before a malfunction or inappropriate diagnosis occurs.

The user should determine the best method for recording and archiving the baseline and periodic checks. In most cases the choice is hard copy.

It is important to maintain good consistent records for inspections that may arise, as well as to detect system performance trends.

System Setup

The user should tailor the tests to their particular needs. It is certainly not necessary to make all checks with all probes. A representative example, with the probes used most often by the customer, should be adequate in judging system performance trends.

Use a gray scale phantom as the scan object for the tests. Commercial phantoms are supplied with its own operator manual. Be familiar with proper phantom operating procedures prior to use for quality assurance evaluations.

1. Adjust image monitor. Brightness and contrast should be set to the normal viewing of a good gray scale image.
2. Check all recording devices for proper duplication of image monitor. Ensure that what is seen is what is recorded.
3. Annotate non-displayed image processing controls.
4. Set TGC slide pots to center (detent) position.
5. Place focal zone marker(s) in area of interest for an optimum image.

Test Procedures

The following are recommended Quality Assurance tests. A brief description of the test, the benefit it provides and steps to accomplish the test are supplied.

The importance of recording scan parameters and consistent record keeping cannot be stressed enough. Reproducibility to monitor system trends is the key to quality assurance evaluations.

Using the system's dual image display format is often very convenient and saves recording media.

Axial distance measurements

Description

Axial measurements are the distance measurements obtained along the sound beam. See *Figure 12-14 • Representative Phantom Example* on page 716 for more information.

Benefit

The accurate measurement of the size, depth and volume of a structure is a critical factor in determining a proper diagnosis. Most imaging systems use depth markers and/or electronic calipers for this purpose.

Method

Axial distance should be measured in the near, mid and far fields as well as in zoom. If necessary, different depths or fields of view can be tested.

Procedure

To measure axial distance:

1. Scan a test phantom with precisely-spaced vertical pin targets. Adjust all scan controls, as necessary, for the best image of the pin targets to typical depths for the probe being used.
2. Press **Freeze** to stop image acquisition and perform a standard distance measurement between the pins at different points in the image. Record all images for archiving.
3. Scan the vertical pins in zoom or at different depth/scale factors.
4. Press **Freeze** to stop image acquisition; repeat the distance measurements between pins and record the images for archiving.
5. Document the measurements for reference and future comparison.

Contact a Service Engineer if vertical measurements differ by more than 1.50% of the actual distance.

Lateral distance measurements

Description

Lateral measurements are distance measurements obtained perpendicular to the axis of the sound beam. See *Figure 12-14 • Representative Phantom Example* on page 716 for more information.

Benefit

The purpose is the same as vertical measurements. Precisely-spaced horizontal pin targets are scanned and results compared to the known distance in the phantom.

Method

Lateral distance should be measured in the near, mid and far fields as well as in zoom. If necessary, different depths of fields of view can be tested.

Procedure

To measure lateral distance:

1. Scan a test phantom with precisely-spaced horizontal pin targets. Adjust all scan controls, as necessary, for the best image of the pin targets from side to side.
2. Press **Freeze** to stop image acquisition and perform a standard distance measurement between the pins at different points in the image. Record all images for archiving.
3. Scan the horizontal pins in zoom or at different depth/scale factors.
4. Press **Freeze** to stop image acquisition; repeat the distance measurements between pins and record the images for archiving.
5. Document the measurements for reference and future comparison.

Contact a Service Engineer if horizontal measurements differ by more than 3mm or 3% of that distance, whichever is greater.

Axial resolution

Description

Axial resolution is the minimum reflector separation between two closely-spaced objects to produce discrete reflections along the axis of the sound beam. It can also be monitored by

checking the vertical size of known pin targets. See *Figure 12-14 • Representative Phantom Example* on page 716 for more information.

Axial resolution is affected by the transmitting section of the system and the probe.

Benefit

In clinical imaging, poor axial resolution displays small structures lying close together as a single dot. This may lead to improper interpretation of the ultrasound image.

Procedure

To measure Axial resolution:

1. Scan a test phantom with precisely-spaced vertical pin targets.
2. Adjust all scan controls, as necessary, for the best image of the pin targets to typical depths for the probe being used.
3. Press **Freeze** to stop image acquisition.
4. Perform a standard distance measurement of the pin vertical thickness at different points in the image. Record all images for archiving.
5. Scan the vertical pins in zoom or at different depth/scale factors.
6. Press **Freeze** to stop image acquisition; repeat the vertical thickness measurements of the pins and record the images for archiving.
7. Document the measurements for reference and future comparison.

Axial resolution should remain stable over time. Contact a Service Engineer if any changes are observed.

Lateral resolution Description

Lateral resolution is the minimum reflector separation between two closely spaced objects to produce discrete reflections perpendicular to the axis of the sound beam. It can also be monitored by checking the horizontal size of known pin targets. See *Figure 12-14 • Representative Phantom Example* on page 716 for more information.

Lateral resolution is dependent upon the beam width produced by the probe. The narrower the beam, the better the lateral resolution.

The beam width is affected by the frequency, degree of focusing, and distance of the object from the face of the probe.

Benefit

Clinically, poor lateral resolution will display small structures lying close together as a single dot. This may lead to improper interpretation of the ultrasound image.

Procedure

To measure lateral resolution:

1. Scan a test phantom with precisely-spaced horizontal pin targets.
2. Adjust all scan controls, as necessary, for the best image of the pin targets from side to side.

3. Press **Freeze** to stop image acquisition and perform a standard distance measurement of the horizontal thickness of a pin at different points in the image. Record all images for archiving.
4. Scan the horizontal pins in zoom or at different depth/scale factors.
5. Press **Freeze** to stop image acquisition; repeat the horizontal thickness measurements of the pins and record the images for archiving.
6. Document the measurements for reference and future comparison.

Pin width should remain relatively constant over time ("1mm). Dramatic changes in pin width may indicate beamforming problems. Contact a Service Engineer if beam width changes consistently over 2 to 3 periodic tests.

Penetration Description

Penetration is the ability of an imaging system to detect and display weak echoes from small objects at large depths. See *Figure 12-14 • Representative Phantom Example* on page 716 for more information.

Penetration can be affected by the system's:

- Transmitter/receiver
- Degree of probe focusing
- Attenuation of the medium
- Depth and shape of reflecting object
- Electromagnetic interference from local surroundings.

Benefit

Weak reflecting echoes are commonly produced from the internal structure of organs. Definition of this tissue texture is important in the interpretation of the ultrasound findings.

Method

Scan a phantom to see how echoes begin to fade as depth is increased. The maximum depth of penetration is the point at which homogeneous material in the phantom begins to lose brightness.

Procedure

To measure penetration:

1. Set the front panel TGC slide pots to their center (detent) position.
2. Gain and acoustic output can be adjusted, as necessary, since these values are displayed on the monitor.
3. Scan a test phantom along the vertical pin targets to typical depths for the probe being used.
4. Perform a standard distance measurement from the top of the image displayed to the point at which homogeneous material in the phantom begins to lose brightness.
5. Document the depth measurement for reference and future comparison.

Contact a Service Engineer if the depth of penetration shifts more than one centimeter (1cm) when using the same probe and same system settings.

Functional resolution

Description

Functional resolution is an imaging system's ability to detect and display the size, shape, and depth of an anechoic structure, as opposed to a pin target. See *Figure 12-14 • Representative Phantom Example* on page 716 for more information.

The very best possible image is somewhat less important than reproducibility and stability over time. Routine tests at the same settings should produce the same results.

Benefit

The data obtained will give a relative indication of the smallest structure the system is capable of resolving at a given depth.

Procedure

To measure functional resolution:

1. Set the front panel TGC slide pots to their center (detent) position.
2. Gain and acoustic output can be adjusted as necessary, since these values are displayed on the monitor.
3. Scan a test phantom with a vertical row of anechoic cyst targets to typical depths for the probe being used.
4. Evaluate the cysts at various depths for a good (round) shape, well-defined borders and no fill in. Remember, TGC slide pots are centered and should remain fixed. This may NOT provide optimal cystic clearing.
5. Document all results for future reference and comparison.

Contact a Service Engineer if a greatly distorted image is obtained.

Contrast resolution

Description

Contrast resolution is the ability of an imaging system to detect and display the shape and echogenic characteristics of a structure. See *Figure 12-14 • Representative Phantom Example* on page 716 for more information.

Specific values measured are less important than stability over time. Routine tests at the same settings should produce the same results.

Benefit

A correct diagnosis is dependent upon an imaging system's ability to differentiate between a cystic or solid structure versus echo patterns from normal surrounding tissue.

Method

A phantom with echogenic targets of different sizes and depths should be used.

Procedure

To measure contrast resolution:

1. Set the front panel TGC slide pots to their center (detent) position. Set dynamic range to 54 db.
2. Gain and acoustic output can be adjusted, as necessary, since these values are displayed on the monitor.
3. Scan a test phantom with echogenic targets at the depths available.
4. Evaluate the echogenic targets for contrast between each other and between the surrounding phantom material. Remember, TGC slide pots are centered and should remain fixed. This may NOT provide an optimal scan image.
5. Document all results for future reference and comparison.

Contact a Service Engineer if the echogenic characteristics or shapes of the targets appear distorted.

Gray Scale photography

Description

Poor photography will cause loss of low level echoes and the lack of contrast between large amplitude echoes. See *Figure 12-14 • Representative Phantom Example* on page 716 for more information.

Benefit

When photographic controls and film processors are properly adjusted, weak echoes, as well as strong echoes, are accurately recorded on film.

Procedure

1. Adjust the camera according to the manufacturer's instructions until the hard copy and video display are equal.
2. Scan the phantom and it's echogenic contrast targets.
3. Make a hard copy photograph of the display and compare it to the image on the video monitor for contrast and weak echo display.
4. Document all results for future reference and comparison.

Contact a Service Engineer if camera cannot duplicate what is on the image monitor.

NOTE

Optimization of brightness/contrast controls on the display monitor is imperative in order to make sure that the hardcopy and monitor look alike.

The display monitor is adjusted first. The hardcopy camera or printer is adjusted to match the display monitor.

Setting up a Record Keeping System

Preparation

The following is needed:

- Quality Assurance binder.

User Maintenance

- Hard copy or electronic file of images.
- Quality Assurance Checklists.
- Display the following information while testing quality assurance:
 - Acoustic Output
 - Gain
 - Depth
 - Probe
 - Dynamic Range
 - Set up new patient to be the name of the test.
- Annotate the following:
 - Any control where its value is NOT displayed.
 - Significant phantom information.

Record Keeping

Complete the following:

1. Fill out the Ultrasound Quality Assurance Checklist for each probe, as scheduled.
2. Make a hard copy or archive the image.
3. Compare images to baseline images and acceptable values.
4. Evaluate trends over previous test periods.
5. File hard copy or electronic file of images and checklist in Quality Assurance binder.

Ultrasound Quality Assurance Checklist

Table 12-16 • Ultrasound Quality Assurance Checklist (Part 1)

Performed By		Date
System		Serial Number
Probe Type	Probe Model	Serial Number
Phantom Model	Serial Number	Room Temperature
Acoustic Output	Gain	Focal Zone
Gray Map	TGC	Depth
Monitor Setting		
Peripheral Settings		
Other Image Processing Control Settings		

Table 12-17 • (Part 2)

Test	Baseline Value Range	Tested Value	Image Hardcopy/ Archived	Acceptable ? Yes/No	Service Called (Date)	Date Resolved
Vertical Measurement Accuracy						

Test	Baseline Value Range	Tested Value	Image Hardcopy/ Archived	Acceptable ? Yes/No	Service Called (Date)	Date Resolved
Horizontal Measurement Accuracy						
Axial Resolution						
Lateral Resolution						
Penetration						
Functional Resolution						
Contrast Resolution						
Gray Scale Photography						

Image Quality Check

Image Quality Check (IQC)

Image Quality Check is intended to facilitate Image Quality checks during Quality Assurance Evaluations. Quality Assurance tests are used to determine whether a scanner is providing the same level of performance year after year.

By using the same settings year after year, this ensures that the data collection is consistent, independently of who performs the test.

This preset only includes fundamental settings for B-Mode. Processing modes like SRI, Harmonics, etc., are turned off.

To do an Image Quality Check (IQC),

1. Activate IQC via **Utility > Imaging Preset Manager > Category** (select the Category first).
2. Click on the plus sign in front of IQC for Service and select **IQC**.
3. Assign IQC to a Touch Panel key by using the right arrow key.
4. Map the IQC to the location you want it to appear on the Touch Panel.
5. Select **Probe**. Then select **IQC**.

Assistance

Supplies/Accessories

Not all features or products described in this document may be available or cleared for sale in all markets.

Contact the distributor, GE HealthCare affiliate or sales representative for approved peripherals. For HCATs, contact your sales person. For field service replacement part numbers (FRUs), these are service replacement part numbers that may be either new or refurbished, please consult the Basic Service Manual. To order these, contact CARES in the US, or call service in Europe and Asia.

The following supplies/accessories have been verified to be compatible with the system:

Peripherals





Table 12-18 • Peripherals and Accessories


Accessory	Illustration/Photo
Onboard Printer installation kit (UP-D898DC)	
Sony UP-D25M Color Printer	
HP 8210 Printer	
Footswitch, USB	
Powervar144k120v MG UPS UPS	N/A
USB barcode reader Honeywell for LE	N/A
USB barcode reader Honeywell for Nezha	N/A
Barcode Reader USBee1000A	N/A

Accessory	Illustration/Photo
Magstripe Card Reader	

Console

Table 12-19 • Console Accessories

Accessory	Illustration/Photo
English Keyboard Assembly	N/A
Protective Cover	N/A
Swedish Keyboard Assembly	N/A
Norwegian Keyboard Assembly	N/A
Greek Keyboard Assembly	N/A
Russian Keyboard Assembly	N/A
German Keyboard Assembly	N/A
French Keyboard Assembly	N/A
Power Cord - North American	
Power Cord - Argentina	
Power Cord - Europe	
Power Cord - UK-Ireland	
Power Cord - Switzerland	
Power Cord - Denmark	
Power Cord - Italy	
Power Cord - Israel	
Power Cord - Japan	
Power Cord - China	

Accessory	Illustration/Photo
Power Cord - Australia/New Zealand	
Power Cord - India	
Power Cord - South Africa	
Power Cord - Brazil	
Power Cord - Taiwan	N/A
High Cabinet installation kit	N/A
Low Cabinet installation kit	N/A
Side Cabinet Installation kit	N/A
Option 5inch bay installation kit	N/A
Small Probe Holder	N/A
VERTICAL TV PROBE HOLDER	N/A
TVTR Probe Holder Assembly	N/A
Probe Cable Hanger	N/A
Ethernet Protection cable	N/A
Option Tray	N/A

Volume Navigation

Table 12-20 • Volume Navigation

Accessory
VNav Needle Tracking Kit - 18/20g or less
VNav VirtuTRAX Starter Kit
VNav Virtual Tracker
VNav NEEDLE TRACKING
VNav Needle Tracking storage insert
VNav eTRAX 12 14G Starter kit
VNav eTRAX 14 16G Starter kit
VNav Active Tracker kit
VNav MR Active Tracker

VNav Shelf Load Description

Shelf load for the VNav wire shelf on the VNav Stand is 10 lbs.

Hardware Options

Table 12-21 • Hardware options

Accessory
CW Doppler
Realtime 4D
Scan on Battery Option Kit
Power Assistant
LOGIQ Wireless
Vscan Air LOGIQ Integration Kit
S-Video Option
Pencil CW
DVD Drive
Large Storage Drawer

ECG Accessories

Table 12-22 • ECG Accessories

Accessory
ECG
ECG Cables, IEC, AHA Style for Americas
ECG Cables, IEC Style
ECG Cable Kit

Probes

Table 12-23 • Probes and Accessories

Probe	Biopsy Guide	V Nav Capable
BE9CS-D	BE9CS Biopsy Kit 742-401 BE9CS Biopsy Kit 742-339	No
C1-6-D Convex	Verza starter Kit Note: ONLY Verza Starter Kit is supported.	No
C1-6VN-D Convex	Verza starter Kit Note: ONLY Verza Starter Kit is supported.	Yes
C2-7-D Convex	C2-7 Biopsy Kit C2-7 Biopsy Kit Stainless	No
C2-7VN-D Convex	C2-7 Biopsy Kit C2-7 Biopsy Kit Stainless	Yes
C2-9-D Convex	Starter Kit	No

Probe	Biopsy Guide	V Nav Capable
C2-9VN-D Convex	Starter Kit	Yes
C3-10-D Convex	N/A	Yes
IC5-9-D Micro Convex Intracavitary	IC5-9-D Reusable Biopsy Guide IC5-9-D Needle guide	Yes
L2-9-D Linear	Verza starter Kit Note: ONLY Verza Starter Kit is supported.	No
L2-9VN-D Linear	Verza starter Kit Note: ONLY Verza Starter Kit is supported.	Yes
L3-9i-D Linear	N/A	No
L3-12-D Linear	L3-12-D Biopsy Kit L3-12-D Multi Angle Biopsy Set	No
L4-20t-D Linear	Verza biopsy starter kit Note: ONLY Verza Starter Kit is supported.	No
L8-18i-D Linear	N/A	Yes
L6-24-D Linear Array	N/A	No
M5Sc-D XDclear Active Matrix Single Crystal Phased Array Transducer	Multi-angle disposable with a reusable bracket	Yes
ML6-15-D Matrix Array Linear	ML6-15 Biopsy Starter Kit	Yes
RIC5-9-D 4D Convex Volume Intracavitary	Single Angle, Reusable Biopsy Kit for RAB Light	N/A
RAB6-D 4D Volume	Starter Kit	N/A
6S-D	N/A	No
6Tc-RS + RS-DLP Trans- esophageal	N/A	N/A
P2D Pencil Probe	N/A	N/A
P6D Pencil Probe	N/A	N/A
P8D Pencil Probe	N/A	N/A

Table 12-24 • Vscan Air

Probe
Vscan Air CL A1 Kit

Probe
Vscan Air CL A2 Kit
Vscan Air CL C1 Kit
Vscan Air CL I1 Kit
Vscan Air CL G1 Kit
VScan Air On Console Charger

Transesophageal Probe Options

Table 12-25 • Transesophageal Probe Options

Option
Bite Guard
Bite Guard Opr. Clip On
TEE Scanhead. Protective Cover
Bite Guard AD.16"27MM
Bite Hole Inducator
Storage Rack

Options

NOTE

Not all options are available in all countries.

Table 12-26 • Options

Options
4D Complete Package
Advanced Abdominal Tools Package
Advanced Contrast Abdominal Imaging
Advanced Contrast Small Parts Imaging
Advanced privacy and security (vulnerability scan)
Advanced probes
Advanced privacy and security (vulnerability scan)
Advanced probes
Advanced SRI-HD Type 2
Auto Abdominal Color Assistant
Auto Abdominal Package
Auto IMT
Auto Preset Assistant
Auto Renal Measure Assistant
B Steer+
B-Flow

Options
Breast Assistant, Powered by Koios DS
Breast Measure Assistant
Breast Productivity
C1-6-D UP2 Biopsy Kit Software Option
Cardiac Strain (Automatic Function Imaging)
Contrast LI-RADS Productivity Package
Coded Contrast Imaging
Color Quantification
Compare Assistant
CW Doppler
DVR
Elastography
Elastography quantification
Hepatic Assistant
OB Measure Assistant
Omni View
Report Writer
Parametric Imaging
Data Streaming
Scan Assistant
Shear Wave Elastography
SonoNT/SonoIT
STIC (Spatio-Temporal Image Correlation)
Strain Elastography
Stress echo
LOGIQ Apps
Thyroid Assistant, Powered by Koios DS
Thyroid Productivity
Tricefy
Tomographic Ultrasound Imaging (TUI)
Ultrasound Guided Attenuation Parameter (UGAP)
VCI Static (Static Volume Contrast Imaging)
VOCAL_II
Volume Navigation
Probe Check
*The LOGIQ Fortis is designed for compatibility with commercially available Ultrasound contrast agents. Because the availability of these agents is subject to government regulation and approval, product features intended for use with these agents may not be commercially marketed nor made available before the contrast agent is cleared for use. Contrast-related product features are enabled only on systems for delivery to an authorized country or region of use.

Gel

Table 12-27 • Gel — Refer to Accessories Catalog

Accessory	Units
Thermasonic Gel Warmer	Holds three plastic bottles (250ml or 8 oz)
Aquasonic 100 Scan Gel	5 liter jug
	250 ml plastic bottles (12/case)
Scan Ultrasound Gel	8 oz plastic bottles (12/case)
	1 gallon plastic jug
	Four 1-gallon plastic jugs

Disinfectant

Table 12-28 • Disinfectant — Refer to Accessories Catalog

Accessory	Units
Cidex Activated Dialdehyde	16/1 quart bottles
	4/1 gallon bottles
	2/2.5 gallon bottles

Ultrasound Probe and Cord Sheath Sets

Table 12-29 • Probe and Cord Sheath Sets — Refer to Accessories Catalog

Accessory	Units
Sterile Ultrasound Probe Sheath Set	20 Per Set
Sterile Ultrasound Cord Sheath Set	20 Per Set
Sanitary Rectal/Vaginal Probe Cover	20 Per Set
Sterile Combination Probe and Cord Cover Set	12 Per Set
Sterile Ultrasound Probe Sheath Set for Wide (2.5 and 3.5) Aperture Sector Probes	20 Per Set

Chapter 13

Advanced Features

In this section

Using 4D.	737
Contrast Imaging.	776
Strain Elastography.	816
Elastography Analysis	821
Shear Wave Elastography.	824
Ultrasound-Guided Attenuation Parameter (UGAP) Option.	843
Continuous Wave Doppler (CWD)	848
Tissue Velocity Imaging (TVI)	849
Quantitative Analysis (QAnalysis).	851
Stress Echo.	866
Cardiac Automated Functional Imaging (Cardiac Strain)	887
Auto EF.	902
ECG.	907
Volume Navigation.	913
Vscan Air™ CL (Option).	956
Breast Productivity Package.	971
Thyroid Productivity Package.	976
Start Assistant.	982
Scan Assistant.	986
Compare Assistant.	1012
OB Measure Assistant.	1020
Auto Renal Measure Assistant.	1021
Auto Preset Assistant.	1025
Auto Abdominal Color Assistant.	1030

Hepatic Assistant.	1034
EZ Imaging.	1040
Anonymize the patient.	1046
Report Writer.	1048
Configuring the Wireless Network.	1088
Device Mgmt.	1096
Smart Device Apps.	1110
Tricefy Uplink.	1117
Digital Expert.	1120
Data Streaming (Option).	1121
Running Scanner Apps.	1125
Service and Applications Support.	1127
Service Desktop.	1133
Battery Power Mode	1134
Magstripe Card Reader.	1143
Footswitch	1144
DVR.	1145

Using 4D



WARNING

Please take extra care on the volume probe surface temperature before and during scanning to avoid heat injury.

When using volume probe for 4D acquiring, end the 4D acquiring once the current exam is complete, or freeze the 4D image when the machine is not in use.

DO NOT keep volume probe working in 4D acquiring mode for a long time, or the probe surface temperature will exceed the normal value.

4D provides continuous, high volume acquisition of 3D images. 4D adds the dimension of “movement” to a 3D image by providing continuous, real-time displays. With 4D, you can apply rendering techniques to smooth out the appearance of an anatomical structure, for example, a baby’s spine.

You can perform the following types of volume acquisitions within the 4D feature:

Table 13-1 • 4D Package Options

4D Type	Description	Acquisition Mode
4D	Designed for continuous volume acquisition of a 3D image.	B, 4D
Static 3D	Designed for single volume acquisition of a 3D image.	B, 3D

Features supported with 4D

The following features are supported with 4D:

- Most B-Mode controls
- Annotations
- Measurements and Calculations


The following post-processing controls are available with 4D:

- CINE
- Zoom

4D Principles of Operation

The acquisition of volume starts with a 2D image using special probes designed for performing 3D sweeps and 4D scans. The volume box defines the region of interest to be used for the volume sweep.

Volume sweep refers to the range of the sweep of the 2D image to be transformed into a rendered, 3D or 4D image. Static 3D acquisition involves a single volume sweep. 4D involves multiple, continuous volume sweeps.



- 1 Central 2D Scan
- 2 Start 2D Scan
- 3 Range of VOI Sweep

Figure 13-1 • Volume Sweep

When initiating a volume sweep, the angle of the volume can be adjusted.


Interactive 3D Rendering

Interactive 3D rendering allows visualization of certain structures, helping to view and analyze different sections of the volume.

Region of Interest (ROI)/Render Box

The Region of Interest (ROI) - also referred to as the Render Box in rendering - contains the section of the volume you want to render. Therefore, objects that are not inside of the box are not included in the render process and are cut out (this is important in surface mode to allow a free line of sight). This may or might not be the entire Volume of Interest (VOI).

The view direction of the ROI can be adjusted.



1	Up/Down
2	Left/Right
3	Front/Back
4	Down/Up
5	Right/Left
6	Back/Front


Figure 13-2 • Render View Directions

Render View

In Render view, only the rendered image displays - no reference images.

Image Orientation

Orientation of Image in Sectional view.



1	Longitudinal
2	Transverse
3	Coronal

Figure 13-3 • Quad Render Visualization Mode

Principle of Sectional Planes

Sectional planes represent three different planes of the same 3D volume. There are three separate planes, A (Longitudinal), B, (Transverse) and C (Coronal).




Figure 13-4 • Illustration of Sectional Planes


The presentation of three orthogonal sectional planes is different from the conventional patient orientation in 2D sonography.

NOTE

Whenever the usual, longitudinal section of the patient to display in field A is selected, the conventional orientation for longitudinal and transverse sections is valid.

Reference Images

Reference images are the individual image displays within the corresponding sectional plane. Reference image A represents the longitudinal view; reference image B the transverse view, and reference image C represents the coronal view.



- 1 Reference Image A (Longitudinal)
- 2 Reference Image B (Transverse)
- 3 Reference Image C (Coronal)

Figure 13-5 • Monitor Display of Reference Images in Sectional View

Orientation Help

When you view a 4D image on the display, it's sometimes difficult to recognize the orientation. To help, the system displays a three-dimensional drawing to illustrate the orientation. This drawing displays ONLY in sectional view.




Figure 13-6 • Reference Image A

For Reference image A, the transducer plane migrates from the FRONT to the REAR through the volume body.




Figure 13-7 • Reference Image B

For Reference image B, the transducer plane migrates from the LEFT to the RIGHT through the volume body.





Figure 13-8 • Reference Image C


For Reference image C, the transducer plane migrates from the TOP to the BOTTOM through the volume body.

Examples of Probe Orientation with Reference Planes




- 1 Image Plane A
- 2 Image Plane B
- 3 Image Plane C

Figure 13-9 • Abdominal Probe Orientation



- 1 Image Plane A
- 2 Image Plane B
- 3 Image Plane C

Figure 13-10 • Small Part Probe Orientation



- 1 Image Plane A
- 2 Image Plane B
- 3 Image Plane C
- 4 Groove

Figure 13-11 • Endocavity Probe Orientation

4D Operational Controls

Control Panel Overview

When entering 3D/4D mode, the behavior of some of the Control Panel buttons changes. For example, in 3D/4D mode, the PW-Mode, CF-Mode and M-Mode buttons are used (along with Depth) to manipulate the Volume of Interest (VOI).



1. User-configurable controls.
2. M key, M-Mode key used to rotate about the X axis.
3. PW key, PW-Mode key used to rotate about the Y axis.
4. CF Key, CF-Mode key used to rotate about the Z axis.
5. Depth key, used as a transverse translation through the image.
6. Width key, used to adjust the size and position of the VOI.
7. Trackball, used to move the VOI. Also, the 4 keys surrounding the Trackball map to additional functionality, as shown on the monitor display.
8. L (left) key, used to begin a 4D acquisition.
9. Freeze or R (right) key, used to freeze a 4D image.

Figure 13-12 • Control Panel Buttons

4D Monitor Display

Imaging parameters are displayed in the upper right-hand portion of the display. The 4D specific parameters are Quality (Q), Volume Angle (A) and Volume Rate (VR). The Status Bar contains instructions on the tasks you can perform at each stage of the 4D imaging process. Remember to take a look at the Status Bar as needed.

4D Touch Panel Overview

Most 4D Touch Panel screens contain some similar controls. Refer to the table below for descriptions of these controls. Controls that are unique to or that contain slightly different functionality are described in the respective sections.

Table 13-2 • Common 4D Touch Panel Controls

Preset Parameter	Description
Tile	The display can be divided into 1, 2, or 4 windows for Render view (Render = On) and 1 or 4 windows for Sectional view (Render = Off).

Preset Parameter	Description
Reset Curve	Resets the three-point curve to a straight line.
Direction	Adjusts the view direction of the ROI.
Visualization	Sectional, Render, VCI, or Tomographic Ultrasound Imaging (TUI). Render view displays one rendered image, or reference image(s) and rendered image.
Focus Position	Adjusts the focal position.
Volume Angle	Sets the range of the volume sweep.
Quality	Balances speed with line density. Max combines the highest density with the slowest speed; Low combines the lowest density with the highest speed.

4D Presets

Real-Time 4D/Static 3D Presets

1. When entering 3D/4D mode, select the **Preset** tab.
2. Select one of the preset settings for data acquisition and display. Presets are defined in the preset file and differ by application.

Table 13-3 • Common 4D Touch Panel Controls

Preset Parameter	Description
Save	Selections: Overwrite, Create New, Cancel. Overwrite. Overwrite the application preset file with the changes you just made. Create New. Create a new user application preset file based upon the current exam category and application. Cancel. Cancel without saving preset parameters.
Pre-defined Preset	Reloads the presets for the selected application.
User1, User2, User3, User4	Used to define new user presets for a given application.

Static 3D Presets

1. When entering 3D/4D mode, select **Static 3D**, then the **Preset** tab.
2. Select one of the preset settings for data acquisition and display. Presets are defined in the preset file and differ by application.


Performing a 4D Scan

Visualizations

4D provides two types of views for displaying and working with images: Sectional, Render, and Tomographic Ultrasound Imaging (TUI).

Sectional View

Sectional view contains one display for each sectional plane.



1 Sectional Image A

2 Sectional Image B

3 Sectional Image C

Figure 13-13 • Monitor Display in Sectional View

Render View

The ultrasound system continuously displays the 4D rendered image.

NOTE

When the tile selection is single, only the rendered 4D image appears. When the tile selection is quad, the sectional images are located in 3 quadrants with the rendered 4D image in the fourth.




Figure 13-14 • Quad Tile Render View

Orientation Help

When viewing a 4D image on the display monitor, it's sometimes difficult to recognize the orientation. To help, the system displays a three-dimensional drawing to illustrate the orientation. This drawing displays **ONLY** in sectional view.




Figure 13-15 • Orientation Help Graphic

Acquiring and Rendering a 4D VOI Starting with a 2D Image

To create a 4D image, start with an optimized 2D image. The 2D image serves as the mid-line for the resulting 4D image.

1. Connect the appropriate 4D-compatible probe, leaving the probes in their respective holders.

NOTE

If the appropriate 4D probe is not connected, the original 3D Touch Panel appears.

2. Obtain a 2D image. Optimize the image as usual.

Entering 3D/4D Mode

In 3D/4D mode, the type of scan to perform is chosen: 4D or Static 3D.

1. Press **3D/4D** to enter 3D/4D mode. The first time 3D/4D is pressed, the system is in B Pre mode.

NOTE

The location of the number of focal zones might change when entering 3D/4D mode, since the number of zones is pre-determined by the default ROI.

The default acquisition mode varies by application. In the OB application, the default acquisition mode is Real-Time 4D. For all other applications, Static 3D is the default acquisition mode. When entering pre-mode, an ROI graphic may appear on the monitor display that defines the initial ROI (Region of Interest) of the volume.




Figure 13-16 • ROI Graphic

2. Press the **Preset** tab. Select one of the preset settings for data acquisition and display. Presets are defined in the preset file and differ by application.

Quick Acquisition Steps

1. Connect the appropriate 4D-compatible probe, leaving the probes in their respective holders. Follow the guidelines in Chapter 3 for connecting probes.
2. Select a 4D probe from the probe indicator.
3. Obtain a 2D image. Optimize the image as usual.
4. Press **3D/4D**. An ROI graphic appears. 4D is selected.
5. Define the Volume of Interest (VOI) to be scanned. Use the **Trackball** to move the VOI and the **Width** button to re-size and re-position the VOI. Only the area defined within the VOI is rendered.
6. Adjust the volume angle and quality. This defines the range of the volume sweep. A small sweep angle results in a lower number of slices with a high volume rate.
7. To begin 4D acquisition, press the **L** key.

You DO NOT have to hold the probe steady during data acquisition.

During data acquisition, you can manipulate the VOI to see different views of the image. To rotate the VOI left or right, use the **PW** control. To rotate the VOI forward or backward, use the **CF** control. To rotate the VOI in a circular motion, use the **M** control.

To return to 3D/4D pre-mode, press **L**.

NOTE

If the volume size is too large, the message "Volume Size Too Big - Quality Degraded" displays in the status bar. The system changes the quality automatically to below the upper limit and displays the proper Quality value in the information window.

8. Set Render to On.
9. To complete the acquisition, press **Freeze** or **R**.
10. Store the image.

4D

4D provides continuous, high volume acquisition of 3D images. Rendering techniques can be applied to smooth out the appearance of an anatomical structure, for example, a baby's face.

4D imaging contains three main viewing modes: Sectional, Render, and Tomographic Ultrasound Imaging (TUI). Sectional mode displays three separate representations of the image: Longitudinal (original 2D image), Transverse (elevational), and Coronal (horizontal). Render mode displays one rendered 4D image.

Acquiring a 4D Volume of Interest (VOI)

Once an optimized 2D image is acquired, a 4D scan can be performed to acquire the 4D image.

During 4D image acquisition:

- Frame Averaging is disabled.
- The transmit frequency can't be changed.

To acquire a 4D VOI:

1. Select **4D**.
2. Make sure the VOI is defined appropriately. If necessary, adjust the volume angle. This defines the range of the volume sweep. A small sweep angle results in a lower number of slices with a higher volume rate.

See Manipulating the Volume of Interest (VOI) for more information.

3. To begin 4D acquisition, press the **L** key. The system will perform continuous sweeps across the VOI. The probe does not have to be held steady during a 4D scan.

To return to 4D pre-mode, press **L**.

4. Set the Render to On.

Sectional VOI Acquisition

Sectional view provides three separate views of the same image: Longitudinal (original image), Transverse (elevational), and Coronal (horizontal).

1. In the 4D tab, Render defaults to On (Render mode). Change Render to Off for Sectional view.
2. To select a reference image, use the Ref Image control on the Touch Panel. The reference image selected contains the focus for control panel keys, allowing you to manipulate or optimize that image.

Table 13-4 • 4D Data Acquisition Parameters

4D Parameter	Description
Restore View	Resets all parameters back to the original values or chosen presets.
Tile	Selections: Single, Quad. You can divide the display into 1 or 4 windows.
Visualization	Sectional, Render, or Tomographic Ultrasound Imaging (TUI). Render view displays one rendered image, or reference image(s) and rendered image.
Ref Image	Use to select the reference image that has focus for use with the control panel keys and Trackball.
Orientation Help	Displays a three-dimensional drawing to illustrate the orientation. Only displays in sectional view.
Volume Angle	Sets the range of the volume sweep.
B Quality	Selections: Max, Hi2, Hi1, Mid2, Mid1, Low. Used to balance speed with line density. Max combines the highest density with the slowest speed. Low combines the lowest density with the highest speed. BQ displays on the display.

Render VOI Acquisition

Rendering allows distinguishing of subtle anatomical detail. All areas of a VOI can be rendered, or just certain regions of the VOI. The region defined for rendering is referred to as the Render Box.

1. Define the area to be rendered. For example, with an image of an entire fetus, only the fetal face may need to be rendered. Therefore, define the fetal face as the VOI.
2. Set Render to On.

Table 13-5 • 4D Data Acquisition Parameters - Render Mode

4D Parameter	Description
Restore View	Select to reset all parameters back to the original values or chosen presets.
Tile	Selections: Single, Dual, Quad. You can divide the display into 1, 2, or 4 windows.
Visualization	Sectional, Render, or Tomographic Ultrasound Imaging (TUI). Render view displays one rendered image, or reference image(s) and rendered image.
3D Orient	When selected, changes the orientation of the image on the monitor display. Selections include: 0 degrees, 90 degrees, 180 degrees, and 270 degrees.

4D Parameter	Description
Ref Image	Use to select the reference image that has focus for use with the control panel keys and Trackball. This control is enabled only if Tile is set to Quad.
Volume Angle	Sets the range of the volume sweep.
Quality	Selections: Max, Hi2, Hi1, Mid2, Mid1, Low. Used to balance speed with line density. Max combines the highest density with the slowest speed. Low combines the lowest density with the highest speed.
Activate Curve	Define a three-point curved surface for the render window using the Trackball.
Reset Curve	Reset the three-point curve to a straight line.
Mix	Selections: 0-100% in increments of 2. Allows you to mix a Rend Mode 1 mode with a Rend Mode 2 mode. Always select two modes.
Lower Threshold	Selections: 0-255. Sets a lower threshold below which weaker echoes are removed.

3. Select the **Render Setting** tab.

The Render Setting tab allows selection and combining of gray-scale and color rendering modes.



HINTS

When using Surface modes, adjust the Lower Threshold to recognize border structures more clearly.

Table 13-6 • 4D (Data Acquisition) Render Parameters

4D Parameter	Description
Direction	The ROI determines the region that is rendered during 4D acquisition. The direction in which this ROI is viewed can be changed. Selections: Up/Down, Down Up, Left/Right, Right/Left, Front/Back, Back/Front.
Gray Map	Displays the gray map selections on the display monitor. Select maps using the Trackball.
Colorize	Displays the tint map selections on the display monitor. Select maps using the Trackball.
Render Mode	Select Gray or Inversion. Selecting Inversion inverts the gray values of the rendered image (e.g., image information that was black becomes white, and vice versa).

4D Parameter	Description
Render 1	Allows combination of render mode values from render mode 1. Select the render map combination from the upper-left portion of the monitor display. Select map combinations using the Trackball. Render Mode 1 Selections: Surface Smooth, Surface Texture, Transp Max, Transp X-ray, Transp Min, HDlive Texture. Surface Smooth - Surface displays in a smoothed texture mode, which means that the gray values of the surface are identical with the gray values of the original 2D scan. Surface Texture - Surface displays in texture mode, which means that the gray values of the surface are identical with the gray values of the original 2D scan. Transp Max. - Displays the maximum intensity of gray values in the ROI. This is helpful for viewing bony structures. Transp X-Ray - Displays the mean value of all gray values in the ROI. Trans Min. - Displays the minimum number of gray values in the ROI. This is helpful for viewing vessels and hollow structures. HDlive Texture - Uses an illumination source that can be positioned by the user around the rendered 3D object on a spherical coordinate. By highlighting structures from the side, the three-dimensional impression can be improved considerably.
Render 2	Allows combination of render mode values from render mode 2. Render Mode 2 Selections: Surface Smooth, Light, Gradient Light, Transp Max, Transp X-ray, Transp. Min., HDlive Smooth. Surface Smooth - Surface displays in a smoothed texture mode, which means that the gray values of the surface are identical with the gray values of the original 2D scan. Light - Surface displays in light mode. Structures in the near field are brighter; structures in the far field are darker. Gradient Light - Surface displays as if it is illuminated from a spot light source. This is helpful if the displayed surface is surrounded by hypoechoic structures (for example, liquids). Transp Max. - Displays the maximum intensity of gray values in the ROI. This is helpful for viewing bony structures. Transp X-ray - Displays the mean value of all gray values in the ROI. Transp Min. - Displays the minimum number of gray values in the ROI. This is helpful for viewing vessels and hollow structures. HDlive Smooth - Smoothed HDlive texture mode. When Render 1 value is set to HDlive Texture, the only available mode for Render 2 is HDlive Smooth. Note: HDlive rendering is not available if VCI is active.
Edit Light	Activates the light editor for the virtual light source in HDlive render mode. User can position the light source by moving the Trackball, A light icon on the main display indicates the direction of the light. Note: The Edit Light parameter is only available for the HDlive rendering modes.
Transparency	Selections: 20 to 250. Sets the transparency of the image. The higher the number, the more transparent the gray scale information.

Manipulating the Volume of Interest (VOI)

The 3D/4D ROI is a tangible anatomical object that can be viewed and manipulated easily using the **Trackball** and other control panel keys.

If the monitor display is in Sectional view, select the desired reference image before manipulating the image.

NOTE

The manipulation examples are with A set as the reference image.

Rotating the 4D VOI Left/Right or Forward/Backward

The VOI can be rotated around the X, Y, and Z axes. To rotate the VOI around the Z axis, turn the **CF** control left/right.




Figure 13-17 • Rotate about Z Axis with CF Control

To rotate the VOI around the Y axis, turn the **PW** left/right.




Figure 13-18 • Rotate about Y Axis with PW Control

Rotating the 4D Image in a Circular Motion

To rotate the VOI around the X axis, turn the **M** control left/right.




Figure 13-19 • Rotate about X Axis with M Control

NOTE

To speed up the rotation, press the PW, CF, or M controls before you turn them. To slow down the rotation, press PW or CF again.

Moving through the VOI

To move through the image to view a particular slice, press Depth.

This allows a displacement of the center of rotation along the intersection lines of the sectional planes A, B, and C. The displacement of the center of rotation leads to the display of parallel sectional images. See *Reference Images* on page 740.

Zooming the Image

Rotate **Zoom** to zoom on the image.

Moving the VOI Position

To move the position of the VOI, move the **Trackball** left, right, up and down, as needed.

Resizing the VOI

To re-size the VOI, use the **Width** control panel button. See the Width section in this chapter.

Stopping 4D Image Acquisition

To stop acquiring a 4D image, press **Freeze** or R in Render view, or just **Freeze** in Sectional view.

4D VOI Post-Processing

Pressing Freeze or R displays one of the following Touch Panel screens, dependent upon whether in Render view or Sectional view.

Volume CINE

The system constantly stores CINE images to play back and review those images. CINE is useful for focusing on images during the specific part of the heart cycle or to view short segments of a scan session.

To activate CINE in 4D:

1. Press **Freeze**.
2. Select the **VolCine** tab.

Table 13-7 • 4D Cine Parameters

Preset Parameter	Description
Loop Mode	Selections include: One Way, BiDirectional (two-way). One Way - plays one loop sequence forward. BiDirectional - plays the sequence forward and backward.
First	Displays the first volume in the CINE loop.
Last	Displays the last volume in the CINE loop.
Run/Stop	Starts and stops the CINE loop.
Loop Speed	Adjusts the CINE loop speed.
Volume by Volume	Used to select an individual volume in the CINE loop.

3. If 4D CINE mode was entered from Render visualization mode, press L to return to Pre-Mode.
If 4D CINE mode was entered from Sectional visualization, press L to return to Pre-Mode.
4. To re-start real-time 4D acquisition, press **Freeze**.

Static 3D

A single sweep, single volume static 3D image can be created.

Performing a Static 3D Scan

1. Connect the appropriate 4D-compatible probe, leaving the probes in their respective holders. Follow the guidelines in Chapter 3 for connecting probes.
2. Select a 4D probe from the probe indicator.
3. Obtain a 2D image. Optimize the image as usual.
4. Press **3D/4D**.
5. Press **Static 3D**. Set Visualization to Render.
6. Set the Volume of Interest (VOI) to be rendered. Use the **Trackball** to move the VOI and **Width** to re-size the ROI.
7. Adjust the volume angle. This defines the range of the volume sweep. A small sweep angle results in a lower number of slices with a high volume rate.
8. Set the probe down on the patient, making sure the probe is held steady. Press the **L** key to start acquisition.

NOTE

During 3D acquisition, no control panel keys are not available, except for 'R.'

NOTE

When the 3D acquisition begins, the Touch Panel appears blank for a brief moment.

Advanced Features

9. Hold the probe steady until the system stops automatically. The acquisition has stopped when the Touch Panel changes to display the Render Setting, 3D Rotational Cine, and Scalpel tabs.

To stop the acquisition manually, press the **R** key.

10. Save the image.

11. To further manipulate the 3D image, press **Static 3D**.

The Touch Panel that displays depends on the visualization mode selected prior to freeze, Sectional, Render, TUI or VCI.

Static 3D Sectional View

Table 13-8 • 3D After Acquisition Parameters - Sectional View

Preset Parameter	Description
Orientation Help	Displays a 3-dimensional drawing to illustrate the orientation. Only displays in sectional view.

Static 3D Render View

Table 13-9 • 3D After Acquisition Parameters - Render View - Page 1

Preset Parameter	Description
Edit/Accept ROI	Selections include Edit, Accept. Edit - Select to adjust the size of the Region of Interest (ROI). Accept - accepts the active 3D image.
3D Orient	When selected, changes the orientation of the image on the monitor display. Selections include: 0 degrees, 90 degrees, 180 degrees, and 270 degrees.

VCI

VCI (Volume Contrast Imaging) allows sweeping smaller slices of data with a higher volume rate. The resulting image shows an average, integrated gray value of the tissue contained within the ROI. VCI improves the contrast resolution and signal/noise ratio. It also reduces image speckle. This may facilitate finding diffuse lesions in organs.

Touch Panel

The data is represented as in Static 3D - Sectional Planes. However, the three planes are VCI renderings (tissue information of a thick slice) computed from the 3D dataset.

Table 13-10 • VCI View

Preset Parameter	Description
Slice Thickness	Select the slice thickness

Static 3D Color

To view Static 3D Color:

1. Acquire the anatomy to view in B-Mode.
2. Activate Color Flow or PDI.
3. Activate 3D/4D. The Pre-3D displays. Select the desired Visualization.

NOTE

4D cannot be selected from Pre-Mode with Color active.

4. Press **L** to render the image.

The color is rendered. You can adjust the Render Mode (default is Glass Body):

- Gray. Gray, no color.




Figure 13-20 • 3D Color - Gray

- Inversion. Inverts the grayscale.
- Color. Displays Color-Flow.

Select the rendering method ([Rend Color1] (Surface only) or [Rend Color2]) and the mixing ratio.




Figure 13-21 • 3D Color - Color

- Glass Body. Displays both Color-Flow and B-Mode.

Select the rendering method ([Rend Gray] or [Rend Color]) and the mixing ratio.




Figure 13-22 • 3D Color - Glass Body

To hide the color on the sectional planes, select **Color Off** from the Touch Panel.

NOTE

If you scan in Render mode and select Tile as Quad, you can hide the color with the Color Off Touch Panel key.

Volume Review

Volume Review enables post-processing the VOI and scrolling through acquisition planes to review all of the frames within the volume set.

In Volume Review, start in Run/Stop so that every frame in the volume is displayed.

NOTE

If the screen is adjusted while in Volume Review, these changes ARE NOT reflected on the Static 3D volume.

Scalpel

Scalpel allows editing/cutting sections of a 3D image. Scalpel is available only on a rendered image.

1. Select **Scalpel**.

Table 13-11 • Scalpel Parameters

Preset Parameter	Description
Cut Mode	Selections: Inside Contour, Outside Contour, Inside Box, Outside Box, Eraser Big, Eraser Small. Inside Contour, Outside Contour - Allows tracing the portion of the image to cut. Trace Outside removes all portions of the image that fall outside the traced region. Trace Inside removes all portions of the image that fall inside the traced region. Inside Box, Outside Box - Displays a box used to define the portion of the image to cut. Outside Box removes all portions of the image that fall outside the box. Inside Box removes all portions of the image that fall inside the box. Eraser Big, Eraser Small - Provides a big and small eraser tool used to define the portion of the image to cut by hand. Available only if Depth is Full.
Cut Depth	Selections: Full, Define. Full -The entire depth of the selected region will be cut. Define - Allows you to define the depth to cut using the Depth control panel knob.
Undo Last	Undoes the last cut only.
Redo	Select to redo scalpel.
Undo All	Undoes all cuts since you entered Scalpel mode.
Done	Applies to User Defined Cut Depth when complete.

2. Select the cut mode.
3. Use the **Trackball** and Set key to define the portion of the image to cut. Press Set to start, move the **Trackball** to define the region, then press Set again to cut the image. The portion is removed.

To undo the last cut, select **Undo Last**.

To undo all cuts in the current session, select **Undo All**.

NOTE

With the cut image displayed, if you attempt to switch to the Static 3D tab to edit the ROI, the following warning message appears: Scalpel changes will be lost. Do you want to continue? [Yes/No].

3D Rotation CINE

3D Rotation CINE allows you to view the 3D image from various angles.

To activate rotation CINE in 3D:

1. Press **Freeze**.
2. Select the 3DRot Cine tab.

Table 13-12 • 3D Rotation Cine Parameters

Preset Parameter	Description
Rotational Angle	Sets the rotational angle of the 3D image over which the CINE loop is played. Typical values are 30, 45, 60, 90, 180 and 360 degrees.
Step Angle	Sets the step angle between individual frames in the CINE loop.

Preset Parameter	Description
Rotation Axis	Sets the axis about which the CINE loop is calculated. Selections X and Y.
Loop Mode	Selections include: One Way, BiDirectional (two-way). One Way - plays one loop sequence forward. BiDirectional - plays the sequence forward and backward.
First	Displays the first volume in the CINE.
Last	Displays the last volume in the CINE.
Run/Stop	Starts and stops the CINE sequence.
Start Angle	Used to select the starting angle in the CINE loop range. The default Start Image is calculated from the rotational angle as: $-1 \times \text{Rotational angle} / 2$. If you adjust the Start Image, the Rotational Angle is re-set to be the value of the adjusted Start Image.
End Angle	Used to select the ending angle in the CINE loop range. The default End Image is calculated from the rotational angle as: $\text{Rotational angle} / 2$. If you adjust the End Image, the Rotational Angle is re-set to be the value of the adjusted End Image.
Image by Image	Used to select an individual image in the CINE loop.

VOCAL

Use VOCAL (Virtual Organ Computer-aided Analysis) to visualize and calculate the volume of anatomical structures, such as a tumor lesion, cysts, and the prostate. VOCAL is available after a Static 3D or Real-Time 4D acquisition.

1. Select **Vocal**. Specify the volume calculation method (**Manual**, **Contour Detect**, **SemiAuto Detect**, or **Sphere**). Select the reference image you want to use to perform the trace by selecting **Ref Image** A, B, or C. Press **Start**.
2. Trace the anatomy using the **Trackball**. Press **Set** to start and end the trace. You must go across the dotted line for the trace to take effect (it turns yellow). The trace is performed on each image slice, separated by the rotational step angle. Rotate the **Rot. Ref** dial until you have completed the total of the required rotations (for example, if you've selected 30 degrees, you need to complete six traces if you've selected Manual). After you've completed the trace target, the **Calc Volume** button is active for you to press. The calculated VOCAL image appears in the lower, right-hand corner of the display. You are now in the edit state.

NOTE

Trace not used for Sphere. For Sphere, set the Poles.

3. Edit as necessary. You can apply a shell, adjust its thickness, navigate through the reference angles, or restart the VOCAL.

VOCAL Touch Panel states are described in the following tables.

Table 13-13 • VOCAL Parameters on Setup Touch Panel

Parameter	Description
Manual	When you select the Manual method, you need to perform a manual trace on each of the rotation angles.
Contour Detect	When you select the Contour Detect method, you need to perform a manual trace on each of the rotation angles.
SemiAuto Detect	When you select the Semi Automatic Detect method, you need to perform a trace on only two rotation angles. The system applies an algorithm to define the traces.
Structure	Structure is only available with the SemiAuto Detect method. Select Hypo, Cystic, or Hyper/Iso.
Sphere	Calculates the volume based on the pole settings.
Pole 1	Adjust the upper contour point (green arrow) of the structure.
Pole 2	Adjust the lower contour point (green arrow) of the structure.
Rotational Step Angle	Specify the angular spacing between contour traces. Typical values are 6, 9, 15, and 30 degrees. The number of planes varies by this formula: 180 degrees / selected rotational step angle.
Ref Image	Use this to select the image you want to use to perform the trace.
Start	Press Start when you're ready to perform the trace.
Rot.Ref #/# Back/Next	Select Next/Back to move to the next image for contour definition in the rotation step.

Table 13-14 • VOCAL Calculate Volume Touch Panel

Parameter	Description
Calc Volume	Press Calc Volume to initiate the VOCAL image calculation.
Clear	Press Clear to remove the trace from the image.
Restart Vocal	Press Restart Vocal to return to the initial VOCAL state.

The Shell Modes allows you to construct a shell or contour “around” the structure of interest, which enables you to distinguish between the contour of the targeted structure and the contours the inside and outside of the structure.

Table 13-15 • VOCAL Parameters on Edit Touch Panel

Parameter	Description
Shell Off	Select Shell Off if you do not want a shell around the VOCAL image.
Inside	Select Inside if you want a shell inside the volume.
Outside	Select Outside if you want a shell outside the volume.
Symmetric	Select Symmetric if you want half of the shell thickness inside and half outside the volume's perimeter.
Shell Thickness	Adjust to vary the thickness of the shell.

Static 3D Render Setting

Table 13-16 • 3D After Acquisition Parameters - Render View - Page 1

Preset Parameter	Description
Edit/Accept ROI	Selections include Edit, Accept. Edit - Select to adjust the size of the Region of Interest (ROI). Accept - accepts the active 3D image.
3D Orient	When selected, changes the orientation of the image on the monitor display. Selections include: 0 degrees, 90 degrees, 180 degrees, and 270 degrees.

Static 3D Volume Review

Volume Review enables post-processing the VOI and scrolling through acquisition planes to review all of the frames within the volume set.

In Volume Review, start in Run/Stop so that every frame in the volume is displayed.

NOTE

If the screen is adjusted while in Volume Review, these changes ARE NOT reflected on the Static 3D volume.

Storing 4D Images

4D images are stored exactly as 2D images are stored. Because 4D images contain more data, they require more space. Pay close attention to the size of the volumes.

- Still image: Store as Raw Data
- CINE: Store as Raw Data when “Enhanced DICOM” is selected for printing 4D.

Tomographic Ultrasound Imaging (TUI)

Tomographic Ultrasound Imaging (TUI) is a visualization mode which presents data as parallel slices (planes) through the dataset. This method of visualization is consistent with CT and MRI. The distance between the different planes can be adjusted.

1. Select TUI as the Visualization mode.
2. Press **L** to start acquisition.
3. If in 4D, press **R** to end the acquisition. This step is not required in Static 3D. The reference image + the number of specified slices appears. The reference image always displays and indicates which slices you are currently viewing as solid lines.




Figure 13-23 • TUI 3x3 Example

- a. The TUI reference image that shows the slice position. This image is orthogonal to the reference image.
- b. The number and green asterisk shows the position of each slice. A green asterisk indicates the center image (A, B or C-plane).
- c. Slice distance displays when the slices are in certain intervals.
- d. A solid line indicates the slice appears on the monitor.

A dotted line indicates the slice did not appear on the monitor.

NOTE

TUI with Color is only available with Static 3D, not with 4D.

4. Adjust the number of slices and slice distance.

You can adjust the number of slices by using the **Slices** rotary. You can adjust the distance between the slices using the **Slice Distance** rotary. Max value is 40mm.

- Move forward/backward through the slices via **Prev./Next Slice**.
- Change the center image via **Ref. Image** if needed (Reference image A, B or C).
- Select **Display Format** from 1X1, 1X2, 2X2 and 3X3.
- The following features are supported in TUI: Zoom, Rotation (X/Y/Z), Trackball (Move the position), Translation and Gain.
- To hide the color, select **Color Off** from the Touch Panel.

You can adjust each slice position with **Adjust Slices**.

1. Press **Adjust Slices** on the Touch Panel.
2. The pointer displays. Select a slice by using the **Trackball**.




Figure 13-24 • Adjust Slice screen (example)

- a. Selected slice displays with yellow border
 - b. Distance from the center image
- Adjust the slice distance from the reference image by using **Slice Position**. Slice Position only affects the selected slice. If you place the pointer on the reference image and rotate **Slice Position**, the position of the reference image and all slices are moved.
 - Adjust the number of right and left slices off the center image by using **Left Slices** or **Right Slices**.
3. After the adjustment is finished, press **Set**. The slice is marked with an “X”. To print this slice, press a print key.

SonoRenderlive

SonoRenderlive helps to find the render start position to easily separate solid tissue in front of the render object.

The SonoRenderlive algorithm “looks” for the transition from solid to liquid tissue and positions the “Render Start” into the liquid area visualized by the green render start line. The render start line is not a straight line but a “free” trace for optimal adaptation to the render object.




Figure 13-25 • Screen Display Off



Figure 13-26 • Screen Display On

Using SonoRenderlive

1. Start the Render Visualization Mode.
2. Press the **SonoRenderlive** Rotary control.
3. To adjust the distance between the render start position and the render object, rotate the **Sensitivity** control below the Touch Panel. A high value indicates a smaller distance.

NOTE

If SonoRenderlive is not used, the Render Start line can also be modified manually. Press the trackball button Curve to activate Curved Render Start and move the trackball to modify the line.

SonoAVC™

SonoAVC helps detect low echogenic objects (eg. cysts) in an organ (eg. kidney) and analyzes their shape and volume. From the calculated volume of the object, an average diameter is calculated. All calculated objects are listed according to size. The calculation results are displayed in the top left corner, with the object numbers color-coded to denote the object on the image. If the mouse cursor hovers over a specific item on the list, the respective object in the image is highlighted (and vice versa). The color of the object is bound to its position on the list.

NOTE

SonoAVC is not applicable for China market.



Figure 13-27 • SonoAVC Display Screen

- d(V): Diameter, calculated as if the object were a perfect sphere
- dx: Length of x-axis of the best fitting ellipsoid
- dy: Length of y-axis of the best fitting ellipsoid
- dz: Length of z-axis of the best fitting ellipsoid
- mean d: Average value of the x-axis, y-axis and z-axis
- V: Volume of the object



Figure 13-28 • SonoAVC Menu

Table 13-17 • Menu

Parameter	Description
Measure Method	- Manual: Each object has to be selected and measured manually - SemiAuto: Each object is selected manually with the system cursor but traced / measured automatically - Auto: Objects are detected automatically (Default)
ROI Shape	- Rounded: The box has an elliptic shape with rounded corners (Default). - Angular: The box has the shape of a rectangle.
Ref Image	Select the reference image to which all image dependent functions like parallel shifts, rotations, etc. are applied.
Start SonoAVC	With Abd application: - Left Kidney - Right Kidney With Gyn application: - Left Ovary - Right Ovary

SonoAVC Main Menu




Figure 13-29 • SonoAVC Main Menu




Figure 13-30 • SonoAVC Trackball Menu

Table 13-18 • Main Menu

Parameter	Description
New Analysis	Starts a new analysis.
Add Manual	Only available when auto or semi-auto is selected. Objects not automatically detected can be added manually, or wrongly detected objects can be removed.
Save	Add the result to database.
Worksheet	Open the worksheet.
Cut/Merge	- Undo All: Undo all edits - Redo: Redo the last edit - Undo: Undo the last edit

Parameter	Description
Display	• None: No visible segmentation • Contour: Segmentation are shown with a contour-line • Filled: Segmentation are filled with color
Ref Image	Select the reference image to which all image dependent functions like parallel shifts, rotations, etc. are applied.
Add/Rem.	Add detected objects manually.
Cut	Cut detected objects.
Merge	Combine detected objects.

Using SonoAVC Manual

1. Acquire 3D volume or reload previously acquired volume.
2. Select **SonoAVC** tab.
3. Switch sect. plane A to full size if not present.
4. Adjust the sect. plane A with x, y, z rotation, parallel shift and zoom to achieve the correct measuring plane.
5. Select **SonoAVC**.
6. Select **Manual** and adjust the ROI shape, if desired.
7. Start the measurement by selecting either Left Ovary or Right Ovary.
8. The green measurement cross appears in the middle of the image.
9. Perform measurements:
 1. Position the start-point of the long diameter with the Trackball and press **Set**.
 2. Position end-point of the long diameter with the Trackball and press **Set**.
The first measurement displays.
 3. Position the start-point of the short diameter with the Trackball and press **Set**.
 4. Position end-point of the short diameter with the Trackball and press **Set**.
The second measurement displays.
 5. To start the next measurement, move the Trackball and repeat step 9.
10. Select **Save** to save the measurements to the current exam.

Semi Auto

1. Acquire 3D volume or reload previously acquired volume.
2. Select **SonoAVC**.
3. Select **SemiAuto** and adjust the ROI shape, if desired.
4. Start the measurement by selecting either Left Ovary or Right Ovary.
5. Measure all desired objects on the current plane:
 1. Position the system cursor over the object to be measured and press **Add/Rem**.
 2. Click the cursor by pressing **Set** on the object.
The selected trace with corresponding color and number is displayed.
 3. Position the system cursor over the next object to be measured.

Advanced Features

4. If the measurement is not finished, repeat step 5.
6. Select **Save** to save the measurements to the current exam.

Auto

1. Acquire 3D volume or reload previously acquired volume.
2. Select **SonoAVC**.
3. Select **Auto** and adjust the ROI shape, if desired.
4. Start the measurement by selecting either Left Ovary or Right Ovary.
The rendered objects and the result list are displayed on the screen.
5. Select **Save** to save the measurements to the current exam.

Add Manual

1. Press **Add Manual** button.
2. The measurement cross appears in the middle of the sect. plane A.
3. Perform measurements:
 1. Position the start-point of the long diameter with the Trackball and press **Set**.
 2. Position end-point of the long diameter with the Trackball and press **Set**.
The first measurement displays.
 3. Position the start-point of the short diameter with the Trackball and press **Set**.
 4. Position end-point of the short diameter with the Trackball and press **Set**.
The second measurement displays.
 5. To start the next measurement move the Trackball and repeat step 3.

OmniView

OmniView provides an arbitrary cross-sectional plane, as opposed to a strict coronal plane. In VCI OmniView, the rendering box is very thin so you can visualize the tissue information of a thick slice. The resulting image shows the average (integrated) gray value of the tissue contained within the narrow box.

VCI OmniView improves the contrast resolution and the signal/noise ratio and therefore facilitates the detection of diffuse lesions in organs. The result is an image with no speckle pattern and a highly improved tissue contrast.

- 4D Real Time
- 3D Static
- STIC

OmniView Workflow

1. Press **3D/4D**.
2. Select **OmniView**.
3. Adjust Volume Angle and B Quality as necessary.

4. Press the right **Set** key to start acquisition.
The volume acquisition starts and the acquired images are displayed.
5. The cross cursor displays on the reference image.
6. Select the appropriate trace mode to display the sectional plane.
 - Line: Select the sectional image by using straight line.
 - Curve: Select the sectional image by using curved line.

Move the cursor to the start point and press **Set**. Move the cursor to the end point and press **Set**. Use the **Trackball** to make a curved line and press **Set**.
 - Trace: Select the sectional image by using any arbitrary curve.
 - Polyline: Select the sectional image by using a continuous line composed of one or more line segments. You can create a polyline by specifying the endpoints of each segment by press **Set**.

OmniView Main Menu

Table 13-19 • OmniView Parameters on Edit Touch Panel

Parameter	Description
View Icon	Show or hide the OmniView Icon.
View Line	Show or hide the OmniView Line.
VCI OmniView	Switch to VCI Omniview.
Trace Mode	Four tracking line methods are available.
Tile	Select a display format (Single, Dual, Quad)
Clear All	All existing lines are deleted and a new line entry is started.
Ref. OmniV	To activate OmniView Ref. Line view.
Ref. Image	Select the Reference image.
OmniV. Rot	Rotates the OmniView line.
Undo	To re-adjust the traced line.
Mix	Selections: 0-100% in increments of 2. Allows you to mix a Rend Mode 1 mode with a Rend Mode 2 mode. Always select two modes.
Slice Thickness	Select the slice thickness

OmniView Control



Figure 13-31 • OmniView Control

Table 13-20 • OmniView Parameters on Line Edit Mode: New

Parameter	Description
Img.	Activate the image.
Redo	Restart the line.
OmniView	Activate OmniView Line
View#	Select the next OmniView image and line as 1, 2, or 3.
Next	Select to activate the next OmniView Ref. Line view.

STIC (Spatio-Temporal Image Correlation)

Overview

With this acquisition method the fetal heart or vascularity can be visualized. It is not a Real Time 4D technique, but a post processed 3D acquisition.

STIC is designed for beating (fetal heart) as well as blood perfused organs. Only STIC can synchronize structures that have a pulsation in Doppler mode, but no visible pulsation in B-Mode.

Data is acquired for a predefined period of time. The acquired images are post processed to calculate a Volume Cine sequence representing one complete heart cycle.

In order to achieve a good result, try to adjust the size of the volume box and the sweep angle to be as small as possible. The longer the acquisition time, the better the spatial resolution will be. The user must be sure that there is minimal movement of the participating persons (e.g., mother and fetus), and that the probe is held absolutely still throughout the acquisition period. Movement will cause a failure of the acquisition. If the user (trained operator) clearly recognizes a disturbance during the acquisition period, the acquisition has to be canceled.

A good STIC data set shows a regular and synchronous beating of the fetal heart or of an artery. Please make sure that the borders of the fetal heart or the artery are smooth and there are no sudden discontinuities. Always adopt a critical attitude to images created in STIC mode.

If the expected frame rate is too low (10Hz) for a good STIC quality, a warning is displayed in STIC mode.

After the STIC acquisition is finished the calculation process starts to calculate the volume cine sequence. If a result is detected by the system, the Volume cine sequence is shown in run mode and the STIC accept menu appears. As soon as the result is accepted the system releases the volume cine mode. If the result is not accepted but canceled, the system switches back to STIC pre mode.

One or more of the following artefacts in the data set indicate a disturbance during acquisition:

- Sudden discontinuities in the reference image B: These are due to the motion of the mother, the fetus or fetal arrhythmia during acquisition.
- Sudden discontinuities in the color display: Motion of the mother, the fetus or fetal arrhythmia affects the color flow in the same way it affects the gray image.

- Fetal heart rate far too low or far too high: After acquisition the estimated fetal heart rate is displayed. If the value does not correspond to the estimations based on other diagnostic methods at all, the acquisition failed and has to be repeated.
- Asynchronous movement in different parts of the image: e.g., the left part of the image is contracting and the right part is expanding at the same time.
- The color does not fit the structures displayed in gray mode: The color is displayed above or below the actual vessel.
- Color “moves” through the image in a certain direction: This artefact is caused by a failure in detecting the heart rate due to low acquisition frame rate. Use higher acquisition frame rate for better result.

NOTE

In all of the above cases the data set has to be discarded and the acquisition has to be repeated.

When is it not allowed to perform the STIC fetal cardio acquisition?

- Severe fetal arrhythmia

Following STIC acquisition modes are available:

- STIC
- STIC CFM
- STIC PDI

Performing STIC Scan

NOTE

STIC can only be used in OB applications.

1. Select a 4D probe from the probe indicator.
2. After obtaining a feasible 2D image of the fetal heart or an artery, press 3D/4D to activate the Volume mode.
3. Select **STIC**.
4. Set the Volume of Interest (VOI) to be rendered. Use the Trackball to move and resize the VOI.

NOTE

Press top Trackball key to switch Pos and Size.

5. Adjust the **Acquisition Time**.
6. Set the volume sweep angle by using **Volume Angle**.

NOTE

In order to acquire a good result, try to adjust the ROI and the sweep angle to be as small as possible. The longer the acquisition time, the better the spatial resolution will be.

7. Hold the probe still and ask the patient not to move.
8. Press the right **Set** key to start the acquisition.

NOTE

When the STIC acquisition begins, the Stop Acquisition button appears. If the user recognizes a movement of the probe, fetus or the patient during the scan, press Stop acquisition to cancel the acquisition.

9. Hold the probe steady until the system stops automatically. Estimated Fetal Heart Rate and Accept/Cancel buttons appears on the Touch Panel.
10. Press **Accept** if the fetal heart rate displayed on the Touch Panel is accepted. STIC image is displayed.
If the result is not accepted, select **Cancel**. The system switches back to STIC-pre mode.

NOTE

If the user (trained operator) clearly recognizes a disturbance during the acquisition period, the acquisition has to be canceled.

Screen Layout

The yellow caution icon and the calculated heart rate are displayed as well.

NOTE

Displayed Heart Rate indicates the heart rate [B/min] calculated from the delta time length per beat. A yellow caution icon indicates that the displayed heart rate is only an estimation. Do not diagnose based on this value.



Figure 13-32 • Screen Layout

STIC Controls

Touch Panel/All hard key and trackball controls are the same as in 4D/Static 3D mode. Therefore only the STIC relevant controls are described here:

Table 13-21 • STIC Pre-mode Touch Panel

Relevant controls	Description
Acq. Time	Only available in STIC. Acquisition time can be set.
Volume Angle	Adjusts the volume angle.

Contrast Imaging

Overview

NOTE

The is designed for compatibility with most commercially available Ultrasound contrast agents. Because the availability of these agents is subject to government regulation and approval, product features intended for use with these agents may not be commercially marketed nor made available before the contrast agent is cleared for use. Contrast-related product features are enabled only on systems for delivery to an authorized country or region of use.



WARNING

Misdiagnosis based on image artifacts

Misdiagnosis in ultrasound contrast images may be caused by several artifacts, most importantly:

Motion artifacts: Give rise to signals independent of contrast presence. This may be caused by patient movement (including respiration) or by probe movement influenced by the operator.

Regional drop outs: Caused by unintentional destruction of the contrast agent, too low concentration of contrast agent, poor acoustic penetration due to rib/lung shadows or the system failing to detect the contrast agent due to erroneous settings induced by the operator.

Tissue harmonics: Creates contrast-like signals independently of the presence of contrast agent.



CAUTION

Cardiac rhythm disturbances during cardiac perfusion studies using gas ultrasound contrast agents have been observed in the diagnostic range of Mechanical Index (MI) values. See the specific package insert for the contrast agent being used for details.



CAUTION

Read and follow contrast agent instructions provided by the manufacturer.

Contrast Imaging Overview

By adjusting the acoustic output, you can enhance either contrast harmonics or stimulated acoustic emission (SAE).

See Chapter 11 for probe availability.

Coded Contrast Imaging utilizes three technologies: Amplitude Modulation (AM), Phase Inversion (PI), and Coded Harmonic Angio (CHA).

- Contrast: Amplitude Modulation and Phase Inversion (AM and PI) is included in the Contrast option.

Table 13-22 • Available Contrast Display Settings

Label	Description	Clinical Use
Low MI B--Single Low MI B-Mode image display mode	Find the tumor before/after contrast agent injection	Tumor detection and characterization
Contrast Only—single contrast image display mode	Use contrast agent nonlinearity to detect contrast agent signal.	Tumor detection and characterization
Single View--Single Contrast enhanced image display with low MI B-Mode image data acquisition mode	Use contrast agent nonlinearity to detect contrast agent signal.	Tumor detection and characterization
Dual View—Real-time side-by-side dual image display with the right side contrast image and the left side B-Mode reference image	Use contrast agent nonlinearity to detect contrast agent signal.	Tumor detection and characterization

Benefits

The contract imaging technique, facilitated by the use of a contrast agent, detects non-linear signals while suppressing linear signals from surround tissue. Blood containing the contrast agent stands out brightly against a dark background of normal tissue.

Clinical Use

Possible clinical uses are to detect and characterize tumors of the liver, kidney, and pancreas and to enhance flow signals in the determination of stenosis or thrombus.

Affect on other controls

The default acoustic output adjusts for contrast imaging and the Power Output key provides more subtle gradations for use while in contrast imaging. When you exit Contrast Imaging, the system returns the acoustic output to its original setting. When you reactivate Contrast Imaging, the system enters the default Contrast Mode.

Most system controls are available (Depth, Zoom, Colorize, etc.). However, some controls are not available (Anatomical M Mode, Rejection, and Suppression).

Controls adjusted while in Contrast Imaging retain these values when you exit Contrast Imaging (except for post-processing controls).

Bioeffect

Activating Contrast Imaging may change the TI and/or MI. Observe the output display for possible effects.

The **Power Output** key has been enhanced to provide more subtle gradations for use while contrast imaging. The Mechanical Index displays values less than 0.1. These values are displayed on CINE Loops and on archived images.

Feature Availability

3D and Volume Navigation are available; Multi Image and LOGIQView are not available.

Technique Availability by Probe

Use the table in Chapter 11 to determine which contrast technique is available by probe.

Mode

Reference Mode

Reference (Ref) Mode is to image the anatomical reference, not the contrast enhancement.

Contrast Mode

There are several contrast imaging techniques. Note that the appropriate imaging technique may vary by agent and application. In other words, the imaging technique is not dedicated for the agent and vice versa.

Contrast MVI Mode

Contrast MVI uses MVI (Micro Vascular Imaging) in Contrast mode. The characteristics of MVI's higher sensitivity and higher spatial resolution can be used in Contrast Imaging.

In addition to the MVI features, Flash, Flash + Capture, Max Enhance, and Contrast Clock can be used.

The choice of AM, PI, CHA mode in Contrast is not in Contrast MVI.

Support probes: C1-6-D, C1-6VN-D, L2-9-D, L2-9VN-D, L3-12-D and ML6-15-D.

Contrast Presets

You access the Contrast Imaging Presets via **Utility > Imaging > Ref or Con tabs**.

Acoustic Output Presets

To set the acoustic output value for Contrast Imaging, press the down arrow next to Acoustic Output (%) on Con tab to select the value.

To set the acoustic output value for the tissue image while in Single and Dual View mode, press the down arrow next to Acoustic Output for Single & Dual View (%) on Ref tab to set the value.

To set the acoustic output value for the tissue image while in Low MI B-Mode, press the down arrow next to Acoustic Output (%) on the Ref tab to set the value.

Time Delay

To set the Trigger time delay, press the down arrow next to Time Delay.

Dual View

To set the default contrast mode, press the down arrow next to Default Mode on Con tab to select the mode.

Additional Presets

To adjust other Contrast Imaging presets (Map, Frame Average, etc.), press the down arrow to adjust the setting. *Imaging Presets* on page 473

Sonazoid™ Contrast Agent

Sonazoid™ is a contrast agent approved in some countries. Refer to local clearances and availability in your respective market(s).

Sonazoid is a different microbubble, compared to Sonovue. You can select the Sonazoid contrast agent via **Utility > Imaging > CON**.

Contrast Controls

Max (Maximum) Enhance

Description

Sets the acoustic output to its maximum setting (100%)

Values

On/Off. When you deactivate Max Enhance, the acoustic output is returned to its previous setting. Max Enhance is deactivated by the system when you turn it off, change probes, or change the contrast technique.

Benefits

This control provides quick transition to High MI imaging. This allows the user a quick one-button push to destroy the agent. Useful when the user is interested in the bubble wash-in characteristics of the anatomy being scanned.

Contrast Clock (Timer)

Description

You can use the Contrast Clock by activating it at the time of injection and deactivating it at the end of the exam.

Two timers, Contrast Clock1 and Contrast Clock2 can be displayed on the bottom, left-hand corner in the image area and info area for several injections.

NOTE

You can also configure the system to perform a countdown for the contrast injection with the **Utility > System > System Imaging > Countdown Time** for Contrast preset.

Values

On/Off. You deactivate the Contrast Clock via the Touch Panel control or by starting a new patient.

Display

There are two areas on the screen where the Contrast Clock displays: on the image and on the lower, left-hand portion of the display. The timer on the image freezes when you freeze the image (the timer updates when you unfreeze the image). However, the timer located on the lower, left-hand portion of the display continues to display over a freeze, probe change, mode change, multi image, and zoom.

The timer also appears on CINE Loops and archived images.

Benefits

Advanced Features

The Contrast Clock measures the time since injection.

You can save the data of contrast clock to an external file by using Export Traces of TIC.

1. Press **Freeze**. Scroll with the Trackball to show Cine tab.
2. Press **TIC Analysis** on the Touch Panel to enter TIC application.
3. Put a ROI on the image.
4. Press Export Traces. Type the file name and store it to the storage device.

Contrast Clock Highlight

Contrast Timer T1 is zoomed and blinking for 'Highlight Time' (sec) from 'Highlight Time' (sec) before each 'Interval' (sec).

See *Table 10-25 • Contrast Clock Highlight* on page 462.



Figure 13-33 • Contrast Clock Highlight

Accumulation

Description

Accumulation enhances the flow in an image.

Values

8 settings: 0=Off, 0.2, 0.4, 0.6, 0.8, 1.6, 3.2, and Infinite.

If Accumulation is turned off, then Frame Averaging is used; if Accumulation value is set, then Accumulation is used.

Availability

Available in Contrast, Color Flow, and PDI.

Benefit

Accumulation detects the maximum signal and holds it for the level specified (1-7).

SRI Usage

Description

You can use SRI-HD for the tissue image and the contrast image independently or together.

Adjusting

You can preset SRI Usage in **Utility > Imaging > Ref** tab.

Visualization

Description

Define the display technique. Only available for Single View or Dual View.

Values

- Contrast. Displays the contrast-enhanced image.
- Tissue. Displays the tissue image.
- Hybrid Contrast. Displays the contrast-enhanced image and the tissue image using Hybrid Map.

Hybrid Map

Description

Select the hybrid map for the Hybrid Contrast visualization in the Dual/Hybrid Display.

Flash

Description

This feature provides a way to expose the higher acoustic power for a specified time duration by pressing a control once.

NOTE

Set the frame numbers to scanned with the higher acoustic power in **Utility > Imaging > Con > Flash Frames**. The frame numbers defined by Flash Frames are applied to both the contrast imaging modes.

- If you select Flash once in the Con menus, the system scans with 100% acoustic output burst pulses for the specified number of frames. The acoustic output then reverts to the original settings.
- When Max Enhance is ON for the contrast imaging modes, the system keeps Max Enhance = ON with no acoustic output change when Flash is selected in the Con menu.

Relationship with other controls

- L + R Simultaneous Display (L: Tissue, R: Active Visualization)
 - Accumulation/Cine Capture
 - Applied on the right side image only.
 - You cannot compare On and Off image using L + R.
 - SRI-HD
 - Applied on both side image.
 - Frame Average
 - Applied on both side image.
- TIC
 - Measured on a recalled CINE, except "Hybrid Contrast".

- “Hybrid Contrast” is disabled and measured on “Contrast” visualization.
- Easy 3D
 - Build volume data from active visualization, except “Hybrid Contrast”.
 - Hybrid map is disabled and forces “Contrast” visualization.
- Advanced 3D/Tru3D
 - Build volume data from active visualization.
 - For the Single/Dual View data set of contrast mode, both reference and contrast data are rendered in individual volume segments. Each volume data is manipulated using Active Data and Visual Data.
- 3D
 - Hybrid Map is disabled and forces “Contrast” visualization.
 - The Cine frames have “Contrast” data only.
- Archive
 - The raw data size increases between two to three times the size compared to the previous raw data. It takes a longer period of time to save as Cine.

Static 3D with Contrast

Static 3D is available while in Contrast imaging mode while imaging with 4D probes.

1. Activate Contrast imaging.
2. Adjust the image, as necessary.
3. Activate the Contrast Clock (if you want the acoustic output set at 100%). Or, you can activate these when you inject the contrast agent.

NOTE

To restart the Contrast Clock, you need to turn it off, then back on.

4. Activate **3D > Static 3D**. Specify the quality, angle, and render mode (Sectional, Render, and VCI Static are supported).
Wait to inject the contrast agent until after you have set up your pre-mode data. Adjust the ROI and other parameters before injecting the agent.

NOTE

Accumulation and Trigger turns off or resets to 0 when you activate Pre-mode.

5. Activate the Contrast Clock and inject the contrast agent, then press the Start Key to activate the Static 3D acquisition as soon as the contrast agent washes in.

The system automatically ‘stores’ the sweep (to internal memory, not to the clipboard).

NOTE

You may want to set up a Fast Key to automate these keystrokes. See Chapter 6 for more information.

- To store this VOI, press a print key. All CINE and volume data loops are saved in 3D when you press P1. Store loops from Static 3D or from the Volume Review tab in order to store the complete volume.

NOTE

You can also configure the system to perform a count down for the contrast injection via **Utility > System > System Imaging**.

- At the completion of the sweep, the Static 3D tab appears.

NOTE

The Contrast Timer on the Volume Review images represents a timestamp of when the image was captured relative to when the timer was started.

- Press the **Start** key to acquire volume data.

CEUS LI-RADS Feature

This feature allows you to create the check sheet which follows ACR CEUS LI-RADS 2017 by pressing CEUS LI-RADS on CINE tab. For additional details, refer to ACR Reporting and Data Systems (RADS) documentation.

Figure 13-34 • CEUS LI-RADS Check sheet

Table 13-23 • CEUS LI-RADS Check Sheet Controls

Preset parameter	Description
CEUS LI-RADS Category (if needed)	Input your evaluation (if needed).
Cancel	Close the sheet.
Print	Store the screen capture.

Time Intensity Curve (TIC) Analysis

Overview

Time Intensity Curve (TIC) enables the user to perform the following analysis:

- Time-Intensity analysis allows instant time-intensity calculation from up to eight regions of interest.
- Curve fitting analysis for research studies of contrast agent concentration rates.

The basic TIC process works as follows:

1. Scan the patient after injecting the contrast agent.
2. Observe the agent flow through the anatomy of interest.
3. When the desired contrast effect has been visualized, freeze the image and select a range of images for analysis.
4. Position an ROI (region of interest) on one of those images where the contrast effect is visible.
5. The system then calculates the mean pixel intensity within that ROI for all frames in the user designated loop and plots the resulting data as a function of time.

You can also choose to fit this data to one of several mathematical functions. The fundamental idea is that the contrast effect flowing through the organ of interest can be modeled mathematically, and details of the wash in and washout of the agent can be gleaned by analyzing the numerical parameters of the mathematical model.

Activating TIC

Starting TIC in cine loop

1. Open an examination and select a contrast cineloop.
2. Select **TIC Analysis** on the Touch Panel.

Starting TIC in live mode

1. Scan and freeze the patient in Contrast live mode.
2. Move the trackball to activate Cine.
3. Select **TIC Analysis** on the Touch Panel..

TIC Analysis for Multiple Clips

When additional contrast data is loaded on the clipboard in TIC mode, the system registers the contrast clips to the Analysis List (up to 10 clips in one TIC analysis session). The system maintains individual TIC settings (such as ROI shape, position, smoothing, motion tracking state) for each clip. You can recall the clip with the “Graph #” button on the Touch Panel. The “Remove Graph” button removes a selected clip from the Analysis List.


The registered clip list and TIC settings are reset by exiting the TIC mode.

Exiting TIC Analysis

There are several methods to exit TIC Analysis.

- Select **Exit TIC Analysis** on the TIC Touch Panel.
- Press **Freeze** to unfreeze and resume scanning.
- Press any other button that returns the system to real-time scanning.


TIC Analysis Screen Description



1. Contrast cineloop window
2. B-Mode cineloop window
3. Analysis window
4. Sample area
5. Time and velocity at cursor position
6. Sample area tools
7. Layout icons
8. Frame marker

Figure 13-35 • TIC Analysis Screen - Graph with dual image layout (example)

Table 13-24 • Cineloop windows

Graphic	Description
	Displays Contrast image data Sample area: Indicates sampling position of the intensity trace. The sample area is color-coded: the first sample area is yellow, the second blue...etc.
	Displays B-Mode data Sample area: Indicates sampling position of the intensity trace. The sample area is color-coded: the first sample area is yellow, the second blue...etc. NOTE B-Mode image is not displayed when cine clip stored in contrast only mode.
 	System menu on Sample Area This menu is displayed by pressing the left Set key when the cursor is placed over a sample area in one of the Cineloop windows. NOTE The system menu is dependent on mode. • Set As Default ROI Size: Displays on elliptical ROI. • Label Sample Area: Sets a descriptive name to the sample area. The label is useful for identification of the sample area when exporting data. • Delete Anchor (On ROI setting as anchor point only) • Copy Sample Area • Copy & Move • Copy & Move (Same Depth) • Move (Same Depth) • Set Start Frame: Set start frame for current ROI to calculate the TIC parameters and fitting curves. • Set End Frame: Set end frame for current ROI to calculate the TIC parameters and fitting curves. • Cancel: exits the System menu. System menu on Image when copy sample area is selected. • Paste Sample Area


Graphic	Description
	Displays time-intensity curve. • Y axis: Intensity scale (logarithmic) (db) or linear acoustic units (AU). • X axis: Time(s) or Dt(s), elapsed time from previous frame. • ECG (where available -- not shown): displays ECG trace (where available). • Frame Marker: the current frame marker and the start and stop markers for the cineloop. • Time at cursor position and velocity at cursor position. • Intensity (dB or AU) at cursor position. • Intensity (dB or AU) at frame marker position (color coded)
	System menu of the analysis window This menu is displayed by pressing left Set key when the cursor is in the Analysis window. Note: The system menu is dependent on mode. • Vertical Auto-Scaling: selects between full unit range or a range according to the maximum and minimum values of the displayed trace(s). Delayed, On, Off. • Vertical Unit: toggles between logarithmic (dB) and linear acoustical units (AU). Vertical unit of previous analysis is retained. • Horizontal scale: Only on a loop including different frame rate. • Line Style: selects between solid line only or solid line with square markers at each data point. • TIC Parameters: The TIC parameters dialog appears. • Curve Fitting Parameters: toggles between Wash-in, Wash-out, Gamma Variate and off. • Gradient: On or Off. • Gradient Plot: Gradient, Gradient Derivative, All or Off (On Graph with Dual Image Layout and Small Data Layout only). • ECG Triggering: Only on a loop including the ECG cycle. • Cancel: exits the System menu.

Table 13-25 • Analysis windows


Graphic	Description
	Displays time-intensity curve. - Y axis: Intensity scale (logarithmic) (db) or linear acoustic units (AU). - X axis: Time(s) or Dt(s) [Diastolic time in seconds], elapsed time from previous frame. - ECG (where available -- not shown): displays ECG trace (where available). - Frame Marker: the current frame marker and the start and stop markers for the cineloop. - Time at cursor position and velocity at cursor position. - Intensity (dB or AU) at cursor position. - Intensity (dB or AU) at frame marker position (color coded)
	System menu of the analysis window This menu is displayed by pressing left Set key when the cursor is in the Analysis window. Note: The system menu is dependent on mode. - Vertical Auto-Scaling: selects between full unit range or a range according to the maximum and minimum values of the displayed trace(s). Delayed, On, Off. - Vertical Unit: toggles between logarithmic (dB) and linear acoustical units (AU). Vertical unit of previous analysis is retained. - Horizontal scale: Only on a loop including different frame rate. - Line Style: selects between solid line only or solid line with square markers at each data point. - TIC Parameters: The TIC parameters dialog appears. - Curve Fitting Parameters: toggles between Wash-in, Wash-out, Gamma Variate and off. - Gradient: On or Off. - Gradient Plot: Gradient, Gradient Derivative, All or Off (On Graph with Dual Image Layout and Small Data Layout only). - ECG Triggering: Only on a loop including the ECG cycle. - Cancel: exits the System menu.

Table 13-26 • Layout icon

Graphic	Description
	You can select layout from following. • Graph with Dual Image Layout • Small Data Layout • Large Data Layout • Everything Layout • Single Image Layout • Dual Image Layout
	Indicates an analysis graph.
	Indicates an image One icon means Contrast image. Two icons mean Contrast and B-Mode.
	Indicates parameters. 10 dots means that all parameters display. 4 dots means that only displays high level parameters.

Table 13-27 • Sample area drawing tool


Graphic	Description
	Creates a sample area based on freehand drawing.
	Creates a sample area with a pre-defined circular/elliptical shape.

Table 13-28 • Trackball Assignment

Graphic	Description
	Press the top Trackball key to toggle the trackball assignment between QA and Scroll. • QA Pointing tool in TIC mode. • Scroll When the cineloop is stopped, enables scrolling through the cineloop. • Menu Press left Set key to display System Menu. • Set

TIC Touch Panel

Table 13-29 • TIC Touch Panel Description

Parameter	Description
Exit TIC Analysis	Exit TIC.
Motion Tracking	Users use TIC to analyze tumor characteristics precisely, without distortion due to patient movement. Motion Tracker enables the system to automatically adjust the ROI's placement across multiple frames in order to accommodate patient breathing or body movements. To activate, press Motion Tracking on the Touch Panel. This starts the calculation to adjust the ROI positions for every image frame. After completion, the ROI graphic on each frame is changed to the ROI with the anchor.
Accumulation	Enhances the flow in an image.
Enable All Frames	Re-enables disabled frames.
Curve Fitting	Toggles between Wash-In, Wash-Out, Gamma Variable and Off.
Smoothing	Smooths the trace displayed by applying a filter over a defined time window. Both the filter type and time window are user-selectable. The type of filter available depends on the analysis signal displayed.
Delete Sample Area	Removes selected sample area from the CINE Loop window and accompanying trace in the Analysis window. The Trackball marker must be pointed at an anchored sample area.
Auto Calc Range	Auto Calc Range enables system to automatically estimate the ROI's start frame and end frame to calculate TIC parameters and fitting curves. The start and end frames for each ROIs are displayed by small squares on the graph. Toggles between Current Sample, All Samples and Reset All. 
First	Move to the first frame of cineloop.
Last	Move to the last frame of cineloop.
Run/Stop	Start/Stop the cineloop review.
Loop Speed	Adjust the cine loop playback speed.
ROI width/height	Move the rotary left/right or up/down to adjust ROI width/height.
ROI tilt angle	Rotate the rotary to adjust ROI tilt angle.
Start Frame	Rotate the rotary to select the start frame and push to set the frame.
End Frame	Rotate the rotary to select the end frame and push to set the frame..
Frame by Frame	Rotate the rotary to review the CINE image frame by frame manually.

Parameter	Description
Disable frame	Push the rotary to disable the selected frame.
Graph #	Select the clip for general TIC analysis. System can register up to 10 clips and the individual parameters for each TIC analysis are maintained in one TIC analysis session.
Remove Graph	Remove the selected graph from the clip list in Graph # button.
Merge Graphs	Referring to Contrast Clock1 in the stored clips, system merges the TIC graphs. Contrast Clock1 and at least 2 ROIs are required for Merge Graphs

Generating a Trace

Up to eight traces can be generated.

About the sample area

The sample area can be in three different states:

- Free sample area: freely moving sample area (QA cursor) before anchoring.
- NOTE**
The free sample area disappears when the QA cursor is moved over a static anchored frame.
- Static sample area: the free sample area is anchored by pressing **Set**.
 - Dynamic anchored sample area: the sample area is anchored in two or more frames (see Manual tracking below). In these particular frames, the sample area is displayed with an anchor. The sample area moves smoothly between the anchored positions when playing/scrolling the cineloop.

Trace from a pre-defined sample area (Ellipse ROI)

- Press the top **Trackball** key until the QA trackball assignment is selected.
- If necessary, select the sample area Ellipse ROI button (shape icon on the monitor display).
- Move the cursor to one of the Cineloop windows using the **Trackball**.
The cursor is changed to a sample area (white circle). A preview of the trace is displayed in the Analysis window.
- Press **Set** to anchor the sample area.
In this frame, the sample area is marked with an anchor. If the cineloop has more than one heart cycle, a sample area will also be anchored in the corresponding frame in the next heart cycle.
The trace is updated accordingly in the Analysis window.

Trace from freehand sample area

- Select the Freehand ROI button (pencil icon on the monitor display).
- Move the cursor to one of the Cineloop windows using the **Trackball**.
- Trace the outline of the desired ROI by moving the caliper with the **Trackball**.
- Press **Set** to anchor the sample area.

The sample area is automatically closed and the trace is updated accordingly in the Analysis window.

Copy, move and paste a Sample Area

To copy and paste the ROI,

1. Move the cursor over the ROI and press the left **Set** key. The system menu displays.
2. Select **Copy sample area**.
3. Move the cursor to the desired location for the copied ROI and press the left **Set** Key. The system menu displays.
4. Select Paste sample area.

To copy and move the ROI,

1. Move the cursor over the ROI and press the left **Set** key. The system menu displays.
2. Select **Copy & move**. Or if you want to move to the same depth as the original ROI, select **Copy & move (same depth)**.
3. Move the copied ROI using the **Trackball**. Press **Set** to fix the position.

To move the ROI,

1. Move the cursor over the ROI and press the left **Set** key. The system menu displays.
2. Select **Move (same depth)**.
3. Move the ROI using the **Trackball**. Press **Set** to fix the position.

Manual tracking of the sample area (dynamic anchored sample area)

The sample area can be moved within the loop to ensure that data in the trace is generated from the same anatomical location during the cyclic motion of the heart.

1. Place a sample area over a region of interest. Note the anatomical location of the sample area.
2. Scroll to a new frame using the **Trackball**.
3. Press the top **Trackball** key until the QA trackball assignment is selected.
4. Move the cursor to the sample area using the **Trackball**.
5. Press **Set**. The sample area is unanchored.
6. Drag the sample area to the corresponding anatomical location in the new frame.

When the sample area is anchored in more than one frame, linear interpolation is performed so that the sample area is smoothly moved between the anchored positions in the selected frames when running the cineloop.

NOTE

In the original frame and this particular frame the sample area is marked with an anchor.

7. Press the top **Trackball** key until the scroll trackball assignment is selected.
8. Using the **Trackball**, scroll through the cineloop and control that the sample area follows the moving anatomical structure.

9. Add anchored sample areas in several frames to obtain a more accurate displacement of the sample area.

Moving a dynamic anchored sample area

1. Freeze the image.
2. Press top Trackball key until the scroll trackball assignment is selected.
3. Using the **Trackball**, browse through the cineloop to display one of the frames where the sample area was anchored.

NOTE

In these frames, the sample area is marked with an anchor.

4. Press top Trackball key until the QA trackball assignment is selected.
5. Move the cursor to the sample area using the **Trackball**.
6. Press **Set**. The sample area is unanchored.
7. Drag the sample area to a new location.
8. Press **Set** to anchor the sample area to the new location.

If you want to move the sample area to the same depth, select Move (same depth) from the System Menu.

Zooming in the Analysis window

To zoom:

1. In the Analysis window, press and hold down the **Set** key while dragging the cursor to define the zooming area.
2. Release the **Set** key.

To unzoom:

1. Press the left **Set** key in the Analysis window. The system menu displays.
2. Select **Unzoom**.

NOTE

Shown only in zoom mode.

Delete a trace

The user can delete all traces at once or one at a time.

1. If necessary, press the top Trackball key until the QA trackball assignment is selected.
2. Move the cursor over one of the sample area. Confirm that cursor is changed to hand icon.
3. Press the **Delete Sample Area** on the touch panel.
4. Select **Current Sample** or **Delete all** as necessary.

NOTE

The corresponding traces for the deleted ROIs are erased from the plot.

Disabling/Enabling the frame

Frame disabling excludes the actual frame from the cineloop display. Frame disabling is available only with contrast data.

Disabling the frame from the frame marker

To disable One Frame :

1. Use the **Trackball** to move the cursor to the frame on the Frame Marker which you want disable.
2. Press **Set** to disable the frame.
3. The frame marker is changed from green to red to indicate the frame has been disabled.

NOTE

The disabled frame is no longer displayed in the reference window when scrolling through CINE memory.

Disabling multi-frames from the frame marker

1. Use the Trackball to move the cursor to the first frame on the Frame Marker.
2. Press and hold down **Set**
3. Move the cursor with the Trackball to the last frame to be disabled and release **Set**.

The marker is turn red and the data from that frame is removed from the trace and any subsequent trace processing.

Disabling ECG triggered frame (where available)


In a multi-cycle acquisition, the user may deselect all frames in all heart cycles but a selected one. This function can be used for example to select a particular systolic frame for each heart cycle.

1. Scroll through the cineloop to identify the cardiac phase to analyze or identify the cardiac phase on the ECG trace (where available).
2. Position the cursor on the analysis window and press left **Set** key. The system menu displays.
3. Select **ECG triggering** (where available).

All frames in all heart cycles are disabled except for the selected and corresponding frames in the other heart cycles.

To enable the frames

1. Select **Enable all frames** on the Touch Panel.
2. All disabled frames are re-enabled.



1. Analysis Window
2. Frame markers axis
3. Enabled frame (Green)
4. Disabled frame (Red)
5. ECG (where available)
6. Current frame

Figure 13-36 • Frame markers

Manipulating the Sample Area

Up to eight ROIs can be saved on the reference image, with the corresponding eight traces plotted simultaneously on the graph. Each ROI display has a different color, and its corresponding trace data is plotted using that same color.

Once eight ROIs have been saved, the system does not automatically generate an active ROI when the cursor is positioned over the displayed reference image.

The saved ROIs can be a mixture of elliptical and freehand ROIs.

When the user repositions an ROI, the old trace data is erased from the plot and the trace data for the new position replotted.

If the ROI position on the last frame of the selected image range is moved, the corresponding ROIs on all frames are repositioned to match the last frame.

The user shall also have the capability of setting separate ROI positions on different frames of the contrast images, and the system shall linearly interpolate the ROI positions for the frames in between the selected frames.

Setting the default sample area shape

1. Place the cursor on the sample area.
2. Press left **Set** key. The system menu displays.
3. Select **Set as Default ROI Size**.
4. The current ROI size is set as the default for subsequent Ellipse ROIs.

Reshaping a Sample Area

To reshape the sample area:



Figure 13-37 • ROI

- Move the rotary up and down to change the height.
- Move the rotary left and right to change the width.
- Rotate the rotary to change the tilt angle.

Labeling a Sample Area

The sample area label is used to identify data associated with the sample area when exporting.

1. Position the cursor on the ROI to label and press the left **Set** key.
2. The ROI system menu displays. Select **Label sample area**. The Label Dialog box displays.
3. Enter a name for the sample area.
4. Select **OK**.

TIC Plot Control

Vertical Unit

When analyzing the contrast data, the Y-axis can be set to display either logarithmic scale (dB) or linear, acoustic units (AU) for both tissue intensity (2D) or Angio intensity data.

To toggle between dB and acoustical display units for the Y-axis. The unit of previous analysis is retained.

- **dB**—The traditional log compressed B-Mode data is used to calculate the time-intensity curve values.
- **Acoustic**—The system reverse the log compression function to provide un-log compressed data for the TIC analysis.

Vertical auto-scaling

The system can be configured to display the full unit range or a range according to the maximum and minimum values of the displayed trace(s) (auto-scaling function). In addition, the auto-scaling function can be set to be live update (updates while the sample area is moved) or delayed (updated when the sample area is anchored).

- **Delayed**—The system automatically rescale the vertical axis of the trace graph only when a new ROI is saved, to account for changing input dynamic range.
- **On**—The system automatically rescale the vertical axis of the trace graph every time the currently selected (active) ROI is moved.
- **Off**—Disable any automatic scaling of the vertical axis. There is user-defined system defaults on the system preset page for the fixed vertical scale to be used for the plot.

Y-Scale

1. When you select “Off” for Vertical Auto-Scaling, Y-Scale dialog displays.
2. Enter maximum and minimum value for vertical scale on the graph.
3. Press **OK**. The vertical scale is updated.

Line Style

- **Solid**—Setting the results in a plotted trace that does not display small boxes at the data points
- **Squares**—Setting the results in a plot where small squares are displayed at each data point, and the squares are linked together by lines.

Horizontal Scale

Set the horizontal unit as time scaling (s) or time interval (dt) between frames.

NOTE

Only on a loop including different frame rate.

Smoothing

The system can smooth the traces displayed by applying a filter over a defined time window. The type of filter available is depending on the analysis signal displayed.

1. Select **Smoothing** on the Touch Panel.

NOTE

When smoothing is turned on, it applies to all traces in the plot window.

2. The smoothing filter list displays. Select the appropriate parameter.

NOTE

When smoothing is turned on, it applies to all traces in the plot window.

Merge Graphs

Multiple graphs can be merged, such as early phase and late phase, for 2 TIC graph plots by using “Merge Graphs” on the Touch Panel. This function requires the stored clips which have the Contrast Clock1 information and at least 2 ROIs (yellow and light blue) for TIC analysis. The horizontal scale of merged graph is based on Contrast Clock1.

The horizontal scale of merged graph is based on Contrast Clock1.

Table 13-30 • Merge Graphs Touch Panel Description

Parameter	Description
Exit	Exit to individual TIC analysis mode. All TIC parameters for registered clips are maintained.



Merged Graph Screen Description



Figure 13-38 • Merged Graph Display

- 1 Merged graph window
- 2 Parameter table

Table 13-31 • Merged graph window

Graphic	Description
	Displays the merged graph on the upper part of monitor. The horizontal scale is based on the Contrast Clock1.
	Displays the parameters on the lower part of monitor. The parameters are the intensity of ROI1 (Yellow), ROI2 (Light blue), Subtraction (ROI1-ROI2) and Standard deviation of ROI1 and ROI2. These are calculated for peak and last five second average of first clip, and average of other clips based on Contrast Clock1 value.
	This menu is displays by pressing the left Set key when the cursor is placed over the merged graph window. Connection: selects the graph connection method. Average, Direct and No Line. Vertical Unit: toggles between logarithmic (dB) and linear acoustic units (AU). Vertical unit of previous analysis is retained.

SaveAs (Save image file and export merged trace data)

The graph image file and trace data can be saved.

1. Select SaveAs.
2. The following dialog displays.
 - Location: Select Location to save.
 - Filename: Enter the file name. (Only Text)
 - Export Trace: The system exports trace data to a csv file if this box is checked.
Note: Name of exported file has same name as the image file saved at the same time.
3. Select OK to save the image and data and return to the merged graph screen.

Trace Measurement Gradient

Gradient is displayed on the screen instead of Intensity (db or AU). The gradient calculates from 7 points (includes previous and next frames).




Figure 13-39 • Gradient

TIC Parameters

Following parameters are automatically calculated and displayed with the graph.

Table 13-32 • TIC Parameters

MGrad	Maximum Gradient
MGT	Max Gradient Time
Grad.	Mean Gradient for the peak intensity.
ArT	Arrival Time
TtoP	Time to Peak
PI	Peak Intensity
TWH	Time Width at Half maximum intensity
TWR	Time Width Ratio for wash-in and wash-out

AUC	Area Under the Curve
WiAUC	Wash-in Area Under the Curve
WoAUC	Wash-out Area Under the Curve
A, B, C, k and MSE	The coefficients and the mean square error for fitting curve equation. These parameters are displayed only on Large Data Layout with curve fitting.




Figure 13-40 • TIC Parameters

Show graph (Gradient plot)

1. Select **Gradient Plot** from the pull-down menu.
2. Select the parameter.
 - Off: A graph plots TIC.
 - Gradient: Two graphs plot TIC and TIC gradient.
 - Unit of Y-axis is dB or AU in case of intensity.
 - The unit is $d(\text{dB})/dt$ or $d(\text{AU})/dt$ in case of the intensity gradient.
 - Gradient values for the current frame are displayed in the upper right corner of the graph.
 - Gradient Derivative: Two graphs plot TIC and TIC gradient derivative.
 - The Y-axis units is $d^2(\text{dB})/dt^2$ or $d^2(\text{AU})/dt^2$ in case of the intensity gradient derivative.
 - Gradient derivative values for the current frame are displayed in the upper right corner of the graph.
 - All: Three graphs plot TIC, TIC gradient and TIC gradient derivative.

Curve Fit

1. Select **Curve Fitting** on the Touch Panel.

2. The Curve Fit selection list displays.




Figure 13-41 • Curve Fit Selection List

- **Off**—Remove the fitted curves from the plot and the fit parameters from the display.
- **Wash-in**—Used to find and estimate the local perfusion rate using the contrast agent. Exponential wash-in is described by the function:
 $Y(t) = A(1-\exp(-kt))+B$, where:
 - A (dB or AU) is the intensity from the contrast agent.
 - B (dB or AU) is the intensity at time $t=0$ (defined as the time of the left marker). This corresponds to the tissue (baseline) signal if no contrast is present at the selected starting point.
 $A + B = \text{contrast} + \text{tissue} = \text{plateau level}.$
 - k (1/s) is a time constant.
- **Wash-out**—Used to find and estimate a local wash-out rate. Exponential wash-out is described by the function:
 $Y(t) = A\exp(-kt)+B$, where :
 - A (dB or AU) is the intensity from the contrast agent.
 - B (dB or AU) is the intensity from the tissue = baseline signal.
 $A + B$ is the initial intensity level.
 - k (1/s) is a time constant.
- Gamma variate
 $Y(t) = At^k\exp(-kt)+B$

Parameters of Gamma curve fitting




Figure 13-42 • Gamma Curve

- tc : Increasing function ($C > 0$) for “Wash-in”.
For larger C , the intensity increases quickly before the peak.
- $\exp(-kt)$: Decreasing function ($k > 0$) for “Wash-out”.
For larger k , the intensity decreases quickly after the peak.
- B : Intercept intensity at $t=0$.
- The peak intensity of the curve is affected by all parameters.
Larger A , larger B , larger C , and smaller k make larger peak. The peak time is calculated by C/k .
- MSE: Mean Square Error
If the MSE is small, the difference of actual data and the fitted curve is small.

Set Start/End Frame for Curve Fit per ROI position

1. Generate TIC and perform a Curve Fit. In this state, the Curve Fit graph is drawn from Cine Start Frame to Cine End Frame for all the ROIs.
2. Push Auto Calc Range on Touch Panel and select All Samples. Then the start and end frames for each ROI are automatically estimated.
If you need to optimize the start and/or end frames,
3. Select the start frame as with Cine or move the cursor to the desired position on the graph and press the right **Set** key.
4. Move the arrow pointer on the ROI (hand cursor appears) and press the left **Set** key. The system menu appears.
5. Select **Set Start Frame** from the menu.
6. Select the End Frame as with Cine or move the cursor to the desired position on the graph and select the right **Set** key.
7. Move the arrow pointer on the ROI (hand cursor appears) and select the unmarked key (the left Set key). The system menu appears.
8. Select **Set End Frame** from the menu. The ROI colored line displays.

9. Repeat the above procedures as necessary. The system retains the start/end frame per ROI while TIC is active. Once the TIC menu is closed, the settings are lost.

Display/Hide Calculation Values

You can select the TIC parameters for Small Data Layout and Everything Layout.

1. Place the cursor on the analysis graph and press left Set key. Select TIC Parameters from the pull-down menu.
2. The TIC Parameters dialog appears.
3. Select a maximum of 6 parameters to display for each TIC.
 - Save As Default : saves as a system preset.
 - Save : saves as temporary.
 - Cancel

NOTE

If you select more than 6 parameters and select Save or Save as default, you will be prompted to select up to 6 parameters.

4. The selected parameter displays for Small Data Layout and Everything Layout. Press the Large Data Layout button to display all parameters; or press the other Layout button to hide all parameters.



Figure 13-43 • Curve Fitting Parameters Dialog

Raw Data store with TIC Setting

You can store the raw data with TIC setting.

1. Run TIC cineloop.

2. During cine running mode, press appropriate print key.

When you recall the raw data clip with TIC data, you can add, delete or modify the analysis measurements on the recalled raw data in TIC mode.

Printing TIC Data

1. Press **Run/Stop** to freeze the image.
The still image can be get when cineloop is stopped by Run/Stop button.
2. The system captures a single still frame which consists of the plot, the reference image and user annotation.

Annotating the TIC Data

The user can annotate both the reference image and the trace plot displays. Use **Comment** key to type the annotation. See Chapter 6 for reference.

SaveAs (Save image file and export trace data)

You can save the image file and trace data.

1. Select **SaveAs**.
2. The following dialog displays.
 - Location: Select Location which to save.
 - Filename: Enter the file name. (Only Text)
 - Export Trace: If you check this box, the system exports trace data to a csv file.

Note: Name of exported file has same name as the image file saved at the same time.
3. Select **OK** to save the image and data and return to the TIC Analysis screen.
 - All displayed ROI traces are saved in the exported file.
 - The fit parameters are included in the trace file if the user has done a curve fit.

NOTE

The Smoothed trace is the one saved if the user has applied a smoothing filter.

NOTE

Only data from the user selected image range is included in the exported trace file.

NOTE

Data for disabled frames are not included in the exported trace file.

NOTE

No trace results are saved in the standard image database.

NOTE

Trace results are not shown on the Worksheet.

TIC cineloop store

When you transfer TIC cineloop to the external server, check “Add Multiframe Data” in **Utility** > **Connectivity** > **Button** and then store the cineloop to the system.



Figure 13-44 • Button page

Parametric Imaging**Overview**

The Parametric Imaging is the application to mainly visualize the contrast agent "arrival time" in colors by processing cine loops with pre-defined threshold value for each pixel.

Parametric Imaging is available for cineloop of B-Mode (includes CHI mode), Contrast mode and B-Flow mode.




Figure 13-45 • Parametric Imaging - Example

NOTE

Parametric Imaging mode is available only if the system has a valid option key.

Typical Workflow

Here is the typical workflow for Parametric Imaging:

1. Enter Contrast Mode.
2. Inject the contrast agent.
3. Store the Contrast images until the contrast agent reaches the anatomy of interest in a very early phase.
4. Activate Cine mode and select the start frame/end frame of calculation if needed.
5. Select background image type as needed.
6. Press **Parametric Image** on the CINE Touch Panel.
Calculate and color encode according to the arrival time to CINE image with motion correction.




Figure 13-46 • Calculating

1. Cine playback stops automatically when the calculation is done. You can playback/stop using Run button as normal cineloop.

NOTE

You can cancel in the process of the calculation.

2. Measure the arrival time and time difference.
3. Save the data.
4. Unfreeze the image.

Background selection

You can select background image depending on the display setting of the contrast image.

- Gray map, Tint map: Black/White Contrast image
- Hybrid visualization: Black/White Reference Image

Table 13-33 • Background selection - Gray map/Tint map





	Gray map	Tint map
		
	Capture On	Capture Off
Parametric Image Map On		
Parametric Image Map Off		
	Background: Contrast image	

Table 13-34 • Background selection - Hybrid Visualization

	Hybrid visualization	
		
	Capture On	Capture Off
Parametric Image Map On		
Parametric Image Map Off		
	Background: Reference Image	

Configuration parameters

You can adjust the following parameters. Press **Parametric Property** on the Cine Touch Panel.

- Color bars
- Transparent
- Gradation
- Capture
- Threshold
- Motion correction method
- Preset Load/Save



Figure 13-47 • Parametric Property window

Color bar

Select the colors for parametric imaging, the numbers of color and the timing for color change.



Figure 13-48 • Color bar window

Adjustment and change color bar

Click on the color bar and select for

- Change color: Click **Change Color**. The color window displays. Select appropriate color and click **OK**.
- Change length of each color: Drag the right border line left or right to change the length.
- Entire length of color bar (length of most right color): Click left or right triangle icon under the color bar to change the length.
- Delete color: Click **Delete**.
Add color bar

Click **Add** and select color.

Save As

1. Click **Save As**. The Save As preset window displays.
2. Click on the color bar which you want save to select.
3. Press **Save**. You can save up to 8 color bars with:
 - Color bars
 - Transparent
 - Gradation
 - Capture

Advanced Features

- Threshold level
- Motion correction method




Figure 13-49 • Save As window

Load the parameters

1. Click **Load...**
2. Select appropriate map and click **Load** on the Load preset window.
3. After you change map, you have to click **Recalculate** to recalculate the arrival time.

NOTE

If you want reload factory default setting, click **Reload Factory Defaults** on the Load preset window.



Figure 13-50 • Load Preset window

Other parameters



Figure 13-51 • Transparent, Gradation, Capture



Figure 13-52 • Threshold, Transparent %, Motion Correction

Transparent

Click the checkbox next to Transparent to bring out the contrast image behind the parametric image. You can change the percentage of transparency.

1. Click More on the Parametric property window.
2. Type the percentage. The higher the number, the more transparent the background contrast image.

NOTE

You can adjust the percentage by CF Volume key.

Gradation

Check the checkbox next to Gradation to smooth the color gradient of adjacent colors.

Advanced Features

Capture	Check the checkbox next to Capture to use the captured image of background image.
Threshold	Type threshold level (0 - 100%) to decide the arrival time.
Motion correction	Select Motion Correction On/Off and correction method. • Off: No motion correction • On: Motion correction on • With rotation: X-Y direction with rotation • No rotation: X-Y direction without rotation • X Only: X direction only

Table 13-35 • Transparent, Gradation, Capture





	• Transparent On • Gradient On • Capture On	
Transparent Off	Gradient Off	Capture Off
		

Table 13-36 • Background image

Capture On	Capture Off
	

Trackball key for Parametric Imaging

Press Upper Trackball key to select Parametric Imaging Control. Parametric Imaging controls are assigned to each Trackball key.




Figure 13-53 • Trackball key map

Map On/Off

Display the Parametric Image On/Off. If you select Map Off, the background image displays temporarily.





Figure 13-54 • Map On/Off

Adjust start



Press **Adjust Start** to change zero point where you place the cursor.

Figure 13-55 • Adjust start

Advanced Features

NOTE

Points before new zero point, change the color same as zero point.

NOTE

The cursor displays on the reference image, too.

Reset start


Press **Reset Start** to reset zero point to the start frame of the cineloop.

Measurement

Arrival time

1. Press **Measure** and move the cursor on the image. The pointer changes to green cross type.
2. The arrival time of the cursor position displays in the measurement result window.

Measurement on the parametric image



Arrival time and Standard deviation

You can calculate the average arrival time and standard deviation of the traced area.




Figure 13-56 • Trace measurement

- Arrival time
 - 2D point caliper
- Arrival time and Standard deviation
 - 2D ellipse, 2D 3 point ellipse, 2D circle, 2D area trace, 2D spline trace
- Arrival time and Standard deviation for dual view
 - 2D dual area trace, 2D dual ellipse, 2D dual spline trace, 2D dual circle

NOTE

You can create measurement preset (Dual trace, etc.) in **Utility > Measure > M&A** tab. See *Adding a Measurement* on page 328 for more information.

Strain Elastography

Description

Strain Elastography shows the spatial distribution of tissue elasticity properties in a region of interest by estimating the strain before and after tissue distortion caused by external or internal forces. The strain estimation is filtered and scaled to provide a smooth presentation when displayed.

Below is an example of Strain Elastography. The image is displayed in dual mode with the color map/bar of the Strain Elastography on the left side and the imaging parameters on the right side of the display below E.

You activate Elastography via the **Elasto** hard key on the Control Panel.




Figure 13-57 • Strain Elastography Example

Using Strain Elastography

The Strain Elastography image is achieved by pulsating the probe while you are scanning the anatomy of interest. Here are some criteria to use:

Handheld elasticity imaging can be very dynamic as the size of distortion depends on the movement of the hand-held probe. To maintain stable and consistent displayed strains, pay attention to the Quality graph. Two forms of feedback are provided. In either form, an ideal manual compression is indicated by a high value feedback. In addition, apply the following post-processing controls: Smoothing, Window, Scaling, and Frame Averaging.



Strain Elastography displays firmer tissue in blue and softer tissue in red. To enhance Blue, increase Hard Compress; to enhance Red, increase Soft Compress on the Touch Panel. To enhance strain elastography contrast, reselect the Color Map.

If you need more resolution, reduce Smoothing, increase Frequency, or reduce Window.

If you need a smoother image, increase Window or Smoothing.

If the images seem too flashy, decrease Frame Reject to 1.0 and Noise Reject to have consistent imaging throughout.

Table 13-37 • Strain Elastography

Using Strain Elastography	
Manual Compression: 1. Press “Elasto” button at the console to activate. 2. Select Strain on the Touch Panel. 3. Adjust the position of the ROI to place the suspicious area at the center. 4. Adjust the size to include surrounding tissue (sample area size = x3 dimension of the lesion per axis). Manual compression depends on the type of probe. 1. Linear probes: Perform slight compressions keeping transducer perpendicular to the skin. Duration: 5 sec or 10 compressions. 2. Convex probes: Turn the patient on his left side more than 90 deg. Pressing with the probe above the lesion, allowing the heart and lungs to create the compressions. 3. Endocavitary probes: Perform soft, angular movement in plane of the probe. Duration: 5 sec or 10 compressions. NOTE Any very soft (vessel, cyst, air) or very stiff (bone) tissue above the lesion or the reference area may cause interference with the compressions. You may want to attempt from a different view. NOTE Keep the lesion within the image and watch the quality graph for consistent high nearly flat (plateau-like) peaks.	Select the frame to analyze Press Freeze. 1. Using the trackball or “frame by frame” knob, select a frame on a plateau of the quality graph (Image 1, A) or when consistent frames with green bars are visualized (Image 1, B).  Image 1: Elastography image with quality graph (A), quality bar (C) and elasto color bar (B). The relation between elastography colors and area's stiffness is given at the elastography color bar (Image 1, B).  The color bar indicates stiffness levels: S (Less Stiff) at the top, transitioning through yellow, green, and blue to H (Stiffer) at the bottom. Arrows point from the text 'Less Stiff' and 'Stiffer' to the corresponding ends of the color bar.

Clinical Applications

Table 13-38 • Clinical Applications

Application	Compare the suspicious lesion with:	Probe
Breast	A fatty area in the breast or the average rounding breast tissue.	• L2-9-D/L2-9VN-D • ML6-15-D • L3-12-D • L6-24-D
Thyroid	A normal area in the parenchyma of the gland.	• L2-9-D/L2-9VN-D • ML6-15-D • L3-12-D • L8-18i-D • L6-24-D
MSK	A normal area in the same part of the body.	• L2-9-D/L2-9VN-D • L8-18i-D • L3-12-D • L6-24-D
Liver	A normal area in the parenchyma of the liver.	• C1-6-D/C1-6VN-D • C2-9-D/C2-9VN-D • L2-9-D/L2-9VN-D
Prostate	A normal area in the parenchyma of the prostate.	• IC5-9-D • BE9CS-D
Uterus	A normal area in the parenchyma of the uterus.	• IC5-9-D

Strain Elastography Controls

Table 13-39 • Strain Elastography Touch Panel Description

Parameter	Description
Axial Smoothing	Controls the smoothness of the strain elastography image in the axial direction. A higher value means a smoother image.
Lateral Smoothing	Controls the smoothness of the strain elastography image in the lateral direction. A higher value means a smoother image.
Window	Controls the RF data segment size for the motion tracking. A higher Window value gives a better signal to noise ratio (SNR) at the cost of axial resolution.

Parameter	Description
Map	Controls the strain elastography maps. Seven different maps are available with various contrast and color schemes, including a grayscale map. Selections: E0-7, E-Gray, and S Map. The E Map calculates the mean strain for the whole ROI and assigns this mean value to a green (center) color. This is best suited for imaging a localized mass compared to the surrounding tissue with external forces (the movement of the hand-held probe). The S Map is useful for imaging diffusely-distributed diseases and local strain changes brought about by internal forces such as the movement of a beating heart and moving vessels. The default strain sensitivity is unique by probe/application. You can adjust the strain scale by adjusting the strain sensitivity tool.
Frame Average	Controls the persistence of the strain elastography images.
Frequency	Controls the transmit frequency.
Strain Sensitivity	If S0 Map (in live) is selected, the user can control Strain Sensitivity (SS).
Soft Compress	Individually controls the image enhancement for the softer than average tissues.
Hard Compress	Individually controls the image enhancement for the harder than average tissues.
Scale	Controls the time interval between consecutive firings. A lower value dictates a higher sensitivity to weak manual motion.
Transparency	High values bring out the tissue behind the strain elastography data. You adjust via the Color Gain control; this imaging parameter appears as a “T” on the right-hand portion of the display.
Biopsy Kit	Biopsy Kit.
Frame Reject	Controls how many frames get rejected due to low quality vertical motion. A higher value means more frames get rejected. A rejected frame has a completely transparent ROI with the B-Mode background showing through.
Noise Reject	Controls how many frames get rejected due to lateral and elevational motion. A higher value means more frames get rejected. A rejected frame has a completely transparent ROI with the B-Mode background showing through.
Line Density	Optimizes B-Mode frame rate or spatial resolution for the best possible image.
Show Quality Graph (restart needed)	Select to display a Quality Graph for Elastography. The higher the level, the higher the data quality for the frames.

Application Parameters

You can set the Quality Bar and Quality Graph on the **Utility > Application > Settings > Elasto** page.

For Quality Bar:

- Check to display a Quality Bar for Elastography.

Advanced Features

For Quality Graph:

- Off (No elasto quality graph is displayed in the image)
- Small, Medium or Large for display size of Quality Graph.

General Imaging Parameters by Application/Probe/Feature

Select the Default Elasto Mode via **Utility > Imaging > General**.

You can specify Strain to be the default setting by application and probe.

1. To specify a default probe per application, select **Utility > Imaging > General**.
2. Select the application.
3. Select the default probe from the pull-down menu.
4. Specify the Default Elasto setting:
 - Strain
 - Shear

Elastography Analysis

Overview

The Elastography Mode detects strains by correlating the echo amplitudes of the tissue when compressed and uncompressed. Different displacement of echoes is an indicator for different stiffness (strain) of the tissue. High strain means that the tissue is softer, low strain means that it is stiffer. Zero is absolutely stiff without any elasticity. Elastography Analysis is a strain ratio comparative tool that enables users to compare the strain of one tissue to the surrounding tissue.

Using Elastography Analysis

1. If not yet in Elastography Mode press **Elasto**.
2. Perform the scan. Proper manual compression/decompression is indicated by a fully green quality bar.
3. Press **Freeze** and scroll the Trackball to bring up the CINE tab on the Touch Panel.
4. Press **Analysis** on the Touch Panel. The Elastography Analysis Touch Panel appears and the monitor screen shows the Elastography Analysis display. The valid Elastography frames are marked green.
5. Adjust the start and end of the CINE Loop within the green frames using the rotary buttons below the Touch Panel (Start Frame/End Frame). Press Set to confirm. .
6. Activate the cursor and move it over the Elastography image on the top left side of the screen. A white sample area appears. By default this sample area is a circle. It will be the reference sample area and should be placed in the normal breast tissue.
7. Position this reference sample area and press **Set**. A yellow plot curve displays the strains over time on the right side of the monitor screen.
8. Move the Trackball again. A new sample area appears (sample area 1) which should be placed in the lesion.
9. Position the sample area and press **Set**. A second plot curve is displayed (blue curve).





Figure 13-58 • Elastography Analysis Example

In total you can create 7 sample areas and 1 reference area. Each sample area can be edited, moved, copied, or deleted. A sample area can also be drawn manually.

1. To edit an existing sample area, move the cursor over the sample area then press the left Trackball control (Menu) to bring up the menu. Select "Set Sample Area Shape" to enter a dialogue window where Height and Width can be adjusted. You can also label, delete, copy, move the sample area by selecting corresponding options in the menu.
2. To draw a sample area manually, select the pencil icon. Then the cursor becomes a cross inside the strain image. Start to draw the sample area by pressing **Set**. Press **Set** again to stop drawing.



3. Both Stiffness and Ratio plots can be displayed: Switch to Ratio plots by pressing the **Ratio** control on the Touch Panel.
4. Press **Exit Analysis** to return to Elastography Mode.

**HINTS**

Use the trackball to scroll the Cine Loop quickly.

NOTE

The maximum strain value in human tissue can be up to 2%.

NOTE

The ratio value indicates how many times the tissue of a sample area is harder or softer than the tissue of the Reference sample area.

Additional Notes for Elastography Analysis

**HINTS**


Limitations:

- This is a relative quantification tool based on freehand manual palpation technology. It cannot show the stiffness by the kPa (kilopascal).
- There is no compatibility among manufacturers regarding the value. It depends on their strain imaging technology and definition of the value.
- Colors indicate degree of stiffness and do not directly correlate to a specific tissue type. Interpretation of what the tissues are and how to apply these ratios clinically is at the discretion of the user.
- Elastography physics dictates that cystic structures will be displayed with a three-layer pattern. This three-layer pattern will start with blue on the factory default map (which corresponds to hard), then progresses to green and then to red (which corresponds to soft). The posterior displacement of elastography patterns also may cause the B-Mode cyst to consist primarily of blue with the green to red being posterior to the B-Mode cyst. You need to be aware of the three-layer pattern of a cyst in elastography. Utilizing Elastography Analysis and setting the sample area in the blue portion of the three-layer pattern on the cyst and then setting the sample area in the “normal” tissue may cause you to misinterpret the Elastography Analysis ratio as the cyst to be hard as compared to the “normal” tissue.

Shear Wave Elastography

Overview

Shear wave elastography on the LOGIQ Fortis is an ultrasound imaging mode in which shear waves are generated in-vivo acoustically via the imaging ultrasound transducer. The motion of the shear waves is then tracked using ultrasound to determine their velocity of propagation, which is a quantifiable indicator of the mechanical properties of the tissue through which it traveled. The steps associated with performing this analysis on the LOGIQ Fortis include correctly placing a user-specific region of interest (ROI) over the anatomy of interest. Next, the user activates the shear wave analysis mode where the shear wave generation and tracking occurs. Up to four (4) sites may be analyzed. After acquiring the data, the user either stores the image or analyzes it via measurement tools which can produce shear wave velocity or stiffness statistics of areas within the ROI.



1. Excitation and imaging transducer
2. Tissue
3. Propagating Wave
4. Focus

Figure 13-59 • Shear wave Probe Characteristics

Intended Uses

Shear wave elastography Intended Uses are:

- Abdomen
- Small Parts
- Musculoskeletal
- Gynecological

- Urological

**CAUTION**

Shear wave elastography IS NOT intended for use for Obstetrical exams.

Configuring Shear Wave

To configure Shear Wave parameters, you need to adjust the following Utility pages:

- ELASTO Imaging Parameters (**Utility > Imaging > Elasto**)
- General Imaging Parameters (**Utility > Imaging > General**)
- System Imaging Parameters (**Utility > System > System Imaging**)
- Measurement Parameters (**Utility > Measure > Advanced > for both Abdomen and Small Parts Breast**)

ELASTO Imaging Parameters

To configure Shear Wave Elastography settings, select **Utility > Imaging > ELASTO**.

- Push Output (%) — The acoustic output of the shear wave push
- Track Output (%) — The acoustic output of the shear wave tracking pulse
- Transparency — The transparency of the shear wave image overlay
- Gain — Gain, as can be manipulated by the CF knob
- Width — ROI Width (values vary by probe)
- Vertical Size (cm) — ROI Vertical Height (values vary by probe)
- Center Depth (cm) — ROI Center Depth (values vary by probe)
- Color Map — Select shear wave Map 0 or shear wave Map 1 (SW0 or SW1) plus other ELASTO Maps (E1, E2, E3, E4, E-GRAY)
- Enter ELASTO DualView by Default — Check to select entering shear wave in DualView.

Shear wave Color Bars:




Figure 13-60 • Shear wave Maps

Advanced Features

The illustration shows the color maps for kiloPascals and velocity (meters per second). The images look the same, regardless of the unit (m/s or kPa). You can preset which color on the map represents stiffness (Red as Hard or Blue as Hard).

General Imaging Parameters by Application/Probe/Feature

Select the Default Elasto Mode via **Utility > Imaging > General**.

You can specify shear wave to be the default setting by application by probe.

1. To specify a default probe per application, select **Utility > Imaging > General**.
2. Select the application.
3. Select the default probe from the pull-down menu.
4. Specify the Default Elasto setting:
 - Shear
 - Strain

System Imaging Parameters

You can set the shear wave Display Units and specify which color on the shear wave map represents stiffness on the **Utility > System > System Display > Display** page.

Set the Shear Elasto Display Units to either of the following settings:

- m/s (meters per second)
- kPa (kiloPascals)

Specify which color on the shear wave map represents stiffness:

- Red as Hard
- Blue as Hard

Measurement Parameters

On the **Utility > Measure > Advanced > Abdominal** page, set the following:

- Shear Measure Size — Sets the Default Diameter size of the shear wave measurement circle
- Shear Measure Fixed Size — Sets to ON to move the shear wave measurement circle with keeping the fixed size.
- Shear Calculation Method — Specify Mean (Mean averages of all of the shear wave points within the measurement circle) or Median (Median sorts, then selects the middle point of all points within the measurement circle)
- Shear Units Determine Folder — When On is specified, the unit specified pre-selects the measurement folder. If m/s is specified as the unit, then the Velocity folder is used; if kPa is specified as the unit, then the Stiffness folder is used

On the **Utility > Measure > Advanced > Small Parts** page, specify the Calculation Method (Mean/Median).

- Shear Measure Size - Sets the Default Diameter size of the shear wave measurement circle.

- Shear Measure Fixed Size - Sets to ON to move the shear wave measurement circle with keeping the fixed size.
- Shear Calculation Method — Specify Mean (Mean averages all of the shear wave points within the measurement circle) or Median (Median sorts, then selects the middle point of all points within the measurement circle)

Activating Shear Wave

To activate shear wave, press **ELASTO** and Touch Panel Shear controls. (1) To position and size the shear wave ROI, or to Start or return to Pre-Mode shear wave, touch the right-most Trackball control. (2) To save the Shear Elasto image, activate the P1 control.



Figure 13-61 • Shear wave Operator Panel Controls

ELASTO -- When you activate ELASTO, the default Elasto Mode appears on the display and Touch Panel. If shear wave is preset as the default Elastography Mode, then Shear Wave Elastography mode appears; if Strain Elastography is set, then Strain Elastography mode appears.

Shear Wave Display

Shear wave displays as follows while performing a measurement:



Figure 13-62 • Shear wave Display with Measurement

1 Measurement Window

2 Shear wave Map

3 Measurement(s)

4 Shear wave ROI

5 Trackball Controls

6 Imaging Parameters

E = Elastography, Shear Wave

Gn = Gain

T = Transparency

SVD = Sample Volume Depth

PO% = Push Output Percentage

TO% = Track Output Percentage

f50-200Hz = Shear Wave Frequency range

Gen/Pen = General/Penetration.

“Gen” appears when viewing an image where general settings were applied.

“Pen” appears when viewing an image optimized for penetration.

“Clock” displays in Low Frame Rate Use cases.

NOTE

The measurements' Median and Inter-Quartile Range (IQR) are displayed by default in the Measurement Window and Worksheet. The Caliper Area, Depth of Caliper Center, and average Quality within the measurement area can also be displayed on the Worksheet, along with how many measurements used for the median calculation.

Selecting Measurements to Display

The following measurements are displayed by default, except those marked with an asterisk (*):

- Shear Stiffness (kPa)
- Standard Deviation within the Shear Stiffness Measurement Area*

- Shear Velocity (m/s)
- Standard Deviation within the Shear Velocity Measurement Area*
- Measurement Area (cm2)*
- Depth of Caliper Center (cm)*
- Average Quality within the measurement area (%)*

*Items not displayed by default can be added via **Utility > Measure**. See below.




Figure 13-63 • Configure Shear Wave Measurements

Checkmark the Shear Wave measurements to be displayed on the monitor and on the Extended Worksheet.



Figure 13-64 • Selecting Default Shear Wave Measurements



Figure 13-65 • Shear Wave Measurement Results Window

Shear (Elasto) [Site 1]							
Med	Stiffness		Velocity				
	2.26 kPa		0.87 m/s				
IQR	0.08 kPa		0.03 m/s				
IQR/Med	3.4 %		2.7 %				
N	6		6				
	Stiffness	Stiffness SD	Velocity	Velocity SD	Area	Depth	Quality
1	2.34 kPa	0.23 kPa	0.88 m/s	0.26 m/s	0.96 cm2	4.86 cm	3 %
2	2.26 kPa	0.21 kPa	0.87 m/s	0.27 m/s	1.04 cm2	4.60 cm	3 %
3	1.98 kPa	0.12 kPa	0.83 m/s	0.20 m/s	1.03 cm2	6.15 cm	3 %
4	2.24 kPa	0.18 kPa	0.86 m/s	0.24 m/s	2.04 cm2	5.82 cm	3 %
5	2.40 kPa	0.23 kPa	0.91 m/s	0.27 m/s	0.72 cm2	4.85 cm	4 %
6	2.25 kPa	0.25 kPa	0.87 m/s	0.29 m/s	1.47 cm2	4.35 cm	3 %
7							
8							

Figure 13-66 • Shear Wave Extended Worksheet

Shear Wave Quality Indicator

The Shear Wave Quality Indicator displays the reliability of the Shear Wave image. A Quality value is computed for every pixel of the Shear Wave image and displayed as an image in a color-coded ROI. Locations with higher quality values have more reliable Shear Wave data.

To View the Shear Wave Quality Image

The Quality Image can only be displayed in Dual Mode. It can be turned on and off using the Quality button on the Shear Elasto page on the Touch Panel.




Figure 13-67 • Quality Image Button

NOTE

You must be in Elasto Dual Mode to select the Quality button.

When the Quality Image display is turned on, the Quality Image is displayed on the left and the Shear Elasto image on the right.




Figure 13-68 • Shear Wave Display with Quality Image Enabled

1Color Bar for Quality Image

2Quality Image

The Color Bar on the left of the Quality Image shows the quality value corresponding to the colors in the Quality Image. Low quality is indicated by the colors at the bottom of the bar, and high quality by colors at the top.

Quality Map

There are two Quality Maps available to choose from on the Shear Elasto screen (under Quality Map), each with a specific color range:

- SWQ0 - Black/red/orange/yellow/white (low to high quality)
- SWQ1 - Red/orange/yellow/green (low to high quality)

Elements of a good Shear Wave image

Elements of a good shear wave image contain the following:

- Good contact
- ROI placed in the middle of image
- Gain at or near factory default
- Uniform color fill-in
- Homogeneous color pattern

Shear Wave Touch Panel

Table 13-40 • Shear wave elastography Touch Panel Parameters

Preset Parameter	Description
Strain or Shear Elasto	Toggle touch key between Strain and shear wave elastography.
Measurement Bias Information	Press to bring up measurement bias information.
Quality	Press the Quality button to open the Quality Indicator screen (while in Dual Mode).
Phantom	Press the phantom button to get good performance when measuring stiff, motion-free phantoms. The 049A phantom is recommended. For more information, refer to the CIRS website at http://www.cirsinc.com/products/all/74/elasticity-qa-phantoms/?details=specs .
Penetration	Press Penetration for instances measuring especially hard tissue (fibrotic or cirrhotic livers or technically difficult tissue in general).
Map	Shear wave Color Map.
Push Output	The acoustic output of the shear wave push.
Track Output	The acoustic output of the shear wave tracking pulse.
Range	Min/Max Velocity, or Stiffness displayed.
Transparency	The transparency of the shear wave image overlay.

Measurement Bias Information Tables

The Measurement Bias Information tables display the bias and precision percentage at different spatial resolutions (Bias/Precision vs Object Size, in millimeters) and at incremental depths (Bias/Precision vs Depth, in centimeters) for each shear wave probe.



Figure 13-69 • C1-6-D/C1-6VN-D Measurement Bias Information (Velocity on the left and Stiffness on the right)



Figure 13-70 • IC5-9-D Measurement Bias Information (Velocity on the left and Stiffness on the right)



Figure 13-71 • L2-9-D/L2-9VN-D Measurement Bias Information (Velocity on the left and Stiffness on the right)



Figure 13-72 • L3-12-D/L3-12VN-D Measurement Bias Information (Velocity on the left and Stiffness on the right)



Figure 13-73 • L8-18-D Measurement Bias Information (Velocity on the left and Stiffness on the right)



Figure 13-74 • ML6-15-D Measurement Bias Information (Velocity on the left and Stiffness on the right)



Figure 13-75 • L6-24-D Measurement Bias Information (Velocity on the left and Stiffness on the right)

Speed Ranges and Accuracy are shown at the top of the Bias/Precision vs Object Size tables.


Using Shear Wave (SW)

There are three shear wave states:

- Pre-shear wave acquisition
Pre-shear wave acquisition is an intermediate mode between B-Mode and shear wave acquisition. During Pre-mode, the previous B-Mode imaging mode is still active.
NOTE
You cannot store images in Pre-Mode.
- Shear wave acquisition
Press the **Start** set key initiates SW acquisition.
- Post-shear wave acquisition (Freeze)
The system displays the acquired SW image and background B-Mode image. The User can cycle through the acquired frames, perform measurements, and annotate the image.

SW images can be stored in both raw data and DICOM format.

Table 13-41 • Using Shear wave



Typical Exam Protocol (Liver)

A recommended shear wave elastography protocol to scan the liver is,

1. Image the right lobe of the liver intercostally in Pre-shear wave Mode.
2. Position and size the ROI.



HINTS

The smaller the ROI, the faster the Frame Rate.

3. Instruct the patient that during the exam they can mostly breathe normally. However, advise the patient that they will need to suspend their breathing mid breath so that you can obtain an optimum image while performing the scan.
4. Adjust ROI as needed, avoiding vessels and fluid-filled structures. It is advisable to avoid rib shadows whenever possible.
5. Start the shear wave acquisition.
6. Freeze the image when desired frame is obtained.
7. Perform the measurement. The system prompts you through the measurement.
8. The system will auto sequence the measurements and walk you through all of the measurements (Abdominal Application Preset), if preset.

NOTE

For Breast measurements, users typically take single measurements to measure the lesion once. Or, users can perform a ratio of two different tissues (one of the lesion and non-lesion tissue).

9. Repeat steps 3 through 9 for the remaining samples.
10. Typically, users obtain ten (10) samples.
11. Once measurements are complete, you can elect to add another site to the exam. To add another site, press **Add Site** on the Touch Panel. Select **Enter Site Name** to add the name for the new site.



Figure 13-76 • Add Additional Shear Wave Site and Enter Site Name

NOTE

If you no longer need the added site, you can select to delete the site via **Delete Site** on the Touch Panel.

12. Once measurements for all sites are complete, select Summary on the Touch Panel to view saved measurements.

**CAUTION**

Excessive manual compression of the underlying tissue with the probe can lead to biased shear wave measurements.

Shear Wave Measurements

The higher the velocity, the stiffer the tissue.

Types of Shear Wave Measurements

There are two types of measurement units for quantifying stiffness:

- Velocity (meters per second)
- Stiffness (kiloPascals)

Measurement used to quantify stiffness: Shear wave imaging measures the velocity of shear waves generated by acoustic radiation force impulse in tissue. The velocity, in units of meters per second (m/s), can be converted to Young's Modulus (stiffness), in units of kiloPascals (kPa), under simplifying assumptions. Velocity or stiffness can be used to quantify the local tissue elasticity.

NOTE

Acoustic radiation force is generated by a transfer of momentum from an acoustic wave to the medium through which it is propagating, caused by absorption and scattering in soft tissue. Impulsive application of focused acoustic beams in tissue can generate shear waves which propagate away from the focal region of the beam.

**CAUTION**

Tissue inhomogeneities and other factors may bias shear wave measurements.

Measurement Analysis

To take shear wave measurements, typically users get ten (10) samples.

1. After you have acquired the desired image, avoiding vessels and fluid-filled structures, press Freeze.
2. Press Measure. Select the measurement (Stiffness or Velocity) unless preset to "Shear Units Determine Folder" is turned on.

NOTE

This step only needs to be done in the Breast Small Parts Application Preset and is not needed in Abdomen.

3. Perform the measurement. The system prompts you through the measurement.

- a. Position the first caliper at the desired location on the ROI. Place the first control point and press Set on the Trackball control.

An ROI ellipse appears and a second caliper.

- b. Adjust the size of the ROI. Adjust ellipse and place the last control point.

NOTE

This step only needs to be done in the Breast Small Parts Application Preset and is not needed in Abdomen.

4. Repeat the scan and measurement. The system will auto sequence the measurements and walk you through all of the measurements, if preset.
5. Measurements are transferred to the Worksheet.

NOTE

The measurements' Median and Inter-Quartile Range (IQR) are displayed by default in the Measurement Window and Worksheet. The Caliper Area, Depth of Caliper Center, and average Quality Percent within the caliper, and Standard Deviation within the caliper can also be displayed by going to **Utility > Measure**, as shown in *Figure 13-63* • *Configure Shear Wave Measurements* on page 829.

Shear Wave Worksheets

There are two types of Worksheets (Overall and Extended).

Shear Elasto /UGAP [Site1]				
Name	Stiffness Site1	Velocity Site1	Attenuation Coefficient Site1	Attenuation Rate Site1
Med	2.26 kPa	0.87 m/s	0.50 dB/cm/MHz	
IQR	0.08 kPa	0.01 m/s	0.02 dB/cm/MHz	
IQR/Med	3.4 %	1.7 %	3.7 %	
N	6	6	5	
1	2.34 kPa	0.88 m/s	0.49 dB/cm/MHz	
2	2.26 kPa	0.87 m/s	0.50 dB/cm/MHz	
3	1.98 kPa	0.81 m/s	0.51 dB/cm/MHz	
4	2.24 kPa	0.86 m/s	0.51 dB/cm/MHz	
5	2.49 kPa	0.91 m/s	0.49 dB/cm/MHz	
6	2.25 kPa	0.87 m/s		
7				
8				

Figure 13-77 • Shear Wave Overall Worksheet

Shear Elastic (Site 1)							
	Stiffness	Velocity					
Med	2.26 kPa	0.87 m/s					
IQR	0.08 kPa	0.03 m/s					
IQR/Med	3.6 %	3.7 %					
N	6	6					
	Stiffness	Stiffness SD	Velocity	Velocity SD	Area	Depth	Quality
1	2.34 kPa	0.23 kPa	0.88 m/s	0.28 m/s	0.96 cm ²	6.86 cm	3 %
2	2.26 kPa	0.21 kPa	0.87 m/s	0.27 m/s	1.04 cm ²	6.40 cm	3 %
3	1.98 kPa	0.12 kPa	0.83 m/s	0.20 m/s	1.03 cm ²	6.25 cm	3 %
4	2.26 kPa	0.18 kPa	0.86 m/s	0.24 m/s	2.04 cm ²	5.82 cm	3 %
5	2.49 kPa	0.23 kPa	0.91 m/s	0.27 m/s	0.72 cm ²	6.85 cm	4 %
6	2.25 kPa	0.25 kPa	0.87 m/s	0.29 m/s	1.47 cm ²	6.35 cm	3 %
7							
8							

Figure 13-78 • Shear Wave Extended Worksheet

Deleting or Excluding Measurements from the Worksheet

Measurements can be deleted from the Extended Worksheet. Once a measurement has been deleted, it cannot be re-added. Deleted measurements show as a blank.

NOTE

You can also choose to delete all Shear Wave measurements, without deleting non-Shear Wave measurements from the worksheet.

Measurements can also be excluded from the Extended Worksheet -- and added back in at a later time. Excluded measurements show as a blank.

To delete/exclude a measurement,

- 1. Highlight the measurement you want to delete/exclude.
- 2. Right click on the Extended Worksheet to bring up the Delete/Excluded pop-up menu.



- 3. Select the appropriate action.

In the example below, measurements from line 4 with the blue rectangle have been excluded (and can be added back in); and measurements from line 2 have been deleted (and cannot be added back in).

Shear Wave Graph (Shear)							
Stiffness				Velocity			
Med	2.41 kPa			0.50 m/s			
IQB	0.39 kPa			0.07 m/s			
IQBMed	16.3 %			6.0 %			
Cor	3			3			
	Stiffness	Stiffness SD	Velocity	Velocity SD	Area	Depth	Quality
1	2.30 kPa	0.38 kPa	0.56 m/s	0.38 m/s	1.25 cm ²	4.66 cm	31.6%
2	2.02 kPa	0.15 kPa	0.44 m/s	0.21 m/s	1.25 cm ²	3.29 cm	15.1%
3	2.41 kPa	0.21 kPa	0.49 m/s	0.27 m/s	1.25 cm ²	5.58 cm	27.1%
4							
5							
6							
7							
8							
9							
10							
11							
12							

Figure 13-79 • Example showing both deleted and excluded measurements

Shear Wave Graph

The Shear Wave Graph displays an average of Shear Wave stiffness and velocity measurements for a patient over a period of time. To see the Shear Wave Graph, select **Worksheet** and then **Graph** on the touch panel.




Figure 13-80 • Shear Wave Graph

Measurement Information

Table 13-42 • Shear Wave Speed Variation and Precision with Depth

Measurement	Units	Useful Range	Variation with Depth	Limitations or Conditions
Velocity	m/s	Shear wave ROI	Absolute range ≤ 0.5 m/s or relative range $\leq 15\%$, whichever is greater, for C1-6-D and C1-6VN-D. Absolute range ≤ 0.5 m/s or relative range $\leq 10\%$, whichever is greater, for all other probes with shear wave elastography.	Normalized SD less than or equal to 5% for multiple repeated measurements over the range of depths for which measurements can be made, limited by shear wave penetration.

The formula for Stiffness: $E = 3 \cdot \rho \cdot c^2$

Where E = Young's modulus of tissue

ρ = density of tissue (assumed to be 1 g/cc)

c = shear wave speed in m/s

NOTE

The conversion from shear wave speed (m/s) to Young's modulus is done under the assumption that the underlying material in which the shear wave propagates is linear, isotropic, incompressible, and homogenous.

Relative Range = (Absolute Range / Actual Velocity)

Normalized SD = (SD) / Average Velocity



CAUTION

The values for shear wave speed and tissue modulus are relative indices intended only for the purpose of comparison with other measurements performed using the system. Absolute values for these measurements may vary among different measurement devices. Use Shear Wave Elastography as a complement to other techniques when making a diagnostic decision.

Shear Wave Elastography Calculation

An Interquartile Range/Median (IQR/Median) Shear Wave Elastography ratio has been added. You can use this ratio to evaluate the reliability of Shear Wave measurements in the liver. This ratio is displayed as a percentage (%) on the worksheet (Display Accuracy to 1 decimal place). Values $<30\%$ are recommended for a reliable Shear Wave Liver exam.

NOTE

The IQR/Median ratio is calculated automatically for both Velocity and Stiffness and displayed by default to the user.

Scanning Hints



HINTS

You may find the following recommendations helpful when performing a Liver shear wave scan:

- Locate right lobe of liver intercostally.
- Place the ROI away from the capsule in an area free of vessels and fluid-filled structures.
- Suspend Patient breathing in mid-breath during the scan
- Position the ROI between 2-5 cm deep for an optimal shear wave scan

You may find the following recommendations helpful when performing a Breast shear wave scan:

- Locate lesion
- Place lesion in center of ROI, including a sufficient amount of the surrounding tissue
- Only compress slightly, if necessary (compression changes the elastic tissue properties).

To increase frame rate:

- Reduce ROI Width
- Turn off Penetration Mode
- Reduce Push Output

To increase penetration:

- Turn on Penetration Mode
- Keep Push and Track Output at 100%
- Place ROI away from edges of image

To reduce artifacts:

- Minimize motion during acquisition
- Ensure there are no vessels within the ROI or near the left or right edge of the ROI. It is advisable to avoid rib shadows whenever possible.
- Keep ROI at least 1cm away from the liver capsule.

Ultrasound-Guided Attenuation Parameter (UGAP) Option

Overview

Ultrasound-Guided Attenuation Parameter (UGAP) measures the attenuation value (i.e. attenuation coefficient [dB/cm/MHz] or attenuation rate [dB/m]) in the liver to evaluate diffuse liver disease. There are four visualizations: B Ref, Color Ref, B/Color Dual, and Qual./Att. Dual. All visualizations measure a representative attenuation value. Color Ref, B/Color Ref, and Qual./Att. Dual measure a representative attenuation value in 2D color map and dual display of B-Mode and 2D color map, respectively.

UGAP Availability

UGAP is available on the C1-6-D and C1-6VN-D probes in the Abdomen (Abd, AbdPen, and AbdDetail) application, and C2-9-D and C2-9VN-D probes in the PedAbd (Pediatric Abdomen) application.

Activation

To activate UGAP, select the appropriate probe Abdominal application, then press BT1 (configured as UGAP). UGAP parameters are displayed as “U” (Frequency and Acoustic Output Percent) in Simple Mode. Unit of attenuation coefficient values are displayed as dB/cm/MHz.




Figure 13-81 • UGAP Display

Measurement and Worksheet

Measure the Coefficient Value

1. Scan to get the appropriate view to measure.
2. Press **Start** (right Set key).

NOTE

The Measurement ROI can be moved using the Trackball; more than one ROI can be placed in one frame. Avoid structures such as vessels to obtain UGAP measurement.



Figure 13-82 • UGAP Study

Measurements can be performed afterwards using CINE or recalled data where multiple frames can be selected to measure attenuation coefficients.

Add a New Site

Once measurements are complete, you can elect to add another site to the exam. To add another site:

1. Press the **Measurement** button on the console.
2. Select **Add Site** on the Touch Panel.
3. Select **Enter Site Name** to add the name for the new site.



Figure 13-83 • Add Additional Shear Wave Site and Enter Site Name

NOTE
If you no longer need the added site, you can select to delete the site via **Delete Site** on the Touch Panel.

Summary and Overall Worksheets

Measurements are transferred to the Summary and Overall Worksheets.

Shear Elastic /UGAP (Site1)				
Name	Stiffness Site1	Velocity Site1	Attenuation Coefficient Site1	Attenuation Rate Site1
Med	2.26 kPa	0.87 m/s	0.50 dB/cm/MHz	
IQR	0.08 kPa	0.01 m/s	0.02 dB/cm/MHz	
IQR/Med	3.4 %	1.7 %	3.7 %	
N	6	6	5	
1	2.34 kPa	0.88 m/s	0.49 dB/cm/MHz	
2	2.26 kPa	0.87 m/s	0.50 dB/cm/MHz	
3	1.98 kPa	0.81 m/s	0.51 dB/cm/MHz	
4	2.24 kPa	0.86 m/s	0.51 dB/cm/MHz	
5	2.49 kPa	0.91 m/s	0.49 dB/cm/MHz	
6	2.25 kPa	0.87 m/s		
7				
8				

Figure 13-84 • UGAP Summary Worksheet



The screenshot shows a software interface titled "UGAP Overall Worksheet". It contains a table with columns for "Parameter", "Value", "sd", "sd2", "sd3", "sd4", "sd5", "sd6", and "Method". The table lists several parameters under the heading "B Mode Measurements":

Parameter	Value	sd	sd2	sd3	sd4	sd5	sd6	Method
B Mode Measurements								
Attenuation (dB/cm/MHz)								
C1	0.01 dB/cm	0.01						1.001
C2	0.05 dB/cm	0.05						1.001
C3	0.07 dB/cm	0.07						1.001
C Median	0.07 dB/cm	0.07						
C IQR	0.03 dB/cm	0.03						
C IQR/Median	3.2 %	3.2						
N	5	5						

Figure 13-85 • UGAP Overall Worksheet

Attenuation Coefficients are displayed (C1, C2, C3), Median, IQR, IQR/Median, and number of measurements (N).

2D Color Map

2D color map is available on Color Ref, B/Color Dual and Qual./Att. Dual. There are two 2D color maps: Attenuation Map and Quality Map. Attenuation Map provides the distribution of attenuation values. Quality Map provides the distribution of the signal quality to support ROI placement.




Figure 13-86 • Attenuation Map




Figure 13-87 • Quality Map

Continuous Wave Doppler (CWD)

Overview

There are two CW Doppler operating modes: Steerable and Non-Imaging.

Allows examination of blood flow data all along the Doppler Mode cursor rather than from any specific depth. Gather samples along the entire Doppler beam for rapid scanning of the heart. Range gated CW allows information to be gathered at higher velocities.

Steerable

Allows viewing of the B-Mode image to position the Doppler cursor to the area of interest while viewing the Doppler spectrum (shown below in the B-Mode image) and listening to the Doppler Audio signal.

Non-Imaging

Provides only Doppler Spectrum and Audio for ascending/descending aortic arch, other hard-to-get-to spaces or higher velocities.

Activating CW Doppler

To activate CW Doppler Mode, press **CW**.

The Steerable CW Doppler spectrum displays along with the B-Mode image. The cursor changes to a Doppler cursor.

You can now position and size the sample volume gate to get a velocity. Use Doppler Audio to listen for when the sample volume gate is positioned over an area of flow.

Update toggles between real time B-Mode with Doppler Mode and real time spectral display.

Exiting CW Doppler

To exit CW Doppler Mode, press **CW**.

Tissue Velocity Imaging (TVI)

Intended Use

Tissue Velocity Imaging (TVI) calculates and color-codes the velocities in tissue. The tissue velocity information is acquired by sampling of tissue Doppler velocity values at discrete points. The information is stored in a combined format with gray scale imaging during one or several cardiac cycles with high temporal resolution.



CAUTION

TVI can be activated on Cardiac Sector and TEE probes only.

Activating TVI

1. Select the desired probe.
2. While in B-Mode, press the **TVI** key located above the PDI control on the Touch Panel. The TVI image and Touch Panel display.

NOTE

To set up TVI on the Imaging Preset page, go to **Utility > Imaging > TVI**.

Optimizing TVI

The use of preset gives optimum performance with minimum adjustment. If necessary, the following controls can be adjusted to further optimize the TVI display:

- To reduce quantification noise (variance), the Nyquist limit should be as low as possible, without creating aliasing. To reduce the Nyquist limit: Reduce the Scale value.

NOTE

The Scale value also affects the frame rate. There is a trade off between the frame rate and quantification noise.

- TVI provides velocity information only in the beam direction. The apical view typically provides the best window since the beams are then approximately aligned to the longitudinal direction of the myocardium (except near the apex). To obtain radial or circumferential tissue velocities, a parasternal view must be used. However, from this window the beam cannot be aligned to the muscle for all the parts of the ventricle.

TVI and TVD

TVI

You can preset all parameters in **Utility > Imaging > TVI**.

The TVI parameters function the same as those described in the Color Flow specific section. The only differences would be that it pertains to tissue velocity rather than the color flow image. In the table below any TVI parameter or parameter specifics are noted.

Table 13-43 • TVI Parameters

Control	Details
Visible Description Adjusting Values	In LIVE/Freeze/Archive, you can display TVI Color with TVI. Select Visible on the Touch Panel. On or Off.
Invert	Color Invert
Baseline	Adjusts the Baseline.
Angle Steer	Steers the angle
Line Density	Optimizes B-Mode frame rate or spatial resolution for the best possible image.
Map	Values: TV1 and TV2.
Frame Average	Averages color frames.
Threshold	High values display more color. Low values limit the color to lower tissue echo (Opposite of Threshold in Color Flow Mode).
Transparency Map	Values 0-5.
Spatial Filter	Values 0 and 1.
Duplex	Allows two modes to be active at the same time.
TVI Gain	Control color transparency. High values display more color; low values display more tissue. This parameter is assigned to the Color Gain control.

TVD

While in TVI, press **PW** to activate Tissue Velocity Doppler.

You can preset all parameters in **Utility > Imaging > TVD**.

Quantitative Analysis (QAnalysis)

Overview

Quantitative Analysis is available for the following CINE loops obtained in the following modes: Tissue Velocity Imaging, Color Flow Mode, and Power Doppler Mode. All of the Quantitative Analysis modes operate similarly, with some variation.

The Touch Panel may be slightly different, for example; and the type of information you quantify varies by mode as well. Please see a summary of each mode below; followed by general instructions on how to perform Quantitative Analysis.

Quantitative Flow Analysis

Provide tools for semi-quantitative assessment of inflammation in joints and vascularization in tumors.

NOTE

There is frame number limitation for color quantification as 400 frames.

Statistics

The LOGIQ Fortis extracts various statistics from the image data within each sample area. The statistics depends on the imaging mode in use.

Press **Statistics** to enable/disable display of statistics of the frame or loop. The statistics are shown only when the loop is stopped.

- Ratio: Ratio of Color (Power) Doppler pixels over total sample area.
- Area (mm²): The size of sample area
- Max Ratio/Time of Max Ratio: Maximum Ratio of Color (Power) Doppler pixels in each sample area, and which frame that occurs in.
- Min Ratio/Time of Min Ratio: Minimum Ratio of Color (Power) Doppler pixels in each sample area, and which frame that occurs in.

Quantification supports two different statistical display formats: Short Form and Long Form.

- Short Form: Ratio and Area.



Figure 13-88 • Color Flow Mode Short Form - Example

- Long Form: Ratio, Area, Max Ratio/Time of Max Ratio, Min Ratio/Time of Min Ratio



Figure 13-89 • Color Flow Mode Long Form - Example

QAnalysis - Tissue Velocity Imaging

Multiple Time -Motion trace display from selected points in the myocardium.

QAnalysis Screen Description




Figure 13-90 • Q-Analysis Screen

Table 13-44 • QAnalysis Screen Description

1.	TVI Cineloop Window Sample Area: Indicates sampling position of the velocity. The sample area is color-coded: the first sample area is yellow, the second green, etc.
2.	B Cineloop Window Sample Area: Indicates sampling position of the velocity. The sample area is color-coded: the first sample area is yellow, the second green, etc.
3.	Analysis Window. • Y axis: Velocity scale (cm/s) • X axis: Time(s) • ECG • Time at cursor position. • Velocity at Cursor position. • Velocity at frame marker position (Color coded)
4.	Sample Area
5.	Time at cursor position and velocity at cursor position. Position the pointer cursor over the analysis window.
6.	Sample Area Tools. • Pencil Icon: Creates a sample area based on freehand drawing. • Shape Icon: Creates a sample area with a pre-defined circular/ellipse shape.

QAnalysis Plot Control

The following controls are user configurable presets which are configurable through the pull-down menu in QAnalysis mode. When using the pull-down menu:

1. Place the cursor over the analysis window and press the left **Set** key. The system menu displays at the cursor position.
2. Select the appropriate parameter.

To switch trace (Analysis signal)

Analysis Signal toggles the trace display between velocity, displacement or gray scale intensity curves.

1. Position the cursor over the plot window and select Analysis Signal from the pull-down menu.
2. Select **Velocity**, **Displacement** or **Grayscale Intensity** as necessary.

Vertical Unit

NOTE

Vertical Unit is only available when Grayscale intensity is selected in Analysis Signal.

When analyzing the data, the Y-axis can be set to display either logarithmic scale (dB) or linear, acoustic units (AU).

To toggle between dB and acoustical display units for the Y-axis.

- **dB**—The traditional log compressed B-Mode data is used to calculate the time-intensity curve values.
- **Acoustic**—The system reverse the log compression function to provide un-log compressed data for the Qanalysis.

Drift Compensation

Drift Compensation compensates drifting of Tissue Tracking curves by either resetting the curve to zero at the tracking start point (cycle resetting) or by linear compensation throughout the cycle (linear compensation).

NOTE

When Displacement is chosen by AnalysisSignal, Drift Compensation is active.

NOTE

Drift Compensation is inactive if ECG data cannot be acquired.

Trace Measurements

Gradient

Select Gradient entry on the pull-down menu that is obtained when the cursor is placed over the plot.

Gradient is displayed on the screen instead of velocity. The gradient calculates from 7 points (includes previous and next frames).

Max Gradient

Displays the time and gradient that becomes the maximum gradient between the CINE start and end frame.




Figure 13-91 • Max Gradient

Selecting QAnalysis Image Range

A range of frames is selected for the QAnalysis in Cine mode (before accessing QAnalysis). Only the frames in this range are used for the QAnalysis.

If a range is not selected prior to accessing the QAnalysis, the system uses the default Cine start and end frames as the default start and stop frames.

1. The first frame in the analysis series is selected by adjusting the **Start Frame** control to the desired frame
OR
using the **Trackball** or the **Frame by Frame** control to select the desired first frame and then selecting the **Start Frame** control.
2. The last frame in the analysis series is selected by adjusting the CINE **End Frame** control to the desired frame
OR
using the **Trackball** or the **Frame by Frame** control to select the desired end frame and then selecting the **End Frame** control.

Activating QAnalysis

1. Scan and Freeze the patient in the desired live mode or recall a desired cine loop from the stored images.

NOTE

QAnalysis is only available when the system is in CINE mode.

NOTE

Images from the current scan session acquired in the desired analysis mode (already in CINE) or from a saved image loop can be used for QAnalysis.

2. **QAnalysis** displays on the Cine Touch Panel.
3. Select **QAnalysis**. The QAnalysis screen and Touch Panel displays. To toggle the trackball function between QA and Scroll, press the top **Trackball** key.

Common QAnalysis Function

Display the System Menu

Place the cursor to the desired position and press the left **Set** key. The system menu displays at the cursor position.

The system menu is dependent on the area which you place the cursor.

Generating a Trace

Up to eight traces can be generated.

About the sample area

The sample area can be in three different states:

- Free sample area: freely moving sample area (QA cursor) before anchoring.

NOTE

The free sample area disappears when the QA cursor is moved over a static anchored frame.

- Static sample area: the free sample area is anchored by pressing Set.
- Dynamic anchored sample area: the sample area is anchored in two or more frames (see Manual tracking below). In these particular frames, the sample area is displayed with an anchor. The sample area moves smoothly between the anchored positions when playing/scrolling the cineloop.

Trace from a pre-defined sample area

1. If the trackball assignment is not on QA, press the top **Trackball** key until QA highlights.
2. If necessary, select the sample area Ellipse sample area button (shape icon on the monitor display).
3. Move the cursor to one of the Cineloop windows using the **Trackball**.
4. Press **Set** to anchor the sample area.

In this frame, the sample area is marked with an anchor. If the cineloop has more than one heart cycle, a sample area will also be anchored in the corresponding frame in the next heart cycle.

The trace is updated accordingly in the Analysis window.

Trace from freehand sample area

1. Select the Freehand sample area button (pencil icon on the monitor display).

2. Move the cursor to one of the Cineloop windows using the **Trackball**.
3. Press and hold down the **Set** key while drawing a sample area using the **Trackball**.
4. Release the **Set** key.

The sample area is automatically closed and the trace is updated accordingly in the Analysis window.

Manual tracking of the sample area (dynamic anchored sample area)

1. Place a sample area over a region of interest. Note the anatomical location of the sample area.
2. Scroll to a new frame using the **Trackball**.
3. Press the top **Trackball** key until the QA trackball assignment is selected.
4. Move the cursor to the sample area using the **Trackball**.
5. Press **Set**. The sample area is unanchored.
6. Drag the sample area to the corresponding anatomical location in the new frame.

When the sample area is anchored in more than one frame, linear interpolation is performed so that the sample area is smoothly moved between the anchored positions in the selected frames when running the cineloop.

7. Press the top **Trackball** key until the scroll trackball assignment is selected.
8. Using the **Trackball**, scroll through the cineloop and control that the sample area follows the moving anatomical structure.
9. Add anchored sample areas in several frames to obtain a more accurate displacement of the sample area.

Moving a dynamic anchored sample area

1. Press the top **Trackball** key until the scroll trackball assignment is selected.
2. Using the **Trackball**, browse through the cineloop to display one of the frames where the sample area was anchored.

NOTE

In these frames, the sample area is marked with an anchor.

3. Press the top **Trackball** key until the QA trackball assignment is selected.
4. Move the cursor to the sample area using the **Trackball**.
5. Press **Set**. The sample area is unanchored.
6. Drag the sample area to a new location.
7. Press **Set** to anchor the sample area to the new location.

If you want to move the sample area to the same depth, select **Move (same depth)** from the System Menu.

Manipulating the Sample Area

Up to eight sample areas can be saved on the reference image, with the corresponding eight traces plotted simultaneously on the graph. Each sample area display has a different color, and its corresponding trace data is plotted using that same color.

Once eight sample areas have been saved, the system does not automatically generate an active sample area when the cursor is positioned over the displayed reference image.

The saved sample areas can be a mixture of elliptical and freehand sample areas.

When the user repositions a sample area, the old trace data is erased from the plot and the trace data for the new position replotted.

If the sample area position on the last frame of the selected image range is moved, the corresponding sample areas on all frames are repositioned to match the last frame.

The user shall also have the capability of setting separate sample area positions on different frames of the contrast images, and the system shall linearly interpolate the sample area positions for the frames in between the selected frames.

Setting the default sample area shape

1. Select **Set sample area shape**. The Information Box displays.



Figure 13-92 • Sample Area Information Box

2. Select Height, Width and Tilt angle.
3. Select **Set as default**. The current sample area size is set as the default for subsequent Ellipse sample areas.

Sample Area Shapes

There are two different methods for determining the shapes of the sample area.

Ellipse sample area

1. Select the ellipse icon (shape icon on the monitor display).
2. When the trackball positions the image display cursor over the reference image(s), an elliptical sample area is automatically generated and displays on the reference image(s).
3. The average velocity value inside the ellipse is calculated for every image in the image analysis range and plotted in the image display area.
4. The last generated or selected ellipse is considered the active sample area, and its trace plot automatically updates as the user repositions it on the reference image. Old traces are erased.
5. When scanning with an elliptical sample area, press **Set** to fix the sample area position and freeze its corresponding trace on the plot. A new active sample area is generated whose position is manipulated by the trackball and whose velocity curve traces will be plotted as before, while the previous sample area and trace remain fixed at the points they were saved at.

NOTE

Elliptical sample areas can be positioned in any manner that keeps their center within the image boundaries. In the case that part of the sample area is outside the image boundary, only data from within the image boundary is used for calculating the mean velocity value.

NOTE

You can change the size of the Ellipse sample area by adjusting the Ellipse control.

Freehand sample area

1. Select Freehand icon (pencil icon on the monitor display).
Use the **Trackball** to position the caliper on the reference image at the start point. Press **Set** to fix the start point.
2. While holding down the **Set** key, trace the outline of the desired sample area by moving the caliper with the **Trackball**.

Reshaping a Sample Area

To reshape the sample area:

1. Position the cursor on the sample area to reshape and press the left **Set** key.
2. The sample area system menu displays. Select **Set sample area shape**.



Figure 13-93 • Sample Area Information Box

3. Adjust Height, Width and Tilt angle.
4. Press **OK**. The selected sample area size changes.

Labeling a Sample Area

The sample area label is used to identify data associated with the sample area when exporting.

1. Position the cursor on the sample area to label and press the left **Set** key.
2. The sample area system menu displays. Select **Label sample area**. The Label Dialog box displays.



Figure 13-94 • Label Dialog Box

Advanced Features

3. Enter a name for the sample area.
4. Select **OK**.

Copy, move, and paste a Sample Area

To copy and paste the sample area,

1. Move the cursor over the sample area and press the left Set key. The system menu displays.
2. Select **Copy sample area**.
3. Move the cursor to the desired location for the copied sample area and press the left Set Key. The system menu displays.
4. Select Paste sample area.

To copy and move the sample area,

1. Move the cursor over the sample area and press the left Set key. The system menu displays.
2. Select **Copy & move**. Or if you want to move to the same depth as the original sample area, select **Copy & move (same depth)**.
3. Move the copied sample area using the **Trackball**. Press **Set** to fix the position.

Deleting a Sample Area

Sample areas and their corresponding traces can be deleted using **Delete Sample Area**.

1. Select **Delete Sample Area**; a pull-down menu displays.



Figure 13-95 • Delete Sample Area pull-down menu

2. Select **Current sample** to delete the currently active sample area.
Select **Delete all** to delete all currently set sample areas and all of their traces.

NOTE

The corresponding traces for the deleted sample areas are erased from the plot.

NOTE

Deleting a sample area causes the sample areas to be deleted from all frames in the analysis loop.

QAnalysis Plot Control

NOTE

Plot Control is available only with TVI and Elastography modes.

The following controls are user configurable presets which are configurable through the pull-down menu in QAnalysis mode. When using the pull-down menu:

1. Place the cursor over the analysis window and press the left **Set** key. The system menu displays at the cursor position.
2. Select the appropriate parameter.

Vertical auto-scaling

The system can be configured to display the full unit range or a range according to the maximum and minimum values of the displayed trace(s) (auto-scaling function). In addition, the auto-scaling function can be set to be live update (updates while the sample area is moved) or delayed (updated when the sample area is anchored).

- **Delayed**—The system automatically rescale the vertical axis of the trace graph only when a new sample area is saved, to account for changing input dynamic range.
- **On**—The system automatically rescale the vertical axis of the trace graph every time the currently selected (active) sample area is moved.
- **Off**—Disable any automatic scaling of the vertical axis. There is user-defined system defaults on the system preset page for the fixed vertical scale to be used for the plot.




Figure 13-96 • Vertical Autoscale Pop-up menu

Line Style

- **Solid**—Setting the results in a plotted trace that does not display small boxes at the data points
- **Squares**—Setting the results in a plot where small squares are displayed at each data point, and the squares are linked together by lines.



Figure 13-97 • Line style Pop-up menu

Smoothing

The system can smooth the traces displayed by applying a filter over a defined time window. The type of filter available is depending on the analysis signal displayed.

1. Select **Smoothing**.

OR

Position the cursor over the analysis window and press the left **Set key**. The System menu is displayed at the cursor position. Select **Smoothing**.

NOTE

When smoothing is turned on, it applies to all traces in the plot window.

2. The smoothing filter list displays. Select the appropriate parameter.

Horizontal Sweep

Horizontal Sweep allows you to increase or decrease the time interval over which to plot the analysis curve.

The default is the user selected image range. If the user has not yet selected a first and last frame, the first and last default frames from the displayed CINE loop are used.

Zooming in the Analysis window

To zoom:

1. In the Analysis window, press and hold down the **Set** key while dragging the cursor to define the zooming area.
2. Release the **Set** key.

To unzoom:

3. Press the left **Set** key in the Analysis window. The system menu displays.
4. Select **Unzoom**.

Disabling/Enabling the frame

NOTE

Frame disable/enable is available only with Elastography, CF, and PDI modes.

Frame disabling excludes the actual frame from the cineloop display.

Disabling the frame from the frame marker

To disable One Frame:

1. Use the **Trackball** to move the cursor to the frame marker to disable.
2. Press **Set** to disable the frame.
3. The frame marker is changed from green to red to indicate the frame has been disabled.

NOTE

The disabled frame is no longer displayed in the reference window when scrolling through CINE memory.

Disabling multi-frames from the frame marker

1. Use the trackball to move the cursor to the first frame marker to disable.
2. Press and hold down **Set**
3. Move the cursor with the Trackball to the last frame to be disabled and release Set.

The marker is turn red and the data from that frame is removed from the trace and any subsequent trace processing.

Disabling a frame from the cineloop window

1. Use the trackball to move the cursor to the cineloop window.
2. Press the left **Set** key. The system menu displays.
3. Select **Disable frame**.

The current frame is disabled and the corresponding frame marker displays red.

Disabling Frames - Auto

When the user enters Elasto Quantification, the useless frames are disabled automatically. The LOGIQ Fortis uses Elastography Index value to find useless frame (low quality frames).

Disabling ECG triggered frame (where available)

In a multi-cycle acquisition, the user may deselect all frames in all heart cycles but a selected one. This function can be used for example to select a particular systolic frame for each heart cycle.


1. Scroll through the cineloop to identify the cardiac phase to analyze or identify the cardiac phase on the ECG trace (where available).
2. Position the cursor on the ECG trace (where available) and press left **Set** key. The system menu displays.
3. Select **ECG triggering** (where available).

All frames in all heart cycles are disabled except for the selected and corresponding frames in the other heart cycles.

To enable the frames

To re-enable all deleted frames:

1. Position the cursor on the Frame Marker line and press the left **Set** key. The system menu is displayed at the cursor position.
2. Select **Enable all frames**.
3. All disabled frames are re-enabled.



1. Analysis Window
2. Frame markers axis
3. Enabled frame (Green)
4. Disabled frame (Red)
5. ECG (where available)
6. Current frame

Figure 13-98 • Frame Markers

Exporting Traces (Saving the Trace Data)

You can save the trace data to an external file in ASCII format, readable in spreadsheet programs.

1. Select **Export Traces** to save the trace data.
2. Specify the following:
 - Location : Select Location which to save.
 - Filename : Enter the filename. (Only Text)
3. Select **OK** to save the data and return to the QAnalysis screen.
 - All displayed sample area traces are saved in the exported file.

NOTE

The Smoothed trace is the one saved if the user has applied a smoothing filter.

NOTE

Only data from the user selected image range is included in the exported trace file.

NOTE

No trace results are saved in the standard image database.

Annotating the QAnalysis Data

The user can annotate both the reference image and the trace plot displays. Use **Comment** key to type the annotation. See Chapter 6 for reference.

Printing QAnalysis Data

Press the appropriate print key in the desired analysis mode mode.

The system captures a single still frame which consists of the plot, the reference image, and user annotation.

Exiting QAnalysis

There are several methods to exit QAnalysis.

- Toggling **Exit QAnalysis** on the QATouch Panel.
- Press **Freeze** to unfreeze and resume scanning.
- Press any other button that returns the system to real-time scanning.

Stress Echo

Introduction

The LOGIQ Fortis Ultrasound system provides an integrated stress echo package, with the ability to perform image acquisition, review, image optimization, and wall segment scoring and reporting for a complete, efficient stress echo examination.

The stress package provides a protocol template for the two types of stress exams (exercise and pharmacological stress).

In addition to preset factory protocol templates, templates can be created or modified to suit your needs.

You can define various quad screen review groups, in any order and combination, that will suit your normal review protocol.

When reviewing stress examination images, the images are viewed at their original image quality, and different post-processing and zoom factors may be applied to the images under review for effective image optimization.

The protocol template may be configured for Continuous Capture.

A stress echo examination consists of three steps:

- Selection of a stress test protocol template
- Image acquisition
- Stress Analysis

NOTE

If WallMotion Segment Score is not displayed on the screen, select the **WallMotion** preset in the **Utility > Measure > M&A > Plot > Available Folders and Measurements**.

Getting started with a stress study

1. After selecting the appropriate application and probe, press the **Protocol** tab on the Touch Panel. The protocol screen displays the layout of the default stress protocol for the current probe. This layout is also known as a template.

Table 13-45 • Protocol Tab

Parameter	Description
Analyze	Display the Analysis screen
Template Editor	Display the template editor screen
Add Level	Add Level to the template
Delete Images	Delete the selected image
Move Image	Move the selected image to the another cell
Sync. Select	Synchronize the selected images.
End CC	End Continuous Capture

Parameter	Description
Begin/Cont.	Begin or continue the acquisition
Template	Display the template list
T1	Display/Hide the timer T1
T2	Display/Hide the timer T2
Cancel	Cancel Stress Echo

2. To use the current template, press **Begin/Cont.** to initiate scanning.
To use another template, press **Template**. The template list displays.



Figure 13-99 • Template List

3. **Trackball** to the desired template and press **Set**.
4. The selected template displays.



Figure 13-100 • Template (Example)

- a. Level
- b. Projection
- c. Current Acquisition (green)

5. Press **Begin/Cont.** to initiate scanning using the new template.

Image acquisition

Images are acquired in a pre-defined order, according to the selected template. The highlighted (green) cell of the template matrix, displayed in the Clipboard window, indicates which view is currently being acquired.

The names of the view and levels for the current cell are displayed on the top left corner of the image area and under the template matrix.

Acquisition Screen

1. Current View Level
2. Timer
3. Template Matrix
4. Current View (Green cell)

Starting acquisition

1. Select the template.
2. Press **Begin/Cont.**
3. Perform a scan that conforms with the view that is highlighted in the template matrix on the Clipboard window.
4. Press the **P1** (Image Store) key.
 - If the actual stress level is configured to preview the cine loop before storing, use the cine loop controls to select the most appropriate heart cycle and if desired adjust the loop markers. Press **P1** again to save the selected cine loop.
or
If you do not want to store the cine loop, press **Freeze** to cancel. Return to the scan screen.
 - If the actual stress level is not configured to preview the cine loop before storing, the system automatically stores the last heart cycle.

Stress levels can be configured for side-by-side display/comparison of the reference loop from baseline or previous level and the loop to acquire.

5. After storing the cine loop, the system automatically highlights the next view in the matrix to be acquired.
6. Repeat previous steps until all required views are completed.
7. If you select Auto Start Analysis on the Template Editor for this template, a dialogue asking "Do you want to start protocol analysis now?" displays when the last acquisition is complete. If you select Yes, the Stress Echo Analysis screen is displayed.

The template used can be configured so that analysis automatically starts by displaying the first protocol group. The wall segment scoring diagrams for each view is displayed in the Parameter window on the left side of the screen.

If the **Protocol** tab is selected during acquisition, the following Touch Panel displays.

Table 13-46 • Protocol Tab during acquisition

Parameter	Description
Stop	Stop Stress Echo.
Pause	Pause Stress Echo. The template matrix continues in display. Even if you press P1, the cine loop does not store to the matrix.
Select Cycles	The Continuous Capture Selection screen is displayed (only available in Continuous Capture mode).
Analyze	Enter Analysis screen.
Template	Enter Template screen.
Add Level	Add level to the template.
T2	Display (Start)/Hide Timer T2.

Selecting a view during acquisition

A fixed protocol is provided for scanning, based on the selected template. The system automatically highlights the next view to be acquired in the template matrix, as images are stored. However, the order of scanning may be changed manually as follows:

Manual selection of a view during acquisition

1. Use the **Trackball** or the **arrow keys** on the alphanumeric keyboard to move the cursor to the cell that represents the view to be acquired.
2. Press **Begin/Cont.** to initiate scanning.
3. Scan and save the selected loop as explained in the previous section.

After storage, the system automatically highlights the next available view to be acquired.

Moving an acquired image

An image can be moved from one cell to another during acquisition.

Procedure 1

1. When in the Protocol screen, press **Move Image**.
2. Use the **Trackball** to move the cursor to the desired image.
3. Press **Set**.
4. Use the **Trackball** to move the cursor to the destination cell.
5. Press **Set**. The image is moved from the source cell to the destination cell.

Procedure 2

1. In the Protocol screen, use the **Trackball** to move the cursor to the cell containing the image to move (source cell).
2. Press and hold down **Set**.
3. With the **Set** key still depressed, move the **Trackball** to the desired cell.
4. Release the **Set** key. The image is moved from the source cell to the destination cell.

If the destination cell contains an image, the images from the source and destination cells is exchanged when moving an acquired image.

Timers

Two timers can be displayed in the Stress mode acquisition screen, beside the template matrix.

Timers

- T1 displays the elapsed time from the start of stress examination.
- T2 starts when entering live scanning on the second stress level.

Both T1 and T2 timers can be manually stopped and restarted during the acquisition.

The display of T1 and T2 is user-configurable.

NOTE

If you activate the Timer in Stress Echo, the T1 timer is displayed in the lower left-hand corner of the image area after exiting Stress Echo.

Continuous Capture mode

Continuous Capture mode enables the user to perform acquisition continuously for all views at any level depending on the selected template configuration. Continuous Capture consists of temporary images and loops acquired during the session in a storage buffer in system memory. To enable best possible use of the limited storage buffer capacity, a Pause/Capture mode is provided, as opposed to the normal Freeze/Scan mode. The Pause mode enables scanning and live display on the screen, without any capture, thereby leaving the buffer available.



WARNING

Leaving a Stress Echo exam PRIOR TO saving portions of the Continuous Capture through “select cycle” or triggering a “Store All” could result in the loss of Continuous Capture data.

The continuous capture acquisition obtained using the P1 print key while in protocol is stored in temporary memory. Loss of power before properly ending the patient or before selecting cycles will result in the loss of this information in certain configurations.

Upon completion of cycle selection or when ending current patient, the user will be prompted as to what they would like to do with this temporary buffer. If they choose “store all,” the information will then be placed in permanent memory on the system’s hard drive.



WARNING

DO NOT change modes or power off the system while in Continuous Capture Mode. Doing so causes images and data acquired during the session to be lost.

To run Continuous Capture, the user has to select a template where this feature is activated.

The buffer bar

When entering a level with Continuous Capture enabled, a buffer bar displays in the window.

The Buffer bar displays the following information:

- The scanning state

- Pause (live scanning without storing)
- Capture (live scanning with storing temporary images and loops to system memory buffer)
- The percentage of the buffer that is filled
- The buffer filling progression showed by a filling gage
- The capturing sessions, reflected by the red lines along the buffer bar



Figure 13-101 • Buffer Bar

Controlling the capture process

When entering a stress level with Continuous Capture enabled, the system is automatically set in Pause mode.

1. Press **P1** to start image capture.
“Capture” is displayed in the buffer bar, the gage starts filling and the percentage of filled memory buffer increases.
2. Press P1 again to stop capture.
“Pause” is displayed in the buffer bar.

When 90% of the memory buffer is filled up, the text display in the buffer bar turns red.

The system enters Freeze mode automatically once the buffer is full and the captured loops display in the Continuous Capture selection screen.

Activating Continuous Capture

1. Do all your pre-stress acquisitions in the Cardiac application.
2. Press the **Protocol** tab to enter the Stress Echo mode. The Protocol screen displays.
3. Press **Template**. The template list displays.
4. Select the template **Exercise 2x4** from the list.
5. Press **Begin/Cont..**
6. Acquire the resting loops in all four views.

NOTE

Use the **P1** key to store the images.

7. Once the fourth loop is acquired, the system enters into a waiting mode where Continuous Capture is in a pause state awaiting the patient to exercise.
8. When the patient is back on the bed, press **P1**. The Continuous Capture acquisition starts.
9. Acquire all your views.
The memory buffer gage increases. When memory exceeds 90%, the percent number turns red.
10. Press **Freeze** to finish.
11. Press **Select Cycle**.
The Continuous Capture selection screen displays.
Refer to the next section if additional image acquisitions are necessary after the buffer is full.




Figure 13-102 • Continuous Capture Selection Screen

12. Assign the cine loops to the four views.
 - a. **Trackball** to the desired loop.
 - b. Press **Set**. A drop-down menu appears with the available choices.
 - c. **Trackball** to the appropriate view.
 - d. Press **Set**.
 - e. Continue these steps until all views are selected.

NOTE

To access additional cycles, use the arrow keys on the lower left portion of the select cycle screen.

- f. Select **Done** when complete. A dialogue window displays, asking whether the entire Continuous Capture acquisition should be saved.




Figure 13-103 • Drop-down Menu




Figure 13-104 • Dialogue Window

13. Press **Delete** to discard the loop or press **Store** to keep the entire loop.
14. Perform Analysis and scoring.

Continuous Capture with additional image acquisition

If the buffer is filled up before all the image acquisitions are done, additional loops can be stored in the clipboard before doing image assignment to the views.

1. Perform the Continuous Capture. *Activating Continuous Capture* on page 871 (Steps 1 to 11).
2. In the Continuous Capture selection screen, press **Select Later**.
The Continuous Capture screen displays.
3. Perform the additional acquisition.
4. In order to resume the stress echo exam and assign loops for the views from the Continuous Capture buffer, press **Protocol**. If not displayed, select the template **Exercise 2x4** from the template list.
5. Click the continuous capture images on the Protocol Template screen.
The Continuous Capture selection screen displays.
6. Assign the cine loops to the view. *Activating Continuous Capture* on page 871 (Step 12 a - f).
7. Press **Delete** to discard the loop or press **Store all** to keep the entire loop.
The normal procedure is to discard the loop. The loop is very big and requires a lot of disk space.
8. Perform Analysis and Scoring.

Postponed image assignment

The assignment of the cine loops to the view can be done on a later stage on a stored Continuous Capture acquisition.

1. Perform the Continuous Capture. *Activating Continuous Capture* on page 871 (Steps 1 to 11).
2. Press **Store all**.
The entire Continuous Capture acquisition is stored. The examination can be ended and the image assignment, analysis and scoring can be done later.
3. Re-open the examination, if necessary.
4. Press **Protocol**. The Protocol screen displays.
5. Click the continuous capture images on the Protocol Template screen.
The Continuous Capture selection screen displays.
6. Assign the cine loops to the view. *Activating Continuous Capture* on page 871 (Step 12 a - f).
7. Select **Done**.
8. Perform analysis and scoring.

Restart capture from the Continuous Capture Selection

- Press **Restart Capture**.
The recording in memory is deleted and the Continuous Capture starts again.

Resume Continuous Capture

- Press **Continue Capture**.
Resumes Continuous Capture recording (only if the Continuous Capture buffer is not full).

Assigning and storing the cine loop

The cine loops captured in the buffer are assigned to the stress protocol views and stored from the Continuous Capture selection screen.




Figure 13-105 • Continuous Capture Selection Screen

- 1 Highlighted loop
- 2 Cycle number and total number of cycles
- 3 Blue Gauge: Position of the highlighted loop within buffer area.
- 4 Red bar: Pause session.
- 5 Navigation Controls: << < > >> (back to first selection, back to previous selection, forward to next selection, and forward to final selection).

Assigning a cine loop to a view

1. Use the **Trackball** to move the cursor to the desired cine loop in order to assign it to a particular view of the stress template.
The frame of the loop is highlighted.
2. Press **Set**.
A pop-up menu displays with the view names of the template.




Figure 13-106 • Loop Assignment

3. Use the **Trackball** to select the required view name.
4. Press **Set**.
The name of the view displays above the timers in the window.
5. Repeat steps 1 through 4 to assign loops to the other views of the level.
6. Press **Done** when complete.
A dialogue window displays asking whether the entire Continuous Capture acquisition should be saved.
7. Press **Delete** to discard the loop or press **Store all** to keep the entire loop.
The normal procedure is to discard the loop. The loop is very big and requires a lot of disk space.

Post Acquisition Features

Post acquisition, you can utilize Raw Data to adjust the following in B-Mode:

- Zoom
- SRI-HD
- Rejection
- Frame Average
- TGC
- Maps
- Dynamic Range
- Gain
- Rotation

You can also take measurements post Stress Echo acquisition.

Analysis

Analysis consists of viewing previously saved loops and assigning scores to each cardiac segment, in order to quantify the function of the muscle or wall segment.

Depending on the protocol configuration, the analysis stage can start manually or automatically after completion of the stress test. In this case, the usual procedure consists of sequentially opening all image groups (if defined) and performing scoring from image to image.

The quad screen is the standard display for comparing heart cycles. The heart cycle loops in the display are synchronized to enable comparison. Each loop in the quad screen can be magnified, using the zoom control.

Image Selection for Analysis

Images can be selected manually or from a pre-defined group in the Protocol screen.

Selection of Images from a Group

If groups of images have been defined in the protocol template, you can select a group of images for analysis and sequentially analyze all images from all groups from within the Analysis screen.

1. In a stress examination, press **Protocol**. A preview of the acquisition displays.
2. Press **Analysis**. A pre-defined group appears in the display with a Wall Segment window on the left.
3. To advance to other groups, use the Trackball to move the cursor to the arrows at the bottom of the Wall Segment window. Select an arrow to advance to another group.

Manual selection of images from Analysis screen

1. When currently in the protocol analysis screen in the Stress analysis quad screen, hold down the **SHIFT** key while performing Steps 2 through 4.
2. Use the **Trackball** to move the cursor to the first image to select in the template matrix.
3. Press **Set**. The frame of the selected loop is in the Stress analysis screen and the next window in the quad screen is automatically selected.
4. Repeat step 2 and 3 to select other images.
5. Depress **SHIFT**.

Manual selection of images in the Protocol screen

1. In a stress examination, press **Protocol**. A preview of the acquisition displays.
2. Use the **Trackball** to move the cursor to the first image to select.
3. Press **Set**. The frame of the selected loop highlights.
4. Repeat Steps 2 and 3 to select other images.
5. Press **Analyze** to open images in the Analysis screen.

Scoring acquired loops

1. After image selection, press **Analyze**.
The Stress Echo analyze screen displays.




Figure 13-107 • Analysis Screen

1. Wall segment diagram
 2. Selected loop (Highlighted frame)
 3. Displayed loops (Highlighted frames)
 4. Exit Wall motion scoring
 5. Change page or enter next image group
2. Use the **Trackball** to move the cursor to a score.
 3. Press **Set**.
The score displays in the relevant segment area in the diagram.
- NOTE**
To edit a score, select it and choose a new score.
4. Repeat step 1 through 3 to score relevant segments.
 5. Press the **Change Page** arrow to display the next group of images.
 6. Repeat step 1 through 3 to score relevant segments on the new loops.




Figure 13-108 • Segment Scoring

NOTE

Since Cine changes into sync mode, subsequent scans are also synchronized. Exit sync mode from the Cine menu.

Editing/Creating template

The stress package provides protocol templates for exercise as well as pharmacological stress examinations.

The user can create new templates or modify existing templates to suit the individual needs. Up to ten projections and fourteen stress levels can be created in a template.

Templates created may be temporary, used only during the current examination, or saved as new templates, for future use and reference.

The editions that may be performed include:

- Adding/Deleting levels and projections.
- Assigning new labels to levels and projections.
- Defining level options.
- Defining new groups.

Templates are edited/created from the Template editor screen.

Entering the Template editor screen

1. Press **Protocol** to enter the stress echo mode.
2. Press **Template**. The template list displays.
3. Use the **Trackball** to select the Template Editor.
4. Press **Set**. The Template Editor screen displays.

OR

- 1. Press **Protocol** to enter the stress echo mode.
- 2. Press **Template Editor** on the Touch Panel. The Template Editor screen displays.

Template Editor screen overview




Figure 13-109 • Template Editor Screen

Table 13-47 • Template

Parameter	Description
Template	Select a pre-defined template from the pull-down menu. The protocol template preview updates accordingly.

Table 13-48 • Protocol Template Preview

Parameter	Description
Protocol Template Preview	• Displays an updated preview of the template accordingly to the settings applied. • To change Projection and Stress level labels, select a pre-defined label from the pull-down menu or press Set in the actual label field and type a new name.

Table 13-49 • Template Settings

Parameter	Description
Template Settings	- Cycles: select the number of cine loop heart cycles to store for each level from the pull-down menu or enter the desired value manually. - Continuous Capture: Checking this parameter enables continuous image acquisition throughout the level. The images acquired are temporarily stored in the system's storage buffer. - Preview of store: Checking this parameter enables review and adjustment of cine loops before store. - Show reference: Checking this parameter displays a dual screen with the reference level (first or previous level) on the left and the live image on the right.

Table 13-50 • Scan Mode

Parameter	Description
Scan Modes	2D, Color, PW (Pulsed Wave Doppler), CW (Continuous Wave Doppler), MM (M-Mode), Color MM, Color PW, Color CW

Table 13-51 • Other options

Parameter	Description
Other Options	- Grid Size: Enter the number of levels and projections for the selected template. - Timers: If you check this parameter, starts T1 and T2 timers automatically. - Auto-start analysis: If you check this parameter, displays the Stress Echo Analysis when the last acquisition is performed. - Show Systole in Analysis: When selected, the systolic part of the cardiac or ECG cycle is only displayed. The whole cycle is not displayed. - Smart Stress: Check Smart Stress to store and automatically reuse a subset of the image acquisition settings from a previous level in the corresponding views in the next level. When smart stress is used with "preview to store" function the timeline modes (PW, CW, M Mode, TDI) will only transfer image acquisition settings from baseline. - Reference image: When Show Reference is selected, selects either corresponding baseline loop or corresponding loop from the previous level to be displayed as reference image during acquisition.

Table 13-52 • Pre-defined groups

Parameter	Description
Pre-defined groups	- Shows the image groups created. - New group: Creates a new image group. Select the desired images on the template preview. - Update group: Edits a selected group after new loop selection on the template preview. - Delete group: Deletes a selected group.

Editing/Creating a template

Selecting a base template to edit

1. Select the base template from the template pull-down menu on the upper left corner.
2. Press **Set**.

The selected template displays in the protocol template preview field, showing the levels, projections and their labels.

Adding/Deleting levels and projections

1. Enter the number of levels and projections in the Grid size field.
The new grid size displays in the protocol template preview field.
2. Press **New Template** to create a new template.
or
Press **Save Template** to update the base template.

Display timers

Check the box(es) to display timer(s) as specified.

NOTE

The timers can also be started or stopped at any time during stress examination by using the **T1** and **T2** Touch Panel key.

Start analysis automatically

Check **Auto Start Analysis** to display the Stress Echo Analysis screen when the last acquisition is performed.

Smart Stress

Check Smart Stress to store and automatically reuse a subset of the image acquisition settings from a previous level in the corresponding views in the next level. When smart stress is used with "preview to store" function the timeline modes (PW, CW, M Mode, TDI) will only transfer image acquisition settings from baseline.

Configuring levels

The following options can be set up for each level:

Number of cycles to be stored in the cine loop:

1. Enter the desired number (1, 2, 3, 4, 5, 6, 7, 8, 9, 10, 15, 20) in the Cycles field.

Continuous Capture

1. Check Continuous Capture if continuous image acquisition throughout the level is desired.

When Continuous Capture is selected, preview of the cine loop and reference display during acquisition are not possible.

Preview of store

1. Check Preview of store if review and adjustment of cine loops before storage is desired.

Show reference

1. Check Show reference if the display of the corresponding reference loop is desired during acquisition (Dual screen mode).

Adding a group

1. In the Protocol template preview field, select the cells to be part of the group.
2. In the Pre-defined group field, press **New group**.
A dialogue box displays to ask the user to enter a name for the new group.
3. Enter the group name.
4. Press **OK**. The new group displays in the pre-defined group field.

Updating an existing group

1. In the Pre-defined group field, select the group to edit.
The selected cells are highlighted in the Protocol template preview field.
NOTE
A selected group is highlighted by a blue frame.
2. Either select a new cell(s) to add to the group or deselect an existing cell(s) to remove from the group.
3. Press **Update group**.
The display in the Protocol template preview field is updated accordingly.

Deleting a group

1. In the Pre-defined group field, select the group to delete.
NOTE
The selected group is highlighted by a blue frame.
2. Press **Delete group**.

Advanced Features

The group is removed from the list in the pre-defined group field.

Specifying Scan Mode for each Projection

Specify the Scan Mode for each Projection: 2D (B-Mode), Color Flow Mode, M-Mode, Color M-Mode, PW Mode, Color PW Mode, CW Mode, or Color CW Mode.

Saving the Template

You can save the template using controls at the bottom of the Template Editor page, or use the controls on the Touch Panel.

Table 13-53 • Template Editor Saving Options

Parameter	Value
New Template	Select this option to create an entirely new template.
Save As Template	If you would like to create a new template based on the existing template with your modifications, select to Save this Template As, and give it a name.
Save Template	Select this option to save the default template with your modifications.
Delete Template	Select this option to delete a template.

Wall Motion Segment Setup

You can set up the following parameters for Wall Motion Segment in the Utility screen (**Utility** > **Measure** > **Advanced** > **Cardiac**).

Table 13-54 • Wall Motion Segment Parameters

Parameter	Value
WMS freeze loop at ES	Specify to freeze the Loop at End Systole
WMS Segment Model	Select 16 or 18 segments
WMS initial scoring	Undefined or Normal
WMS scoring legend	ASE, ASIA or European

Utility Application Settings for Protocol

Table 13-55 • Protocol Parameters

Parameter	Description
Show Protocol Tab	Show/Hide the Protocol tab for that preset (Bicycle Normal, Bicycle Sporty, Contrast Pharmacological, Pharmacological 4x4, Pharmacological 8x5, Exercise 2x4, Exercise 2x4 B, Pharmacological US 4x4, or User-configured).
Template	Select the default template.

Report

If you set up the Wall Motion Analysis field on the Report, you can insert the results. Select Report to view either the Bull's Eye or Cut Plane Report.



Figure 13-110 • Bull's Eye Report Sample




Figure 13-111 • Cut Planes Report Example



HINTS

Select either Bull's Eye or Cut Plane on the **Reports > Adult Template**.



Cardiac Automated Functional Imaging (Cardiac Strain)

Cardiac Strain

Cardiac Automated Function Imaging (Cardiac Strain) is a decision support tool for global and regional assessment of the LV systolic function. Cardiac Strain calculates the myocardial tissue deformation based on feature tracking on B-Mode grey scale loops.

Cardiac Strain is performed on the standard TTE apical views, apical long-axis (APLAX), 4-chamber (A4CH) and 2-chamber (A2CH), as well as on the standard apical views acquired with a TEE probe.

Data Requirements

Data Types

All Cardiac Strain analyses are available on raw data, which also is the recommended data format for all AFI processing.

Views

Automated Function Imaging of the Left Ventricle can be performed on one or three of the standard views.

If a complete analysis of all three views is performed, the result is presented as a Bull's eye display showing color coded and numerical values for peak systolic full wall longitudinal strain, PSS (Peak Systolic Strain), TTP (Time To Peak global longitudinal strain) and traces.

If the analysis of one view alone is performed, the result is presented as a Quad layout, with two screens to evaluate the ROI and tracking, a strain curves graph and a Color Anatomical M-Mode (CMM).

Recommended Probes (RAW Data only)

Table 13-56 • Recommended probes

Application	Probe
Cardiac	M5Sc-D, 6Tc-RS
Pediatric	6S-D



CAUTION

Cardiac Strain is only recommended for adult cardiac images acquired with the M5Sc-D and 6Tc-RS probes. Cardiac Strain is only recommended for pediatric cardiac images acquired with the 6S-D probe.

Image characteristics

- Images acquired with Texture have different speckle characteristics and might therefore influence the results of speckle tracking. For this reason, it will not be possible to perform Cardiac Strain analysis of images where Texture was enabled.
- A frame rate higher than 40 frames per second is recommended. Lower frame rates are accepted by the tool (37 fps for raw data). Higher frame rates will benefit the workflow by making it easier to achieve satisfactory speckle tracking. A higher frame rate is recommended for high heart rates.
- The entire myocardium should be visible.
- A depth range that includes the entire heart chamber of interest (LV) should be used.
- Acquisitions must be made without the use of contrast agents.

Acquisition

Create an exam, connect the ECG device and make sure to obtain a stable ECG trace. The apical views may be acquired sequentially in 2D mode.

Sequential Acquisition

Acquire 2D grey scale cineloops of an APLAX, A4CH and A2CH view.

NOTE

It is recommended to acquire all three apical views sequentially to get similar heart rate in all views.

Heart rate and cycles

- Heart rate variability between the recordings should not be greater than 30%.
- If the acquisition has more than one heart cycle, the analysis will by default be done on the second to last heart cycle.
- The system should be configured to store 100 ms before and after each heart cycle.

Configuring the tool

It is possible to configure some of the AFI controls and analysis settings.

To access the configuration menu, press **Measure** on the control panel. Then select the AFI study on the Touch Panel to configure the AFI tool designed for analysis on the Left Ventricle.

Configurable Workflow Controls

NOTE

The presets in R2 and R3 are incompatible with R4.



Figure 13-112 • AFI Auto Processing Configuration

It is possible to configure some of the AFI controls to modify the workflow slightly.

Press Utility > Scanner Apps. Select the AFI tab. The following parameters may be adjusted:

- Autoprocessing timeout - The time the operator has to keep the trackball still before tracking starts automatically.
- BE Mode - Colormaps that shall be available for the Bulls Eye.
- ROI Method - Select whether fully automatic or 3-Click ROI shall be the tool default.
- YoYo - When adjusting the ROI, select whether a limited number of frames around the selected ROI frame should be looped back and forth to ease ROI adjustment.
- AVC Stage mode - Select whether to always open the AVC Selection stage. If “Auto” is selected, upon analysis of the APLAX view, the Results stage displays after the ROI stage. “Auto” AVC or Event timing is used by default. It is still possible to re-visit the AVC Selection stage to edit the chosen AVC.
- Prioritize event timing - Select “yes” to use “event timing” values for AVC if there are some in the current exam and AVC Stage mode is on “Auto”.
- Positive Peak Rule - If On, then a positive strain value is shown if the maximum positive peak exceeds 30% of the maximum negative peak, resulting in a blue segment in the Bull's Eye. If Off, then the negative strain value is shown regardless of the size of the positive peak.

Configurable analysis settings

- BE Segment model - Select the preferred Bulls Eye standard (ASE/AFI, 17/18 segment models).
- Strain reference layer - Select whether the strain values shall be calculated based on endocardial or full wall tracking.

Reference Layer for Strain

The tool supports calculation of strain parameters based on either endocardial deformation or myocardial/full wall deformation.



CAUTION

Measurements based on endocardial and myocardial analysis are not comparable.

- Endocardial deformation is assessed by performing analysis of the endocardial part of the ROI trace. Measurements derived from endocardial deformation are suffixed Endo, such as GPeakSysSL(Avg)_Endo
- Myocardial/full wall deformation is assessed by performing analysis of the whole ROI trace. Measurements derived from myocardial deformation are suffixed Full, such as GPeakSysSL(Avg)_Full

It is possible to change the reference layer preference using the Config screen, see *Auto EF* on page 902.

NOTE

If attempting to reprocess an analysis using a different reference layer a warning displays that the values in the analysis will change.

Starting the Tool

Starting Cardiac Strain from Sequential Acquisition (RAW data)

1. Open the desired exam for the Cardiac Strain analysis and select one of the apical images to use for the analysis.
2. Press **Measure** on the Control panel and select the Cardiac Strain study.

If the identification of a suitable triplet failed, or if the probe is a transesophageal probe, the tool starts up in the Select View stage.

NOTE

Automatic identification of the suitable triplets is based on view recognitions stored in the files (during Image Store), clipboard location, frame rate and heart rate proximities. The View Recognition algorithm was trained on 9720 recordings and verified on 1361 recordings. Algorithm was able to provide a classification in 98% of cases and had accuracy of 99% for the remaining.

Starting AFI from Simultaneous Acquisition (Tri-plane Acquisition)

1. Open a Tri-plane acquisition.
2. Press **Measure** on the Control Panel and select the AFI study. The AFI application is started displaying the APLAX view.

NOTE

When performing AFI on all three apical views, it is recommended to start with the APLAX view. This allows manual adjustment of the Aortic Valve Closure (AVC) event timing that is used in the calculation of the longitudinal systolic strain in all apical views.

Navigating the Tool

Perform the analysis by following an on-screen guided workflow



Figure 13-113 • A Typical AFI Workflow Menu

- 1 Stage previously visited
- 2 Active stage
- 3 Stages not yet accessible

The Workflow menu (or stage menu) is located on the right side of the tool. It contains a set of stages the operator should visit sequentially. It is possible to re-visit already viewed stages. Click on their title to navigate to any available stage. Stages that are not accessible given the tools state are greyed out.

The stages of an AFI analysis on the left ventricle are:

- Select View: indicates to the tool which view is analyzed.
- Define ROI: draws the ROI where speckle tracking takes place.
- Set AVC (APLAX only): defines the AVC time when processing the APLAX view.
- Results: examines the results of the analysis.

Most stages require interaction with the operator, and hints on what is expected are given below the stage title.

	By clicking the Reset button, the stage is reset and all user entered information in that stage is cleared. Any automatic procedures are re-run.
	If a stage is labeled, it means the stage is complete and nothing more needs to be done in that tool stage.

When available, press **Approve** and select next to return to the Select View stage and perform the analysis for a new view.

When available, press **Approve & Exit** to approve the analysis and store the reviewed measurements to worksheet.

At any time, click **Cancel** to exit the tool without saving any analysis.

Cardiac Strain on the APLAX View

1. The first stage when launching the tool is the Select View stage.
 - Perform any Left-Right flip corrections.
 - Select Cycle and adjust Cine markers as appropriate.
 - Annotate the view by clicking one of the view labeling buttons (A4CH, A2CH, APLAX).
2. An automatic ROI is generated when entering the Define ROI stage. The ROI may be edited by clicking and dragging on the endocardial and epicardial contours. See *ROI Adjustment* on page 895.

NOTE
If Cardiac Strain is performed on pediatric exams, the automatic ROI is disabled. See section about 3-click ROI in *To Create a New ROI* on page 895.
3. When satisfied with the ROI, either stop moving the cursor and wait for automatic processing to start or click on the Process button, Results for A4CH/A2CH, or Set AVC for APLAX. Now the system performs feature tracking to get a temporal ROI trace. On completion, it proceeds to the next stage.
4. (APLAX only) After the ROI edit stage, the system enters the Set AVC stage. Select one of the AVC setting strategies to verify AVC time. *Event Timing* on page 898 On completion, the Cardiac Strain tool now proceeds to the Results stage.
5. The Cardiac Strain tool is now displaying the Results stage in a Quad layout. Now the tracking quality must be inspected and verified. The tracked ROI is divided into segments. The tracking quality for each segment is automatically evaluated and applied

to reject segments for which the tracking is assumed to be unreliable. Segments that have been rejected do not have values in the segmental result ROI to the lower left in the Quad view, but instead shows an X. The strain trace (upper right), Curved Anatomic M-mode section (lower right) and parametric overlay on the dynamic ROI (upper left) are disabled for rejected segments. By clicking on a segment in the segmental result ROI it is possible to override the automatic quality assessment to either show or hide segmental values. The tracking for each segment must be visually controlled and validated as described below.




Figure 13-114 • Results Stage in Quad Layout

Once the tracking quality has been controlled for all segments, you may choose to click:

- Reprocess – To completely reprocess the view currently being reviewed; return to the “Define ROI” stage of the workflow.
- Approve and Select Next - Proceed to the Select View stage where the next view to process from the clipboard must be manually selected.
- Approve and Exit – To exit the tool and store performed measurements. No segmental strain results are stored in this case.

In case of a multi-cycle recording, the system automatically launches the second to last cycle for analysis. If during analysis, the operator wants to switch to another cycle, that is possible by entering the Select View stage and change cycle using the Cycle button.

Cardiac Strain on A4-Ch and A2-Ch views

The procedure for Cardiac Strain on Apical 4-chamber and 2-chamber views is similar to the one used in the APLAX view.

Perform the steps 1, 2, 3 and 5 from the APLAX procedure.

Perform the tracking validation (*Tracking Validation* on page 904) and, optionally, ROI adjustment (*ROI Adjustment* on page 895) procedures.



CAUTION

If the APLAX view was not analyzed first, the strain values displayed in the Quad screen for the A4CH/A2CH are temporarily labeled and may be different after APLAX has been analyzed. The reason for this is that for the A4CH and A2CH views, the AVC time is automatically set based on strain curve peaks (Auto mode). If, during APLAX analysis, the AVC time is manually set, the globally applied AVC time becomes different, causing segmental results to change.



CAUTION

If AVC mode is set to Auto, the final value of AVC time is not available before all three views have been analyzed. Thus, strain values displayed in Quad screen of the two views analyzed first are temporarily labeled. The reason for this is that the Auto-AVC calculation derived from all three views is most accurate and may be different from the intermediate AVC calculations used for each view.

AutoEF Layout

The results screen for A4CH and A2CH views also has an AutoEF layout. In this layout, the system presents automatically generated end systolic and end diastolic traces used to calculate ejection fraction (EF), stroke volume (SV), cardiac output (CO), as well as volumes. See *Auto EF Measurements* on page 902.

NOTE

If both A4CH and A2CH have been analyzed, the system also calculates biplane Simpson EF, SV, CO and volumes.

NOTE

The EF values (including volume values) provided when running the AFI tool appear in the Worksheet, Report and DICOM SR in exactly the same way as EF values provided when running the dedicated AutoEF tool. If AutoEF is performed before AFI, there will be two instances of the EF values in the Worksheet.

Tracking

The accuracy of the measurements provided by AFI rely on good tracking quality.



CAUTION

Poor tracking quality may lead to incorrect measurement results. The tracking for each segment must be visually controlled and validated.

Poor tracking quality could result from a variety of causes. The common causes for bad tracking are:

- Erroneous placement of the basal points when defining the border. If the basal points are placed too far from the annular region, the border segments at the annular base will not move together with the underlying 2D image throughout the entire heart beat.
- Erroneous placement of the apex point when defining the border. The point should be placed so that the resulting border trace covers mainly the endocardium. If the apex point is placed too high, the border trace will mainly cover the epicardium resulting in poor tracking.

- Too narrow ROI width. Narrowing the ROI too much will result in poor tracking due to lack of tissue data in the ROI.
- Too much clutter. Images with too much static clutter will result in poor tracking.

Inspect each segment and make sure that the center line is moving together with the underlying 2D image. Use the various results layouts to examine the tracking quality (e.g. Quad layout).

The following can help examining the tracking quality:

- Turn off the color overlay by selecting **Color Flow**.
- Reduce playback speed by using the Speed slider (or the rotary on the Touch Panel).
- Use the Single layout to get a larger view of the dynamic mesh (especially in difficult cases).

If the tracking needs to be improved for some segments, modify the ROI or create a new ROI.

ROI Adjustment

If the automatic ROI is not optimal (resulting in poor tracking), either adjust the ROI or create a new ROI as described below.

To Create a New ROI

Selecting the correct frame

The system automatically displays a frame where the endocardial border is usually clearly visible. To use another frame, while in Define ROI stage, pause the playback by pressing Stop. Then, use the Frame slider (or rotary) to select a different frame for ROI definition.

Creating a New Automatic ROI

To create a new automatic ROI, click the Reset stage button. This relaunches the automatic segmentation.

Creating a New Manual ROI

Sometimes the automatic ROI may fail to capture the correct ROI. In this case, a 3-Point ROI alternative is provided.

To generate an ROI by this method, when in Define ROI stage, select **3-Click**. The AFI tool prompts to click 3 landmarks. Use care to select proper landmarks, as displayed close to the pointer.

NOTE

The 3-Point ROI is the standard ROI method for TEE data and Pediatric exams.

When the third landmark is selected, a ROI is generated and can optionally be edited with the automatic ROI (see *Creating a New Automatic ROI* on page 895).

**CAUTION**

Faulty landmark selection may cause segment values to be swapped in the final results.

NOTE

If the ROI needs to be adjusted, make sure to make the changes immediately after the ROI displays, before the auto processing of the ROI begins.

NOTE

Auto processing of the ROI starts after a given time delay that is configurable.

NOTE

The Yo-yo function is turned on to help find correct location for the points.



Figure 13-115 • Defining an ROI






Guidelines When Re-creating the ROI



Correct ROI definition is crucial to get good tracking. By not following the ROI definition guidelines, the tool accuracy may be reduced. See use cases below for common pitfalls.

**HINTS**

Make sure to follow the recommendations when placing the three points (see below).

Table 13-57 • Tips for Re-creating the ROI

Tips	Correct	Incorrect
Base 1. Correct position of the base points. 2. The ROI extends into the aortic tract.	 An echocardiogram image showing a correct base ROI. The ROI is defined by three points (cyan, green, and blue) forming a triangle. A green checkmark is in the top right corner.	 An echocardiogram image showing an incorrect base ROI. The ROI is defined by three points (cyan, green, and blue) forming a triangle. A red 'X' is in the top right corner. The ROI extends into the aortic tract.
Apex 1. Correct position of the Apex points. 2. The apex point is placed too high. The ROI is extending beyond the epicardium.	 An echocardiogram image showing a correct apex ROI. The ROI is defined by three points (cyan, green, and blue) forming a triangle. A green checkmark is in the top right corner.	 An echocardiogram image showing an incorrect apex ROI. The ROI is defined by three points (cyan, green, and blue) forming a triangle. A red 'X' is in the top right corner. The apex point is placed too high, and the ROI extends beyond the epicardium.
Apex 1. Correct position of the Apex points. 2. The apex point is too high; the ROI is extending beyond the epicardium.	 An echocardiogram image showing a correct apex ROI. The ROI is defined by three points (cyan, green, and blue) forming a triangle. A green checkmark is in the top right corner.	 An echocardiogram image showing an incorrect apex ROI. The ROI is defined by three points (cyan, green, and blue) forming a triangle. A red 'X' is in the top right corner. The apex point is too high, and the ROI extends beyond the epicardium.

Tips	Correct	Incorrect
Bulges 1. Correct ROI. 2. ROI should not be bulging or follow the papillary muscle. To edit the ROI, see <i>ROI Adjustment</i> on page 895.	 A cardiac strain image showing a correct ROI (red outline) that follows the endocardial border without bulging or following the papillary muscle. A green checkmark is in the top right corner.	 A cardiac strain image showing an incorrect ROI (red outline) that bulges and follows the papillary muscle. A red X is in the top right corner.
General The left ventricle must be visible through the entire cycle. 1. End systole frame: the entire left ventricle displays. 2. End diastole frame: the annulus is not displayed.	 A cardiac strain image showing a correct ROI (red outline) that covers the entire left ventricle. A green checkmark is in the top right corner.	 A cardiac strain image showing an incorrect ROI (red outline) that does not cover the entire left ventricle. A red X is in the top right corner.

Event Timing

Timing information may be crucial to accurate diagnosis. The most important event timing is the aortic valve closure (AVC), since it is part of the definition of the end systolic strain parameter.

Determination of the AVC timing by the system is as follows, depending on the situation:

- An automatic AVC estimate determined by the temporal contraction of all LV segments (Strain curves) is used.
- From the APLAX view, adjust the estimated AVC timing. The adjusted AVC timing is then used in the other apical views when running Cardiac Strain on these views.

This option is only available from the APLAX view.

AVC Timing Adjustment (APLAX only)

After tracking is performed, the system enters the Set AVC stage. The following options for AVC settings are provided:

- Automatic: AVC is set automatically based on the ROI tracking using the time of peak negative strain.

- Manual: Set the AVC time manually. The frame can be selected by the track ball.



Figure 13-116 • Parametric Systolic Strain APLAX View

Inspecting Results

After completed tracking, and for APLAX completed AVC timing adjustment, the system displays the Quad screen layout for tracking validation (see page 8-26) and inspection of results for this view.

The screen contains the following result displays:

- Tracked ROI: A dynamic display of the tracked ROI to be used for tracking validation. The tracked ROI has a texture overlay indicating strain values according to the colormap. The overlay may be turned off/on by clicking the Color button.

NOTE

Rejected segments will not have texture overlay.

- Segmental result ROI: A static display indicating the peak strain value per segment. Rejected segments display with an X instead of a value. Clicking a segment changes its rejection status (*Tracking Validation* on page 904).
- Strain traces: A static display indicating the segmental and global strain traces with time. The peak values are indicated.

NOTE

The peak values can be adjusted by clicking and dragging the peak markers with the mouse.



CAUTION

Segments and peak detection should be checked to make sure that non-physiological traces are excluded in the calculation of indices.

- Curved anatomic M-mode: A static display indicating a curved anatomic M-Mode along the center of the ROI. The M-mode has a parametric overlay indicating peak strain values.

NOTE

Rejected segments will not have texture overlay.

NOTE

The Segmental result ROI, Strain traces and Curved anatomic M-mode are using color codes to link the different segments.

Peak Detection

The peak systolic strain detection for each segment can be verified and manually adjusted as required. To adjust the peak detection:

1. Press Traces. The Bull's Eye and Traces screen displays (see *Exiting AFI* on page 900)
2. To change the peak marker position on a curve:
 - Click on the peak marker (square point) on one of the curves, move the peak marker to a new position and click again to fix the point.
 - Place the cursor on a segment in the Bull's Eye. The corresponding curve is highlighted. Click on the segment to select the corresponding peak marker and move it to a new position.

The peak type is determined by the selected color map (systolic for PSS, global for PSI and TTP). The position of the AVC marker can also be checked in the Bull's Eye and Traces screen. If needed, the APLAX view should be reprocessed to change the AVC time.

Exiting AFI

At any time, the tool can be cancelled by pushing the Cancel button. The plugin closes and any unsaved data is lost.

When in the Results stage the operator may click the Approve and Exit button. The behavior depends on the number of views analyzed:

- If all three views have been analyzed, the tool goes to the Bull's eye only layout and prompts whether the results should be stored. If Yes is selected, measurements are transferred to the worksheet, and a result file with a screenshot of the Bull's Eye is generated.
- If only one or two views have been analyzed, the tool enters the Quad layout of the most recently processed view and prompts whether the results should be stored. If Yes

is selected, measurements are transferred to the worksheet and a result file with a screenshot of the Quad screen is generated.

Auto EF

Auto EF Measurements

Automated Ejection Fraction (AutoEF) is a semi-automatic measurement tool used for measurement of the global EF (Ejection fraction). The AutoEF tool is used as an optional decision support tool.

The AutoEF tool is derived from a 2D speckle tracking algorithm which tracks and calculates the myocardial tissue deformation based on feature tracking on 2D grey scale loops.

AutoEF is performed on either one or both apical 4-chamber or 2-chamber views, in any order.

The result is presented as Ejection Fraction value, calculated by Simpson MOD for each view and MOD Bi-plane Ejection Fraction for the whole LV. All values are stored to the worksheet when approved.

Table 13-58 • Recommended Probes

Application	Probe
Cardiac	M5S-c, 6Tc-RS
Pediatric	6S-D

Acquisition

1. Create an exam, connect the ECG device and make sure to obtain a stable ECG trace.
2. Acquire B-Mode cine loops of an Apical 4 chamber view (4-ch) and an Apical 2 chamber view (2-ch).

Acquisition Requirements

- The frame rate should be between 37 and 80 frames per second. A higher frame rate is recommended for high heart rate.
- The ultrasound system should be configured to store at least 100 ms before and after each heart cycle.
- If the acquisition has more than one heart cycle, the analysis will be done on the second last heart cycle.
- The entire myocardium should be visible.
- A depth range that includes the entire left ventricle should be used.
- Acquisitions must be made without the use of contrast agents.

Starting AutoEF

1. Recall any one of the stored views and press **Measure**.

2. Select **AutoEF** either in Measure menu or on Touch Panel.

The tool will launch and start up in the Select View stage.



CAUTION

AutoEF is only recommended for adult cardiac images acquired with the M5Sc-D probe. The measurement accuracies of the 2D Auto EF measurement values reported in the Advanced Reference Manual are verified with M5Sc-D probe.



Figure 13-117 • Select View Stage

AutoEF on the A4CH view

1. When in Define ROI stage: Verify that the view annotation shown to the upper left of the screen is correct. If it is not, either:
 - Click the Select View stage button to reannotate to the correct view and proceed analyzing that image.
 - Click on an A4CH image in the clipboard. This discards analysis of the current loop and replaces it with the one selected from clipboard. The tool starts in the Select view stage where it should be annotated as A4CH.

Pay attention to the left/right orientation of the image by comparing the LV wall names with a visual inspection of the image. If the image orientation is wrong:

- Go back to the Select view stage.
- Press Left-Right Flip.
- Verify the view by annotating it as A4CH again

NOTE

Alternatively, AutoEF can be exited, the image inverted and AutoEF started again.

2. An automatic endocardial ROI is generated when entering the Define ROI stage. The ROI may be edited by clicking and dragging on the endocardial and epicardial contours. *ROI Adjustment* on page 895.
3. When satisfied with the ROI, either stop moving the cursor and wait for automatic processing or click on the Process button. Now the system tracks the ROI with time. On completion, it proceeds to show the results in the EF result screen. The tracking must be visually controlled and validated as described below.

EF Results

The Results stage opens with a multi frame EF result layout.

- The running loop is shown on the left. A green dotted line marks the inner border of the chamber. In case of poor tracking, the system automatically displays parts of the border in red.

Advanced Features

- The frames with the maximal volume (ED) and minimal volume (ES) display on the right side.
 - Press EF Dual to only display the ED and ES frames.
- The End Diastolic volume (EDV), the End Systolic Volume (ESV) and the resulting Ejection Fraction (EF) display. Results for each view are summarized in a table on the right side.

Tracking Validation

- Inspect the ROI traces for the end systole and end diastole.
- If the tracking results are visually correct, select **Approve and Exit** to exit the tool and store the values to the worksheet, so they can be used in a report, or continue with **Approve and next view**.

Possible Causes of Poor Tracking

Poor tracking quality could result from a variety of causes. The common causes for bad tracking are:

- Erroneous placement of the basal points when defining the border. If the basal points are placed too far from the annular region, the border segments at the annular base will not move together with the underlying 2D image throughout the entire heart beat.
- Erroneous placement of the apex point when defining the border. The point should be placed so that the resulting border trace covers mainly the endocardium. If the apex point is placed too high, the border trace will mainly cover the epicardium resulting in poor tracking.
- Too much clutter. Images with too much static clutter will result in poor tracking.

Tracking Correction

The following can be done if tracking needs correction:

- Press EF dual to display ES and ED frames side-by-side.
- Adjust ES frame and ED frame controls if different frames need to be selected for ES and ED.
- Edit misaligned points on the endocardial border trace (see *Editing the Endocardial Border Trace* on page 905).
- Create a new endocardial border trace (*To Create a New ROI* on page 895)

Trace Adjustment of the Endocardial Border

If the automatic endocardial border detection is not optimal the user can either adjust the trace or create a new trace as described below.

**CAUTION**

Poor tracking quality may lead to incorrect measurement results. The tracking must be visually controlled and validated.

Editing the Endocardial Border Trace

- Enter the Define ROI stage.
- Adjust the trace by moving the cursor over the endocardial border trace, select an anchor point and drag it to a new location. The shape of the endocardial border trace updates accordingly.

In the upper right of the screen there is a pictogram indicating the core features of a good APLAX AutoEF ROI.




Figure 13-118 • Moving an anchor on the trace

To Create a New ROI Trace

The system automatically displays a frame where the endocardial border is usually clearly visible. To use another frame, while in Define ROI stage, pause the playback by pressing Stop. Then, use the Frame slider (or rotary) to select a different frame for ROI definition.

To create a new automatic ROI, click the Reset stage button. This relaunches the automatic segmentation.

Sometimes the automatic ROI may fail to capture the correct ROI. In this case, a 3-Point ROI alternative is provided.

To generate a ROI by this method, when in Define ROI stage, click on the button 3-Click. The AutoEF tool will now prompt to click 3 landmarks. Follow the indications displayed on the screen when placing the three points.

When the third landmark is selected, a ROI is generated and can optionally be edited as for the automatic ROI.

NOTE

If the endocardial border needs to be adjusted make sure to make the changes immediately after the ROI displays, before the auto processing of the ROI begins.

NOTE

The timing when auto processing of the ROI will start is configurable (from Config).

NOTE

The Yo-yo function is turned on to help find correct location for the points.



Figure 13-119 • Tracing the Endocardial border

Exiting AutoEF

At any time, the tool can be cancelled by pushing the Cancel button. The plugin will close and any unsaved data is lost.

When in the Results stage the Approve and Exit buttons may be selected. The tool enters the dual layout of the most recently processed view and prompts whether the results should be stored. If yes, measurements are transferred to the worksheet and a result file with a screenshot of the dual screen is generated.

ECG

Overview

A physiological input panel is available for the LOGIQ Fortis. This panel has inputs for ECG, PCG and auxiliary signals.

The physiological module consists of three channels:

1. ECG and Resp.
2. PCG
3. AUX

AUX is capable of handling external ECG signals from other diagnostic ECG devices.

The scanned image that is displayed is synchronized with the ECG and PCG traces. In Doppler or M-Mode, the traces are synchronized to that particular mode's sweep.

Approved accessory cables provide the proper signals to the Physiological Panel.



CAUTION

There will be a slight time gap between the ECG signal and the Doppler waveform when the selected PRF is low (less than 1.0 kHz).



CAUTION

To avoid skin burns in surgical use, do not place ECG electrodes in current path between Electrosurgical Unit (ESU) active and dispersive electrodes. Keep ESU cables away from ECG leads.

To display the ECG Signal on the monitor, go to **Utility > Imaging > General > ECG Display**.



Figure 13-120 • Optional Physiological Input Panel



WARNING

- Do not use with a defibrillator except with DEFIBRILLATION PROOF APPLIED PARTS. Only the ECG connection port is a defibrillator proof applied part.
- DO NOT USE the physiological traces of the ultrasound system for diagnosis and monitoring in lieu of ECG.
- Only approved and recommended peripherals and accessories should be used.
- After the defibrillator stimulates the patient, the ECG requires 4 to 5 seconds recovery time.

ECG Cable/Respiration

The ECG Cable is a modular cable consisting of two different cable parts:

- Single cable with a system connection at one end and a cable splitter at the other.
- A triple color-coded electrode cable to be inserted into the splitter device. Each electrode cable hooks up to the appropriate stick-on electrode by a color-coded clip type connector.

The color-coding of the electrodes follows one of two standards that are common in different parts of the world. The cable splitter device has a drawing defining the color codes, names and body location for the two standard color codes.



IEC (Europe, Asia, Rest of World)

1R: Red

2L: Yellow

3F: Green



AHA (USA)

1RA: White

2LA: Black

3LL: Red

Table 13-59 • ECG Color Code Cable

IEC (Europe, Asia, ROW)		AHA (USA)		Position of the human body surface
Electrode Mark	Color Code	Electrode Mark	Color Code	
R	Red	RA	White	Right Arm
L	Yellow	LA	Black	Left Arm
F	Green	LL	Red	Left Leg

Physiological Trace Monitor Display

The scanned image is synchronized with the ECG and PCG trace. In Doppler or M-Mode, the traces are synchronized with that particular mode's sweep.

The user can control the gain, position and sweep speed of the traces using the Touch Panel controls.




Figure 13-121 • Physiological Trace Monitor Display

1. PCG
2. AUX
3. ECG
4. Auto Heart Rate Display
5. Respiration

ECG Touch Panel

The ECG Touch Panel provides for control of the physiological input signals.

Without the ECG option, the ECG Touch Panel is not displayed.

Table 13-60 • ECG parameters

Parameter	Description
Sweep Speed	Change the speed of the trace. The sweep speed of the physio signal on the B-Mode image can be set independent of the timeline (Doppler and M-Mode) sweep speed. Value: 1 - 16.
ECG Lead	ECG Lead Pattern: 1, 2, 3 1: RA (-) to LA (+) (Right, Left, or Lateral) 2: RA (-) to LF (+) (Superior Inferior) 3: LA (-) to LF (+) (Superior Inferior) NOTE RA=Right Arm; LA=Left Arm; LF=Left Foot

Parameter	Description
Timer Trigger	Enables intermittent imaging based on a timer. NOTE If Timer Trigger is turned on, the ECG Trigger is set to None. Value: On or Off.
ECG Trigger	Enables intermittent imaging based on the ECG. The trigger location(s) relative to the R trigger are set with the Delay Time key. Adjusting: Press ECG Trigger and select one of the options and adjust the delay time using the Delay Time key. - ECG Trig 1 specifies the delay (ms) from R-wave to triggered frame. - ECG Trig 2 specifies the delay from R-wave to second frame. - Both activates ECG Trig 1 and ECG Trig 2 simultaneously. Trig 2 must be greater than Trig 1 for dual triggering (Both) to be active. NOTE If other than None is selected, Timer Trigger is turned off. Value: None, Trig1, Trig2, and Both.
AUX Display	Provides the ability to turn the AUX trace on for display on the monitor. Adjusting: When the key is selected, the AUX trace toggles between on and off. Value: On or Off
PCG Display	Provides the ability to turn the PCG (phonocardiogram) trace on for display on the monitor. Adjusting: When the key is selected, the PCG trace toggles between on and off. Value: On or Off
ECG Display	Provides the ability to turn on the ECG trace and Auto Heart Rate for display on the monitor. Adjusting: When the key is selected, the ECG trace and Auto Heart Rate toggles between on and off. Value: On or Off
Resp Display	Provides the ability to turn on the respiration tracer. Adjusting: When the key is selected, the respiration trace toggles between on and off. Value: On or Off
Resp Lead	Respirometer Lead Pattern: 1, 2 RA (-) to LA (+) (Right, Left or Lateral) 1. RA (-) to LF (+) (Superior Inferior) NOTE RA = Right Arm, LA = Left Arm, LF = Left Foot

Parameter	Description
ECG Trigger Period	The control specifies the number of heart cycles (R-waves) that are skipped between ECG triggers. The default is 1 or no skipping; 2= skip 1 cycle. Adjusting: Adjust the corresponding control. Value: 1 - 30
Delay Time	In ECG Trigger Mode: If only ECG Trig1 or ECG Trig2 is selected via the ECG Trigger key, the Delay Time key controls the R-Delay time of the active trigger. If both triggers are selected (Both), press this key to toggle ECG Trig1 and ECG Trig2 and rotate the key to change the delay time. Once the trigger is set, the snap shot image is displayed each time the update line passes the active trigger(s). In Timer Trigger Mode: Rotating the knob changes the delay time between images. NOTE Delay time may be different for Trig1/2 (0-2 seconds) and Timer Trigger. Adjusting: Adjust the corresponding control. Value: 0.10 - 10.00
AUX Gain/Position	Allows for the amplitude control of the AUX trace or allows for the vertical positioning of the AUX trace on the image display. Adjusting: Press the knob to toggle between Gain and Position. The default is Gain.
PCG Gain/Position	Allows for the amplitude control of the PCG trace or allows for the vertical positioning of the PCG trace on the image display. Adjusting: Press the knob to toggle between Gain and Position. The default is Gain.
ECG Gain/Position	Allows for the amplitude control of the ECG trace or allows for the vertical positioning of the ECG trace on the image display. Adjusting: Press the knob to toggle between Gain and Position. The default is Gain.
Resp Gain/Position	Allows for the amplitude control of the respiration trace or allows for the vertical positioning of the respiration trace on the image display. Adjusting: Press the knob to toggle between Gain and Position. The default is Gain.

Volume Navigation

Introduction

**DANGER**

DO NOT use the Volume Navigation feature on any patient relying on life-sustaining electronic equipment, such as a pacemaker or defibrillator. Failure to follow this instruction could lead to interference with patient electronic device(s).

**CAUTION**

When performing interventional procedures, remember that the pre-acquired dataset is not live and should not be used as sole guidance for interventional procedures.

Using a position sensor attached to the probe, or using a probe with V Nav Inside, Volume Navigation Fusion (V Nav) lets you import a pre-acquired DICOM volume dataset, register the location of the live Ultrasound image with the 3D volume image, and then use the position sensing system to simultaneously show the live Ultrasound image side-by-side with the corresponding multi-planar reformatted (MPR) slice from the pre-acquired dataset.

NOTE

DICOM datasets must have positional data to allow for V Nav to treat the datasets as a volume. Most CT, MR, and PET datasets have this information as do some X-Ray Angiography datasets.

In addition, you can use V Nav as a type of “GPS” positioning marker to track an anatomy of interest.

The V Nav Trackers allow you to use a V Nav sensor to track where the needle/needle tip is inside the body.

V Nav is available in B-Mode, Color Flow, Elastography, PDI, and Contrast Modes; it is not available while in 3D/4D or when timeline modes are active. Biopsy capability is available while in V Nav. Dual Caliper is available while in V Nav.

You can load Auto Sweep data for V-Nav, after the data stored as volume data in Easy 3D/Advanced 3D.

V Nav displays in Split Screen, with the Ultrasound Image on the left side of the display and the 3D Dataset on the right side of the display.



Figure 13-122 • V Nav Example

V Nav-Specific Definitions

3D Datasets: Computed Tomography (CT)-, or Magnetic Resonance Imaging (MR), Positron Emission Tomography (PET), XA (X-Ray Angiography), hand held SPECT datasets (NM) and Invenia (US) series stored in DICOM format. Ultrasound datasets acquired using Tru3D and AutoSweep.

Registration: Linking the 2D Ultrasound image to the 3D dataset. You can register the 3D dataset to the Ultrasound image via two techniques, Parallel Plane Registration or Point Registration.

Auto Registration: Auto-registration is available when fusing to a Tru3D dataset. When entering V Nav directly from a Tru3D acquisition, the Tru3D volume is automatically loaded into V Nav and automatically registered.

Parallel Plane Registration: This registration technique requires that the user mark a plane parallel to the 3D dataset slices and one common anatomical point, referred to as a Translation Point since it performs an x, y, and z correction at the point.

Image-Based Registration: This registration technique is available that automatically aligns CT/MR volume with the Ultrasound live image.

Point Registration: This registration technique requires that the user mark three or more anatomical point pairs. Performing a Plane Registration first may make the process of matching the anatomical point pairs easier.

Anatomical Point Pair: Marking the same anatomical location in the 2D Ultrasound and the 3D dataset creates an anatomical point pair.

Lock: After you have set the registrations between the 2D image and 3D dataset by either plane or point registration, you lock this registration in place. This means that the image content of the B-Mode image and the corresponding cut plane through the 3D dataset remains the same when you move the probe.

Magnetic Distortion: A magnetic field is generated by the position sensing system. It can become distorted in the presence of ferrous or highly conductive metal.

Environmental Quality: On-screen visual quality map that indicates distortion and proximity to the transmitter. Monitoring and controlling these two variables contribute to the highest quality environment.

RMSD, Root Mean Square Deviation: Root Mean Square Deviation, which equals the 'goodness' of fit. After the user has completed a Point Registration, there's a numeric

indicator to describe the goodness of fit. The lower the number, the better the fit. <1 to 10 would be good, 30-40 not so good.

Window Levelling: Balances the brightness/contrast in the 3D Dataset.

Position Sensing System: Consists of four components: Transmitter, Receiver (2), Probe Bracket, Probe/Receiver Cable Clips.

GPS Marker: Used in GPS to track the position of an anatomical structure while scanning or while performing a biopsy.

Virtual Tracker: The V Nav Virtual Tracker allows you to attach a sensor to the shaft of the needle (away from the tip). The position of the needle tip and the projected needle path are projected onto the 2D Ultrasound image.

Active Tracker: The Active Tracker is a device that consists of at least four markers that show up in a volume imaging dataset such as CT or MR. The markers are in known positions relative to each other such that the orientation of the Active Tracker can be uniquely determined based on the marker positions in the image. The device also holds a position sensor in a known position relative to the markers. Holes are provided in the Active Tracker so that its location can be marked on the body.

Needle Tip Tracker: The V Nav Needle Tip Tracker allows you to insert a V Nav tracker inside the Needle to track the projected path and needle tip location inside the body.

Clinical vs. Research: You can specify which V Nav controls you want to appear on the Touch Panel. You can set these via **UtilityApplication > Imaging Controls > Clinical Controls V Nav**.

Environment Quality

In order to obtain best results when registering, setting GPS markers, it is important to have the highest quality environment (minimal distortion and proximity to the transmitter).



Figure 13-123 • Environmental Indicator

An Environmental Quality reading for each sensor is provided via the Environmental Quality indicator (located in the upper, left-hand corner of the display).

Each sensor has a quality reading from 1 (low quality) to 7 (high quality). In addition, the distance between the two probe sensors is used to adjust the environmental quality reading presented to the user.

If needle tracking is being used, an overall environmental quality reading is provided.

Advanced Features

The Overall Environmental Quality reading is based on the environmental quality readings of the probe sensors and the needle sensor as well as the distance between the needle tip and the transducer face.

The environmental quality readings update in real time and are retained so that they can be displayed during cine scrolling.

Moving the cursor over the Environmental Quality indicator for each sensor provides the following information:

Sensor 1 displays a quality reading.

Sensor 2 displays a quality reading and a Distortion reading between the 2 probe sensors in [mm]. The magnetic distortion is represented as a measured distance between the two sensors relative to the expected distance between the two sensors.

The Needle Sensor displays a quality reading.

The Overall indicator displays an overall Quality reading, and a Probe to Tip distance in cm.

Calibration

To calibrate V Nav, contact GE HealthCare Service.

Configuration

You can specify the maximum retrospective loop length to store via **Utility > Application > Print controls > Time span**.

Accuracy

Accuracy depends on precise registration, minimal distortion, and controlling the patient's breathing during a scan.

Load the Volume Dataset

If you are using media to load the 3D dataset, you may want to store the CT, MR, PET, XA, DX, NM or Invenia Dataset on a USB flash drive, USB HDD or DVD so that you can clearly label the media in a way that identifies the patient information stored on the disk.

We recommend that you load the 3D Dataset onto the Ultrasound system prior to starting the V Nav exam to ensure that it loads properly and to ensure that the quality is acceptable.

Ultrasound volume datasets need to have been acquired using Tru3D, i.e., a volume dataset that was previously acquired using a position sensor.

Query/Retrieve

NOTE

For Query/Retrieve to find a patient, the patient MUST have a Patient ID.

Query



Figure 13-124 • V Nav Query

1. Press **Patient** and select **Data Transfer**. The Data Transfer screen displays.
2. Select Q/R.
3. Select the Query/Retrieve server from the Transfer From pull-down menu.
4. Type Patient name in the name field.
5. Press **Query** in the Transfer From section.
6. All studies for that patient and that modality will list.

Retrieve



Figure 13-125 • V Nav Retrieve

- 1 Select the specific series or study which use in V-Nav exam.
- 2 Press **Transfer**.

NOTE

Transfer time will vary based on network and file size.

Load DICOM Volume Dataset

Load a pre-acquired 3D DICOM volume dataset,

- Exit and re-enter V Nav and retain the 3D Dataset by pressing the V Nav control.

OR,

Discard the pre-acquired 3D Dataset,

- Press Exit and Clear on the Touch Panel.

From CD/USB



Figure 13-126 • Loading Dataset From CD/USB

1. If applicable, insert the media containing the pre-acquired 3D dataset. Volume datasets that can be loaded from disk may include CT, MR, PET, XA, DX, CBCT, and GE ABUS (US).
2. Select VNav. Press **Load** and select **From CD/USB**.
3. Double click on the DICOMDIR or database file (or single click and select the Open button).
4. Navigate to the location of the 3D dataset. A list of patients and a list of image series for each patient displays. Highlight the Series you want to import and select **Import Series**.
5. The Check Patient Data screen appears. Confirm that the patient information matches the patient associated with the current exam. After verifying patient information, press Accept (or Reject). You can also specify to display Patient Data by checking the Display Patient Data box.

If you turn on “Display Patient Data”, Patient name, ID and birth date are displayed on the Volume data. These remain on the live scanning image.

NOTE

If Hide Patient Data preset or Hide Date Time preset (Utility->System->General) is set to "On Store", then the patient data is removed from the volume dataset prior to image storage and then redisplayed after image storage.

6. B-Mode image displays on the left side and the loaded volume data displays on the right side.

From Database

NOTE

Volume datasets that can be loaded from the database: CT, MR, PET, XA, DX, CBCT, and GE ABUS (US).

1. Press **Load** and select **from Database**.
2. Select the desired volume from the Image History screen and press **Load**.
3. The Check Patient Data screen appears. Confirm that the patient information matches the patient associated with the current exam. After verifying patient information, press Accept (or Reject). You can also specify to display Patient Data by checking the Display Patient Data box.

If you turn on "Display Patient Data", Patient name, ID and birth date are displayed on the Volume data. These are remained on the live scanning image.



Figure 13-127 • Check Patient Data Dialog

NOTE

If Hide Patient Data preset or Hide Date Time preset (Utility->System->General) is set to "On Store", then the patient data is removed from the volume dataset prior to image storage and then redisplayed after image storage.

NOTE

If you load a Tru3D volume data, Check Patient Data is not displayed.

NOTE

If you Load From Database and want to cancel out, use the Exit button on the Image History screen.

4. B-Mode image displays on the left side and the loaded volume data displays on the right side.

Load Multi Volume Datasets

You can load multiple datasets (up to 10 datasets) of same study at the same time:

When loading multiple volume data at the same time, they must all be configured in the same direction (eg, Axial plane).

Sometimes there are multiple acquisitions stored within a single series. When you select Load All, these are loaded as one dataset (if they are not overlapping) or as two datasets (if two or more datasets are overlapping).

1. Load dataset From CD/USB, check data of the Select Series pop-up.



Figure 13-128 • Select Series pop-up

Load dataset from database, select one image from each exam and press **Load** in the Image History screen.



Figure 13-129 • Image History screen

2. V Nav screen and multiple dataset menu are displayed.



Item	Description
1-5	Dataset List (CT and PT in this example)
1	Display Color
2	Active Dataset Indicator
3	Active Dataset
4	Name of dataset
5	Display Weighting
6-12	Control Area
6	Single Display
7	Multiple Overlapping Display
8	Volume-to-Volume Registration
9	Compare Volume
10	Color Control
11	Delete
12	Expand/Collapse Dataset list

Figure 13-130 • Multiple Dataset Menu

Using the Touch Panel control, you can adjust Touch Panel Controls.



- 1 Dataset number and its Weight (%)
- 2 Toggle Up/Down: Select the dataset.
- 3 Push: Fix the selected dataset.
- 4 Rotate: Weight Dataset. You can weight the overlay amount of the volume dataset.

Figure 13-131 • Touch Panel Controls

NOTE

When you select Save Volume on the Touch Panel, all volumes are saved to a single file on the Clipboard, including GPS Markers.

Load additional volume datasets

Whenever a dataset is loaded and you want to load another dataset From CD/USB, from clipboard or from a database or if you generate a Tru3D dataset and then switch then back to VNav, the system asks you if you want to add the dataset as an additional dataset or to replace the existing dataset(s).

If you choose to add the dataset, the necessary adaption of the datasets with regard to size and resolution will be performed automatically at the end of the load.

Easy Access to PACS

The ultrasounds system is set up to easily receive exams from PACS.

When "Allow Trusted Devices" is enabled (checked), the following occurs:

- Communications are allowed from PACS (or any defined DICOM device) to the ultrasound system.
- Unsolicited inbound DICOM conversations from undefined devices are prevented to the ultrasound system.

When "Allow Trusted Devices" is disabled (unchecked), communication from external devices is allowed but not regulated.

These are differences between Query/Retrieve (Q/R) and Easy Access to PACS:

- Q/R is a PULL function from the ultrasound system to PACS. The query is initiated from the ultrasound system to PACS and then PACS sends the requested information.
- Easy Access to PACS is a PUSH function from PACS to the ultrasound system. PACS initiates action to send information to the ultrasound system.

To enable Easy Access to PACS when using trusted devices:

1. Navigate to Connectivity > Connection Manager > Scanner.
2. Select Add.
3. Under Properties, in Name, enter the device name.
4. Under Properties, in IP Address, enter the IP address for the device.
5. Repeat steps 3 and 4 for any additional devices.
6. If Allow Trusted (Configured) Devices Only (this is default state) is checked, it applies to all created devices.

The ultrasound system responds to TCP/IP ping or DICOM Echo only from devices with a defined DICOM source IP address. Because ONLY trusted devices are allowed to communicate to the ultrasound system, all those devices intending to push images need to be added under the Device tab.

Be aware that the user operation takes precedence. If the same exam or patient image pushed from PACS is in use in the ultrasound system, no updates to the image clipboard will be reflected until after the exam is ended.

Always refer to the DICOM Job Spooler to identify specific failed operations.

The ultrasound system displays DICOM image transmission status at the bottom of the screen as:

- Incoming DICOM Images: In Progress
- Incoming DICOM Images: Completed
- Incoming DICOM Images: Failed - Check spooler

If "Allow Trusted (Configured) Devices Only" is not checked, the devices sending to the ultrasound system do not need to be configured for Easy Access to PACS.

Setting Up V Nav sensors

Positioning Rules

**CAUTION**

When using fusion or GPS Markers, it is important to place the transmitter so that its location does not move relative to the working space. If you rearrange the working space (patient or transmitter location), you need to redo the registration or replace the GPS marker.

**CAUTION**

Place the transmitter such that the entire area of interest is in the front hemisphere of the transmitter. The front hemisphere generates the magnetic field used by the system.

The transmitter needs to be placed such that the face of the transmitter opposite of the cable (the field side) faces the working area

To avoid magnetic distortion, the following is recommended:

- Remove all ferrous or highly conductive metal from the vicinity of the transmitter and the working area.
- If the patient's bed contains ferrous or highly conductive metal, place the transmitter 20cm or more above the bed so that the central portion of the magnetic field sits well above the bed.
- To check for magnetic distortion, place a phantom that can be scanned from multiple directions in the planned working space. Using the 2D Marker on the GPS control on the Touch Panel, select Point. Scanning the phantom, place the Windows Pointer on a specific point in the phantom and then scan the same point from different directions. The distance between the graphical point (green cross) and actual structure is an indication of the amount of error. Larger than expected discrepancies could be the result of metal distortion or the transmitter field could be facing the wrong direction or be too far away from the working area.

Setting up the Position Sensing Apparatus

To set up the position sensing apparatus,

**CAUTION**

For accurate image registration, ensure that you connect the two probe position sensors to the probe as described in the instructions below. Match the numbers when you attach the sensors connectors to the sensor unit.

1. Attach the receiver probe bracket to the probe (refer to *Multi Angle Biopsy Guide Assembly* on page 661). Ensure that it is attached securely.
2. Fix the position sensor to the bracket.




Figure 13-132 • Set up the position sensor

1. Position the sensor's nub to the bracket's indentation.
2. Push the sensor into the bracket.
3. Tilt the sensor to position it in the bracket.
4. Tilt the sensor until positioned firmly in the bracket.
3. Connect the receiver cable from the Position Sensor to the front of the Ultrasound system.



Figure 13-133 • Volume Navigation Connections

1. V Nav Receiver Connector (2 cables)
2. V Nav Needle Tip Tracker/Virtual Tracker Sensor Connector
3. V Nav Active Tracker Connector
4. V Nav Transmitter Indicator
5. V Nav Transmitter Connector
4. Trace back each cable physically to insure the cable in position 1 on the probe is connected to position 1 on the console and that the cable in position 2 on the probe is connected to position 2 on the console.

5. Use the probe/receiver cable clips to hold the receiver cables to the probe cable.
6. Attach the transmitter cable to the front of the Ultrasound system where the transmitter label is shown.



CAUTION

Place the transmitter such that the entire area of interest is in the front hemisphere of the transmitter (indicated by the '1' in the illustration above). The front hemisphere generates the magnetic field used by the system.

7. Now that all the cables are connected, press the **V-Nav** key. The light above the transmitter cable blinks until the position sensing system has been initialized by the system. Once initialization is complete, the light remains green.

NOTE

Error message displays if the connection is incorrect.

8. Place electromagnetic transmitter next to patient with front toward scanning area of interest.

NOTE

Be sure patient and sensors are on the best position to start the exam.

NOTE

Patient must be in the same position as in the previous acquired exam.



CAUTION

If you need to connect/disconnect sensors, first use the V Nav key to exit V Nav. If you need to connect/disconnect the transmitter, make sure the light on the V Nav position sensing system is off. If it is on, enter V Nav and use the Exit & Clear Touch Panel control to leave V Nav.



HINTS

For information on how to set up the V Nav Needle Tip Tracker, see *Setting up the V Nav Needle Tip Tracker* on page 949

For information on how to set up the V Nav Virtual Tracker, see *Setting up the V Nav Virtual Tracker* on page 951.

V-Nav Inside Probe

V Nav Inside is a feature where the V Nav sensor is not external to the probe but is part of the probe.

Probes with V Nav Inside transmit/receive automatically after you activate V Nav.

Activate V-Nav

1. Begin an exam and optimize the image.
2. Press **V Nav** to activate Volume Navigation.
3. Perform Volume Navigation.

4. Press **V Nav** to exit V Nav. The following actions also exits V Nav: switching probes, recalling an image from the clipboard, starting a new patient, ending an exam, or selecting Exit and Clear from the Touch Panel.

V Nav Controls



Figure 13-134 • V Nav Controls

1. V Nav On/Off
2. Touch Panel and Joystick Rotary Controls
3. 3D Dataset X Axis Rotation Control
4. 3D Dataset Y Axis Rotation Control
5. 3D Dataset Z Axis Rotation Control
6. Windows Pointer
7. Comment
8. Measure
9. Zoom and Depth
10. P1
11. Left Set
12. Right Set
13. Trackball
14. Bottom Trackball key locks the V Nav Plane
15. Left/Right Key: Switches Display Mode

V Nav Touch Panel

V Nav Setup Touch Panel Menu

Prior to using V Nav, set up the following parameters:

Table 13-61 • V Nav Volume Touch Panel descriptions

VNav Parameter	Description
Load	Loads the 3D Dataset From CD/USB (CD, DVD, or USB storage device) or From Database.
Save Volume	Saves current volume, including any GPS markers shown in the volume.
Image-Based Registration	The system automatically aligns CT/MR volume with the real time ultrasound live image.
Patient Data Check	Confirms patient's data
Reference Sensor	Specify which sensor you are using.
Orientation	Select Axial, Sagittal, or Coronal
Orientation Markers	Places orientation markers on the image.
Display Mode	Select Ultrasound only, 3D only, or Split Screen (both)
Clear/Reset	Allows you to clear plane registration(s), point registration(s) or reset all, which clears all registrations and resets the 3D Dataset to its initial position.
Exit and Clear	Clears the current 3D Dataset and exits V Nav; the 3D dataset is no longer available.
Virtual Tracker	Setup and select the Virtual Tracker. [Appears when the Virtual Tracker is present.]
Active Tracker	Select the Active Tracker sensor.
Window/Level (or Center/Width) Rotary	Set the brightness/contrast for the 3D Dataset. To set the brightness, adjust the Level; to set the contrast, adjust the Window.

V Nav Touch Panel Menus**Table 13-62 • V Nav Touch Panel 1 Descriptions**

V Nav Parameter	Description
Scan	Specifies which image is moving with probe motion: Ultrasound (US), 3D, or Both
Registration	Select the type of Registration, Plane, Point/All, Point/Best 3, or None. Auto Registration is an additional choice if a Tru3D Ultrasound dataset is loaded using auto registration.
Save Current Registration	Select to save the current registration
Control Priority	Specifies the priority of shared image controls: Ultrasound image or 3D Dataset.
Restore Registration	Select to restore a stored registration.
Display Mode	Select Ultrasound only, 3D only, or Split Screen (both)
Clear/Reset	Allows you to clear plane registration(s), point registration(s) or reset all, which clears all registrations and resets the 3D Dataset to its initial position.
Overlay	Displays the 3D dataset superimposed on top of the Ultrasound image.
Overlay Brightness	Adjusts the intensity level of the 3D Dataset.
Overlay Weight	Adjusts the intensity level of the 3D Dataset.

V Nav Parameter	Description
Show Needle Tip	Shows or hides the needle graphics. [Appears when the Needle Tracker is present.]

Table 13-63 • V Nav Touch Panel 2 Descriptions

V Nav Parameter	Description
Save Volume	Saves current volume, including any GPS markers shown in the volume.
Measure Accuracy	Press Measure Accuracy, then select a point on both the 2D Ultrasound image and on the 3D Dataset to tell you the distance as an indication of accuracy.
Calibration Delete	Used for calibration purposes, which is only available with a service key and performed by GE HealthCare Service.
Store Registration to File	Press to save the current registration to a file. This is useful when you want to load another dataset with the same geometry. For example, both a T1 and T2 weighted MR series may be available.
Read Registration from File	Press to read a stored registration from a file.
Read Markers from File	Press to recall saved GPS markers from a file.
Show Scan Area	Shows or hides the Ultrasound scan area on the 3D Dataset.
Cut Mode	Select the cut mode: Cubic Plane, the 3D dataset slice in reference to the whole volume of data; Overview, fixed orientation slice; or Detailed, a slice in the same orientation as the current Ultrasound image.
Exit and Clear	Clears the current 3D Dataset and exits V Nav; the 3D dataset is no longer available.

DICOM and Image Storage

Still Images

To store a still image, press **Freeze**, and then press P1. If you want to CINE back to find an image, you may need to set the Priority on the Touch Panel to Ultrasound.

Loops

To store a loop after you have frozen the image, CINE back, change the Priority to Ultrasound (if necessary), indicate the start and end points, Run the CINE Loop, Press **P1**.

Image-by-Image

To store individual images to a USB memory stick or external hard drive via the USB port, recall the image from the Clipboard, select **Menu > SaveAs**. Type in the file name. We recommend that you select Image Only, Jpeg compression, Quality level 99. For CINE Loops, save as WMV. For stills, save as type Jpeg. Then press **Save**. Repeat for as many images as desired. When you take the USB device to another computer, the images are stored in the Export directory.

Entire Exam

You can use MPEGvue to store images for the entire exam. Insert the media (CD, DVD, USB device). Select **Patient > Data Transfer > MPEGvue > To**. Select the patients (multiple patients can be stored using the Ctrl+Shift keys). Then press **Transfer**. The still images are stored in Jpeg format and Loops are stored in MPEG format. Images are stored in

the Images directory, unless you select RunMeToStart. Then the images are displayed in a special viewer.

Save Volume

The system can save the volume and associated GPS markers as an image file in the exam.

The Save Volume key on the Touch Panel stores using the settings of whichever print key has Copy to Dataflow associated with it.

A saved volume can be loaded from the database and used as the pre-acquired volume.



HINTS

The Volume is stored in Raw Data format. If you send the volume to PACS without the raw data, the volume will not be available if you later Query/Retrieve it from PACS.

Measurements and Comments

You can measure distance, angle, circumference, or area on the 3D Dataset. You can annotate and perform any system measurement on the Ultrasound image.

Using V-Nav

Magnetic field range of V-Nav Transmitter

The position of the V-Nav sensor can be detected correctly within the range shown below. Make sure that the V-Nav sensor is within this range.

Check the Environmental Quality indicator on the monitor for the detection accuracy while scanning.

NOTE

Remove the metal product from the region of the magnetic field to increase accuracy of position detection.




Figure 13-135 • Magnetic Field Range from the Transmitter

From the center of the transmitter,

1. 30cm above and below
2. 28cm left and right
3. 20 - 66cm forward

Perform V-Nav Exam

In order to perform a V-Nav exam, you need to have acquired the 3D Dataset and set up the position sensing apparatus.

The workflow for V-Nav is to:

1. Import the dataset if using Query/Retrieve.
Query/Retrieve on page 916
2. Perform the Setup (attach the sensors) as needed.
Setting Up V Nav sensors on page 924

NOTE

Probes with V-Nav Inside transmit/receive automatically after you activate V-Nav.

3. Begin an exam; optimize the image.
4. Enter V-Nav.
5. Load the dataset.
6. Perform the registration.

NOTE

Confirm that the plane orientation of the ultrasound image is the same as that of the loaded volume data before starting registration.

7. Perform the exam.
8. Take measurements and store images via the Print key.



HINTS

Since Volume Navigation uses magnetic field to acquire probe position information, it may be affected by surrounding metal products.

If the image is flicked or misaligned, check the surroundings, such as the bed, the IV stand or the accessory, and move away those metal products from the magnetic field. If you use the bed with metal frame, it is desirable to be able to secure the distance from the metal frame using a thick mattress.

Ending an Exam

Press **V-Nav**, select **End Exam**, select **End Current Patient**, then select **Store all**. The system is ready for the next patient.

Perform the registration

Orient the Dataset and Pick a Slice

You can manipulate the 3D Dataset the same way you adjust any 3D volume.

Manipulate the dataset to pick a slice that most closely matches the area where you will hold the probe initially.

1. After the dataset has loaded, use the Zoom rotary to see the dataset in a larger size.
2. Use the Orientation key to select Axial, Sagittal, or Coronal to easily view the Dataset from different orientations.

You want the orientation that best matches the orientation of the patient.

NOTE

If the data appears unacceptably blurred or smeared in one of these views, the dataset was not acquired with the appropriate settings for viewing the data in 3D.

3. Pushing the X rotary rotates the dataset 90 degrees about the X-axis. Dialing the X rotary rotates the dataset about the X-axis in smaller increments. Same for Y and Z.

Note which points are identifiable in both the live Ultrasound image and in the 3D dataset. Remember that a point needs to be identifiable in all three dimensions. A point along a surface is often difficult to identify in three dimensions; but intersections of vessels, calcifications, or centers of structures can often be identified clearly in three dimensions.

Registering, Locking, and Translating the Image with the 3D Dataset

After you've loaded the 3D Dataset, you need to link the 2D Ultrasound image to the 3D dataset via Registration. You can register the 3D dataset to the Ultrasound image by Parallel Plane Registration or Point Registration.

Lock Plane: Align the section of displayed volume data with the vertical and horizontal orientations of the ultrasound scan plane. Match the state of inspiration and expiration with the loaded volume data and setting the probe orientation correctly to reduce the misalignment during the examination.

Lock Point: The position in the depth direction is adjusted at Lock Point. Synchronous display starts after setting Lock Plane. Set the reference point (Landmark) on the ultrasound and volume data. It can be set repeatedly.

NOTE

Confirm that the plane orientation of the ultrasound image is the same as that of the loaded volume data before starting registration.

NOTE

DO NOT move the patient once you have registered the image; if you move the patient, you lose registration.

Select the type of registration on the Touch Panel.

NOTE

When switching probes in V Nav, the registration is maintained. You can register with one probe, then scan with any probe.

Parallel Plane Registration

A plane registration constitutes a plane lock and a translation point.

To perform a Parallel Plane Registration,

1. Select **Registration > Plane** from the Touch Panel.
2. Hold the probe parallel to the acquired (imported) 3D dataset, being careful not to tilt, twist, or rotate the probe.
3. Press the **Lock Plane** (bottom Trackball key) to lock the plane as parallel.
4. Define a translation point, by marking a common point in each image (Ultrasound and 3D Dataset) using the Windows Pointer and Trackball key labelled **Lock Point**. This allows the system to do an x, y, and z correction at that point.
5. Move the probe and ensure that this registration is correct.
6. Press **Save Current Registration** on the Touch Panel to save this registration. You can save up to five (5) registrations.

Adjusting the Registration

To adjust the registration, you could try several things.

1. Save the registration. It's always a good idea to save the registration first.
2. Update the translation point in the area of interest.
3. Turn on Overlay and use the Z Rotary to adjust the 3D overlay relative to the Ultrasound image.
4. Holding the probe still during the following steps: move the windows pointer over the overlay image, select the Overlay XY Trackball control, move the overlay relative to the Ultrasound image, press the Overlay XY key to lock in the adjustment.
5. Manually adjust the parallel plane lock. Use the scan key on the Touch Panel to select only the 2D Ultrasound or the 3D dataset. Adjust the probe to make the scanned image match the locked image and then press **Lock Plane**.

6. Start over by using the **Clear/Reset > All control** on the Touch Panel.

Point Registration

To perform a Point Registration,

1. Select **Registration > Point** from the Touch Panel.
2. Find a common point on both the 3D Dataset and Ultrasound image. Use the Windows Pointer to select each point, then press Set Point each time. Do this 3 times.

NOTE

You can select either a Point/Best 3 or Point/All. If you identify more than three point pairs, Point/Best 3 takes the best 3 point pairings. Point/All uses all point pairs that you identify when calculating the Point Registration.

3. Move the probe and ensure that this registration is correct.
4. Press **Save Current Registration** on the Touch Panel to save this registration. You can save up to five (5) registrations.

NOTE

Point and Plane registrations are stored separately.

To switch between Point and Plane Registration, change the registration via the Touch Panel.



HINTS

Performing a plane registration is not necessary to do a point registration, but is recommended because it simplifies the process as the probe can easily be used to search for the point pairs in both the Ultrasound image and in the 3D Dataset.



HINTS

When using a point pair to complete or update a registration, registration error is minimized at that point and will increase further away from that point. Therefore, registration accuracy near the anatomy of interest may be improved by marking the last point pair, or adding a point pair, in close proximity to the anatomy of interest.

Auto Registration with Tru3D

Auto-registration is available when fusing to a Tru3D dataset.

When entering V Nav directly from a Tru3D acquisition, the Tru3D volume is automatically loaded into V Nav and automatically registered.

When a Tru3D data set from the current exam is loaded into V Nav from the clipboard or by loading from the database, the user shall be prompted to automatically register to the data set.

NOTE

Tru3D auto registration is possible if the magnet and patient have not been moved since the Tru3D data set was acquired. If the magnet or the patient have moved since the Tru3D dataset was acquired, you need to perform a parallel plane or point registration.

Auto Registration appears as a choice on the Registration control only if auto registration was initially used. You can manually select a different registration method even if Auto registration was initially selected when the volume was loaded.

Overlay

Use the Overlay control on the Touch Panel to overlay the 3D Dataset image onto the Ultrasound image. Use the Overlay Brightness and Overlay Weight joysticks to adjust the intensity level of the 3D Dataset.

By pointing at the overlay (Overlay XY) with the Windows pointer and clicking on it with the Left Set Key, the overlay can be dragged via the Trackball in X and Y directions in an effort to achieve a better registration. To lock in the adjustment, press the Left Set Key again.

By pointing at the overlay with the Windows Pointer and clicking on it with the Right Set Key (Overlay Z), the overlay can be dragged via the Trackball in the Z direction in an effort to achieve a better registration. To lock in the adjustment, press the Left Set Key again.

One technique to improve registration is to adjust the overlay in X and Y as stated above and to also do any necessary Z axis rotation as well. Then, turn the probe 90 degrees and repeat these same adjustments. These steps can be repeated iteratively.



HINTS

In an area with bony structures that show up bright with the 3D Dataset, applying an overlay with a low percentage can be a nice way to show bone structure with an Ultrasound tissue image.



HINTS

When performing Overlay XY or Overlay Z adjustments, it is important to hold the probe still from the time the adjustment key is pressed until it is pressed a second time to lock in the change.

Volume to Volume Registration

Datasets can come from different exams, can be from different orientations, and can be from different imaging modalities. Because datasets are not registered to one another, the system allows auto or manual volume-to-volume registration

To register multiple volumes,

1. Load multiple volumes (not restricted to one exam).
2. Clicks on the “link” icon. The “Select volumes for registration” pop-up window appears.
3. Select the primary and secondary moving volumes. The primary volume is used for registration, the secondary volumes are transformed using the registration parameters of the primary volume.

For example, If the user has a CT1 volume and a CT2 + PET volume loaded, they would select the CT1 as the fixed and the CT2 as primary, and the PET as the secondary moving volumes. After the registration they will have CT2 + PET registered to CT1.

4. Specify if the result should be added as a new registered volume or overwrite the moving dataset. You can also create the result as an additional volume (‘Add registered volume’ is not an option if 5 volumes already exist).

A user can perform two types of manual registration: a registration based off of a single common point pair identified in each volume (Translation); or a registration based off of three or more common point pairs identified in each volume (Registration).

5. Choose Manual, Auto, Semi Auto or Indirect volume to volume registration.



Figure 13-136 • Select volumes for registration

NOTE

If you select Auto registration, also choose the type (bone, tissue or some combination). If you select Manual registration, identify three point pairs between the two volumes. Only one point pair is usually needed for datasets with different orientations from the same exam.

During manual volume to volume registration, the user is able to scroll through either the fixed volume (shown on the left) or the moving volume (shown on the right) by moving the Windows Pointer over the image, pressing the trackball key labeled Scroll Z and then using the trackball to perform the scrolling.

NOTE

During manual volume to volume registration, the user is able to move the volume up and down or left and right on the screen by moving the Windows pointer over the image, switching from Primary to Secondary with the top trackball key, pressing the trackball key labeled Move XY and then using the trackball to perform the XY movement.

When doing manual registration, the left volume is the fixed volume and the right volume is the moving volume. The right trackball key is used to scroll through the datasets and find anatomical points while the left trackball key is used to mark the points. Once a single point pair is marked, the bottom trackball key is labeled Apply Translation which is all that is needed to register the moving volume (oblique, e.g.) to the fixed volume (axial, e.g.).


NOTE

A single point pair is sufficient if the two datasets have the same orientation but are shifted with respect to each other.


Once three point pairs are marked, the bottom trackball key is labeled Calculate Registration which then registers the moving volume to the fixed volume.

937

Three point pairs or more are needed if the datasets do not have the same orientation or are at oblique angles with respect to each other.



Once three point pairs are marked, the bottom trackball key is labeled Calculate Registration which then registers the moving volume to the fixed volume.



1. No point pairs identified
2. One point pairs identified
3. Three or more point pairs identified

If you load the volumes beforehand, the calculation will start immediately when you press Image-Based Registration button.

The liver vasculature must be included in both the 3D ultrasound and CT/MR volume in order for the registration to succeed.

1. Press **VNav** to enter Volume Navigation.
2. Press **Image-Based Registration** on the Touch Panel.

Select **Acquire** to acquire a new 3D ultrasound volume or **Load** to load an existing 3D ultrasound volume according to a navigate window.

NOTE

Need sensor information for 3D ultrasound volume.



Figure 13-139 • Navigate window for 3D US Volume

3. After 3D ultrasound volume is loaded, 3D US loaded icon turns green. Press **Next** on IBR-Process window.
4. Load portal phase volume of CE-CT or hepatobiliary phase volume of EOB-MR from CD/USB or Database according to a navigate window.



Figure 13-140 • Navigate window for CT/MR volume

5. After CT/MR volume is loaded, CT or MR loaded icon turns green. Press **Next**.
6. If multiple volumes are loaded, select volumes as below. Press **Auto** to start the calculation.



Figure 13-141 • Select volumes

- When the calculation is done, LOGIQ Fortis displays the aligned ultrasound live image and CT/MR image.




Figure 13-142 • Image-Based Registration

Compare Volume

You can compare pre-registered volumes in a side by side display.

- Press **Volume Compare** button in the multi volume menu.
- The dialog for selecting left and right side volumes displays. Multiple selection is possible.

NOTE

Volumes need to be from the same group of pre-registered volumes.



Figure 13-143 • Select volumes for comparison

3. The selected volumes is shown side by side and any interaction (scroll, zoom, rotation etc.) is applied to both sides simultaneously.




Figure 13-144 • Compare Volume

Advanced GPS Markers

"GPS" -- Position Marking V Nav

Position markers can be 'placed' on a stationary anatomical structure within the body and then tracked with the existing position sensing system. The marker position is projected onto the current Ultrasound image in a graphical way that indicates the relative position of the point compared to the image. GPS markers track a particular point in space.

Controls

Selecting the GPS tab switches trackball control to GPS markers with 2D Marker or 3D Marker activated.

Table 13-64 • GPS Controls

Preset Parameter	Description
2D Marker	Allows the user to select between a Point Marker and a Target Marker.
3D Marker	Allows the user to select between an Ellipsoid and Spherical Marker.
Show All Markers	Shows all markers, independent of Active State.
Show Active Markers	Shows/Hides all markers with Active State "On".
Active State	An individual GPS marker can be shown/hidden using the Active State control (turn to select, push to toggle state).
Delete	Rotate the control to select a specific GPS Marker to delete; then push the control to delete the marker.
Delete Markers	Deletes all active GPS Markers following the user's confirmation.
Mark Needle Tip	Mark Needle Tip allows a 2D or 3D GPS Marker (whichever is selected) to be put at the current tracked needle tip location. The marker is not attached to the needle and will not move as the needle moves.
Show Distance	Toggles between showing and hiding all defined distances.
Attach to Needle	If the 3D Marker key is selected and needle tracking is active, the "Attach To Needle" key initiates a 3D GPS Marker whose center (in the case of a sphere) or whose first long axis point (in the case of an ellipsoid) is placed at the needle tip.
3D Edit	The 3D Edit key is available only if a 3D GPS Marker is currently selected via the Delete or Active State rotaries. Selecting the Edit key causes the Edit GPS Touch Panel menu to appear.
Dist Start/Dist End/Clear	Rotating the Distance Start and Distance End rotaries cycles through the list of GPS markers including the Needle Tip if needle tracking is active. When two different GPS markers are identified on the Distance Start and Distance End Rotaries, the Distance End rotary is labeled "Define" or "Clear". If the distance is not already defined, the label is "Define". If the user then pushes the rotary, the distance between the Start and End GPS Markers is defined, and the label switches to "Clear". The initial value of the Distance Start rotary is the first GPS marker. The initial value of the Distance End rotary is the needle tip if needle tracking is active or the 2nd GPS marker if not. In the case of a 3D GPS Marker, it's center point is used for all distance calculations.

2D GPS Markers

GPS Markers track a particular point in space. As you adjust the probe/image position, the point is projected onto the current image. If the point intersects the current image, it is displayed as a green cross. As the point gets further away from the current image, the cross turns into a bigger and bigger square. The color of the square indicates the direction; red represents one direction and blue represents the other direction. You can use 'GPS' to track an anatomical marker on the Ultrasound image or for biopsy needle guidance, for example.

You can use GPS Markers with an imported dataset, or independently of an imported dataset.

There are two types of GPS Markers: Point and Target. A Point Marker assigns a number. Each point marker you set has a unique number. A Target Marker only allows one target

point. If you try to place another target marker, the target moves to the newly-indicated position.

To place a Target or Point Marker, use GPS Marker to specify Target or Point. Place the Windows Pointer in the location where you want to set the marker, then press the Set GPS Key.

3D GPS Markers




Figure 13-145 • 3D GPS Marker example

NOTE

When using GPS Markers with a registered 3D dataset, a point marked in either image shows up in both images. GPS Markers are stored with the image.

NOTE

All 3D GPS Markers are appended with the letter "V" (for volume).

NOTE

The inner surface of the 3D GPS marker and margin contours can be colored using the colors of the marker/margin. This color is transparently overlaid on top of the US image below.

Type of Markers

- Ellipsoid

The user places one GPS Marker (Set GPS on the right trackball key) at one end of the desired long axis. The point is marked with a marker named LA that behaves like a point marker. The distance between this first long axis point and the current cursor position is displayed in the status bar to allow the user to mark an ellipsoid with a defined long axis length. The user then places a second GPS Marker (Set GPS on the right trackball key) at the other end of the desired long axis. The ellipsoid will be drawn and the long axis endpoints will be marked with crosses that will be displayed only if they are hit by the plane.


- Spherical


The user places the GPS Marker (Set GPS key on right trackball key) at the center point of the anatomical structure of interest. The center of the sphere is marked with a cross, that is only displayed if the plane hits the center.

Editing a 3D Marker

Upon creating a 3D GPS marker, the Edit GPS menu appears on the Touch Panel.

Table 13-65 • Edit GPS Marker - Sphere/Ellipsoid

Preset Parameter	Description
Done	Return to main GPS Touch Panel.
Reposition	Set Key on the image will reposition the sphere when Reposition key is on (otherwise Set Key creates a new sphere).
Select	Select attributes that were previously saved.
Save As	Save and name the current sphere attributes (size, margin, colors and transparency).
Color	Change color (yellow, Red, orange, blue, purple, pink or white).
Margin Color	Change margin color.
Margin Dist.	Adjust margin size.
 Diameter Margin	
Diameter	Adjust sphere diameter.
Inner Alpha/Margin Alpha	Adjust transparency of main area and margin in percent i.e. between 0 (= only ultrasound image) and 100 (= only marker/margin color).
Detach from Needle	The Detach From Needle key is only available if the 3D GPS Marker being edited is attached to the needle. Pressing the key causes the 3D GPS Marker to be removed from the needle and placed at the current location of the tracked needle tip. It will no longer move with the needle tip.
 The length of the red ellipse is the long axis of the ellipsoid. The diameter of the blue circle is the short axis of the ellipsoid.	

Preset Parameter	Description
Shift/Rotate (Ellipsoid)	Shift along long axis or rotate about end of long axis.
Short Axis/Long Axis (Ellipsoid)	Adjust short and long axis lengths.
Show Ablation Needle Path	For ellipsoids, the user can display the needle path and tip position. 

NOTE

Default of the parameters is configurable in **Utility > System > System Imaging > V Nav 3D Marker**.

Move XY GPS allows the 3D Marker to be dragged to a new location in the current image plane.

Repos. GPS allows the 3D Marker to be recentered on the current image plane.



Figure 13-146 • Trackball key for GPS Marker

NOTE

The initial attributes of a 3D GPS marker match the last 3D GPS marker that was created.

NOTE

While on the Edit Menu, the 3D GPS marker being edited is shown in its normal color while all other 3D GPS markers are dimmed. After exiting the Edit Menu, all markers are returned to their normal color appearance.

Saving and Selecting a 3D GPS Marker

To save a 3D GPS Marker,

1. For markers that are used often, such as kill zones associated with a particular ablation needle, select **Save As** from the Edit GPS Touch Panel.
2. The Save 3D GPS Marker pop-up menu appears. Name the GPS Marker.




Figure 13-147 • Saving a 3D GPS Marker

To select a 3D GPS Marker,

1. Select **Select** from the Edit GPS Touch Panel. The Select 3D GPS Marker pop-up menu appears.




Figure 13-148 • Select 3D GPS Marker

2. Highlight the desired Sphere or Ellipsoid, then press **Select**.

Attach to Needle/Detach from Needle

The user attaches (and detach) 3D GPS markers to (from) a tracked needle tip and places a 2D GPS marker at the location of a tracked needle tip.

If the 3D Marker key is selected and needle tracking is active, the **Attach To Needle** key initiates a 3D GPS Marker whose center (in the case of a sphere) or whose first long axis point (in the case of an ellipsoid) is placed at the needle tip.

In the case of an Ellipsoid, it's long axis is aligned with the needle direction, i.e. both long axis points are placed on the needle axis. The 3D GPS Marker is not in a fixed location, but is attached to the needle and therefore moves with the needle tip.

The key is only available if needle tracking is enabled and the 3D Marker key is selected.

To detach the 3D Marker from the needle, the user can either press the **Detach from Needle** button on the GPS Edit page or the Mark Needle Tip on the GPS page.

Mark Needle Tip

Mark Needle Tip allows a 2D or 3D GPS Marker (whichever is selected) to be put at the current tracked needle tip location. The marker is not attached to the needle and will not move as the needle moves.

NOTE

You must use Tip Tracking or Virtual Tip Tracking.

1. Place a needle.
2. Before extracting the needle, select **Mark Needle Tip** to get a 2D or 3D GPS marker at the needle tip.




Figure 13-149 • Mark Needle Tip - Example

GPS Distance

The user displays the distance between two GPS markers or a GPS marker and a tracked needle.

- Between any two GPS markers (2D Point, 2D Target, 3D Sphere, 3D Ellipsoid)
- Between any GPS Marker and the tracked needle tip (Tip-Tracked, virtual tip-tracked)




Figure 13-150 • GPS Distance - Example

Procedure

1. After marking GPS locations, rotate Dist. Start to select 1st marker.

NOTE

Cycle through the list of GPS markers including the Needle Tip if needle tracking is active.

2. Rotate Dist. End to select 2nd Marker. Push Define to display the distance.

NOTE

If the distance is not already defined, the label is "Define". If the user then pushes the rotary and the distance between the Start and End GPS Markers is defined, the label switches to "Clear".

In the case of a 3D GPS Marker, it's center point is used for all distance calculations.

When distances are defined to be displayed, they are displayed in the same location as 3D measurement results. When regular 3D measurements are displayed, the distance measurements are not shown. The label of a distance measurement is the numerically lowest GPS Marker followed by the numerically larger GPS Marker. If the needle tip is one of the markers, it is listed second. Some examples are "1-2V", "3X-Needle", and "5-7P".

V Nav Trackers

V Nav supports the V Nav Needle Tip Tracker, the Virtual Tracker and Active Tracker.



Figure 13-151 • Volume Navigation Connections

1. V Nav Receiver Connector (2 cables)
2. V Nav Needle Tip Tracker/Virtual Tracker Sensor Connector
3. V Nav Active Tracker Connector
4. V Nav Transmitter Indicator
5. V Nav Transmitter Connector

NOTE

Pink probe brackets and sensors are necessary when using the Needle Tracker or Virtual Tracker



HINTS

- Remember to exit V Nav with the V Nav control BEFORE disconnecting and connecting the V Nav Trackers.
- With the V Nav Trackers attached, tracking can be turned on/off using the Show Needle Tip button.
- When using V Nav Tracking, it is important to have high Environmental Quality readings.
- When Needle Tracking is turned on, the Needle Tracking graphics show the:
 - Needle Tip and projected path on the live image
 - The intersection point with the current image when not in plane
 - Markers to assist with guiding the needle in plane with the current image
- The combination of Fusion and Needle Tracking is available once the volume dataset is registered. By selecting the Needle in Plane button, the volume dataset can show the needle in plane rather than the same plane as the live Ultrasound image.
- You can push the Z-Rotation Rotary to rotate the needle path plane about the needle axis (coarse adjustments). Rotating the Z-Rotation Rotary can be used for fine adjustments.

Setting up the V Nav Needle Tip Tracker

The V Nav Needle Tip Tracker allows you to insert a V Nav tracker inside the Needle to track where the needle tip is inside the body. The Needle Tip Tracker makes use of a needle assembly (combination of stylet and sheath) where the stylet has a hollow core in which a position sensor receiver may be placed, thereby placing the sensor near the tip of the stylet. The Needle Tip Tracker tip is annotated with an “N.”

To set up the V Nav Needle Tip Tracker, you need the CIVCO eTRAX Needle System Starter Kit.

1. The Kit includes the eTRAX Sensor as well as 5 sterile packages. Each sterile package contains a needle assembly (combination of stylet and sheath) and a cover for the sensor handle and cable.



Figure 13-152 • Needle Tip Sensor and Cable Close-up

2. The Needle Sensor fits inside the needle assembly. Prior to performing the procedure, you need to cover the sensor handle and cable.
3. Advance the sensor all the way into the needle assembly.
4. The Needle assembly locks into place with the Sensor handle.



CAUTION

For illustration purposes only, the needle sensor is shown without a cover. Always place a cover over needle sensor to protect patients and users from cross-contamination.

5. Using proper sterile technique, insert needle sensor through opening in cover, taking care to minimize bending of needle sensor.
6. Extend cover over needle handle.
7. Inspect cover to ensure there are no holes or tears.
8. After you have the Needle Assembly placed at the desired anatomical location, you can unscrew the stylet from the sheath and remove the stylet (and sensor). The procedure needle can be placed in the sheath to perform the procedure.
9. Remove the Sensor and Needle
10. The site is now ready for the procedure (core biopsy, ablation, etc.).

NOTE

Dispose of the needle assembly and sensor cover. The Needle sensor should not be disposed.

NOTE

You DO NOT need to sterilize the Needle Sensor. Instead, clean and disinfect the needle sensor the same way you would a probe, taking care not to bend the sensor.

NOTE

The Tracker must be used in conjunction with a probe that has a V Nav Bracket. You need three (3) sensors: two (2) sensors attached to the probe and one (1) Tracker sensor. However, if you are using a V Nav Inside probe, no sensors need to be attached to the probe and only the sensor for the tracker is needed.

V Nav Virtual Tracker (Part of V Nav Option)

The V Nav Virtual Tracker allows you to attach a sensor to the shaft of the needle (away from the tip). The position of the needle tip and the projected needle path are projected onto the 2D Ultrasound image. The virtual needle tip is annotated with a "V."

Setting up the V Nav Virtual Tracker

To set up the Virtual Tracker, you need the CIVCO VirtuTRAX Starter Kit.



Figure 13-153 • V Nav Virtual Tracker

1. The kit includes the VirtuTRAX Sensor as well as 5 sterile packages. Each sterile package contains a sensor bracket, a cover for the sensor, and rubber bands for the cover.
 Insert sensor inside cover. Place covered sensor in sensor bracket. Push down on sensor. Push down until sensor snaps into place.
2. Insert needle into insertion point of the bracket, making sure that the arrow on the side of the bracket points toward the needle tip. Rotate knob to tighten needle holder to adjust it to the needle's gauge.
3. Assembled V Nav Virtual Tracker.
4. Attach Virtual Sensor to the system. Connect V Nav sensors to the system.
5. You are ready to calibrate the Virtual Tracker. See *Calibrating the V Nav Virtual Tracker* procedure below.

6. To track the needle, the system needs to know the distance from the sensor bracket to the needle tip.
7. To provide the distance, select the V Nav Virtual Tracker control. Details are provided below.



8. On the display the virtual needle tracker appears like the needle tip tracker except that the needle tip is annotated with a "V."

NOTE

Dispose of the sensor bracket and sensor cover. The sensor should not be disposed.

NOTE

You DO NOT need to sterilize the sensor. Instead, clean and disinfect the sensor the same way you would a probe.

NOTE

The Tracker must be used in conjunction with a probe that has a V Nav Bracket. You need three (3) sensors: two (2) sensors attached to the probe and one (1) sensor attaching to the needle sensor bracket.



CAUTION

Because the sensor is away from the tip of the needle, the system cannot detect needle bending. When the needle bends, the projection of the needle trajectory and the representation of the needle tip will vary with respect to the actual needle trajectory and tip position.

Calibrating the V Nav Virtual Tracker

Method 1. You enter the needle length.

1. Select Virtual Tracker.
2. Enter a device name and specify a length. Press OK (Length is measured from the tip of the needle inserted to the far surface of the sensor bracket, measured along the needle shaft).
3. Check the length by placing the needle tip on the center of the probe face.
4. There should be a 'V' on the probe face in the image.

Method 2. You define the needle length via the image.

1. Select Virtual Tracker.
2. Enter a device name, and click on Define via Image. Press OK in the dialog.
3. Set the tip on the center of the probe face, move windows cursor to that point on the image and press Set.
4. The tip becomes marked with a 'V' on the image, initially on the probe face.

Existing Virtual Tracker devices can be selected and their calibrations modified if needed.

Active Tracker

The Active Tracker is a device that consists of four markers that show up in a volume imaging dataset such as CT or MR. The markers are in known positions relative to each other such that the orientation of the Active Tracker can be uniquely determined based on the marker positions in the image. The device also holds a position sensor in a known position relative to the markers. Holes are provided in the Active Tracker so that its location can be marked on the body.

When the Active Tracker sensor is attached to the Active Tracker device, it can be used as a reference sensor. This means that all tracking is done in relationship to this sensor rather than in relationship to the transmitter.



Figure 13-154 • Left, Active Tracker - Right, CT of Active Tracker




Figure 13-155 • MR Active Tracker

The Active Tracker adds sensor-based auto registration for CT/MR. An Active Tracker device, which can hold a sensor, is placed on the patient during the CT/MR scan and

then detected in the CT/MR images to facilitate an auto registration. CT/MR uses the same Active Tracker device as a Reference Sensor, meaning that tracking is done relative to the sensor attached to the Active Tracker. Provides some breathing compensation, and allows registration and GPS markers to be maintained in case the transmitter/patient moves.




Figure 13-156 • Active Tracker Setup and Positioning

1. Connect the following sensors to the V Nav Module. V Nav Probe Connectors (into Slots 1 and 2). Active Tracker Connector. Transmitter Connector.
2. Ensure that the probe has the V Nav attachment in place, along with the two V Nav Sensors.
3. Attach the CT/MR Active Tracker to the sensor.
4. Position the Active Tracker on the patient.

Using the Active Tracker

Perform CT/MR exam.

1. Apply sensor.

For an MR Exam, avoid having the MR coil touch the MR Active Tracker.

Place the MR Active Tracker where there is a gap in the coil.

The MR Active Tracker must be in the area being scanned during the MR. It is good to place the Active Tracker on a rigid area, such as the xyphoid, and good to place it in the vicinity of the intervention, but still out of the way.

Use pads (available in the MR suite) to keep the coil from touching the MR Active Tracker.

2. Perform CT/MR scan.
Scanning the patient,
3. Use the Active Tracker to hold the sensor.

4. Tape the sensor to the body. First, clean the area, then peel and stick the sensor to the body.
5. Choose the reference sensor on the Touch Panel.

NOTE

If you choose the Active Tracker for Auto Registration, the sensor is automatically selected as the reference.

6. Load MR/CT dataset.

With Sensor and Active Tracker

In order to acquire an ultrasound volume and auto register live ultrasound to it at a future time, perform the following steps:

1. Use a probe with VNav position sensors attached to or built into the ultrasound probe.
2. Attach the Active Tracker device to the patient and connect the active tracker sensor to the device.
3. Using a body marking pen, mark the position of the Active Tracker on the body using a body marking pen and the four holes in the base of the Active Tracker.
4. Enter Tru3D and select the acquisition as "With Sensor + Active Tracker"
5. Perform a Tru3D sweep as usual and store it as usual.

Entering VNav directly from Tru3D or entering VNav and then loading the saved Tru3D data set will allow automatic registration of the live ultrasound to the Tru3D data set. If you want to automatically register to the Tru3D data acquired with Active Tracker at a future time, perform the following steps:

6. Follow steps 1 and 2 above except when placing the Active Tracker device use the four body marks to recreate the same position as before.
7. Enter VNav and load the Tru3D data set that was previously acquired with the active tracker attached. Perform sensor-based auto registration in the same way as for CT/MR.
8. Manually adjust the registration as needed.

Vscan Air™ CL (Option)

LOGIQ supports Vscan Air™ - The Vscan Air is a battery-operated, wireless, general-purpose diagnostic handheld ultrasound imaging system which consists of a dual-headed probe.

More information about the Vscan Air CL, care and handling, can be found in the Vscan Air™ User Manual on the GE HealthCare Customer Documentation Portal website at: <http://gehealthcare.com/usermanual>

The Vscan Air CL acquires and forms the ultrasound image, sends the image data in real time through Wi-Fi to the ultrasound system, and the console displays the image. LOGIQ console provides the needed software to use the console as the display and the UI control unit.



Figure 13-157 • Vscan Air and Ultrasound System as Display and User Interface

The Vscan Air CL probe supports B mode and Color flow mode. Currently, PW and M mode are not supported on LOGIQ systems.

Table 13-66 • Clinical Applications - Additional Capabilities and Features

	Clinical Applications	Additional Capabilities and Features

Curved probe	• Abdomen (incl.Pleural) • OB/GYN • Pediatric • Peripheral Vascular • Musculoskeletal	Overwrite and create application presets Retain FOV Easy3D LOGIQ View LOGIQ Apps ATO (B Mode) Auto Doppler (Colorflow Mode - Linear) Rawdata (B & CF) Scan Assistant Compare Assistant Measurement Package (by application) Annotation package (by application) Body Pattern (by application) Imaging controls on Utility page Imaging display (dual/quad, B CF simultaneous) Print / Recall / change order Patient page/Exam Category
Linear probe	• Abdomen (incl.Pleural) • OB/GYN • Small Parts • Pediatric • Peripheral Vascular • Musculoskeletal	

Prepare the Vscan Air CL Probe for Use



CAUTION

The ultrasound system with Vscan Air CL is not intended for ophthalmic use or any use causing the acoustic beam to pass through the eye.



CAUTION

Only use GE HealthCare provided chargers or a Qi compliant charger with the Vscan Air CL probe. Failure to use a compliant charger may result in the probe not charging, or possible damage to the probe.



CAUTION

Scanning stops within 10 seconds if the Wi-Fi link to the Vscan Air CL probe is lost. Ensure the Wi-Fi connection is sufficient to avoid loss of image and delay of care.



CAUTION

The probe will not begin a scan if the Vscan Air CL probe battery is critically low. Ensure probe is sufficiently charged before beginning scan to avoid delay of care.



CAUTION

Scanning stops within 10 seconds when the Vscan Air CL probe battery becomes critically low. Ensure probe is sufficiently charged before beginning scan to avoid delay of care.

NOTE

Ensure only one Vscan Air CL is around the console. The ultrasound system will try to pair with the VScan Air CL with strongest signal.

NOTE

To guarantee the maximum performance with Vscan Air CL on the ultrasound system, check the environment and avoid other Wi-Fi devices.

NOTE

Make sure the ultrasound system is connected to an access point (if needed) BEFORE pairing with Vscan Air CL. Connecting to an access point after pairing with Vscan Air CL will result in significant performance degradation.

NOTE

Connect to the network infrastructure through an ethernet connection, if possible, to avoid any possible Wi-Fi interference.

NOTE

The VScan Air CL must be activated from a mobile (Android, iOS) device first through the Vscan Air app. The ultrasound system does not provide a workflow for activating a new VScan Air CL.



CAUTION

The Vscan Air CL may lose imaging capability if the probe firmware is incompatible with the console (or mobile device). Use caution to ensure console compatibility when performing Vscan Air CL probe firmware updates. Check Utility > Connectivity > Vscan Air under “Special Devices” to check which versions are compatible with the system.

Vscan Air On Console Charger



CAUTION

The Vscan Air On Console Charger must ONLY be used to charge the Vscan Air CL probe. DO NOT charge a mobile phone or any other device with the Vscan Air On Console Charger. Failure to follow these instructions may damage the charger or other device.

FCC Statement

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

Warning: Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

Note: This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation.

This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

(1) Reorient or relocate the receiving antenna. (2) Increase the separation between the equipment and receiver. (3) Connect the equipment into an outlet on a circuit different from that to which the receiver is connected. (4) Consult the dealer or an experienced radio / TV technician for help.

Radiation Exposure Statement

This equipment complies with FCC radiation exposure limits set forth for an uncontrolled environment. This equipment should be installed and operated with minimum distance of 20 cm between the radiator and your body.

IC Statement

This device complies with RSS-216 of Industry Canada. Operation is subject to the condition that this device does not cause harmful interference.

This device contains license-exempt transmitter(s)/receiver(s) that comply with Innovation, Science and Economic Development Canada's license-exempt RSS(s). Operation is subject to the following two conditions:

- (1) This device may not cause interference.
- (2) Any changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

RF exposure statement: The equipment complies with IC Radiation exposure limit set forth for uncontrolled environment. This equipment should be installed and operated with minimum distance 20cm between the radiator and your body.

Install the Vscan Air On Console Charger

1. Install the probe holder onto the probe charger by sliding down until it latches into place.



Figure 13-158 • Install Probe Holder on Charger

2. Select the probe holder location on the system to install the charger assembly and insert the probe charger post through the holder location.
3. Thread the nut onto the bottom end of the charger mounting post and tighten the nut by hand until the charger is secured firmly to the system.
4. Connect the USB charging cable to a USB port under the control panel (see *Figure 13-159 • LOGIQ Fortis USB Ports* on page 960).



Figure 13-159 • LOGIQ Fortis USB Ports

5. Place the Vscan Air CL probe into the charger to begin charging.



Figure 13-160 • Vscan Air On Console Charger Installed (with Probe)

Charge the Vscan Air CL Probe

To charge the Vscan Air CL probe, place the probe inside of the plastic probe holder on the probe charger. Reference *Table 13-67 • Vscan Air On Console Charger LED Charging Status Indicator* on page 960 for details on the battery charging status.

Table 13-67 • Vscan Air On Console Charger LED Charging Status Indicator

LED Color	Charging Status
Blue	Charger initially plugged in - no probe on charger.
Blinking Green	Probe battery charging.
Blinking Green and Static Blue	Probe battery fully charged - probe on charger.
Blue	No probe on charger.

LED Color	Charging Status
Orange	Charging failure. Probe battery not full, charging paused.

Pair Vscan Air CL to the Ultrasound System

To pair the Vscan Air CL to the ultrasound system:

1. Turn the ultrasound system on.
2. Press and hold the Vscan Air CL power button for approximately two seconds, while watching the probe LED lights.

NOTE

Do not continuously hold the button for longer than 5 seconds, or the probe will shut down after booting up.

The LED lights will first briefly display the battery level (green, yellow, or orange), then the display power up (two blue lights). Release the Vscan Air CL power button when you see the power up light.

3. Touch the first probe connector (Vscan Air CL shares the probe port menu with the CW-only probe) to start the pairing process.
4. The LED lights will enter the booting up state (two white lights blinking alternately), followed by the searching state (two white lights blinking synchronously). **Tap** the power button to allow connection during searching state.
5. Pairing will take 5-30 seconds to complete. When pairing is complete, the two LED lights will shine steady blue, indicating the connection is established. At the same time, the Vscan Air CL probe icon will appear on the ultrasound system Touch Panel (see *Figure 13-161 • Vscan Air CL Touch Panel Icon* on page 961), and the probe battery and Wi-Fi icons will also show on the ultrasound system main display (see *Figure 13-162 • Vscan Air CL Battery and Wi-Fi Icons* on page 962).




Figure 13-161 • Vscan Air CL Touch Panel Icon



- 1 Battery Level
- 2 Wi-Fi Signal Strength

Figure 13-162 • Vscan Air CL Battery and Wi-Fi Icons

Vscan Air CL Battery Indicator

Table 13-68 • Vscan Air CL Battery Indicator

Icon	Description
	Battery charged 90-100%
	Battery charged 60-80%
	Battery charged 40-60%
	Battery charged 20-40%
	Battery charged <20 - prepare to recharge the a batter.

Wireless Connection Quality Indicator

A wireless probe has a limited inherent risk of a disrupted connection due to various factors that could lead to loss of real time imaging. The 'Wireless Connection Quality Indicator' provides a visual indication of the connection quality between the probe and the ultrasound system during scanning. In *Figure 13-163 • Wi-Fi Connection Quality Indicator Icons - Left: Compromised, Right: Good* on page 963, the left image indicates a compromised Wi-Fi connection and the image on the right indicates a good Wi-Fi connection. An unstable connection may result in loss of image quality or slow image update during real time imaging.



Figure 13-163 • Wi-Fi Connection Quality Indicator Icons - Left: Compromised, Right: Good

Vscan Air CL Temperature Indicator and Thermal Management

When the Vscan Air CL operating temperature is increased, the thermal management system inside the probe may automatically decrease frame rates and/or reduce image width, to keep the probe temperature within optimal functional levels to support continuous scans up to 50 minutes. There are five thermal management levels (0-4). Level 0 is the initial state when starting with a cool probe, and at Level 4 the probe temperature reaches the maximal allowed level. At Level 4 a user notification will appear on screen and the probe will automatically shut down.

The 'Probe temperature indicator' tracks and displays changes in the operating temperature of the probe during scanning. Factors affecting probe temperature are:

- **Transducer:** The curved array transducer gets warm more quickly than linear array due to higher power consumption.
- **Preset:** Certain presets like Abdominal and Cardiac have higher power requirements, depending on the image settings, causing the probe to warm up faster than other presets.
- **Mode:** Operating in Color Flow mode warms up the probe faster than Black and white mode.
- **Length of scan:** The duration of continuous scanning.
- **Ambient temperature:** Higher ambient temperatures can cause the probe to warm faster.

NOTE

The probe temperature and related user notifications are independent of the probe battery status.



Figure 13-164 • Thermal Status Icons - Level 0 to 4

Scanning

To begin scanning with the Vscan Air CL:

1. Select the Vscan Air CL probe icon on the Touch Panel.




Figure 13-165 • Vscan Air CL Touch Panel Icon

2. Select the **Curved** or **Linear** tab on the Touch Panel.



Figure 13-166 • Vscan Air CL Curved and Linear Applications




Figure 13-167 • Vscan Air CL Curved and Linear Applications

3. Select a preset on the Touch Panel. The Vscan Air CL probe begins scanning and the system displays the images.




Figure 13-168 • Vscan Air CL Image Display on Ultrasound System



HINTS

The power button on the Vscan Air CL probe can be configured as a Freeze/Unfreeze button during scanning.

Preparing for a Guided Procedure with Vscan Air CL

A wireless probe has a limited inherent risk of a disrupted connection due to various factors that could lead to loss of real time imaging.



WARNING

If a temporary, unexpected disruption to real time imaging is determined to have a severely negative adverse effect on the patient's health, outweighing the benefits of using an ergonomic wireless probe at the point of care, it is recommended to consider using a wired ultrasound device for the specific procedure guidance.



CAUTION

When performing a guided procedure or a freehand biopsy (without a biopsy guide), it is the user's responsibility to use the appropriate equipment. Ensure that the needle (especially the needle tip) is always visible in the ultrasound image during the entire procedure.



CAUTION

Always use only B-Mode when performing guided procedures or a freehand biopsy.

Prepare Vscan Air CL probe and check Wi-Fi connection prior to a procedure

Prior to setting up for a guided procedure, it is recommended to prepare the Vscan Air CL probe to be in optimal condition, and if assess if a stable Wi-Fi connection can be maintained during the procedure. Follow the below steps for the assessment.

Use a Cool Probe

Ensure the probe is sufficiently cooled down after any previous scanning. Disruption or reduced scan quality may occur due to the probe overheating if a lengthy procedure is anticipated. Leaving the probe on a desk (outside the pocket or the case) after it is powered off will cool it down faster. It should take approximately 30-60 minutes to get to a reasonably cool state, depending upon how warm it was from the previous scan, and the ambient temperature. If accelerated cooling of the probe is required, place the probe, while turned off, in front of a fan, run under cold water, or apply a cooling pack.

Ensure the Probe is Sufficiently Charged

Ensure the probe battery is sufficiently charged before beginning a procedure. Battery levels of 50% and above are recommended before starting. The battery level of the probe can be checked on the top left of the imaging screen after connecting with the system. A green battery icon indicates sufficient battery level.



Figure 13-169 • Battery Icon - Fully Charged

Check Wi-Fi Environment

Whenever possible, check the Wi-Fi connection between the probe and ultrasound system in the environment where the procedure is expected to be performed. This will help detect any unexpected challenges before the actual procedure.

Select Small Parts from the Linear presets menu (Small Parts is the most Wi-Fi-challenging preset due to its high image data rate) and turn up the B-Mode gain. Ensure that the connection quality indicator is steadily green on the imaging screen. Observe the noise pattern, and the random movement of noisy pixels should not appear to pause occasionally.



Figure 13-170 • Wi-Fi Icon

The probe temperature indicator is also visible on the right upper corner of the image screen. A gray colored thermometer icon confirms a cool probe.



Figure 13-171 • Probe Temperature Indicator

Disconnect the Vscan Air CL Probe

To disconnect the Vscan Air CL probe from the ultrasound system, simply turn off the probe by pressing and holding the power button until the LED lights appear purple, which indicates the probe is powering down.

You can also shut down the Vscan Air CL probe by selecting the Vscan Air power button on the Touch Panel, or by selecting the battery icon on the title bar and selecting **Yes** in the Vscan Air Battery Icon dialog box.




Figure 13-172 • Vscan Air Power Button




Figure 13-173 • Vscan Air Battery Icon Dialog Box

Supported Features

Most features on the ultrasound system (measurement, annotation, image tag, etc.) are supported on the Vscan Air CL probe.

Table 13-69 • Features Available on the Vscan Air

Feature Name
Scan Mode: B Mode
Scan Mode: CF Mode
Overwrite and create application presets
Retain FOV
Easy3D
LOGIQ View
LOGIQ Apps
ATO (B Mode)
Auto Doppler (Colorflow Mode - Linear)
Rawdata (B & CF)

Feature Name
Scan Assistant
Compare Assistant
Measurement Package (by application)
Annotation package (by application)
Body Pattern (by application)
Imaging controls on Utility page
Imaging display (dual/quad, B CF simultaneous)
Print / Recall / change order
Patient page/Exam Category

Charger Cleaning and Disinfection

For approved charger cleaners and disinfectants, refer to "Cleaning and Disinfecting the System" in Chapter 12 "User Maintenance" of the LOGIQ Fortis User Manual 5855997-1EN.



CAUTION

Avoid using ALCOHOL (ISOPROPANOL) 70% on the Vscan Air probe holder and charger. ALCOHOL (ISOPROPANOL) 70% may compromise the durability of the probe holder and charger.



CAUTION

To prevent cross-contamination, clean and disinfect the Vscan Air probe holder and charger after every exam.

To clean and disinfect the Vscan Air probe holder and charger:

1. Remove the Vscan Air probe from the charger probe holder and place in a safe location.
2. Remove the clear probe holder from the charger (see *Figure 13-174 • Remove Vscan Air Probe Holder* on page 969).



Figure 13-174 • Remove Vscan Air Probe Holder

3. Clean the probe holder and the charger with an approved cleaning agent.
4. Disinfect the probe holder and the charger with an approved disinfectant; allow to dry for the manufacturer's recommended contact time.
5. Reassemble the probe holder on the charger.

Regulatory Requirements



CAUTION

The Qi compliant wireless charger supplied as an accessory with the product is verified for use with the Vscan Air CL probe. The wireless charger is considered to be information technology equipment that does not affect the basic safety or essential performance of the Vscan Air CL product. The wireless charger is compliant to the IEC/EN 62368-1 standard, which applies to audio/video, information and communication technology equipment.

The Vscan Air CL probe complies with regulatory requirements of the European Directive 93/42/EEC concerning medical devices.

NOTE

Regulatory information regarding the Vscan Air can be found from Utility > Admin > Vscan Air.

NOTE

All applications may not be supported with the different variations of the Vscan Air (e.g. Vscan Air SL – Sector/Linear). Check which VScan Air products are supported from Utility > Admin > Vscan Air.

Product-Vision Air Wireless Charger Model: VPC-10000 Input: 5V 2.1A



The ultrasound system does not control any acoustic output for the Vscan Air CL. For acoustic output reporting tables for the Vscan Air CL, refer to the manual provided by the

The ultrasound system does not control any acoustic output for the Vscan Air CL. For acoustic output reporting tables for the Vscan Air CL, refer to the manual provided by the manufacturer of the Vscan Air CL probe or on <http://gehealthcare.com/usermanual>.

NOTE

These acoustic output reporting tables are produced according to IEC 62359 Ed.2.

NOTE

The Acoustic Output tables are in English only.

Breast Productivity Package

Overview

There are three features in the Breast Productivity Package:

- **Breast Lesion M&A** — includes lesion measurement folders, show features, summary, etc.
- **Breast Measure Assistant** — contains the Auto Contour feature. It also has measurements related to Breast (distance to nipple, ratio).
- **Breast Assistant, Powered by Koios DS** — breast lesion analysis option.

NOTE

Koios DS is not applicable for China market.

Breast Lesion M&A

Breast Lesion M&A allows you to document up to 30 breast lesions for each breast. Lesion Height/Width/Length, Distance to Nipple and A/B Ratio are available. Distance to Nipple allows you to enter the value (this is not a calculated measurement).

ACR BI-RADS® lesion classification can be notated via Show Features and Show Assessment.

The Breast Measure Assistant (Auto Contour) feature can also be used to automatically detect and outline the breast lesion.

Worksheet and Summary Worksheets show all the documented right/left breast lesions.

From the Small Parts Application Preset, select the Breast Application. Next, select the Right/Left Lesion (Select RtSide/LtSide below the Touch Panel).

Table 13-70 • Breast Lesion M&A Touch Panel Controls

Preset Parameter	Description
Position	Specify the position of the lesion: Clock position 1-12 O'Clock, Areolar, SubAreolar, Axillary, or "-" (default).
Segment	Specify A, B, C, None, or "-" (default).

Preset Parameter	Description
Show Features	Press to activate the Show Features notations. To add notations for each feature, position the Trackball to the right of each feature and press Set. This brings up the available notations. Move the Trackball to highlight a notation and press Set to select a notation. The notation will then appear next to the feature. If a Feature has an asterisk next to it (*), then you can select multiple notations -- select all that apply and then select 'Done.' These features are displayed on the Trackball. Below is a list of each Feature with its possible notations: • Shape: Oval, Round, Irregular, None (-) • Orientation: Parallel, Not Parallel, None (-) • Margin: Circumscribed, Indistinct, Angular, Microlobulated, Spiculated, None (-) • Echo Pattern: Anechoic, Hyperechoic, Complex, Hypoechoic, Isoechoic, Heterogeneous, None(-) • Posterior Features: No posterior features, Enhancement, Shadowing, Combined Pattern, None(-) • Associated features: Architectural distortion, Duct changes, Skin thickening, Skin retraction, Edema, Absent, Internal vascularity, Vessels in rim, Soft, Intermediate, Hard, None(-) • Calcifications: Calcifications in a mass, Calcifications outside of mass, Intraductal calcifications, None(-) • Special Cases: Simple Cyst, Clustered microcysts, Complicated cysts, Mass in or on skin, Foreign body including implants, Lymph nodes-intramammary, Lymph nodes-axillary, Vascular abnormalities AVMs, Vascular abnormalities, Mondor disease, Postsurgical fluid collection, Fat Necrosis, None(-)
Show Assessment	Specify the ACR BI-RADS Assessment: None (-), 0, 1, 2, 3, 4a, 4b, 4c, 5, 6. A comment field is available directly below the ACR BI-RADS Assessment.
Return	Press to return to the previous Touch Panel.
Lesion #	Indicates which lesion you are viewing (Lesion # of Total Number of Lesions). Press the left/right arrow to move from lesion to lesion.
L	Lesion Length
H	Lesion Height
W	Lesion Width
Distance to Nipple	Used to manually enter the distance the lesion is from the nipple.
Auto Contour (HxL)	Press to activate the Auto Contour feature, using the height and length.
Auto Contour (HxW)	Press to activate the Auto Contour feature, using the height and width.
Rt or Lt A/B Ratio	Right or Left Lesion A/B Ratio, measured by Area or Diameter.
Composition	Specify the composition of the lesion: None (-), Homogeneous background echotexture-fat, Homogeneous background echotexture-fibroglandular, or Heterogenous background echotexture.
Delete Lesion	Press to delete this lesion.

Worksheet and Summary Worksheets

Worksheets and Summary Worksheets are provided for all documented Breast Lesions.



To move to the next page, select the **Page Change** control beneath the Touch Panel.

Figure 13-176 • Breast Lesion Worksheet and Summary

NOTE

Only defined features are displayed on the Summary. To display the undefined features, select **Show Undefined Features** at the bottom of the Summary Worksheet.

Breast Measure Assistant (Auto Contour)

You can request that the system trace/outline the border of a breast lesion using Breast Measure Assistant (Auto Contour). You do this by setting the Region of Interest (ROI) around the lesion; the system can then measure the lesion by drawing the contour around it.

To automatically detect the breast lesion on the display,

1. Press **Measure**.
2. Press **Auto Contour (HxW)** on the Touch Panel.
3. Place the Cursor in the center of the lesion and press **Set**. Size the ROI around the lesion. Use the Trackball to resize the ROI.
 - To increase the size of the circle, move the Trackball down and to the right.
 - To decrease the size of the circle, move the Trackball up and to the left.

NOTE

Include the entire lesion, even if additional surrounding tissue is included.

4. Press **Set** on the Trackball. A trace appears around the lesion.

NOTE

Multiple breast lesion traces may be generated by the system. To cycle through the generated contours, use the Select Contours rotary on the Touch Panel.

5. Inspect the generated contour for accuracy. If edits are necessary, execute steps 6-7 to edit the contour prior to accepting the measurement. Otherwise, skip to step 8.
6. To edit the selected contour, move the Trackball to appropriately size the edit region and then press **Set** on the Trackball.

7. The blue portion of the contour can be edited by moving the Trackball to the portion of the contour you want to edit.

NOTE

The Caliper closest to the cursor enables editing.

NOTE

To limit the horizontal/vertical editing capabilities, you can set a preset via **Utility > Measure > Advanced > Small Parts > Restrict Breast Contour Caliper Edit**.

8. After you have completed your edits, press **Done** on the bottom Set Key or press **Print** to accept the measurement.

Breast Assistant, Powered by Koios DS (not available in all countries)

Breast Assistant, Powered by Koios DS is a Breast Lesion Analysis Option. Koios DS is integrated with the LOGIQ Fortis via DICOM. Koios DS is configured similar to a DICOM Service. The user can accept/dismiss analysis results. If accepted, these results are included in the DICOM Structured Report.

To perform Breast lesion analysis using Koios DS:

1. In B-Mode in the Breast Application, select the Breast Productivity Package Lesion # then press **Koios** via the Touch Panel.




Figure 13-177 • B-Mode Touch Panel in Breast Application

NOTE

Koios DS analysis requires a frozen dual B-Mode image with a Length, Width, and Height measurement.

2. Measure the breast lesion (Length, Width, Height) over 2 orthogonal scan planes; or measure the lesion using Auto Contour.
3. Using the Trackball control, press **Analyze**.
4. Koios DS analyzes the lesion (a pop-up appears, "Analyzing").



Figure 13-178 • Koios DS Analysis

5. A pop-up appears with the result of the Koios DS risk assessment and its alignment to an ACR BI-RADS® or U1-U5 risk category.
6. If the assessment is accepted, it becomes part of the exam including being displayed on the exam's Summary Report and included in the DICOM SR data associated with the exam. If the assessment is dismissed, the results are not stored as part of the exam.

NOTE

For additional details of the Koios DS assessment, refer to Koios Medical's Koios DS product documentation.

Thyroid Productivity Package

Overview

A Thyroid Productivity Package is available.

NOTE

Koios DS is not applicable for China market.

Table 13-71 • Thyroid/Parathyroid/Lymph Node/Nodule Touch Panel Controls

Preset Parameter	Description
Side	Specify the side: Right, Left, Isthmus.
Worksheet/Summary	Select to view the Worksheet/Summary Worksheet.
Add#1, Add#2, etc.	Cycles through the available lesions, or adds a new lesion/node/nodule, etc.
Rt/Lt Thyroid Rt/Lt Parathyroid Rt/Lt/Isthmus Lymph Node Rt/Lt/Isthmus Nodule	To initiate a Left/Right Thyroid/Parathyroid or Left/Right/Isthmus Lymph Node/Nodule, select the corresponding folder on the Touch Panel. Length, Height, and Width are available for all thyroid measurements. The Cortical Thickness measurement is available for the Lymph Node. Show Features is available for all thyroid measurements.
Location	Parathyroid: Specify Upper Gland or Lower Gland Lymph Node: Supraclavicular fossa, Lower cervical, Middle cervical, Upper cervical, Parotid, Submandibular, Submental, Posterior triangle Nodule: - Location A: Upper, Lower, Mid, None - Location B: Lateral, Medial, Midline, None
Show Features - Overall Thyroid	Press to activate the Show Features notations. To add notations for each feature, position the Trackball to the right of each feature and press Set. This brings up the available notations. Move the Trackball to highlight a notation and press Set to select a notation. The notation will then appear next to the feature and on the Summary Worksheet. Below is a list of each Feature with its possible notations by measurement type: - Overall Thyroid (Top Level Touch Panel) - Resected: Totally, Partially, None (-) - Appearance: Within normal limits, Abnormal, Symmetric, Asymmetric R>L, Asymmetric L>R, None (-) - Comment


Preset Parameter	Description
Show Features - Lt/Rt Thyroid / Parathyroid and/ Lt/Rt/Isthmus Lymph Node / Nodule	Press to activate the Show Features notations. To add notations for each feature, position the Trackball to the right of each feature and press Set. This brings up the available notations. Move the Trackball to highlight a notation and press Set to select a notation. The notation will then appear next to the feature and on the Summary Worksheet. Below is a list of each Feature with its possible notations by measurement type: • Lt/Rt Thyroid • Resected: Totally, Partially, None (-) • Echogenicity: Homogeneous; Coarse; Heterogeneous; Hashimoto, Classic; Hashimoto, Probable; None (-) • Vascularity: Normal, Increased, Decreased, None (-) • Size: Normal, Enlarged, Small, None (-) • Comment • Isthmus Comment • Lt/Rt Parathyroid -- Upper/Lower Gland • Visibility: Visualized, Not Visualized, None (-) • Comment • Lt/Rt/Isthmus Lymph Node • Appearance: Within normal limits, Suspicious, Pathologic, None (-) • Composition: Cystic, Complex, Solid, None (-) • Vascularity: Normal, Increased hilar, Increased non-hilar, None (-) • Comment

Preset Parameter	Description
Show Features - Lt/Rt Thyroid / Parathyroid and/ Lt/Rt/Isthmus Lymph Node / Nodule	• Lt/Rt/Isthmus Nodule The selections for each category/feature and associated points match the ACR® TI-RADS™ (Thyroid Imaging Reporting & Data System) risk-stratification system as published in 2017 by ACR*. Once one or more values are assigned for each feature, the points are summed and the corresponding ACR TI-RADS level is determined. Additional information including feature assignment guidelines and criteria for fine needle aspiration or follow up ultrasound can be found at the American College of Radiology website at: https://www.acr.org/-/media/ACR/Files/RADS/TI-RADS/TI-RADS-chart.pdf?la=en • Composition (choose one): Cystic (0 point), Spongiform (1 point), Mixed cystic and solid (1 point), Solid (2 points), - • Echogenicity (choose one): Anechoic (0 point), Hyperechoic (1 point), Isoechoic (1 point), Hypoechoic (2 points), Very hypoechoic (3 points), - • Shape (choose one): Wider-than-tall (0 point), Taller-than-wide (3 points), - • Margin (if more than one type, choose the most suspicious one): Smooth (0 point), Ill-defined (0 point), Lobulated (2 points), Irregular (2 points), Extra-thyroidal extension (3 points), -. Note: If more than one type, choose the most suspicious. • Echogenic foci (choose all that apply): None (0 point), Comet (0 point), Macrocalcifications (1 point), Peripheral calcifications (2 points), Punctate echogenic foci (3 points), - • Comment • TI-RADS level: TR1, TR2, TR3, TR4, TR5 - TR1 - 0 points - Benign - TR2 - 2 points - Not Suspicious - TR3 - 3 points - Mildly Suspicious - TR4 - 4-6 points - Moderately Suspicious - TR5 - 7+ points - Highly Suspicious NOTE Although it is possible for a nodule to be awarded zero points and hence be characterized as TR1, all other nodules merit at least two points because a nodule that has a mixed cystic and solid composition (one point) will also gain at least one more point for the echogenicity of its solid component. • Koios DS Assessment * Tessler FN, Middleton WD, Grant EG, Hoang JK, Berland LL, Teefey SA, Cronan, JJ, Beland MD, Desser TS, Frates MC, Hammers LW, Hamper UM, Langer JE, Reading CC, Scoutt LM, Stavros AT. ACR Thyroid Imaging, Reporting and Data System(TI-RADS): White Paper of the ACR TI-RADS Committee. J Am Coll Radiol. 2017May;14(5):587-595. doi: 10.1016/j.jacr.2017.01.046. Epub 2017 Apr 2. PubMed PMID:28372962.

Preset Parameter	Description
Return	Press to return to the previous Touch Panel.
H	Height
W	Width
L	Length
Isthmus AP	Used to measure the Isthmus anterior to posterior distance.
Cortical Thickness	Cortical thickness of the lymph node.
Delete	Press to delete this anatomy.

Worksheet and Summary Worksheets

Worksheets and Summary Worksheets are provided for all documented Thyroid anatomies.



To move to the next page, select the **Page Change** control beneath the Touch Panel.

Figure 13-179 • Thyroid Worksheet

NOTE

Only defined features are displayed on the Summary Report. To display the undefined features, select **Show Undefined Features** at the bottom of the Summary Worksheet.

NOTE

To exit back to the previous measurement screen, press **Set** on exit.

NOTE

To exit back to the scan screen, press **Worksheet/Summary** on the Touch Panel.

Thyroid Assistant, Powered by Koios DS (not available in all countries)

Thyroid Assistant, Powered by Koios DS is a Thyroid Nodule Analysis Option. Koios DS is integrated with the LOGIQ Fortis via DICOM. Koios DS is configured similar to a DICOM Service. The user can accept/dismiss analysis results. If accepted, these results are included in the DICOM Structured Report.

To perform Thyroid nodule analysis using Koios DS:

1. In B-Mode in the Thyroid Application, select the Thyroid Productivity Package Nodule # then press **Koios** via the Touch Panel.




Figure 13-180 • B-Mode Touch Panel in Thyroid Application

NOTE

Koios DS analysis requires a frozen dual B-Mode image with a Length, Width, and Height measurement.

2. Measure the thyroid nodule (Length, Width, Height) over 2 orthogonal scan planes; or measure the nodule using Auto Contour.
3. Using the Trackball control, press **Analyze**.
4. Koios DS analyzes the nodule (a pop-up appears, "Analyzing").



Figure 13-181 • Koios DS Analysis

5. A pop-up appears with the result of the Koios DS risk assessment and its alignment to an ACR TI-RADS® risk category.
6. If the assessment is accepted, it becomes part of the exam including being displayed on the exam's Summary Report and included in the DICOM SR data associated with the exam. If the assessment is dismissed, the results are not stored as part of the exam.

NOTE

For additional details of the Koios DS assessment, refer to Koios Medical's Koios DS product documentation.

Start Assistant

Introduction

Start Assistant automatically saves exam settings as Exam Mappings when the first image is acquired during an exam. Saved Exam Mappings are automatically recalled when a Worklist item is loaded, keying off the Worklist Exam Description.

Start Assistant has two functional modes:

- On: Use Exam Description (default mode)
- On: Use Scan Assistant only




Figure 13-182 • Start Assistant Modes

NOTE

An Exam Description or Scan Assistant program entry must be present for Start Assistant to save the Mapping.

On: Use Exam Description Mode

In On: Use Exam Description mode, Start Assistant saves and loads the following exam settings:

- Exam Description - Populated from the Worklist item, the Exam Description is the default key by which the exam Mapping is saved.
- Category
- Scan Assistant
- Preset

- Probe

If a new Patient is entered manually (without using a Worklist), and no Exam Description is entered, Scan Assistant will become the key by which the Mapping is saved.

On: Use Scan Assistant only Mode

In On: Use Scan Assistant only mode, the exam description is ignored and the selection of the Preset and Probe is based on the Scan Assistant protocol only. The Exam Category and Scan Assistant program are not automatically selected and Start Assistant Editor does not show the Exam Description column. Manual entry of an Exam Description for table entry is prevented.

Saved Start Assistant Exam Mappings can also be added, edited or deleted with the Start Assistant Mapping Editor - *Start Assistant Mapping Editor* on page 983

NOTE

When using Start Assistant Mapping Editor in “On: Use Scan Assistant only” mode, the Exam Description field will not appear on the Start Assistant Mapping Editor, Add Start Assistant Mapping or Edit Start Assistant Mapping screens.

Start Assistant Mapping Editor

The Start Assistant Mapping Editor can be accessed from the **Utility > System > General** screen by selecting “Create/Edit” under the Start Assistant section (see *Figure 13-183 • Start Assistant Create/Edit* on page 983).




Figure 13-183 • Start Assistant Create/Edit

On the Start Assistant Mapping Editor screen you can Add, Edit or Remove a Start Assistant map (see *Figure 13-184 • Start Assistant Mapping Editor* on page 984).




Figure 13-184 • Start Assistant Mapping Editor

NOTE

When using Start Assistant Mapping Editor in “On: Use Scan Assistant only” mode, the Exam Description field will not appear on the Start Assistant Mapping Editor, Add Start Assistant Mapping or Edit Start Assistant Mapping screens.

Add a Start Assistant Mapping

Select Add from the Start Assistant Mapping Editor screen to add a new mapping (see *Edit a Start Assistant Mapping* on page 984).




Figure 13-185 • Add Start Assistant Mapping

Edit a Start Assistant Mapping

Highlight an existing mapping on the Start Assistant Mapping Editor screen and select Edit or double-click the item to edit a mapping (see *Delete a Start Assistant Mapping* on page 985).




Figure 13-186 • Edit Start Assistant Mapping

Delete a Start Assistant Mapping

Highlight one or more existing mappings and select **Delete** to delete.

Save and Exit Start Assistant Mapping Editor

Before exiting Start Assistant Mapping Editor, select **Save** to save any changes you have made.

Select **Exit** to exit Start Assistant Mapping Editor.

Scan Assistant

Introduction

Scan Assistant provides an automated exam script that moves you through an exam step-by-step. This allows you to focus on performing the exam rather than on controlling the system and can help you to increase consistency while reducing keystrokes. The system automatically invokes the correct mode and imaging parameters, advances to the next step in an exam, annotates the image, initiates measurements, and assigns the measurements to the worksheet/report.

NOTE

Scan Assistant Creator (off-board) only supports 64-bit Windows.

Availability

The following additional imaging parameters and preferences are available for use in a Scan Assistant program: Contrast, Contrast Clock, CW Doppler, Dual on Freeze, Depth, Color Scale, PW Doppler Scale, PW Sample Volume size, and Flow Model Selection.

You can initiate one or more manual Doppler measurements/calculations.

Body Patterns are available for use during a Scan Assistant program. You can turn a Body Pattern on/off, select a particular Body Pattern graphic, and specify the position of the probe mark on the Body Pattern graphic.

The footswitch can be used with Scan Assistant. You can map Pause/Resume, Previous Step, and Next Step to the footswitch.

The “Always Use Doppler Cursor” preset, available on the **Utility > System > General** page, allows all PW Doppler steps to start with full screen 2D image plus mode cursor. You can specify the Store Order in Scan Assistant to set the Reading Order for the radiologist. The Learn Probe attribute can be set to learn and change the probe for the user in the middle of the exam.

Scan Assistant Definitions

Scan Assistant definitions:

- Scan Assistant Manager. Available via the **Utility > Scan Assistant** page to import/export Programs created via the Scan Assistant Creator and to assign Programs to a user/exam category.
- Import. Used to load Programs created via the Scan Assistant Creator on to the LOGIQ Fortis.
- Export. Used to move Programs from one LOGIQ Fortis system to another LOGIQ Fortis.
- Scan Assistant Creator. Used to create Scan Assistant Programs.

Scan Assistant Description



Figure 13-187 • Scan Assistant Display Description

- 1 Program name, completed steps/out of total number of steps, and step description area.
- 2 Program step status (Complete/Incomplete), step number, step name. A checkmark indicates that this step has been completed. You can also manually check the box to bypass this step.
- 3 This column indicates the mode or when a measurement needs to be made.
- 4 This column indicates that the action moves the Program to the next step.
- 5 Active step The box is green when the program is active or yellow when it is paused.
- 6 Navigation: Stop, Pause, Pause/Resume. Edit (Pencil Icon). Also available via the left/right keyboard arrow keys. Stop also allows the program to be stopped, restarted, or a new program selected.

Setting up Scan Assistant

To set up Scan Assistant,

1. Import the Scan Assistant Program created using the Scan Assistant Creator or exported from another LOGIQ Fortis program.
Insert the media with the saved Program from the Scan Assistant Creator or exported program from another LOGIQ Fortis.
Press **Utility > Scan Assistant**.
Select Import from the Scan Assistant Manager page.
In the Source field at the top of the Import Programs pop-up, select the media that the Program is stored on.

Highlight the Program(s) to be imported. If a folder is highlighted, all programs in the folder are selected.

Select Import. The Program(s) you selected are stored to the LOGIQ Fortis. You can add it to the exam category and user.

2. Assign the imported Program to the exam category and user. Under Program Selections on the right-hand side of the Scan Assistant Manager page, specify the Exam Category and User for this Program. You can select All Users, or a specific user. If you specify All Users, all users will have the ability to use this Program while in the specified exam category, unless the user has his/her own list defined.
3. Select the imported Program from Available Programs -> Custom Programs on the left-hand side of the page. Then press the right arrow button to move the imported Program to the exam category and user selected above.
4. The Program list you created in **Utility > Scan Assistant** is visible in the Program field on the Patient menu. Pressing **New Patient** erases any patient data and Scan Assistant program you entered. First press **New Patient** and then set up patient data, Scan Assistant program, and finally, press **Register Patient**.

You can access the Scan Assistant Creator to edit the exam's program from the imaging display via the Creator Icon located at the bottom, left-hand corner of the Scan Assistant Program monitor on the display. You can activate the Scan Assistant Creator from the image screen, make edits, and then run Scan Assistant to test your changes.

NOTE

If you edit the program after you have already stored several images, and your edits change the number of program steps, you are prompted to Restart or Continue the Scan Assistant program.

NOTE

If you edit the program after you have already completed several steps, checkmarked steps remain checkmarked, even if you insert a new step between checkmarked steps. If this is not correct, you can edit the checkmarks or restart the program.

Using Scan Assistant

After you have set up Scan Assistant, the Program is active when you exit the Patient menu. The Program is located on the left-hand side of the display and as you can see in the example below, the annotation for the first step has been automatically noted on the image, ready for you to scan the specified anatomy.



Figure 13-188 • Scan Assistant Display

1. Follow the steps indicated in the Program: image/measure the appropriate anatomy.
2. Perform the indicated trigger to move to the next step in the Program.

NOTE

The footswitch can be used with Scan Assistant. You can map Pause/Resume, Previous Step, and Next Step to the footswitch.

3. To pause or unpause Scan Assistant, press the pause button on the display or press the left/right arrow on the keyboard.
4. To stop or restart a Program, press the Stop icon at the bottom of the Scan Assistant Program. A dialog pops up. This dialog lets you restart the current Program, start another Program, or stop Scan Assistant.
5. To skip a step or move to a certain step, press the up/down arrows on the keyboard or select the step you want to move to using the Trackball and Set keys.

Reference Images Overview

Reference images can be attached to a step of Scan Assistant protocol. The reference image is displayed in an extra window on the scan screen during scanning with the protocol.

Any image (JPEG/DICOM) can be attached to (or removed from) a protocol with Scan Assistant Creator.

Using Reference Images

A reference image window appears when a protocol with a reference image attached is active. The window is positioned over the preview window and can be moved anywhere on the screen (the system will remember the last image position).




Figure 13-189 • Reference Image Window

1. Minimize window by clicking minimize button in upper left window corner.
2. Change window size by dragging window corners.
3. Drag to move window anywhere on screen.

Inserting a Step

1. To insert a pre-defined step to the active Scan Assistant Program, press the Plus Sign (+) on the Scan Assistant Display.



Figure 13-190 • Insert Steps

The Insert Steps Pop-Up Menu appears.

1. Select the steps to be inserted via the pull-down Scan Assistant menu.



Figure 13-191 • Select from pull-down menu

1. Press Insert. All the steps in the inserted program are added immediately after the active step in the current program.

To undo inserted steps, use Edit button to edit the program and delete the recently inserted steps.

NOTE

After the exam is completed, the system removes the temporarily saved Scan Assistant program with inserted steps.

NOTE

The inserted steps is not adapt the annotation and body pattern from the active step.

Breast Measurement Auto Insert Steps

To add a program to be inserted whenever the Add Lesion key is selected:

1. Press the Plus Sign (+) on the Scan Assistant Display.




Figure 13-192 • Insert Steps

2. Select the "Add Lesion" tab.
3. Identify a breast lesion while using a breast scan assistant program. Freeze on the lesion and press "Add Lesion" on the measurement menu.
4. Complete the length measurement of the lesion. Unfreeze the image.
5. Upon unfreeze, the predefined program is inserted into the existing program and the next step is started.



Figure 13-193 • “BreastLesionAssess”

In this example, BreastLesionAssess contains four (4) steps, as indicated within the brackets on the figure above.

Exporting Scan Assistant Programs to Another LOGIQ Fortis

Exporting Scan Assistant Programs allows them to be imported to another LOGIQ Fortis or to be edited offline with the Scan Assistant Creator tool. To export a Program,

1. Insert the media to save the Program to.
2. Press **Utility > Scan Assistant**.
3. Select **Export** from the Scan Assistant Manager page.
4. In the Source field at the top of the Export Programs pop-up, select the media that the Program is to be stored on.
5. Specify the Program Directory using the drop-down menu if the desired Program Directory already exists on the media. If not, or if you want to export the Program to a new Program Directory, type a new Program Directory name in the field.
6. Highlight the Program(s) to be exported. If a folder is highlighted, all programs in the folder are selected.
7. Select **Export**. The Program(s) you selected are stored to the media. You can now import it to a new LOGIQ Fortis.

Scan Assistant Creator

Scan Assistant Creator is used to build customized Programs that can be imported onto the LOGIQ Fortis. These Programs automate many of the steps normally performed manually by

the user, thereby reducing the number of user actions and the amount of time to perform an exam.

The Scan Assistant Creator tool can be used both on the scanner and as an off scanner tool.




Figure 13-194 • Tool Layout Overview

- 1 Menu
- 2 Toolbars
- 3 Steps
- 4 Step Attributes
- 5 Rule Checking

Help

Help is available via the F1 key.

File Handling

When using Scan Assistant Creator off the scanner, it is very important to organize the programs in a way that will make it easy to import the programs onto the scanner. Each Program is a computer file. While these computer files can be copied, pasted and deleted like any other computer file, the Program files are only viewable using the Scan Assistant Creator.

File Extensions

Factory defined Programs have an .ep (exam Program) extension while user-defined Programs have an .uep (user exam Program) extension. Both factory and user-defined Programs can be read into the Scan Assistant Creator, but only user-defined Programs are

created. If a factory Program is read into the Scan Assistant Creator and then edited, it is saved as a user-defined Program.

Off-Scanner Directory Structure

The Scan Assistant Creator organizes the Programs in a directory structure that allows easy importing into the system. In order to be imported, all Programs must be stored in a LOGIQ_SCAN_ASSISTANT Programs Directory. Within this directory, one or more user-specified directories are created. Within each of these user-specified directories are the category directories (VAS, ABD, etc.) that hold the actual Programs.

The dialog in the figure below allows the user to specify the location of the LOGIQ_SCAN_ASSISTANT directory (root directory) and to either select an existing User Program Directory or create a new one.




Figure 13-195 • Directory Structure

Exporting Programs from the System

Factory or user-defined Programs on the system are easily exported for editing with the Scan Assistant Creator.

On the system:

1. Insert a USB storage device (or CD/DVD).
2. Select **Utility** -> **Scan Assistant**.
3. Select **Export**.
4. Select the media type and specify a directory. If a directory is specified that already exists, the Export adds the Programs along with any existing Programs. If the names of Programs are the same, use the resulting dialog to decide how to continue.
5. Select the Program to be exported and export them.

On the computer with the Scan Assistant Creator installed:

6. Insert the USB storage device (or CD/DVD) used above.
7. Copy the LOGIQ_SCAN_ASSISTANT directory from the USB storage device (or CD/DVD) to the hard drive. The hard drive directory that you copy to is the root directory. If you want to work with the Programs directly on the USB storage device, this step can be skipped.
8. Either open a Program by double-clicking it or selecting File '-> Open from the Scan Assistant Creator.

Importing Programs to the System

Programs created with the Scan Assistant Creator are easily imported to the system.

On the computer with the Scan Assistant Creator installed:

Copy the complete LOGIQ_SCAN_ASSISTANT directory from the computer hard drive to a USB device (or CD/DVD). The LOGIQ_SCAN_ASSISTANT directory needs to be at the top level (not in a subdirectory) on the USB device (or CD/DVD).

Table 13-72 • Directory Structure Example

LOGIQ_SCAN_ASSISTANT		
	MyUserNameDirectory	
		ABD
		CARD
		GYN
		OB
		PED
		SMP
		UR
		VAS

On the system:

1. Insert the USB device (or CD/DVD).
2. Select **Utility -> Scan Assistant**.
3. Select **Import**.
4. Select the media type.
5. Select the Programs to be imported and import them. If you attempt to import Programs that already exist with the same name, use the resulting dialog to decide how to continue.

Sharing Programs

To share a program with someone else, the file can be sent via e-mail as an attachment or copied onto a media. If the person receiving the program has the Scan Assistant Creator tool installed, open the file and use "Save As" to save it to an appropriate directory.

If the person receiving the program does not have the Scan Assistant Creator tool installed, the program can still be loaded onto a scanner by creating the following structure at the top level directory on a media device, copying the file to one of the category directories and then importing the protocol onto the scanner.

Table 13-73 • Media Directory Structure

LOGIQ_SCAN_ASSISTANT	
	User Program Directory (Any user name)
	Category Directories (e.g. ABD, CARD)

To share an entire portfolio of programs with someone else, the entire user program directory can be zipped. Make sure to set the options to include subfolders and to include relative path information. On the receiving end, the user can unzip the directory into a LOGIQ_SCAN_ASSISTANT directory.

Exporting the Scan Assistant Creator to a PC

To export the Scan Assistant Creator to a PC,

1. Insert a USB Flash Drive in a USB port on the Control Panel.
2. Press **Utility** -> **Scan Assistant**.
3. Press **Export**.
4. Place a checkmark in the Export Scan Assistant Creator Installation.
5. Press **Export**.

Creating and Editing Programs

To access Scan Assistant Creator on the LOGIQ Fortis:

1. Select **Utility** > **Scan Assistant**.
2. Select **Creator** on the Scan Assistant Manager tab on the monitor.

Creating New Programs

1. Select File -> New.
2. Before creating a New Program, select **Single Step** or **Multi Step** in the Toolbar.




Figure 13-196 • File Toolbar

3. Proceed to add/update your settings for the Step: Step Name, Instructions, etc.
4. Once finished, highlight the finished Step.
5. Select **Edit** -> **Copy**
6. In the Toolbar along the left, select **Insert Step Before Selected** or **Insert Step After Selected**.



Figure 13-197 • Insert Step

7. Highlight the copied step and proceed to edit accordingly.
8. Proceed to follow the same procedure to add more steps to your Program.
9. When done, select **Check** to verify your Steps.



Figure 13-198 • Rule Check and results area

10. The results are listed as to whether the Scan Assistant Rule Check Passed or if any Issues were detected. Issues found when running the check do not mean the Program is unusable.

NOTE

The rule check may report an unequal number of left and right steps. This may or may not be the expected result. If a change is made in response to the rule check results, a new rule check can be run to see if the issue has been resolved.

Editing Programs

When editing Programs, changes can be made at both the step level and the step attribute level. Steps can be added, inserted, moved, deleted, copied and pasted. Step attributes can be modified for a given step or across multiple steps.

Editing Steps

The step toolbar allows steps to be inserted, moved up and down, and deleted. For steps to be moved, one or more consecutive steps must be selected.

When the last step in a Program is selected, the Enter key automatically appends a new step to the end of the Program and selects the new step. When the Enter key is pressed on any other step, the next step is activated. The up and down arrows can also be used to move between steps.



- 1 Insert Step above selected step (Ctrl+I)
- 2 Insert Step below selected step
- 3 Move selected step(s) up (Ctrl+Up Arrow)
- 4 Move selected step(s) down (Ctrl+Down Arrow)
- 5 Delete selected step(s)
- 6 Edit Toolbar (Undo, Redo, Cut, Copy, Paste and Paste Special)
- 7 Edit Menu (same toolset as the Edit Toolbar)

Figure 13-199 • Step Toolbar and Edit Toolbar and Menu

When selecting multiple steps for Cutting or Copying, the Shift + Left Mouse and Ctrl + Left Mouse key combinations can be used.

Paste Special

The Paste Special control allows copied steps to be pasted with some modification. Select the desired Conversion and select **Paste**. An added feature to the Paste Special control is the Define Conversions function, which is used to define the text that is converted. An

example is shown in the figure below. If an exact case match is found, it is used for the conversion. If there is a match, but with a different case, it is used only if there is not an exact case match.

There are 3 user-defined conversions that can be edited and named. These user-defined conversions can also be used to perform a find and replace capability.




Figure 13-200 • Define Conversions Dialog

Editing Step Attributes

To edit a step attribute, select the step attribute and edit it, such as picking from a drop down menu, checking or unchecking a box, or typing in text. To edit multiple steps in a Multi Step view, select the step attribute of choice and do the following actions:

- Set the value of the step attribute as desired.
- Left click (and release) in the square in the bottom right-hand corner of the attribute.
- Drag to highlight the other steps to be changed in the same way.
- Left click (and release) again.

This will take the original step's content and copy it down the highlighted steps. This is also available when multiple attributes are selected within the same step.

To edit multiple steps in the single step view, highlight the multiple steps that you want to edit and then change the step attribute. If a step attribute is highlighted in green, this indicates that its current value varies across the selected steps.

If a step attribute is not editable, it may be because the attribute requires a different attribute to be set a particular way in order to become enabled. These dependencies are outlined below.

Table 13-74 • Step Attribute Dependencies

Step Attribute	Dependency
PDI	Color step attribute must be checked
Color / Dop Steer	Color or PW step attribute must be checked
Measure 1	Measure Trigger must not be set to None
Measure 2	Measure 1 must be set
Measure 3	Measure 1 and Measure 2 must be set

Editing the Current Program on the Scanner

If you are currently using a Scan Assistant program and choose to edit that program while using it (by selecting the pencil icon at the bottom of the Scan Assistant steps), the program will be reloaded when scanning is restarted. If the number of steps is changed, the checkmarks that were in place before editing are cleared. If the number of steps in the program has not changed, the checkmarks that were in place before editing are maintained.

The current program can be restarted at any time by selecting the Stop button on the Scan Assistant navigation window and selecting restart.

Opening Existing Programs

Multiple Programs can be open at the same time by selecting File -> Open. Each Program will open within the primary Scan Assistant Creator window. Finding the Program file (.ep or .uep) and opening the file automatically opens the file in the Scan Assistant Creator.

To switch between Programs, the title banner of the window is selected or the Program is selected from the Window Menu. An asterisk indicates that the Program has been edited but not saved.

With multiple Programs open, steps copied from one Program can be inserted into another Program via the paste or paste special features.

Saving Programs

Programs are saved via Save or Save As.

When Saving a Program, the Scan Assistant Creator provides an opportunity to run a rule check on the Program before saving.

NOTE

The name of a Program is appended with an asterisk (*) when the Program has been changed, but those changes have not yet been saved

Views


A Program is made up of a series of steps. Each step is made up of various step attributes. The step and step attribute data can be viewed in many ways using the Scan Assistant Creator. The different ways to look at the data are called Views.

Single Step Views

There are two Single Step views: Basic and All. The Basic view shows the most common attributes of the selected step. The All view shows all of the attributes of a given step.

For both views, the step names are shown on the left with the active step highlighted. The step attributes appear on the right and are separated into four groupings:

- General attributes at the top
- Imaging and Comment attributes on the left
- Measure attributes on the right.



- 1. Steps
- 2. Step Attributes

Figure 13-201 • Basic Single Step View

Multi Step Views

Multi Step views show certain step attributes for all the steps in a Program. There are six Multi Step views: General, Comment, Scan, Measure, Custom and All.



1. Steps
2. Step Attributes

Figure 13-202 • Multi Step Scan View

Customizing Multi Step Views and HTML Export

The contents of the Multi Step Views and the HTML Export is configurable via the Customize Menu or the small downward pointing arrow next to the word "Custom" on the View Toolbar Menu.




Figure 13-203 • Customize Menu

The Language selection allows the Scan Assistant Creator language to be configured.

The column widths of the steps and step attributes are customizable. The desired width is set by selecting and dragging the line separating column headers. These adjustments are remembered for the next time the Scan Assistant Creator is used.

The locations of the toolbars are customizable. The location is set by selecting and dragging the toolbar gripper as shown in the figure below. The toolbars can be placed at the top, left, right or bottom of the Scan Assistant Creator.




- 1 Toolbar Gripper

Figure 13-204 • Gripper used for Toolbar placement

Customize Multi Step Views Dialog Multi Step View

In the Customize Multi Step Views dialog, each tab represents a different Multi Step view. Within a tab, the checked boxes are the step attributes that are displayed in that Multi Step view. The views are independent of one another, so if a step attribute is desired in both views, it needs to be selected in both views.



HTML Export

The HTML Export feature allows a Program to be stored in a file format (*.mht) that is compatible with Windows Internet Explorer. This file is useful for printing the Program or viewing the Program, but is not useful for editing the Program. HTML Export is available via the File Menu.

An example of how a Program looks in HTML format is shown below.

#	Step Name	Optional	Admission On	Isolation/Route	Comment #1	Location #1	Comment #2	Location #2	Comment #3	Location #3	Color Mode	RF Trigger Mode	W Mode	Real	Barometric	Virtual Camera	External Beam	FDI	Color / Deep View	AB	LOGIQView
1	01 Thyroid Trans Med	No	Front		01 Thyroid Trans Med	Bottom Center			01 Thyroid Trans Med	Right Angled, Bottom Center	No	No	No	Off	Default	Default	Default	Default	Center	No	No
2	02 Thyroid Trans Med	No	Front		02 Thyroid Trans Med	Bottom Center			02 Thyroid Trans Med	Right Angled, Bottom Center	Yes	No	No	Off	Default	Default	Default	Default	Center	No	No
3	03 Thyroid Long Med	No	Front		03 Thyroid Long Med	Bottom Center			03 Thyroid Long Med	Right Angled, Bottom Center	No	No	No	Off	Default	Default	Default	Default	Center	No	No
4	04 Lymphatic	No	Front		04 Lymphatic	Bottom Center			04 Lymphatic	Right Angled, Bottom Center	No	No	No	Off	Default	Default	Default	Default	Center	No	No
5	05 Thyroid Trans Med	No	Front		05 Thyroid Trans Med	Bottom Center			05 Thyroid Trans Med	Right Angled, Bottom Center	No	No	No	Off	Default	Default	Default	Default	Center	No	No
6	06 Thyroid Trans Med	No	Front		06 Thyroid Trans Med	Bottom Center			06 Thyroid Trans Med	Right Angled, Bottom Center	Yes	No	No	Off	Default	Default	Default	Default	Center	No	No
7	07 Thyroid Long Med	No	Front		07 Thyroid Long Med	Bottom Center			07 Thyroid Long Med	Right Angled, Bottom Center	No	No	No	Off	Default	Default	Default	Default	Center	No	No
8	08 Thyroid Trans	No	Front		08 Lymphatic	Right Left	Right Angle, Transmed Left Lobe	Bottom Left		Bottom Left	No	No	No	Off	Default	On	Default	Default	Center	No	No

Keyboard Navigation

In addition to moving the windows pointer and selecting an item, there are several keyboard controls to help navigate through the Views.

Table 13-75 • Keyboard Program Navigation

Keyboard Entry	Step Selected	Step Attribute Selected
Enter	Moves to next step. If on the last step, it creates a new step and selects it.	Single Step View: Varies based on the step attribute selected. Multi Step View: Moves to same step attribute in the next step. If on the last step, it creates a new step and selects it.
Tab	Single Step View: Moves to next step. On the last step it moves to the first step attribute. Multi Step View: Moves to the first step attribute.	Moves to next step attribute. If last attribute for the step, moves to next step.
Alt+Tab	Single Step View: Moves to previous step. On the first step it moves to the last step attribute. Multi Step View: Moves to the last step attribute of the previous step.	Moves to next step attribute. If first attribute for the step, moves to previous step.
Up Arrow	Moves to previous step	Single Step View: Varies based on the step attribute selected. Multi Step View: Moves to same step attribute in the previous step.

Advanced Features

Keyboard Entry	Step Selected	Step Attribute Selected
Down Arrow	Moves to next step	Single Step View: Varies based on the step attribute selected. Multi Step View: Moves to same step attribute in the next step.
Left Arrow	No action	Single Step View: Varies based on the step attribute selected. Multi Step View: Moves to the previous step attribute.
Right Arrow	Single Step View: No action. Multi Step View: Moves to the first step attribute.	Single Step View: Varies based on the step attribute selected. Multi Step View: Moves to next step attribute.
Page Up	Scrolls to previous page of steps	Single Step View: Varies based on the step attribute selected. Multi Step View: Moves to same step attribute in the previous step.
Page Down	Scroll to next page of steps	Single Step View: Varies based on the step attribute selected. Multi Step View: Moves to same step attribute in the next step.

Scan Assistant Features

Scan Assistant allows the user to program the steps in an exam and to program certain attributes for each step. The attributes are what give the Scan Assistant Program behavior. The tables below provide the names of all attributes along with a description.

General Attributes

Table 13-76 • General Attributes

Attribute Name	Description
Store Order Specify	Used to enable the Store Order Definition dialog so that the Store Order can be set
Store Order	Specifies the Store Order number associated with the Step
Step Number	Number of the step that appears in the Scan Assistant Navigation menu
Step Name	Name of the step that appears in the Scan Assistant Navigation menu
Advance On	- Print: Advance to the next step and go live after Print / Image Store (e.g P1 key). This can be a single image store or a loop store. - Print & Unfreeze: Advance to the next step after Print / Image Store (e.g. P1 key) and unfreeze. This can be a single image store or a loop store. - User Selection: Advance to next step only after next step is manually selected (e.g. down arrow)
Instructions	User notes displayed in the Scan Assistant Navigation menu when the step is active

Attribute Name	Description
Optional	Optional: An optional step is given a check mark during Program execution even if no image is acquired Mandatory: A mandatory step is given a check mark only if an image is acquired for the step
Learn Probe	On or Off. Learn and change the probe for the user, when selected.

Comment Attributes

Table 13-77 • Comment Attributes

Attribute Name	Description
Comment 1, Comment 2	User annotation associated with the step. When editing in a Multi Step View, use Alt+Enter to create a new line.
Location 1, Location 2	Choose where the annotation is located on the image area for single or dual screens.
BP 1,2	Blank: Body Pattern not specified. Selected Body pattern graphic with or without probe position will be set, if selected
BP Specify	Used to enable the Body Pattern Selection dialog so that the Body Pattern graphic can be selected and probe position can be set
BP Clear	Clears BP 1, BP 2 defined for the step
BP Probe	BP Probe mark set by Scan Assistant, when selected
BP X/Y/Rot	Displays the X/Y/Rot information of the Probe mark set from Body Pattern Selection dialog

Imaging Mode and Imaging Preference Attributes

The probe and application associated with a program is not configurable. Instead, the scanner remembers the last probe and application used for a given Scan Assistant program and automatically selects them the next time the program is started.



Figure 13-205 • Imaging Mode and Imaging Preference Attributes

Table 13-78 • Imaging Mode Attributes

Attribute Name	Description
Color, PDI, MVI, PW, CW, M Mode, B Flow, B Flow Color, 3D, LOGIQView, Contrast, Contrast Clock, Elasto (Strain Elastography) and SWE (Shear Wave Elastography)	- On - when selected, the mode is on. - Off - when not selected, the mode is off.
Dual	- Off - Dual screen is not in use. - Left Active - Dual screen is active and the left image is the active image. - Right Active - Dual screen is active and the right image is the active image. - DualView (simul) - DualView is active (both left and right images are live).

NOTE

B Flow Color is not supported by the system. It still exists in Scan Assistant Creator because other products that offer B Flow Color use the Scan Assistant Creator program.

Imaging Preferences work slightly different than other attributes. For example, if an abdomen Program has 20 steps and all steps have the Harmonics attribute set to Default, then Scan

Assistant will not affect the harmonics setting. Now, assume that steps 10-12 are gallbladder steps and that the harmonics attribute has been set to on for these steps. When transitioning into this group of steps (step 9 to step 10), harmonics will be turned on (or remain on if it was previously on). If harmonics is then manually turned off in step 10 then Scan Assistant will not turn it back on when advancing to step 11. In other words, a group of consecutive steps with the same Imaging Preference are treated as a group by Scan Assistant and not as individual steps.

Table 13-79 • Imaging Preference Attributes

Attribute Name	Description
Harmonics	- On - when selected - Off - when unchecked - Default - Not specified so Scan Assistant does not set this attribute.
Virtual Convex	
CrossXBeam	
Color/Doppler Steer	- Left - Color/Doppler steered to the left - Center - Color/Doppler not steered - Right - Color/Doppler steered to the right
Flow Model	- Specified Flow Model is selected - Aorta, Renal, Penetration, Slow Flow, Med Flow, Fast Flow - Default - Not specified so Scan Assistant does not set this attribute.
Depth	2.0 to 36.0. - Default - Not specified so Scan Assistant does not set this attribute.
Color Scale	1 to 200. - Default - Not specified so Scan Assistant does not set this attribute.
Doppler Scale	
SV	1 to 16. - Default - Not specified so Scan Assistant does not set this attribute.
B-Flow Sensitivity	1.0 to 50. - Default - Not specified so Scan Assistant does not set this attribute.
B-Flow Color Scale	0.02 to 1.5. - Default - Not specified so Scan Assistant does not set this attribute.
Color Baseline	0 to 100. - Default - Not specified so Scan Assistant does not set this attribute.
Doppler Baseline	5 to 95. - Default - Not specified so Scan Assistant does not set this attribute.

Measurement Attributes



Figure 13-206 • Measurement Attributes

Table 13-80 • Measure Attributes

Attribute Name	Description
Side	When selected, the side measurement qualifier is set to: - Rt - Right side of the body - Lt - Left side of the body - None - Not used (neither Right nor Left) - Default - Side not specified so Scan Assistant does not set Side
Location	When selected, the location measurement qualifier is set to: - Prox - Proximal - Mid - Middle - Dist - Distal - None - Not used - Default - Location not specified. Scan Assistant does not set Location
Dop Vessel Assign	Various Doppler measurement Vessel folders Specifies the Vessel folder to assign auto calcs to. The assignment happens when the image is stored / printed (e.g. P1 key).
Auto Calcs	When selected, the Auto Calcs state is set to: - Frozen - Live - Off - Default - Auto Calcs state not specified. Scan Assistant does not set Auto Calcs state.
Auto Calc Params	- Various Auto Calc parameters - specifies the auto calc parameters to be used. - Default - Auto Calc parameters are not specified. Scan Assistant does not set the Auto Calc parameters.
Auto Calc Specify	Used to enable the Auto Calcs Parameter Selection dialog so that the Auto Calc Params attribute can be set
Auto Calc Default	Used to set the Auto Calcs Params attribute to Default.

Attribute Name	Description
Fetus	When selected, the fetus measurement qualifier is set to: • A - Fetus A • B - Fetus B • C - Fetus C • D - Fetus D • Default - Fetus not specified so Scan Assistant does not set Fetus
B/M Measure Trigger	When selected, the following action initiates the "Measure 1" attribute: • Measure Key • Freeze Key • Image Store - when the Measure key is manually selected or the image is stored. This is used to store / print an image and then measure on it and then store it again. Therefore, the Advance On Print attribute is ignored on the first store / print when the Measure Trigger attribute is set to Image Store. • None - Measurements are not triggered by Scan Assistant. The "Measure 1" attribute is ignored.
B/M Measure 1	Various 2D or M-Mode measurements Specifies the first 2D or M-Mode measurement to be initiated. The point at which the measurement is initiated is based upon the Measure Trigger attribute.
B/M Measure 2	Various 2D or M-Mode measurements Specifies the second 2D or M-Mode measurement to be initiated after the measurement associated with the Measure 1 attribute is completed.
B/M Measure 3	Various 2D or M-Mode measurements Specifies the third 2D or M-Mode measurement to be initiated after the measurement associated with the Measure 2 attribute is completed.
Double Print - Meas	• On - If an Image Store / Print (e.g. P1 key) is performed on an image with measurements, the image is stored / printed two times, once with the measurements and once without. • Off - No special Store / Print behavior.
Double Print - Color	• On - If an Image Store / Print (e.g. P1 key) is performed on an image with color, the image is stored / printed two times, once with color and once without. If double print on color and double print on measurements are both configured to be on, the image is stored / printed two times, once with the measurements and once without. • Off - No special Store / Print behavior.
View Worksheet	• On - The worksheet is turned on • Off - The worksheet is not turned on

Measurements

Because there are many measurements available on the LOGIQ Fortis and because the measurement package is highly customizable, there is some special handling for

Advanced Features

measurements. The Measurement attributes affected by this special handling are Measure 1-3 and Vessel.

Selecting a Measurement Package

The Measurement selection menu can be used to specify which measurement package is to be used for the Program. The Measurements packages are organized by category and subcategory. The choices for the Measure 1-3 and Vessel attributes are limited by the selection of category and subcategory. To select a subcategory, select a category, move to the subcategory list and select a subcategory. To select all subcategories for a given category, select the category and then reselect the Measurements menu item to remove the menu.

A single Program is not allowed to have measurements from multiple categories, but it may have measurements from multiple subcategories.




Figure 13-207 • Measurements Menu

User-Defined Measurements

User-defined subcategories and individual measurements can be used with the Scan Assistant feature. To do so, the Scan Assistant Creator needs to know about the user-defined measurements on the system.

On the system, use the Scan Assistant utility menu to Export Programs to a USB storage device (or CD/DVD). On the export menu, check the "Export user config data" checkbox to store the user-defined measurement information to the Program User Directory on the media. The name of the file is UserConfigSystemFile.res. When this file exists in the Program User Directory, then it is used. Otherwise, the default file installed with the Scan Assistant Creator is used.

Rule Checking

Scan Assistant Creator allows Programs to be checked. During a rule check, Scan Assistant Creator applies a series of rules against the Program being checked and reports any inconsistencies between the rules and the Program. This rule check is intended to find potential issues in the Program before it is tested on a LOGIQ Fortis. Issues found when running the check do not mean the Program is unusable. It means that if there happens to be a problem with running the Program, the first place to look would be at the Issue noted when you ran the Check.

For example, if there is a step name Right Kidney and the Measurement Location is set to "Left", the rule check would report this inconsistency.

Running a Rule Check

The "Check" button below the Program area is used to initiate a rule check. The results are displayed in the Rule Check Results window to the right of the button. A rule check is also initiated when attempting to save a Program that has not previously passed a rule check.



Figure 13-208 • Rule Check Button and Rule Check Results

Rule Check Results

If the issue is specific to a particular step, double-clicking on the issue number in the Rule Check Results window selects the step associated with the issue. The results are intended to be potential issues and therefore may be ignored at the discretion of the user. For example, the rule check may report that there are an unequal number of left and right steps. For a particular Program, this may be the expected result. If a change is made in response to the rule check results, a new rule check can be run to see if the issue has been resolved.

Compare Assistant

Overview

Compare Assistant allows you to perform serial scans on a patient when you can compare the images from a patient's previous exam(s) to the patient's current exam.

NOTE

In order to reload imaging parameters from a previous exam, the image must have been stored as raw data.

NOTE

Side-by-side scanning comparison is present whether or not you are able to import the scanning parameters.

In Comparison Mode the system can automatically reload scanning parameters from a previous exam performed on a archived as raw data and allow for side-by-side scanning for image comparison. This allows you to use consistent scanning parameters from exam to exam on the same patient and may assist in assessing the progress of the patient.

Compare Assistant is available in B-Mode, Harmonics, Contrast, and in Color Flow and Power Doppler Imaging Modes.

In B-Mode the following parameters can be transferred from the Compared Image to the Active B-Mode Image: Gain, Depth, Frequency, CrossXBeam, Virtual Convex (linear probes), SRI, Frame Averaging, Map, Dynamic Range, Acoustic Output, Harmonic state, Focal Zone Number and Position, Width, and Line Density.

In Color Flow/PDI Mode the following parameters can be transferred from the Compared Image to the Active Color Flow/PDI Mode image: Gain, ROI Size/Position, Frequency, Frame Averaging, Packet Size, Flow Model, Scale (PRF), Wall Filter, Spatial Filter, Acoustic Output, Invert, Threshold, Sample Volume, Color Map, Virtual Convex (linear probes), and Line Density.

Image parameter copy is supported for the following additional modes: Contrast (supports depth, focal zone position, acoustic output, and frequency), Elasto (supports ROI size, ROI position, frequency, and scale), and B-Flow (supports depth, frequency, focal zones, and sensitivity).

Annotations and Body Patterns can also be transferred from a patient's previous exam to the patient's current active exam.

Up to four (4) exams, three (3) other exams and the current active exam, can be compared. Each exam is displayed on the bottom of the display, with Tabs indicating each exam. Tabs are shown in chronological order with the most recent exam displayed on the left.

Compare Assistant is not supported with V Nav, 3D/4D, Quantitative Analysis, Stress Echo, or LOGIQView.

Only measurements taken during the active exam are transferred to the current exam's Worksheet; previous exam measurement information is not included on active worksheets.

In order to utilize the Compare Assistant tool, images should have RawDICOM data. The RawDICOM data is saved locally when available. Some scanning modes do not provide raw data. If the raw data associated with the comparison image cannot be displayed in dual

image format, the DICOM image data is used to display the image. RawDICOM data is typically removed when images are sent to PACS; therefore, images that are loaded using Query/Retrieve will likely be standard DICOM only.

NOTE

Images from some other 'unsupported' modes (PW Mode, CW Mode, and M-Mode) can be recalled in Compare Assistant, but the tool will not attempt to copy over imaging parameters unless it is one of the types listed above (it will still attempt annotation copy). Imaging parameter copy will also not be attempted unless the current probe is the same as the probe used to capture the original image and RawDICOM data for the original image is available.

Workflow example

NOTE

If you use DICOM images for comparison, check the "Compare" checkbox on the DICOM Image folder and press Compare on the Image History page.

1. Place the cursor on the Compare Assistant icon and press Set to activate Compare Assistant.
2. Place the cursor on the date tab of the desired comparison exam and press **Set**.
3. Place the cursor on the desired comparison image and press **Set**.
4. Start scan and freeze the image at the same position to make a comparison.

Use Copy Setting Icon as necessary. Comparison Image parameters transfer to the Active Image based on your preset (automatic or manual)

5. Place the cursor on the Compare Assistant icon and press Set to deactivate Compare Assistant.



Figure 13-209 • Compare Assistant Screen (example)

Advanced Features

1. Active Image
2. Comparison Image
3. Active Exam clipboard
4. Date tab of comparison exam
5. Comparison exam clipboard

Setting Up Compare Assistant

Compare Assistant allows you to set the following parameters:

- Side to place the compared image (left/right)
- Date settings for the compared image
- Imaging and Annotation parameters to be copied from the patient's Compared to the Current image
- Clipboard, Active Image, and Image History Label Layout and Color

You can set these parameters on the System Imaging page via **Utility > System > System Imaging**.



Figure 13-210 • Setting Up Compare Assistant Parameters

Table 13-81 • Compare Assistant Parameters

Parameter	Settings
Comparison Image Side	• Left • Right (Default)
Comparison Image Date	• All Dates (Default): The date is always displayed on the comparison image. • Different Date: Date only displayed when the date of the comparison image is different than the active exam date. • None: No comparison image date is displayed.

Parameter	Settings
Copying Settings	You can set the default Compare settings on the System Imaging Utility page or you can select the Compare setting via the On-screen controls. Comparison Image parameters transfer to the Active Image based on your preset (automatic or manual). Automatic Settings are copied from the Comparison Image to the current image as soon as the comparison image is loaded. - Automatic: Imaging & Annotations (Default): Imaging Parameters, Annotations, and Body Patterns copied - Automatic: Imaging Only: Only Imaging Parameters copied - Automatic: Annotation Only: Only Annotations and Body Patterns copied Manual Settings are copied from the Comparison Image to the current image when you manually select the control. - Manual: Imaging & Annotations: (Default) Imaging parameters, Annotations, and Body Patterns copied - Manual: Imaging Only: Only Imaging Parameters copied - Manual: Annotations Only: Only Annotations and Body Patterns copied Off: No parameters copied.
NOTE Copy settings are restored to the default whenever Compare Assistant is turned on and the active exam or current patient has changed since the last time it was on.	
Image Label Layout: Clipboard Active Images Image History	- No Label 1-Line Label (Default) 2-Line Label
Image Label Color: Clipboard Active Images Image History	- Bright/Soft White - Bright/Soft Yellow - Bright/Soft Red - Bright/Soft Orange - Bright/Soft Blue - Bright/Soft Purple

To set Compare Assistant print parameters, select **Utility > Connectivity > Button > Advanced**. You can set the system to print the Comparison image only, to store the New Image only; or to store both the Comparison and New Image.

Compare Assistant Controls


The following controls can be used while in Compare Assistant:

Table 13-82 • Compare Assistant Controls

Control	Description
	1. Image Display Area Icon 2. Worksheet 3. Active Image Screen 4. Compare Assistant On/Off. Select the Compare Assistant Icon to activate Compare Assistant. 5. Copying Settings 6. Save As Menu 7. Next/Previous Image, or Press [Ctrl]+[Next Image Arrow] to start a Slide Show.
Keyboard Arrow Keys	You can also use the Arrow Keys on the Keyboard to move to the next/previous clipboard image.
Operator Panel L/R Keys	If comparison mode is on and the single image display is active, to enter dual comparison display press the L (Left) or R (Right) key. If dual comparison display is on and the dual comparison display is active, to enter single image display, press the L (Left) or R (Right) key.
Freeze Key	To deactivate Comparison Mode, select the comparison image side while in Dual Comparison Display and press Freeze.
Print Key	When you store an image while in Dual Comparison Display, the system automatically switches to the active exam side prior to storing the image.

Table 13-83 • Compare Assistant icon and Copy Setting icon

	Compare Assistant On. Copy to Right.		Compare Assistant On. Copy to Left.
	Off		Off
	Automatic: Imaging and Annotations		Automatic: Imaging Only

	Automatic: Annotation Only		Manual: Imaging & Annotations
	Manual: Imaging Only		Manual: Annotations Only
	Off. No parameters copied.		Disabled.

NOTE

Compare Assistant icon is disabled when the system does not have Compare Assistant Option.

Comparison Clipboard

In Comparison Mode two (2) image clipboard areas display images from the active exam and the selected comparison exam. The Comparison Clipboard can display up to four (4) exams, including the active exam. Each exam has a tab that you click on to select the desired comparison or active exam. The active exam is always displayed (either on the Left or Right side, depending on your preset).



Figure 13-211 • Comparison Clipboard

The comparison exam(s) are identified by the date of the exam on the tab; the active exam is identified by the word, “Active” and by the asterisk (*). You can activate the Image History screen by clicking on the “...” tab.

Each image clipboard has its own scroll bar that can be used to scroll through all the images.

The highlighted image in the clipboard is the image that’s displayed along with the exam’s active image.

Query/Retrieve

You can use DICOM Query/Retrieve to search for a patient’s previous exams. When you select a patient on the Patient Screen, select the **Query** button. This automatically registers the patient (if needed) and initiates a query via the Data Transfer page. Select the exams you wish to be transferred and press the **Transfer** button.

After the retrieve completes, the system automatically returns to the Patient Screen with the same patient and exam active as when you selected to Query this patient.




Figure 13-212 • Patient Exam View

You can also push the data from the PACS device to the system.

Activating Compare Assistant

Compare Assistant displays an image associated with the active exam (live, frozen, or CINE) and the other side contains an image from the comparison exam.

Comparison Image parameters transfer to the Active Image based on your preset (automatic or manual). You can also use the Display Copying control adjacent to the clipboard to copy settings from the comparison image to the active image.

NOTE

If an imaging parameter is not available for the current probe, a message in the status bar indicates that this parameter is not available.

You activate Compare Assistant via the Comparison Mode Key adjacent to the Image Clipboard and On-screen Trackball Controls or you can activate Compare Assistant via the Patient Image History screen by selecting the image and pressing **Compare**.



Figure 13-213 • Image History

At the start of an exam the system automatically checkmarks the three (3) most recent exams, excluding active exams and exams without any images as the comparison exams. You can adjust this by placing a checkmark next to the exam you wish to include in Compare Mode (the system automatically deselects the oldest checked exam). To activate Comparison Mode, select which specific image you would like to compare to the current image from the clipboard on the Image History screen. Select the image and press the **Compare** button.

Exiting Compare Assistant

To exit Compare Assistant Mode, select the Comparison Mode Key adjacent to the Image Clipboard, or change the exam Application Preset, change active exam, change patient, or recall an image from the active clipboard.

OB Measure Assistant

Overview

The user can request the system to trace/outline the borders of specified OB measurements using the OB Measure Assistant feature. You can auto-detect the Biparietal Diameter, Head Circumference, Abdominal Circumference, Femur Length and Humerus Length.

To automatically trace the fetal anatomy on the display,

1. Press **Measure Assist Settings** on the OB Touch Panel to set OB Measure Assistant parameters.

Table 13-84 • OB Measure Assistant Touch Panel

Touch Panel Settings	Description
Default	When you first activate the Measure Assistant Touch Panel, the following measurements are highlighted: BPD, HC, AC, and FL. This means that OB Measure Assistant is active for all these measurements.
There are four (4) BPD Auto Custom Settings:	
1. BPD Highlighted	BPD and HC OB Measure Assistant is generated one at a time. BPD is auto generated--> Edit/Set--> HC is auto generated--> Edit/Set--> Done
2. BPD + AutoSet BPD	BPD and HC generated at the same time. BPD and HC are auto generated--> Edit/Set HC--> Edit/Set BPD--> Done.
3. BPD + BPD Only	Only the BPD is auto generated--> Edit/Set--> Done.
4. No Highlight	User must perform these measurements manually

2. Press the measurement (BPD, HC, AC, FL or HL). The trace is auto generated on the display.
3. To edit the selected measurement graphics:
 - Move the Trackball to appropriately size the edit region of interest
 - Edit the Calipers via Cursor Select.
 - Adjust the circumference using the Ellipse key.

Press **Set** on the Trackball to complete the measurement.

NOTE

Specify the mode to measure the BPD: Outer to Inner, or Outer to Outer via **Utility > Measure > Advanced > Obstetrics > Measure Assistant BPD Method**.

If the system was not able to detect the anatomy automatically, measurement calipers are displayed in the center of the screen in an editable state and a message is displayed on the status bar.

Auto Renal Measure Assistant



CAUTION

Auto Renal Measure Assistant is not applicable when measuring pediatric kidneys or renal transplants.

Auto Renal Measure Assistant maps Touch Panel selections to perform measurements based on pre-selected options.

NOTE

Prior to using, the Auto Renal Measure Assistant must be installed and activated.

NOTE

The maximum reduction in keystrokes (equaling time savings) is achieved while performing measurement(s) on a transverse renal view image using Auto Renal Measure Assistant when a **standard transverse renal view image is acquired with the renal hilum at a 90-degree angle.**

Availability

Auto Renal Measure Assistant is limited to processing the non-transplant and non-pediatric renal images scanned by the ultrasound system with the C1-6-D or C1-6VN-D probe using B-Mode in the Abdominal application, during a routine renal examination.

Auto Renal Configurations

Options can be chosen from the Auto Renal Option dropdown list on the **System Measure** screen (see *Figure 13-214 • Auto Renal Configurations* on page 1021).



Figure 13-214 • Auto Renal Configurations

The options under Auto Renal Configurations are:

- **L:** Selecting the RenalL button on the Touch Panel performs only the Renal Length Measurement on the Longitudinal Renal View.
- **LxH:** Selecting the RenalL or RenalH button on the Touch Panel performs both the Renal Length and Renal Height measurements on the Longitudinal Renal View.
- **LxH W:** Selecting the RenalL button or RenalH button on the Touch Panel performs both the Renal Length and Renal Height measurements on the Longitudinal Renal View at the same time. Selecting RenalW performs Renal Width measurements on the Transverse Renal View.
- **L HxW:** Selecting the Renal L button performs Renal Length measurements on the Longitudinal Renal View. Selecting the Renal H or Renal W button performs Renal Height and Renal Width measurements on the Transverse Renal View.
- **LxH HxW:** Selecting the Renal L or Renal H button performs Renal Length and Renal Height measurements on the Longitudinal Renal View. Selecting the Renal W button performs Renal Height and Renal Width measurements on the Transverse Renal View.

NOTE

If only the “Renal Length” measurement button is selected, only the RenalL measurement is taken.

Activate Measure Assistant

1. Select **Measure**.
2. Select Renal tab.
3. Select **Measure Assist** (see *Perform Measurements* on page 1023).




Figure 13-215 • Measure Assist

Perform Measurements

The options chosen in the Auto Renal Option dropdown list on the **System Measure** screen determine the measurements taken when selecting RenalL, RenalH or RenalW on the Renal measurement Touch Panel screen (see *Auto Renal Configurations* on page 1021).

NOTE

Measure Assist must be selected to use Auto Renal Measure Assistant functionality.

1. Select Renal Measurement. The system automatically places the caliper(s) based on the selection.
2. To edit the calipers, use Edit Caliper in the lower right corner of the image display (see *Figure 13-216 • Active Caliper and Edit Caliper Trackball Button* on page 1023). The active caliper is highlighted. Select the active caliper using the Edit Caliper trackball control, and use the main system trackball to change the active caliper position. (See the status bar messages for editing guidance.)
3. Finalize the measurement by selecting the Done Trackball button or the Print button.

The measurement is populated in the Worksheet.



Figure 13-216 • Active Caliper and Edit Caliper Trackball Button

- 1 Active Caliper
- 2 Edit Caliper Trackball Button

Summary of Results

1. The Longitudinal model for length measurements has average accuracy of 96.45% with 95% CI of $\pm 1.26\%$ and average absolute error of 0.35cm at 95% CI of ± 0.12 cm.

Advanced Features

2. The Transverse model for width measurements has average accuracy of 92.94% with 95% CI of $\pm 3.02\%$ and average absolute error of 0.38cm at 95% CI of ± 0.14 cm.
3. The Transverse model for width measurements has average accuracy of 93.13% with 95% CI of $\pm 3.63\%$ and average absolute error of 0.37cm at 95% CI of ± 0.14 cm.

Auto Preset Assistant

This AI-based framework recognizes some anatomy by analyzing live B-Mode images. The system then automatically adjusts the B-Mode model (user preset) or suggests the user change the preset manually. While live scanning, the system analyzes the B-Mode image at regular time intervals. When the anatomy is recognized for a specific model (user preset), the associated B-Mode model is enabled.

This feature works on L2-9-D, L2-9VN-D, ML6-15-D, L4-20t-D and L3-12-D Linear probes and supports nine B-Mode Anatomical areas: Abdomen, Breast, Carotid, Leg, MSK, Scrotal, Thyroid, Air and Other.



WARNING

This feature uses AI based computer algorithms to determine the anatomical area. There is an inherent risk that these algorithms may sometimes give suboptimal or incorrect results. Verify that automatically selected presets are appropriate for the scanned anatomy type.

NOTE

This view recognition algorithm was trained on 14,333 images and validated on 2,601 images which were independent data sets for the training algorithm.

Based on the validation test, Minimum Acceptable Algorithmic Performance was confirmed to 80% accuracy.

Auto Preset Assistant Configuration in Imaging Preset Manager

The Imaging Preset Manager can incorporate the Auto Preset selections, or manually added presets.

1. Select **Utility > Imaging Preset Manager > Auto Preset** in Category.
2. Available Imaging Presets on the left side includes all available presets (including the “Auto Preset” selections).




Figure 13-217 • Auto Preset Assistant Configuration

For example, in *Figure 13-217 • Auto Preset Assistant Configuration* on page 1025, if the confidence percentage result meets the MSK preset, the preset that is automatically selected

is “MSK Gen.” If no preset was defined, the system does not recommend or automatically change any presets.

In case that probes that do not support preset as factory default, but this feature supports, configuration assigned a different preset by factory default, tentatively.

They are customizable, but have limited functionality.

Auto Preset Assistant Configuration in System Imaging

1. Select **Utility > System > System Imaging** on the Touch Panel.
2. “Auto Preset Assistant” configuration column is shown in Utility System Imaging menu.
3. Turn on check box of Enable in Auto preset Assistant in Utility System Imaging menu.
4. See *Table 13-85 • Auto Preset Assistant* on page 1026 for detail parameters.



Figure 13-218 • Configuration for Auto Preset Assistant

Table 13-85 • Auto Preset Assistant

Parameter	Description
Enable	Select to enable the Auto Preset Assistant feature.
Default Auto Mode in new patient	<i>Off</i>: the system does not automatically change to the recommended preset. <i>Automatic</i>: the system automatically changes to the recommended preset when the confidence percentage is higher than predefined Threshold. <i>Air detected</i>: The system automatically changes to the recommended preset when the confidence percentage is higher than predefined Threshold and is immediately recognized as Preset of Air.
Turn off Auto after changing preset	If selected, once the preset changes, no automatic preset changes are allowed.
Automatically Retain Field of View	Selecting this ensures that the Imaging Parameters shown in <i>Table 10-154 • Retain Field of View</i> on page 598 stay constant over Probe and Preset changes by Auto Preset Assistant.

Check before automatic change	Requires confirmation before the preset changes.
Recommendation without specifying preset	If there are presets to recommend, the “Check your preset” message displays.
Show detected view	A tool displays the detected view result.

Auto Preset Assistant Configuration in User Defined Trackball Set Key Menu

In addition to making this function work automatically, it is possible to confirm the recommended preset using the Trackball Set key. These operations can be assigned to the trackball sub-set key in the User Configurable Key menu. The B-Mode preset can be changed to the recommended preset without moving hands from the operator panel.

1. Select **Utility > System > User Configurable Key** menu.
2. Select **Auto Preset** in the B/BF/Con category of the User Defined Trackball Set Key menu.




Figure 13-219 • User Defined Trackball Set key

Table 13-86 • User Defined Trackball Set key


Parameter	Description
AUTO	On: Automatically changes to the recommended preset.  Off: Does not automatically change to the recommended preset. 
Change	Displayed if there is a suggested preset change. If preset should change to the recommended preset, press <i>Change</i>

Using Auto Preset Assistant

1. Perform the required setup in the configuration menu.
2. Select the Linear probe and start a B-Mode scan.
3. Scan the same plane continuously for more than one second.
4. During live scan, Auto Preset Assistant uses computer algorithms to estimate which of the 9 B-Mode presets is closest to the current scan plane.
5. If the confidence percentage result exceeds the predefined Threshold of recommendation and the confidence level is different from the current preset, a message displays to press the Change key to change to the appropriate preset. The Change key is then assigned to the right side of the Sub-set key.



6. If the “Change” key is pressed while the above message is displayed, the system changes from the current to the proposed preset.
7. If the returned confidence result exceeds the predefined threshold of automatic change and inferencing is different from the current preset, a system message similar to “Changing to Thyroid Preset” appears. The preset change algorithm is automatically executed.



This feature only starts if certain conditions are met:

- Simplex B-Mode live images
- When there is a link between the preset and the confidence preset
- Depth > 2 cm
- Does not work with a contrast application
- Does not work in write zoom
- Does not work with Scan Assistant

This algorithm may display suboptimal or incorrect results, especially in the following cases.

- When the target or probe is moving significantly.
- In case of poor visualization due to probe contact, deep attenuation, artifacts, etc..
- When it is not a typical scan cross section. For example, including tumors, cysts, or diffuse changes.
- Affected by B-Mode display settings such as Compress and Gain, or user adjustments for post-processing.
- When multiple different organs coexist. Algorithms tend to make more important decisions based on image information close to the center.
- MSK anatomical area is limited to only scan views that contain bone structures of knee and elbow.
- LEG (Arm) anatomical area is limited to only scan views that contain vessel structures of Leg and Arm.

- Scan views including both Thyroid tissue and Carotid short axis are sometimes recognized as buffered anatomical area (Thyroid/Carotid Axial). For that model class, presets between Thyroid and Carotid will not change.
- The system does not actively change presets if a preset within same categories is already selected. For example, if MSK sup and UEA are selected, do not automatic change to MSK gen or LEA, respectively.

Auto Abdominal Color Assistant

This AI-based framework recognizes some anatomy by analyzing live B-Mode images. When the user enters Color Doppler, the system automatically adjusts to a flow shortcut based on a confidence level. While live scanning, the system analyzes the B-Mode image at regular time intervals. When the anatomy is recognized for a specific model (user preset) in B-Mode and Color mode is entered and updated, the associated flow shortcut is activated.

This feature works on C1-6-D, C1-6VN-D, C2-7-D, C2-7VN-D, C2-9-D, C2-9VN-D, C3-10-D, RAB6-D, L2-9-D, L2-9VN-D, L3-12-D, ML6-15-D and M5Sc-D probes supporting abdomen applications (Abdomen, Abdomen Detail, Abdomen Penetration, Abdomen Biopsy and Renal). The algorithm supports five abdominal areas: “Aorta,” “Liver/IVC/Spleen,” “Gallbladder/Urinary Bladder,” “Kidney” and “Pancreas.”



WARNING

This feature uses AI based computer algorithms to determine the abdominal model presets. There is an inherent risk that these algorithms may sometimes give suboptimal or incorrect results. Verify that automatically selected Flow models are appropriate for the scanned anatomy type.

NOTE

This view recognition algorithm was trained on 11,478 images and validated on 1,183 images which were independent data sets for the training algorithm.
Based on the validation test, the Minimum Acceptable Algorithmic Performance was confirmed to 80% accuracy.

Auto Abdominal Color Assistant Configuration for System Imaging

- 1. Select **Utility > System > System Imaging** on the Touch Panel.
- 2. “Auto Abdominal Color Assistant” configuration column is shown.
- 3. Turn on check box of Enable in Auto Abdominal Color Assistant.
- 4. See *Table 13-87 • Auto Abdominal Color Assistant* on page 1030 for parameter details.



Figure 13-220 • Configuration for Auto Abdominal Color Assistant

Table 13-87 • Auto Abdominal Color Assistant

Parameter	Description

Enable	To use this feature, turn on check box of Enable Auto Abdominal Color Assistant.
Auto Flow Model selection	Turn on check box, selecting recommended Flow model automatically at the timing of entering color mode.
Turn off Auto after manual Flow Model selection	If the Flow Model is selected by manual operation from the Touch Panel, automatic change function will be turned off tentatively. After selecting a new patient or probe, it becomes valid again.
Show detected view	A tool displays the detected view result.

Auto Abdominal Color Assistant Configuration in User Defined Trackball Set Key Menu

In addition to making this function work automatically, it is also possible to accept user confirmation operations based on the confidence results. These operations can be assigned to the trackball sub-set key in the Configuration Menu below. The Flow model can be changed while still touching the operator panel.

1. Select Utility > System > User Configurable Key menu.
2. Select “Auto Preset” check in CF/TVI/PDI category of User Defined Trackball Set Key menu.



Figure 13-221 • User Configurable Key for Auto Flow Model

Table 13-88 • User Defined Trackball Set Key

Parameter	Description
Reset CF	System turns off any Flow model. This can be easily reverted if the Flow model proposed by the algorithm and automatically modified does not meet expectations.
Change	Displayed if there is a proposed Flow model. Press to change to the proposed Flow model.

Auto Flow Model Selection Configuration in Imaging Menu

Each confidence preset can be freely associated with any of the four pre-prepared flow models. Each flow model name and parameters can be customized.

1. Select **Utility > Imaging > CF**-tab.
2. The left side of “Auto Preset - Flow Model” shows bundled view confidence preset as Aorta, Liver/Spleen/IVC, Renal, Gall/Urinary Bladder and Pancreas. Select the corresponding Flow Model from the column of right side.




Figure 13-222 • Configuration for Auto Preset Flow Model

Using the above sample settings as an example, when the confidence algorithm estimates the Aorta Preset, the “Flow Model 1” is automatically or semi-automatically selected by entering and updating color mode. For the Pancreas Preset (“-” is selected), the system does not recommend or automatically change any Flow model. For the Liver/Spleen Preset (“None” is selected), the system turns off any Flow model and changes to the default setting. The settings of these four pre-prepared flow models (Flow Model 1...4) can be confirmed and adjusted on the setting screen (see Figure 13-223 • Setting screen example - Auto Abdominal Color Assistant on page 1032).




Figure 13-223 • Setting screen example - Auto Abdominal Color Assistant

Using Auto Abdominal Color Assistant

1. Perform the required setup in the configuration menu.
2. Select a probe that supports abdominal applications and start the B-Mode scan.
3. Scan the same plane continuously for more than three seconds.
4. By entering color mode, the Auto Abdominal Color Assistant uses computer algorithms to estimate which of the 5 abdominal areas is closest to the current scan plane.
5. If the returned confidence result exceeds the predefined threshold of recommendation and inferencing is different from the current Flow model, a system message similar to “Recommend changing to Aorta flow model” appears on the bottom of the display (information area). The “Change” key is then assigned to the right side of the Sub-set key.
6. If the “Change” key is pressed while the above message is displayed, the system changes from the current to the proposed Flow model.
7. If the returned confidence result exceeds the predefined threshold of automatic change and inferencing is different from the current Flow model, a system message similar to “Changing to Aorta flow model” displays. The preset change algorithm is automatically executed.



This feature only starts if certain conditions are met:

- Using supported applications or an application created based on one of them.
- When color or PDI mode is activated from the B-Mode Live scan.
- When there is a link between the Flow model and the confidence preset.
- When there is live information of B-Mode for more than 3 seconds. This feature does not work if Color Flow is activated immediately after starting Auto Flow mode Model, or if the color mode is turned on/off in a shorter time frame.
- Does not work with a contrast application or other flow modes (only with Color and PDI).
- Does not work with Scan Assistant.

This algorithm may display suboptimal or incorrect results, especially in the following cases:

- When the target or probe is moving significantly.
- In case of poor visualization due to probe contact, deep attenuation, artifacts, etc.
- When it is not a typical scan cross section. For example, including tumors, cysts, or diffuse changes.
- Aorta anatomical area is limited only long view, not working for Transverse view.
- Affected by B-mode display settings such as Compress and Gain, or user adjustments for post-processing.
- When multiple different organs coexist. Algorithms tend to make more important decisions based on image information close to the center.
- The algorithm uses only information for the entire B mode, not the position of the color ROI or blood flow information. Therefore, depending on the location or size of CF ROI, algorithm may sometimes give suboptimal or incorrect results.

Hepatic Assistant

Overview

Hepatic Assistant is supported on C1-6-D and C1-6VN-D (Abd (Abdomen), AbdPen (Abdomen Penetration) and AbdDetail (Abdomen Detail)) to measure SWE and UGAP simultaneously.

See *Shear Wave Elastography* on page 824 and *Ultrasound-Guided Attenuation Parameter (UGAP) Option* on page 843 for detailed workflows.

Hepatic Assistant Availability

Hepatic Assistant is supported on Abdomen application of C1-6-D and C1-6VN-D (Abd (Abdomen), AbdPen (Abdomen Penetration) and AbdDetail (Abdomen Detail)) to measure SWE and UGAP simultaneously.

Activate Hepatic Assistant

1. Configure presets for Hepatic Assistant in **Utility > Imaging > UGAP > Abdomen/Abdomen Pen/Abdomen Detail** - C1-6 (Hepatic Assistant) if needed.
 - **Visualization**
 - UGAP/SWE (A/E)
 - UGAP Q/SWE (A Qual./E)
 - B/SWE & UGAP/UGAP A (**BE > A/Att.**)
 - SWE Q/SWE & UGAP Q/UGAP (**E Qual./E > A/Qual.**)
 - SWE Q/SWE & UGAP Q/UGAP (**Qual./E > Qual./Att.**)
 - **SWE Display Side:** Right or Left.
 - **Frames to store:** UGAP frames to store for Auto Measurement Multi can be configured in increments of 10 from 0 to 300 (0, 10, 20, etc.).
 - **UGAP ROI Display in PreMode:** On or Off.
 - **Pause before beginning UGAP acquisition:** The Pause functionality can be selected On or Off on the Touch panel prior to entering UGAP when using Multi-Measurement configuration.



Figure 13-224 • Presets for Hepatic Assistant

2. Assign Hepatic Assistant to any User Defined key in **Utility > System > User Configurable key**. Press **Save**.
3. Select C16-D/C1-6-VN-D probe and Abd, AbdPen or AbdDetail for application.
4. Press assigned User Defined key to display Pre Mode.




Figure 13-225 • Pre Mode

5. Move SWE ROI to desired position, if needed.
6. Press **Start** to begin SWE and UGAP live data acquisition.
The UGAP measurement is activated when the SWE measurement is completed.
 - The LOGIQ Fortis acquires one SWE frame data and one UGAP frame data for Auto Measurement Single.
 - The LOGIQ Fortis acquires SWE frames data until the user presses **Freeze**, and the specified number of UGAP data frames for Auto Measurement Multi.
 - The LOGIQ Fortis acquires SWE data frames until the user presses **Freeze**, and UGAP data frames until the user presses **Store** (when Auto Measurement is set to Off).
7. Press the assigned User Defined Key to exit Hepatic Assistant.

Visualization

The below Visualizations are selectable in Utility.

- Dual view of UGAP and SWE




Figure 13-226 • Example - UGAP/SWE (A/E)

- Dual view of UGAP Quality Map and SWE




Figure 13-227 • Example - UGAP Q/SWE (A Qual./E)

- Dual view of B-Mode and SWE to UGAP and UGAP Attenuation Map




Figure 13-228 • Example - B/SWE & UGAP/UGAP A (B/E " A/Att.)

- Dual view of SWE Quality Map and SWE and dual view of UGAP Quality Map and UGAP




Figure 13-229 • Example - SWE Q/SWE & UGAP Q/UGAP (E Qual./E " A/Qual.)

- Dual view of SWE Quality Map and SWE to UGAP Quality Map and UGAP Attenuation Map




Figure 13-230 • Example - SWE Q/SWE & UGAP Q/UGAP (Qual./E -> Qual./Att.)

EZ Imaging

Overview

EZ Imaging provides an efficient workflow to complete exams with minimal operator button input. EZ Touch Panel allows the operator quick access to change model, flow modes and Doppler modes without searching through multiple pages or many different parameters.

Quick Patient Change allows you to start a new exam without entering the patient page.

Activate EZ Touch Panel

1. Check **EZ Touch Panel Page** in Utility > System > System Imaging > EZ Settings.
2. Select “By Probe” (Default) or “By Category” for MyPreset Shortcuts.



Figure 13-231 • EZ Settings

3. EZ Touch Panel appears on the first page of B, flow and Doppler mode tabs.



Figure 13-232 • B-Mode - EZ Touch Panel (Example)

- 1 MyPreset Shortcuts - Change model quickly on the Touch Panel.
- 2 Image Shortcut - Change imaging frequency (Res > Gen > Pen). (Frequencies are not editable.)
- 3 Essential Control: Virtual Convex (or Max Angle), Advanced SRI
- 4 MyPreset Shortcut: Last used or current model displays in the 4th position.
- 5 Essential Control (Configurable) - Configurable essential controls display in the 3rd and 4th positions. A desired control can be assigned in the 3rd and 4th positions.



Figure 13-233 • CF/PDI/MVI/TVI - EZ Touch Panel (Example)

- 1 Flow mode buttons - Change Flow Technology quickly on the Touch Panel.
- 2 Flow Shortcut
 - CF/PDI: Displays available existing shortcuts.
 - MVI/TVI: No shortcut.
- 3 Essential Control

- CF/PDI/MVI: Virtual Convex, Invert, Simultaneous
- TVI: Map Compress and Map

4 Essential Control (Configurable) - Configurable essential controls display in the 4th and 5th positions. A desired control can be assigned in the 4th and 5th position.



Figure 13-234 • B-Flow - EZ Touch Panel (Example)

- 1Flow mode buttons - Change Flow Technology quickly on the Touch Panel.
- 2Flow Shortcut - Displays flow shortcut to change Sensitivity/PRI
- 3Essential Control - Virtual Convex, Advanced SR
- 4Essential Control (Configurable) - Configurable essential controls display in the in the 4th and 5th position. A desired control can be assigned in the 4th and 5th position.




Figure 13-235 • Doppler Mode - EZ Touch Panel (Example)

- 1Doppler mode buttons - Change Doppler Technology quickly on the Touch Panel (if both PW and CW are supported).
- 2Essential Control (Non-Configurable)

- Invert, Sweep Speed, SV Length (for PW)
- Wall Filter, Sweep Speed, Colorize (for CW)

3 Essential Control (Configurable) - Configurable essential control displays in the 4th position. You can assign a desired control in the 4th position.

4. Switch to conventional Touch Panel by selecting to turn the Touch Panel page.



Figure 13-236 • Switch between EZ Touch Panel and Conventional Touch Panel

MyPreset Shortcuts “By Probe”



Figure 13-237 • By Probe

- TOP 3 models on My Preset for the probe is displayed on the Touch Panel.
- * (Asterisk) indicates the preset appears on the MyPreset Shortcut.

MyPreset Shortcuts “By Category”



Figure 13-238 • By Category

- TOP 3 models in same exam category on My Preset for the probe is displayed on the Touch Panel.

Quick Patient Change

1. Complete current patient exam.
2. Press the BT key assigned to Quick Patient Change.
3. All Patient data is stored for the current patient and the Patient ID for the next patient is generated automatically.

NOTE

If you turn off Auto Patient Archive, a Confirmation Dialog for Data Archive appears. Select Store All, Delete All, or Cancel.



Figure 13-239 • Confirmation Dialog for Data Archive

4. An information dialog displays to inform the user that the patient has been created.



Figure 13-240 • Information Dialog

5. Start the exam for the new patient. The B-Mode screen displays with the default preset for the connected probe.

Anonymize the patient

Overview

The LOGIQ Fortis offers an option to extract all measurements and DICOM tags from a selected patient in the Patient List when that patient is not active. This option makes the exam(s) anonymous and attaches this information to a newly created Anonymous Patient.

The LOGIQ Fortis warns the user on possible data loss, data mismatch and data that cannot be copied anonymously, including:

- Image data may have patient information burned into the pixel data. This information will not match the updated patient data in the DICOM tags.
- Patient identification on image pixel data which the user annotated. This information cannot be removed.
- Links to Reports will be lost.
- Non-ultrasound exams will not be anonymous.

Anonymize the patient

1. Ensure that you are logged in as an Administrator on the system.
2. Select the Local Archive dataflow.
3. Select the patient from the Patient View list and left click to bring up the “Anonymize” pop-up menu.
4. A confirmation dialog displays. Select **OK**.
5. The “Anonymize Patient” dialog displays. All the fields inherit the values from the original patient’s exam, except for Patient ID.

Fill in the Patient information fields.

Press **OK** to continue.

NOTE

If the newly entered Patient ID is not unique in the database, the Patient ID turns to red and an error message displays on the status bar.

- **Empty Other DICOM Tags:** Check to clear all DICOM Tags.
- **Manage DICOM Tags:** Manage DICOM tags dialog displays. Select DICOM tags individually.




Figure 13-241 • Manage DICOM Tags

Report Writer

Introduction

The LOGIQ Fortis enables the generation of patient reports based on the examination performed and the analyses that were made during the exam. The reports are generated using the data stored in the system with pre-selected templates.

You may edit a report while performing the exam; customize, delete, or add measurements; and save changes until you use the Store command. Once Stored, the reports are read-only.

It is recommended that the data be saved often, and then carefully reviewed before the report is Stored. Use the worksheet to facilitate the review and adjust data before storing a report. The final report can be printed on a standard printer.

Creating a report

Reports summarize the data obtained in the examination. They can contain data, images, and cine loops.

Once generated, the report can be viewed, images can be added, and the patient's personal data can be modified. The examination data itself CANNOT be changed.

GE Medical Systems

Name : GE Test Patient ID : 123-4567 Date : 07/09/2004

Comments

Test exam

Figure 13-242 • Report Example

Activating the Report

1. Select **Reports** on the Touch Panel.
2. The system displays the default report for the current application on the monitor.

The information entered during the examination is automatically filled in the appropriate fields (e.g. demographic, diagnosis, comments).

The preview image appears when the cursor is over the clipboard image.

NOTE

The template is the skeleton of your report. It is composed of different objects that can be customized by the user.

3. Adjust the **Page Change** control to move down the page.



Figure 13-243 • Report Page Example

Table 13-89 • Report Button Controls

Button	Description
Print	Prints out the report to the default printer.
Store	Stores the report page into Archive as CHM* file.
Retrieve	Retrieves the report page from Archive. Stored Date/Time is appended to the name of stored report.
Template	Selects template from the list of selected applications.
Designer	Accesses template editor screen.
Save As	Exports the report page to storage media as CHM or PDF format.
Delete	Deletes the report page from Archive.
Worksheet	Accesses Worksheet Page.
Graph	Accesses OB Graph page (applies only to OB).
Anatomical Page	Accesses Anatomical Survey page (applies only to OB).
*CHM is a compressed HTML help file.	

Selecting another template

You can select another template for the current patient:

1. Select **TEMPLATE** at the bottom of the monitor display or the Touch Panel.
2. A list of available templates and exam categories (Abdomen, Adult, Carotid, etc.) displays.
3. Select the desired template using the **Trackball/Trackball** and press **Set**.
The selected template displays on the monitor.

NOTE

If you select another exam category, the template list of the selected category displays. Select the desired template.



Figure 13-244 • Available Template list

4. Select the desired template name and press **Set**.
5. The report changes to the selected template.

Factory Templates

The system has factory templates for each application. You can modify these templates or create user-defined templates. You need to save revised/new templates with unique names.

A template may include one or more of the following:

- Measurements
- Worksheet or Vessel Summary Images
- Anatomical Surveys or Biophysical Profiles
- Anatomical Graphics
- Graphs
- Images areas
- Score Boxes

NOTE

Additional factory templates can be added from the **Utility > Report** menu (OB for multiple gestation, Renal, etc.).

Editing a Report

Entering the hospital address

When using a factory template, the area for the hospital information is usually placed in the upper portion of the report.

To make a new area, see *Fixed Text* on page 1077 for more information.

To modify the factory template:

1. Select **Reports**.

2. Select **Designer**.
3. Double-click on the area for the hospital information in the template. The Fixed Text dialogue box displays.




Figure 13-245 • Fixed Text Dialogue Box

4. Make changes as necessary.
 - a. Enter the hospital address in the text area.
 - b. Modify Box Properties (box width, box border line width, text align, box height, box left margin, and font).
5. Select **OK**.
6. Save the template.

To keep the same template name:

 - Select **Save** from the File menu, and press **Set**. The Save Template dialog box opens.
 - Select **Yes**. The template retains the same name and appends “[user]”. For example, OB23-Basic[user].

To save the template with a new name:

 - Select **Save As** from the File menu, and press **Set**. The Save Template As dialog box opens.
 - Enter the name of the new template, and press **Set**. The template receives the new name and appends “[user]”. For example, NewReport[user].
7. Exit the Report Designer. The report with the hospital address displays.

Inserting the hospital logo

When using a factory template, the area for the logo is usually placed in the upper left portion of the report.

To make a new area, see *Fixed Text* on page 1077 for more information.

To modify the factory template:

1. Save the preferred hospital logo in either a jpeg or bmp format on a CD.

NOTE

Label the logo with a unique name (e.g. HospitalNameLogo.bmp).

2. Insert the CD into the CD/DVD drive.
3. Select **Reports**.
4. Select **Designer**.
5. Double-click the GE HealthCare Logo so that the frame is highlighted. The logo box displays.



Figure 13-246 • Logo Box

6. Select **Import Logo** (1). Select the CD drive first and then the hospital logo.
7. Select **OK**. The hospital logo displays in the logo list (2). Click the logo to select.

NOTE

Scroll the logo list using the left/right arrow key (3).

8. Modify Appearance (4).
9. Select **OK**.
10. Save the template.

To keep the same template name:

- Select **Save** from the File menu, and press **Set**. The Save Template dialog box opens.
- Select **Yes**. The template retains the same name and adds “[user]”. For example, OB23-Basic[user].

To save the template with a new name:

- Select **Save As** from the File menu, and press **Set**. The Save Template As dialog box opens.
- Enter the name of the new template, and press **Set**. The template receives the new name and adds “[user]”. For example, NewReport[user].

11. Exit the Report Designer. The template with the hospital logo displays.

NOTE

If a different logo prints on the report, rename the logo image which you want on the report and insert it into the report template again.

Changing the Archive Information

When using a factory template, the Archive Information is usually placed below the hospital name and logo.

The contents of the Archive Information is inserted through the related page automatically. If you want to change the description, such as Information or Comments that was entered in the patient menu:

1. Select the yellow text to be changed, e.g. Information or Comments.
The area where the description was entered (e.g. Patient menu) displays.
2. Change the existing data as necessary.
3. Select **Report** to return to the report.




Figure 13-247 • Patient Information Area (Example)

Modifying the displayed objects of Archive Information

1. Select **Designer**.
2. Double click on the area for the Archive Information in the template. The Archive Information Box displays.



Figure 13-248 • Archive Information Box

3. Click the checkboxes to select and deselect the objects.
Objects with checkmarks will appear in the report template.
4. Select Box Properties to change the font, font size, font color, or box size, and select **OK**.
5. Select **OK** to return to the Report Designer.
6. Save the template.

To keep the same template name:

- Select **Save** from the File menu, and press **Set**. The Save Template dialog box opens.
- Select **Yes**. The template retains the same name and adds “[user]”. For example, OB23-Basic[user].

To save the template with a new name:

- Select **Save As** from the File menu, and press **Set**. The Save Template As dialog box opens.
- Enter the name of the new template, and press **Set**. The template receives the new name and adds “[user]”. For example, NewReport[user].

7. Select **File > Exit** to leave the Report Designer.

Entering free text

You can enter free text to the report using the alphanumeric keyboard.

The factory template terms their text area as “Summary or Comments”.

1. Move the cursor to the text field and press **Set**.

NOTE

You can enter the text only to the field set as free text in the Report Designer.

NOTE

DO NOT enter “%s” in a free text field and then try to edit/save the template in the Report Designer.

2. Type the text.

Inserting Text

1. Select **Designer**.
2. Move the cursor where the text is to be inserted and press **Set**.
3. Select the Text Field from the Insert menu. The Text Field dialog box displays.




Figure 13-249 • Text Field Dialogue Box

4. Select the appropriate display items:
 - Ref. Reasons: Retrieves this information from the Direct Report
 - Comments: Retrieves this information from the Comment field of the patient screen and the Exam Comment field of the worksheet.
 - Diagnosis: Retrieves this information from the Direct Report
 - Free Text 1 - 8
5. Type the heading Text.
6. Modify box properties, the heading text and font, and data.
7. Select **OK** or Cancel.
8. Save the template.

To keep the same template name:

- Select **Save** from the File menu, and press **Set**. The Save Template dialog box opens.
- Select **Yes**. The template retains the same name and adds “[user]”. For example, OB23-Basic[user].

To save the template with a new name:

- Select **Save As** from the File menu, and press **Set**. The Save Template As dialog box opens.
- Enter the name of the new template, and press **Set**. The template receives the new name and adds “[user]”. For example, NewReport[user].

Inserting an image to the report

Some factory templates include an image area. If you want to insert or modify the image area, see *Image Display Fields* on page 1071 for more information.

To insert images from clipboard into the image field of the report:

1. Move the cursor to the desired image on the clipboard.

NOTE


The preview image appears when the cursor is over a clipboard image.

2. Press and hold down the **Set** key and drag the selected image to the report by using the **Trackball** or double click the **Set** key on the desired image.
3. To move images between image areas, press and hold down the **Set** key and using the **Trackball**, drag the selected image to the new location.
To remove an image from the report, press and hold down the **Set** key and using the **Trackball**, drag the select image back to the clipboard.

Measurement result section

Measurement results for the current patient display automatically if you have the measurement section in the report template.

The factory template has an appropriate measurement result area. If you want to insert or modify the measurement area, see *Measurements* on page 1075 for more information.



Measurements	
UL-L	7.87 cm
UL-H	4.56 cm
UL-W	4.50 cm
UL Vol	85.71 ml
Rt Ov-L	5.88 cm
Rt Ov-H	4.82 cm
Rt Ov-W	3.10 cm
Rt Ov Vol	48.87 ml
Lt Ov-L	4.95 cm
Lt Ov-H	4.22 cm
Lt Ov-W	2.98 cm
Lt Ov Vol	32.65 ml
Rt Fe-D1 01	6.00 cm
Rt Fe-D2 01	5.15 cm
Rt Fe-D3 01	2.99 cm
Rt Fe-Mean 01	4.72 cm
Rt Fe Vol 01	48.44 ml
Lt Fe-D1 01	1.56 cm
Lt Fe-D2 01	1.12 cm
Lt Fe-D3 01	0.80 cm

Figure 13-250 • Measurement section

Inserting the worksheet

You can insert the worksheet (like you can insert an image) to the image display field. To set an image display field in the report template, see *Image Display Fields* on page 1071 for more information.

1. Display the worksheet on the monitor display.
2. Save the worksheet using the **Print** key.
3. Press **Report**.
4. Drag the worksheet into the report.
 - a. Move the cursor to the desired worksheet on the clipboard.
 - b. Press and hold down the **Set** key. Use the **Trackball** to drag the selected worksheet into the Image Display Field.
 - c. Release **Set**.

NOTE

You can also move the cursor to the desired worksheet on the clipboard, double-click the worksheet, move the cursor to the Image Display Field, and select **Set**.

5. The worksheet displays on the report.

NOTE

You can double-click the worksheet in the report to change the background color to white to save ink during printing. Double-click the worksheet again to return the worksheet to the original color.

Placing objects side-by-side

If you want to place images, the image and comment, anatomical graphic and comment, etc. side-by-side, you must first place a table, which has two (or more) columns, into the report template.

1. Select **Report**.
2. Select **Designer** to display the Report Designer.
3. Place the cursor where you want to insert the object.
4. Select **Table** from the Insert menu. Insert Table box displays.



Figure 13-251 • Insert Table Box

5. Set the number of columns to 2 (or more, as required) and change the table parameters, if needed. Select **OK**.

NOTE

If you do not need a table border, set the Border to 0. Add additional rows if required.

6. Place the cursor in the column and select the desired items from the Insert menu (e.g. logo, image, free text). Specify those items.
7. Repeat step 6 for each column as required.
8. Save the template.

To keep the same template name:

- Select **Save** from the File menu, and press **Set**. The Save Template dialog box opens.
- Select **Yes**. The template retains the same name and adds “[user]”. For example, OB23-Basic[user].

To save the template with a new name:

- Select **Save As** from the File menu, and press **Set**. The Save Template As dialog box opens.
- Enter the name of the new template, and press **Set**. The template receives the new name and adds “[user]”. For example, NewReport[user].

You can insert the images in the order preferred, by row or column, on the factory templates.
Inserting the Table on page 1066

Accessing Worksheet, OB Graph and Anatomical Survey Pages

If the Worksheet, OB Graph, and/or Anatomical Survey pages have been saved for the current patient, you can access these pages from the report page.

NOTE

OB Graph and Anatomical Survey pages applies to OB, GYN and Abdomen.

1. Select either **Worksheet**, **Graph** or **Anatomical Page** on the Touch Panel.

NOTE

There is also Fixed Text set up as hyperlinks for these pages. Cursor to the fixed text and press **Set**.

2. The system displays the appropriate page (Worksheet, OB Graph or Anatomical Survey) with the corresponding Touch Panel.
3. Select **Report** to return to the Report page.

Storing the Report

1. Select **Store**.
The Report is saved as a CHM file to Archive.

NOTE

The archived report cannot be edited; therefore, it is recommended that the data is carefully reviewed before the report is saved.

Retrieving an Archived Report

1. Select **Retrieve**. The Retrieve menu displays.




Figure 13-252 • Retrieve Menu (Prefix "User1\" may not appear)

2. Select the desired report and press **Set**.

NOTE

The retrieved report cannot be edited.

Deleting a Report from Archive

1. Select **Delete**. The Retrieve menu appears on the screen.



Figure 13-253 • Delete Reports Menu (Prefix "User1\" may not appear)

2. Select the report to delete and press **Set**.

Printing the Report

To preview the Print Layout before printing, see *Preview the Print Layout* on page 1064 for more information.

NOTE

Double-click the worksheet and/or image in the report to change the background color to white. A white background will save ink during printing. Double-click the worksheet or image again to return to the original color.

1. Select **Print** to print out the report.
The Report is printed on the default printer.

Exporting the Report to Media

1. Select **More**.
2. Select **Save As**.
The Save As dialog box appears on the screen.
3. Enter the Report title and select the file format.
4. Select the media to export the Report. The system supplies a name (numeric DICOM UID, unique identifier).



Figure 13-254 • Save as Dialog Box

5. Select **Save**.

Exiting the report

1. Select **Store** to save the report.

NOTE

If the user is working on a report and leaves the report screen for any reason, all information added to the report is automatically saved without loss of data.

2. Select another key to close the report page.

Designing Your Own Template

Template Designer

You can design and create your own customized template from a blank template page, or you can use an existing template (factory or user-defined) and save the changes.

Display the desired template and select **Designer** to open the Template Designer page.

The screenshot shows a software interface for a Report Designer. At the top, there are fields for 'Hospital name', 'Address', 'City/State/Zip code', and 'Phone'. Below this is a section titled 'Patient Information' containing fields for 'Name', 'Birthdate', 'Ref. Physician', 'Patient Id', 'MRN', 'Ref. Physician', 'Date', 'GR', 'DOB', and 'Operator'. There are also sections for 'Indications' and 'Comments'. At the bottom, there is a 'Measurements' section with a table structure for data entry.

Figure 13-255 • Report Designer

File Menu

Table 13-90 • File Menu

	Description
New	Creates a new template. A blank template appears.
Save	Overwrites the existing template.
Save As	Saves as a new name.
Page Setup	Enters Print Layout screen.
Print Preview	Executes print preview.
Exit	Exits Report Designer page.

Create a new template

To design a new template without using a pre-existing factory template:

1. Select **Designer** to open the Report Designer.
2. Select **New** from the File menu, and press **Set**.
The blank template displays.
3. Create the report template as needed.
4. Select **Save** from the File menu, and press **Set**.
The Save Template As dialogue box displays.
5. Enter a template name and click **OK**.
6. To exit Report Designer, select **Exit** from the File menu, and press **Set**.
 - Yes: Saves changes and exits Report Designer.
 - No: Does not save changes and exits Report Designer.

- Cancel: Returns to Report Designer.

Create a new template and save as a factory template name

To design a new template by modifying an existing factory template and keeping the same name of the factory template:

1. Select and display the existing factory template.
2. Select **Designer** to open the Report Designer.
3. Modify the report template as needed.
4. To save changes, select **Save** from the File menu, and press **Set**.
The Save Template dialog box displays.
 - Yes: Saves changes.
 - No: Does not save changes.
 - Cancel: Returns to Report Designer.

NOTE

The template name displays in the template list, retains the same name, and adds “[user]”. For example, “OB23-Basic[user]”. You do not lose the original factory template.

5. To exit Report Designer, select **Exit** from the File menu, and press **Set**.
 - Yes: Saves changes and exits Report Designer.
 - No: Does not save changes and exits Report Designer.
 - Cancel: Returns to Report Designer.



HINTS

Save changes frequently as you modify your template. Saving often reduces the risk of losing all your changes.

Create a new template and save with a new name

To design a new template by modifying or copying an existing factory template and saving it with a new name:

1. Select and display the existing factory template.
2. Select **Designer** to open the Report Designer.
3. Modify the report template as needed.
4. Select **Save as** from File menu and press **Set**.
The Save Template As dialog box displays.
5. Type the new template name and click OK.
6. Select **Exit** from the File menu and **Set**.
7. The Report Designer closes and returns to the Report Page.

NOTE

The template receives the new name and adds “[user]”. For example, NewReport[user].

Page Setup

1. Modify the factory template as necessary in **Designer**.
2. Select **Page Setup** from File menu and press **Set**.
3. Change the paper size or orientation to fit the print layout, as necessary.

To define the header and footer for the printed report, type text and enter the required variables listed in the table below. Select "Different for first page" and enter a specific header/footer for that page.

Table 13-91 • Variable and Definition

Variable	Definition	Variable	Definition
{pid}	Patient ID	{prt}	Current time (printing time)
{pnm}	Patient name	{cp}	Current page
{pbd}	Patient date of birth	{tp}	Page count
{exd}	Examination date	{c}	Subsequent text is centered
{prd}	Current date (printing date)	{r}	Subsequent text is right aligned.
{inm}	Institution name		

NOTE

Default is left aligned. Report will appear as black ink on white background.

4. Select **OK** or Cancel.



Figure 13-256 • Page Setup with Header Example

Preview the Print Layout

1. Select **Template** to choose the Report Template.
2. The Print Preview screen displays.

If changes need to be made, select **Close** to exit the Preview page. Modify the template or return to the Report and modify the contents.

Edit Menu

Table 13-92 • Edit Menu

	Description
Delete	Deletes the selected object from the report template.
Undo	Restores the previous state(s) of the report template.

Deleting a template object

1. Select the object to be deleted.
2. Select **Delete** from the Edit menu, and press **Set**. The object is deleted from the template.

Undoing the operation

1. Select **Undo** from the Edit menu, and press **Set**.
2. Repeat as required.

Insert Menu

Table 13-93 • Insert Menu

	Description
Page Break	Inserts a Page Break.
Table	Inserts a Table.
Logo	Inserts a Logo Bitmap File.
Archive Info	Inserts Archive Information.
Anatomical Graphics	Selects anatomical graphics by category to be inserted into a field.
Anatomical Survey	Selects OB, GYN or Abdomen.
Image	Inserts the image display field to the template.
Wall Motion Analysis	Selects Cut Planes, Bull's Eye, or Score Table Box.
OB/GYN	Selects OB Graph, Bar Graph or Anatomy.
Small Parts	Selects Breast or Thyroid.
Measurements	Inserts the measurement display field in the template.
Text Field	Edits text field.
Fixed Text	Enters any comments as Fixed Text.

Inserting the Page Break

1. Place the cursor where the Page Break is to be inserted and press **Set**.
2. Select Page Break from the Insert menu and press **Set**. The page break line displays on the template.

NOTE

To edit the page break line, select the line and double click the **Set** key.

Inserting the Table

1. Place the cursor where the table is to be inserted and press **Set**.
2. Select Table from the Insert menu and press **Set**. The Insert Table dialog box displays.



Figure 13-257 • Insert Table Dialog

3. Specify each parameter as required.

NOTE

To set the table border as not visible, set “Border” parameter to 0 (zero)

4. Select **OK** to insert the table or Cancel.

NOTE

To insert/delete a row/column from the table or access table properties, double click the **Set** key in any empty area inside the table. A table menu appears with those options.

Inserting Images in a Table

You can choose the order in which images are inserted into tables: by row (default) or by column.

Image Order by Row

The system default inserts images in the cells of the first row, then to the next row.




Figure 13-258 • Image Order—Row Preference (System Default)

1. Follow the instruction for inserting a table. When specifying parameters, specify:
No. of Columns=2; No. of Rows=2
2. After inserting the table, insert an image box in each cell of the table.
 - a. Move the cursor to the first cell and select **Insert > Image**.

- b. Repeat this step for each cell in the table.

After the template is saved and you are working in the ReportWriter, when you select images to be inserted in the table, they are placed in the default order.

Image Order by Column

If you prefer to have the image placement by column, images are inserted in each cell of the first column, then the next column.




Figure 13-259 • Image Order—Column preference

In order to achieve the column preference, you need to create a table with 2 columns and 1 row. In each cell of this table, you need to insert another table.

1. Follow the instructions for inserting a table. When specifying parameters, specify:
No. of Columns=2; No. of Rows=1
2. After inserting the table, create a table inside each of the existing table's cells.
 - a. Move the cursor to the left column's cell and press **Set**.
 - b. Select Table from the Insert menu and press **Set**.
 - c. When specifying parameters, specify:
No. of Columns=2; No. of Rows=1; Width=290 pixels.
Select **OK**.
 - d. Repeat steps a-c for the next column.
3. Insert an image box to each table cell.
 - a. Move the cursor to the first cell and select **Insert > Image**.
 - b. Repeat this step for each cell in the 2 tables.

After the template is saved and you are working in the ReportWriter, when you select images to be inserted in the table, they are placed with your column preference.

Inserting Logos

1. Place the cursor where you want to insert the logo and press **Set**.
2. Select Logo from the Insert menu and press **Set**. The Logo Box displays.



Figure 13-260 • Logo Box

3. Select a logo that you want to insert (1), or import a bmp or jpg file from the removable media (2). Scroll the images using the arrow key (3). Specify the appearance (4).
4. Select **OK** to insert the logo or Cancel.

Changing a logo:

1. Place the cursor on the logo to be changed and press **Set** twice. The Logo Box displays.
2. Select a different logo. If the desired logo is not shown, select Import Logo to import a different logo.
3. Specify the appearance.
4. Select **OK** or Cancel.

Inserting Archive Information

Archive information contains all the objects from the different information menus (Patient, Exam, and Site Information). This box accumulates different information menu selections that can be grouped together and displayed in one table.

1. Place the cursor where you want to enter the archive information and press **Set**. If you use a factory template, double click on the current archive information area to display the Archive Information Box.
2. Select **Archive Info** from the Insert menu and press **Set**. The Archive Info Box displays.



Figure 13-261 • Archive Information Box

3. Type the Heading, select a heading link from the pull-down menu, and select the parameters you want to display in the report.
4. Select Box Properties to change the Font, Alignment, Appearance, etc.

NOTE

To set the same font to all fields, select **Set All fields**.



Figure 13-262 • Table Properties

5. Select **OK** or **Cancel**. The contents of the Archive Information is inserted to the related page automatically.



Figure 13-263 • Patient (Archive) Information Example

Editing displayed Archive Information:

1. Select **Designer**.
2. Move the cursor to Archive Information field to be edited.
3. Press **Set** twice. The Archive Information Box displays.
4. Edit the heading, the Heading Link and Information parameters, as necessary.
5. Select **OK** to save or Cancel.

Anatomical graphics

1. Place the cursor where you want to insert the Anatomical Graphics and press **Set**.
2. Select **Anatomical Graphics** from the Insert menu.
3. Select the desired category and press **Set**. The graphic box displays.



Figure 13-264 • Anatomical Graphics Box Example

4. Select the graphic to be inserted to the template or import a bmp or jpg file from the removable media. Scroll the images using the arrow key.
5. Select Appearance.
6. Select **OK** or Cancel.

Image Display Fields

1. Place the cursor where you want to insert the image.
2. Select image from the Insert menu and press **Set**. The Ultrasound Image Box displays.



Figure 13-265 • Ultrasound Image Box

3. Type the Heading text, modify the box properties, and change the heading text font, as necessary.

NOTE

For no heading, type a Space in the Heading text.

To keep the monitor image appearance, the ratio of width to height (W:H) should be 4:3. So, basically 640:480 for large images and 300:225 for two side-by-side images.

4. Select **OK** or **Cancel**.

Cardiac Studies Wall Motion Analysis

1. Place the cursor where you want to insert the wall motion analysis and press **Set**.
2. Select Wall Motion Analysis from the Insert menu.
3. Select and set up the desired parameter.

- Bull's Eye



Figure 13-266 • Bull's Eye Dialog Box




Figure 13-267 • Bull's Eye Report Example

- Cut Planes

The Cut Planes dialog box parameters are similar to the Bull's Eye Dialog Box shown previously.




Figure 13-268 • Cut Planes Report Example

- Score Table Box



Figure 13-269 • Score Table Dialog Box

OB/GYN (OB and GYN Only)

The OB Graph, Bar Graph and Anatomy can be entered into the Report.

1. Place the cursor where you want to insert the graph or anatomy and press **Set**.
2. From the Insert menu, select OB/GYN. The selection menu displays.
3. Select the appropriate item as necessary. A dialog box displays.
 - OB Graph



Figure 13-270 • OB Graph Dialog Box

1. Select the Measurement and Fetus Number.
2. Check Fetus Trending and Fetus Compare, if appropriate.
3. Modify the Layout, if necessary.
4. Select **OK**.

- Bar Graph



Figure 13-271 • Bar Graph Dialog Box

1. Select the exam and fetus number.
2. Modify the Layout, if necessary.
3. Select **OK**.

NOTE

The Bar Graph already contains default application measurements.

- Anatomy




Figure 13-272 • Anatomy Dialog Box

1. Type the Heading.

2. Select qualifiers from the pull-down menu.
3. Select **Add all** to copy all measurements to the right column.
4. Check the box in front of the measurement you need in the left column and select **Add**. The select measurements copy to the right column.
5. To remove measurements you do not need, check the boxes in front of those measurements in the right column, and select **Remove** or **Remove all**.
6. If you want to modify the properties, select Box Properties and set required parameters.

Measurements

Insert a field to display the measurements. The measured parameters displayed in the measurement display field are configured.

1. Place the cursor where you want to insert the measurement and press **Set**.
2. Select Measurements from the Insert menu and press **Set**. The Measurements Box displays.




Figure 13-273 • Measurement Box

3. Type the Heading text, select the Filter Criteria and measurements from the tree, as necessary.
4. Select **OK** or Cancel.

Text Fields

1. Place the cursor where you want to insert the text and press **Set**.
2. Select Text Field from the Insert menu and press **Set**. The Text Field dialog box displays.




Figure 13-274 • Text Field Dialog Box

3. Type the Heading Text. If you do not need the heading, type a space.
4. Select Display item.
 - Ref.Reason: Reason for Referral.
 - Comments: Gets information from the Comment field of the Patient screen and the Exam Comment field of the Worksheet.
 - Diagnosis.
 - Free Text: 1 - 8
5. Specify the border of the Text Field and Font as necessary.
6. Select **OK** or Cancel.

The text is saved automatically into the corresponding area selected on this dialog box.

Editing an existing text field:

1. Move the cursor to the Text Field to be edited.
2. Press **Set** twice. The Text Field dialog box displays.
3. Edit the heading, the settings, or font, as necessary.
4. Select **OK** or Cancel.

Fixed Text

- 1. Place the cursor where you want to insert the fixed text and press **Set**.
- 2. Select Text Field from the Customize menu and press **Set**. The Fixed Text dialog box displays.




Figure 13-275 • Fixed Text Dialog Box

- 3. Type the text (e.g. hospital information, report title, or table title) and specify the border and font.
- 4. Select **OK** or Cancel.

Editing existing Fixed Text:

- 1. Move the cursor to the Fixed Text to be edited.
- 2. Press **Set** twice. The Fixed Text dialog box displays.
- 3. Edit the text, the border or font, as necessary.
- 4. Select **OK** or Cancel.

Customize Menu

Table 13-94 • Customize Menu

	Description
Page Color	Changes the template color.
Preference	The Preference menu for Archive Information field displays.

Advanced Features

Page Color

1. To change the page color, select Page Color from the Customize Menu and press **Set**. The Color dialog box displays.



Figure 13-276 • Color Dialog

2. Choose the desired color or create a new color.
3. Select **OK** or Cancel.

Setting Preferences

To set preferences for the Archive Information:

1. Select Preferences from the Customize menu and press **Set**. The Preference Box displays.




Figure 13-277 • Preferences Box

2. Select the information to be modified and set the desired preferences.
3. Select **OK** or Cancel.
4. Save the template.

Direct Report

Direct Report

You can use Direct Report to enter Comments, Diagnosis, and Referral Reasons at any time during the examination that will be part of the final report. The comments are reflected on the Report if the Report is configured for those parameters.

1. Select **Direct Rep.** on the measurement Summary window. The Direct Report displays on the left side of the monitor display.



Figure 13-278 • Measurement Summary Window

2. The Direct Report displays on the left side of the screen.



Figure 13-279 • Direct Report

1 Select the type of information

NOTE: Comments entered under Diagnosis appear in the Clinical Diagnosis section of the final report. Comments entered under Referral Reasons appear in the Referral Diagnosis section of the final report.

2 Create/insert pre-defined text

3 Text field

4 List of measurements completed

The measurement results display on the Measurement Overview field.

Double Click: inserts value only for selected line, e.g. 5.98 cm

Shift + Double Click: inserts whole line for selected line, e.g. BPD 5.98 cm

5 Exits the Direct report

3. Select the appropriate parameter and type the free text with the alphanumeric keyboard or use Insert Text.

NOTE

You can configure the pre-defined text at the Utility Report screen.

- a. Select **Insert Text** to display the Insert Text Window.

- b. Use the **Trackball** to select the text to be inserted.
- c. Press **Set**. The selected text displays on the Direct Report.




Figure 13-280 • Full Insert Text Window

- New: Enters the new text
 - Edit: Edits the existing text
 - Delete: Deletes the existing text
 - Close: Closes the insert text window
 - More>>: Displays the Full Insert Text Window
 - <<Less: Minimizes the insert text window
 - Move up/Move down: Moves the text up or down
4. Move the cursor over the measurement result displayed in the overview window, and double click the **Set** key.
 5. Select **Done** at the bottom of Direct Report to exit.
If you configure the field of comment, diagnosis, referral reasons or Measurement on the Report, the text and/or measurement results entered in the Direct Report are automatically displayed on the Report.



Figure 13-281 • Direct Report and Report (Example)

Report Presets

Utility Report Page

You can edit the report template, diagnosis code, and text on the Utility Report page.

Templates

Left Column: The list of all templates (Factory Default, User defined, etc)

Right Column: The list of templates displayed on the template list.




Figure 13-282 • Report Template Tab

- To insert the template on the template list:
 1. Select the application which you want to insert into the template from the pull-down menu above the right column.
 2. Select the category (categories) and/or the template(s) in the left column by the check the box.
 3. Select the right arrow to copy the template to the right column.
- To remove the template from the template list but not from the system):
 1. Select the template in the right column.
 2. Use the left arrow to remove the template from the right column.
- To edit the template or to make a new template:
 1. Enter **Utility > Report > Template** tab.
 2. Select the appropriate template in the left column.
 3. Select **Edit Template**. The Template Designer page displays.
 4. Edit the template and save or save as with a new name.

If you use Save As with a new name, the new template is added to the left column.
Designing Your Own Template on page 1061
- To delete the template:
 1. Select the template to be deleted.

2. Press **Delete**.

- To export the template:



HINTS

Export templates to removable media (CDs, DVDs, USBs) so, at a later time, you can import those templates to a different system or a system with a different software version. Export only works on templates, not data.

1. Insert the removable media in the drive.
2. Move the cursor to “Export Templates” and press **Set**. The available user-defined templates display in the Export Templates window.




Figure 13-283 • Export Templates

3. Select the template(s) to be exported.

NOTE

To select multiple templates, use the **Ctrl** or **Shift** keys.

4. Select the desired removable media under the Select Target Device field.
5. Select **OK**.
6. Press **F3** to eject the media.

- To import the template:



HINTS

Import templates from a different system or a system with a different software version. Import only works on templates, not data.

1. Insert the removable media with the report template(s) to be imported.
2. Select **Import Template**. The Import Template window displays.




Figure 13-284 • Import Templates

3. Select the Source Device from the pull-down menu. Select **OK**.
4. Press **F3** to eject the media.

NOTE

Imported templates are stored in the User defined templates\General directory.

- To move the template from the left column to the right, or from the right to the left:
 1. Select the template to be moved.
 2. Select the Right Arrow or Left Arrow button.
- To move the template up or down in the right column:
 1. Select the template to be moved.
 2. Press the Up Arrow or Down Arrow button.

Comment Texts

You can edit the comment text on the Comment Texts tab.




Figure 13-285 • Comment Texts Tab

- **New:** Enters the new comment.
- **Edit:** Edits the existing comment.
- **Delete:** Deletes the existing comment.
- **Move up/Move down:** Moves the comment up or down.

Backup/Restore Report Templates

Backup moves templates to removable media (CDs, DVDs, USBs).

Restore moves templates, that were backed up onto media, to a similar system or a system with the same software version. For example, it allows templates to move from one LOGIQ Fortis to another LOGIQ Fortis.

To backup the report template:

1. Select **Utility**.
2. Select **System** and select the Backup/Restore tab.
3. Select the media.
4. Check User Defined configuration box of the Backup field.
5. Select **Backup**.
6. Select **Save** and eject the media.

To restore the report template:

1. Insert the media.

2. Select **Utility**.
3. Select **Restore**.
4. After the system reboot, select **Utility** and **Report**.
5. Select the Template tab.
6. Select the appropriate template (*Templates* on page 1082)

Configuring the Wireless Network

Wireless LAN (WLAN)

A Wireless Network (WLAN) is available on the LOGIQ Fortis. When the WLAN is active, an icon appears in the status bar to indicate whether the WLAN is installed or disconnected. See *Network and Spooler Status Icons* on page 1090 for a description of the network icons.

Wireless LAN (WLAN) Specifications

The Wireless LAN (WLAN) supports the following network protocols:

Supported Standards

- Conforms to IEEE 802.11 ac/a/b/g/n/h/i/k/n/r/u/v/w/ac/ax/ Wi-Fi with Bluetooth 5.2 Standard
- FCC Part 15 Class B
- Japan's Radio Law
- Radio Equipment Directive 2014/53/EU (RED)
- Canada Requirements

Security Type (Encryption)

- None/Open
- WPA2/3-Enterprise
- WPA2/3-Personal

Encryption

- None
- WEP
- AES

Internet Protocol

- The LOGIQ Fortis supports IPv4 or IPv6.

Intel declaration of conformity

You can find the Intel Declaration of Conformity at this web link:

<https://www.intel.com/content/dam/support/us/en/documents/network-and-i-o/wireless-networking/9260NGW-EU-RED-DoC.pdf>

Authentication

- None
- Microsoft PEAP using MSCHAP v2
- Cisco EAP/LEAP/PEAP
- Username/password
- Limited support for server certificate verification using pre-installed trusted root CAs

- Client certificates not supported
- User authentication/confirmation prompts not supported

Antenna

- 2 integrated internal wireless antennae

Radio Data Rate

- 1, 2, 5.5, 6, 9, 11, 12, 18, 24, 36, 48, 54, 08, 140, 246, and 300 Mbps (Auto Rate Sensing)

Frequency

- 2.4 GHz to 2.5 GHz (CCK and OFDM Modulation)
- 4.9 GHz to 5.9 GHz (OFDM)

Country Compliance

The WLAN subsystem automatically detects location and adjusts the output to be compliant to country regulations. The following is a partial list of countries that the product is compliant with:

- United States: FCC Part 15 Class B
- Canada: Industry Canada Radio Standards Specification (RSS) license-exempt
- European Union: Radio Equipment Directive (RED) 2014/53/EU
- Japan: Japan Radio Law
- Brazil: National Telecommunications Agency (Anatel)
- Malaysia: Malaysian Communications and Multimedia Commission (MCMC)
- Singapore: Infocomm Development Authority of Singapore (IDA)

Connecting to the WLAN

To connect the ultrasound system to the WLAN,

1. Press **Utility > Connection Manager > Network > Wireless > Configure wireless network**.
The Wireless Network Configuration tool appears. If enabled, Wireless Networks broadcasting in your area appear in the list.
2. If necessary, check the box for “Enable Wireless Connection”.
3. Select the wireless network you want to use or set up.
4. Press Connect from the bottom of the Configuration tool. If prompted, enter the correct settings for this wireless network.

NOTE

If the WLAN fails to connect, review and/or recreate the Wireless connection on the Security Tab.

Network and Spooler Status Icons

The following icons identify network and spooler statuses:

Table 13-95 • Network and Spooler Status Icons

Network and Spooler Status Icons				
Ethernet Active 	Ethernet Error 	Ethernet Active Spooler Active 	Ethernet Active Spooler Error 	Mobile
Mobile Error 	Mobile Bluetooth 	Mobile Wifi 	Spooler Active 	Spooler Active Error
Spooler Inactive 	Spooler Inactive Error 	Wifi 1 Bar 	Wifi 2 Bars 	Wifi 3 Bars
Wifi 4 Bars 	Wifi Alert 	Wifi Spooler Active 	Wifi Spooler Error 	Wifi Error

Adding a Wireless Network

To add a WLAN profile (even for a network which is not yet available),

1. Press **Utility > Connection Manager > Network > Wireless > Configure wireless network**. The Wireless Network Configuration tool appears. Available Wireless Networks appear.
2. Select the Security tab.
3. Select Add...
4. Obtain and enter the correct information for each wireless network setting:
 - a. Enter a Network Name (SSID), pre-select the security type, then press **OK**. A new window appears so you can enter the settings for this network.
 - b. On the connection page, check the appropriate boxes based on how you want the ultrasound system to connect to this network.

The ultrasound system attempts to connect to available wireless networks based on the options you enable. If multiple networks are available, connection attempts begin with the network appearing topmost on the list.

- c. Select the Security page.
- d. Select the Security Type from the available options.

NOTE

Dialog boxes vary, depending on the Security Type selected.

- e. Select the Encryption Type from the available options.
- f. For Personal networks, enter the Network key.
- g. For Enterprise networks, choose the network authentication method and press Settings.

Detailed settings for Enterprise networks are complex and must be correct for the connection to succeed. Ensure you have all the required information/settings and work with your site's IT Network Administrator as needed.

NOTE

Credentials must be entered at this time. The ultrasound system does not support automatic prompting for user credentials.

- For Microsoft PEAP networks, press Advanced and enter User authentication credentials (username and password).
- For Cisco networks, enter user credentials in the PEAP Properties dialog after pressing the Settings button.

NOTE

IMPORTANT: User credentials are not validated until you attempt to connect to the network.

- h. When you have correctly entered all required settings, press OK in the Wireless Network Properties dialog.

Your settings will be validated and you may be prompted about certain settings which are not recommended or supported.

Examples: "Automatic connection to unencrypted networks is not recommended. Reminder to enter user credentials for Enterprise networks."

- i. If changes are necessary, press Customize. Make corrections, then press OK. Settings will be validated again.
5. After you have filled in all the required information, press OK. To cancel adding this profile, press Cancel.

Removing a WLAN

To remove a WLAN profile (even for a network which is not available),

1. Press **Utility > Connection Manager > Network > Wireless > Configure wireless network**. The Wireless Network Configuration tool appears. Available Wireless Networks appear.

2. Select the Security tab.
3. Select Remove.

Customizing Wireless Network Settings

To customize an existing WLAN profile,

1. Press **Utility > Connection Manager > Network > Wireless > Configure wireless network**. The Wireless Network Configuration tool appears. Available Wireless Networks appear.
2. Select the Security tab.
3. Select Customize...
4. Obtain and enter the correct information for each wireless network setting:

- a. Enter a Network Name (SSID), then press **OK**.
A new window appears so you can enter the settings for this network.
- b. On the connection page, check the appropriate boxes based on how you want the ultrasound system to connect to this network.

The ultrasound system attempts to connect to available wireless networks based on the options you enable. If multiple networks are available, connection attempts begin with the network appearing topmost on the list.

- c. Select the Security page.
- d. Select the Security Type from the available options.

Dialog boxes vary, depending on the Security Type selected.

- e. Select the Encryption Type from the available options.
- f. For Personal networks, enter the Network key.
- g. For Enterprise networks, choose the network authentication method and press Settings.

Detailed settings for Enterprise networks are complex and must be correct for the connection to succeed. Ensure you have all the required information/settings and work with your site's IT Network Administrator as needed.

NOTE

Credentials must be entered at this time. The ultrasound system does not support automatic prompting for user credentials.

- For Microsoft PEAP networks, press Advanced and enter User authentication credentials (username and password).
- For Cisco networks, enter user credentials in the PEAP Properties dialog after pressing the Settings button.

NOTE

IMPORTANT: User credentials are not validated until you attempt to connect to the network.

- h. When you have correctly entered all required settings, press **OK** in the Wireless Network Properties dialog.

Your settings will be validated and you may be prompted about certain settings which are not recommended or supported.

Examples: "Automatic connection to unencrypted networks is not recommended. Reminder to enter user credentials for Enterprise networks."

- i. If changes are necessary, press **Customize**. Make corrections, then press **OK**. Settings will be validated again.
5. After you have filled in all the required information, press **OK**. To cancel adding this profile, press **Cancel**.

Setting an IP Address

To set an IP Address (Static or Dynamic), select **Utility > Connection Manager > Network > Wireless > Configure wireless network**.

Refreshing a WLAN

Refreshes the list of available Wireless Networks. To refresh the Wireless Network,

1. Press **Utility > Connection Manager > Network > Wireless > Configure wireless network**. The Wireless Network Configuration tool appears. Available Wireless Networks appear.
2. Press **Refresh** from the bottom of the Configuration tool.

Managing Connectivity to a Wireless Network

You can control which wireless networks can be connected to and which networks are preferred over other networks.

- Networks appearing on the Security page, Preferred Wireless Networks list are listed in decreasing preference.
The network listed at the top of the list is the most preferred, and therefore the most likely to get connected.
The network listed at the bottom of the list is the least preferred, and therefore the least likely to get connected.
- Networks with the check box unchecked for automatic connection will never connect unless you manually press the Connect button.
- Non-broadcasting networks will never connect unless the corresponding checkbox is checked; or, you manually press the Connect button.
- Once connected to a network, the system will not automatically switch over to a more preferred network unless the check box is checked.

Monitoring the WLAN

If there are wireless network communication problems, you can monitor the wireless network events.

To monitor Wireless Networking events,

1. Press **Utility > Connection Manager > Network > Wireless > Configure wireless network**. The Wireless Network Configuration tool appears. Available Wireless Networks appear.
2. Select the Monitor Tab.

WLAN Diagnostics

If the wireless network is connected, you can run diagnostics to determine how well, or poorly, the network itself is working. The diagnostic information displayed can help pinpoint causes of networking problems. Tests which pass are shown in green; tests which fail are shown in red.

To run diagnostics for the Wireless Network,

1. Press **Utility > Connection Manager > Network > Wireless > Configure wireless network**. The Wireless Network Configuration tool appears. Available Wireless Networks appear.
2. Select the Diagnostics Tab.
3. Select **Run Diagnostics**.

Repairing the WLAN

Occasionally you may need to repair a WLAN that has lost its connection to the ultrasound system. To repair the Wireless Network,

1. Press **Utility > Connection Manager > Network > Wireless > Configure wireless network**. The Wireless Network Configuration tool appears. Available Wireless Networks appear.
2. Select the Diagnostics Tab.
3. Select **Repair**.

NOTE

DO NOT cancel the Repair operation after you have selected to repair the Wireless LAN connection.

Available WLAN Channels

The available WLAN channels show availability of wireless connect points that the scanner can talk to. Each channel supports a finite number of users and has limited signal strength. This may effect the ability to connect, the throughput and the connection dropping out.

To check the available WLAN channels,

1. Press **Utility > Connection Manager > Network > Wireless > Configure wireless network**. The Wireless Network Configuration tool appears. Available Wireless Networks appear.
2. Select the Properties Tab.
3. Select **Available Channels....**

Disconnecting from the WLAN

To disconnect from the Wireless Network,

1. Press **Utility > Connection Manager > Network > Wireless > Configure wireless network**. The Wireless Network Configuration tool appears.
2. Select **Disconnect**.

Device Mgmt

Overview

Device Mgmt is a remote device management tool that enables bi-directional management capabilities on the device.

Device Mgmt allows Cloud management of system preset configurations to a fleet of systems on network, as well as one to one system preset configuration Cloud backup and restore.

For Device Mgmt connectivity configuration and activation, see *Device Mgmt* on page 531.

NOTE

For Cloud operation please refer to Device Mgmt online user manual after sign-up at <http://AVURI.gehealthcare.com/signup>

Device Mgmt Fleet Management



Figure 13-286 • Device Mgmt Fleet Management

An account on the Cloud with access to Device Mgmt is required.

1. Upload preset configurations for fleet of systems
2. On the Cloud, assign configuration to fleet of systems
3. Receive installation notification (on each system in the fleet)
4. Install preset configuration (on each system in the fleet)

Device Mgmt Cloud Backup/Restore




Figure 13-287 • Device Mgmt Cloud Backup/Restore

An account on the Cloud with access to Device Mgmt is required.

1. Backup to the Cloud
2. On the Cloud, assign configuration to restore to originating system
3. Receive restore notification
4. Restore on system

NOTE

Restoring to a system other than the originating system is not allowed.

NOTE

Cloud Backup jobs are queued and processed in the background and do not disturb user operation.

Upload For Fleet and Manual Backup (Cloud Backup)

To upload preset information from the master system for the fleet:

1. Go to **Utility > System > Backup/Restore**.
2. Select **Upload For Fleet**.
3. Set Comment to identify configuration on the Cloud.
4. Press **Upload**.
5. The system displays the result in the status bar and the Notification dialog.



Figure 13-288 • Upload For Fleet

NOTE

Upload For Fleet can share the preset configuration across Fleet, while Cloud Backup can restore only to the originating device.

NOTE

Upload For Fleet and Cloud Backup copies all “User Defined Configuration” to the Cloud.

NOTE

The user must have Device Mgmt Administration Group Rights to begin the Upload For Fleet.

Backup Automatically (Local and Cloud Backup)

1. Check **Backup Automatically** under Local and Cloud Backup.
2. Press **Save**.
3. User Defined Configuration files are automatically backed up to Cloud when configuration is changed and then saved.

NOTE

Cloud Backup requires activation. Even if it is not activated, files will automatically copy to Local Backup.

NOTE

If the administrator (ADM) login password is reset and reconfigured, a manual backup should be created to save the ADM password change.

Manual Backup (Cloud Backup)

To backup user defined configuration for Cloud manually:

1. Go to **Utility > System > Backup/Restore**.
2. Select **Cloud** under Backup To/Restore From.
3. Check **User Defined Configuration** under Backup
4. Press **Backup** under Backup.
5. Set Comment to identify fleet configuration on the Cloud.
6. Press **Backup**.
7. The system displays the result in the status bar and Notification dialog.

Assign Configuration to Fleet or to the Device (on Cloud)

Installation assignment for configuration files uploaded to the Cloud is performed on the Cloud (not on the device).

- Upload For Fleet - Configuration files can be assigned to the Fleet (group of devices).
- Cloud Backup - Configuration files can be assigned only to the device where they were backed up from.

NOTE

Configuration files that have been uploaded to the Cloud with Upload For Fleet and Cloud Backup are displayed in different lists on the Cloud.

Receive Installation Notification

When a configuration file is assigned for installation from the Cloud, the system displays a pop-up dialog notifying that a download is available by selecting the download icon in the status bar. The pop-up appears only at the following operation points, to avoid interrupting an exam:

- New Patient/End Current Patient
- Unlock Lock-screen
- Log off/Log on

The download icon also appears in the status bar.



Figure 13-289 • Notification Dialog and Installation Icon

NOTE

The user must have Configuration Group Rights to begin the configuration download and installation. If the user does not have Configuration Group Rights, a Warning message is displayed (see *Figure 13-292 • Advanced Configuration Group Rights Warning* on page 1101).

Installation Dialog

Selecting the Installation Icon in the status-bar initiates the Installation Dialog with three options:

- Install
- Delay Install Until Shutdown
- Download Only
- Discard



Figure 13-290 • Installation Icon




Figure 13-291 • Installation Dialog

NOTE

Discard and Download Only are available only when the Installation is marked as optional on the Cloud. Otherwise, no Discard or Download Only buttons appear in the Installation Dialog.

NOTE

Installation requires Advanced Configuration Group Rights. If the user does not have Advanced Configuration Group Rights, selecting the Installation Icon displays a Warning.




Figure 13-292 • Advanced Configuration Group Rights Warning

Install

1. Select the Installation Icon in the status bar.



Figure 13-293 • Installation Icon

2. Select **Install**.




Figure 13-294 • Install

3. Select **Continue** to acknowledge. The configuration installation begins and the result is displayed.



Figure 13-295 • Continue Install

4. Select **OK** to restart.



Figure 13-296 • Select to Restart System

Delay Install Until Shutdown

1. Select the Installation Icon in the status bar.



Figure 13-297 • Installation Icon

2. Select **Delay Install Until Shutdown**.




Figure 13-298 • Select Delay Install Until Shutdown

3. Select **Continue** to acknowledge. Continue the exam.



Figure 13-299 • Continue Delay Install Until Shutdown

NOTE

Icon changes to “Delay Install Until Shutdown”.



Figure 13-300 • Delay Install Until Shutdown Icon

4. Press Power button to display Exit dialog.
5. Select **Shutdown**. Installation begins automatically and the result is displayed.



Figure 13-301 • Shutdown to Begin Installation

6. When installation is complete the system will shutdown automatically.

NOTE

If the installation fails for any reason, a notification will be displayed after the system restarts.

Discard

1. Select the Installation Icon in the status bar.



Figure 13-302 • Installation Icon

2. Select **Discard**.

NOTE

Discard is available only when the Installation is marked as optional on the Cloud. Otherwise, no Discard button appears in the Installation Dialog.




Figure 13-303 • Select Discard

3. Select **Continue** to acknowledge.



Figure 13-304 • Continue Discard

Detailed Restore from Cloud

Download Only from Cloud

NOTE

Because Download from Cloud runs in the background, the download is already completed when selecting Install icon.

1. Select the Install icon.



Figure 13-305 • Install icon

2. Select **Download Only** which creates local backup of the preset.




Figure 13-306 • Select Download Only

NOTE

If installation of the preset is set as Mandatory, then the dialog enables only the Install or Delay Install until Shutdown.

3. Information displays when the download is completed.

Download completed. To install the downloaded configuration, go to **Utility > System > Backup/Restore** and select Cloud as the Location.



Figure 13-307 • Download Completed

Detailed Restore From Cloud

1. Go to **Utility > System > Backup/Restore**.
2. Select **Cloud** as Location of Restore From
3. Select categories and press **Detailed Restore** under Detailed Restore of User Defined.

Restore from Cloud

1. Go to **Utility > System > Backup/Restore**.
2. Select **Cloud** as Location of Restore From
3. Select categories and press **Restore** under Restore.

Cancel a Failed Job

1. Select Notification Icon.



Figure 13-308 • Notification Icon

2. The notification dialog displays on the monitor. Select a failed job and press **Cancel Selected Job**.




Figure 13-309 • Select Failed Job

3. Select **Continue** on the dialog.



Figure 13-310 • Warning Dialog

NOTE

Successful jobs are removed automatically after a period of time. Failed jobs are deleted automatically after 2 weeks.

Icon and Notification

One of the following icons is displayed in the status bar:

Table 13-96 • Device Mgmt Icons

Icon	Description
	Activated/Not Activated

Icon	Description
	Unread/Read Error message
	Unread/Read Warning message
	Unread/Read information

When there is a notification, clicking on the icon displays a message (see Icon Message example below).




Figure 13-311 • Icon Message

Smart Device Apps

LOGIQ Apps

Smart Device LOGIQ Apps are available for Android devices (both a phone and tablet) via Bluetooth: LOGIQ Remote Control App and LOGIQ Photo/Barcode Reader App.



Figure 13-312 • Smart LOGIQ Apps Devices

NOTE

The communication between the LOGIQ Apps and the ultrasound system is Bluetooth which is a wireless, secure transmission. At no time is patient data stored on the device hosting the LOGIQ Apps. For the LOGIQ Photo App, photos are never written to a file, but rather directly transferred to the LOGIQ ultrasound system. Once on the ultrasound system they can be included with the diagnostic images sent to the reviewing/reading physician.



DANGER

DO NOT use the Smart Device Apps if the patient has a life-sustaining device, such as a pacemaker or defibrillator. Failure to follow this instruction could lead to interference with patient electronic device(s).


NOTE


DO NOT use the USB Ports on the ultrasound system to charge LOGIQ Apps devices as these devices are not approved devices to connect to a medical device.


Connecting the Device

To link to the device:

Table 13-97 • Linking the Device to the LOGIQ Fortis

Steps	On LOGIQ Fortis	On Device
1. Press the Mobile Icon on the LOGIQ Fortis that's located on the display. 		
2. On the LOGIQ Fortis, the QR code appears on the display. 3. On the device, press the LOGIQ Remote App.		
4. Follow the instructions on the device. a. Press "Connect." b. Scan the QR Code with the device.		

Steps	On LOGIQ Fortis	On Device
5. Place the device in front of the LOGIQ Fortis as directed. “Scan QR code with your mobile application to pair and connect.” The LOGIQ Fortis and the device are now ready to pair.		
6. The device displays the Bluetooth pairing request, “Confirm passkey is ##### to pair with the LOGIQ Fortis.” Press OK.		
7. The Bluetooth icon on the LOGIQ Fortis is now paired to the device. 8. The LOGIQ Remote application appears on the device. The Phone App and Tablet App displays are shown to the right.		

Steps	On LOGIQ Fortis	On Device
To switch to the LOGIQ Photo App, press the Camera icon located on the top of the remote device's display. Note: The Camera Icon only appears if LOGIQ Photo App is already installed.		
To disconnect the LOGIQ Fortis from the device, click on the Bluetooth Connected to the Device icon on the bottom of the display and select to disconnect. Or, You can press Disconnect on the Device.		
To Unpair a Device from the LOGIQ Fortis, press Unpair Bluetooth Devices via Utility > Connection Manager > Advanced Settings > LOGIQ Apps.		

To completely break the association between the LOGIQ Fortis and the device (unpair the device from the LOGIQ Fortis):

1. Access the **UtilityConnection ManagerAdvanced SettingsLOGIQ Apps** screen.
2. Select **Unpair Bluetooth Devices**.

Assigning LOGIQ Apps to a BT Key

To assign LOGIQ Apps to a User Define key, select “LOGIQ Apps” to a User Defined key (restart needed)” via **Utility > System > User Configurable Key**.



Figure 13-313 • Assign LOGIQ Apps to a User Defined Key


Using the Apps

Table 13-98 • Using Smart Device LOGIQ Apps


Tablet/Phone Remote Control and Photo/Barcode reader Apps	
1. LOGIQ Remote Control App 2. LOGIQ Photo/Barcode Reader App	
	

Tablet/Phone Remote Control and Photo/Barcode reader Apps

LOGIQ Remote Control App:



Via the Tablet LOGIQ Remote Control App, you will be able to adjust the following controls: Depth, Gain (B-Mode, Color Flow Mode, PW Mode), Left/Right, Measure, Comment, Clear, Trackball, Freeze, Store, Color Flow Scale, and Cursor.



Via the Phone LOGIQ Remote Control App, you will be able to adjust the following controls: Activate PW, CF, or B-Mode, adjust the Cursor, Freeze the image, and sent images to the P1 control.

Photo/Barcode Reader App



You can use the LOGIQ Photo App to capture photos that will go to the patient's exam and to use as a barcode reader to scan in patient exam information.

Note: The photo does not go directly into the exam; instead the image must be stored from the LOGIQ Fortis.

Note: The barcode scanner scans the barcode into the Patient ID field on the Patient menu.

Barcode - Input Mode Off

Enter the Patient ID using the keyboard.

Patient ID

Scan the barcode for the Patient ID or enter the Patient ID using the keyboard.

Complexation

Patient demographics can be entered by scanning the barcode, or manually, using the keyboard.

To enter patient demographics manually with the keyboard (instead of the barcode), select **Cancel**.

1. Enter a string in the Input Data field by scanning from a barcode or typing with the keyboard.
2. Scan a sample barcode. The following items can be included in the barcode:

- Patient ID
- First Name, Last Name, Middle Name
- Birth Year, Month, Day

The character length for Year is four characters, Month is two characters and Day is two characters. The Year, Month and Day should always be provided together.

- Gender

3. Configure the Start and End position for each item.

NOTE

If the barcode does not contain information for an item, configure the Start and End position as "0."

For example, if the scanned barcode is "000001LastNameFirstName191990101F," the configuration and results display as seen in figure below.

The screenshot shows the 'Barcode' configuration window in the LOGIQ Fortis software. The 'Input Mode' is set to 'Barcode'. The 'Input Data' field contains the scanned barcode: "000001LastNameFirstName191990101F". Below this, a table displays the configuration for each item and the corresponding results extracted from the barcode.

Item	Start	End	Result
Patient ID	1	2	00
Order ID	3	4	01
Last Name	5	12	LastName
First Name	13	23	FirstName
Middle Name	24	31	MiddleName
Birth Year	32	35	1990
Birth Month	36	37	10
Birth Day	38	39	01
Gender	40	40	F

At the bottom of the window, there are radio buttons for 'Male' (selected) and 'Female', and a 'Save' button.

Tricefy Uplink

Introduction

Tricefy is a cloud-based image viewer and a platform to archive, collaborate, and share. The Tricefy DICOM server may be used in the way any DICOM server on the product may be used, i.e., Print keys, Send To, etc. The corresponding DICOM destinations can be used via the Print keys. An internet connection is necessary for uploading data to Tricefy.

NOTE

To engage in service offerings with Tricefy, confirm that your country has entered into an agreement with Tricefy.

NOTE

Tricefy DOES NOT support IPV6 and cannot be used in conjunction with Tricefy.

NOTE

Tricefy is not applicable for China market.



Figure 13-314 • Tricefy Account

The tab displays a general info area with a short description of Tricefy, an uplink ID, and an info badge for how to configure Print buttons on system setup (only available when an account is active). Furthermore the account area displays controls and information about the Tricefy account while the test connection area enables you to test a connection and gives detailed information about it.

Uploading Exam Information to the Tricefy Cloud

Enabling Tricefy

1. To enable Tricefy, fill out the information on **Utility > Connectivity > Tricefy**. As soon as this option is enabled, relevant Tricefy items are displayed (e.g. email text field, Activate button,.... and options within the system setup to share or store data to Tricefy via the print key or status bar icons (described below).
2. Enter your email address to register and click Activate account. (If you want to deactivate an active account, click Deactivate account.)
3. Depending on the account status, different account information is displayed:
 - green badge containing the account info, email address, account name, customer name and account status- account activated
 - blue badge - information about a disconnected account
 - orange badge - connection to Tricefy™ failed (due to timeout,...)
4. For testing the connection click Activate. Depending on the connection status, different information is displayed:
 - green badge - connection ok
 - blue badge - instruction for testing the connection (only displayed as long as Activate is not pressed)
 - orange badge - connection failed
5. Upon success activation, you will receive an Uplink ID.

NOTE

When tricefy is successfully activated an image storage and Query//Retrieve service will be created with port number 8104.

Configuring Tricefy

1. Set up the Tricefy Service via **Utility > Connectivity > DICOM..** Set the DICOM Image Storage and Query/Retrieve ports to Port 8104. Refer to the DICOM sections in this chapter and in Chapter 10 for more information. Verify the service.
2. Set up a Dataflow for Tricefy. Refer to the DICOM sections in this chapter and in Chapter 10 for more information. Verify the Dataflow.
3. Configure the Print buttons via **Utility > Connectivity > Print Flow**. Refer to the DICOM sections in this chapter and in Chapter 10 for more information. Verify the Print button(s).

Using Tricefy

You can send a patient's exam to the Tricefy Image Storage service via the Exam Tab.

1. Select the patient's exam and press the **Send To** button. Select the Tricefy Image Storage service. Press **OK**.
You can check the job status by pressing **F4** (DICOM Job Spooler).

You can confirm the exam is on the Tricefy Server via a web browser. Navigate to <https://tricefy4.com>. Log in with your account information. The patient and image(s) you sent are on the Tricefy cloud.

- To Query patient information on the Tricefy cloud, on the Patient page, go to the Data Transfer page. Select the Q/R radial button, then press **Query**. A list of patients on the Tricefy cloud is displayed. Select the patient you want to retrieve, then select Transfer.

Tricefy Icons

Tricefy icons are explained below.

Table 13-99 • Tricefy Icons

Connection Icons	Store Icons	Share Icons
Connected to Tricefy. Ready to Store and Share 	Upload Image to Tricefy 	Share Image with Patient via Tricefy. 
Disconnected from Tricefy. 	Upload CINE to Tricefy 	Share CINE with Patient via Tricefy. 
Registration incomplete. 	Upload 4D Image to Tricefy 	
	Upload 3D Image to Tricefy 	

Tricefy Activation

Refer to *Tricefy* on page 523 for more information.

Digital Expert

Digital Expert Remote Training

Digital Expert is a purchasable option which enables remote training between users and the GE HealthCare Clinical Applications team, through scheduled and real-time sessions. Digital Expert is a self-contained, app-based companion solution, designed to pair with GE HealthCare medical devices. Digital Expert utilizes Intelith software for remote connectivity.

NOTE

This tool is not for clinical diagnostics purposes.

NOTE

Digital Expert is not applicable for China market.

Digital expert also enables customer to customer Enterprise Collaboration, where users can connect with each other within their own network. This provides on-demand access for guidance and consultation for clinicians from in-house experts.

Refer to the user manual included with Digital Expert for information on setting up and using Digital Expert.

Data Streaming (Option)

NOTE

The Data Streaming option key should be installed to enable streaming live/recall/CINE ultrasound data.

The system has the capability to stream live/recall/CINE ultrasound image data over the network connection to enabled devices. The data stream will contain grayscale, color map, geometry, view settings (flip/rotate/reverse), probe and system information, VNav position information and CT/MR registration information with ultrasound data. No patient information is transferred with the streamed data.

The following table represents the types of data streamed in different modes on the ultrasound system.

Data Type	Modes
2D Image	B Mode, B-Flow Mode, Contrast Mode

NOTE

It is recommended to use a 1 Gbps network connection for Data Streaming. The required bandwidth often lies in the 100-300 Mbps range. Usage of a 100 Mbps network leads to dropped frames and the risk of latency buildup.

User Setup for Data Streaming

Only a member of the group **"ReceiveStreaming"** will have permission to receive streamed data. See example of users who can receive streaming data below.



Figure 13-315 • User Setup for Data Streaming

If LDAP authentication is enabled, at least one of the LDAP groups that the user belongs to must be mapped to the group **"ReceiveStreaming"** to allow the device to receive streaming data.

Enable Data Streaming

1. Press **UTILITY+** on the Touch Panel.
2. Press **Utility** on the Touch Panel.
3. Press **Connectivity** on the Touch Panel.
4. Select **Data Streaming** on SmartConnect.
5. Turn on **Enable Streaming**.
6. Enter the Port No (or use the default port).



Figure 13-316 • Data Streaming settings

Starting Data Streaming

1. On the device, connect to the ultrasound system with the IP address, Port Number, user account and password currently used to log on to the ultrasound system.
2. The ultrasound system displays a PIN code. Enter the PIN code on the device.
3. The device and the ultrasound system establish the Data Streaming connection.

NOTE
Once the device establishes the connection, the ultrasound system stores the connection as a certificate. The PIN code is NOT necessary in the next connection.



Figure 13-317 • Certificates

NOTE
The certificate is valid for one year. After the certificate expires, a new PIN code displays on the ultrasound system at the next connection attempt.

NOTE
When the certificate for the device is no longer needed, the certificate can be revoked.

1. Press **APPS+** on the Touch Panel.
2. Press **Utility** on the Touch Panel.
3. Press **Connectivity** on the Touch Panel.
4. Select **Data Streaming** on SmartConnect.
5. Select the certificate.
6. Click Revoke Selected.

Data Streaming in Process

Data streaming is initiated and terminated from an enabled device (e.g. a smart phone/tablet or PC connected to the same network and capable of receiving the streamed data over a custom protocol). An enabled device can only receive data when streaming has been enabled on the system.

Streaming status is indicated on the screen as follows:

Indication (Icon)	Status
	• The streaming device is connected. • The system is in freeze mode, or PW/CW when 2D is not live. • Data is NOT streaming.

Indication (Icon)	Status
	- The streaming device is connected. - Data is streaming.
	- The streaming device is connected. - The system is in a mode not supported in Data Streaming. - Data is NOT streaming.




Figure 13-318 • Data Streaming is in active (example)

Data Streaming can be stopped by the following operations:

- Changing a patient or ending an exam when **Close Stream on Patient/Exam Change** is turned on.
- Pressing the **Pointer Key** to display the pointer on the monitor. Hover the pointer over the Streaming status icon and select the icon.

Running Scanner Apps

App Launchpad

“Apps” in the Utility+ tab lists third-party apps together in a panel. This panel is called App Launchpad and it displays third-party software scanner apps (as buttons in a grid view) which are installed for use with the LOGIQ software. The App Launchpad allows the user launch a scanner app and perform measurements and analysis the app supports.

Scanner Apps

A scanner app is a separate software application having a particular intended use and performance as designed and offered by its legal manufacturer. Qualified apps are separately purchased and installed through the GE HealthCare Marketplace website.

The offered apps are designed and manufactured by a responsible third-party legal manufacturer and verified to be compatible with this ultrasound system. In the App Launchpad the installed apps are represented by app name, name of the legal manufacturer and license information. For the app user instructions, please check the third-party user manuals.

The user can explore more apps verified to be compatible with this ultrasound system, and thus offered through the GE HealthCare Marketplace website. Apps are available on the subscription-based model and can be ordered from the webpage. Orders must be made by an individual with a user account that is enabled to purchase software on behalf of the organization. The app will download and install automatically after completing an order on the webpage, provided that the ultrasound system is connected to the internet. A working internet connection is necessary to acquire the license when an app is activated. The ultrasound system may search for an internet connection on a regular basis to check the user's subscriptions.

Starting a Scanner App from the App Launchpad

1. Recall a B-Mode image or CINE loop from clipboard.
2. Press UTILITY+ > Apps tab.




Figure 13-319 • Accessing the App Launchpad

All compatible Scanner Apps that are already installed on the ultrasound system will be visible on the Touch Panel. Scanner Apps without a valid license will be greyed out.

NOTE

Only Scanner Apps compatible with the chosen image mode will be visible.




Figure 13-320 • Example Scanner Apps with Various License Status

3. Press on the desired available Scanner App to start it.

Service and Applications Support

GE HealthCare Connect Guide

The GE HealthCare Connect Guide helps the user follow the necessary steps to setup the InSite ExC connectivity. To use the GE HealthCare Connect Guide to begin the GE HealthCare Backoffice Connectivity Troubleshooting wizard:

1. Select the Insite icon at the bottom of the display screen to open up the InSite ExC Icon Menu.



Figure 13-321 • InSite ExC Icon Menu with Connection Unconfigured

2. Select **GE Connect Guide** from the menu to open the GE HealthCare Backoffice Connectivity Troubleshooting wizard.

GE Backoffice Connectivity Troubleshooting Wizard

Use the following GE Backoffice Connectivity Troubleshooting wizard screens to setup the GE Backoffice connection.

1. Agent Configuration - The Serial ID is displayed (not editable). Verify CRM ID; edit if necessary and save changes. Select **Next**.




Figure 13-322 • Agent Configuration

2. Network Connection - Ensure you are connected to the Network with a wired or wireless connection. Select **Next**.




Figure 13-323 • Network Connection

NOTE

If you are not connected to the network, see *Configuring Connectivity* on page 497.

3. Agent Proxy Configuration - Ensure proxy configuration details and credentials are accurate. Edit if necessary, save changes and select **Next**.




Figure 13-324 • Agent Proxy Configuration

4. DNS Configuration - Ensure DNS is configured correctly for the system. Select **Next**.




Figure 13-325 • DNS Configuration

5. Network Firewall - Ensure the network firewall in the facility is configured correctly to allow GE InSite communication. Select **Next**.



Figure 13-326 • Network Firewall

6. Summary Page - The Summary page will check the agent configuration and connection status for all configuration steps. When the process completes (may take up to five minutes) the status for each step is listed as "OK" or "Not OK." If any step is reported as "Not OK" select the step from the menu on the left and correct the information on that screen, then proceed to the Summary page again for a connection status update.



Figure 13-327 • Summary Page

Support Requests

InSite ExC

InSite ExC is your direct link with a GE HealthCare Online Service Engineer or Applications Support Engineer by creating a Request for Service via the InSite ExC link at the bottom of the display screen.

NOTE

The Insite connection must already be configured with the GE HealthCare backoffice for the InSite ExC Menu to appear. See *GE HealthCare Connect Guide* on page 1127 for more information.



Figure 13-328 • InSite ExC Menu

Types of InSite ExC Service

1. Contact GE HealthCare - Opens a service dispatch with GE HealthCare Service.
2. Connect Clinical Lifeline - Sets the system in a state of readiness for Virtual Console Observation.
3. Connect To GE HealthCare - Initiates a connection to the GE HealthCare Backoffice.

Initiating a Request for Service (RFS)

To initiate an RFS:

1. Position the Windows pointer on top of the GE HealthCare InSite ExC icon at the bottom of the display.
2. Press the Right Trackball Set Key. Select Contact GE HealthCare. This opens the RFS screen which sends a service dispatch directly to the Remote Service or Applications Team after you fill in the following information:
 - Last Name, First Name, Phone Number
 - Select correct Problem type
 - Choose Problem area
 - Type in Problem description
3. After you have completed filling in all of this information, press **Send** to initiate the Request for Service.

You can confirm if your RFS has been sent as well as RFS's automatically sent by the system. The ultrasound system can automatically submit a Request for Service. These are displayed on the Machine Queue.

In addition, you can use the Users screen to identify your institution's point of contact for service dispatches.

Initiating a Technical or Clinical Support Request

When the Remote Applications or Service person reaches you, they will ask you to you click on the InSite icon and choose either Connect to GE HealthCare or Connect Clinical Lifeline,

1. Position the Windows pointer on top of the GE HealthCare InSite ExC icon at the bottom of the display.
2. Press the Left Trackball Set Key. This opens the following pop-up:
 - Connect to GE HealthCare,
 - Connect Clinical Lifeline, or
 - Cancel
3. Select the option the GE HealthCare representative asks you to select.

NOTE

Selecting Connect to GE HealthCare changes the polling time from 15 minutes to 15 seconds so that your call can be answered as quickly as possible. Selecting Connect Clinical Lifeline also activates disruptive mode.

InSite ExC icons appear differently, depending on their state:

Table 13-100 • InSite Icons

GE HealthCare InSite - Connected - Disruptive Mode - Enabled	
GE HealthCare InSite - Connected - Disruptive Mode - Disabled	

GE HealthCare InSite - Idle - Disruptive Mode - Enabled	
GE HealthCare InSite - Idle	
GE HealthCare InSite - Not Configured For more info on the configuring the connection to the GE HealthCare Backoffice, see <i>GE HealthCare Connect Guide</i> on page 1127.	

InSite ExC Definitions

Here are definitions for the different InSite ExC states:

Virtual Console Observation (VCO). Allows Technical Support to control the ultrasound system functionality remotely.

Disruptive. Allows GE HealthCare's Technical Support person to connect to your system via VCO, to run diagnostics directly on your ultrasound system, and to collect system logs. When the system is in Disruptive Mode, the icons are red. There are two disruptive states. If you see a telephone with a clock, then the system is in Disruptive, Not Connected Mode. If you see a telephone with GE HealthCare, then the system is in Disruptive, Connected Mode.

Non-Disruptive. Allows GE HealthCare's Technical support person to look around on your system, but cannot perform any service-related functions, depending on whether InSite has connected or not connected. There are two Non-Disruptive states. If you see a black and white icon, InSite ExC is activated, but not open for Technical Support access. If you see a yellow icon, InSite ExC is activated and the Technical Support person can look around on your system, but cannot perform any service-related functions.

Connected. InSite ExC is connected.

Not Connected. InSite ExC is not connected.

NOTE

When Disruptive mode has been activated or a diagnostic has been run, the message, "Service Mode is Activated. Reboot required before patient use" appears in red at the bottom of the display. It is recommended that you reboot the system before use. Make sure you disable disruptive mode before rebooting or the message will not be cleared.

Exiting InSite ExC

To exit InSite ExC:

1. The GE HealthCare Technical Support person will exit Disruptive Mode and VCO, then ask you to reboot your system.
2. Reboot your ultrasound system before patient use.

For more information on InSite ExC refer to the Basic Service Manual.

Service Desktop

Overview

The Administrator (and others granted the “Access Service Desktop” privilege) can access the Service Desktop to:

- Access the Service Home Page.
- Perform the following Utilities
 - Change Password
 - Data Transfer
 - Delete Files
 - SSA License
 - Gather Logs
 - View Third Party Licenses
 - Disruptive Mode Utility
- View Installed Options
- InSite Agent Configuration



For more information on Service Desktop, refer to the Basic Service Manual.

Figure 13-329 • Service Desktop

Accessing the Service Desktop

1. Select **Utility** and go to Page 2 of the Utility Menus.
2. Select **Service**.

Battery Power Mode

Overview

The LOGIQ Fortis supports either the Power Assistant (PA) battery Option or the Scan on Battery option (ScoB).

NOTE

Only GE HealthCare Service personnel have access to the batteries. Please contact the technical service department or your local GE HealthCare representative for replacement.

Power Assistant Battery Option

Overview

With Power Assistant (PA), the system power is maintained by the batteries when AC power is lost or the power cable is unplugged. PA can reduce system startup time for portable exams. The system will be able to run for at least 20 minutes after AC power is disconnected.

When the ultrasound system enters PA upon AC power loss, there is an initial temporary message stating that the system is “entering battery mode,” followed by a message that the system is “running on battery.”

If an operation was interrupted by AC power loss, the system displays a notification letting you know which operations were not completed successfully. In this case, you will need to redo the operation that was interrupted, such as an Export, Import, Query Retrieve, etc.

While in PA battery power mode, the system cannot scan or perform post-processing, but can quickly start up once the AC power is re-established. See *Recovering from Power Assistant mode* on page 1135

In PA mode, the following operations are allowed:

- Media Format, Finalize, Verify
- DICOM Image Storage, Performed Procedure, Print, Storage Commitment, Query of Q/R, Worklist.

NOTE

The Power Assistant Label is affixed to the Main Display Panel Rear Cover when the Power Assistant option is installed on the system.




Figure 13-330 • Power Assistant Label

Before starting Power Assistant Mode



CAUTION

The following operations must be completed before unplugging the AC cable.

- Any access to the external media, such as CD, DVD, USB HDD via Export, Import, MPEGvue, Save As, EZBackup.
- DICOM Retrieve from Query/retrieve
- DVR recording should be stopped.
- Volume Navigation should be exited by Exit and Clear Control.
- System brightness and backlighting is also dimmed/stopped.
- Processing 3D/4D images in live.

NOTE

If the system is switched to Power Assistant Mode when any of the operations above are in progress, the system terminates the operation.

When in Power Assistant Mode

In Power Assistant Mode, most of the console devices, such as the keyboard and printers, are turned off to minimize battery usage.



CAUTION

When in Power Assistant Mode, DO NOT perform the following operation:

- Connect and disconnect probes.
- Remove peripheral devices, such as the Volume Navigation unit.
- Eject or insert CD/DVD media and the external USB storage device.

Failure to follow these instructions could lead to exposure to unexpected acoustic energy or electrical interference/loss of patient data.

NOTE

To shut down the system while in Power Assistant Mode, plug in the power cable, exit from Power Assistant Mode, and then perform regular shut down operations.

Recovering from Power Assistant mode

The system recovers from Power Assistant Mode when AC power is less than 10 seconds reconnected to the LOGIQ Fortis and the main breaker is on. The system returns to full functionality in 10 seconds.

Scan on Battery Option (ScoB)

Overview

Scan on Battery Option allows the user to scan using battery power (AC unplugged).

While running on the battery power, the LOGIQ Fortis can scan or perform post-processing and switch to Power Saving Mode.

With a fresh battery that is fully charged, the LOGIQ Fortis stays powered for approximately 50 minutes.

Before starting Scan on Battery



CAUTION

The following operations must be completed before unplugging the AC cable.

- Any access to CD/DVD.
- DVR recording to CD/DVD.

It is recommended to complete the following before unplugging the AC cable.

- Any access to the external media, such as USB HDD via Export, Import, MPEGvue, Save As, EZBackup.
- Any access to the Network device, such as DICOM transfer or Network Storage.
- Volume Navigation should be exited by Exit and Clear Control.
- Processing 3D/4D images in live mode.

Before starting Scan on Battery, go to Utility -> System > General page and select the appropriate value for “Auto Switch to Power Saving Mode” and “Freeze when AC is Unplugged” (Default: Off).




Figure 13-331 • Extended Battery for Scanning

Auto Switch to Power Saving Mode

- Select 15, 30 minutes or 1 hour to automatically switch to Power Saving Mode after the time specified on battery operation.
- Select “Never” to never switch to Power Saving Mode.
- Select “Always” to always switch to Power Saving Mode.

Freeze when AC is Unplugged

- On: When selected, the system goes into Freeze mode when the AC cable is unplugged.

Starting Scan on Battery

When the AC cable is unplugged from AC power, the system automatically transitions to Scan on Battery Mode.

NOTE

If the system does not transition to Scan on Battery Mode, contact the technical service department or your local GE HealthCare representative.

**WARNING**

Do not use Scan on Battery Mode when you need to handle (transfer/ export/ import) patient data. (Export/Import/ Backup/ Burn Media, Dicom Transfer, etc.) The system can unexpectedly shutdown when the battery capacity is too low and result in patient data loss. Please note the message in the status bar when Scan on Battery is activated. It is strongly recommend to use AC power when transferring, importing or exporting patient data.

System is using battery power. Make sure the battery has enough capacity to prevent potential data loss

Figure 13-332 • Warning message

Recommendation

Plug in AC power if possible, keep using the battery power can impact to the battery life.

When the battery capacity becomes low

The LOGIQ Fortis displays a dialog for the user to guide to switch to Power Saving Mode when the battery capacity becomes low. User can either keep scanning until the system shuts down or go to Power Saving Mode to extend time before shutdown.

Switch to Power Saving Mode

**WARNING**

The system switches to Power Saving Mode automatically when the capacity is low.

You can switch to Power Saving Mode by selecting “Switch to power-saving” in the battery status popup anytime you want.

Switch to power-saving
Battery Condition : Healthy - Discharging
98% Charged

The LOGIQ Fortis allows going back to Scan on battery mode by pressing the **Power** button or **Freeze** button.

Figure 13-333 • Battery status pop-up



Figure 13-334 • Dialog during Power Saving Mode

View current battery status

When the system is running under Power Assistant/Scan on Battery, battery icons appear in the title bar and on the status bar, indicating how much charge is left in the battery.

Battery Status Icon on the status bar

The system displays the battery's current charge and posts a notification message on the display when battery power needs attention.

Detailed battery information is available by clicking on the icon and display the popup, which includes the battery's current capacity displaying.





Figure 13-335 • Battery status pop-up

NOTE

When the Front Battery LED is blinking, the system is in standby mode and the batteries are charging.

Table 13-101 • Battery status icons

Icon	Description
No icon	AC Plugged; no battery present
	AC Unplugged; battery is fully charged (81 - 100%)
	AC Unplugged; battery is partially charged (61 - 80%)

Icon	Description
	AC Unplugged; battery is partially charged (31 - 60%)
	AC Unplugged; battery is partially charged (0 - 30%)
	AC Plugged; battery is fully charged (81 - 100%) and charging
	AC Plugged; battery is partially charged (61 - 80%) and charging
	AC Plugged; battery is partially charged (31 - 60%) and charging
	AC Plugged; battery is partially charged (0 - 30%) and charging
	Battery error icon When an abnormal status of the battery is detected, the system displays the following icon: • Battery Temperature error • Communication error or battery charge error NOTE If this icon displays, do not initiate Power Assistant. Contact the technical service department or your local GE representative.

Battery status Icon in the title bar

Fuel Gauge icon appears in the title bar.




Figure 13-336 • Fuel Gauge icon

Table 13-102 • Battery status icons (AC unplugged)

Icon	Description	Icon	Description
	91 - 100%		41 - 50%
	81 - 90%		31 - 40%
	71 - 80%		16 - 30%
	61 - 70%		0 - 15%
	51 - 60%		

NOTE
To avoid an unintended shutdown of the system, and to avoid risk of losing patient data, it is strongly recommended to connect to AC power when the fuel gauge icon is yellow or red.

Battery charging

NOTE
Whether the LOGIQ Fortis is on or off, the battery will charge as long as the system is connected to AC power via the power cable, and the break is on.

A fan will run while the battery is charging. The fan stops automatically when the battery is done charging.

While the battery is charging, the front battery LED will blink. When the battery is fully charged, the LED will stop blinking and remain on.



Figure 13-337 • LED location of the battery

Approximate charging time (from empty to full)

- Power Assistant battery: About 2.5 hours
- Scan on Battery: About 2.5 hours

Refreshing the battery

To maintain battery life and accuracy of the fuel gauge, and to avoid unexpected shutdowns, it is recommended to refresh the battery every 6 months with the following procedure:

1. Plug the AC cable to the wall outlet and turn on the circuit breaker.
2. Wait until the battery is fully charged. It takes about 2.5 hours to fully charge the battery (depend on the remaining battery capacity).
3. Wait at least 1 hour.
4. Remove all probes.
5. Turn on the system.
6. Unplug the AC cable, letting the system run on battery until it automatically shuts down. It may take at least 30 minutes (Power Assistant Battery)/ 50 minutes (Scan on Battery) to complete shutdown.
7. Wait at least 5 hours.
8. Plug the AC cable to the wall outlet and turn on the circuit breaker.
9. Wait until the battery is fully charged. It takes about 2.5 hours to fully charge the battery.

Battery Deterioration

When the system detects battery deterioration, the following dialog displays:




Figure 13-338 • Battery Life Message

If this message appears, the ultrasound system disables both Power Assistant and Scan on Battery.

Contact the technical service department or your local GE representative.

Battery Disposal

Lithium Ion

Used batteries must be disposed of properly and as chemical waste. They cannot be treated as regular waste. Contact your building administration for proper disposal.

NOTE

When removing a defective battery, ensure that it is disposed of in accordance with local regulations. Alternatively, provide it to GE HealthCare for proper disposal.

Magstripe Card Reader

Magstripe Magnetic Card Reader

The Magstripe Card Reader can be attached to the LOGIQ Fortis by connecting it to any LOGIQ Fortis USB port.



Figure 13-339 • Magstripe Card Reader

The Magstripe Card Reader can be used in Patient screen, Worklist and Image Display.

- Patient screen: Patient ID, First Name, Last Name, Middle Name, Perf. Physician, Ref. Physician
- Worklist: Patient ID, Search String
- Image Display: Comments (Annotation)

Footswitch

Wired Footswitch

You can attach this Footswitch to the system by connecting it to any USB port on the system.



This is a 3-pedal Footswitch. You can configure its functionality via the **Utility > Applications > Footswitch** parameters.

Figure 13-340 • Footswitch and USB Cable



CAUTION

When using the Footswitch, DO NOT hold down the footswitch pedal. Press and release the Footswitch pedal. Pushing and holding down the pedal behaves the same way as pushing and holding down a key on the keyboard.

DVR

Setting up the DVR

To set up the Digital Video Recorder (DVR), configure the LOGIQ Fortis PRIOR TO using the DVR since the system MUST BE restarted after updating these configuration parameters.

- Set up following parameters in **Utility > System > Peripherals**.
 - Media: DVD, USB
Pre-format USB device as NTFS; pre-format DVD-R/-RW/+RW as NTSC or Pal.
 - DVD Format: NTSC, PAL
For DVD-RW, NTSC is preferred; for DVD+RW, Pal is preferred.
 - Picture Quality: HQ, SP, LP, EP
 - Microphone Level: 1 to 5
 - DVD Chapter Record Interval (sec.): 15, 30, 60, 120
 - USB Playback Skip Interval (sec.): 15, 30, 60, 120
- If needed, assign [DVR Record/Pause] to a Print key in following steps.
Add [Video Capture] and select [DVR Record/Pause] as the Type, in **Utility > Connectivity > Service**.
Assign the [Video Capture] to the Print Button, in **Utility > Connectivity > Button**.
- Restart the system.

DVR functionality is now set to record the scan.

Using the DVR



WARNING

Loss of patient data may occur during an AC failure. Ensure that you are not using Power Assistant and that you do not disconnect AC power while saving an exam to the DVR.

- Create a patient record or open an existing one.
- Press the Video-assigned key.
- Insert a DVD media or USB storage, which is selected as DVR media (**Utility > System > Peripherals**).

NOTE

There must be a formatted DVR-R disk installed in the system for Software DVR to record.

NOTE

The DVD formatted using **Utility > Connectivity > Removable Media** is for data archiving; it cannot be used for DVR recording.

- Press **Record** to start recording.

The status bar updates with the recording information. While recording, you can only pause or stop the recording and cannot perform any other DVR functionalities.

NOTE

Recording automatically pauses when you access the Utility screen.

NOTE

When more than one USB storage is connected, the DVR records to the first drive letter, in alphabetical order.

NOTE

The system does not start the recording while the following operation is in progress: Data Transfer, EZBackup, Utility screen.

5. Press **Record** or **Pause** to pause/resume recording.
6. Press **Stop** on the Touch Panel to stop recording.
When you stop recording, the DVR writes the title data. During this period of time, the busy state icon is displayed in the status bar (green arrow circling) and you cannot operate any DVR functionalities.

NOTE

It may take up to 2 minutes for the DVD-RW/DVD+RW to stop recording.

NOTE

The DVR automatically stops recording before the media fills up.

7. Press **Eject** on the Touch Panel or the **F3** key to eject.
8. Press **Scan** tab to return to scanning.

Recording Functionality - DVD Media

Support media

DVD-R, DVD+RW, DVD-RW

Support video format

NTSC or PAL

Support Picture Quality

- HQ (Record length: approx. 70 minutes)
- SP (Record length: approx. 90 minutes)
- LP (Record length: approx. 140 minutes)
- EP (Record length: approx. 280 minutes)

Number of Titles and Chapters

You can create up to 99 titles on a disk and 92 chapters in a title.

While recording to DVD, each chapter is automatically created at a specific interval that is configured (**Utility** > **System** > **Peripherals**).

When you stop recording, the DVR writes the title. When you end an exam (e.g. press New Patient, End Current Patient, ...), the DVR automatically stops recording and creates a title.

Volume label

The system automatically labels following string to DVD media: YYYYMMDD_XX

- YYYY - Year (4 digit)
- MM - Month (2 digit)
- DD - Day (2 digit)
- XX - Serial number of the day (2 digit)

NOTE

Volume label of DVD+RW/-RW can be edited via the Utility -> Connectivity -> Removable Media page.

Video title

The title name on the DVD is titled as follows: <Date>_<Time>_<Patient Last name>_<Patient First name>

NOTE

If the patient's name is not provided, then a serial number is used as follows: <Date>_<Time>_<serial number>.

Compatibility

DVD media recorded on the LOGIQ Fortis is compatible with the following players:

- Player on LOGIQ Fortis

NOTE

DVD media recorded on previous versions of the LOGIQ Fortis is playable on the system.

- DVD video player on a PC
- Commercial DVD video player

Recording Functionality - USB Storage Supported media

USB-HDD (NTFS only), USB Flash Device (NTFS only)

NOTE

You cannot record or play FAT (exFAT, FAT32, etc) on a USB Device. Please format the USB device to the NTFS file system on a PC.

Supported format

MPEG video file

Recording to USB storage creates an mpeg video file (*.mpg) (\LOGIQ_Series_DVR\folder) for each title. The file size can be up to 2GB.

NOTE

The mpeg video file does not have a chapter.

Support Picture Quality

- HQ (Record length: approx. 30 minutes)

Advanced Features

- SP (Record length: approx. 40 minutes)
- LP (Record length: approx. 60 minutes)
- EP (Record length: approx. 130 minutes)

NOTE

The maximum size of each video file: 4GB.

Video file name

The mpeg video file name is titled as follows: <Date>_<Time>_<Patient Last name>_<Patient First name>..

When the patient's name is not provided, then the video file name is as follows, <Date>_<Time>_<serial number> where the serial number is the number of video files on the media plus one.

Recovery (DVD only)

While recording to DVD (DVD-R or DVD+RW/-RW), if a power failure occurs, the system will automatically start recovery of the media right after the system restarts.

NOTE

If the recording has not finished before the power failure, the last title may be unreadable. Recovery copies the last cached title to the media.

If the DVD media in the tray does not have sufficient free space, the following Recovery message appears (instead of the system starting auto recovery):

"Last title was not burned to DVD. Free space on the media is too small. Insert blank DVD for the recovery. Do you want to start recovery?" Select "Yes," "Remind me later," or "Delete Data."

- Remind me later

If you select "Remind me later", you can cancel the Recovery operation. But if you press the Record button, the recovery menu displays again and you cannot start to record.

Select "Yes" (Recover) or "Delete Data" to enable the recording.

- Yes (Recover)

If you select "Yes", the following dialogue displays ("Insert blank DVD for the recovery."). Insert the blank media and press "OK".

During recovery, the message "Recovery of DVR media is in progress" displays in the status bar.

- Delete Data

If you select "Delete Data", you get the following confirmation menu ("Are you sure you want to delete unburned DVR data?").

OK: Deletes the unburned cached data.

Cancel: Goes back to the Recovery menu.

Video Touch Panel Menu

DVR Touch Panel menus:



Figure 13-341 • Video Touch Panel

Table 13-103 • Touch Panel Keys

Touch Panel Key	Explanation
Skip (I<</>>I)	DVD media: Skip to the previous/next chapter. USB storage: Skip specific time to forward/backward.
4x Skip (4x I<</>>I 4x)	DVD media: Skip four chapters forward/forward. USB storage: Quad time interval skip to backward/forward.
Search	Press Search to start the Search dialog. Select Title, adjust offset then press Search Counter to start the Search. The title name is as follows: <Date>_<Time>_<Patient's Last name>_<Patient's First name>
Record (Red circle with Media Inserted)	Start recording.
Pause (II)	Pauses and resume Recording and Playback. NOTE The Freeze key works the same as the Pause button, when in Playback.
Open Media Drawer	Press to open the Media Drawer.
Eject	Eject the DVD media/USB storage device. Also you can eject media pressing the F3 key.
Play (> with Media Inserted)	Play a video from DVD or USB storage device. NOTE A DVD-R ONLY plays back AFTER the entire disc has been finalized. NOTE A DVD-RW/+RW plays back after finalizing each session.
Stop (Box)	Stop playback or recording.

Touch Panel Key	Explanation
Format (Media Inserted with Red X)	Reformats a DVD+RW/-RW. NOTE Format deletes all the content on the media.
Finalize (Media Inserted with Green Checkmark)	Finalize a DVD-R. NOTE DVD+RW/-RW is automatically finalized when you stop recording.

DVR Display Icons

Displays the following icons in the status bar:



Figure 13-342 • DVR Disk Status

Table 13-104 • DVR icons

1. DVR status			
	Recording		Stopping
	Pause Recording		Pause Playback
	Playback		Busy
	No media		
2. DVD Media status Displays device availability and available free space.			
	DVD+RW/-RW is in the tray		DVD-R is in the tray
3. USB storage status Displays the device availability and the free space size.			
	USB Storage is attached		
4. Current Title Number. Displays while recording			

5. Time counter.

Special Characters

- % Stenosis
 - generic measurement 279

Numerics

- 3D/4D imaging
 - introduction 737
 - manipulating the volume of interest 752
 - operational controls 743
 - principles of operation 737
- 3D/4D presets, changing 494
 - Advanced 496
 - Render Setting 495
 - ROI (Region of Interest) 495
 - Scalpel 496
- 3D/4D presets, overview 494

A

- A/B Ratio
 - Doppler generic measurement 289
 - generic measurement 283
 - M-Mode generic measurement 285
- Acceleration time (AT)
 - OB/GYN vessel measurement 290
- Acceleration, measuring 290
- accessory
 - connector panel 81
- Acclimation time 144
- Accumulation, B-Flow 179
- accuracy
 - clinical calculation 686
 - clinical measurement 684
- acoustic output
 - default levels 61
- Activating Continuous Capture 871
- Active Images 409
- active images, description 146
- Admin
 - overview of Utility screen 563
- Admin screen
 - Logon 572
 - System Admin 568

- Admin screen ()
 - Users 571
- administrator
 - specifying system 571
- AFI
 - AVC Timing Adjustment 898
- AFI, see Amniotic Fluid Index (AFI) 351
- air filter
 - removing 711
- ALARA (as low as reasonably achievable), bioeffects 39
- Alpha Hip, pediatric measurement 399
- Amniotic Fluid Index (AFI), measuring 351
- Analysis
 - Stress Echo 877
- Angle, B-Mode generic measurement 283
- annotating an image
 - text overlays 249
 - using the annotation library 249
 - using typed words 253
- Annotation mapping
 - Breast lesion 480
- Annotations Libraries Presets Menu 476, 479
- annotations, presets 478
- anonymize 1046
- Anonymize Patient
 - patient data 1046
- application presets
 - user-defined 594
- Applications
 - setting presets 487
- Applications Preset Menu 485
- area measurements
 - ellipse 293
 - spline 294
 - trace 293, 294
- AUA
 - OB worksheet 310
- audio, speakers 81
- Auto Calcs
 - modifying 208
- Auto Calcs, modifying 208
- Auto sequence
 - using 325

Auto Vascular Calculation, see also Manual
Vascular calculations
activating 386
Auto Vascular Calculation, see also Manual
Vascular calculations
setting up calculation parameters 386
AutoEF
Tracking validation 904

B

B Steer plus, adjusting 182
B-Mode imaging
intended uses 170
optimizing 170
scanning hints 170
typical exam protocol 170
B-Mode measurements
Echo Level 296
Gynecology exam 371
B-Mode measurements, general 292
B-Mode measurements, generic
% Stenosis 279
A/B Ratio 283
Angle 283
Volume 280
B-Mode measurements, mode
circumference and area (ellipse) 293
circumference and area (spline trace) 294
circumference and area (trace) 294
distance 292
back view, console 78
Background, B-Flow 179
backing up data
EZBackup/Move, see EZBackup/Move,
using 425
Backup and Restore Preset Menu 467, 599
Basic Service Manual 24
Battery status 1138
Battery, refresh 1141
biological hazards 43
Bladder volume, measuring 393
body pattern application libraries, selecting
485
Body Patterns 254
General tab 484
Body Patterns Preset Menu 484
Breast lesion
Annotation mapping 480

C

calculations
OB worksheet 311
selecting 273
urology 393
calipers, description 274
Capture Recon., B-Flow 237
Capture, B-Flow 180
Care and maintenance
cleaning the system 703
footswitch 713
operator controls 708
system cabinet 706
Caution icon, defined 39
Change the display format 135
CINE capture 236
Enhancement 237
CINE loop
previewing only 406
storing and previewing 406
storing without previewing 406
CINE mode 231
synchronizing loops 233
using 232
circuit breaker
description 89
circumference measurements
ellipse 293
spline 294
trace 293, 294
Clinical
calculation accuracy 686
measurement accuracy 684
Color Flow imaging
activating 195
intended uses 195
M-Mode 204
optimizing 195
Power Doppler 200
Compare Assistant
Control 1015
Set up 1014
Workflow 1013
Connectivity
configuring 497
presets 497
console
front view 78
moving 140

- console ()
 - transporting 143
 - wheels 141
- Continuous Capture
 - Stress Echo 870
- contraindications 28
- Contrast Clock Highlight 780
- Contrast Imaging
 - overview 776
- Control panel
 - replacing key caps 713
 - replacing key lamps 713
- Control Panel
 - description 113
- controls
 - annotation function 121
 - display function 121
 - keyboard 115
 - measurement function 121
 - mode function 121
 - operator 113
 - print function 121
 - Touch Panel 117
- CUA
 - OB worksheet 310
- Curved Anatomical M-Mode 194
- CW Doppler
 - activating 848
 - exiting 848
 - non-imaging 848
 - overview 848
 - steerable 848

D

- D/S ratio 289
- Danger icon, defined 39
- Data Transfer, description 146
- dD Ratio, pediatric measurement 399
- device labels 65
- devices
 - acceptable 59
- DICOM Viewer 436
- Digital Expert 1120
- disinfecting probes 646
- disinfecting solutions, probes 646
- distance measurement
 - general 292
- Doppler Auto Calc Average Cycle, using 299
- Doppler measurements, generic

- Doppler measurements, generic ()
 - A/B Ratio 289
 - D/S ratio 289
 - Heart Rate 289
 - Pulsatility Index (PI) 288
 - Resistive Index (RI) 288
- Doppler measurements, mode
 - TAMAX and TAMEAN 297
 - time interval 297
 - velocity 297
- Doppler Mode
 - generic study 287
- Doppler Mode, general measurements 296
- Doppler Mode, PW
 - intended uses 205
 - optimizing 205
 - typical exam protocol 205
- dual image mode, see split-screen imaging 121
- DVD Multi drive
 - location 79

E

- Echo level measurements 296
- editing
 - patient information 163
 - user-defined calculations 333
- EFW growth percentile
 - OB worksheet 311
- electrical
 - configurations 76
- electrical hazard 43
- electromagnetic compatibility (EMC) 48
- ellipse measurement, general 293
- EMC (electromagnetic compatibility) 48
- End diastole (ED)
 - OB/GYN vessel measurement 290
- Endometrium thickness (Endo), GYN exam 371
- Enhancement 237
- environmental requirements 76
- Environmental requirements
 - probes 623
- equipment safety 43
- Erasing
 - measurements 273
- exam
 - definition of terms 272
 - deleting 164

Index

exam ()

- OB 150

- workflow 272

explosion hazard 43

EZBackup/Move, using 425

F

Fast Key 262

Federal law (USA), requirements 29

Fetal growth bar graph 359

Fetal growth curve graph

- description 354

- multiple fetus 361

- quad view 355

- selecting 355

Fetal trending

- fetal growth curve graph 357

- multiple fetuses 361

Fetus

- entering number of 360

- selecting on an OB worksheet 310

Fetus Compare

- multiple fetus 361

Flow Volume 301

folders, measurement

- adding 326

Follicle measurements, GYN exam 371

Footswitch 1144

freezing an image 230

front panel, location 84

front view, console 78

FV 301

G

Gels, coupling 646

General imaging

- changing presets 474

generic studies and measurements 278

Generic study

- Doppler mode 287

GYN exam

- endometrium thickness 371

- follicle measurements 371

- ovaries measurements 372

- uterus measurements 372

Gynecology exam 371

- B-Mode measurements 371

H

hazards 644

- biological 620

- electrical 620

hazards, safety symbols 39

hazards, types

- biological 43

- electrical 42, 43

- explosion 43

- mechanical 42

- smoke and fire 43

Heart Rate

- Doppler generic measurement 289

- M-Mode generic measurement 286

Hip Dysplasia, pediatric measurement 397

I

Image acquisition

- Stress Echo 868

Image History, description 146

Image Zoom 239

images

- deleting 164

- recalling from clipboard 419

- reviewing 409

Imaging Presets 473

Imaging presets, changing

- General 474

InSite, using 1130

Intravessel ratio, calculating 389

K

keyboard

- special keys 115

L

labeling probes 71

Libraries Preset Menu 483

Lock the monitor 124

log on procedures

- defining 572

LOGIQView 181

Logon

- Admin screen 572

M

- M-Mode
 - CAMM 194
- M-Mode imaging
 - intended uses 191
 - optimizing 191
 - typical exam protocol 191
- M-Mode measurements, generic
 - % Stenosis 285
 - A/B Ratio 285
 - Heart Rate 286
- M-Mode measurements, mode
 - time interval 305
 - tissue depth 304
- M-Mode, general measurements 304
- managing images
 - media handling tips 403, 439
 - media requirements 403
- Manual Vascular Calculation 388
- Measurement
 - Tool
 - Copy, Past&Move 275
- Measurement & Analysis screen
 - accessing 318, 362, 363
- measurement controls, location 268
- Measurements
 - erasing 273
 - general instructions 273
 - OB
 - AFI 351
- measurements, general 267
- measurements, generic
 - overview 278
- measurements, types
 - amniotic fluid index (AFI) 351
- measurements, using
 - adding 326
 - automatically starting in workflow 325
 - calipers 274
 - changing 326
 - deleting 334
 - selecting in different category 270
- Minimum diastole (MD)
 - OB/GYN vessel measurement 290
- MO drive
 - location 79
- Mode Cursor, displaying
 - Doppler Mode 208
- monitor

- monitor ()
 - speakers 81
- moving the system 139
 - during transport 143
 - precautions 140
 - wheels 141
- Multiple fetuses 360
 - fetal trending 361
 - identifying 360
 - on OB worksheet 361
- My Desktop
 - set up 134

N

- network status indicator, location 84
- new patient
 - scanning 153

O

- OB
 - graph 353
 - identifying multiple fetuses 360
 - Patient data 151
 - studies 349
- OB exam
 - preparing 150
 - starting 151
- OB graphs
 - fetal growth bar graph 359
 - patient data 356
 - viewing 353, 354
- OB measurements, types
 - amniotic fluid index (AFI) 351
- OB multigestational 360
- OB worksheet
 - AUA 310
 - calculations 311
 - CUA 310
 - EFW growth percentile 311
 - multiple fetuses 361
 - patient data 310
 - selecting ultrasound age 310
- OB/GYN vessel measurements 290
 - acceleration 290
 - acceleration time 290
 - end diastole 290
 - minimum diastole 290
 - peak systole 290

Index

- Operator controls 708
- optimizing images
 - B-Mode 170
 - Color Flow 195
 - Doppler, PW 205
 - M-Mode 191
- options
 - system 568
- Orientation Help, 3D/4D imaging 747
- ovaries, measurement 372
- P**
- past exam
 - entering patient data 358
- Patient Cardiac and ECG Connections
 - location 85
- Patient data
 - deleting 164
 - OB 151
 - OB graphs 356
 - OB worksheet 310
 - searching 162
- patient exam
 - reviewing 409
- patient safety 41
- Patient Screen
 - active images 146
 - Data Transfer 146
 - Image History 146
- Peak systole (PS)
 - OB/GYN vessel measurement 290
- Peak systole/end diastole ratio, Doppler
 - generic measurements 289
- pediatric exam
 - calculations 397
- pediatric measurements, types
 - alpha HIP 399
 - dD ratio 399
 - hip dysplasia 397
- peripherals
 - connector panel 81
 - location 78
- peripherals, digital
 - setting up 444
- Phantoms 716
- Plot Both, fetal trending 357
- Plot Control
 - line style 861
 - vertical auto scaling 861

- portable exam
 - performing 452
- power
 - power up sequence 90
- Power 88
 - circuit breaker 89
 - connection
 - USA 88
 - Cord 139
 - On/Off 90
 - shut down 100
 - switch, location 90
- Power Assistant 1134
- Power Doppler imaging 200
- prescription device, caution 29
- Preset Program menu
 - Acoustic Output
 - Fetal Exposure 151
 - Prudent Use 151
- presets
 - organizing folders and measurements 321
- presets, changing
 - Anotations/Libraries 476
 - Body Patterns 483–485
 - Connectivity 497
 - Imaging 473
 - System 456
- presets, overview 454
- previous exam data
 - entering manually 358
- Probe handling and infection control 622
- probe orientation, 3D.4D imaging 742
- probes
 - cable handling 109
 - connecting 107
 - coupling gels
 - coupling gels, probes 646
 - deactivating 110
 - disconnecting 111
 - disinfecting 646
 - environmental requirements 623
 - ergonomics 606
 - labeling 71
 - planned maintenance 649
 - probe orientation 616
 - safety 618
 - using protective sheaths 620
 - storing 111
- Prostate volume, measuring 394

prudent use 39
 PS/ED or ED/PS Ratio, measuring 289
 Pulsatility Index (PI), Doppler generic measurement 288

Q

QAnalysis 821, 851
 Drift Compensation 854
 Exit 865
 generating a trace 856
 manipulating the sample area 857
 Plot Control 853, 861
 Select Image Range 855
 smoothing 862
 QAPlot Control
 horizontal sweep 862
 Quality Assurance 715
 baselines 717
 frequency of tests 715
 Introduction 715
 periodic checks 717
 phantoms 716
 record keeping 723
 system setup 718
 test descriptions 718
 typical tests 715

R

real-time 4D imaging
 acquiring data 749
 stopping acquisition 754
 Record keeping 723
 reference images, 3D/4D 740
 Renal volume, measuring 393
 Render view, 3D/4D imaging 746
 Report
 Wall Motion Analysis 885
 Report Writer report templates, customizing 1061
 Report Writer reports
 activating 1049
 creating 1048
 Direct Report, using 1079
 editing 1051
 factory templates 1051
 inserting images 1057
 Resistive Index (RI), Doppler generic measurement 288

Results window 273

S

S/D Ratio, Doppler measurement 289
 safety
 electromagnetic compatibility (EMC) 48
 equipment 43
 hazards 39, 43, 61, 620
 biological 644
 labels 65
 patient 41
 acoustic output hazard 42
 electrical hazards 42
 mechanical hazards 42
 patient identification 41
 patient training, ALARA 42
 personnel 43
 precaution icons, defined 39
 precaution levels, defined 39
 probes 618
 handling precautions 622
 Scalpel, 3D/4D imaging 758
 Scan Assistant
 Availability 983, 986
 Definitions 986
 Description 987
 Exporting Programs 992
 Setting Up 987
 Using 988
 Scan Assistant Creator
 Attributes 1004
 Features 1004
 File Handling 993
 Help 993
 Keyboard
 Navigating 1003
 Measurements 1009
 Overview 992
 Programs
 Editing 997
 Views 999
 Scoring
 Stress Echo 878
 Sectional view, 3D/4D imaging 746
 Sensitivity/PRI, B-Flow 180
 Shear wave imaging
 Activating Shear wave 827
 Configuring Shear wave 825
 Intended Uses 824

Index

- Shear wave imaging ()
 - Measurement Bias Information Tables 832
 - Overview 824
 - Scanning Hints 842
 - Shear wave Measurements 837
 - Typical Exam Protocol (Liver) 836
 - Using Shear wave 835
 - site requirements, before the system arrives 76
 - small parts exam
 - thyroid measurements 344
 - speakers, description 81
 - Speckle Reduction Imaging-High Detection (SRI-HD) see SRI-HD, adjusting 175
 - Spectral Doppler, see Doppler Mode, PW 205
 - Speed of Sound 188
 - Split Screen 242
 - split-screen imaging 121
 - SRI-HD, adjusting 175
 - starting an OB exam 151
 - Static 3D imaging 755
 - acquiring data 755
 - storage area
 - location 79
 - Strain Elastography 816
 - Stress Echo
 - create a template 879
 - studies
 - OB 349
 - study
 - adding 326
 - definition 272
 - deleting 334
 - generic 278
 - organizing 321
 - system
 - acclimation time 144
 - electrical configurations 76
 - environmental requirements 76
 - options 568
 - power down 100
 - System Admin
 - Admin screen 568
 - System cabinet 706
 - System Imaging Preset Menu 461
 - System Measure Preset menu 464
 - System presets, changing
 - General 456
 - System About 472
 - System presets, changing ()
 - System Imaging 461
 - System Measure 464
 - System Peripherals 468
 - System/Backup and Restore 467, 599
 - System presets, overview 456
 - System/General Preset Menu 456
- ## T
- TAMAX and TAMEAN, Doppler mode measurement
 - manual trace 297
 - Test Patterns
 - overview 130
 - TGC, adjusting 186
 - Thyroid measurements 344
 - TIC analysis
 - activating 784
 - curve fitting 800
 - exiting 785
 - generating a trace 791
 - manipulating the sample area 795
 - overview 784
 - smoothing 797
 - TIC Plot Control
 - horizontal scale 797
 - line style 797
 - overview 796
 - vertical auto scaling 796
 - Time Gain Compensation, see TGC, adjusting 186
 - Time interval
 - Doppler mode measurement 297
 - M-Mode measurement 305
 - Time-Based Store 489
 - Timers
 - Stress Echo 870
 - Tissue depth, M-Mode measurement 304
 - Tissue Velocity Imaging 849
 - Trace measurement, general 293, 294
 - TVD 850
 - TVI 850
- ## U
- Ultrasound age
 - selecting on OB worksheet 310
 - Ultrasound-Guided Attenuation Parameter (UGAP) Option 843

- Unlock the monitor 124
- urology exam, preparing 272, 342
- urology measurements
 - bladder volume 393
 - pelvic floor 395
 - prostate volume 394
 - renal volume 393
- USB Quick Save 442
- user IDs
 - defining 571
- User label
 - set up 134
- user-defined calculations, editing 333
- user-defined measurement, adding 328
- Users
 - Admin screen 571
- Uterine cavity, amniotic fluid index (AFI) 351
- Uterus measurements 372
- Utility screens
 - Admin 563
 - connectivity 497

V

- Vascular measurements, selecting 389
- Vascular worksheets
 - saving and printing 310, 314
 - vessel summary 312
- Velocity, Doppler measurement 297
- Vessel summary, vascular exam 312
- Volume measurement, B-Mode generic measurement 280
- Volume Navigation Connections/Indicator
 - location 86

W

- Warning icon, defined 39
- wheels, console 141
- word wrap, annotation 253
- Worksheet
 - viewing 306
- Write Zoom, activating 239

Z

- zooming an image
 - bioeffects 240

