



Benannt durch/Designated by
 Zentralstelle der Länder
 für Gesundheitsschutz
 bei Arzneimitteln und
 Medizinprodukten
 www.zlg.de
 BS-MDR-099



Product Service

EU Quality Assurance Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex XI Part A
 (Class I Devices in sterile condition, with measuring function or reusable surgical instruments)

No. G21 086533 0008 Rev. 00

Manufacturer:

Jinhua Jingdi Medical Supplies Co., Ltd

Building 2, Ditian Function
 Xiaoshun Town, Jindong Zone
 321000 Jinhua City, Zhejiang
 PEOPLE'S REPUBLIC OF CHINA

SRN Manufacturer - CN-MF-000009653

Authorized Representative:

MedPath GmbH
 Mies-van-der-Rohe-Strasse 8, 80807 Munich, GERMANY

The Certification Body of TÜV SÜD Product Service GmbH certifies that the manufacturer has established, documented and implemented a quality management system as described in Article 10 (9) of the Regulation (EU) 2017/745 on medical devices. Details on device categories covered by the quality management system are described on the following page(s).

The Report referenced below summarises the result of the assessment and includes reference to relevant CS, harmonized standards and test reports. The conformity assessment has been carried out according to Annex XI Part A of this regulation with a positive result.

As applicable the involvement of the notified body is limited to the aspects relating to:

- establishing, securing and maintaining sterile conditions,
- conformity of the devices with the metrological requirements,
- reuse of the device, in particular cleaning, disinfection, sterilization, maintenance and functional testing and the related instructions for use.

The certified quality assurance system is subject to periodical surveillance by TÜV SÜD Product Service GmbH. All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with. For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:G21 086533 0008 Rev. 00

Report No.: SH2275501

Valid from: 2023-08-18

Valid until: 2028-08-17

Christoph Dicks
 Head of Certification/Notified Body

Issue date: 2023-08-18



Benannt durch/Designated by
 Zentralstelle der Länder
 für Gesundheitsschutz
 bei Arzneimitteln und
 Medizinprodukten
 www.zlg.de
 BS-MDR-099



Product Service

EU Quality Assurance Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex XI Part A
 (Class I Devices in sterile condition, with measuring function or reusable surgical instruments)

No. G21 086533 0008 Rev. 00

Classification: Class I
Device Group: M04 - SPECIAL DRESSINGS
Device Properties: MDS 1005.1 - Ethylene Oxide sterilization

Classification: Class I
Device Group: T02 - PROTECTIVE CLOTHING AND DRAPES (EXCLUDING
 PERSONAL PROTECTIVE EQUIPMENT - PPE)
Device Properties: MDS 1005.1 - Ethylene Oxide sterilization

The validity of this certificate depends on conditions and/or is limited to the following: ./.

Revision History:

Rev.	Dated	Report	Description
00	2023-08-18	SH2275501	Initial issuance