

Test Report issued under the responsibility of:





IEC 60601-1 Medical electrical equipment			
Part 1: General requirements for basic safety and essential performance			
Report Reference No	N40P0001		
Date of issue:	Mar 23, 2020		
Total number of pages	145		
CB Testing Laboratory	SGS Germany GmbH, CRS Munich		
Address:	Hofmannstrasse 50, 81379 Munich, Germany		
Applicant's name:	ADLINK Technology GmbH		
Address:	Ulrichsberger Str. 17		
	94469 Deggendorf, Germany		
Test specification:			
Standard:	IEC 60601-1:2005 (Third Edition) + CORR. 1:2006 + CORR. 2:2007 +		
	A1:2012 (or IEC 60601-1: 2012 reprint)		
Test procedure:	CB Scheme		
Non-standard test method:	N/A		
Test Report Form No	IEC60601_1K		
Test Report Form(s) Originator.:	UL(US)		
Master TRF:	2015-11		
Copyright © 2015 Worldwide System for Conformity Testing and Certification of Electrotechnical			

Copyright © 2015 Worldwide System for Conformity Testing and Certification of Electrotechnical Equipment and Components (IECEE), Geneva, Switzerland. All rights reserved.

This publication may be reproduced in whole or in part for non-commercial purposes as long as the IECEE is acknowledged as copyright owner and source of the material. IECEE takes no responsibility for and will not assume liability for damages resulting from the reader's interpretation of the reproduced material due to its placement and context.

If this Test Report Form is used by non-IECEE members, the IECEE/IEC logo and the reference to the CB Scheme procedure shall be removed.

# This report is not valid as a CB Test Report unless signed by an approved CB Testing Laboratory and appended to a CB Test Certificate issued by an NCB in accordance with IECEE 02.

#### General disclaimer:

The test results presented in this report relate only to the object tested.

This report shall not be reproduced, except in full, without the written approval of the Issuing CB testing laboratory. The authenticity of this Test Report and its contents can be verified by contacting the NCB, responsible for this Test Report.



Test item description:	Medical Panel Computer	
Trade Mark		
	Leading EDGE COMPUTING	
Manufacturer	ADLINK Technology GmbH	
Model/Type reference	MLC 8 series (MLC8-21, MLC8-23, MLC8-27)	
Ratings:	100 – 240 Vac; 1,5 A – 0,75 A; 50/60 Hz	

Testing procedure and testing location:			
CB Testing Laboratory:	SGS Germany GmbH, CRS Munich		
Testing location/ address:	Hofmannstrasse 50, 81379 Munich, Germany		
Tested by (name, function, signature):	Stefan Koschke Qualification Engineer		
Approved by (name, function, signature):	Katja Blaesing Qualification Engineer		
Testing procedure: CTF Stage 1:			
Testing location/ address:			
Tested by (name, function, signature)			
Approved by (name, function, signature):			
Testing procedure: CTF Stage 2:			
Testing location/ address:			
Tested by (name, function, signature)			
Witnessed by (name, function, signature):			
Approved by (name, function, signature):			
Testing procedure: CTE Stage 3:			
Testing procedure: CTF Stage 4:			
Testing location/ address			
Tested by (name, function, signature)			
Witnessed by (name, function, signature):			
Approved by (name, function, signature):			
Supervised by (name, function, signature):			



List of Attachments:			
			Page:
Photographs		121	
Test equipment and additional test documentation			125
Attachments N	National Deviations		126
Canadian Nati	onal Differences		126
Japan Nationa	al Differences		131
Korea Nationa	Il Differences		142
USA - Differer	ices		144
- IEC 60601-1	-6 Test Report No : N40P0002 of SGS	Germany GmbH (	kent in file)
- IEC 60601-1	-2 Test Report No : N40P0010 of SGS	Germany GmbH (	kept in file)
- IEC 60601-1	-2 Test Report No : N40P0014 of SGS	Germany GmbH (	kept in file)
Summary of t	easting.		
Summary of t	lesting.		
Tests perforn	ned (name of test and test clause):	Testing location:	
Clause	Tests	SGS Germany Gmb	H, CRS Munich
4.11	Power Input	Hofmannstrasse 50	
5.7	treatment	81379 Munich	
7.1.2	Legibility of markings	Germany	
7.1.3	Durability of Marking Test		
8.4.3	ME equipment intended to be connected to a power source by a plug		
8.6.4	Impedance and current-carrying capability of protective earth connections		
8.7	Leakage currents and Patient auxiliary currents		
8.7.4.7	Patient Leakage Current – measurements made under special test conditions		
8.8.3	Dielectric Strength		
8.9.4	Measurement of Creepage Distances and Air Clearances		
11.1	Maximum temperature during normal use		
11.6.3	Spillage test		
11.6.6	Cleaning and disinfection of ME EQUIPMENT and ME SYSTEMS		
11.8	Interruption of the power supply / supply mains to ME equipment		
13.2	Single fault conditions		
15.3.2	Push test		
15.3.3	Impact test		
15.3.4	Drop test		
Summary of compliance with National Differences			



⊠ National Differences

CA, US, CH, JP, KR

Explanation of Codes: CA=Canada, CH=Switzerland, JP=Japan, KR=Korea, US=United States of America



#### Copy of marking plate

The artwork below may be only a draft. The use of certification marks on a product must be authorized by the respective NCBs that own these marks.



Product identification label





Mains input fuse label



GENERAL INFORMATION				
Test item particulars (see also Clause 6):				
Classification of installation and use :	Stationary			
Device type (component/sub-assembly/ equipment/ system):	Equipment			
Intended use (Including type of patient, application location) :	The All-in-One panel computers of the MLC 8 series are devices compliant to medical purpose and are intended to display, monitor and store data accumulating while processing medical and / or patient data in medical environments			
Mode of operation:	Continuous			
Supply connection	Appliance coupler			
Accessories and detachable parts included:	None			
Other options include				
Testing				
Date of receipt of test item	Jul 08, 2019			
Date(s) of performance of tests	Jul 08, 2019 to Nov 06, 2019			
Possible test case verdicts:				
- test case does not apply to the test object	N/A			
- test object does meet the requirement:	Pass (P)			
- test object was not evaluated for the requirement	N/E (collateral standards only)			
- test object does not meet the requirement:	Fail (F)			
Abbreviations used in the report:				
- normal condition: N.C.	- single fault condition: S.F.C.			
- means of Operator protection: MOOP	- means of Patient protection: MOPP			
General remarks:				
"(See Attachment #)" refers to additional information appended to the report. "(See appended table)" refers to a table appended to the report. The tests results presented in this report relate only to the object tested. This report shall not be reproduced except in full without the written approval of the testing laboratory. List of test equipment must be kept on file and available for review. Additional test data and/or information provided in the attachments to this report.				
Throughout this report a $oxtimes$ comma / $oxtimes$ point is used as the decimal separator.				
This document is issued by the Company subject to its General Conditions of Service (www.sgsgroup.de/agb). Attention is drawn to the limitations of liability, indemnification and jurisdictional issues established therein.				
This document is an original. If the document is submitted digitally, it is to be treated as an original within the meaning of UCP 600.				
Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.				





The MLC 8 series devices can be identified via the assembled LCD panel size and power input. Sub variants are generated via different CPUs, optional interfaces and extensions such as WLAN/Bluetooth modules or various storage medium capacities. Sub variants are not reflected by key codes but tracked via serial number and device database entries.

All variants are available with incorporated PSU connected to mains

#### **Tested Configuration:**

Component	Туре	<b>Identification No</b>	lss.	Serial No	Comment
Medical Panel Computer	MLC 8-21			MLC8-21-0001-19	
Medical Panel Computer	MLC 8-23			MLC8-23-0001-19	
Medical Panel Computer	MLC 8-27			MLC8-27-0001-19	

#### Conditions of Acceptability:

- 1. The system is to be installed, switched on and maintained only by suitably trained and qualified personnel in accordance with the installation instructions provided with the EUT.
- 2. The maximal acceptable operating temperature is +30 °C
- 3. The EUT is reliable connected to ground. For testing of clause 8.6.4 a detachable cord with 2 m length has been used (refer to table 8.10), different length shall be considered in end use application
- 4. Manual checked in English; for placing the product on other markets it should be in local language.
- 5. The device does not offer any essential performance itself. Essential performance must be evaluated in final end user application.
- 6. Any mechanical provisions to mount the displays are not part of this investigation and shall be considered in final end user application
- 7. All connected IT equipment must comply with appropriate standards and shall be considered in end use application
- 8. Performance of installed software and connection to an IT network (Chapter 14 of this standard) shall be considered in end use application



Test Report issued under the responsibility of:





### TEST REPORT IEC 60601-1-2

## Medical electrical equipment –

### Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests

Report Number:	N40P0014
Date of issue:	Feb 24, 2020
Total number of pages	85
Name of Testing Laboratory preparing the Report:	SGS Germany GmbH, Consumer and Retail Hofmannstrasse 50 81379 Munich, Germany
Applicant's name:	ADLINK Technology GmbH
Address:	Ulrichsberger Str. 17
	94469 Deggendorf
	Germany
Test specification:	Germany
Test specification: Standard:	Germany
Test specification: Standard: Test procedure:	Germany IEC 60601-1-2:2014 CB
Test specification: Standard: Test procedure: Non-standard test method	Germany IEC 60601-1-2:2014 CB N/A
Test specification: Standard: Test procedure: Non-standard test method: Test Report Form No	Germany IEC 60601-1-2:2014 CB N/A IEC60601_1_2E_EMC
Test specification: Standard: Test procedure: Non-standard test method: Test Report Form No: Test Report Form(s) Originator:	Germany IEC 60601-1-2:2014 CB N/A IEC60601_1_2E_EMC UL(US)

# Copyright © 2017 IEC System of Conformity Assessment Schemes for Electrotechnical EQUIPMENT and Components (IECEE System). All rights reserved.

This publication may be reproduced in whole or in part for non-commercial purposes as long as the IECEE is acknowledged as copyright owner and source of the material. IECEE takes no responsibility for and will not assume liability for damages resulting from the reader's interpretation of the reproduced material due to its placement and context.

If this Test Report Form is used by non-IECEE members, the IECEE/IEC logo and the reference to the CB Scheme procedure shall be removed.

This report is not valid as a CB Test Report unless signed by an approved CB Testing Laboratory and appended to a CB Test Certificate issued by an NCB in accordance with IECEE 02.

General disclaimer:

The test results presented in this report relate only to the object tested and documents reviewed. This report shall not be reproduced, except in full, without the written approval of the Issuing CB Testing Laboratory. The authenticity of this Test Report and its contents can be verified by contacting the NCB, responsible for this Test Report.

Test item description Me	Medical Panel Computer			
Trade Mark				
Manufacturer Al	urer ADLINK Technology GmbH			
Model/Type reference Mi	LC 8 series:			
- N	1LC8-23			
Ratings 10	0 – 240 V ac; 50/60 Hz; Class 1			
Responsible Testing Laboratory (as	applicable), testing procedure and testing location(s):			
CB Testing Laboratory:	SGS Germany GmbH, Consumer and Retail			
Testing location/ address	Hofmannstrasse 50 81379 Munich, Germany			
Tested by (name, function,				
signature)	Sperling			
Approved by (name, function, signature)	91/21			
	Wössner Abh Wall			
Testing procedure: CTF Stage 1:				
Testing location/ address				
Tested by (name, function, signature)				
Approved by (name, function, signature)				
Testing procedure: CTF Stage 2:				
Testing location/ address				
Гested by (name, function, signature)				
Witnessed by (name, function, signature)				
Approved by (name, function, signature)	Approved by (name, function, signature)			

	Testing procedure: CTF Stage 3:	
	Testing procedure: CTF Stage 4:	
Test	ing location/ address	
Test sign	ed by (name, function, ature)	
Witn sign	essed by (name, function, ature)	
App sign	roved by (name, function, ature)	
Supe sign	ervised by (name, function, ature)	

List of Attachments (including a total number of pages in each attachment): None			
Summany of testing:			
Tasta performed (neme of test and subslause):	Testing lesstion:		
Conducted EMISSIONS (7.3)	SGS Gormany		
Radiated EMISSIONS (7.3)	Consumer and Retail FMC-I ab		
Disturbance Power EMISSIONS (7.3)	81379 Munich. Germany		
Harmonic Currents (7.2.1)	·····		
Voltage Fluctuations and Flicker (7.2.3)			
Electrostatic Discharges (8.9)			
RF Electromagnetic Fields (8.9)			
Proximity fields from RF wireless communications EQUIPMENT (8.10)			
Electrical Fast Transients (8.9)			
Surge (8.9)			
Conducted Disturbances Induced by RF Fields (8.9)			
Voltage Dips and Interruptions (8.9)			
Note: delete tests not conducted in the list above.			
Summary of compliance with National Differences:			
List of countries addressed			
$oxedsymbol{\boxtimes}$ The product fulfils the requirements of IEC 60601-1-2: 2014 (Fourth Edition)			