

# Operation Manual

## ERO•SCAN

Screening and Diagnostic Version





## Table of Contents

<b>1</b>	<b>Introduction .....</b>	<b>4</b>
1.1	Intended Use Statement.....	4
1.2	Indications for Use Statement .....	4
1.3	Contraindications of Use Statement.....	4
1.4	Features .....	5
1.5	Description .....	6
<b>2</b>	<b>For Your Safety .....</b>	<b>7</b>
2.1	Reading this Operation Manual.....	7
2.2	Customer Responsibility.....	8
2.3	Manufacturer’s Liability.....	8
2.4	Regulatory Symbols .....	9
2.5	General Precautions .....	10
2.6	Electrical and Electrostatic Safety .....	10
2.7	Electromagnetic Compatibility (EMC).....	12
2.8	Battery Safety and Capacity .....	13
<b>3</b>	<b>Warranty, Maintenance and After-Sales Service .....</b>	<b>14</b>
3.1	Warranty.....	14
3.2	Maintenance.....	14
3.3	Cleaning and Disinfection Recommendations.....	15
3.4	Safe Use of Disposables .....	16
3.5	Disposables – Eartips.....	16
3.6	Disposables – Probe Tubes .....	17
3.7	Troubleshooting .....	18
3.8	Recycling/Disposal.....	21
<b>4</b>	<b>Unpacking and Hardware Orientation .....</b>	<b>22</b>
4.1	Unpacking the System .....	22
4.2	Hardware Orientation .....	24
4.3	Handling the Micro-Probe.....	25
4.4	Battery Charging .....	26
4.5	Establishing a PC-Connection.....	27
4.6	Using the Thermal Printer HM-E200 .....	28
4.7	Storage.....	29
<b>5</b>	<b>Operating the Device .....</b>	<b>30</b>
5.1	Getting Started with the ERO•SCAN.....	30
5.2	Indication Lights .....	31
5.3	Control Panel .....	32
5.4	Turning On the Device .....	32

5.5	Turning Off the Device .....	33
5.6	Main Menu .....	33
5.7	Testing Procedure .....	34
5.8	Interpreting Test Results .....	40
5.9	Managing Test Results.....	44
5.10	Settings.....	47
<b>6</b>	<b>Technical Data.....</b>	<b>60</b>
6.1	ERO•SCAN Hardware.....	60
6.2	Connections and Pin Assignment .....	65
6.3	Calibration Values .....	66
6.4	Electromagnetic Compatibility (EMC).....	66
6.5	Electrical Safety, EMC and Associated Standards .....	70
6.6	Configurations and Test Protocols .....	71
6.7	Flowcharts.....	72

Title: Operation Manual ERO•SCAN Screening and Diagnostic Version

Date of issue/last revision: 16/07/2024



MAICO Diagnostics GmbH  
Sickingenstr. 70-71  
10553 Berlin  
Germany  
Tel.: + 49.30.70 71 46-50  
Fax: + 49.30.70 71 46-99  
E-mail: sales@maico.biz  
Internet: www.maico.biz

All available operation manuals can be found in the download center on the MAICO homepage:

Germany:



<https://www.maico-diagnostics.com/german/support/resources/>

International:



<https://www.maico-diagnostics.com/support/resources/>

## Copyright © 2024 MAICO Diagnostics

All rights reserved. No part of this publication may be reproduced or transmitted in any form or by any means without the prior written permission of MAICO. The information in this publication is proprietary to MAICO.

## Compliance

MAICO Diagnostics is an ISO 13485 certified corporation.

**Caution for the USA:** Federal Law restricts this device to sale by or on the order of a licensed medical professional.

## Trademark Notice

HearSIM™ is an unregistered trademark of OtoAccess A/S for use in the USA.

OtoAccess is a trademark of OtoAccess A/S registered in the USA and Europe.

Sanibel Supply is a trademark of Interacoustics A/S registered in the USA and Europe.

# 1 Introduction

This Section offers you important information about:

- the intended use of the device
  - indications and contraindications of use
  - features and benefits
  - a description of the device
- 

## 1.1 Intended Use Statement

The ERO•SCAN Hearing Test System is indicated for testing of cochlear function in infants, children, and adults by measuring otoacoustic emissions (OAEs). This instrument is suitable for use in all settings, including hospitals, schools, physician's offices, and audiologist practices. Factory defined protocols allow for simple screening measurements and user customizable protocols allow for diagnostic evaluations. The ERO•SCAN is intended to be used by hearing healthcare professionals (i.e. ENT doctors, audiologists) and/or technicians, neonatal nurses and school nurses who have been trained by a hearing healthcare professional.

## 1.2 Indications for Use Statement

The OAEs are generated by a series of clicks that are directed into the ear canal.

Otoacoustic emissions are low level audio-frequency sounds that are produced by the cochlea as part of the normal-hearing process. Available evidence suggests that otoacoustic emissions are generated by the cochlea's outer hair cells and that the presence of OAEs is an indication that the outer hair cells are viable. Clinical evidence indicates that these emissions normally occur with normal hearing, or at most, mild hearing loss (usually 30-40 dB HL). The majority of hearing-impaired individuals will be identified by a simple OAE test.

## 1.3 Contraindications of Use Statement

Testing should not be performed on patients with one of the following symptoms without a medical doctor's approval:

- Recent stapedectomy or other middle ear surgery
- Discharging ear
- Acute external auditory canal trauma
- Discomfort (e.g., severe otitis externa)
- Occlusion of the external auditory canal

Visual inspection for obvious structural abnormalities of the external ear structure and positioning as well as the external ear canal should be performed before testing.

## 1.4 Features

### 1.4.1 General Information About the ERO•SCAN

The ERO•SCAN features:

- Screening and diagnostic measurements with TEOAE and/ or DPOAE
- Fast automatic OAE screening with Pass/Refer results and graphical displays
- 2 predefined protocols for screening version, 5 DP and 3 TE protocols for diagnostic devices
- High noise immunity for operation in normal clinical environment
- Lightweight, small ear probe
- Sharp, colored OLED display
- Wireless printing
- Various software solutions available

### 1.4.2 Licenses

The ERO•SCAN is available in versions (each with or without printer):

- ERO•SCAN Screener DPOAE
- ERO•SCAN Screener TEOAE
- ERO•SCAN Screener DPOAE + TEOAE (2 DP and 2 TE protocols with fix parameters)
- ERO•SCAN Diagnostic DPOAE
- ERO•SCAN Diagnostic TEOAE
- ERO•SCAN Diagnostic DPOAE + TEOAE (5 DP and 3 TE protocols (4 DPOAE and 2 TEOAE are customizable))

### 1.4.3 Printing Options

Printing test results from the ERO•SCAN is accomplished in a variety of ways:

- Print directly from ERO•SCAN using the optional wireless thermal printer that is available from MAICO.
- Transfer test data into a PC Software and print results using your standard printer attached to the PC.

### 1.4.4 PC-Software

The ERO•SCAN can be connected to the following PC software:

- MAICO Sessions Standalone
- MAICO Sessions with OtoAccess® Database
- MAICO Sessions with Noah Database
- MAICO Sessions with your existing Practice Management Software via GDT or XML interface
- HearSIM™ Software with OtoAccess® Database (ERO•SCAN Screener only)

## **1.5 Description**

### **1.5.1 General**

The purpose of the ERO•SCAN test system is to provide a rapid measurement and documentation of Distortion Product Otoacoustic Emissions (DPOAEs) or Transient Evoked Otoacoustic Emissions (TEOAEs) at several frequencies.

The ERO•SCAN is available as a Screening or Diagnostic version.

### **1.5.2 TEOAE**

Transient Evoked Otoacoustic Emissions (TEOAE) technology uses a click stimulus to screen patients' ears for cochlear hearing loss. The emissions are clearly related to the stimulus and therefore can be measured via a sensitive microphone placed in the patient's ear canal. The responses can be divided into frequency bands for assessment.

### **1.5.3 DPOAE**

Distortion product otoacoustic emissions (DPOAE) technology uses pairs of pure tones presented in sequence to screen patients for cochlear hearing loss. The emissions are clearly related to the stimulus and therefore can be measured via a sensitive microphone placed in the patient's ear canal.

### **1.5.4 Sensitivity and Specificity**

Sensitivity and specificity of this type of device are based on the test characteristics defined by the user and may vary depending on environmental and operating conditions. The presence of otoacoustic emissions suggests normal outer hair cell function, which in turn correlates to normal hearing. However, a passing result using this device is not an indication that the full auditory system is normal. Thus, a PASS result should not be allowed to override other indications that hearing is not normal. A full audiologic evaluation should be administered if concerns about hearing sensitivity persist. A REFER test result should not be assumed to be an indicator of a lack of auditory function, however, it should be followed with full audiologic diagnostic testing.

## 2 For Your Safety

This Section offers you important information about:

- how to read the operation manual
- where to spend special attention
- the customer responsibility
- the explanation of all regulatory symbols used
- important cautions and warnings that have to be considered during the whole time handling and operating your device

### 2.1 Reading this Operation Manual

This operation manual contains information pertinent to the use of the ERO•SCAN system including safety information as well as maintenance and cleaning recommendations.

It is highly recommended that users read the operation manual in its entirety prior to use of the ERO•SCAN device on a patient.



**READ THIS ENTIRE OPERATION MANUAL BEFORE ATTEMPTING TO USE THIS SYSTEM!**

Use this device only as described in this operation manual.

All images and screenshots are only examples and may differ in appearance from the actual device settings.

In this manual the following two labels identify potentially dangerous or destructive conditions and procedures:



**WARNING**

The **WARNING** label identifies conditions or practices that may present danger to the patient and/or user.



**CAUTION**

The **CAUTION** label identifies conditions or practices that could result in damage to the equipment



The information sign displays alternative documents or sections in this operation manual that provide more detailed information.

**NOTE:** Notes help you identify areas of possible confusion and avoid potential problems during system operation.

## 2.2 Customer Responsibility

All safety precautions given in this operation manual must be observed at all times. Failure to observe these precautions could result in damage to the equipment and injury to the operator or subject.

The employer should instruct each employee in the recognition and avoidance of unsafe conditions and the regulations applicable to his or her work environment to control or eliminate any hazards or other exposure to illness or injury.

It is understood that safety rules within individual organizations vary. If a conflict exists between the material contained in this manual and the rules of the organization using this device, the more stringent rules should take precedence.



This product and its components will perform reliably only when operated and maintained in accordance with the instructions contained in this manual, accompanying labels, and/or inserts. A defective product should not be used. Make sure all connections to external accessories are snug and secured properly. Parts which may be broken or missing or are visibly worn, distorted, or contaminated should be replaced immediately with clean, genuine replacement parts manufactured by or available from MAICO.

**NOTE:** Customer responsibility includes proper maintenance and cleaning of the device.



- Section 3.2 Maintenance
- Section 3.3 Cleaning and Disinfection Recommendations

Breach of the customer responsibility can lead to limitations of Manufacturer's Liability and Warranty.



- Section 2.3 Manufacturer's Liability
- Section 3.1 Warranty

**NOTE:** In the unlikely case of a serious incident, inform MAICO as well as the competent authority of the Member State in which the user is established.

## 2.3 Manufacturer's Liability

Usage of the device in a way deviant from the intended use will lead to a limitation or termination of the manufacturer's liability in case of damage. Improper use includes failure to follow the operating instructions, operation by unqualified personnel, and unauthorized modifications to the equipment.

## 2.4 Regulatory Symbols

The following Table 1 explains the symbols used on the device itself, on the packaging and the accompanying documents including the Operation Manual.

**Table 1 Regulatory Symbols**

REGULATORY SYMBOLS	
SYMBOL	DESCRIPTION
	Serial number
	Date of manufacture
	Manufacturer
	Caution, consult accompanying documents
	Warning, consult accompanying documents
	Information sign (reference for more detailed information)
	Return to authorized representative, special disposal required
	Reference number
	Medical Device
	UDI information: (01) GTIN (Global Trade Item Number), (11) Date, (21) Serial number
	Applied part type B according to IEC 60601-1
	Refer to operation manual (mandatory)
	Keep away from rain
	Transport and storage temperature range
	Transport and storage humidity limitations
	Transport and storage atmospheric pressure limitations
	Voltage transformer
	Do not reuse
	CE label with notified body ID
	FCC marking with ID
	Non-ionizing electromagnetic radiation
	Label Marking of Radio Equipment based on Certified Type
	Direct Current (DC)
	ETL listed mark
	Underwriters Laboratories, Inc. Label
	Logo

## 2.5 General Precautions



Before starting a measurement make sure, that the device works properly.

Use and store the device indoors only. For operation, storage and transport conditions see table in Section 6.1.



Do not drop or otherwise cause undue impact to this device. If the device is dropped or otherwise damaged, return it to the manufacturer for repair and/or calibration. Do not use the device if any damage is suspected.



Equipment is not user repairable. Repairs must be performed by a qualified service representative only.

No modifications of the equipment are allowed by anyone other than a qualified MAICO representative. Modification of the equipment could be hazardous.

No parts of the equipment can be serviced or maintained while in use with the patient.



Use of accessories, transducers and cables other than those specified or provided by the manufacturer of this equipment could result in increased electromagnetic emissions or decreased electromagnetic immunity of this equipment and result in improper operation.

The list of accessories, transducers and cables can be found in:



Section 6.4 Electromagnetic Compatibility (EMC)



Do not immerse the device in any fluids. Should the user suspect fluids have contacted the system components or accessories, the device should not be used until deemed safe by a MAICO certified service technician.

## 2.6 Electrical and Electrostatic Safety

This device complies with part 15 of the FCC Rules. Operation is subject to the condition that this device does not cause harmful interference.



This icon indicates that applied parts of the device conform to IEC 60601-1 Type B requirements.



In case of emergency, disconnect the device from the computer.

In Case of Emergency



In Case of Emergency

In case of emergency, disconnect the device from power supply. Position the device in such a way that it can be easily disconnected from the power supply at any time.



Do not use the device if the power supply unit is damaged.

Data transfer to the PC can be done via wireless connection or USB connection.

To learn how to safely establish a connection via USB with a power supplied PC or laptop (medical device/nonmedical device) or to a battery-driven laptop see:

 Section 4.5 Establishing a PC Connection



This equipment is intended to be connected to other equipment thus forming a Medical Electrical System. External equipment intended for connection to signal input, signal output or other connectors shall comply with the relevant product standard e.g. IEC 62368-1 for IT equipment and the IEC 60601-series for medical electrical equipment. In addition, all such combinations – Medical Electrical Systems – shall comply with the safety requirements stated the general standard IEC 60601-1, edition 3, clause 16. Any equipment not complying with the leakage current requirements in IEC 60601-1 shall be kept outside the patient environment i.e. at least 1.5 m from the patient support or shall be supplied via a separation transformer to reduce the leakage currents. Any person who connects external equipment to signal input, signal output or other connectors has formed a Medical Electrical System and is therefore responsible for the system to comply with the requirements. If in doubt, contact qualified medical technician or your local representative.



If the device is connected to a PC (IT equipment forming a system) ensure not to touch the patient while operating the PC. Do not touch the patient and the printer at the same time.



A Separation Device (isolation device) is needed to isolate the equipment located outside the patient environment from the equipment located inside the patient environment. In particular such a Separation Device is required when a network connection is made. The requirement for the Separation Device is defined in IEC 60601-1 clause 16.



If the device is connected to a PC (IT equipment forming a system) assembly and modifications shall be evaluated by qualified medical technician according to safety regulations in IEC 60601-series.



Do not touch the contacts of the device and the patient at the same time.

If the device is connected to a PC (IT equipment forming a system) do not touch the patient and the IT equipment at the same time.

The consequence of not following this warning could be a too high leakage current to the patient.



The device is not intended for operation in areas with an explosion hazard. Do NOT use the device in a highly oxygen rich environment, such as a hyperbaric chamber, oxygen tent, etc. If the device is not used turn it off and disconnect it from the power supply.

Never short-circuit the terminals.



To avoid the risk of electric shock, this equipment must only be connected to the medical power supply originally delivered by MAICO. Using another power supply can also lead to electrical damage on the device.



Prevent cable breakage: cables must not be bent or buckled.




Do not open the case of the ERO•SCAN device. Refer servicing to qualified personnel.

## 2.7 Electromagnetic Compatibility (EMC)



The device fulfills the relevant EMC requirements. Avoid unnecessary exposure to electromagnetic fields, e.g. from mobile phones etc. If the device is used adjacent to other equipment it must be observed that no mutual disturbance appears.

Also refer to EMC consideration in:

 Section 6.3 Calibration Values



The ERO•SCAN has been verified by an independent laboratory to conform to international standards for EMC (electromagnetic emissions and immunity). The user is advised to avoid installation and use of this device in proximity with other devices or equipment that may emit or be susceptible to electromagnetic interference, including mobile phones. If the device is used adjacent to other devices or equipment, the user is instructed to verify that no disturbance is found in the operation of this or other equipment in proximity. It may be necessary to take mitigation measures, such as reorienting or relocating the ERO•SCAN or shielding the location.

## 2.8 Battery Safety and Capacity

### 2.8.1 Battery Safety



#### Explosion hazard

The internal battery must be only replaced by an authorized service representative. Damage to the electronics resulting from an attempt to change the battery by someone other than an authorized representative will not qualify for repair under the product warranty.

### 2.8.2 Battery Capacity

The capacity of the battery will degrade over time with repeated charging/discharging cycles. The need to replace the battery due to diminishing capacity depends on usage patterns.

To extend battery capacity, do not allow the battery to fully discharge. To learn how to charge the device correctly see:



Section 4.4      Battery Charging

## 3 Warranty, Maintenance and After-Sales Service

This Section offers you important information about:

- warranty conditions
  - maintenance
  - cleaning and disinfection recommendations
  - accessory and replacement parts
  - recycling and disposal of the device
- 

### 3.1 Warranty

The MAICO ERO•SCAN is guaranteed for at least 1 year.

This warranty is extended to the original purchaser of the device by MAICO through the distributor from whom it was purchased and covers defects in material and workmanship for a period at least 1 year from date of delivery of the device to the original purchaser.

The device shall only be repaired and serviced by your distributor or by an authorized service center. Opening the device case voids the warranty.

In the event of repair during the guarantee period, please enclose evidence of purchase with the device.

### 3.2 Maintenance

To ensure that the device works properly, it must be checked and calibrated at least once every 12 months.

The service and calibration must be performed by your distributor, or a service center authorized by MAICO.

When returning the device for repairs or calibration it is essential to send the acoustic transducers with the device. Please include a detailed description of faults. In order to prevent damage in transit, please use the original packaging when returning the device.

## 3.3 Cleaning and Disinfection Recommendations

### 3.3.1 General



Also, check-out our training videos:

ERO•SCAN | OAE | MAICO Training | Hearing Screening Diagnostic Tests - Cleaning the device

<https://youtu.be/oVSkbMbuKrc?si=2Z5YXalG1vP9sBby&t=246>

It is recommended that parts (device and accessories) which come in direct contact with the patient be subjected to standard cleaning and disinfecting procedure between patients.

Recommendations for cleaning and disinfection of MAICO device presented in this document are not intended to replace or contradict policies in effect or procedures required for infection control at the facility.

If there is not a high infection potential, MAICO recommends:

- Before cleaning always turn off and disconnect the device from power supply.
- Remove disposable ear tips or probes prior to disinfection.
- For cleaning use a lightly dampened cloth with soap water solution.
- Disinfect the device and its accessories by wiping the surfaces with dampened disinfection wipes. Follow the instructions on the specific disinfection product.
  - Wipe before and after each patient
  - After contamination
  - After infectious patients



To avoid damage of the device and its accessories, please mind the following:

- Do not autoclave or sterilize the device or probes.
- Do not use the device in the presence of fluid that can come into contact with any of the electronic components or wiring.

Should the user suspect fluids have contacted the system components or accessories, the device should not be used until deemed safe by a MAICO certified service technician.

Do not use hard or pointed objects on the device or its accessories.

For more detailed cleaning recommendations see the following Sections and follow the instructions on the items that are relevant for your system.

**NOTE:** Long-term exposure to any disinfecting agents has the potential to alter the material properties of the plastic housing and labeling of the device.

Always follow the safety and disposal guidelines given by the manufacturer of cleaning and disinfectant chemicals.

### 3.3.2 Cleaning and Disinfecting the Screen

Use a lens cleaning or microfiber cloth to clean the screen.

Disinfect the screen of the device by wiping the surfaces with dampened disinfection wipes.

### 3.4 Safe Use of Disposables



Figure 1

Operating the ERO•SCAN requires the use of ear tips (Figure 1) and probe tubes.



Ear tips are intended for single use only. They must be discarded after use. They cannot be cleaned.



If you use the same ear tips for different patients, you enhance the risk of bacterial cross contamination which can cause serious infections!

Replace the ear tips with unused ones at the end of each patient's examination and discard the used ones.

**IMPORTANT NOTE:** All disposable supplies included with the ERO•SCAN are produced by Sanibel Supply. The system has only been tested using disposables supplied by Sanibel Supply. Use of other supplies could alter the behavior and results obtained with the device and is not recommended. Sanibel disposables are latex, DEHP and BPA free and have been tested for biocompatibility. Data sheets are available upon request.

### 3.5 Disposables – Ear Tips

#### 3.5.1 Applying Ear Tips

The ERO•SCAN comes with a box of disposable ear tips that fit a variety of ear canal sizes. The probe tube must have an ear tip attached before inserting it into an ear canal.

**NOTE:** See Section 5.7.2 on how to properly insert a probe into the patient's ear canal.



Figure 2

Choose an ear tip that is appropriate for the patient's ear volume. Push the ear tip onto the probe until it is flush against the base of the probe (Figure 2). Twisting the ear tip slightly while pushing it onto the probe is recommended. Be sure the ear tip is fully seated on the probe.



Figure 3

There shall be no gaps between the ear tip and the collar of the probe head (Figure 3).

### 3.5.2 Removing Ear Tips




Figure 4

To remove the ear tip, grasp the ear tip at the base using the ear tip removal tool and pull it smoothly straight off the probe tube (Figure 4). If you do not have a Removal tool, grasp the ear tip at the base with your fingers and twist it while pulling off the probe tube. Grasping the base of the ear tip will prevent the probe tube from being inadvertently pulled out of the probe head along with the ear tip.

### 3.6 Disposables – Probe Tubes



Also, check-out our training videos:

ERO•SCAN | OAE | MAICO Training | Hearing Screening Diagnostic Tests - Changing the probe tube

<https://youtu.be/oVSkbMbuKrc?si=3tteec6YDRB4Lp4&t=202>

#### 3.6.1 Safe Use of Probe Tubes

Probe tubes should be replaced when clogged.



CAUTION

Do not attempt to clean probe tubes. If the probe tube is re-used after it was removed from the probe it can damage the probe head since it will not sit as tight as before.

#### 3.6.2 Probe Tube Removal

Use the Probe Tube Removal Tool for replacement of probe tubes.



Figure 5




Figure 6

1. Place the front hole of the Probe Tube Tool over the end of the probe, as shown. The probe should be seated against the face of the tool and snap in place.
2. Squeeze the tool closed and hold it.
3. Twist the tool a couple of times while holding the tool closed and pulling away from the probe (Figure 5).
4. The probe tube will pull out from the probe. Discard the probe tube (Figure 6).

### 3.6.3 Applying a New Probe Tube



Figure 7

1. Place a new tube into the hole on the top of the tool.
2. Align the probe with the probe tube (Figure 7).
3. Push the probe onto the probe tube until the probe contacts the face of the probe tube tool and a snapping sound is heard (Figure 8).
4. Slide the tool off leaving the tube in place in the probe.



Figure 8

## 3.7 Troubleshooting

If problems occur while working with your ERO•SCAN, you can usually easily solve them yourself. Follow the instructions in Table 2 for general problems or Table 3 if display messages are shown.

Contact MAICO for service if a problem persists.

Table 2 Troubleshooting


Problem	Explanation
<b>Device does not turn on</b>	<ul style="list-style-type: none"> <li>• Press the ▼<b>DOWN</b> arrow for a full second (the yellow LED (<b>TEST</b>) illuminates).</li> <li>• Connect the charger as shown in Figure 13 in Section 4.4. Confirm that the blue LED (<b>CHARGE</b>) is illuminating in a slow blink pattern. Wait at least 10 minutes and then attempt to turn on the device.</li> </ul>
<b>Test does not start</b>	<ul style="list-style-type: none"> <li>• Select a different sized ear tip.</li> <li>• Reposition the probe.</li> <li>• Change the probe tube.</li> <li>• Verify that the ear tip is sealed in the ear canal via feedback from the PROBE CHECK screen.</li> <li>• Try if the device starts in your own ear with the proper ear tip. If the test does not start or if the AutoStart tones sound unusual, replace the probe tube.</li> </ul>

Problem	Explanation
<b>Results are not printed (HM-E 200)</b>	<ul style="list-style-type: none"> <li>• Check the printer status. Turn the printer on (wake from sleep mode) by pressing the <b>power/MODE</b> button.</li> <li>• If the printer does not turn on, plug in the power supply unit to charge the battery.</li> <li>• Make sure the printer has paper.</li> <li>• If paper comes out of the printer, but nothing is printed on the paper, the paper roll may have been inserted the wrong way round.</li> <li>• Run a self-test: Make sure the printer is turned off. Press and hold the <b>FEED</b> button and simultaneously press the <b>power/MODE</b> button to turn on the device and run the self-test.</li> </ul>
<b>Display is frozen and device does not respond to button presses</b>	<ul style="list-style-type: none"> <li>• Press and hold the <b>▼DOWN</b> arrow button for 10 seconds to force the device to turn off. The device should function normally the next time it is turned on.</li> </ul>
<b>Numerical results show a line rather than a number</b>	<p>Dashes are displayed when there is excessive noise while testing.</p> <ul style="list-style-type: none"> <li>• Reduce the noise and retest.</li> </ul> <p>Dashes are displayed in the DP and SNR column when the signal level is at a minimum of 5 dB over target level.</p> <ul style="list-style-type: none"> <li>• Reinsert the probe and retest.</li> </ul>

**Table 3 Display Messages**

Display Message	Explanation
<b>Attach Probe</b>	<p>Probe is not detected.</p> <ul style="list-style-type: none"> <li>• Check that the probe connector is fully seated in the socket.</li> <li>• Disconnect and reconnect the probe.</li> <li>• Restart the device.</li> </ul>
<b>Device not Responding</b>	<p>The printer is not responding to queries from the device.</p> <ul style="list-style-type: none"> <li>• Check printer status.</li> <li>• Awaken printer from sleep mode.</li> <li>• Charge printer battery if necessary.</li> </ul>

Display Message	Explanation
<b>Due For Service</b>	<p>Indicates that the calibration of the device is recommended. The message appears upon the calibration due date set in the device. Message appears during device startup.</p> <ul style="list-style-type: none"> <li>• Have the device calibrated by a service technician authorized by MAICO.</li> </ul>
<b>Fit Error Cannot Obtain L</b>	<p>For a DP test, the desired level (L1 or L2) cannot be obtained within allowable limits.</p> <ul style="list-style-type: none"> <li>• Refit the probe and retry the test.</li> <li>• Replace the probe tube.</li> </ul>
<b>Fit Error Too High</b>	<p>For a DP test, the level of the calibration tone is too high.</p> <ul style="list-style-type: none"> <li>• Refit the probe and retry the test.</li> <li>• Replace the probe tube.</li> </ul>
<b>Fit Error Too Low</b>	<p>For a DP test, the level of the calibration tone is too low. User should refit the probe and retry the test.</p> <ul style="list-style-type: none"> <li>• Refit the probe and retry the test.</li> <li>• Replace the probe tube.</li> </ul>
<b>Limit Error</b>	<p>Overflow error during the calculation of the DFTs for a DP test.</p> <ul style="list-style-type: none"> <li>• Repeat the test.</li> <li>• Restart the device.</li> </ul>
<b>Memory Almost Full</b>	<p>Saved tests are within 5 tests of the maximum limit.</p> <ul style="list-style-type: none"> <li>• Print or transfer test result to avoid interruption in testing.</li> </ul>
<b>Memory Full!</b>	<p>The maximum saved test limit is reached.</p> <ul style="list-style-type: none"> <li>• Clear the memory before starting a new test.</li> </ul>
<b>Power Low!</b>	<p>The battery charge level is too low for operation.</p> <ul style="list-style-type: none"> <li>• Charge the battery before starting a new test.</li> </ul>
<b>Printer Error</b>	<p>Indicates a problem with the printer. Check the printer status.</p>
<b>Printer Paper Out!</b>	<p>Indicates that printer paper has run out.</p> <ul style="list-style-type: none"> <li>• Replace the paper roll.</li> </ul>

Display Message	Explanation
<b>Time/Date Error</b>	<p>The clock is checked during power on to ensure it has not lost time and been reset. In the case of clock reset, this message is shown.</p> <ul style="list-style-type: none"> <li>Set the correct date/time.</li> </ul>
<b>Wireless Device Not Found</b>	<p>The paired wireless device cannot be detected. The device may be turned off or too far away.</p> <p>Paired to Printer (HM-E 200):</p> <ul style="list-style-type: none"> <li>Check that the printer is turned on.</li> <li>Move closer to the printer.</li> <li>Turn the printer off and on again.</li> <li>Try again.</li> </ul>
<b>Wireless Error #xxx</b>	<p>There is an error condition with the wireless device.</p> <ul style="list-style-type: none"> <li>Check if the wireless connection to the printer is active.</li> <li>Attempt to connect to wireless device again.</li> </ul>
<b>Wireless Not Configured</b>	<p>Printing has been attempted, but no wireless device is paired with the ERO•SCAN.</p> <ul style="list-style-type: none"> <li>Pair the ERO•SCAN with the wireless device.</li> </ul>
 Section 5.10.2.2 Wireless Device Pairing	

### 3.8 Recycling/Disposal



Within the European Union it is illegal to dispose of electric and electronic waste as unsorted municipal waste. According to this, all MAICO products sold after August 13, 2005, are marked with a crossed-out wheeled bin. Within the limits of Article (9) of DIRECTIVE 2012/19/EU of the European Parliament and of the Council of 4 July 2012 on waste electrical and electronic equipment (WEEE), MAICO has changed their sales policy. To avoid additional distribution costs we assign the responsibility for the proper collection and treatment according to legal regulations to our customers.

Non-European countries

Outside the European Union, local regulations should be followed when disposing of the product after its useful life.



Batteries may explode or cause burns, if disassembled, crushed, or exposed to fire or high temperatures.

## 4 Unpacking and Hardware Orientation

This section provides information on:

- **unpacking the system**
  - **becoming familiar with the hardware inclusive connections**
  - **system assembly**
  - **using the printer**
  - **how to power the ERO•SCAN**
  - **how to store the device**
- 

### 4.1 Unpacking the System

#### **Check Box and Contents for Damage**

- It is recommended that you unpack your ERO•SCAN carefully making sure that all components are removed from the packaging materials.
- Verify that all components are included as shown on the packing list included with your shipment.
- If any component is missing, contact your distributor immediately to report the shortage.
- If any component appears to be damaged in shipment, contact your distributor immediately to report it. Do not attempt to use any component or device that appears to be damaged.

#### **Reporting Imperfections**

Notify the carrier immediately if any mechanical damage is noted. This will ensure that a proper claim is made. Save all packaging material, so the claim adjuster can inspect it as well.

#### **Report Immediately any Faults**

Any missing part or malfunction should be reported immediately to the supplier of the device together with the invoice, serial number, and a detailed report of the problem.

#### **Keep Packaging for Future Shipment**

Save all the original packaging material and the shipping container, so the device can be properly packed if it needs to be returned for service or calibration.

The ERO•SCAN comes with different components (see Table 4 and Table 5). The availability of configurations with the following components are country-specific. Contact your local distributor for more information.

**Table 4 List of Components**

List of Components
ERO•SCAN Device incl. Battery
Micro USB Power Supply Unit for Charging the Lithium-Ion Battery (UES12LCP-050160SPA)
Micro-Probe*
Ear Tip Removal Tool
Probe Tube Removal Tool
Printer HM-E200 Kit (Includes 2 Rolls of Thermal Paper, Printer Power supply Unit/Charger with Plug Adapters (5 V/1.6 A) UES12LCP-050160SPA) and Battery Pack
Carry Case
Operation Manual**
Quick Guide**

\*Applied parts according to IEC 60601-1

\*\*As download from the download center – see accompanying leaflet

## Disposables Supplied

**NOTE:** MAICO strongly recommends using Sanibel ear tips for reliable results.

**Table 5 Disposables**


Disposables
Ear Tip Kit (120 pcs) including replacement probe tubes and removal tool

## 4.2 Hardware Orientation

The ERO•SCAN system consists of the following components (configuration-dependent):

1. ERO•SCAN device
2. Micro-Probe
3. disposable ear tips (for single use only)
4. probe tubes
5. Micro-USB cable
6. Thermal Printer

Figure 9 shows the ERO•SCAN device.



**Figure 9**

### 4.3 Handling the Micro-Probe



Also, check-out our training videos:

ERO•SCAN | OAE | MAICO Training | Hearing Screening Diagnostic Tests - How to set up the ERO•SCAN

<https://youtu.be/oVSkbMbuKrc?si=z1n1K4kPOQbDYBCR&t=7>

#### The Micro-Probe System




Figure 10

The Micro-Probe houses speaker and microphone which produce test stimuli and measure the sound pressure level (SPL) present in the sealed ear canal.

The Micro-Probe system (Figure 10) consists of:

- 1 = probe body
- 2 = probe tube
- 3 = ear tip

Handling and replacement of the probe tube and the ear tips is explained in:



Section 3.4 Safe Use of Disposables



Section 3.6 Disposables – Probe Tubes

#### Connecting the Micro-Probe with the ERO•SCAN



Figure 11

Turn off the ERO•SCAN and insert the Micro-Probe connector into the socket on the top of the ERO•SCAN (Figure 11). The plug fits in only one direction. A MAICO logo is on the probe connector and aligns with the device control panel.

**NOTE:** Misalignment of the plug and socket can cause damage. The plug and socket should be visually inspected prior to each installation of the probe. If damage is observed, contact MAICO Diagnostics.

### 4.4 Battery Charging




Figure 12

The ERO•SCAN is powered by an integrated rechargeable lithium-ion battery (1700 mAh) providing 15 hours (1000 tests, minimum) of operation between full charging. The battery status is indicated by the battery icon shown in the upper right corner of the **Main Menu** (Figure 12).

Full battery charge is represented by a full battery symbol on the display and reduces to an empty battery in increments corresponding to the discharge of the battery.

**NOTE:** Battery life varies depending on the product configuration.

The maximum capacity of this battery decreases with time and usage.

For maximum battery life, do not completely deplete the battery, but charge when 5 % to 10 % of the battery remains.



Figure 13

The Micro-USB port on the bottom of the device is used for charging via the power supply unit. Connect the charger as shown in Figure 13.

**NOTE:** Misalignment of the plug and socket can cause damage. The plug and socket should be visually inspected prior to each installation of the charging cable. If damage is observed, contact MAICO Diagnostics.



Figure 14

The blue LED (**CHARGE**) (Figure 14) provides a visual indication of the battery status during operation or charging. The status is indicated as follows:

Two fast blinks followed by a pause	Battery is low. Charging is needed.
Slow blinking	Charging is in progress.
Fast blinking	Indicates an error. Contact your local representative or MAICO for help.
Steady	Battery is fully charged.

### 4.5 Establishing a PC Connection via USB

Data transfer to a PC can be done via USB connection. If the device is used with office equipment that is not medical electrical equipment (ME equipment) itself (see Table 3, PC Connection 1), make sure to establish the PC connection in one of the following ways (see Table 3, PC Connection 2, 3 or 4).



**WARNING**

Make sure you use only office equipment with the device that is medical electrical equipment itself or meets the requirements of IEC 62368-1. If nonmedical electrical equipment is used within the patient environment (1.5 m from the patient as defined in IEC 60601-1) an isolation transformer must be used (exception: a battery-driven laptop is used).

**Table 3 PC Connections**

PC CONNECTIONS	
<p><b>PC Connection 1:</b> ME equipment – ME equipment</p>	<p><b>PC Connection 2:</b> ME equipment – Non-ME equipment</p>
<p><b>PC Connection 3:</b> ME equipment – Non-ME equipment</p>	<p><b>PC Connection 4:</b> ME equipment – Laptop (battery-driven)</p>

### 4.6 Establishing a Wireless PC Connection

If you want to use the wireless connection to MAICO Sessions, see:



Section 5.10.2.2 Wireless Device Pairing

## 4.7 Using the Thermal Printer HM-E200

### 4.7.1 Connecting the Thermal Printer to ERO•SCAN

The connection of the ERO•SCAN and the thermal printer is made via wireless pairing.

 Section 5.10.2.2 Wireless Device Pairing

### 4.7.2 Powering the Thermal Printer



Figure 15

The thermal printer is powered by a Lithium-ion battery. Use the micro-USB power supply unit delivered by MAICO to power the thermal printer (Figure 15).

### 4.7.3 Insert Paper Rolls into the Thermal Printer

The printer indicates that it has run out of paper by displaying the message **"Out of paper"** on the screen and the blue LED (ERROR) flashes (Figure 16).

Open the printer by pressing the small latch button (Figure 17).

Insert the paper roll into the printer with the paper end placed towards the open cover. Hold the paper end in place and close the cover. Turn the printer on and press the feed button on the left side so that the printer can properly align the paper with the print head (Figure 18).



Figure 16



Figure 17



Figure 18

#### 4.7.4 Using the Thermal Printer

##### Power on

Push **power/MODE button**  for 2 s to power on or off.

Acoustical feedback will be presented upon power on and power off.

---

**NOTE:** Selecting **Print** on the ERO•SCAN when the printer is off will result in an error message. The printer must be turned on and in close proximity of the ERO•SCAN for printing to proceed.

---

##### Self-test

Make sure the printer is turned off. Press and hold the **FEED** button and simultaneously press the **power/MODE** button to turn on the device and run the self-test.

The printer will be ready to receive data after the self-test finishes.

##### Paper feed

When powered press the **FEED** button. Paper will feed as long as the button is pressed.

---

**NOTE:** Reorder paper from MAICO or your local distributor.

---

#### Connecting the Thermal Printer to ERO•SCAN

The connection of the ERO•SCAN and the printer is made via wireless pairing from the **Settings** menu.



Section 5.10.2.2 Wireless Device Pairing

---

**NOTE:** It is possible to pair 6 devices with one printer. Do not have several printers powered on and within range while searching.

---

#### 4.8 Storage

When the ERO•SCAN is not in use, store it in the carry case or in a location where it will be safe from damage to the screen or other sensitive parts. Store according to the recommended temperature conditions.



Section 6.1 ERO•SCAN Hardware – Device SPECIFICATIONS –  
Transport & Storage environment

## 5 Operating the Device

This Section provides information on:

- how to get started with the ERO•SCAN
- the device layout
- the function keys
- performing the measurement methods of Audiometry
- patient management
- documentation of results
- changing settings in the user menu

### 5.1 Getting Started with the ERO•SCAN

#### 5.1.1 Use of Equipment After Transport and Storage

Make sure the device is functioning correctly before use. If the device has been stored in a colder environment (even for shorter time) allow the device to become acclimatized. This can take a long time depending on the conditions (like environmental humidity). You can reduce the condensation by storing the device in its original packaging. If the device is stored under warmer conditions than the use conditions, no special precautions are required before use. Always ensure proper operation of the device by following routine check procedures for audiometric equipment.

#### 5.1.2 Where to Setup



Also, check-out our training videos:

ERO•SCAN | OAE | MAICO Training | Hearing Screening Diagnostic Tests - How to set up the test environment for hearing screening

<https://youtu.be/oVSkbMbuKrc?si=M22QKbhHRyXRgLDH&t=67>

The ERO•SCAN should be operated in a quiet room, so that the audiometric examinations are not influenced by outside noises. Ambient sound pressure levels in an audiometric test room shall not exceed the values specified in the norm ISO 8253-1 or ANSI S3.1.

#### 5.1.3 Noise Sources

When the noise level exceeds the noise rejection limit of the device, the orange 'NOISE' light will appear. It is common for the 'NOISE' light to appear while testing. The light will appear infrequently if the noise level in the ear canal is low, and it will appear more often if the noise level in the ear canal is high. Otoacoustic emissions are very low-level sounds. Any noise in the ear canal at the time of testing can mask this emission. This noise can come from a variety of sources.



Section 5.2 Indication Lights

The largest source of noise can come from the patient. This is biological noise, such as movement, coughing, sucking, talking, etc. The patient must be calm and not move or talk. Ambient noise in the test environment can also be a large source of noise during the test. A properly sealed ear tip can block a large amount of this noise, but performing the testing in a relatively quiet environment is recommended.







## 5.2 Indication Lights



The device has 4 indication lights (LED, Figure 19) that help you to know the actual state of the device. Table 7 give explanation to the indication lights.

Figure 19

Table 7 Indication Lights

Indication Light	Explanation
<p><b>NOISE / Red</b></p> 	<p>The noise level measured during the test exceeds a nominal threshold of 55 dB SPL. Test duration will be prolonged.</p> <p> Section 5.1.3 Noise Sources</p> <p>Also used to indicate some error conditions and when the result of test is <b>Refer</b>, <b>Noisy</b>, or <b>No Seal</b>.</p>
<p><b>TEST / Orange</b></p> 	<p>The selected test is being performed. This indication light remains lit throughout the test procedure.</p>
<p><b>READY / Green</b></p> 	<p>The device is ready to perform a test.</p>
<p><b>CHARGE / Blue</b></p> 	<p>Shows the battery recharging function and battery status. The rate of illumination provides a means of identifying the status of the charging function.</p> <p> Section 4.4 Battery Charging</p>

## 5.3 Control Panel



Figure 20

The ERO•SCAN uses 4 buttons to control all functions of the device (Figure 20). These buttons are arranged in a directional cursor format. The arrows on the keypad (**◀LEFT**, **▶RIGHT**, **▲UP**, and **▼DOWN**) correspond to the arrows that are used on the screen. The screen will indicate which button to push by showing the appropriate arrow.

---

**NOTE:** The **▲UP** key will always bring you back to either the previous menu or the **Main Menu**. The **▲UP** key will also access the print command from the **Main Menu**.

---

## 5.4 Turning On the Device



Figure 21

To turn on the ERO•SCAN press the **▼DOWN** key located below the display (Figure 20). The yellow light (**TEST**) will appear briefly just above the display screen. The green light (**READY**) will remain on indicating the device is ready for use. The Splash Screen (Figure 21) appears briefly. It shows

- the device version – Screening Version (SCR), Screener Plus (SC+), Standard (STD), or Combo (CMB)
- the software version (e.g., 105.05) and
- the serial number (for example ME1234567)




Figure 22

If the battery is sufficiently charged, the ERO•SCAN powers on and automatically checks the date and time. If there are no date/time errors detected, the **Main Menu** is shown (Figure 22).

**NOTE:** If the date or time shown is wrong you can change it in the device settings menu.



Section 5.10.1 Clock and Date Settings

## 5.5 Turning Off the Device

### Manual Turn-Off

Press **▲UP** to turn off the device.

### Automatic Shutdown

The ERO•SCAN has an automatic shutdown feature, designed to prolong battery life. The device automatically shut down after 1 minute (default) of inactivity. To turn it back on, press the **▼DOWN** key.

**NOTE:** You can change the time for automatic shutdown in the device settings menu.



Section 5.10.2.4 Auto Shutdown Time

## 5.6 Main Menu

Figure 23 gives explanation to the **Main Menu**.




Figure 23

## 5.7 Testing Procedure

### 5.7.1 Calibration and Test

The ERO•SCAN automatically performs a calibration prior to the start of each test. During calibration test signals are presented to the ear canal to calibrate the levels of the frequencies to be tested. After calibration the test phase starts automatically.

**NOTE:** The testing procedure can be stopped. Press: **▲UP** to stop testing.

No record of an aborted test is saved in the internal memory.

The ERO•SCAN allows the user to select from two options for viewing the results:

- the **Graph SNR** view
- the **Graph Value** view.




Figure 24

#### Graph SNR View

The **Graph SNR** view (Figure 24) shows the signal-to-noise ratio for each DP test frequency or TE test band.

The set of bars display the emissions measured. This can be reviewed after a measurement. Each column represents one test frequency (DP) or frequency band (TE).




Figure 25

#### Graph Value View

The **Graph Value** view shows the absolute emission and noise levels for each DP test frequency or TE test band (Figure 25).



Section 5.10.2.8 Graph Style (Settings)


### 5.7.2 Selecting the Test Protocol



Also, check-out our training videos:

ERO•SCAN | OAE | MAICO Training | Hearing Screening Diagnostic Tests - How to handle the ERO•SCAN

<https://youtu.be/oVSkbMbuKrc?si=9IMn5PFhrfBqzquv&t=28>




The currently selected protocol is shown on the **Main Menu** (Figure 26). To select another protocol press:

▼ **CHANGE** to proceed to the next screen (Figure 26).

◀ **CHANGE** ▶ to select another protocol (Figure 27).

Figure 26



▲ **UP** to return to the **Main Menu** and start testing OR

**SETUP** ▼ to enter the **Setup** menu.

Figure 27

**NOTE:** Some protocols of the Diagnostic Version are customizable.



Section 6.6 Configurations and Test Protocols

Section 5.10.3 Advanced Settings Menus (Diagnostic Version only)

### 5.7.3 Preparing for Testing



**WARNING**

Keep in mind the indications and contraindications of use given in



Section 1.2 Indications for Use Statement

Section 1.3 Contraindications of Use Statement

During the testing process it is important that the patient is calm and relaxed. This is often difficult to achieve when testing babies. The following suggestions can help you prepare a infant for the testing process:

Hearing screening is most successfully and efficiently performed on a quiet, sleeping infant. If the infant is awake but quiet or sucking intermittently, testing is possible though the test time may be affected. If the infant is crying, moving, or sucking vigorously and constantly then the test will be prolonged and the chance of a Refer result will be increased. In this case it would be best to terminate the screening and return when the infant is sleeping.

Screening can be performed when the infant is lying in a crib, in a car seat or is being held by the screener or parent. The key is to make the infant comfortable and quiet for the screening. Swaddling the infant in a blanket with the arms wrapped inside is recommended. This will calm the infant and keep the infant from interfering with the screening device components.

**Otoscopic examination**

Otoscopic examination of the patient’s ear canals should be performed prior to testing. Excessive cerumen or vernix in the ear canals may interfere with the test and give invalid or incomplete results. Patients with excessive cerumen, debris, or foreign bodies in the ear canals should be referred to a qualified professional for removal of the blockage prior to testing.

Place the patient in a position that will allow easy access to the ear canal. Use the shirt clip on the probe to secure the probe to clothing or bedding. The patient should remain still and quiet while the test is being performed.

**Attaching a Probe Tube and an Ear Tip**



**WARNING** If you insert the probe tip into the patient's ear without having an ear tip applied, the adapter can scratch the patient's ear.

Always apply an ear tip before you insert the OAE probe tip into the patient’s ear!



**WARNING** If you use the same probe tube and ear tip for different patients, you enhance the risk of bacterial cross contamination which can lead to infections.

Always use a new probe tube and a new ear tip for a new patient and discard the used ones.



Select an ear tip that is appropriate for the patient's ear. Apply the ear tip onto the probe tip

**Figure 28**

**NOTE:** Use MAICO’s special tools for attaching and/or removing the ear tips and probe tubes:



- Section 3.4 Safe Use of Disposables
- Section 3.6 Disposables – Probe Tubes

## Placing the Probe in the Ear Canal



Also, check-out our training videos:

[ERO•SCAN | OAE | MAICO Training | Hearing Screening Diagnostic Tests - How to position the probe](#)

<https://youtu.be/oVSkbMbuKrc?si=SLsNUOgam-2sNd8l&t=86>



Figure 29

Insert the ear tip into the patient's first ear to test. For newborn hearing screening do this by pulling gently down and out on the patient's ear lobe to open the ear canal, for older patients pull the patient's ear lobe up and back. Hold the probe and aim and twist (gently) the ear tip into the ear canal. The fit of the ear tip should be secure; not superficial. Release the earlobe. You should not hold the probe during the measurement since this can cause acoustic noise (Figure 9).

### 5.7.4 Starting a Test (Probe Check)



Also, check-out our training videos:

[ERO•SCAN | OAE | MAICO Training | Hearing Screening Diagnostic Tests - OAE Measurement](#)

[https://youtu.be/oVSkbMbuKrc?si=yOjDOiFzQi7EC\\_0w&t=126](https://youtu.be/oVSkbMbuKrc?si=yOjDOiFzQi7EC_0w&t=126)

**IMPORTANT NOTE:** To test children with Pressure Equalizer (PE) tubes, the Probe Check needs to be omitted. Proceed as follows to disable the **AutoStart** function:

- Hold down the **<LEFT or RIGHT>** arrow keys for 3 seconds until the green LED (**READY**) turns off.

Once the key is released, the ERO•SCAN will calibrate and test as normal.

### Probe Check





To start a test, insert the probe into the ear and select either the **<LEFT or RIGHT>** arrow key to indicate which ear will be tested.

After the test ear is selected, the **PROBE CHECK** screen is shown. This screen shows if

- you have selected a proper ear tip size for the patient's ear canal volume.
- if you have placed the probe and the ear tip well to obtain a seal.

See Table 8 for the various status displays on the **PROBE CHECK** screen.

Table 8 PROBE CHECK Screen

Screen	Explanation
 <p>Figure 30</p>	The ear canal volume of the patient is too large or the probe is not inserted into the ear or in place properly (Figure 30).
 <p>Figure 31</p>	Volume is too small for the test to begin or probe tip is blocked (Figure 31).
 <p>Figure 32</p>	The ear canal volume is in the target area for testing, but the bars stay grey until the seal is also detected (Figure 32).
 <p>Figure 33</p>	The ear canal volume is in the target area for testing and a seal is detected. The test starts automatically ( <b>AutoStart</b> ) as soon as the probe fit is stable (Figure 33).

When a seal is obtained, the device will automatically begin testing (**AutoStart**), and the yellow LED (**TEST**) lights up during the test procedure.

If the probe check is not successful, try the following:

1. Select an ear tip that fits the ear canal of the patient better.
2. Try once more to place the probe (see description above).

If you can still not start the test proceed as follows:

- Turn off the device.
- Change the probe tube.
- Change the ear tip.
- Make sure that the probe connector is seated properly in the socket.
- Try again to start a test.

### 5.7.5 TEOAE and Pause Function

The ERO•SCAN incorporates a pause function within the TEOAE test when a leak is detected. When the detection occurs, all three LED lights at the top of the device (Noise, Test and Ready) will illuminate to display the device is in pause mode. The pause mode will wait for up to 30 seconds. This allows the tester to reposition the probe without starting the test all over. If the seal is not able to be obtained within the 30 seconds a **No Seal** error will display and the probe should be repositioned, and test restarted.

5.7.6 Viewing Test Results



Also, check-out our training videos:

ERO•SCAN | OAE | MAICO Training | Hearing Screening Diagnostic Tests - How does the test result look with ERO•SCAN

[https://youtu.be/oVSkbMbuKrc?si=\\_ePk7qDKtGB11uEu&t=177](https://youtu.be/oVSkbMbuKrc?si=_ePk7qDKtGB11uEu&t=177)




Figure 34

When testing is complete, a result screen (e.g. Figure 34) appears. The result screen shows the tested ear as well as one of the following test results:

- Pass** The patient passed the screening.  
No further action necessary.

---

- Refer** The patient did not pass the screening.  
Review the test result (see below) and consider whether the test should be repeated.

---

- Noisy** Excessive noise was present during the test.  
Reduce noise and repeat the test.  
  
i Section 5.1.3 Noise Sources

---

- No Seal** A seal was not maintained throughout the test.  
Try another ear tip size and retest.

---

- FIT ERR** Probe placement in the ear canal is inadequate to produce target stimulus intensities.  
  
Try to achieve a better probe fit and retest.

**NOTE:** The results of the test are automatically saved in the memory as soon as the test is complete. The results are saved even if the device turns off or the batteries are temporarily depleted.

i Section 5.9.2 Saving Results




Figure 35

Reviewing Test Results

It is possible to review the test results in two ways:

- **Review on the screen:** Press **Review** to see the graph view of the test result (Figure 34 and e.g. Figure 35).
- **Review the printout:** see:

i Section 5.8.2 Interpreting Printed Results

## 5.8 Interpreting Test Results

### 5.8.1 Understanding the Test Result Screen

The ERO•SCAN shows the test results as a graph. The screen is being generated and shown during the test and can be reviewed after the test is complete (see Section 5.7.6).

The ERO•SCAN allows the user to select from two options for viewing the results.

- **SNR graph style:** shows the signal-to-noise ratio for each DP test frequency or TE test band.
- **Value graph style:** shows the absolute emission and noise levels for each DP test frequency or TE test band.

**NOTE:** The graph style can be changed in the **Settings**.



Section 5.10.2.8

Graph Style

### SNR Graph View




Figure 36 shows the SNR bar graph view. You see the signal-to-noise ratio (SNR) test results which are displayed as the emissions and noise floor are measured. Each column represents one DP test frequency or TE frequency band. The height of each column represents the SNR measured in dB.

Figure 36

If you have selected a protocol with a **Pass** criterion, you see a horizontal green line at the decibel level corresponding to the SNR required for a pass. Green vertical bars represent a **Pass** result, a yellow bar is a **Refer** result at the frequency band.

**NOTE:** Diagnostic protocols can display green or purple bars. Green is displayed when a **Pass** criterion is activated and set in the device.



Section 5.10.3.2

Advanced Options for DPOAE Testing –  
 Setting the PASS SNR Level  
 and  
 Setting the Number of Frequencies for PASS

## Value Graph View




Figure 37

Figure 37 shows the **Value graph** view for right ear. Red circle ● symbols represent the absolute emission levels at each DP test frequency or TE frequency band. For the left ear, dark blue X symbols represent the absolute emission levels at each DP test frequency or TE frequency band. White upside-down triangles (i.e. ▽) represent the noise floor at each DP test frequency or TE frequency band.

**Only for DPOAE protocols:** Additional display features include the grey shaded area representing the Boys Town Norms. The green line represents the minimum amplitude setting has been turned on. Both of these settings are optional to be displayed and are independent from one another. These settings are defaulted off.

### 5.8.2 Interpreting Printed Results

Results from the ERO•SCAN can be printed in multiple ways. It is up to the examiner to select the appropriate method for their practice needs. The printing options include:

- **Thermal Printout:** An optional wireless printer can be purchased for immediate printing to a thermal printer.
- **Software:** You can transfer the test results to your PC software and print from it.



Respective Software Manual

5.8.2.1 Understanding the Thermal Printout (DPOAE)

The following information is provided for each test (Figure 15):




Figure 38

1. Manufacturer Logo
2. The software version number (e.g.: V105.05)
3. The time and date of the test, based on the setting of the internal clock; if the clock is set correctly, this time and date will be correct
4. The test number (if operating in **Save 500** mode) (e.g.:001)
5. The protocol selected (e.g.: DP 2.0 - 5.0)
6. The averaging time used for this test (e.g.: 4 sec avg.)
7. Device/Probe serial number (SN)
8. The ear selected (Right or Left)
9. A PASS/REFER indication if there is a criterion set for the selected protocol
10. Graphic display of results
11. Printout notification when Minimum Amplitude is 'On' (e.g.: MIN\*)
12. The f2 frequency in kHz (e.g.: 2.0, 3.0, 4.0, 5.0)
13. SPL of presented tones (L1, L2)
14. The level of the emission in dB SPL (DP)
15. The noise floor in dB SPL (NF)
16. The signal-to-noise ratio (SNR) = DP – NF
17. A "P" indicates that the frequency passed based on the criterion settings within the selected protocol

5.8.2.2 Understanding the Thermal Printout (TEOAE)

The following information is provided for each test (Figure 39 – TEOAE printout with **Value Graph** view, Figure 40 – TEOAE Printout with **SNR Graph** view).



Section 5.10.2.8 Graph Style




Figure 39




Figure 40

1. Manufacturer Logo
2. The software version number (e.g.: V105.05)
3. The time and date of the test, based on the setting of the internal clock; if the clock is set correctly, this time and date will be correct
4. The test number (if operating in **Save 500** mode) (e.g.:001)
5. The protocol selected (e.g.: TE 64s)
6. The test time used to complete the test (e.g.: 4 sec avg.)
7. Device/Probe serial number (SN)
8. The ear selected (Right or Left)
9. A PASS/REFER indication if there is a criterion set for the selected protocol
10. Graphic display of results (SNR or Value graph)
11. Printout notification when Minimum Amplitude is 'on'
12. The frequency band center (F)
13. SPL peak equivalent of presented click (L)
14. The level of the emission in dB SPL (TE)
15. The noise floor in dB SPL (NF)
16. The signal-to-noise ratio (SNR) = TE – NF
17. A "P" indicates that the frequency passed based on the criterion settings within the selected protocol

## 5.9 Managing Test Results

### 5.9.1 General

Dependent on the configuration there are different possibilities to manage test results. It is possible to delete test results, print the session directly with the thermal printer or transfer the data to a PC for further processing.

Users have the option of printing to the thermal printer or transferring results to the PC.

### 5.9.2 Saving Results

The ERO•SCAN automatically saves the results of completed tests in the non-volatile memory (meaning tests are saved even if the battery is temporarily discharged). However, the ERO•SCAN is not intended for long-term storage of test results.

Users are strongly encouraged to print/transfer all test results at the completion of testing to avoid potential loss of data.

How the test results are saved depends on the save mode.



Section 5.10.2.5 Save Mode/Storing Test Results

Choose between:

**Save L/R mode:** only the most recent test results for the left and right ear are saved for printing and/or transfer to a PC.

**Save 500 mode:** The last 500 test results are saved for printing and/or transfer to a PC.

If patient data are transferred from a connected database, patients can be selected on the ERO•SCAN using the **<LEFT or RIGHT>** arrow keys. You may use "**Unnamed**" if the patient is not found in the device and use it for saving results and transfer to the database.

If no patients are transferred to the device the tests are automatically numbered from 1 to 500. It is important to keep record of the test numbers for each patient.

### 5.9.3 Deleting Test Results

The ERO•SCAN holds data in non-volatile memory. Deletion of test results depend on the saving mode.



Section 5.9.2 Saving Results

You can delete test data in the following ways:

### Automatic Deletion

**Save L/R Mode:** A single test for the **Left** ear and a single test for the **Right** ear are saved. Test data are deleted as soon as a new test for the left or right ear is started.

**Save 500 Mode:** Test data are deleted as soon as new Patient Names are uploaded from the PC Software to the ERO•SCAN (a warning is provided that data will be deleted).

**Printing:** As soon as test data are printed via the thermal printer or transferred to the PC software all tests are marked for deletion. Final deletion of results occurs with the start of a new test.

### Manual Deletion

To learn how to manually delete test results in the ERO•SCAN device see:



Section 5.10.2.3 Clearing Test Results

## 5.9.4 Transferring Test Results to a PC

It is possible to transfer data from the ERO•SCAN to the following PC software:

- MAICO Sessions Standalone
- MAICO Sessions with OtoAccess® Database
- MAICO Sessions with Noah Database
- MAICO Sessions with your existing Practice Management Software via BDT/GDT interface
- HearSIM™ Software with OtoAccess® Database (only results of the test protocols TE 32s, TE 64s, DP 4s and DP 2s)

### Connecting the ERO•SCAN to a PC

Connection to the PC software is achieved by using the provided Micro-USB to USB-A cable. If you use MAICO Sessions, wireless connection is also possible. This requires pairing with the PC.



Section 5.10.2.2 Wireless Device Pairing

Before transferring data to a PC make sure that you have installed the PC software properly according to the separately delivered operation manual.

Before establishing the PC connection via USB you will have to consider the recommendations given in Section 4.2.4 in case the ERO•SCAN is connected to a non-medical device.

Make sure to set the ERO•SCAN device to the Main Menu before transferring results to a connected software.

---

**NOTE:** The actual transfer process to the PC software depends on the used Software product.



Respective Software Manual

---

5.9.5 Printing Test Results

5.9.5.1 Printing to a Thermal Printer

**NOTE:** Make sure you have successfully paired your ERO•SCAN device with the printer before trying to start the print process.



Section 5.10.2.2 Wireless Device Pairing



Figure 41




Figure 42

To start the printing process, press:

- ▲ **UP** in the **Main Menu** to enter the **Connect to printer** screen (Figure 41).
- ◀ **CONNECT** ▶ to connect to the printer. Printing process starts as soon as the printer is found. The screen shows **PLEASE WAIT** (Figure 42).
- ▲ **CANCEL** to cancel the printing process if wanted.

**NOTE:** All printed test results are marked for deletion but will continue to be stored in memory until a new test is started at which time all tests in the memory will be erased. This allows the user to reprint the tests if printing is unsuccessful (for example, the paper runs out before printing is complete).

5.9.5.2 Printing After Data Transfer to PC

The actual printing process from the PC software depends on the used Software product. Printing is possible directly from the following PC Software products.

- MAICO Sessions
- OtoAccess® Database
- HearSIM™
- Your Practice Management Software



Respective Software Manual

## 5.10 Settings

### 5.10.1 Clock and Date Settings

**NOTE:** For changing the **Clock Mode** see:



Section 5.10.2.7 Clock Mode

You may want to change the date and/or time

- if you use your ERO•SCAN the first time.
- if you travel to another time zone.
- due to seasonal time change.
- the device's battery has completely discharged so that the device has automatically reset date and time.

**IMPORTANT:** The clock should be set prior to testing, as changing it after tests are saved will not change the date on the printout (i.e., whatever date was previously in memory will be the date on the printout).

To enter the clock menu proceed as follows:

▼ **CHANGE** to enter the **Protocol Selection Menu**.

▼ **SETUP** to enter the **Clock menu**.




Figure 43




Figure 44




Figure 45

The **Clock menu** shows the currently set date and time (Figure 43). Press:

▼ **CHANGE** to proceed to the next screen. (briefly)

**NOTE:** If you press the ▼ **CHANGE** too long, you will access the advanced settings.

◀ **LEFT** or **RIGHT** ▶ to set the **Month** (Figure 44), **Day** or **Hour**.

▼ **NEXT** to proceed to the next screen.

◀ **LEFT** or **RIGHT** ▶ or to set the **Minute** (Figure 45).

▼ **DONE** to save the new date and time settings and return to the **Main Menu**.

## 5.10.2 General Device Settings

### 5.10.2.1 General

The ERO•SCAN allows the user to change many of the device settings or functions. These settings include:

- Wireless Device Pairings
- Clearing Test Results
- Auto Shutdown Time
- Minimum Amplitude Value
- Save Mode
- Clock Mode
- Language
- Reset to Default Settings

#### Access the General Device Settings Menu

▼ **CHANGE** to access the **PROTOCOL** menu.

▼ **SETUP** to access the **CLOCK** menu.

▼ **CHANGE** to enter the **WIRELESS DEVICE** menu.

(hold for 3 s) Press and hold for 3 s until the green LED (READY) turns off.



Figure 20

**NOTE:** It is also possible to set advanced options for the DPOAE and TEOAE tests.



Section 5.10.3 Advanced Settings Menus (Diagnostic Version only)

## 5.10.2.2 Wireless Device Pairing

You can pair the ERO•SCAN

- with a thermal printer for printing the test results directly.
- with a PC for data transfer to the PC Software.

### Pairing with a Wireless Thermal Printer




Figure 47

The **WIRELESS DEVICE** menu allows the user to pair the ERO•SCAN with a wireless printer (Figure 47).

- <DISCOVER>** to search for wireless devices (approx. 15 s). Display shows “**Please wait**” and yellow LED (TEST) flashes.
- ▲CANCEL** to cancel discovery if wanted.



Figure 48

Discovered devices are displayed. A compatible thermal printer is displayed with “**PRT-XX-XX**” (e.g., PRT-e4-8c) (Figure 48).

- <CHANGE>** to select the printer.
- ▼PAIR** to pair with the printer.




Figure 49

If the display shows “**WIRELESS PAIRED**”, the pairing process was successful.

**NOTE:** If you cannot pair the device with the thermal printer or if any error messages are displayed, see


 Section 3.7 Troubleshooting

- ▼MainMenu** to exit the wireless pairing menu.

**Pairing with a PC**

The **WIRELESS DEVICE** menu allows the user to pair the ERO•SCAN to A PC.

Activate wireless connection on the PC (Figure 50).



**Figure 50**



**Figure 51**

Search for the PC (Figure 47):

**<DISCOVER>** to search for wireless devices (approx. 15 s). Display shows “**Please wait**” and yellow LED (TEST) flashes.

**▲CANCEL** to cancel discovery if wanted.




**Figure 52**

Discovered devices are displayed. Select the PC (Figure 48).


**<CHANGE>** to select the PC

**▼PAIR** to pair with the PC.



**Figure 53**

**On the PC:** Enter the PIN **1234** and press **Allow** (Figure 53).



**Figure 54**

If the connection has succeeded, press **Close** (Figure 54).




Figure 55

If the display shows “**WIRELESS PAIRED**”, the pairing process was successful.

**NOTE:** If you cannot pair the device with the PC or if any error messages are displayed, see



Section 3.7 Troubleshooting

▼ **MainMenu** to exit the wireless pairing menu.

### 5.10.2.3 Clearing Test Results




Figure 56

The **TEST RESULTS** menu allows the user to clear the test results stored in the device (Figure 56).

**NOTE:** After printing or transferring the test data to the PC software, all tests saved in memory will be permanently deleted as soon as a new test is started. It is not necessary to manually clear the results using this menu.

◀ **CLEAR** ▶ to clear test results.

◀ **YES or NO** ▶ to verify clearing or to cancel.

### 5.10.2.4 Auto Shutdown Time




Figure 57

The **POWER OFF** Menu allows you to set a time after which the device shall turn off automatically to prolong battery life (Figure 57).

**NOTE:** The device will immediately turn off after printing regardless of this setting.

◀ **CHANGE** ▶ to select a value between 0.5 and 4 min.

▼ **NEXT** to proceed to the next screen.

5.10.2.5 Save Mode/Storing Test Results




Figure 58

The ERO•SCAN can store either the most recent test results for the right and left ear (Figure 58) or the last 500 tests (Figure 59).

**<CHANGE** to select **SAVE L/R TESTS** or **SAVE 500 TESTS**.

**▼NEXT** to proceed to the next screen.



Figure 59

When using the 500 Tests mode with the ERO•SCAN, there are two ways of operating.

When a Software is used to transfer patient names to the device, 50 patients can be transferred to the device. The ERO•SCAN will show the names in an alphabetical order. You can cycle through the names using the left and right arrow. A patient named "**Unnamed**" is always included at the beginning of the ERO•SCAN list for instances when a patient is being tested but the patient name was not transferred to the ERO•SCAN.

When no patient names are transferred to the device, the ERO•SCAN will automatically number each test from 1 to 500.

**NOTE:** When 495 tests have been saved, the user will be warned that the memory is almost full. When the ERO•SCAN device reaches 500 saved tests, it will not allow any further testing. At this point either the results must be printed, transferred to the PC software, or they must be cleared from memory.

5.10.2.6 Minimum Value (MIN VALUE)




Figure 60

With the **MIN VALUE** setting (Figure 60) the user is allowed to include the minimum OAE response amplitude to the **PASS/REFER** criterion for a single frequency. When the **MIN VALUE** setting is set between -10 dB SPL and + 5 dB SPL the OAE response for the frequency under test needs to be at an equal or higher level than the **MIN VALUE** set to get a **PASS** result. If **MIN VALUE** is set to **OFF**, the **PASS** decision is based on the SNR alone for the current frequency measured. Default **MIN VALUE** setting is -10 dB SPL.

**<CHANGE>** to select a value between **-10 dB SPL** to **5 dB SPL** or to set it **OFF**.

**▼NEXT** to proceed to the next screen.

**5.10.2.7 Clock Mode**




The **CLOCK MODE** menu allows the user to change the clock from a 24-hour mode to a 12-hour mode (Figure 61).


- <CHANGE>** to select **12 Hour** or **24 Hour**.
- vNEXT** to proceed to the next screen.

Figure 61

**5.10.2.8 Graph Style**




The user can select between two options of displaying the result. (Figure 62). The SNR graph displays the Signal-to-noise ratio as a function of frequency. The value graph view shows absolute OAE levels and noise levels as a function of frequency.

 Section 5.7.1 Calibration and Test

- <CHANGE>** to select **GRAPH SNR** or **GRAPH VALUE**.
- vNEXT** to proceed to the next screen.

Figure 62

**5.10.2.9 Boys Town Norms**



The user can choose if the Boys Town<sup>1</sup> will be displayed as comparative normative data if the Graph view setting is set to Value Graph (Figure 63).

- <CHANGE>** to select **NORMS ON** or **NORMS OFF**.
- vNEXT** to proceed to the next screen.

Figure 63

**5.10.2.10 Reversed Frequency**



The reversed frequency setting enables the user to decide on the order of frequencies tested. ON will lead to measure in a descending order (high to low freq.) whereas OFF will lead to measuring in an ascending order (low to high freq.) This setting is only applicable for DPOAE testing (Figure 64).

- <CHANGE>** to select **REV FREQ ON** or **REV FREQ OFF**.
- vNEXT** to proceed to the next screen.

Figure 64

<sup>1</sup> Gorga, M.P., Neely, S.T., Ohlrich, B., Hoover, B., Redner, J. And Peters, J. (1997). "From laboratory to clinic: a large scale study of distortion product otoacoustic emissions in ears with normal hearing and ears with hearing loss." Ear & Hearing, 18, 440-455

**5.10.2.11 Auto Stop**



Figure 65

The activated **AUTO STOP** function stops the DPOAE test automatically as soon as the required number of passed frequencies is reached (for time saving reasons) (Figure 65).

**<CHANGE>** to select **AUTO STOP ON** or **AUTO STOP OFF**.

**▼NEXT** to proceed to the next screen.

**5.10.2.12 Language**



Figure 66

The **LANGUAGE** menu allows the user to select among several display languages (Figure 66).

**<CHANGE>** to select the display language.

**▼NEXT** to proceed to the next screen.

**5.10.2.13 Reset to Default**



Figure 67

In the **RESET TO DEFAULT** menu, you can reset the device settings to the factory settings (Figure 67).

**NOTE:** By resetting the device settings, you delete all existing test results and reset all system and protocol settings. This procedure also separates the ERO•SCAN from connected wireless devices.

**<RESET>** to reset the device settings to default.

**<YES or NO>** to confirm or stop the reset.

**▼NEXT** to proceed to the next screen.

**5.10.3 Advanced Settings Menus (Diagnostic Version only)**

**5.10.3.1 Customizing Advanced Settings**

The Advanced Options menu permits modification of the test parameters and pass criterion for the customizable protocols. Changes to the protocol should be made only by qualified personnel, usually the administrator. If you are not familiar with the use of these variables, do not attempt to change the protocols. Changes to any of these characteristics may yield test results that differ from those obtained in other test modes.

---

**IMPORTANT NOTE:** Any change of the parameters Level, SNR for PASS, Number of Frequencies for PASS and Minimum Value has an influence on the Sensitivity and Specificity of the PASS and REFER decision. If the criterion for PASS is set too loose,

---

there is a risk, that the test result is a PASS even if a hearing loss is present. If you choose other settings than the default values, you are responsible.




- Section 5.10.3      Advanced Settings Menus (Diagnostic Version only)
- Section 5.10.2.6      Minimum Value (MIN VALUE)

The ERO•SCAN comes with pre-programmed protocol settings. Test protocol changes are saved in the non-volatile memory so the settings will be retained even when the battery is discharged temporarily.



Section 6.6    Configurations and Test Protocols



When a change is made within the Advanced Options menu to a default protocol, an \* is placed in the protocol name (Figure 68).

Figure 68

To change a parameter of a DPOAE or TEOAE Test Protocol proceed as follows:

▼ <b>CHANGE</b>	to enter the <b>PROTOCOL</b> menu.
◀ <b>CHANGE</b> ▶	to select the DPOAE/TEOAE protocol (DP 4s and TE 64s protocols are not customizable).
▼ <b>SETUP</b>	to enter the <b>Clock menu</b> .
▼ <b>CHANGE</b> (press for 3 s)	to enter the <b>Setup menu</b> . Press and hold for 3 s until the green LED (READY) turns off.
▼ <b>NEXT</b> (press for 3 s)	to enter the DPOAE/TEOAE menu. Press and hold for 3 s until the green LED (READY) turns off.
▼ <b>NEXT</b>	to select a protocol parameter to change.
◀ <b>CHANGE</b> ▶	to change a parameter.
▼ <b>DONE</b>	to save the parameters and exit the TEOAE menu.

**NOTE:** If you push the ▼ **DOWN** arrow key without holding it for 3 seconds, you will scroll through date and time, etc., rather than accessing the custom menus.

5.10.3.2 Advanced Options for DPOAE Testing

Selecting the Level of Primary Tones




Figure 69

In the **LEVEL** menus (**LEVEL L1**, **LEVEL L2**) you can set the level for the primary tones (L1, L2) (Figure 69).

**<CHANGE>** to set the level L1 to a value between **40 dB SPL** and **70 dB SPL** (1 dB steps).

**▼NEXT** to proceed to the **LEVEL L2** menu.

**<CHANGE>** to set the level L2 to a value between **40 dB SPL** and **70 dB SPL** (1 dB steps).

**▼NEXT** to proceed to the next screen.

Setting the Averaging Time




Figure 70

A longer averaging time increase measurement duration and result quality. A shorter averaging time decreases measurement duration and result quality (Figure 70).

**<CHANGE>** to set the averaging time to **0.5 s**, **1.0 s**, **2.0 s** or **4.0 s**.

**▼NEXT** to proceed to the next screen.

Setting the PASS SNR Level




Figure 71

To provide a PASS/REFER determination for each test, the PASS SNR must be set. The PASS SNR defines the ratio of how much higher the level of the OAE response has to be compared to the noise. This requirement is used in combination with the number of frequencies to determine an overall PASS/REFER for each test (Figure 71).

**<CHANGE>** to select a value between **3 dB** and **10 dB**.

**▼NEXT** to proceed to the next screen.

### Setting the Number of Frequencies for PASS




Figure 72

The user can adjust the number of frequencies that have to fulfill the **SNR PASS** criterion (and optionally the **MIN VALUE** criterion) to generate an overall PASS result on the screening measurement. If the defined number of FREQS FOR PASS is not reached, the measurement will result in a REFER.

Setting the **FREQS FOR PASS** to **0** disables the functionality and no PASS/REFER result will be displayed after measurement (Figure 72).

**NOTE:** If this function is disabled the bars of the SNR graph are displayed in purple, if activated in green (PASS) or in yellow (REFER).

**<CHANGE>** to select a number between **0** and the maximum number of frequencies of the selected protocol.

**vNEXT** to proceed to the next screen.

### Reset Protocol



Figure 73

This **RESET TO DEFAULT** menu allows you to reset the settings of the currently selected protocol to the factory settings (Figure 73).

**NOTE:** If you use this function, you only reset the currently selected protocol. The device settings and settings of other protocols remain unaffected.

**<RESET>** to reset the settings of the currently selected protocol to default.

**vNEXT** to proceed to the next screen.

### Save Protocol




Figure 74

The **SAVE PROTOCOL** menu allows you to save the protocol settings made or leave the menu without saving (Figure 74).

**<SAVE>** to save the new protocols settings.

**vDONE** to exit the menu.

**NOTE:** If you press **vDONE** before saving you lose the settings made.

5.10.3.3 Advanced Options for TEOAE Testing

Selecting the Averaging Time




Figure 75

A longer averaging time increase measurement duration and result quality. A shorter averaging time decreases measurement duration and result quality (Figure 75).

**<CHANGE>** to set the averaging time to **4 s, 8 s, 16 s, 32 s or 64 s**.

**▼NEXT** to proceed to next screen.

Setting the PASS SNR Level




Figure 76

To provide a **PASS/REFER** determination for each test, the **PASS SNR** must be set. The PASS SNR defines the ratio of how much higher the level of the OAE response has to be compared to the noise. This requirement is used in combination with the number of frequencies to determine an overall **PASS/REFER** for each test (Figure 76).

**<CHANGE>** to select a value between **3** and **10** (dB).

**▼NEXT** to proceed to next screen.

Setting the Number of Frequencies for PASS




Figure 77

The user can adjust the number of frequencies that have to fulfill the SNR PASS criterion (and optionally the MIN VALUE criterion) to generate an overall PASS result on the screening measurement. If the defined number of **FREQS FOR PASS** is not reached, the measurement will result in a REFER.

Setting the FREQS FOR PASS to 0 will disable the functionality and no PASS/REFER result will be displayed after measurement (Figure 77).

**<CHANGE>** to select a number between **0** and **6**.

**▼NEXT** to proceed to the next screen.


## Reset Protocol



This **RESET TO DEFAULT** menu allows you to reset the settings of the currently selected protocol to the factory settings (Figure 78).

**NOTE:** Using this function, you only reset the currently selected protocol. The device settings and settings of other protocols remain unaffected.

Figure 78




**◀RESET▶** to proceed to the **RESET PROTOCOL?** screen (Figure 79).

**▼NEXT** to confirm reset of the protocol and proceed to the next screen.

Figure 79

## Save Protocol



The **SAVE PROTOCOL** menu allows you to save the protocol settings made or leave the menu without saving (Figure 80).

**◀SAVE▶** to save the new protocols settings.

**▼DONE** to exit the menu.

**NOTE:** If you press **▼DONE** before saving you will lose the settings made.

Figure 80

## 6 Technical Data

This section offers you important information about

- the ERO•SCAN hardware specifications
- the pin assignment
- calibration values
- electromagnetic compatibility (EMC)
- electrical safety, EMC and associated Standards
- Configurations and Test Protocols

### 6.1 ERO•SCAN Hardware



The ERO•SCAN is an active, diagnostic medical product according to the class IIa of the Medical Device Regulation (EU) 2017/745.

#### General Information About Specifications


The performance and specifications of the device can only be guaranteed if it is subject to technical maintenance at least every 12 months.

MAICO Diagnostics puts diagrams and service manuals at the disposal of authorized service companies.

#### STANDARDS

<b>Device Safety</b>	IEC 60601-1:2005 + Cor. :2006 + Cor. :2007 + A1:2012/ ANSI/AAMI ES60601-1: 2005 / A2:2010/ CAN/CSA-C22.2 No. 60601-1:14 Internally powered, Type B Applied Parts
<b>EMC</b>	EN 60601-1-2:2014
<b>Calibration</b>	ISO 389-2 / ISO 389-6
<b>Test Signal</b>	IEC 60645-3:2007
<b>OAE</b>	IEC 60645-6:2009, Type 2

## DEVICE SPECIFICATIONS

<b>Operation environment</b> 	Temperature	+15 °C to +35 °C / + 59 °F to +95 °F
	Relative Humidity	30 % to 90 % (non-condensating)
	Ambient Pressure	30 % to 80 % recommended 98 kPa to 104 kPa
<b>Transport &amp; Storage environment</b>	Storage Temperature	0 °C to 50 °C, 32 °F to 122°F
	Transport Temperature	-20 °C to + 50 °C / -4 °F to +122 °F
	Storage and transport rel. humidity	10 % to 95 % (non-condensating)
<b>Altitude rating</b>	Max. operating altitude	2000 m / 6561 ft. above sea level
<b>Warm-up Time</b>		< 5 s
<b>Boot-up time</b>		< 1 min
<b>Dimensions</b>		65.5 mm x 31.2 mm x 146.0 mm 2.58 in x 1.23 in x 5.78 in
<b>Weight</b>		176 g (6.2 oz.) 204 g (7.2 oz.) (with probe)
<b>Display</b>	Display Size	42.7 mm x 33.4 mm / 1.7 in x 1.3 in
	Resolution	160 x 128
<b>Mode of operation</b>		Continuous
<b>Connectors</b>	OAE	HDMI Connector for connection to the Micro-Probe
	USB	Micro USB
<b>Data Interfaces</b>	PC connection	Micro USB, wireless data transfer to MAICO Sessions
	Printing	Wireless connection
	Frequency Range	2402 MHz to 2480 MHz
	Effective radiated power	Up to +12 dBm
	Modulation	GFSK, $\pi/4$ -DQPSK or 8-DPSK
<b>User Feedback</b>	Acoustical	Integrated speaker
	Visual	Color display and LED
<b>User Interface</b>		OLED display to provide user information and progress of measurement Control Panel with Membrane-Type Push Buttons
<b>Language Settings</b>		English, Japanese, German, French, Spanish, Russian, Polish, Turkish, Portuguese, Italian, Chinese, Korean, Arabic

## DEVICE SPECIFICATIONS

<b>Battery</b>	Type	BAKTHNP120, Lithium-Ion battery rechargeable
	Capacity	3.7 V 1800 mAh
	Expected Life Time	Depending on use – typically > 3 years
	Battery Life	500 tests per charge, minimum 15 hours on-time
<b>Memory</b>		500 tests can be stored
<b>Connectors / Communications</b>		Integrated USB communication capability for battery charging and communication with PC-based database programs HDMI Connector for connection to the Micro-Probe Integrated wireless Class 2 + EDR with SPP Protocol for communication with optional printer

## POWER SUPPLY UNIT

Model No.	UES12LCP-050160SPA
Input	100 to 240 V AC, 50/60 Hz, 0.5 A
Output	5.0V DC, 1.6A MAX
Safety	IEC 60601-1, Class II

## DPOAE

<b>Stimulus</b>	Frequency range	1500 Hz to 12 kHz
	Frequency accuracy	< 1 %
	Default frequencies	See Section 6.6 Configurations and Test Protocols
	Nominal frequency	F2
	F2/F1 Ratio	1.2
	Level range	40 dB SPL to 70 dB SPL
	Level accuracy	± 1.5 dB
	Default level (L1/L2)	65 dB SPL / 55 dB SPL with in-ear calibration
	Transducer	Micro-Probe
<b>Recording</b>	Maximum test time	Depends on protocol and device settings
	Artifact rejection level	55 dB SPL
	Probe fit check	Low frequency level loss
	Residual noise	RMS measurement in frequency domain
	Display	SNR bars or line diagram with OAE and noise level per frequency
<b>Pass criteria</b>	SNR criteria	6 dB
	#Freq for pass	3

## TEOAE

<b>Stimulus</b>	Stimulus type	Non-Linear click (according to IEC 60645-3)
	Default level	83 dB peSPL (peak-to-peak calibrated), auto in-ear calibration
	Level tolerance	± 3 dB
	Click rate	Approx. 61 /s
	Transducer	Micro-Probe
<b>Analysis bands</b>	Frequency range	700 Hz to 4000 Hz
	Default center frequencies	1.5, 2, 2.5, 3, 3.5, 4
<b>Recording</b>	Maximum test time	64 s, depends on averaging time of the selected protocol
	Maximum noise level	55 dB SPL
	Averaging method	Bayesian weighted averaging
	Display	SNR view / value graph view
<b>Pass criteria</b>	SNR criteria	Depends on protocol
	Response criteria	Depends on protocol
	#Freq bands	Depends on protocol


## TRANSDUCER


<b>Micro-Probe</b>	Microphone System	-20 dB SPL at 2 kHz (1 Hz bandwidth)
	Noise	-13 dB SPL at 1 kHz (1 Hz bandwidth)
	Supported tests	DPOAE, TEOAE
	Cable length	110 cm / 43 in
	Connector	HDMI
	Probe tip	Replaceable
	Weight (incl. cables)	28 g / 1.00 oz

**PRINTER**

<b>Thermal printer</b>	Type	HM-E200
	Connection	Wireless
	Battery	7.4 V rechargeable Li-polymer battery, 1300 mAh
	Weight	234 g / 8.3 oz
	Paper	Paper Rolls
	Paper size	Thermal paper 57.5 mm ± 0.5 mm (width)
	Printing time	<5 seconds per test result
<b>Power supply unit</b>	Type	UES12LCP-050160SPA
	Input	100 to 240 V AC, 50/60 Hz, 0.5 A
	Output	5.0V DC, 1.6A MAX
	Safety	IEC 60601-1, Class II

## 6.2 Connections and Pin Assignment

ERO•SCAN CONNECTOR	
MICRO USB B (IN)	
	1. +5 VDC
	2. Data -
	3. Data +
	4. ID
	5. Ground

PROBE CONNECTOR			
TYPE A RECEPTACLE HDMI (FEMALE)			
			
Pin 1	Rcvr +	Pin 11	Unused
Pin 2	Rcvr Shield	Pin 12	Unused
Pin 3	Rcvr -	Pin 13	Unused
Pin 4	Reserved	Pin 14	Reserved
Pin 5	Shield	Pin 15	Comm Power
Pin 6	Reserved	Pin 16	Comm Data
Pin 7	Mic Power +	Pin 17	Ground
Pin 8	Mic Shield	Pin 18	+3.3V
Pin 9	Mic Out	Pin 19	Ground
Pin 10	Mic Power -	-	-

### 6.3 Calibration Values

Output Frequency (Hz)	Minimum Frequency (Hz)	Maximum Frequency (Hz)	Minimum Magnitude (dB SPL)	Maximum Magnitude (dB SPL)
732.4	727	737	83	93
1037.6	1033	1043	85	95
1464.8	1460	1470	88	98
2075.2	2070	2080	92	102
2929.7	2925	2935	92	102
4150.4	4145	4155	85	95
5859.4	5855	5865	76	86

Frequency [Hz]	IEC 60711, RA-0045
732.4	88.0
1037.6	90.0
1464.8	93.5
2075.2	97.8
2929.7	97.8
4150.4	90.6
5859.4	81.9

### 6.4 Electromagnetic Compatibility (EMC)

This device is suitable in hospital environments except for near active HF surgical equipment and RF shielded rooms of systems for magnetic resonance imaging, where the intensity of electromagnetic disturbance is high.

ESSENTIAL PERFORMANCE for this device is defined by the manufacturer as:

- This device does not have an ESSENTIAL PERFORMANCE.
- Absence or loss of ESSENTIAL PERFORMANCE cannot lead to any unacceptable immediate risk. Final diagnosis shall always be based on clinical knowledge.

Use of this device adjacent to or stacked with other equipment should be avoided because it could result in improper operation. If such use is necessary, this device and the other equipment should be observed to verify that they are operating normally.

Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the ERO•SCAN, including cables specified by the manufacturer. Otherwise, degradation of the performance of this equipment could result in improper operation.

This device is in compliance with IEC 60601-1-2:2014+AMD1:2020, emission class B group 1

*NOTICE:* There are no deviations from the collateral standard and allowances uses.

*NOTICE:* All necessary instruction for maintaining compliance with regard to EMC can be found in the general maintenance section in this instruction. No further steps required.

## Electromagnetic Compatibility (EMC)

Portable and mobile RF communications equipment can affect the **ERO•SCAN**. Install and operate the **ERO•SCAN** according to the EMC information presented in this chapter.

The **ERO•SCAN** has been tested for EMC emissions and immunity as a standalone **ERO•SCAN**. Do not use the **ERO•SCAN** adjacent to or stacked with other electronic equipment. If adjacent or stacked use is necessary, the user should verify normal operation in the configuration.


The use of accessories, transducers and cables other than those specified, with the exception of servicing parts sold by MAICO as replacement parts for internal components, may result in increased EMISSIONS or decreased IMMUNITY of the device.

Anyone connecting additional equipment is responsible for making sure the system complies with the IEC 60601-1-2 standard.

Guidance and manufacturer's declaration - electromagnetic emissions		
The <b>ERO•SCAN</b> is intended for use in the electromagnetic environment specified below. The customer or the user of the <b>ERO•SCAN</b> should assure that it is used in such an environment.		
Emissions Test	Compliance	Electromagnetic environment - guidance
RF emissions CISPR 11	Group 1	The <b>ERO•SCAN</b> uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF emissions CISPR 11	Class B	The <b>ERO•SCAN</b> is suitable for use in all commercial, industrial, business, and residential environments.
Harmonic emissions IEC 61000-3-2	Not Applicable	
Voltage fluctuations / flicker emissions IEC 61000-3-3	Not applicable	

Recommended separation distances between portable and mobile RF communications equipment and the <b>ERO•SCAN</b> .			
The <b>ERO•SCAN</b> is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the <b>ERO•SCAN</b> can help prevent electromagnetic interferences by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the <b>ERO•SCAN</b> as recommended below, according to the maximum output power of the communications equipment.			
Rated Maximum output power of transmitter [W]	Separation distance according to frequency of transmitter [m]		
	150 kHz to 80 MHz $d = 1.17\sqrt{P}$	80 MHz to 800 MHz $d = 1.17\sqrt{P}$	800 MHz to 2.7 GHz $d = 2.23\sqrt{P}$
0.01	0.12	0.12	0.23
0.1	0.37	0.37	0.74
1	1.17	1.17	2.33
10	3.70	3.70	7.37
100	11.70	11.70	23.30
For transmitters rated at a maximum output power not listed above, the recommended separation distance $d$ in meters (m) can be estimated using the equation applicable to the frequency of the transmitter, where $P$ is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.			
<b>Note 1</b> At 80 MHz and 800 MHz, the higher frequency range applies.			
<b>Note 2</b> These guidelines may not apply to all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.			

<b>Guidance and Manufacturer's Declaration - Electromagnetic Immunity</b>			
The <b>ERO•SCAN</b> is intended for use in the electromagnetic environment specified below. The customer or the user of the <b>ERO•SCAN</b> should assure that it is used in such an environment.			
<b>Immunity Test</b>	<b>IEC 60601 Test level</b>	<b>Compliance</b>	<b>Electromagnetic environment - guidance</b>
Electrostatic Discharge (ESD) IEC 61000-4-2	± 8 kV contact ± 15 kV air	± 8 kV contact ± 15 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be greater than 30%.
Electrical fast transient/burst IEC61000-4-4	± 2 kV for power supply lines 100kHz repetition frequency ± 1 kV Line-to-line 100kHz repetition frequency	Not applicable ± 1 kV Line-to-line	Mains power quality should be that of a typical commercial or residential environment.
Surge IEC 61000-4-5	± 1 kV Line-to-line ± 2 kV Line-to-ground	Not applicable	Mains power quality should be that of a typical commercial or residential environment.
Voltage dips, short interruptions and voltage variations on power supply lines IEC 61000-4-11	0% <i>UT</i> for 0.5 cycle 0 % <i>UT</i> for 1 cycle and 70% <i>UT</i> for 25/30 cycles Single phase: at 0°	Not applicable	Mains power quality should be that of a typical commercial or residential environment. If the user of the <b>ERO•SCAN</b> requires continued operation during power mains interruptions, it is recommended that the <b>ERO•SCAN</b> be powered from an uninterruptable power supply or its battery.
Power frequency (50/60 Hz) IEC 61000-4-8	30 A/m	30 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or residential environment.
<b>Note:</b> <i>UT</i> is the A.C. mains voltage prior to application of the test level.			

Guidance and manufacturer's declaration — electromagnetic immunity			
The <b>ERO•SCAN</b> is intended for use in the electromagnetic environment specified below. The customer or the user of the <b>ERO•SCAN</b> should assure that it is used in such an environment.			
Immunity test	IEC / EN 60601 test level	Compliance level	Electromagnetic environment – guidance
Conducted RF IEC / EN 61000-4-6	3 Vrms 150kHz to 80 MHz	3 Vrms	Portable and mobile RF communications equipment should be used no closer to any parts of the <b>ERO•SCAN</b> , including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.  <b>Recommended separation distance:</b> $d = 1,2\sqrt{P}$
	6 Vrms in ISM bands 150kHz to 80 MHz 80 % AM at 1 kHz	6 Vrms	
Radiated RF IEC / EN 61000-4-3	3 V/m 80 MHz to 2,7 GHz 80 % AM at 1 kHz	3 V/m	$d = 1,2\sqrt{P}$ 80 MHz to 800 MHz $d = 2,3\sqrt{P}$ 800 MHz to 2,7 GHz  Where $P$ is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and $d$ is the recommended separation distance in meters (m).  Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey, <sup>a</sup> should be less than the compliance level in each frequency range. <sup>b</sup>  Interference may occur in the vicinity of equipment marked with the following symbol:  
NOTE1 At 80 MHz and 800 MHz, the higher frequency range applies NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.			
<sup>a)</sup> Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the <b>ERO•SCAN</b> is used exceeds the applicable RF compliance level above, the <b>ERO•SCAN</b> should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the <b>ERO•SCAN</b> . <sup>b)</sup> Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3 V/m.			

This device is in compliance with IEC 60601-1-2:2014, emission class B group 1.

**NOTE:** There are no deviations from the collateral standard and allowances uses

**NOTE:** All necessary instruction for maintaining compliance with regard to EMC can be found in the general maintenance section in this instruction. No further steps required.

To ensure compliance with the EMC requirements as specified in IEC 60601-1-2, it is essential to use only the following accessories (see Table 9).

**Table 9 EMC Requirements – Accessories**

ITEM	MANUFACTURER	MODEL
Micro-Probe	MAICO	-

## 6.5 Electrical Safety, EMC and Associated Standards

- IEC 60601-1:2005 + Cor. :2006 + Cor. :2007 + A1:2012: Medical Electrical Equipment, Part 1 General Requirements for basic safety and essential performance
- ANSI/AAMI ES60601-1: 2005 / A2:2010: Medical Electrical Equipment, Part 1 General Requirements
- CAN/CSA-C22.2 No. 60601-1:14 Medical Electrical Equipment, Part 1 General requirements for basic safety and essential performance
- IEC/EN 60601-1-2:2014 Medical Electrical Equipment, Part 1 - Electromagnetic Compatibility - Requirements and Tests
- General Safety and Performance Requirements of the current REGULATION (EU) 2017/745
- DIRECTIVE 2011/65/EU of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment (RoHS)
- DIRECTIVE 2012/19/EU of the European Parliament and of the Council of 4 July 2012 on waste electrical and electronic equipment (WEEE)

## 6.6 Configurations and Test Protocols

### DPOAE Protocols

DEVICE VERSION	NAME	NUMBER OF FREQ.	F2 FREQ. [KHZ]	L1/L2 [DB SPL]	AVERAGING TIME PER FREQ	PASS SNR	NUMBER OF FREQ. FOR PASS
Screening	DP 4s	4	2, 3, 4, 5	65/55	4 s	6 dB	3
	DP 2s	4	2, 3, 4, 5	65/55	2 s	6 dB	3
Diagnostic	DP 4s	4	2, 3, 4, 5	65/55	4 s	6 dB	3
	DP 2.0-5.0	4	2, 3, 4, 5	65/55**	4 s**	6 dB**	3**
	DP 1.5-6.0	6	1.5, 2, 3, 4, 5, 6	65/55**	4 s**	6 dB**	0**
	DP 1.6-8.0	12	1.6, 2, 2.5, 3.2, 3.6, 4, 4.5, 5, 5.6, 6.3, 7.1, 8	65/55**	4 s**	6 dB**	0
	DP 1.5-12	12	1.5, 2, 3, 4, 5, 6, 7, 8, 9, 10, 11, 12	65/55**	4 s**	6 dB**	0

### TEOAE Protocols

DEVICE VERSION	NAME	NUMBER OF FREQ.	F2 FREQ. [KHZ]	AVERAGING TIME PER FREQ	PASS SNR	NUMBER OF FREQ. FOR PASS
Screening	TE 64s	6	1.5, 2, 2.5, 3, 3.5, 4	64 s	4 dB	3
	TE 32s	6	1.5, 2, 2.5, 3, 3.5, 4	32 s	4 dB	3
Diagnostic	TE 64s	6	1.5, 2, 2.5, 3, 3.5, 4	64 s	4 dB	3
	TE1.5-4.0	6	1.5, 2, 2.5, 3, 3.5, 4	64 s**	4 dB**	3**
	TE0.7-4.0	6	0.7, 1, 1.4, 2, 2.8, 4	64 s**	4 dB**	0**

\*\* Customizable parameters:

L1/L2:

DP: 40 to 70 dB SPL

Average time:

DP: 0.5, 1.0, 2.0 or 4.0 s

TE:

4, 8, 16, 32 or 64 s

Pass SNR:


DP and TE: 3 dB to 10 dB

Number of Frequencies for PASS:

DP and TE: 0 (no PASS/REFER indication) to Number of Frequencies of the corresponding protocol


## 6.7 Flowcharts

### 6.7.1 Test Operation Flowchart



6.7.2 Setup Menu Flowchart

**NOTE:** DPOAE/ TEOAE Menu are only accessible in the ERO•SCAN Diagnostic version.



Specifications are subject to change without notice



MAICO Diagnostics GmbH

Sickingenstr. 70-71

10553 Berlin

Germany

Tel.: + 49 30 / 70 71 46-50

Fax: + 49 30 / 70 71 46-99

E-mail: [sales@maico.biz](mailto:sales@maico.biz)

Internet: [www.maico.biz](http://www.maico.biz)