

CERTIFICATE of Conformity



according to Directive 2006/42/EC
on safety of machinery

Certificate No.: M2 69250753 0001

Test Report No.: 28228762 002

Certificate Holder: Celitron Medical Technologies Kft.
Szent László út 36.
2600 Vác
Hungary

Manufacturer: Celitron Medical Technologies Kft.
Szent László út 36.
2600 Vác
Hungary

Product: Integrated Sterilizer & Shredder

Identification: Type designation: ISS AC-575 SD
Serial number: 43150033

Sterilization temperature: 134°C-137°C
Technical datas: 3AC 400V; 50Hz; 3×63A; 10kA; 27kW; 150L

The Sterilizer is a Stream Sterilizer with an integrated shredder for the treatment of medical waste in hospitals and clinics.

This Certificate does not attest conformity of a hazardous waste treatment technology.

Tested according to: 2006/42/EC - Annex I
EN 60204-1:2006+A1

Date of Validity: 2025.09.16.

This certificate refers to the above mentioned product. This is to certify that the test sample is in conformity with the requirements of Annex I of the Directive. The Certificate Holder is entitled to use this certificate in connection with the EC Declaration of Conformity in accordance with Annex II of the Directive. This certificate does not imply assessment of the series-production of the product and does not permit the use of a TÜV Rheinland mark of conformity. The certificate is valid up to the date of validity provided that the test specifications and construction of the product remain unchanged during the period of validity.

Date of Issue:

Budapest, 2020.09.17.



TÜV Rheinland InterCert Kft. – Product Certification Body — H-1132 Budapest, Váci út 48/A-B — www.tuv.hu
The CE marking may be used if all relevant and effective EC directives are complied with.