

Quality Assurance Certificate

Product: RIDASCREEN® Clostridium difficile Toxin A/B

REF	Art. No.:	C0801
LOT	Lot:	12045E
	Expiry:	2016-05

Control		OD _{450 nm}	OD _{450/620 nm}	QC specification	
				OD _{450 nm}	OD _{450/620 nm}
Control	+	2,574	2,492	> 0.8	> 0.8
Control	-	0,038	0,001	< 0.200	< 0.160

Content	LOT	
Control	+	13514
Control	-	14105
Conjugate 1		14025
Conjugate 2		11035
Plate		15394
Stop		12404
Diluent		14145
Substrate		15354
Wash		11394

R-Biopharm AG, Darmstadt, Germany, certifies that this batch has been approved by the Quality Assurance Department and conforms to specifications.

R-Biopharm AG
Quality Assurance Department

sign. N. Stork-Heininger
Quality Control Manager

Date: 2015-04-08

Remark: This document is created electronically and therefore valid without a signature.

Quality
Management
System certified
ISO 9001
ISO 13485
DQS, Germany

R-Biopharm AG
Postal address: An der neuen Bergstraße 17
64297 Darmstadt, Germany
Corporate seat: Pfungstadt
Phone: +49 (0) 61 51 - 81 02-0
Fax: +49 (0) 61 51 - 81 02-40
www.r-biopharm.com

**Chairman of supervisory
board:** Dietrich Mollat
Board of directors:
Dr. Ralf M. Dreher (CEO),
Dr. Carsten Bruns,
Jochen Hirsch,
Dr. Peter Schubert

Reg.No.:
Amtsgericht Darmstadt,
HRB 8321
Tax No.:
07242 / 02 888
VAT Reg. No.:
DE 111 657 409

