

# Certificate of Conformity

<b>VWR Catalogue Number</b>	612-3744
<b>Description</b>	Volumetric flasks, tall form, class A, with 100 mL Capacity, 14/23 N/S, 0.100 ± mL Tolerance
<b>Lot Number</b>	01.20
<b>Date of Manufacture (yyyy-mm-dd)</b>	2020-01-01
<b>Country of Origin</b>	Manufactured in India
<b>Date of Issue (yyyy-mm-dd)</b>	2022-07-19

## Quality System Compliance

Products are manufactured under the **ISO 9001:2015** standard and calibrated in **ISO 17025:2017** certified Laboratory. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

## QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

## Product Specifications

The product is in conformance with the requirements of the following standards and regulations:

Technical standards	EN ISO 1042
Calibration procedure	ISO 4787
<b>Material</b>	Borosilicate 3.3 Glass DIN ISO 3585:1998 TYPE-1, CLASS A, and meets ASTM E-438, TYPE-1, CLASS-A Standard.
<b>Sterilization</b>	Hot air sterilization, in the oven up to 140 °C Steam sterilization, in an autoclave 121 °C/ 20 min 134 °C/ 10 min.
ATP Assay	Not applicable.
DNase & RNase Free	Not applicable.

<b>BSE/TSE</b>	No use of any raw material produced from or substances derived from animal origin. The manufacturing process of the product does not use any ingredient of animal origin and no material derived from or exposed to animals affected by or under quarantine for transmitting Animal Bovine Spongiform Encephalopathy/Spongiform Encephalopathy (BSE)/(TSE).
<b>Cytotoxicity</b>	Not applicable.
<b>Non-Pyrogenic Statement</b>	Not applicable.
<b>Latex Statement</b>	Not applicable.
<b>BPA Statement</b>	Not applicable.
<b>DEHP Statement</b>	Not applicable.
<b>RoHS</b>	Not applicable.
<b>REACH Statement</b>	Not applicable.
<b>Storage Conditions</b>	Stored at room temperature in normal warehouse condition.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.  
This document has been produced electronically and is valid without a signature