



Product Service

Attestation of Conformity

No. N8A 083766 0007 Rev. 00

Holder of Certificate: Jinan Biobase Biotech Co., Ltd.
North Side of Jiawang Road
Mingshui Economic Development District
250200 Zhangqiu, Jinan City
PEOPLE'S REPUBLIC OF CHINA

Product: Laboratory Equipment
(Microbiological Safety Cabinet)

Brand: BIOBASE

Model(s): BSC-3FA2, BSC-4FA2

Parameters:

Rated Voltage:	230V~
Rated Frequency:	50Hz
Rated Power:	1050W for BSC-3FA2, 1150W for BSC-4FA2
Protection Class:	Class I
Class of MSC:	II

Tested according to: EN 61010-1:2010/A1:2019
EN 12469:2000

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2014/35/EU relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. For details see: www.tuvsud.com/ps-cert

Test report no.: 64111200354101

Date, 2020-08-12


(Bin Li)

Page 1 of 1

After preparation of the necessary technical documentation as well as the EU declaration of conformity the required CE marking can be affixed on the product. The declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany

TUV®