

### **GENERAL FEATURES OF ECO PHARMACY REFRIGERATORS**

- · Anti corrosive powder painted steel outer body
- · High density polyurethane for insulation
- · Double glazed glass aluminum door with magnetic gasket
- $\cdot$  Temperature measurement accuracy with the precision of 0,1°C
- · Lockable doors
- · Maximum air circulation with special shelf design
- · Position-adjustable shelves
- Defrosting feature
- · High temperature, low temperature, and audible alarm system for door open (prob) faults
- Extra external probe input (optional)
- Easy to move by swivel castors with breaks

#### WARNING

Our cabinets are designed to operate within the temperature ranges of 20-25°C and relative humidity ranges of 40-55 rh, which are ideal hospital ambient conditions. For different ambient conditions, we have optional solutions.



## ECO250ASI



## Manufactured following the DIN58345 standard.

		2000 mm
Exterior Dimensions WxDxH (mm)	590x620x1540	
Interior Dimensions WxDxH (mm)	490x465x1091	
Packed Dimensions WxDxH (mm)	614x660x1664. 5	
Net Volume (L)	225	
Gross Volume (L)	240	1500 mm
Net Weight (kg)	81	1300 11111
Gross Weight (kg)	87	
Insulation Thickness (mm)	50	
Operating Temperature	0 / +15 °C	
The Set Temperature	4 °C	
Voltage	220 / 240V	1000 mm
Frequency	50 Hz	
Power Consumption	1.8 kWh	
Ampere (A)	1.2	
Refrigerant Fluid	R600a	
Outer Body Color	RAL9003	500 mm
Shelves	3 pcs	
Container Loading 20"	27	
Container Loading 40"	57	
High Container Loading 40"	57	
Truck Loading 13.6 Meters	80	
		0 mm



### **OPTIONS**

- Extra external probe input
- Outer body color
- Side sticker with logo dressing
- Block door

### **Satisfying Standards**

As with other Coolermed refrigerators and freezers, the EC02500 cabinet provides optimal conditions to protect valuable, sensitive products such as medicines and vaccines with its defrosting feature,

adjustable shelf system, and special shelf design ensuring maximum air circulation.





#### EU DECLARATION OF CONFORMITY



Manufacturer ÖZTAŞ GLOBAL SOĞUTMA TEKNOLOJİLERİ VE METAL SANAYİ ve TİCARET

ANONIM

Address ŞİRKETİ

ORGANIZE SAN. BÖLGESİ 4. KISIM, AHMET NAZİF ZORLU BULV. NO:32 MANİSA

Product FREEZER

Tuna MC60ASI - M170ASI- M260ASI - M390ASI - M400ASI -M660ASI - M900ASI- ECO380ASI CLASS - ECO670ASI CLASS-

M1600ASI

The product described above is in conformity with the following EU Directives

•2014/35/EU REGULATION ON THE ELECTRICAL EQUIPMENT DESIGNED FOR USE WITHIN CERTAIN VOLTAGE LIMITS

•2014/35/EU

ELECTROMAGNETIC COMPATIBILTY REGULATIONS.

•2006/42/EU

MACHINERY DIRECTIVE

•93/42/EEC •2011/65/EU MEDICAL DEVICE DIRECTIVE ROHS DIRECTIVE

The products described above, corresponding to this, bear the CE-mark.

The products descried above are in conformity with the following harmonized standards:

SAFETY STANDARDS

EN ISO 61010-1:2010

Safety requirements for electrical equipment for measurement, control and laboratory use -

Part 1: General requirements

EN ISO 61010-2-011:2017

Safety requirements for electrical equipment for measurement, control, and laboratory use -

Part 2-011: Particular requirements for refrigerating equipment

**EMC** 

EN ISO 61326-1:2013

Electrical equipment for measurement, control and laboratory use - EMC requirements -

Part 1: General requirements

Öztaş Global Soğutma Teknolojileri ve Metal Sanayi ve Ticaret Anonim Şirketi Keçiliköy Osb Mahallesi Ahmet Nazif Zorlu Blv. No:32 Yunusemre / Manisa Mesir V.D. 707 051 5339 - Tel: 0236 213 12 12 - Fax: 0236 213 12 08
www.coolermed.com.tr / info@coolermed.com.tr

Ersin ÖZTAS

GENERAL MANAGES

GLOBAL SOĞUTMA TEKNOLOJİLERİ VE METAL SANAYİ TİCARET ANONİM ŞİRKETİ Keçilikiy OSB Mah. Ahmel Nazif Zorlu Bulv. No:32 Yunusemo

Keçilikoy OSB Mah. Ahmet Nazif Zorlu Bulv. No:32 Yunusemma Meşir V.D. 707 051 63 39 MANISA Tel: 0236/213 12 12 Pbx; Fax; (0236) 213 12 08





## **CERTIFICATE**

This certificate is granted to the organization,

### ÖZTAŞ GLOBAL SOĞUTMA TEKNOLOJİLERİ VE METAL SANAYİ TİCARET ANONİM ŞİRKETİ

Keçiliköyosb Mahallesi Ahmet Nazif Zorlu Bulvarı No:32 Yunusemre / Manisa / Türkiye

Branch: Muradiye Mahallesi Sanayi Caddesi Öztaş Ltd. Şti. Apartmanı

No:26 Yunusemre / Manisa / Türkiye

CUTTING, SHAPING, WELDING, SALES AND MARKETING OF STEEL AND OTHER METAL MATERIALS- DESIGN, MANUFACTURING SALES AND MARKETING OF COMMERCIAL TYPE REFRIGERATOR, MEDICAL TYPE REFRIGERATOR AND INDUSTRIAL TYPE REFRIGERATOR CABINETS

EA 17, 18

according to the scope,

### ISO 9001:2015

to certify that Quality Management System in accordance with standard's clauses is established and being implemented.

First Date of Issue : 01.03.2019

Date of Issue : 28.02.2022

**Certificate Period** : 3 Years


**Reissue Due Date : 28.02.2023** 

Certificate No : 01.19.7859.113419.D

First Quality Crtification
System Certificate Approved
Maltepe / İstanbul / Türkiye







FQC, India & Head Office FQC First Quality Certification Pvt. Ltd. SCO 37, Sector-12, Panchkula - 134109, Haryana, India T: +91 172 6535975 // F: +91 172 4416931 www.fqcert.com info@fqcert.com FQC, Turkey & Critical Location
FQC Uluslar Arası Belgelendirme ve Eğit. Hiz. A.Ş.
Cevizli Mah. Tansel Cad. No:12/18 Kat:4 D: 28-29 Maltepe / İstanbul / Türkiye
T: +90 216 444 21 41 / +90 216 457 69 08 F: +90 216 457 98 69
www.fqcglobal.org info@fqcglobal.org

Bu belge, müşterinin FQC'nin kurallarına ve sözleşme şartlarına uyduğu sürece geçerlidir. Sertifika geçerlilik durumu FQC' nin internet sitesinden takip edilir. This document shall remain valid as long as the customer obeys FQC rules and terms of the contract. Certificate validity may be checked on FQC website.

# **SZUTEST**

## CERTIFICATE



Medical Devices Quality Management System
CERTIFICATE NO: 31928401

### ÖZTAŞ GLOBAL SOĞUTMA TEKNOLOJİLERİ VE METAL SANAYİ VE TİCARET A.Ş.

Keçiliköy OSB Mahallesi Ahmet Nazif Zorlu Bulvar No: 32 YUNUSEMRE MANİSA/TÜRKİYE

### EN ISO 13485:2016

Design, Manufacturing, Sales-Marketing, Technical Service Activities of Medical Type Refrigerator (Blood Storage, Plasma Storage, Vaccine-Serum Kit Storage, Ultra Low Temperature Medical Type Freezer)

Approves that the Medical Devices Quality Management System implemented for above scope.

Issue Date

11.10.2019

**Expiry Date** 

10.10.2022

Revision Date/No

14.03.2022 / 2





epocycleneral Manager

The certificate inquiry is made by reading the QR codes by mobile devices, providing necessary information on http://public.szutest.com.tr or by using BDS No on https://tdbs.turkak.org.tr.