



Technical Specification No

Infusion device under pressure device with manometer and blood & fluid warmer device:

- In Manufacturer: Smithworks Medical
- Country of Origin : USA
- Model: Soft Sack

| Technical Specification | Yes | No | Remarks |
|---|-----|----|--------------------------|
| Configuration | | | |
| Technical specifications | | | |
| Blood&fluid warmer device | | | |
| ☐ Portable heating system for infusion solutions with supply at 12 V. | X | | ☐ leaflets , user manual |
| • portable electric device , power supply 12 V, | X | | ☐ leaflets , user manual |
| • for minimum 3 bags 1 or 6 bags 0.5 l. | | | |
| • with transport isothermic bag , thermal insulation for 2 hours minimum | | | |
| • Allows the heating of minimum 3 solution bags of 1 l each or 6 bags of 0.5 l each. | X | | ☐ leaflets , user manual |
| • A carrying bag, thermally isolated, with shoulder strap has to be included. | X | | ☐ leaflets , user manual |
| • The thermal isolation is efficient for 2 hours from its disconnection from the power supply. | X | | ☐ leaflets , user manual |
| | X | | ☐ leaflets , user manual |
| Pressure infusion device. | | | |
| • Infusion cuffs , adjustable pressure, built-in safety valve to maintain pressure on drug or blood bags. | X | | ☐ leaflets , user manual |
| • Cuffs cotton maded , with hangers for hanging on the infusion stand, washable at up to 60 ° C | X | | ☐ leaflets , user manual |
| • high pressure manometer included, chrome-plated aluminum with Ø 49 mm minimum , graded scale from 0 mm Hg to 300 mm Hg minimum , which allows monitoring of blood pressure during infusion, Latex Free Pressure Pressure. | X | | ☐ leaflets , user manual |
| • with the following capacity: 1000 ml, | X | | |

| | | | |
|---|------------|-----------|-------------------------------|
| | X | | ☐ leaflets , user manual |
| | X | | ☐ leaflets , user manual |
| Certification | | | |
| ☐ ISO 9001 | X | | ☐ certificate attached |
| ☐ ISO 13485 | X | | ☐ certificate attached |
| ☐ CE/CE Declaration of Conformity 93/42/EEC | X | | ☐ certificate attached |
| ☐ FDA | X | | ☐ certificate attached |
| Technical Specification | Yes | No | Remarks |
| Other services included in offer | | | |
| ☐ transport to the final customer | X | | ☐ bidder declaration attached |
| ☐ instalation and put in function | X | | ☐ bidder declaration attached |
| ☐ training for the technical staff for final customer | X | | ☐ bidder declaration attached |
| ☐ training for the medical staff for final customer | X | | ☐ bidder declaration attached |
| warranty for 24 months | | | |
| GUARANTEE | | | |
| ☐ 24 months guaranteed from signing the receiving minutes | X | | ☐ bidder declaration attached |
| SERVICE AND MAINTENANCE | | | |
| ☐ the bidder has technical facilities necessary for the service both for ambulances as well as for medical equipment according to the general guarantee conditions and the user guide of the manufacturer | X | | ☐ bidder declaration attached |
| ☐ Maximum time for technical intervention – 48 hours. | X | | ☐ bidder declaration attached |
| ☐ Maximum time for remedial measures, in total 72 | X | | ☐ bidder declaration attached |

| | | | |
|---|------------|-----------|--|
| hours. | X | | <input type="checkbox"/> bidder declaration attached |
| <input type="checkbox"/> The technical service and current repair will be performed without waiting in a queue. | X | | <input type="checkbox"/> bidder declaration attached |
| <input type="checkbox"/> Temporary replacement of the equipment has to be ensured according to the periods mentioned above. | | | |
| <input type="checkbox"/> During the guarantee period, at the reasonable request of the user, repair, adjustments and maintenance of medical equipment ,according to the specifications in the manufacturers' guides, will be done free of charge. | X | | <input type="checkbox"/> bidder declaration attached |
| <input type="checkbox"/> The spare parts and workmanship are for free, except from the consumables for vehicles established by the manufacturer | | | |
| Technical Specification | Yes | No | Remarks |
| AVAILABILITY OF SPARE PARTS | | | |
| <input type="checkbox"/> the bidder will ensure the availability of spare parts, accessories and consumables for all offered positions on the market of the Republic of Moldova for at least 10 years from putting into operation, free of charge or against payment, as follows: | X | | <input type="checkbox"/> bidder declaration attached |
| o spare parts free of charge, including the workmanship for the 36 months of guarantee period | X | | <input type="checkbox"/> bidder declaration attached |
| o for the rest of the period – against payment for at least 10 years period is fulfilled. | X | | <input type="checkbox"/> bidder declaration attached |
| <input type="checkbox"/> The bidder will submit the pricelist for the spare parts and the maintenance program to the Ministry of Health, Labour and Social Protection as beneficiary | X | | <input type="checkbox"/> bidder declaration attached |
| GUIDES | | | |
| <input type="checkbox"/> Technical Guide and User Guide. | X | | <input type="checkbox"/> bidder declaration attached |
| <input type="checkbox"/> All guides will be available in Romanian /Russian. | X | | <input type="checkbox"/> bidder declaration attached |