



贝恩医疗设备（广州）有限公司

地址：中国广州市经济技术开发区东区骏成路 10 号 邮编：510760

Bain Medical Equipment(Guangzhou) Co., Ltd.

Add: No.10, Juncheng Road, Eastern Section, Guangzhou Economic & Technological Development District, Guangzhou, China 510760

Tel: 86-20-82265249 Fax: 86-20-32067500

DECLARATION OF CONFORMITY

Date: 2023-03-13

To whom it may concern:

We, BAIN MEDICAL EQUIPMENT (GUANGZHOU) CO., LTD. with its location at No.10, Juncheng Road, Eastern Section, Guangzhou Economic & Technological Development District, Guangzhou, China 510760, hereby declare that the following **Fistula Needle REF Codes** have the specs as per table below:

Brand	REF Code	Specification
DORA	BAIN-A.V.F-026AG	Artery (fistula) - 16G, type "butterfly", with artery color marking, needle length: 25 mm, tube length: 150 mm, 360° rotating, with red clip.
DORA	BAIN-A.V.F-026VG	Vein (fistula) - 16G, type "butterfly", with vein color marking, needle length: 25 mm, tube length: 150 mm, 360° rotating, with blue clip.

We, BAIN MEDICAL EQUIPMENT (GUANGZHOU) CO., LTD, under our sole responsibility fully assure that there should be no doubt about the quality and performance of **DORA** brand products (the REF numbers and declared specs are mentioned in table above).

Kind regards,

For and on behalf of
BAIN MEDICAL EQUIPMENT (GUANGZHOU) CO., LTD
贝恩医疗设备（广州）有限公司


Authorized Signature

Zhang Peng,

Executive Deputy Factory Director of

Bain Medical Equipment (Guangzhou) Co., Ltd