

S/N: 008764

No.: **ICR/VC/HBX240502****Name and address of Applicant**

Jiangsu Rooe Medical Technology Co.,Ltd
NO.19, Renmin East Road, YangShe Town, Zhangjiagang City (Cathay Pacific Plaza) Z1310,
China

Name and address of manufacturer:

Jiangsu Rooe Medical Technology Co.,Ltd
NO.19, Renmin East Road, YangShe Town, Zhangjiagang City (Cathay Pacific Plaza) Z1310,
China

Product name:

Integrated Autoclave with Shredder

Product types:

MCDS Series

Product trademark:

ROOETECH

Verification was carried within following scope:

Information on the Declaration of Conformity:

Result:

Legislation:**Standard:**

- | | |
|--------------------|---|
| ✓ MD [2006/42/EC] | EN ISO 12100:2010
EN 60204-1:2018 |
| ✓ EMC [2014/30/EU] | EN IEC 61000-6-4:2019
EN IEC 61000-3-2:2019+A1:2021
EN IEC 61000-6-2:2019 |

The assessment process has been carried out in accordance with individual rules and conditions agreed with the applicant.
Evaluation has been carried out in accordance with:

Test report:

CTICAJG18240295051437DR
CTICAJG15240295051436AR

Tests conducted by:

CTIC Testing Group (Guangdong) Co., Ltd.

Issue date:

16.05.2024

Expiration date:

15.05.2029

Remarks:

- VoC was issued on voluntary basis and does not imply meeting all essential requirements listed in Declaration of Conformity.
- For introducing this product on European market may be needed EC/EU-type examination conducted by appropriate Notified Body.



ICR Co. Ltd.
www.icrqa.com
www.icrpolska.com
cert@icrqa.com

A handwritten signature in black ink, appearing to read 'Keateng Jim'.

CEO, ICR Co., Ltd.

