









## Medical electric drill RJ-PD (ND-2011)

IMAGE	Technical data
	<p><b>Medical electric canulated drill RJ-PD ND-2011</b></p> <p><b>Working voltage: 14.4V</b></p> <p><b>Adjustable speed: 0-1000rpm</b></p> <p><b>Noise: ≤45dB; Drill chuck detachable.</b></p> <p><b>Output power: 4N.m Output Voltage:100W</b></p> <p><b>Sterilization method: autoclavable 135° C</b></p> <p><b>Handpiece weight: 1420g,with battery 1840g</b></p>
	<p><b>Battery charger: PJCG001</b></p> <p><b>Input voltage: AC110V-60Hz and 220V,50HZ</b></p> <p><b>Output voltage: 14.4V</b></p> <p><b>Battery charger status: light indicator.</b></p> <p><b>Charging time: 2-3 hours.</b></p> <p><b>It can charge one battery each time.</b></p>
	<p><b>Re-chargeable battery PJDC1440-2psc</b></p> <p><b>Voltage: 14.4V 1800mAh</b></p> <p><b>Fully charged working time: 20 min non-stop</b></p>
	<p><b>Aseptic battery transfer kit PJXD001</b></p>

	<p><b>Key PJYS001-1pcs</b></p>
	<p><b>Sterilization case PJDB006-1 psc</b></p>
	<p><b>Drill bits: 5 pcs RJ-RP-TDB</b></p>
	<p><b>K wire adapter RJ-RP-KWA1- 1 psc</b></p>

**Date: June 14, 2024**

**Signature:**