

EC Certificate
Directive 93/42/EEC Annex V
Production Quality Assurance
Medical Devices

Registration No.: DD 60144068 0001

Report No.: 12018179 052

Manufacturer: OLYMPUS MEDICAL SYSTEMS CORP.
2951 Ishikawa-cho,
Hachioji-shi, Tokyo
192-8507 Japan


Products: Sterile Endotherapy Devices used in conjunction with Endoscopes, Sterile Non Active Instruments used in conjunction with Endoscopes and Sterile Non Active Instruments used in conjunction with Medical Ultrasound Diagnostic Imaging Systems
Replaces Approval, Registration No.: DD 60123877 0001

Expiry Date: 2024-05-26

The Notified Body hereby declares that the requirements of Annex V of the directive 93/42/EEC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex V, section 4 of the aforementioned directive. For placing on the market of class IIb and class III devices covered by this certificate an EC type-examination certificate according to Annex III is required.

Effective Date: 2019-10-31

Date: 2019-10-31


Notified Body
M. Aihara
M.Sc. M. Aihara

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg
TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC concerning medical devices with the identification number 0197.