



Nanchang Micare Medical Equipment Co.,Ltd

Address: No.666 Yaohu West 5th Road,Hi-Tech Zone,Nanchang, Jiangxi, China
Email address: sales3@micare.cn Website: www.micare-med.com/www.surgicallight.com
MP: +86-18979105197 Contact:Eric sang

ATT: Tender No. ocds-b3wdp1-MD-1752741146369
Achiziționarea Dispozitivelor medicale (lămpi chirurgicale)
conform neccesităților instituțiilor medico-sanitare publice (repetat)

DECLARATION

We, Nanchang Micare Medical Equipment Co.,Ltd, address: Building 16th No.666 Yaohu West 5th Road, Hi-Tech Zone, Nanchang, Jiangxi, China, do hereby confirm the below specification of the customized surgical light model POWERLED E700/700 according to the requirement of the tender, version of the Catalog Customized Version 2025-07-30, offered by our distributor MedPunkt SRL.

Specification	POWERLED E700/700
Voltage	95~245V · 50/60HZ
Illuminance at distance of 1 m(LUX)	60,000-160,000/60,000-160,000
Brightness Adjustable	0-100% (10 Steps)
Lamp Head Diameter	700MM/700MM
Quantity Of LEDS	80PCS/80PCS
Color Temperature Adjustable	3,500-5,500K
Color rendering index Ra	> 96
Color rendering index R9(Red)	98
Light Field Size Adjustable	150-350MM
Air Flow Design	YES
Endoscopy Mode	YES
Endoscopy Mode LEDS	8PCS/8PCS
Illumination For Endo-Mode	10%
LED service life	80,000hrs
Main Material	Aluminum
Rotating Angle Of Arm	>540°
Power Consumption	105 W
Vertical movement (MM)	≥ 800
Control panel on each head	Touch Screen
Light head protection rating	IP54+FireProof
Depth of illumination L1+L2	1400MM
Working distance (MM)	700 -1400

南昌迈柯尔医疗器械有限公司
NANCHANG MICARE MEDICAL EQUIPMENT CO.,LTD
Authorized Signature(s)

Antibacterial coating	YES
Light head freedom of movement	$\geq 4^\circ$
Back Up Battery (4-6 hrs)	Optional
Monitor With Extra Arm (22/27 Inches)	Optional
Internal/External Sony Camera (20X)	Optional
Removable sterilizable handles that provide adjustment of light intensity	YES
Ceiling Height (M)	3 - 4

The performances, features and parameters of our surgical lights can be customized according to the requirements of our clients and not all information is placed on the website of our company.

南昌迈柯尔医疗器械有限公司
NANCHANG MICARE MEDICAL EQUIPMENT CO.,LTD

Signed and Stamped:

Name: Eric Sang

Title: Sales Manager.....



Authorized Signature(s)

Date: July 30th, 2025