

CHIRANA T. Injecta, a.s.

Nám. Dr. Schweitzera 194

916 01 Stará Turá

Slovenská republika / Slovak Republic

IČO: 36794619

IČ DPH/VAT No: SK2022402822

Banka/Bank:

Sídlo/Branch:

SWIFT/BIC:

Č.ú./Account No/IBAN:

To whom it may concern

Maximum pressure declaration

"We, manufacturing company Chirana T. Injecta, a.s. located in Nam. Dr. Schweitzera 194, 916 01 Stará Turá, Slovak Republic, hereby declare following:

The table below shows the maximum pressure values in 3 different pressure units that some of our products can withstand:

ITEM	MAXIMUM PRESSURE			standard
	[kpa]	[bar]	[psi]	
Transfusion set (Chirahem)	50	0,5	7,2	ISO 1135-4
Pressure infusion set (Chiraplus P)	200	2	29	ISO 8536-8
Gravity infusion set (Chiraplus G)	20	0,2	2,9	ISO 8536-4
Extension line (Chiraline)	200	2	29	ISO 8536-9
IV cannula STANDARD/SAFETY (Chiraflex)	2060	20,6	300	ISO 10555-1 / ISO 10555-5
3-way stopcock standard (Chiraway)	450	4,5	65	ISO 8536-10
3-way stopcock lipid resistant (Chiraway)	550	5,5	79	ISO 8536-10



Branislav Munk
Sales & Marketing Manager
CHIRANA T.Injecta, a. s.