

## DECLARATION OF CONFORMITY

### CytoBrite Slide Incubation System

SciGene  
1287 Reamwood Ave  
Sunnyvale, CA 94089 USA



Declares that the above referenced product(s) meets the essential requirements of the following European Union Directives by using the relevant standards shown below to indicate compliance.

#### EMC Directive 2004/108/EC

EN 61326-1 2013

Electrical equipment for measurement, control and laboratory use to include:

|               |      |         |
|---------------|------|---------|
| EN 55011      | 2010 | Class A |
| EN 61000-3-2  | 2009 |         |
| EN 61000-3-3  | 2013 |         |
| EN 61000-4-2  | 2009 |         |
| EN 61000-4-3  | 2006 |         |
| EN 61000-4-4  | 2012 |         |
| EN 61000-4-5  | 2006 |         |
| EN 61000-4-6  | 2014 |         |
| EN 61000-4-8  | 2010 |         |
| EN 61000-4-11 | 2004 |         |

#### LVD Directive 2006/95/EC

IEC 61010-1 2013

Safety requirements for measurement, control and laboratory use  
3rd Edition Part 1: General requirements

#### RoHS Directive 2011/65/EU

Terry Gill

Name of Authorized Representative

Sunnyvale, California, USA

Place of Issue

Director of Product Manufacturing

Title of Authorized Representative

March 19, 2015

Date of Issue

Signature of Authorized Representative