

BOEN HEALTHCARE CO., LTD.

CERTIFICATE OF ANALYSIS

Report No.: GM03-BN23

Product Name		Microscope Slide	Specification	7105
Order Quantity		20000BXS	Test Quantity	25bxs
Lot. No.		20230401	Mfg. Date	20230401
Exp. Date		20260331	Report Date	2023.05.15
Inspection Standard		JB/T 8230.3-1997		
No.	Technical Requirement			Conclusion
1	The size of microscope slide is 25.4 x 76.2mm (1"x 3").			Meet the requirement
2	The Thickness of microscope slide is 1.1mm±0.05mm.			Meet the requirement
3	The microscope slide should be colorless and transparent.			Meet the requirement
4	The surface of microscope slide should be smooth and thickness should be uniformity.			Meet the requirement
5	The microscope slide should be clean, no mist, fingerprint or water stains.			Meet the requirement
6	The edges of microscope slide should be smooth and should be polished complete.			Meet the requirement
7	After boiled test, the microscope slide should be no blur, bubble or crack.			Meet the requirement
Conclusion		According to the requirements of JB/T 8230.3-1997, concluded that this batch pass.		

Inspector: Leo Sun
Date: 2023.05.15



Checker: Irene Qiao
Date: 2023.05.15