DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

| MANUFACTURER: | CONTEC MEDICAL SYSTEMS CO., LTD No.112 Qinhuang West Street, Economic & Technical Development Zone, Qinhuangdao, Hebei Province, PEOPLE'S REPUBLIC OF CHINA | |
|---|--|--|
| MEDICAL DEVICE: | Electrocardiograph ECG1200G | |
| CLASSIFICATION - ANNEX IX: | Class II a,Rule 10 | |
| CONFORMITY ASSESSMENT ROUTE: | Annex II excluding chapter 4 | |
| WE, (CONTEC MEDICAL SYSTEMS CO., LTD) HEREWITH DECLARE THAT THE STATED MEDICAL DEVICES. MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES; INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EEC ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURE. | | |
| STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED. | | |
| NOTIFIED BODY: | TÜV SÜD PRODUCT SERVICE GMBH RIDLERSTR 65, D-80339 M NCHEN, GERMANY | |
| IDENTIFICATION NUMBER: | C € ₀₁₂₃ | |
| (EC) CERTIFICATE(S): | G1 050972 0050 Rev.02 | |
| EC REP EUROPEAN REPRESENTATIVE: | Shanghai International Holding Corp. GmbH(Europe) Eiffestrasse 80, 20537 Hamburg Germany | |
| START OF CE-MARKING: 2 | 012-04-20 (Date or Lot or serial number) | |
| PLACE, DATE OF DECLARATION: | QINHUANGDAO, 2019-07-23 | |
| SIGNATURE: | President | |

TF-CE110613-09

Page 1 of 2

Ver: H

DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

Appendix: list of (harmonised - EN) standards

| No. | Reference | Title of Standard |
|-----|------------------------------|---|
| 1 | IEC 60601-1: 1988 | Medical electrical equipment; Part 1: General requirements |
| | +A1:1991+A2:1995 | for safety |
| 2 | IEC 60601-1-6:2006 | Medical electrical equipment Part 1-6: General requirements for safety - Collateral Standard: Usability |
| 3 | IEC60601-2-25: 1993 + | Medical electrical equipment – Part 2-25: Particular |
| | A1:1999 | requirements for the safety of electrocardiographs |
| 4 | EN60601-1-4:1996 +A1:1999 | Medical electrical equipment; Part 1: General requirements for safety –4 Collateral standard: Programmable electrical medical systems |
| 5 | IEC 60601-1-2: 2007 | Medical electrical equipment Part 1: General requirements for safety -2 Collateral standard: Electromagnetic compatibility - Requirements and tests |
| 6 | IEC62304:2006 | Medical device software Software life cycle processes |