

CERTIFICATE OF ANALYSIS

4408399 GeneScan™ - 600 LIZ™ Size Standard v2.0

Packaging Lot: 3499750

Expiry Date: 04.11.2027 (DD.MM.YYYY)

Storage: at 5±3°C in the dark

Note: **For Research Use Only. Not for use in diagnostic procedures.**

To verify the conformity of the manufacturing process, kits were visually inspected for correct assembly and labeling.

Filling lots for components in package:

Lot	Quantity	Description
3488077	2 x 200 µL	GeneScan™ - 600 LIZ™ Size Std. v2.0

QUALITY CONTROL

Parameter	Method	Requirement	Result
Use test	Product is tested for the correct number of peaks and the absence of extra peaks on a multi-capillary Applied Biosystems Genetic Analyzer.	Correct peak number and absence of extra peaks	Pass
Functional test	Product is tested for overall average peak height on a multi-capillary Applied Biosystems Genetic Analyzer.	2500-14000 RFU	Pass

ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by: **J. Žilinskienė**

