

Technical Data Sheet

Vacuum flask 3202

For MAYOR AUTOHEMOTHERAPY with medical ozone

Part	Material	Regulation	Norm
Glas	Sodium silicate glas type III	USP XX III EP III DIN ISO 719	DIN ISO 719
Plug	Latexfree synthetic isoprene caoutchouc siliconized,	EP III DAB	
Crimping cap	Aluminium / PP		DIN ISO 8536-7
MPS- System	PP	FDA-Regulations § 177.1520 „Olefin Polymers“ DIN EN ISO 10993 BgVV	DIN EN ISO 10993
Sodium Citrate	Tri-Na-dihydrate citrate dihydrate: 31,3 g Citric acid: 0,192 g Aqua ad iniectionem ad 1000 ml		GMP
Hanger	PP	BgVV EP III	
Label	flask		ISO 15223-1
Label	Package		ISO 15223-1