

Lot 1. Oto-emission screening apparatus

Specification requested by customer		Specifications proposed by the manufacturer
MODEL code from catalog		EROSCAN
Description	Intended for the recording of TEOEA and DPOEA acoustic otoemisis	yes
Parameter	Specification	
Type	portable	yes
Testing, method	TEOEA and DPOEA	yes
Number of frequencies for each EOA type	minimum 3	yes
Average test time (1 ear)	10-20 seconds	yes
maximum output level	≤90 dB SPL	yes
Internal noise level	≤30 dB SPL	yes
Test mode	Automatic, manual	automatic
Indicator of the correctness of the insertion of the probe into the external auditory canal	yes	yes
Battery	yes	yes
Food	220V, 50Hz	yes
PC connection, USB port	yes	yes (PC connection software optional)
Accessories	battery charger	yes
	A set of test earbuds in different sizes	yes
	extension cord (elastic tube) for olive insertion	not applicable
CE, ISO, catalog and brochure and full technical documentation		provided