

Certificate of Analysis

ThermoFisher
SCIENTIFIC

Anode Buffer Container 3500 Series

Product No. 4393927
Lot No. 2501001
Date of Manufacture 26FEB2025
Expiration Date 25APR2026

| TEST | SPECIFICATION | RESULT |
|---|-------------------|------------|
| PHYSICAL INSPECTION Material is clear, colorless and free of particulates | | |
| Appearance - Conformance | | Pass |
| pH TEST Measured at 25.0°C ± 0.3°C | | |
| pH Value | 7.6 - 8.2 | 7.946 |
| CONDUCTIVITY TEST Measured at 25.0°C ± 0.1°C | | |
| Conductivity | 1500 - 2000 µS/cm | 1777 µS/cm |

For Research Use Only.
Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system. Site:
Bedford, MA, USA

Corporate Legal Entity
Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com

Kim Kheav
Quality Assurance
Issued 06MAR2025