

BOEN HEALTHCARE CO., LTD.

Certificate of Analysis

Report No.: GM03-BN23

Product Name	Safety Blood Lancet	Specification	21G, 23G
Lot. No.	23052632	Mfg. Date	2023-05-04
Exp. Date	2028-05-03	Order Quantity	20000pcs
Sampling Quantity	13pcs	Inspection Date	2023.05.15
Inspection Standard	Factory Standard	Report Date	2023.05.15

Item	Technical Requirement	Result	Conclusion
Appearance	The lancet shall be clean, no foreign impurities.	Comply	Qualified
	The lancet shall be no injection defects such as burr, fly edges, bubble, etc.	Comply	Qualified
Lancet	The needle should be smooth, no defects.	Comply	Qualified
	The needle is sharp.	Comply	Qualified
Firmness	All parts shall be connected firmness.	Comply	Qualified
Sterile	The lancet shall be sterile.	Comply	Qualified
Conclusion	Meet the requirement		

Inspector: Tom Li
Date: 2023.05.15

Checker: Leo Sun
Date: 2023.05.15

