



## EC DECLARATION OF CONFORMITY

|                              |   |
|------------------------------|---|
| MANUFACTURER:                | DIESSE DIAGNOSTICA SENESE SPA<br>STRADA DEI LAGHI 39<br>53035 MONTERIGGIONI (SIENA),<br>ITALY |
| EUROPEAN REPRESENTATIVE:     | //  |
| GENERIC NAME                 | AUTOMATIC INSTRUMENT FOR ESR ANALYSIS   |
| PRODUCT:<br>CODE:            | <b>CUBE 30 touch</b><br><b>10395</b>  |
| TECHNICAL DATA:              | 110-230 Vac (50-60 Hz); Pwr: 100 VA   |
| CLASSIFICATION:              | IVD NOT IN ANNEX II OR SELF-TESTING IVD   |
| CONFORMITY ASSESSMENT ROUTE: | ANNEX APPLIED N° III EXCLUDING (6)<br>ESSENTIAL REQUIREMENTS OF ANNEX I                       |

WE HEREWITH DECLARE THAT THE DESIGN, TYPE OF MANUFACTURE OF THE IN VITRO MEDICAL DIAGNOSTIC DEVICE DESCRIBED ABOVE AND THE VERSION DISTRIBUTED ON THE MARKET, CONFORMS TO THE 98/79/EEC DIRECTIVE RELEVANT TO THE IN VITRO MEDICAL-DIAGNOSTICS DEVICES (IVD)

### THIS CERTIFICATE WILL LOSE ITS VALIDITY IN THE EVENT OF:

- MODIFICATIONS MADE TO THE MACHINE IN QUESTION WITHOUT OUR AUTHORIZATION
- INCORRECT USE OF THE INSTRUMENT
- TECHNICAL INTERVENTIONS PERFORMED BY UNAUTHORIZED PERSONNEL
- INSTALLATION OF NON-ORIGINAL SPARE PARTS.

### THE PRODUCT CONFORMS, AS A WHOLE AND IN ITS PARTS, WITH THE FOLLOWING STANDARDS AND THEIR AMENDMENTS:

|                     |   |
|---------------------|---|
| EN 61010-1:2010     | SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - PART 1: GENERAL REQUIREMENTS  |
| EN 61010-2-081:2015 | SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - PART 2-081: PARTICULAR REQUIREMENTS FOR AUTOMATIC AND SEMI-AUTOMATIC LABORATORY EQUIPMENT FOR ANALYSIS AND OTHER PURPOSES |
| EN 61010-2-101:2015 | SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - PART 2-101: SAFETY REQUIREMENTS FOR IN VITRO DIAGNOSTIC (IVD) MEDICAL EQUIPMENT   |

EN 61326-1:2013 ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL AND LABORATORY USE - EMC REQUIREMENTS - PART 1: GENERAL REQUIREMENTS

EN 61326-2-6:2013 ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL AND LABORATORY USE - EMC REQUIREMENTS - PART 2-6: PARTICULAR REQUIREMENTS - IN VITRO DIAGNOSTIC (IVD) MEDICAL EQUIPMENT

**AND THEREFORE MEETS THE ESSENTIAL REQUIREMENTS OF THE FOLLOWING COMMUNITY DIRECTIVES AND THEIR AMENDMENTS:**

LOW VOLTAGE DIRECTIVE (2014/35/EU)

ELECTROMAGNETIC COMPATIBILITY DIRECTIVE (2014/30/EU)

RESTRICTION OF HAZARDOUS SUBSTANCES IN ELECTRICAL AND ELECTRONIC EQUIPMENT DIRECTIVE – ROHS2 (2011/65/EU)

NOTIFIED BODY: NOT NECESSARY

(EC) CERTIFICATE: N.A.

START OF CE-MARKING: FEBRUARY 2018

REVISION: 5

PLACE, DATE OF ISSUE: MONTERIGGIONI, 10 DECEMBER 2021

EXPIRY DATE: 25 MAY 2022

**THE PRESENT EC DECLARATION OF CONFORMITY IS ISSUED UNDER THE SOLE AND EXCLUSIVE RESPONSIBILITY OF THE MANUFACTURER.**

SIGNATURE:



CHIARA MUZZI  
REGULATORY AFFAIRS MANAGER

THIS IS THE CERTIFIED COPY OF THE ORIGINAL DOCUMENT STORED IN ARCHIVE OF DIESSE DIAGNOSTICA SENESE SPA

ISSUED: MONTERIGGIONI, 10/12/2021



MAGDALENA STOCZKO  
REGULATORY SUPERVISOR