Certificate of Analysis



Cathode Buffer Container CE-SeqStudio

Product No. A33401

Lot No. **2501157**

Date of Manufacture 06MAR2025

TEST	SPECIFICATION	RESULT
PHYSICAL INSPECTION Material is clear, colorless and free of particulates		
Appearance - Conformance		Pass
pH TEST Measured at 25.0℃ ± 0.3℃		
pH Value	7.6 - 8.2	7.954
CONDUCTIVITY TEST Measured at 25.0℃ ± 0.1℃		
Conductivity	1500 - 2000 μS/cm	1785 μS/cm

For Research Use Only.

Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system. Site: Bedford, MA, USA

Corporate Legal Entity Life Technologies 5781 Van Allen Way Carlsbad, CA, USA 92008 www.thermofisher.com For inquiries, contact us at cofarequests@thermofisher.com Elsa Qirollari
Quality Assurance
Issued 07MAR2025