

杭州奥泰生物技术股份有限公司 Hangzhou AllTest Biotech Co.,Ltd	文件号 Document No.: ZTC-QC-005-R-003
传染病、心肌、肿瘤类 COA The Infectious Disease、Cardiology、Tumor COA	生效日期 Effective Date: 2018 年 07 月 02 日

Certificate of Analysis

Product Name: COVID-19 Antigen Rapid Test(Nasopharyngeal Swab)

Catalog No.: ICOV-502

Batch No.: COV2010042-S

Quantity:10PCS

Expiry Date: 2022-10

Date of Sampling: 2020-10-13

Date of Analysis: 2020-10-13

Other information:

Buffer Lot:20100027,EXP:2022-10

QC Item		QC Criterion	QC Result	Conclusion
Physical	Appearance	Good	Good	Pass
Functional Performance	Positive Sample	Positive	100% Positive	Pass
	Negative Sample	Negative	100% Negative	Pass

Others	/
---------------	---

Final QC Conclusion:	This batch of product met the QC Criteria.
-----------------------------	--

QC supervisor: *Freeman.zheng*

Date: 2020.10.13

