

CERTIFICATE OF ACCREDITATION

SENTECHKOREA

Accreditation No. : KC23-428

Corporation Registration No. : 174611-0032736

Address of Laboratory : 21-6, Jimok-ro 75beon-gil, Paju-si, Gyeonggi-do, Republic of Korea

Date of Initial Accreditation : August 30, 2023.

Validity of Accreditation : August 30, 2023. ~ August 29, 2027.

Scope of Accreditation : Attached Annex

Date of issue : August 30, 2023.

This calibration laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to Joint ISO-ILAC-IAF Communiqué).



CHIN CHONGWOOK

Head

Korea Laboratory Accreditation Scheme

901. Chemical analysis

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Breath alcohol analyzers C ₂ H ₅ OH	90101	(0 ~ 0.050) % BAC (0.050 ~ 0.100) % BAC	0.002 % BAC 0.003 % BAC	Standard Gas / KSQ-G-01