# Sırt Tahtası (Plastik) (ES-510)

## Spine Board (Plastic)





- · Vakumlama sistemi ile sert plastikten imal edilmiştir.
- 160 Kg. taşıma kapasitesine sahiptir.
- Çocuk ve yetişkinler için 2 farklı tiptedir.
- · Yetişkin tipinde 3 adet, çocuk tipinde 2 adet emniyet kemeri vardır.
- · Made of hard plastic with vacuum system.
- · 160 Kg. maximum carrying capacity.
- Two different types for children and adults.
- 3 seat belts for adult type, 2 for children type.

Ağırlık	Uzunluk <i>Lenght</i>	Genişlik <i>Width</i>
Weight		
5.1 Kg.	1840 mm.	425 mm

A Zivelala	ocuk için / For child Uzunluk	Genişlik
Ağırlık	Ozuniuk	
Weight	Lenght	Width
3.5 Kg.	1250 mm.	425 mm

# Kurtarma Yeleği (ET-370) Body Splint

- Omurga, baş, boyun kırık zedelenmelerinde, hastanın güvenli ve hareket ettirilmeden kurtarma operasyonları ve taşınmasında kullanılır.
- Renk kodlu, kolay açılıp kilitlenebilir kayışları sayesinde karanlıkta ve acil müdahale gerektiren anlarda bile hatasız olarak kullanılabilir.
- · Taşıma çantası bulunmaktadır.



- Usage for rescue operations and carrying of patient safely in cases of spine, head, neck fractures and injures.
- Can be used faultlessly by means of easy opening and lockable belts with color codes even in the darkness and in the events requiring emergency intervention.
- Accompanying carrying case.

## **DECLARATION OF CONFORMITY**



EMS MOBIL SISTEMLER A.Ş..

1. Organize Sanayi Bölgesi Kırım Hanlığı Caddesi No: 9

Sincan-Ankara /TURKEY

Tel : +90 312 267 18 67 - Faks : +90 312 267 18 66

**Product** 

: Life Vest (KED)

**Product Code** 

: ET 370

Classification

: Class I- Rule I

We herewith declare with this document that compliance evaluation was applied upon the above mentioned products as per the Council Directive 93/42/EEC Medical Devices Regulation Annex 7 for Medical Device Directive

### **General Applicable Directives**

Medical Device Directive council directive 93/42/EEC of June 1993 concerning medical devices. (MDD 93/42/EEC) Compliance Declaration evaluation was applied as per Annex 7.

#### Standarts

TS EN 980:2011 Symbols for use in the labeling of medical devices

TS EN 1041+A1:2015 Information supplied by the manufacturer of medical devices

TS EN ISO 14971:2013 Medical devices - Application of risk management to medical devices

Signature

1. Gradel / Aanay Belgesi Krimhanlığı (Jariel Ab. 9 Sincan/ANKAR/ Tel: 0312 26/// Fax: 0312 267 18 60 Mersis 047500550050001. Sincan de Dairesi: 475 005 5005

Name Position : Mevlüt Şahin: General Manager

Date

: 15,05,2016

First Release Date: 05.07.2009 Rev. No: 01