

Camp Medica Distribution S.R.L. - No. 29 Stanei Street, Bucharest, Romania
phone: +4021-450 58 90
e-mail: export@campmedica.ro
http : //www.campmedica.ro

"See Now" 6 DOA panel
(AMP,BAR,BZD,COC,OPI,THC) - Urine
For in vitro Diagnosis Use
Product Code: SN 7.X.6

CE

INTRODUCTION

The "See Now" 6DOA panel (AMP,BAR,BZD,COC,OPI,THC) test is intended for the qualitative detection of the presence of AMP,BAR,BZD,COC,OPI,THC and their metabolites in urine at or above the cutoff level of:
300 ng/ml AMP; 300ng/mL BAR; 300 ng/mL BZD; 300ng/mL COC; 300 ng/mL OPI; 50 ng/mL THC. The device is designed for professional use. This assay provides only a preliminary result. Clinical consideration and professional judgment should be applied to any drug of abuse test result, particularly in evaluating a preliminary positive result. To obtain a confirmed analytical result, a more specific alternate chemical method is needed..

SPECIMEN COLLECTION AND STORAGE

- Urine specimen may be collected at any time in a clean, dry container without preservatives.
- If specimen cannot be assayed immediately, they can be stored at 2-8°C for up to 72 hours prior to testing or frozen at -20°C for longer period of time.
- Specimens should be equilibrated to room temperature before testing if they were refrigerated or frozen.

TEST PROCEDURE

- Remove the test device from pouch when ready to perform the test. Label the test device with patient or control identification.
- Take out the bottom part of the plastic device to release the strips.
- Immerse the strips into the urine and take care about the instructions from the label of the plastic device.
- Take the strip out after 10 sec. Lay the strips of the panel on a flat, clean, dry, non-absorbent surface.
- Read the results at 10 minutes. Ensure that the background of the test area is white before interpreting the result

INTERPRETATION OF RESULTS

Positive

Only one color band appears at the control region. No apparent band at the test region. This indicates that drug presence is above the cutoff concentration.



Negative

Two distinct color bands appear at the control and test regions. This indicates that there is no drug in the sample or drug presence is below the cutoff concentration.



Invalid

No visible band at the control region. Repeat with a new test kit. If test still fails, please contact the distributor with the lot number.

Note: A faint line at the test region indicates the drug in sample is near the cut-off level for the test. These samples should be re-tested or confirmed

with a more specific method before a clinical determination is made.

STORAGE AND STABILITY

The test kit can be stored at temperature (2 to 30°C) in the sealed pouch to the date of expiration. The test kit should be kept away from direct sunlight, moisture and heat.

PRECAUTION

- FOR IN VITRO DIAGNOSTIC USE ONLY**
- Don't use it after the expiration date.
- The test device should not be reused.

PERFORMANCE CHARACTERISTICS

Sensitivity AMP

The "See Now" Amphetamine test strip detects amphetamine and its metabolites in urine at concentrations equal to or greater than 300 ng/ml.

Sensitivity BAR

The "See Now" Barbiturate test strip detects barbiturate and its metabolites in urine at concentrations equal to or greater than 300 ng/ml.

Sensitivity BZD

The "See Now" Benzodiazepine test strip detects Benzodiazepine and its metabolites in urine at concentrations equal to or greater than 300 ng/ml, which is suggested by the National Institute on Drug Abuse (NIDA) for the immunoassay method.

Sensitivity COC

The "See Now" Cocaine test strip detects Cocaine and its metabolites in urine at concentrations equal to or greater than 300 ng/ml.

Sensitivity OPI

The "See Now" Opiate Test Strips detects opiates and its metabolites in urine at concentrations equal to or greater than 300 ng/ml.

Sensitivity THC(Marijuana)

The "See Now" THC Test Strips detects THC and its metabolites in urine at concentrations equal to or greater than 50ng/ml.