

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**

AB-0583-T

21003762-
İNG

02-21

| REQUIRED TESTS | RESULT | COMMENTS |
|---|--------|----------|
| PHYSICAL PROPERTIES | | |
| Tensile Strength / Dry | P | |
| Tensile Strength / Wet | P | |
| Bursting Strength / Dry | P | |
| Bursting Strength / Wet | P | |
| Water Permeability | P | |
| Lint and Other Particles Generation From Nonwoven | P | |
| MICROBIOLOGICAL TESTS | | |
| Wet-Bacterial Penetration | P | |
| Dry-Bacterial Penetration | P | |
| Microbial Cleanliness (Bioburden) | P | |

P: Pass
F: Fail
R: Refer to retailer technologist.
Test results were evaluated according to EN 13795-1:2019 Standard Performance Properties Critical and Less Critical limit values

REMARK: Original samples are kept for 3 months and all technical records are kept for 5 years unless otherwise specified. If requested, measurement uncertainty will be reported. But unless otherwise specified, measurement uncertainty is not considered while stating compliance with specification or limit values. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95 %. The declaration of conformity was given in accordance with the Simple Acceptance Decision Rule. Tests marked (*) in this report are not included in the accreditation schedule.



This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

**ZENMED TIBBİ CİHAZLAR
VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD.ŞTİ.**
Mahmutbey Mah. 2567.Sk. No:1 İş Yeri No:11,21
Bağcılar / İSTANBUL / TÜRKİYE
Tel: 0212 445 45 25 Tic. Sic. No: 794704
Güneşli VD: 897 067 1942