



# Declaration of Conformity



**Manufacturer Name:** TSI Incorporated

**Address:** 500 Cardigan Road  
Shoreview, Minnesota 55126  
USA

**Telephone:** +011 800-874-2811

**Web:** [www.tsi.com](http://www.tsi.com)

*TSI hereby declares under our sole responsibility that the following apparatus as originally delivered:*

**Product Description:** AEROTRAK®+ Portable Airborne Particle counter

**Model Number(s):** A100 series and all variations of accessories

*Complies with the essential requirements of the following applicable European Directives and displays the CE Mark:*

Electromagnetic Compatibility (EMC) Directive 2014/30/EU

Radio Equipment Directive (RED) 2014/53/EU

Low-Voltage Directive (Safety) 2014/35/EU

RoHS Directive 2011/65/EU

*Conformity is assessed in accordance to the following standards:*

**EMC:** EN 61326-1:2013, Class A (Emissions and Immunity)  
EN 301 489-1 V2.2.3

**RED:** EN 300 330 V2.1.1

**LVD:** EN 61010-1: 2010 + A1:2019

**RoHS:** EN 63000:2018

16-Aug-2022

Date

Thomas Jacobson, VP of Engineering

**European Contacts:**

**TSI GmbH – Aachen**  
Neuköllner Strasse 4  
52068 Aachen  
GERMANY  
**Telephone:** +49-241-52303  
**Fax:** +49-241-5230349  
**E-mail:** [TSIGmbH@tsi.com](mailto:TSIGmbH@tsi.com)

**TSI Instruments Ltd.**  
Stirling Road  
Cressex Business Park  
High Wycombe, Buckinghamshire  
HP12 3ST  
UNITED KINGDOM  
**Telephone:** +44 (0) 149 4 459200  
**E-mail:** [tsiuk@tsi.com](mailto:tsiuk@tsi.com)

**TSI France Inc.**  
Hotel technologique  
BP 100  
Technopôle de Château-Gombert  
13382 Marseille cedex 13  
FRANCE  
**Telephone:** +33 (0)1 41 19 21 99  
**Fax:** +33 (0)1 47 86 00 07  
**E-mail:** [tsifrance@tsi.com](mailto:tsifrance@tsi.com)