

Letter of Declaration

We, **Zhongke Meiling Cryogenics Co.,Ltd**, located No.1862 Zishi Road,Hefei City,Anhui,China, hereby declare that

The temperature uniformity for Pharmacy refrigerator YC-315L and YC-525L is 1.1°C, the noise level is 45dB(A),

On behalf of Zhongke Meiling Cryogenics Company Limited

中科美菱低温科技股份有限公司
ZHONGKE MEILING CRYOGENICS COMPANY LIMITED
Vincent Cao

Vincent Cao

Sales Manager

Issued Date: 2023-10-13