

## Specificații tehnice

[Acest tabel va fi completat de către ofertant în coloanele 2, 3, 4, 6, 7, iar de către autoritatea contractantă – în coloanele 1, 5.]

Numărul procedurii de achiziție : ocds-b3wdp1-MD-1664371848792 din 28.09.2022						
Obiectul achiziției: Echipamente medicale						
Denumirea serviciului	Denumirea modelului serviciului	Țara de origine	Producătorul	Specificarea tehnică deplină solicitată de către autoritatea contractantă	Specificarea tehnică deplină propusă de către ofertant	Standarde de referință
1	2	3	4	5	6	7
<b>Bunuri</b>						
Gloves, examination, n/s, pwd free, latex, size S, M, L piece	Gloves, examination, n/s, pwd free, latex, size S, M, L piece	Malaiezia	LONGCANE INDUSTRIES SDN BHD	Single use, ambidextrous examination gloves, made from natural latex, powder-free and non-sterile. Disposable, single-use. Available in sizes small (S), Medium (M) and large(L). Can come in several colors. Do not apply force when putting on, keep fingernails short when using. Dispose carefully after use. Packed in <b>boxes of 100 gloves.</b> Minimum thickness 0.05mm, minimum 230mm total length, tested standard: <b>EN 455.</b>	Single use, ambidextrous examination gloves, made from natural latex, powder-free and non-sterile. Disposable, single-use. Available in sizes small (S), Medium (M) and large(L). Colour white. Do not apply force when putting on, keep fingernails short when using. Dispose carefully after use. Packed in <b>boxes of 100 gloves.</b> Minimum thickness 0.05mm, minimum 230mm total length, tested standard: <b>EN 455..</b>	ISO. Test report, Dec of conformity

Semnat: \_\_\_\_\_ Numele, Prenumele: Cobzari-Turcan Rodica

În calitate de: Administrator Ofertantul: "Aelo Grup" SRL

Adresa: Mun.Chisinau, str.Petru Movila 23/5, ap (of)3