

Certificate of Analysis

Certificat d'Analyse

applied
biosystems
by Thermo Fisher Scientific

PrioCHECK™ Porcine SVDV Ab Kit



440 Tests



7610205

LOT

S240901L



2024-09-02



2025-09-30

KIT CONFIGURATION CONFIGURATION DU KIT

COMPONENT DESCRIPTION

Description du composant

	REF	UNIT	LOT	
1: Test Plate	7610220	5 plates	S240901L201	2025-09-30
2: Conjugate (30x)	7610217	1.5 mL	S240901L212	2025-09-30
3: Conjugate Additive (30x)	7610215	1.2 mL	S240901L250	2025-09-30
4: Washing Fluid (200x)	7610301	60 mL	S240901L238	2025-09-30
5: Dilution Buffer	7610218	60 mL	S240901L248	2025-09-30
6: Demineralized Water	7610303	10 mL	S240901L218	2025-09-30
7: Blocking Reagent	7610214	5x 2.0 mL	S240901L254	2025-09-30
8: Reference Serum 1	7610211	0.5 mL	S240901L226	2025-09-30
9: Reference Serum 2	7610212	0.5 mL	S240901L228	2025-09-30
10: Reference Serum 3	7610213	0.5 mL	S240901L230	2025-09-30
11: Chromogen (TMB) Substrate	7588925	60 mL	S240901L214	2025-09-30
12: Stop Solution	7610302	60 mL	S240901L216	2025-09-30
Additional kit contents	-	5 plate sealers	-	-

INSTRUCTIONS FOR USE



English
Spanish

MAN0013854 Rev A.0
MAN0013857 Rev A.0

Notices d'utilisation

German
French

MAN0013855 Rev B.0
MAN0013856 Rev A.0

QUALITY CONTROL RESULTS RESULTATS DE CONTROLE QUALITE

Repeatability Répétabilité: CV < 10 %

Components	Specification	Result	Conclusion
Composants	Spécification	Résultat	Conclusion
Dilution Buffer	OD ₄₅₀ max > 1.000	2.126	PASSED
Reference Serum 1	OD ₄₅₀ < 0.300	0.186	PASSED
Reference Serum 2	PI > 50%	69%	PASSED
Reference Serum 3	PI < 50%	18%	PASSED

SUPPLEMENTAL INFORMATION - Available upon request to thermofisher.com/support

Informations complémentaires - Disponibles sur demande auprès de thermofisher.com/support

Additional quality control data can be provided upon request.

The aforementioned batch meets all specifications.

Le lot susmentionné est conforme aux spécifications.



Prionics Lelystad B.V.
Platinastraat 33
8211 AR Lelystad
The Netherlands

www.thermofisher.com/support

Anna Costerbaan

Quality Assurance

Date: 23-Sep-2024